

Calibration Certificate of Combined Gauge		Certificate No. : SAF-CAL2024-42			
		Date of Issue : 01-03-2024			
		Date of Calibration : 01-03-2024			
		Next Calibration Due : 31-05-2024			
Name and Address of Customer: M/s. Super Auto Forge Private Limited Vallam - Oragadam		Description of Gauge / Instrument :			
		Range/Size : 70/26/29 mm			
		Identification No. : CG-0291-27-1			
		SI No. : -			
		Make : -			
Environmental Condition					
Temperature	20±2°C	Humidity	30% to 75%		
Standard Ref / SOP No.		- / -			
Master Used for Calibration					
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability
Calibration results					
Parameter	Nominal Size (mm)	Permissible Size (mm)		Observed Size (mm)	Cal Status
		Min/Max	Wear Limit		
Go Size	70.0	70.02/70.05	70	70.028	ok

Remarks :

- Acceptance Remarks : accepted
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★