

| Calibration Certificate of | | | | | | | Certificate No. : CAL2024-51 Date of Issue : 02-03-2024 | | | |
|---|--------|-----------------|------------------|---------------------|-----------------|-------------|---|---------------------|------------|--|
| Combined Gauge | | | | | | | Date of Calibration: 02-03-2024Next Calibration Due: 01-04-2024 | | | |
| Name and Address of Customer: M/s. Super Auto Forge Private Limited Vallam - Oragadam | | | | | | Range/Size | Identification No. : 1145 SI No. : 2568 | | | |
| Environmental Condition | | | | | | | | | | |
| Temperature | | | 20±2°C | | | Hui | Humidity | | 30% to 75% | |
| Standard Ref / SOP No. | | | | | | | IS 3456 / Saf-02 | | | |
| Master Used for Calibration | | | | | | | | | | |
| SiNo | IMTE N | lo. It | Item Description | | Certificate No. | | Valid Upto | Traceability | | |
| 1 | test | | Bridge Gauge | | 25666 | | 01-04-2024 | Tespa Tools Pvt Ltd | | |
| Calibration results | | | | | | | | | | |
| Parameter | | Nominal Size () | | Permissible Size () | | ble Size () | Observed | Size () | Cal Status | |
| | | | | Min/Max | | Wear Limit | Observed Size () | | Cai Status | |

Remarks:

- Acceptance Remarks : accepted
- The Expanded uncertainty is \pm 0.002 mm at 95.45% confidence level with coverage factor k=2.
- This certificate is digitally signed and does not require any seal or signature

 $\star\star\star$ END OF THE CERTIFICATE $\star\star\star$