

Calibration Certificate of Plunger Dial Gauge			Certificate No. : SAF-CAL2024-27 Date of Issue : 01-02-2024 Date of Calibration : 01-03-2024 Next Calibration Due : 30-04-2024		
Name and Address of Customer: M/s. Super Auto Forge Private Limited Tvm - Thirumudivakkam			Description of Gauge / Instrument : Range/Size : 0.08 mm Least Count : - Identification No. : III-DI-0.08-1 SI No. : - Make : -		
Environmental Condition					
Temperature		20±2°C		Humidity	
				30% to 75%	
Standard Ref / SOP No.			IS 3459 / Saf-05		
Master Used for Calibration					
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability
1	I-GB-88-1	Slip Gauge Set	undefined	07-11-2025	undefined
Calibration results					
Parameter	Nominal Size (mm)	Permissible Error (mm)	Observed Error (mm)	Cal Status	
-	0.08	-0.005/0.005	0.002	ok	

- Remarks :
- Acceptance Remarks : accepted
 - The Expanded uncertainty is ± 0.005 mm at 95.45% confidence level with coverage factor k=2.
 - This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★