| Calibration Certificate of | | | | | Certificate No. : SAF-CAL2024-26 Date of Issue : 29-02-2024 | | | | |
|---|----------|-------------------|------------------|-----------------------|--|----------------------------------|------------|--------------|--|
| Comparator Stand | | | | | Date of Calibration : 29-02-2024 Next Calibration Due : 14-03-2024 | | | | |
| Name and Address of Customer: M/s. Super Auto Forge Private Limited | | | | | Description of Gauge / Instrument : | | | | |
| Vallam - Oragadam | | | | | | itification No. : TEST-1 | | | |
| | | | | | SI No. Make | : 147 : Alfa | | | |
| Environmental Condition | | | | | | | | | |
| Т | emper | ature | 20±2°C | | Humidity | | 30% to 75% | | |
| Standard Ref / SOP No. | | | | | | IS 3456 / Saf-06 | | | |
| Master Used for Calibration | | | | | | | | | |
| SiNo | IMTE No. | | Item Description | | Certificate No. | Valid | Upto | Traceability | |
| 1 | I-G | B-88-1 | Slip Gauge Set | | undefined | 07-11 | -2025 | undefined | |
| Calibration results | | | | | | | | | |
| Parameter | | Nominal Size (mm) | | Permissible Size (mm) | | Observed Size (mm) Cal Sta | | Cal Status | |
| | | | | Min/Max | | Observed Oize (IIIII) Odi Status | | | |
| Flatness | | 200X200 | | 0/0.005 | | 0.005 | | ok | |

Remarks:

- Acceptance Remarks : accepted
- The Expanded uncertainty is \pm 0.006 mm at 95.45% confidence level with coverage factor k=2.
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★

Metric Calibrated By