



Calibration Certificate of Position Gauge			Certificate No. : SAF-CAL2024-07 Date of Issue : 27-02-2024 Date of Calibration : 27-02-2024 Next Calibration Due : 26-06-2024		
Name and Address of Customer: M/s. Super Auto Forge Private Limited Vallam - Oragadam			Description of Gauge / Instrument : Range/Size : 25.6 mm Identification No. : Test SI No. : - Make : Metric		
Environmental Condition					
Temperature		20±2°C		Humidity	
				30% to 75%	
Standard Ref / SOP No.			- / -		
Master Used for Calibration					
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability
1	P-01	2D Height Gauge		26-02-2025	
Calibration results					
Parameter	Nominal Size (mm)	Permissible Size (mm)		Observed Size (mm)	Cal Status
		Min/Max	Wear Limit		
-	25.6	25.6/25.62	-	25.619	ok
Diameter	30	30.0/30.02	-	30.019	ok

Remarks :

- Acceptance Remarks : accepted
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★

Sumathy
Calibrated By

Sumathy P
Approved By