



|  |                          |                              |   |                           |                     |
|--|--------------------------|------------------------------|---|---------------------------|---------------------|
| <b>Calibration Certificate of<br/>Plain Plug Gauge</b>   |                          |                              | <b>Certificate No.</b> : SAF-CAL2024-46<br><b>Date of Issue</b> : 02-03-2024<br><b>Date of Calibration</b> : 02-03-2024<br><b>Next Calibration Due</b> : 01-06-2024     |                           |                     |
| <b>Name and Address of Customer:</b><br>M/s. Super Auto Forge Private Limited<br>Tvm - Thirumudivakkam |                          |                              | <b>Description of Gauge / Instrument :</b><br><b>Range/Size</b> : 50.6/50.8 mm<br><b>Identification No.</b> : TEST-001<br><b>SI No.</b> : 45465<br><b>Make</b> : Metric |                           |                     |
| <b>Environmental Condition</b>   |                          |                              |   |                           |                     |
| <b>Temperature</b>   |                          | 20±2°C                       |   | <b>Humidity</b>           |                     |
|  |                          |                              |   | 30% to 75%                |                     |
| <b>Standard Ref / SOP No.</b>  |                          |                              | - / -   |                           |                     |
| <b>Master Used for Calibration</b>   |                          |                              |   |                           |                     |
| <b>SiNo</b>  | <b>IMTE No.</b>          | <b>Item Description</b>      | <b>Certificate No.</b>  | <b>Valid Upto</b>         | <b>Traceability</b> |
| 1  | I-CMM-1                  | CMM                          | -   | 27-03-2024                | -                   |
| <b>Calibration results</b>   |                          |                              |   |                           |                     |
| <b>Parameter</b>   | <b>Nominal Size (mm)</b> | <b>Permissible Size (mm)</b> |   | <b>Observed Size (mm)</b> | <b>Cal Status</b>   |
|  |                          | <b>Min/Max</b>               | <b>Wear Limit</b>   |                           |                     |
| Go Dia   | 50.6                     | 50.61/50.63                  | 50.6  | 50.622                    | ok                  |

Remarks :

- Acceptance Remarks : accepted
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★

Jagan

Calibrated By

Approved By