



Calibration Certificate of Template Gauge			Certificate No. : CAL2024-49 Date of Issue : 02-03-2024 Date of Calibration : 02-03-2024 Next Calibration Due : 01-04-2024		
Name and Address of Customer: M/s. Super Auto Forge Private Limited Vallam - Oragadam			Description of Gauge / Instrument : Range/Size : 50.2/50.6 mm Identification No. : C-0224-10-4 SI No. : - Make : -		
Environmental Condition					
Temperature		20±2°C		Humidity	
				30% to 75%	
Standard Ref / SOP No.			IS 3455 / Saf-01		
Master Used for Calibration					
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability
1	I-CMM-1	CMM		27-03-2024	
Calibration results					
Parameter	Nominal Size ()	Permissible Size ()		Observed Size ()	Cal Status
		Min/Max	Wear Limit		

Remarks :

- Acceptance Remarks : accepted
- The Expanded uncertainty is ± 0.001 mm at 95.45% confidence level with coverage factor k=2.
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★

Venkatesan
Calibrated By

Vignesh S
Approved By