



Calibration Certificate of Plain Plug Gauge			Certificate No. : CAL2024-54 Date of Issue : 07-03-2024 Date of Calibration : 06-03-2024 Next Calibration Due : 05-06-2024		
Name and Address of Customer: M/s. Metric Measurement Technologies Chennai - Ambattur			Description of Gauge / Instrument : Range/Size : 12.10/12.50 mm Identification No. : MMT-208 SI No. : - Make : -		
Environmental Condition					
Temperature		20±2°C		Humidity : 30% to 75%	
Standard Ref / SOP No.			IS 3455 / SOP-01		
Master Used for Calibration					
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability
1	MMT-02	CMM		20-02-2025	Carl Zeiss
Calibration results					
Parameter	Nominal Size (mm)	Permissible Size (mm)		Observed Size (mm)	Cal Status
		Min/Max	Wear Limit		
Go Dia	12.10	12.11/12.13	12.10	12.119	ok
Nogo Dia	12.50	12.50/12.52	-	12.519	ok

Remarks :

- Acceptance Remarks : accepted
- The Expanded uncertainty is ± 0.002 mm at 95.45% confidence level with coverage factor k=2.
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★