

Calibration Certificate of Template Gauge		Certificate No. : SAF-CAL2024-30 Date of Issue : 29-02-2024 Date of Calibration : 29-02-2024 Next Calibration Due : 19-03-2024			
Name and Address of Customer: M/s. Super Auto Forge Private Limited Vallam - Oragadam		Description of Gauge / Instrument : Range/Size : 50.2/50.6 mm Identification No. : C-0224-10-4 SI No. : - Make : -			
Environmental Condition					
Temperature	20±2°C	Humidity	30% to 75%		
Standard Ref / SOP No.		IS 3455 / Saf-01			
Master Used for Calibration					
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability
Calibration results					
Parameter	Nominal Size (mm)	Permissible Size (mm)		Observed Size (mm)	Cal Status
		Min/Max	Wear Limit		
-	50.6	50.61/50.64	50.6	50.62	ok

- Remarks :
- Acceptance Remarks : accepted
 - The Expanded uncertainty is ± 0.001 mm at 95.45% confidence level with coverage factor k=2.
 - This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★