



<b>Calibration Certificate of Plain Plug Gauge</b>			<b>Certificate No.</b> : SAF-CAL2024-45 <b>Date of Issue</b> : 02-03-2024 <b>Date of Calibration</b> : 02-03-2024 <b>Next Calibration Due</b> : 01-06-2024		
<b>Name and Address of Customer:</b> M/s. Super Auto Forge Private Limited Tvm - Thirumudivakkam			<b>Description of Gauge / Instrument :</b> <b>Range/Size</b> : 50.6/50.8 mm <b>Identification No.</b> : TEST-001 <b>SI No.</b> : 45465 <b>Make</b> : Metric		
<b>Environmental Condition</b>					
<b>Temperature</b>		20±2°C		<b>Humidity</b>	
				30% to 75%	
<b>Standard Ref / SOP No.</b>			- / -		
<b>Master Used for Calibration</b>					
<b>SiNo</b>	<b>IMTE No.</b>	<b>Item Description</b>	<b>Certificate No.</b>	<b>Valid Upto</b>	<b>Traceability</b>
1	I-CMM-1	CMM	-	27-03-2024	-
<b>Calibration results</b>					
<b>Parameter</b>	<b>Nominal Size (mm)</b>	<b>Permissible Size (mm)</b>		<b>Observed Size (mm)</b>	<b>Cal Status</b>
		<b>Min/Max</b>	<b>Wear Limit</b>		
Go Dia	50.6	50.61/50.63	50.6	50.620	ok

Remarks :

- Acceptance Remarks : accepted
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★

Jagan

Calibrated By

Approved By