

Calibration Certificate of Analog Micrometer			Certificate No. : SAF-CAL2024-02 Date of Issue : 18-11-2021 Date of Calibration : 19-02-2024 Next Calibration Due : 18-11-2022		
Name and Address of Customer: M/s. Super Auto Forge Private Limited Vallam - Oragadam			Description of Gauge / Instrument : Range/Size : 25-50 mm Identification No. : 67215777 SI No. : - Make : -		
Environmental Condition					
Temperature		20±2°C		Humidity	
				30% to 75%	
Standard Ref / SOP No.			IS 2967 / WI-89		
Master Used for Calibration					
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability
1	I-GB-01	Slip Gauge Set	undefined	27-12-2024	undefined
Calibration results					
Parameter	Nominal Size	Permissible Error (mm)	Observed Error (mm)	Cal Status	
Error	25-50	-0.002/0.002	0.002	ok	

- Remarks :
- Acceptance Remarks : accepted
 - The Expanded uncertainty is ± 0.002 mm at 95.45% confidence level with coverage factor k=2.
 - This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★