

Multimodal Cyber-Physical Telemetry Complex for Neuroendocrine Stress Regulation & Autonomic Stabilization

Technical Case Study & Architecture Specification | Additional Documents & Work Examples

Author: Serhii Khylevych (Principal Systems Architect) • **Email:** hillmtz@gmail.com • **Phone:** +380 67 500 8897 • **LinkedIn:** [linkedin.com/in/serhii-khylevych-aa152322/](https://www.linkedin.com/in/serhii-khylevych-aa152322/)

1. EXECUTIVE SUMMARY & CLINICAL ENGINEERING RATIONALE

Problem Statement: Prolonged exposure to active warfare, kinetic threats, and severe operational strain in Ukraine induces chronic dysregulation of the Hypothalamic-Pituitary-Adrenal (HPA) axis and Autonomic Nervous System (ANS) among military veterans and specialized civilians. This condition triggers persistent hypercortisolemia (causing unproductive hepatic gluconeogenesis and cognitive exhaustion), profound vagal withdrawal (suppressed Heart Rate Variability), and neuromotor hyperarousal.

Engineering Objective: To architect, build, and validate an end-to-end, deterministic cyber-physical research and digital therapeutics stand that integrates continuous biochemical sensing, sub-millisecond electrophysiology, high-frequency neuromuscular telemetry, and adaptive stimulus control into a closed-loop biofeedback architecture. The system synchronizes heterogeneous sensor streams within a **1-millisecond master synchronization window**, establishing quantitative, verifiable proof of vegetative recovery and cortisol normalization.

2. COMPREHENSIVE SYSTEM ARCHITECTURE (PLANTUML MODEL)

The complex is architected on an open **Linux (Debian 13 x86_64) real-time platform**, replacing closed-box proprietary architectures to achieve unconstrained kernel-level access to hardware buses, low-jitter timer clocks, and named IPC pipes. Connectors are routed with strictly orthogonal right-angled connections for maximal clarity and architectural rigor across multi-stage data flows:




Figure 1: Multimodal Cyber-Physical Telemetry Complex Architecture — End-to-End System Flow (Right-Angled Connectors, 1 ms Master Hub)

3. MULTI-MODAL SENSOR MATRIX & TELEMETRY SPECIFICATIONS

Channel	Sensor / Source	Physical Interface	Frequency	Precision	Physiological Parameter
CH 1: ECG / Autonomic	Polar H10 Chest Belt	BLE GATT 0x180D (Char 0x2A37)	Event-driven (~1 Hz)	0.976 ms (1/1024 s)	R-R Intervals, RMSSD, SDNN, Baevsky Stress Index, Cardiac Phase
CH 2: Metabolic / Endocrine	LinX CGM Interstitial Sensor	BLE → Mobile → Wi-Fi REST API	1 sample / 60 sec	0.1 mmol/L	Interstitial Glucose, Cortisol-driven Gluconeogenesis Rate (dG/dt)
CH 3: Neuromotor & Tremor	DualShock 4 v2 Gamepad	USB Raw HID / Linux evdev	250 Hz (every 4 ms)	16-bit analog axis	Sympathetic Micro-Tremor (8–12 Hz), Aim Jitter, Trigger Muscle Tension
CH 4: Behavioral Marker	PCsensor FootSwitch	USB Keyboard HID Scan Code	Event-driven (<10 ms)	Discrete boolean	Subjective Cognitive Event Tagging, Hands-Free Push-to-Talk
CH 5: Acoustic & Vocal	Logitech Zone Vibe 100	Bluetooth mSBC / PipeWire	48 kHz, 16-bit PCM	Spectral Hz / dB	Think-Aloud Protocol vocal strain, F0 pitch jitter, voice envelope
CH 6: Contextual Stimulus	Virtual Engine + F4SE	Named Pipe / IPC Stream	Event-driven (<1 ms)	String event IDs	State transitions (OnCombatStateChanged, OnHit, QuestMilestones)

4. SIGNAL PROCESSING & MATHEMATICAL METHODOLOGY

4.1. Autonomic Regulation & Heart Rate Variability (HRV) Processing

Electrocardiographic signals from the Polar H10 provide hardware-level R-wave peak detection. The system executes mathematical transformations over sliding temporal windows (N = 300 beats):

A. Parasympathetic / Vagal Restoration Metric (RMSSD)

The Root Mean Square of Successive Differences assesses parasympathetic innervation via the vagus nerve:

$RMSSD = \sqrt{[(1 / (N - 1)) \times \sum_{i=1}^{N-1} (RR_{i+1} - RR_i)^2]}$

Clinical Interpretation: Chronic wartime stress suppresses RMSSD (<20 ms). Active vegetative stabilization and immersion induce an ascent toward >55 ms, demonstrating effective vagal braking against neuroendocrine alarm.

B. Baevsky Stress Index (SI / Index of Regulatory Systems Tension)

Evaluates sympathetic activation and cardiac stabilization under operational load:

$SI = AMo / (2 \times Mo \times MxDMn)$

Where Mo (Mode) is the most frequent R-R interval in seconds, AMo is the mode amplitude (% of intervals in the mode bin, bin width = 50 ms), and MxDMn is the variation range (RR_{max} – RR_{min}). Baseline resting homeostasis = 50–150; acute stress > 500–1200.

C. Neurocardiological Cardiac Phase Gating

- Systolic Window (0 – 350 ms from R-peak):** Ventricular contraction triggers aortic arch baroreceptors, producing central cortical sensory gating (attenuated cerebral processing).
- Diastolic Window (350 ms – next R-peak):** Cardiovascular relaxation; period of maximal cognitive responsiveness and psychomotor speed.

4.2. Continuous Glucose Monitoring (dG/dt) as Cortisol Proxy

Cortisol stimulates hepatic gluconeogenesis and suppresses peripheral glucose utilization. Because direct blood cortisol sampling is invasive and delayed, continuous interstitial glucose rate-of-change (dG/dt) serves as a continuous metabolic indicator:

$dG/dt(t) = [G(t) - G(t - \Delta t)] / \Delta t,$ with physiological lag model: $G_{blood}(t) \approx G_{interstitial}(t) + \tau \times (dG/dt)(t), \quad \tau \approx 12 \text{ min}$

Acute neuroendocrine stress spikes dG/dt > +0.08 mmol/L/min without muscular exertion. Successful stabilization flattens dG/dt into a stable plateau (–0.01 to +0.02 mmol/L/min).

4.3. Neuromuscular Tremor Spectral Analysis (8–12 Hz)

Analog stick coordinates sampled at 250 Hz undergo Short-Time Fourier Transform (STFT) with 500 ms sliding Hanning windows to isolate physiological tremor from intentional motor input:

$E_{tremor}(t) = \int_{8 \text{ Hz}}^{12 \text{ Hz}} |X(t, f)|^2 df$

Surges in 8–12 Hz power signify acute adrenergic (epinephrine/norepinephrine) discharge, preceding heart rate acceleration by 2–4 seconds.

4.4. Unified 1-Millisecond Telemetry Record (JSONL / Parquet)

```
{
  "timestamp_epoch_ms": 1789636266491,
  "iso_time": "2026-09-17T12:11:06.491Z",
  "cardiac": {
    "is_r_peak": true,
    "rr_interval_ms": 842.8,
    "cardiac_phase": "diastole",
    "hrv_rmssd_running": 56.4,
    "stress_index_baevsky": 98.2
  },
  "glucose": {
    "interstitial_mmol": 5.4,
    "slope_dG_dt": -0.015,
    "hpa_axis_status": "quiescent"
  },
  "neuromotor": {
    "stick_coords": [14200, 310],
    "aim_tremor_power_8_12hz": 142.6,
    "trigger_l2_tension": 0.85
  },
  "stimulus_context": {
    "event_type": "OnCombatStateChanged",
    "threat_id": "Target_Neutralized",
    "latency_to_action_ms": 184
  }
}
```

5. INDUSTRIAL & SYSTEMS ENGINEERING CROSS-POLLINATION

5.1. Direct Transferability to Getinge Applikon Bioprocessing Systems

- **Liquid-Contact Biochemical Sensing & Drift Compensation:** The algorithms developed for CGM micro-dialysis sensor drift, enzymatic saturation, and fluid-phase impedance directly translate to online monitoring of glucose, lactate, dissolved oxygen (DO), and optical biomass in Getinge stirred-tank and single-use bioreactors.
- **Closed-Loop Control (Cascade / PID):** The closed-loop biofeedback architecture regulating cognitive/visual stimuli based on physiological feedback mirrors bioreactor process automation (closed-loop feeding, gas-mixing manifolds, thermal regulation).
- **Data Integrity & Regulated Standards (GAMP 5 / 21 CFR Part 11):** The monotonic high-resolution time-series recording, audit-trail generation, and tamper-evident telemetry logs comply directly with pharmaceutical software validation requirements.

5.2. Direct Transferability to Xsens (Movella) Motion Tracking & Sensor Fusion

- **Multi-Rate Sensor Fusion:** Coupling high-rate interrupts (250 Hz motorics, 1000 Hz cardiac events) with slow biochemical metrics utilizes state-estimation and Kalman filtering techniques identical to fusing high-rate MEMS IMUs (gyros/accels) with low-rate GPS/GNSS.
- **Deterministic Low-Jitter Clock Synchronization:** The hardware-level Linux daemon eliminates clock drift across heterogeneous wireless/wired interfaces, directly applicable to Xsens MVN motion capture suits and MTi-series navigation units.
- **Vibrational Noise Decoupling:** The spectral isolation of 8–12 Hz physiological tremor translates directly to decoupling high-frequency engine vibrations and structural acoustic noise from MEMS inertial sensors.

6. CONCLUSION

This project demonstrates the capacity to conceptualize, architect, and execute a complex, safety-critical cyber-physical bioengineering complex under extreme constraints. By uniting rigorous systems engineering principles (V-model, formal ICDs, deterministic real-time scheduling) with advanced physiological and biochemical instrumentation, it serves as a verified portfolio example of cross-domain systems architecture leadership.