

SOMATOM go.Top - Exam Protocols *syngo* CT VA40

The dose values stated in the exam protocol list are estimations and provide orientation and overview. For individual exam protocols, the valid dose values are displayed at the CT scanner.

The descriptions below refer to the corresponding column headings of the exam protocol tables on the following pages.

CARE kV IQ level

'CARE kV IQ level' specifies the relative image quality that the system would achieve under specific reference conditions with regards to the patient's size and anatomic shape and standard tube voltage of 120 kV, independent of the specific scanner conditions.

Eff. mAs

'Eff mAs' indicates the values for operation with deactivated CARE Dose4D. The dose values are defined as a typical adult (70 - 80 kg or 155 - 180 lbs) for adult protocols, or as a typical child (5 years, appr. 20 kg or 45 lbs) for child protocols.

CTDIvol [mGy]

Measured in a 16 cm phantom for head modes and in a 32 cm phantom for body scans.

CTDIvol [mGy]; DLP [mGy*cm]

The indicated dose values 'CTDIvol', and 'DLP' refer to the indicated settings for operation with deactivated CARE Dose4D, regardless if CARE Dose4D is activated by default or not.

Shaped Filter

Displays the setting of the filter in collimator, standard or narrow.

Absolute Start Delay [s]

The delay time of all ranges of an auto range always refers to 0.

Iterative Recon (SAFIRE or ADMIRE)

It is an iterative reconstruction and the Strength can be adjusted to the customer's preference.

For each Siemens Healthineers exam protocol [factory], the **Decision Trees** and the **Recon Strategies** are listed. The scan strategies can be combined with patient characteristics and categories or patient questions into a decision tree in the myExam Cockpit workspace in the Exam Designer.

For each Siemens Healthineers exam protocol [factory], the **Contrast Protocols** are listed. The contrast protocols are displayed in the Contrast Protocols workspace in the Exam Designer, where you can set up the injection phases and contrast agent information for CARE Contrast CT examinations.

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit			
Brain [factory] (Adult)	>Indications: stroke, hemorrhage, cerebral atrophy, hydrocephalus, inflammation, wo CM >Scan modes: RoutineSpiralAdultHead >Decision Tree: Brain contrast medium necessary [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard					2.0																	
			Topogram 0.60 Tr20 sag																																			
		Brain		Brain	RoutineSpiralAdultHead	Brain contrast medium necessary [factory]	64 x 0.60	Off			120	265		(16 cm) 49.7	660	Standard	0.55		1.00		2.0									Tr20								
			Brain 5.00 Hr40 I2 Brain 5.00 Hr60 I2																				5.00	5.00	Brain contrast medium necessary [factory].wo CM; Brain contrast medium necessary [factory].w CM	None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2							
Brain Multi Phase [factory] (Adult)	>Indications: stroke, tumors, cerebral atrophy, hydrocephalus, inflammation,w and wo CM >Scan modes: RoutineSpiralAdultHead >Contrast Protocol: Brain [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard					2.0																	
			Topogram 0.60 Tr20 sag																																			
		Non Contrast		Brain	RoutineSpiralAdultHead		64 x 0.60	Off			120	265		(16 cm) 49.7	660	Standard	0.55		1.00		2.0																	
			Non Contrast Brain 5.00 Hr40 I2 Non Contrast Brain 5.00 Hr60 I2																				5.00	5.00		None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2							
		Venous Phase Brain		Brain	RoutineSpiralAdultHead		64 x 0.60	Off			120	265		(16 cm) 49.7	660	Standard	0.55		1.00		180.0	Brain [factory]																
			Venous Phase Brain 5.00 Hr40 I2																				5.00	5.00		None	None		Hr40	On	2							
Brain DE Hemorrhage Post Intervention [factory] (Adult)	>Indications: distinguish residual contrast agent from brain hemorrhage post intervention >Scan modes: RoutineSpiralAdultHeadTwinspiral (max. 100 cm)	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard					2.0																	
			Topogram 0.60 Tr20 sag																																			
		Brain		Brain	RoutineSpiralAdultHeadTwinspiral (max. 100 cm)		64 x 0.60	Off						(16 cm) 48.8	691	Standard	0.55		1.00		2.0																	
			Brain 5.00 Hr40 I2 Brain 5.00 Hr60 I2																				5.00	5.00		None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2							
			Brain 3.00 Qr40 I2 ax																				3.00	2.00		DE Parallel Ranges	None	MPR	Qr40	On	2							
			Brain 1.50 Qr40 I2																				1.50	1.00			None		Qr40	On	2							
Brain Perfusion [factory] (Adult)	>Indications: stroke, hemorrhage, ischemia, malperfusion, aneurysm, w and wo CM >Possible application: CT Neuro Perfusion >Scan modes: RoutineSpiralAdultHead, DynamicSpiralAdultHead, PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioNeck >Contrast Protocol: Brain Perfusion [factory], CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard					2.0																	
			Topogram 0.60 Tr20 sag																																			
		Non Contrast Brain		Brain	RoutineSpiralAdultHead		64 x 0.60	Off			120	265		(16 cm) 49.7	660	Standard	0.55		1.00		2.0																	
			Non Contrast Brain 5.00 Hr40 I2 Non Contrast Brain 5.00 Hr60 I2																				5.00	5.00		None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2							
			Non Contrast Brain 5.00 Hr40 I2 ASPECT Non Contrast Brain 4.00 Hr40 I3 BrainHem																				5.00	5.00		ASPECT Score	None		Hr40	On	2							
		Perfusion Brain		Brain	DynamicSpiralAdultHead		64 x 0.60	Full	78	Soft Tissue with Contrast	70	132		(16 cm) 133	1464	Standard	0.50		0.33		2.0	Brain Perfusion [factory]																
			Perfusion Brain 5.00 Hr36 3.39s - 47.38s																				5.00	3.00					Hr36	Off	2							
		Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					7.0																	
			Topogram 0.60 Tr20 cor																																			
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																	
			Premonitoring 10.00 Br36																																			
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]																
			Monitoring 10.00 Br36																																			
		CTA Brain and Neck		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	145	Auto (Vascular)	100	115		(32 cm) 6.26	270	Standard	1.20		0.33			CTA [factory]																
			CTA Brain and Neck 3.00 Hv36 I2 sag CTA Brain and Neck 3.00 Hv36 I2 cor																																			
			CTA Brain and Neck 3.00 Hv36 I2 ax																																			

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Brain Perfusion Prolonged [factory] (Adult)	>Indications: stroke, hemorrhage, ischemia, malperfusion, aneurysm, w and wo CM >Possible application: CT Neuro Perfusion >Scan modes: RoutineSpiralAdultHead, DynamicSpiralAdultHead, PremonitoringSequence Adult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngio Neck >Contrast Protocol: Brain Perfusion [factory], CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard						2.0															
			Topogram 0.60 Tr20 sag																											Tr20							
		Non Contrast		Brain	RoutineSpiralAdultHead		64 x 0.60	Off				120	265		(16 cm) 49.7	660	Standard	0.55		1.00		2.0															
			Non Contrast Brain 5.00 Hr40 I2																					5.00	5.00			None	Impression Multi Recons (2)		Hr40	On	2				
			Non Contrast Brain 5.00 Hr60 I2																											Hr60	On	2					
			Non Contrast Brain 5.00 Hr40 I2 ASPECT																						5.00	5.00		ASPECT Score	None		Hr40	On	2				
			Non Contrast Brain 4.00 Hr40 I3 BrainHem																					4.00	4.00			Brain Hemorrhage	None			On	3				
		Perfusion Brain		Brain	DynamicSpiralAdultHead		64 x 0.60	Full	78	Soft Tissue with Contrast	70	132			(16 cm) 159	1751	Standard	0.50		0.33		2.0	Brain Perfusion [factory]														
			Perfusion Brain 5.00 Hr36 3.39s - 63.51s																						5.00	3.00					Hr36	Off	2				
		Topogram			ScanplanningTopoAdultSn		6 x 0.60					Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						7.0														
			Topogram 0.60 Tr20 cor																												Tr20						
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with	100	26			(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0															
			Premonitoring 10.00 Br36																						10.00				None			Br36					
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with	100	26			(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]														
			Monitoring 10.00 Br36																						10.00				None			Br36					
		CTA Brain and Neck		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	145	Auto (Vascular)	100	115			(32 cm) 6.26	270	Standard	1.20		0.33				CTA [factory]													
	CTA Brain and Neck 3.00 Hv36 I2 sag																						3.00	3.00			None	Orientation Multi Recons (3)	MIP Thin	Hv36	On	2					
	CTA Brain and Neck 3.00 Hv36 I2 cor																																				
	CTA Brain and Neck 3.00 Hv36 I2 ax																																				
Inner Ear [factory] (Adult)	>Indications: inflammation, tumorous processes of pyramids, acoustic neuroma, pathologies of the mastoid process, cerebellopontine angle tumors, cochlea implants, trauma, wo CM >Scan modes: RoutineSpiralAdultHeadSn	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard						2.0															
			Topogram 0.60 Tr20 sag																												Tr20						
		Inner Ear		Brain	RoutineSpiralAdultHeadSn		64 x 0.60	Full	110	Auto (Bone/Calcium)	Sn 130	501		(16 cm) 17.6	147	Standard	0.55		1.00		2.0																
			Inner Ear 1.50 Hr60 I3 ax																					1.50	1.50			None	Impression Multi Recons (2)	MPR	Hr60	On	3				
			Inner Ear 1.50 Hr40 I3 ax																											Hr40	On	3					
			Inner Ear 0.80 Hr60 I3 cor R																						0.80	0.80		Parallel Ranges	None	MPR	Hr60	On	3				
	Inner Ear 0.80 Hr60 I3 cor L																						0.80	0.80		Parallel Ranges	None	MPR	Hr60	On	3						
Sinus [factory] (Adult)	>Indications: sinusitis, mucocoele, pneumatization, polyposis, tumor, corrections, trauma, wo CM >Scan modes: RoutineSpiralAdultHeadSnXcare >Decision Tree: Sinus sinusitis or soft tissue [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard						2.0															
			Topogram 0.60 Tr20 sag																												Tr20						
		Sinus		Brain	RoutineSpiralAdultHeadSnXcare	Sinus sinusitis or soft tissue [factory]	64 x 0.60	Full	50	Bone/Calcium	Sn 100	581		(16 cm) 5.51	62.7	Standard	0.55		1.00		2.0																
			Sinus 3.00 Hr60 I3 ax																					3.00	3.00		Sinus sinusitis or soft tissue [factory].low dose	None	Impression Multi Recons (2)	MPR	Hr60	On	3				
			Sinus 3.00 Hr40 I3 ax																											Hr40	On	3					
			Sinus 3.00 Hr60 I3 cor																						3.00	3.00		Sinus sinusitis or soft tissue [factory].low dose	Parallel Ranges	None	MPR	Hr60	On	3			
			Sinus 3.00 Hr40 I3 ax																						3.00	3.00		Sinus sinusitis or soft tissue [factory].standard dose	None	Impression Multi Recons (2)	MPR	Hr40	On	3			
			Sinus 3.00 Hr60 I3 ax																											Hr60	On	3					
	Sinus 3.00 Hr40 I3 cor																						3.00	3.00		Sinus sinusitis or soft tissue [factory].standard dose	Parallel Ranges	None	MPR	Hr40	On	3					
Orbit [factory] (Adult)	>Indications: fracture, tumor, trauma, wo CM >Scan modes: RoutineSpiralAdultHeadSnXcare	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard						2.0															
			Topogram 0.60 Tr20 sag																												Tr20						
		Orbit		Brain	RoutineSpiralAdultHeadSnXcare		64 x 0.60	Full	110	Auto (Bone/Calcium)	Sn 100	1,227		(16 cm) 11.6	97.5	Standard	0.55		1.00		2.0																
			Orbit 2.00 Hr60 I3																						2.00	2.00		None	Impression Multi Recons (2)		Hr60	On	3				
			Orbit 2.00 Hr40 I3																											Hr40	On	3					
			Orbit 2.00 Hr60 I3 cor R																						2.00	2.00		Parallel Ranges	None	MPR	Hr60	On	3				
			Orbit 2.00 Hr60 I3 cor L																						2.00	2.00		Parallel Ranges	None	MPR	Hr60	On	3				
			Orbit Hr40 I3 cor Radial																									Radial Ranges	None	VRT	Hr40	On	3				

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit			
Dental [factory] (Adult)	>Indications: display and measurement of the bone structures of the upper and lower jaw as the basis for planning in oral surgery, wo CM >Possible application: CT	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard						2.0																
			Topogram 0.60 Tr20 sag																																			
		Dental		Brain	RoutineSpiralAdultHead		64 x 0.60	Full	65	Bone/Calcium	120	50		(16 cm) 9.38	59.0	Standard	0.55		1.00		2.0																	
			Dental 1.00 Hr60 I3																					1.00	0.70		None	None		Hr60	On	3						
			Dental 5.00 Hr60 I3 ax																					5.00	5.00		None	None	MPR	Hr60	On	3						
Neck [factory] (Adult)	>Indications: masses, tumor, metastases, lymphoma, lymph nodes in the larynx or pharynx region, w/wo CM >Scan modes: RoutineSpiralAdultNeck >Decision Tree: Neck contrast medium necessary [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.06	1.62	Standard						7.0																
			Topogram 0.60 Tr20 sag																																			
		Neck		Neck	RoutineSpiralAdultNeck	Neck contrast medium necessary [factory]	64 x 0.60	Full	155	Non-Contrast	120	119		(32 cm) 10.7	213	Standard	0.80		1.00		7.0																	
			Neck 5.00 Br40 I3 sag																					5.00	5.00	Neck contrast medium necessary [factory].wo CM; Neck contrast medium necessary [factory].w CM	None	Orientation Multi Recons (3)	MPR	Br40	On	3						
			Neck 5.00 Br40 I3 cor																																			
			Neck 5.00 Br40 I3 ax																																			
Neck DE Virtual Non Contrast [factory] (Adult)	>Indications: masses, tumor, metastases, lymphoma, lymph nodes in the larynx or pharynx region, w CM >Scan modes: RoutineSpiralAdultNeckTwinbeam >Decision Tree: Neck VNC TwinBeam DE possible [factory] >Contrast Protocol: Neck [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.06	1.62	Standard						7.0																
			Topogram 0.60 Tr20 sag																																			
		Venous Phase Neck		Neck	RoutineSpiralAdultNeckTwinbeam	Neck VNC TwinBeam DE possible [factory]	64 x 0.60	Full	155	Auto (Soft Tissue with Contrast)	AuSn 120	362		(32 cm) 10.1	224	Standard	0.45		0.33		45.0	Neck [factory]																
			Venous Phase Neck 5.00 Br40 I3 sag																				5.00	5.00	Neck VNC TwinBeam DE possible [factory].TwinBeam DE; Neck VNC TwinBeam DE possible [factory].Single Energy	None	Orientation Multi Recons (3)	MPR	Br40	On	3							
			Venous Phase Neck 5.00 Br40 I3 cor																																			
			Venous Phase Neck 5.00 Br40 I3 ax																																			
			Venous Phase Neck 3.00 Qr40 I3 sag																					3.00	2.00	Neck VNC TwinBeam DE possible [factory].TwinBeam DE	DE Parallel Ranges	Orientation Multi Recons (3)	MPR	Qr40	On	3						
			Venous Phase Neck 3.00 Qr40 I3 cor																																			
			Venous Phase Neck 3.00 Qr40 I3 ax																																			
			Venous Phase Neck 1.50 Qr40 I3																						1.50	1.00	Neck VNC TwinBeam DE possible [factory].TwinBeam DE	SPP	None		Qr40	On	3					
Shoulder [factory] (Adult)	>Indications: evaluation of joint cavities, trauma, dislocations, orthopedic indications, shoulder joint, wo CM >Scan modes: RoutineSpiralAdultShoulder >Decision Tree: Side [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard						2.0																
			Topogram 0.60 Tr20 cor																																			
		Shoulder		Shoulder	RoutineSpiralAdultShoulder	Side [factory]	64 x 0.60	Full	130	Auto (Bone/Calcium)	120	100		(32 cm) 8.96	152	Standard	0.80		1.00		2.0																	
			Shoulder 5.00 Br60 I3 ax																					5.00	5.00	Side [factory].recon both; Side [factory].recon right; Side [factory].recon left	None	None	MPR	Br60	On	3						
			Shoulder 3.00 Br60 I3 sag R																				3.00	3.00														
			Shoulder 3.00 Br60 I3 cor R																																			
			Shoulder 3.00 Br60 I3 ax R																																			
			Shoulder 3.00 Br40 I3 sag R																					3.00	3.00	Side [factory].recon both; Side [factory].recon right	Parallel Ranges	Orientation Multi Recons (3)	MPR	Br40	On	3						
			Shoulder 3.00 Br40 I3 cor R																																			
			Shoulder 3.00 Br40 I3 ax R																																			
			Shoulder 3.00 Br60 I3 sag L																						3.00	3.00	Side [factory].recon both; Side [factory].recon left	Parallel Ranges	Orientation Multi Recons (3)	MPR	Br60	On	3					
			Shoulder 3.00 Br60 I3 cor L																																			
			Shoulder 3.00 Br60 I3 ax L																																			
			Shoulder 3.00 Br40 I3 sag L																						3.00	3.00	Side [factory].recon both; Side [factory].recon left	Parallel Ranges	Orientation Multi Recons (3)	MPR	Br40	On	3					
			Shoulder 3.00 Br40 I3 cor L																																			
	Shoulder 3.00 Br40 I3 ax L																																					

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit				
Thorax [factory] (Adult)	>Indications: masses, tumor, metastases, lymphoma, lymph nodes, tuberculosis, empyema, infection, early visualization of pulmonary nodules, screening of the lung, w/wo CM >Scan modes: RoutineSpiralAdultThorax >Decision Tree: Thorax multi [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																		
			Topogram 0.60 Tr20 cor																																				
		Thorax		Thorax	RoutineSpiralAdultThorax	Thorax multi [factory]	64 x 0.60	Full	80	Non-Contrast	120	62		(32 cm) 5.55	184	Standard	1.20		0.33		6.0								Tr20										
			Thorax 5.00 Br40 I3 cor																					5.00	5.00	Thorax multi [factory].wo CM wo X-CARE; Thorax multi [factory].wo CM w X-CARE; Thorax multi [factory].wo CM fast acquisition wo X-CARE; Thorax multi [factory].w CM wo X-CARE; Thorax multi [factory].w CM w X-CARE; Thorax multi [factory].w CM fast acquisition wo X-CARE	None	Orientation Multi Recons (2)	MPR	Br40	On	3							
			Thorax 5.00 Br40 I3 ax																					5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3							
			Thorax 5.00 BI60 I3 sag																																				
			Thorax 5.00 BI60 I3 cor																																				
			Thorax 5.00 BI60 I3 ax																																				
	Thorax 10.00 BI60 I3 cor																							10.00	10.00	Thorax multi [factory].wo CM wo X-CARE; Thorax multi [factory].wo CM w X-CARE; Thorax multi [factory].wo CM fast acquisition wo X-CARE; Thorax multi [factory].w CM wo X-CARE; Thorax multi [factory].w CM w X-CARE; Thorax multi [factory].w CM fast acquisition wo X-CARE	None	None	MIP Thin	BI60	On	3							
	Thorax 1.00 BI60 I3																							1.00	0.70	Thorax multi [factory].wo CM wo X-CARE; Thorax multi [factory].w CM wo X-CARE; Thorax multi [factory].w CM fast acquisition wo X-CARE	Lung CAD	None		BI60	On	3							
Thorax Dynamic [factory] (Adult)	>Indications: abnormalities, upper airway obstruction, trachea bronchomalacia, pulmonary air trapping, abnormal diaphragmatic motion, tumor invasion, routine scan in inspiration w CM, dynamic scan during full breathing wo CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultThorax, DynamicSequenceAdultBody >Decision Tree: Thorax breathhold and X>CARE [factory] >Contrast Protocol: Thorax [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																		
			Topogram 0.60 Tr20 cor																																				
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	120	19		(32 cm) 1.30	1.30	Narrow		0.0	0.33	Full 0.33	2.0								Tr20									
		Monitoring	Premonitoring 10.00 Br36																					10.00				None		Br36									
				Vascular Body																																			
			Monitoring 10.00 Br36																						10.00				None		Br36								
		Thorax		Thorax	RoutineSpiralAdultThorax	Thorax breathhold and X-CARE [factory]	64 x 0.60	Full	80	Auto (Soft Tissue with Contrast)	120	62		(32 cm) 5.55	184	Standard	1.20		0.33				Thorax [factory]																
			Thorax 5.00 Br40 I3 cor																						5.00	5.00	Thorax breathhold and X-CARE [factory].standard acquisition wo X-CARE; Thorax breathhold and X-CARE [factory].standard acquisition w X-CARE; Thorax breathhold and X-CARE [factory].fast acquisition wo X-CARE	None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Thorax 5.00 Br40 I3 ax																																				
			Thorax 5.00 BI60 I3 sag																						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3						
			Thorax 5.00 BI60 I3 cor																																				
			Thorax 5.00 BI60 I3 ax																																				
			Thorax 10.00 BI60 I3 cor																						10.00	10.00	Thorax breathhold and X-CARE [factory].standard acquisition wo X-CARE; Thorax breathhold and X-CARE [factory].standard acquisition w X-CARE; Thorax breathhold and X-CARE [factory].fast acquisition wo X-CARE	None	None	MIP Thin	BI60	On	3						
			Thorax 1.00 BI60 I3																						1.00	0.70	Thorax breathhold and X-CARE [factory].standard acquisition w X-CARE; Thorax breathhold and X-CARE [factory].fast acquisition wo X-CARE	Lung CAD	None		BI60	On	3						
		Dynamic in free breathing		Abdomen	DynamicSequenceAdultBody																																		
			Dynamic in free breathing 5.00 BI60 I3 2.25s - 9.67s																							5.00	5.00												
		Dynamic in free breathing		Abdomen	DynamicSequenceAdultBody																																		
			Dynamic in free breathing 5.00 BI60 I3 2.25s - 9.67s																							5.00	5.00												
		Dynamic in free breathing		Abdomen	DynamicSequenceAdultBody																																		
			Dynamic in free breathing 5.00 BI60 I3 2.25s - 9.67s																							5.00	5.00												
Dynamic in free breathing		Abdomen	DynamicSequenceAdultBody																																				
	Dynamic in free breathing 5.00 BI60 I3 2.25s - 9.67s																							5.00	5.00														

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Thorax and Abdomen [factory] (Adult)	>Indications: masses, tumor, lymphoma, lymph nodes of the lung, liver, pancreas, gall bladder, spleen, kidneys, w CM >Scan modes: RoutineSpiralAdultAbdomenFlexdose >Decision Tree: Thorax and Abdomen contrast medium necessary [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Thorax and Abdomen		Abdomen	RoutineSpiralAdultAbdomenFlexdose	Thorax and Abdomen contrast medium necessary [factory]	64 x 0.60	Full	145	Non-Contrast	120	91		(32 cm) 8.14	510	Standard	0.80		0.33		6.0																
			Thorax and Abdomen 5.00 Br40 I3 cor																				5.00	5.00	Thorax and Abdomen contrast medium necessary [factory].wo CM; Thorax and Abdomen contrast medium necessary [factory].w CM	None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Thorax and Abdomen 5.00 Br40 I3 ax																					5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3					
			Thorax and Abdomen 5.00 BI60 I3 sag Lung																																		
			Thorax and Abdomen 5.00 BI60 I3 cor Lung																																		
			Thorax and Abdomen 5.00 BI60 I3 ax Lung																																		
	Thorax and Abdomen 1.00 BI60 I3																							1.00	0.70	Thorax and Abdomen contrast medium necessary [factory].wo CM; Thorax and Abdomen contrast medium necessary [factory].w CM	Lung CAD	None		BI60	On	3					
Thorax and Abdomen DE Virtual non Contrast [factory] (Adult)	>Indications: visualize contrast agent uptake in thorax, liver, kidneys or other soft tissues, generate Virtual Non Contrast images, w CM >Scan modes: RoutineSpiralAdultAbdomenTwinbeamFlexdose >Decision Tree: Thorax and Abdomen VNC TwinBeam DE possible [factory] >Contrast Protocol: Thorax [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Venous Phase Thorax and Abdomen		Abdomen	RoutineSpiralAdultAbdomenTwinbeamFlexdose	Thorax and Abdomen VNC TwinBeam DE possible [factory]	64 x 0.60	Full	145	Auto (Soft Tissue with Contrast)	AuSn 120	257		(32 cm) 7.14	458	Standard	0.30		0.33		60.0	Thorax [factory]															
			Venous Phase Thorax and Abdomen 5.00 Br40 I3 cor																				5.00	5.00	Thorax and Abdomen VNC TwinBeam DE possible [factory].TwinBeam DE; Thorax and Abdomen VNC TwinBeam DE possible [factory].Single Energy	None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Venous Phase Thorax and Abdomen 5.00 Br40 I3 ax																					5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3					
			Venous Phase Thorax and Abdomen 5.00 BI60 I3 sag Lung																																		
			Venous Phase Thorax and Abdomen 5.00 BI60 I3 cor Lung																																		
			Venous Phase Thorax and Abdomen 5.00 BI60 I3 ax Lung																																		
			Venous Phase Thorax and Abdomen 3.00 Qr40 I3 cor																						3.00	3.00	Thorax and Abdomen VNC TwinBeam DE possible [factory].TwinBeam DE	Parallel Ranges	Orientation Multi Recons (2)	MPR	Qr40	On	3				
			Venous Phase Thorax and Abdomen 3.00 Qr40 I3 ax																																		
			Venous Phase Thorax and Abdomen 3.00 Qr40 I3 cor																						3.00	2.00		DE Parallel Ranges	Orientation Multi Recons (2)	MPR	Qr40	On	3				
			Venous Phase Thorax and Abdomen 3.00 Qr40 I3 ax																																		
			Venous Phase Thorax and Abdomen 1.50 Qr40 I3																						1.50	1.00		SPP	None		Qr40	On	3				
			Venous Phase Thorax and Abdomen 1.00 BI60 I3																						1.00	0.70	Thorax and Abdomen VNC TwinBeam DE possible [factory].TwinBeam DE; Thorax and Abdomen VNC TwinBeam DE possible [factory].Single Energy	Lung CAD	None		BI60	On	3				
Thorax and Abdomen Multi Phase [factory] (Adult)	>Indications: masses, tumor, lymphoma, lymph nodes of the lung, liver, pancreas, gall bladder, spleen, kidneys, w CM, 2 phases preconfigured >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultThorax, RoutineSpiralAdultAbdomen >Contrast Protocol: Thorax [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard					6.0																
			Topogram 0.60 Tr20 cor																																		

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit				
Lung Screening [factory] (Adult)	>Indications: masses, tumor, metastases, lymphoma, lymph nodes, tuberculosis, empyema, infection, early visualization of pulmonary nodules, screening of the lung, w/wo CM >Scan modes: RoutineSpiralAdultThoraxSn >Decision Tree: Lung screening [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																		
			Topogram 0.60 Tr20 cor																																				
		Lung		Thorax	RoutineSpiralAdultThoraxSn	Lung screening [factory]	64 x 0.60	Full	15	Non-Contrast	Sn 100	179		(32 cm) 0.90	31.4	Standard	0.80		0.33		6.0									Tr20									
			Lung 5.00 Br40 I3 cor																					5.00	5.00	Lung screening [factory].low dose wo X-CARE; Lung screening [factory].low dose w X-CARE; Lung screening [factory].high resolution wo X-CARE; Lung screening [factory].high resolution w X-CARE	None	Orientation Multi Recons (2)	MPR	Br40	On	3							
			Lung 5.00 Br40 I3 ax																																				
			Lung 5.00 BI60 I3 sag																					5.00	5.00	Lung screening [factory].low dose wo X-CARE; Lung screening [factory].low dose w X-CARE; Lung screening [factory].low dose w X-CARE	None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3							
			Lung 5.00 BI60 I3 cor																																				
			Lung 5.00 BI60 I3 ax																																				
			Lung 2.00 BI64 I3 sag																																				
			Lung 2.00 BI64 I3 cor																																				
			Lung 2.00 BI64 I3 ax																																				
	Lung 10.00 BI60 I3 cor																							10.00	10.00	Lung screening [factory].low dose wo X-CARE; Lung screening [factory].low dose w X-CARE; Lung screening [factory].high resolution wo X-CARE; Lung screening [factory].high resolution w X-CARE	None	None	MIP Thin	BI60	On	3							
	Lung 1.00 BI60 I3																							1.00	0.70	Lung screening [factory].high resolution wo X-CARE; Lung screening [factory].high resolution w X-CARE	Lung CAD	None		BI60	On	3							
Lung Low Dose [factory] (Adult)	>Indications: early visualization of pulmonary nodules, screening of the lung, wo CM >Scan modes: RoutineSpiralAdultThoraxSn >Decision Tree: Lung Low Dose breathhold and X>CARE [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																		
			Topogram 0.60 Tr20 cor																																				
		Lung		Thorax	RoutineSpiralAdultThoraxSn	Lung Low Dose breathhold and X-CARE [factory]	64 x 0.60	Full	15	Auto (Non-Contrast)	Sn 100	179		(32 cm) 0.90	33.0	Standard	1.20		0.33		6.0																		
			Lung 5.00 Br40 I3 sag																					5.00	5.00	Lung Low Dose breathhold and X-CARE [factory].standard acquisition wo X-CARE; Lung Low Dose breathhold and X-CARE [factory].standard acquisition w X-CARE; Lung Low Dose breathhold and X-CARE [factory].fast acquisition wo X-CARE	None	Orientation Multi Recons (3)	MPR	Br40	On	3							
			Lung 5.00 Br40 I3 cor																																				
			Lung 5.00 Br40 I3 ax																																				
			Lung 5.00 BI60 I3 sag																						5.00	5.00	Lung Low Dose breathhold and X-CARE [factory].standard acquisition wo X-CARE; Lung Low Dose breathhold and X-CARE [factory].standard acquisition w X-CARE; Lung Low Dose breathhold and X-CARE [factory].fast acquisition wo X-CARE	None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3						
			Lung 5.00 BI60 I3 cor																																				
			Lung 5.00 BI60 I3 ax																																				
			Lung 10.00 BI60 I3 cor																							10.00	10.00	Lung Low Dose breathhold and X-CARE [factory].standard acquisition wo X-CARE; Lung Low Dose breathhold and X-CARE [factory].fast acquisition wo X-CARE	None	None	MIP Thin	BI60	On	3					
			Lung 1.00 BI60 I3																							1.00	0.70	Lung Low Dose breathhold and X-CARE [factory].standard acquisition w X-CARE; Lung Low Dose breathhold and X-CARE [factory].fast acquisition wo X-CARE	Lung CAD	None		BI60	On	3					
Abdomen [factory] (Adult)	>Indications: masses, tumor, lymphoma, lymph nodes of the liver, pancreas, gall bladder, spleen, kidneys, intestine, w/wo CM >Scan modes: RoutineSpiralAdultAbdomen >Decision Tree: Abdomen contrast medium necessary [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																		
			Topogram 0.60 Tr20 cor																																				
		Abdomen		Abdomen	RoutineSpiralAdultAbdomen	Abdomen contrast medium necessary [factory]	64 x 0.60	Full	145	Non-Contrast	120	112		(32 cm) 10.0	321	Standard	0.80		0.50		6.0																		
			Abdomen 5.00 Br40 I3 cor																						5.00	5.00	Abdomen contrast medium necessary [factory].wo CM; Abdomen contrast medium necessary [factory].w CM	None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Abdomen 5.00 Br40 I3 ax																																				

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit					
Abdomen Multi Phase [factory] (Adult)	>Indications: masses, tumor, lymphoma, lymph nodes of the liver, pancreas, gall bladder, spleen, kidneys, intestine, w and wo CM, 3 phases preconfigured >Scan modes: RoutineSpiralAdultAbdomen, PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAbdomen >Contrast Protocol: Abdomen [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																			
			Topogram 0.60 Tr20 cor																																					
		Non Contrast Liver		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	120	112		(32 cm) 10.0	221	Standard	0.80		0.50		6.0										Tr20									
			Non Contrast Liver 5.00 Br40 I3 cor																				5.00	5.00			None	Orientation Multi Recons (2)	MPR	Br40	On	3								
			Non Contrast Liver 5.00 Br40 I3 ax																																					
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.50	Full 0.50	2.0																			
		Monitoring	Premonitoring 10.00 Br36	Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.50	Full 0.50	10.0	Abdomen [factory]		10.00				None		Br36										
			Monitoring 10.00 Br36																					10.00				None		Br36										
		Arterial Phase Abdomen		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Full	145	Auto (Soft Tissue with Contrast)	120	112		(32 cm) 10.0	321	Standard	0.80		0.50					Abdomen [factory]		10.00				None		Br36								
			Arterial Phase Abdomen 5.00 Br40 I3 cor																					5.00	5.00			None	Orientation Multi Recons (2)	MPR	Br40	On	3							
Arterial Phase Abdomen 5.00 Br40 I3 ax																																								
Venous Phase Abdomen		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Full	145	Auto (Soft Tissue with Contrast)	120	112		(32 cm) 10.0	221	Standard	0.80		0.50			75.1	Abdomen [factory]																			
	Venous Phase Abdomen 5.00 Br40 I3 cor																					5.00	5.00			None	Orientation Multi Recons (2)	MPR	Br40	On	3									
	Venous Phase Abdomen 5.00 Br40 I3 ax																																							
Abdomen DE Virtual Non Contrast [factory] (Adult)	>Indications: visualize contrast agent uptake in liver, kidneys or other soft tissues, generate Virtual Non Contrast images, w CM >Scan modes: RoutineSpiralAdultAbdomenTwinbeam >Decision Tree: Abdomen VNC TwinBeam DE possible [factory] >Contrast Protocol: Abdomen [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																			
			Topogram 0.60 Tr20 cor																																					
		Venous Phase Abdomen		Abdomen	RoutineSpiralAdultAbdomenTwinbeam	Abdomen VNC TwinBeam DE possible [factory]	64 x 0.60	Full	145	Auto (Soft Tissue with Contrast)	AuSn 120	316		(32 cm) 8.78	300	Standard	0.30		0.33		75.0	Abdomen [factory]								Tr20										
		Venous Phase Abdomen 5.00 Br40 I3 cor																					5.00	5.00	Abdomen VNC TwinBeam DE possible [factory].TwinBeam DE; Abdomen VNC TwinBeam DE possible [factory].Single	None	Orientation Multi Recons (2)	MPR	Br40	On	3									
			Venous Phase Abdomen 5.00 Br40 I3 ax																																					
		Venous Phase Abdomen 3.00 Qr40 I3 cor																					3.00	2.00	Abdomen VNC TwinBeam DE possible [factory].TwinBeam DE	DE Parallel Ranges	Orientation Multi Recons (2)	MPR	Qr40	On	3									
Venous Phase Abdomen 3.00 Qr40 I3 ax																																								
Venous Phase Abdomen 1.50 Qr40 I3																						1.50	1.00			SPP	None		Qr40	On	3									
Liver Perfusion [factory] (Adult)	>Indications: hemorrhage, ischemia, malperfusion, aneurysm, w and wo CM >Scan modes: RoutineSpiralAdultAbdomen, DynamicSpiralAdultBody >Contrast Protocol: Liver Perfusion [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					15.0																			
			Topogram 0.60 Tr20 cor																																					
		Non Contrast Liver		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	120	112		(32 cm) 10.0	321	Standard	0.80		0.50		15.0										Tr20									
		Non Contrast Liver 5.00 Br40 I3 cor																					5.00	5.00			None	Orientation Multi Recons (2)	MPR	Br40	On	3								
			Non Contrast Liver 5.00 Br40 I3 ax																																					
		Perfusion Liver		Abdomen	DynamicSpiralAdultBody		64 x 0.60	Full	85	Soft Tissue with Contrast	80	154		(32 cm) 114	2169	Standard	1.00		0.33		15.0	Liver Perfusion [factory]																		
Perfusion Liver 5.00 Br40 16.61s - 59.20s																						5.00	3.00					Br40	Off	3										
Kidney DE Stones [factory] (Adult)	>Indications: query and analyze kidney stones, wo CM >Scan modes: RoutineSpiralAdultAbdomenTwinSpiral (max. 100 cm) >Decision Tree: Kidney TwinSpiral DE possible [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																			
			Topogram 0.60 Tr20 cor																																					
		Kidney Stones		Abdomen	RoutineSpiralAdultAbdomenTwinSpiral (max. 100 cm)	Kidney TwinSpiral DE possible [factory]	64 x 0.60	Full	145	Auto (Non-Contrast)				(32 cm) 9.62	317	Standard	0.80		0.33		11.0										Tr20									
		Kidney Stones 5.00 Br40 I3 cor																					5.00	5.00	Kidney TwinSpiral DE possible [factory].TwinSpiral DE; Kidney TwinSpiral DE possible [factory].TwinSpiral DE	None	Orientation Multi Recons (2)	MPR	Br40	On	3									
			Kidney Stones 5.00 Br40 I3 ax																																					
		Kidney Stones 1.50 Qr40 I3 sag																					1.50	1.00	Kidney TwinSpiral DE possible [factory].TwinSpiral DE	DE Parallel Ranges	Orientation Multi Recons (3)	MPR	Qr40	On	3									
		Kidney Stones 1.50 Qr40 I3 cor																																						
		Kidney Stones 1.50 Qr40 I3 ax																																						
Kidney Stones 5.00 Qr40 I3 cor																						5.00	5.00	Kidney TwinSpiral DE possible [factory].TwinSpiral DE	None	None	MPR	Qr40	On	3										

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit				
Colon Low Dose [factory] (Adult)	>Indications: early visualization of diverticles, polyps, nodules, screening of the colon, wo CM >Scan modes: RoutineSpiralAdultAbdomenSn	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						6.0																	
			Topogram 0.60 Tr20 cor																																				
		Supine Colon		Abdomen	RoutineSpiralAdultAbdomenSn		64 x 0.60	Full	40	Non-Contrast	Sn 100	422		(32 cm) 2.13	95.3	Standard	0.80		0.50		6.0									Tr20									
			Supine Colon 5.00 Br40 I3 cor																					5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3							
			Supine Colon 5.00 Br40 I3 ax																																				
			Supine Colon 1.50 Br40 I3			ScanplanningTopoAdultSn		6 x 0.60					Sn 100		60	SPR (32 cm) 0.01	0.49	Standard											None	None		Br40	On	3					
			Topogram																				6.0																
				Topogram 0.60 Tr20 cor																																			
			Prone Colon		Abdomen	RoutineSpiralAdultAbdomenSn		64 x 0.60	Full	20	Non-Contrast	Sn 100	235		(32 cm) 1.19	53.1	Standard	0.80		0.50		6.0																	
				Prone Colon 5.00 Br40 I3 cor																						5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3					
	Prone Colon 5.00 Br40 I3 ax																																						
		Prone Colon 1.50 Br40 I3																									None	None		Br40	On	3							
Whole Body [factory] (Adult)	>Indications: degenerative changes, trauma, osteolysis, masses, tumor, lymphoma, lymph nodes of the torso region, w/w/o CM >Scan modes: RoutineSpiralAdultAbdomen >Decision Tree: Whole Body soft tissue or bone [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard						6.0																	
			Topogram 0.60 Tr20 cor																																				
		Whole Body		Abdomen	RoutineSpiralAdultAbdomen	Whole Body soft tissue or bone [factory]	64 x 0.60	Full	145	Non-Contrast	120	112		(32 cm) 10.0	622	Standard	0.80		0.50		6.0																		
			Whole Body 5.00 Br40 I3 cor																					5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3							
			Whole Body 5.00 Br40 I3 ax																																				
			Whole Body 5.00 BI60 I3 sag Lung																						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3						
		Whole Body 5.00 BI60 I3 cor Lung																																					
			Whole Body 5.00 BI60 I3 ax Lung																																				
			Whole Body 1.00 BI60 I3																																				
			Whole Body 5.00 Br60 I3 sag																						5.00	5.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3						
Whole Body 5.00 Br60 I3 cor																																							
	Whole Body 5.00 Br60 I3 ax																																						
	Whole Body 5.00 Br40 I3 ax																																						

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit			
Pelvis DE Bone Marrow [factory] (Adult)	>Indications: fracture, visualisation, and characterization of the bone marrow composition in the pelvis, wo CM >Scan modes: RoutineSpiralAdultPelvis Twinspiral (max. 100 cm) >Decision Tree: Pelvis TwinSpiral DE possible [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard						6.0																
			Topogram 0.60 Tr20 cor																																			
		Pelvis		Pelvis	RoutineSpiralAdultPelvisTwinspiral (max. 100 cm)	Pelvis TwinSpiral DE possible [factory]	64 x 0.60	Full	170	Auto (Non-Contrast)					(32 cm) 11.3	258	Standard	0.80		0.50		11.0																
			Pelvis 5.00 Br60 I3 sag																					5.00	5.00	Pelvis TwinSpiral DE possible [factory].TwinSpiral DE possible [factory].Single Energy	None	Orientation Multi Recons (3)	MPR	Br60	On	3						
			Pelvis 5.00 Br60 I3 cor																																			
			Pelvis 5.00 Br60 I3 ax Pelvis 5.00 Br40 I3 ax																						5.00	5.00	Pelvis TwinSpiral DE possible [factory].TwinSpiral DE possible [factory].Single Energy	None	None	MPR	Br40	On	3					
			Pelvis 1.50 Qr40 I3 ax																					1.50	1.00	Pelvis TwinSpiral DE possible [factory].TwinSpiral DE	DE Parallel Ranges	None	MPR	Qr40	On	3						
Hip [factory] (Adult)	>Indications: evaluation of joint cavity, trauma, dysplasia, necrosis of the head of the hip joints, congruence evaluations, orthopedic indications, wo CM >Scan modes: RoutineSpiralAdultPelvis >Decision Tree: Side [factory]; Metal implant [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard						2.0																
			Topogram 0.60 Tr20 cor																																			
		Hip		Pelvis	RoutineSpiralAdultPelvis	Side [factory]; Metal implant [factory]	64 x 0.60	Full	170	Bone/Calcium	120	131		(32 cm) 11.7	141	Standard	0.80		1.00		2.0																	
			Hip 3.00 Br60 I3 ax																					3.00	3.00	Side [factory].recon both; Side [factory].recon right; Side [factory].recon left; Metal implant [factory].wo iMAR; Metal implant [factory].w iMAR	None	Impression Multi Recons (2)	MPR	Br60	On	3						
			Hip 3.00 Br40 I3 ax																																			
			Hip 3.00 Br60 I3 cor R																					3.00	3.00	Side [factory].recon both; Side [factory].recon right; Metal implant [factory].wo iMAR; Metal implant [factory].w iMAR	Parallel Ranges	None	MPR	Br60	On	3						
			Hip Br40 I3 cor Radial R																																			
			Hip 3.00 Br60 I3 cor L																						3.00	3.00	Side [factory].recon both; Side [factory].recon left; Metal implant [factory].wo iMAR; Metal implant [factory].w iMAR	Parallel Ranges	None	MPR	Br60	On	3					
			Hip Br40 I3 cor Radial L																																			
			Hip 3.00 Br40 I3 ax																					3.00	2.00	Side [factory].recon both; Side [factory].recon right; Side [factory].recon left; Metal implant [factory].w iMAR	None	None	MPR	Br40	On	3						
C-Spine [factory] (Adult)	>Indications: prolapse/hernia, tumors, degenerative changes, bone tumor, trauma, osteolysis of the cervical spine, wo CM >Scan modes: RoutineSpiralAdultNeck >Decision Tree: C>Spine disc or bone [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.06	1.62	Standard						2.0																
			Topogram 0.60 Tr20 sag																																			
		C-Spine		Neck	RoutineSpiralAdultNeck	C-Spine disc or bone [factory]	64 x 0.60	Full	195	Non-Contrast	120	150		(32 cm) 13.4	242	Standard	0.80		1.00		2.0																	
			C-Spine 2.00 Br40 I3 ax																					2.00	2.00	C-Spine disc or bone [factory].standard dose	None	Impression Multi	MPR	Br40	On	3						
			C-Spine 2.00 Br60 I3 ax																									None	MPR	Br40	On	3						
			C-Spine 2.00 Br40 I3 ax Spine																					2.00	2.00		Spine Ranges	None	MPR	Br40	On	3						
			C-Spine 2.00 Br40 I3 sag Spine																																			
			C-Spine 2.00 Br60 I3 ax																					2.00	2.00	C-Spine disc or bone [factory].low dose	None	Impression Multi	MPR	Br40	On	3						
			C-Spine 2.00 Br40 I3 ax																										None	MPR	Br60	On	3					
			C-Spine 2.00 Br60 I3 ax Spine																					2.00	2.00		Spine Ranges	None	MPR	Br60	On	3						
			C-Spine 2.00 Br60 I3 sag Spine																																			
			C-Spine Br40 I3 sag Radial																																			
T-Spine [factory] (Adult)	>Indications: prolapse/hernia, degenerative changes, osteolysis, trauma, tumors of the thoracal spine, wo CM >Scan modes: RoutineSpiralAdultSpine >Decision Tree: T> and L>Spine disc or bone [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard						2.0																
			Topogram 0.60 Tr20 sag																																			
		T-Spine		Spine	RoutineSpiralAdultSpine	T- and L-Spine disc or bone [factory]	64 x 0.60	Full	285	Non-Contrast	120	219		(32 cm) 19.6	353	Standard	0.80		1.00		2.0																	
			T-Spine 3.00 Br40 I3 ax																					3.00	3.00	T- and L-Spine disc or bone [factory].standard dose	None	Impression Multi	MPR	Br40	On	3						
			T-Spine 3.00 Br60 I3 ax																										None	MPR	Br60	On	3					
			T-Spine 3.00 Br40 I3 ax Spine																					3.00	3.00		Spine Ranges	None	MPR	Br40	On	3						
			T-Spine 3.00 Br40 I3 sag Spine																																			
			T-Spine 3.00 Br60 I3 ax																																			
			T-Spine 3.00 Br40 I3 ax																																			
T-Spine 3.00 Br60 I3 ax Spine																																						
T-Spine 3.00 Br60 I3 sag Spine																																						
			T-Spine Br40 I3 sag Radial																																			

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit			
L-Spine [factory] (Adult)	>Indications: prolapse/hernia, tumor, degenerative changes, trauma, osteolysis of the lumbar spine, wo CM >Scan modes: RoutineSpiralAdultSpine >Decision Tree: T> and L>Spine disc or bone [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard					2.0																	
		L-Spine	Topogram 0.60 Tr20 sag	Spine	RoutineSpiralAdultSpine	T- and L-Spine disc or bone [factory]	64 x 0.60	Full	285	Non-Contrast	120	219		(32 cm) 19.6	353	Standard	0.80		1.00		2.0									Tr20								
			L-Spine 3.00 Br40 I3 ax																					3.00	3.00	T- and L-Spine disc or bone [factory].standard dose	None	Impression Multi Recons (2)	MPR	Br40	On	3						
			L-Spine 3.00 Br60 I3 ax																											Br60	On	3						
			L-Spine 3.00 Br40 I3 ax Spine																						3.00		3.00	Spine Ranges	None	MPR	Br40	On	3					
			L-Spine 3.00 Br40 I3 sag Spine																						3.00		3.00	Spine Ranges	None	MPR	Br40	On	3					
			L-Spine 3.00 Br60 I3 ax																						3.00		3.00	T- and L-Spine disc or bone [factory].low dose	None	Impression Multi Recons (2)	MPR	Br60	On	3				
			L-Spine 3.00 Br40 I3 ax																								Br40	On				3						
			L-Spine 3.00 Br60 I3 ax Spine																						3.00	3.00	Spine Ranges	None				MPR	Br60	On	3			
			L-Spine 3.00 Br60 I3 sag Spine																							3.00	3.00	Spine Ranges	None	MPR	Br60	On	3					
			L-Spine Br40 I3 sag Radial																								3.00	3.00	Radial Ranges	None	VRT	Br40	On	3				
Spine DE Bone Marrow [factory] (Adult)	>Indications: fracture, visualisation, and characterization of the bone marrow composition in the spine, wo CM >Scan modes: RoutineSpiralAdultSpineTwinspiral (max. 100 cm) >Decision Tree: Spine TwinSpiral DE possible [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard					2.0																	
		Spine	Topogram 0.60 Tr20 sag	Spine	RoutineSpiralAdultSpineTwinspiral (max. 100 cm)	Spine TwinSpiral DE possible [factory]	64 x 0.60	Full	285	Non-Contrast				(32 cm) 18.7	355	Standard	0.80		1.00		6.0									Tr20								
			Spine 3.00 Br60 I3 ax																					3.00	3.00	Spine TwinSpiral DE possible [factory].TwinSpiral DE; Spine TwinSpiral DE possible [factory].Single Energy	None	Impression Multi	MPR	Br60	On	3						
			Spine 3.00 Br40 I3 ax																								Br40			On	3							
			Spine 3.00 Br60 I3 ax Spine																						3.00		3.00			Spine Ranges	None	MPR					Br60	On
			Spine 3.00 Br60 I3 sag Spine																						3.00	3.00	DE Parallel Ranges	None	MPR	Qr40	On	3						
	Spine 1.50 Qr40 I3 ax																						1.50	1.00	Spine TwinSpiral DE possible [factory].TwinSpiral DE	DE Parallel Ranges	None	MPR	Qr40	On	3							
Spine DE Metal [factory] (Adult)	>Indications: loosening of metal in the spine, wo CM >Scan modes: RoutineSpiralAdultSpineTwinspiralMetal (max. 100 cm) >Decision Tree: Spine TwinSpiral metal DE possible [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard					2.0																	
		Spine	Topogram 0.60 Tr20 sag	Spine	RoutineSpiralAdultSpineTwinspiralMetal (max. 100 cm)	Spine TwinSpiral metal DE possible [factory]	64 x 0.60	Full	285	Auto (Non-Contrast)				(32 cm) 16.6	312	Standard	0.60		0.50		2.0									Tr20								
			Spine 3.00 Br60 I3 ax																					3.00	3.00	Spine TwinSpiral metal DE possible [factory].Single Energy	None	Impression Multi Recons (2)	MPR	Br60	On	3						
			Spine 3.00 Br40 I3 ax																								Br40			On	3							
			Spine 3.00 Qr60 I3 ax																						3.00		3.00			None	Impression Multi Recons (2)	MPR					Qr60	On
			Spine 3.00 Qr40 I3 ax																											Qr40	On	3						
			Spine 3.00 Qr60 I3 ax Spine																						3.00	3.00	Spine TwinSpiral metal DE possible [factory].TwinSpiral DE; Spine TwinSpiral metal DE possible [factory].Single Energy	Spine Ranges	None	MPR	Qr60	On	3					
			Spine 3.00 Qr60 I3 sag Spine																						3.00	3.00		Spine Ranges	None	MPR	Qr60	On	3					
			Spine Qr40 I3 sag Radial																																			
			Spine 3.00 Qr40 I3 ax																						3.00	3.00	Spine TwinSpiral metal DE possible [factory].TwinSpiral DE	DE Radial Ranges	None	VRT	Qr40	On	3					
			Spine 1.50 Qr40 I3																						1.50	1.00		SPP	None		Qr40	On	3					
Osteo [factory] (Adult)	>Indications: quantitative assessment of vertebral bone mineral density (BMD) in the diagnosis and follow-up of osteopenia and osteoporosis, wo CM >Scan modes: RoutineSequenceAdultOsteo2d	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard					2.0																	
		Osteo	Topogram 0.60 Tr20 sag	Osteo	RoutineSequenceAdultOsteo2d		1 x 10.00	Manual kV	80	Auto (Non-Contrast)	80	219		(32 cm) 5.68	5.68	Standard		0.0	1.00	Full 1.00	2.0									Tr20								
			Osteo 10.00 Or40 Evaluation																					10.00				None		Or40								
		Osteo		Osteo	RoutineSequenceAdultOsteo2d		1 x 10.00	Manual kV	80	Auto (Non-Contrast)	80	222		(32 cm) 5.75	5.75	Standard		0.0	1.00	Full 1.00	14.0																	
			Osteo 10.00 Or40 Evaluation																					10.00				None		Or40								
		Osteo		Osteo	RoutineSequenceAdultOsteo2d		1 x 10.00	Manual kV	80	Auto (Non-Contrast)	80	247		(32 cm) 6.40	6.40	Standard		0.0	1.00	Full 1.00	32.0																	
			Osteo 10.00 Or40 Evaluation																					10.00				None		Or40								
Hand [factory] (Adult)	>Indications: trauma, disorders of the hand bones, wo CM >Scan modes: RoutineSpiralAdultExtremitiesSn >Decision Tree: Side [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard					2.0																	
		Hand	Topogram 0.60 Tr20 cor	Extremities	RoutineSpiralAdultExtremitiesSn	Side [factory]	64 x 0.60	Full	75	Auto (Bone/Calcium)	Sn 100	898		(32 cm) 4.53	102	Standard	0.40		1.00		2.0									Tr20								
			Hand 3.00 Br60 I3 sag R																					3.00	3.00	Side [factory].recon both; Side [factory].recon right	None	Orientation Multi	MPR	Br60	On	3						
			Hand 3.00 Br60 I3 cor R																																			
			Hand 3.00 Br60 I3 ax R																																			
			Hand 3.00 Br40 I3 ax R																																			
			Hand 3.00 Br60 I3 sag L																						3.00	3.00	Side [factory].recon both; Side [factory].recon left	None	Orientation Multi	MPR	Br60	On	3					
			Hand 3.00 Br60 I3 cor L																																			
	Hand 3.00 Br60 I3 ax L																																					
	Hand 3.00 Br40 I3 ax L																						3.00	3.00	Side [factory].recon both; Side [factory].recon left	None	None	MPR	Br40	On	3							

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Charact eristic	Scan Mode	Decision Trees	Acquisi tion [mm]	CARE Dose4D & CARE kV	CAR E kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy* cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickn ess [mm]	Incre ment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Stren gth	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Elbow [factory] (Adult)	>Indications: trauma, disorders of the elbow joint, wo CM >Scan modes: RoutineSpiralAdultExtre mitiesSn >Decision Tree: Side [factory]	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard						2.0															
			Topogram 0.60 Tr20 cor																																		
		Elbow		Extremiti es	RoutineSpiral AdultExtremiti esSn	Side [factory]	64 x 0.60	Full	75	Auto (Bone/Calci um)	Sn 100	898		(32 cm) 4.53	56.6	Standard	0.40		1.00		2.0																
			Elbow 2.00 Br60 I3 sag R																				2.00	2.00	Side [factory].recon both; Side [factory].recon right	None	Orientatio n Multi	MPR	Br60	On	3						
			Elbow 2.00 Br60 I3 cor R																																		
			Elbow 2.00 Br60 I3 ax R																																		
			Elbow 2.00 Br40 I3 ax R																					2.00	2.00	Side [factory].recon both; Side [factory].recon right	None	None	MPR	Br40	On	3					
			Elbow 2.00 Br60 I3 sag L																					2.00	2.00	Side [factory].recon both; Side [factory].recon left	None	Orientatio n Multi	MPR	Br60	On	3					
			Elbow 2.00 Br60 I3 cor L																																		

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Charact eristic	Scan Mode	Decision Trees	Acquisi tion [mm]	CARE Dose4D & CARE kV	CAR E kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy* cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickn ess [mm]	Incre ment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Stren gth	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit					
Knee DE Metal [factory] (Adult)	>Indications: loosening of metal in the knee joint, wo CM >Scan modes: RoutineSpiralAdultExtre mitiesTwinspiralMetal (max. 100 cm) >Decision Tree: Side [factory]	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard					2.0																			
			Topogram 0.60 Tr20 cor																																					
		Knee		Extremiti es	RoutineSpiralAdultExtre mitiesTwinspiralM etal (max. 100 cm)	Side [factory]	64 x 0.60	Full	75	Non-Contrast					(32 cm) 5.36	102	Standard	0.60		0.50		2.0																		
			Knee 3.00 Qr60 I3 sag R																					3.00	3.00	Side [factory].recon both; Side [factory].recon right	None	Orientatio n Multi Recons (3)	MPR	Qr60	On	3								
			Knee 3.00 Qr60 I3 cor R																																					
			Knee 3.00 Qr60 I3 ax R Knee 3.00 Qr40 I3 ax R																						3.00	3.00	Side [factory].recon both; Side [factory].recon right	None	None	MPR	Qr40	On	3							
			Knee 0.80 Qr40 I3 cor R Knee 0.80 Qr40 I3 sag R																					0.80	0.50	Side [factory].recon both; Side [factory].recon left		DE Parallel Ranges	Orientatio n Multi Recons (2)	MPR	Qr40	On	3							
			Knee 0.80 Qr40 I3 R Knee 3.00 Qr60 I3 sag L Knee 3.00 Qr60 I3 cor L Knee 3.00 Qr60 I3 ax L																									SPP	None		Qr40	On	3							
			Knee 3.00 Qr40 I3 ax L Knee 0.80 Qr40 I3 cor L Knee 0.80 Qr40 I3 sag L																				3.00	3.00	Side [factory].recon both; Side [factory].recon left		None	None	MPR	Qr40	On	3								
																							0.80	0.50		DE Parallel Ranges	Orientatio n Multi Recons (2)	MPR	Qr40	On	3									
			Knee 0.80 Qr40 I3 L																						0.80	0.50	SPP	None		Qr40	On	3								
		Brain CTA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the vertebral arteries, COW, w CM >Scan modes: PremonitoringSequence Adult2d, MonitoringSequenceAdul t2d, RoutineSpiralAdultAngio Head >Contrast Protocol: CTA [factory]	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.06	1.62	Standard					2.0																	
					Topogram 0.60 Tr20 sag																																			
				Premonitori ng		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																	
				Monitoring	Premonitoring 10.00 Br36																					10.00				None		Br36								
					Monitoring 10.00 Br36	Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]																
				CTA Brain	Monitoring 10.00 Br36																					10.00				None		Br36								
CTA Brain 3.00 Hv36 I3 cor	Head/Vas cular Head				RoutineSpiral AdultAngioHea d		64 x 0.60	Full	105	Auto (Vascular)	90	83		(16 cm) 7.31	111	Standard	1.50		0.33				CTA [factory]			3.00	3.00	Brain CTA TwinBeam DE possible [factory].recon both; Side [factory].recon left	None	Orientatio n Multi Recons (3)	MIP Thin	Hv36	On	3						
CTA Brain 3.00 Hv36 I3 sag																																								
CTA Brain 3.00 Hv36 I3 ax																							10.00																	
Brain DE CTA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the vertebral arteries, COW, w CM >Scan modes: PremonitoringSequence Adult2d, MonitoringSequenceAdul t2d, RoutineSpiralAdultAngio HeadTwinbeam >Decision Tree: Brain CTA TwinBeam DE possible [factory] >Contrast Protocol: CTA [factory]			Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.06	1.62	Standard					2.0																	
					Topogram 0.60 Tr20 sag																																			
				Premonitori ng		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																	
				Monitoring	Premonitoring 10.00 Br36																					10.00				None		Br36								
					Monitoring 10.00 Br36	Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]																
				CTA Brain	Monitoring 10.00 Br36																					10.00				None		Br36								
					CTA Brain 3.00 Hv36 I3 cor	Head/Vas cular Head	RoutineSpiral AdultAngioHea dTwinbeam	Brain CTA TwinBeam DE possible [factory]	64 x 0.60	Full	105	Auto (Vascular)	AuSn 120	262		(16 cm) 14.4	233	Standard	0.35		0.33				CTA [factory]			3.00	3.00	Brain CTA TwinBeam DE possible [factory].TwinBeam DE; Brain CTA TwinBeam DE possible [factory].Single Energy	None	Orientatio n Multi Recons (3)	MIP Thin	Hv36	On	3				
					CTA Brain 3.00 Hv36 I3 sag																																			
		CTA Brain 3.00 Hv36 I3 ax																																						
			CTA Brain 3.00 Qr36 cor																						3.00	3.00	Brain CTA TwinBeam DE possible [factory].TwinBeam DE	Parallel Ranges SPP	None	MIP Thin	Qr36	Off	3							
	CTA Brain 1.50 Qr40 I3																						1.50	1.00			None	None		Hv36	On	3								
Brain DSA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the vertebral arteries, COW, w and wo CM >Scan modes: RoutineSpiralAdultHead, PremonitoringSequence Adult2d, MonitoringSequenceAdul t2d, RoutineSpiralAdultAngio Head >Contrast Protocol: CTA [factory]	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.11	2.87	Standard					2.0																			
			Topogram 0.60 Tr20 sag																																					
		Non Contrast Brain		Brain	RoutineSpiral AdultHead		64 x 0.60	Full	36	Auto (Non-Contrast)	90	51		(16 cm) 4.49	59.6	Standard	0.55		1.00		2.0																			
		Premonitori ng	Non Contrast Brain 1.50 Hv36 I3																					1.50	1.00			None	None		Hv36	On	3							
			Premonitoring 10.00 Br36	Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																			
		Monitoring	Premonitoring 10.00 Br36																					10.00				None		Br36										
			Monitoring 10.00 Br36	Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]																		
		CTA Brain	Monitoring 10.00 Br36																					10.00				None		Br36										
CTA Brain 1.50 Hv36 I3	Head/Vas cular Head		RoutineSpiral AdultAngioHea d		64 x 0.60	Full	105	Auto (Vascular)	90	83		(16 cm) 7.31	111	Standard	1.50		0.33				CTA [factory]																			

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit				
Brain Dynamic CTA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the vertebral arteries, COW, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, DynamicSpiralAdultHeadAngio >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.06	1.62	Standard						2.0																	
			Topogram 0.60 Tr20 sag																																				
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0									Tr20									
			Premonitoring 10.00 Br36																					10.00				None			Br36								
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]																	
			Monitoring 10.00 Br36																					10.00				None			Br36								
	Dynamic CTA Brain		Head/Vascular Head	DynamicSpiralAdultHeadAngio		64 x 0.60	Full	105	Auto (Vascular)	90	83		(16 cm) 120	2475	Standard	1.00		0.33				CTA [factory]																	
	Dynamic CTA Brain 3.00 Hv36 I3 15.67s - 42.44s																						3.00	3.00						Hv36	On	3							
Brain and Neck CTA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the carotid and vertebral arteries, COW, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioNeck >Contrast Protocol: Brain and Neck CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						7.0																	
			Topogram 0.60 Tr20 cor																																				
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																		
			Premonitoring 10.00 Br36																					10.00				None			Br36								
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Brain and Neck CTA [factory]																	
			Monitoring 10.00 Br36																					10.00				None			Br36								
		CTA Brain and Neck		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	145	Auto (Vascular)	100	115		(32 cm) 6.26	270	Standard	1.20		0.33				Brain and Neck CTA [factory]																
			CTA Brain and Neck 3.00 Bv36 I3 sag																					3.00	3.00			None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3						
			CTA Brain and Neck 3.00 Bv36 I3 cor																																				
			CTA Brain and Neck 3.00 Bv36 I3 ax																																				
			CTA Brain and Neck 3.00 Bv36 I3 cor																						3.00	3.00			Parallel Ranges	None	MIP Thin	Bv36	On	3					
			CTA Brain and Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00			Parallel Vascular Ranges	None	MPR	Br40	On	3					
	CTA Brain and Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00			Parallel Vascular Ranges	None	MPR	Br40	On	3							
Brain and Neck DE CTA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the carotid and vertebral arteries, COW, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioNeckTwinbeam >Decision Tree: Neck CTA TwinBeam DE possible [factory] >Contrast Protocol: Brain and Neck CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						7.0																	
			Topogram 0.60 Tr20 cor																																				
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																		
			Premonitoring 10.00 Br36																					10.00				None			Br36								
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Brain and Neck CTA [factory]																	
			Monitoring 10.00 Br36																					10.00				None			Br36								
		CTA Brain and Neck		Neck	RoutineSpiralAdultAngioNeckTwinbeam	Neck CTA TwinBeam DE possible [factory]	64 x 0.60	Full	145	Auto (Vascular)	AuSn 120	353		(32 cm) 9.81	434	Standard	0.45		0.33				Brain and Neck CTA [factory]																
			CTA Brain and Neck 3.00 Bv36 I3 sag																					3.00	3.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE; Neck CTA TwinBeam DE	None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
			CTA Brain and Neck 3.00 Bv36 I3 cor																																				
			CTA Brain and Neck 3.00 Bv36 I3 ax																																				
			CTA Brain and Neck 3.00 Bv36 I3 cor																						3.00	3.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE; Neck CTA TwinBeam DE possible [factory].Single Energy	Parallel Ranges	None	MIP Thin	Bv36	On	3				
			CTA Brain and Neck 3.00 Qr36 cor																						3.00	3.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE	Parallel Ranges	None	MIP Thin	Qr36	Off	3				
	CTA Brain and Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE; Neck CTA TwinBeam DE possible [factory].Single Energy	Parallel Vascular Ranges	None	MPR	Br40	On	3						
	CTA Brain and Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE; Neck CTA TwinBeam DE possible [factory].Single Energy	Parallel Vascular Ranges	None	MPR	Br40	On	3						
	CTA Brain and Neck 1.50 Qr40 I3																						1.50	1.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE	SPP	None		Qr40	On	3						

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit					
Brain and Neck DSA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the carotid and vertebral arteries, COW, w and wo CM >Scan modes: RoutineSpiralAdultAngioNeck, PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioNeck >Contrast Protocol: Brain and Neck CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						7.0																		
			Topogram 0.60 Tr20 cor																																					
		Non Contrast Brain and Neck		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	46	Auto (Non-Contrast)	100	54		(32 cm) 2.94	127	Standard	1.20		0.33			7.0								Tr20										
			Non Contrast Brain and Neck 1.50 Bv36 I3																					1.50	1.00			None	None		Bv36	On	3							
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																			
			Premonitoring 10.00 Br36																					10.00					None		Br36									
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Brain and Neck CTA [factory]										None		Br36						
			Monitoring 10.00 Br36		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	145	Auto (Vascular)	100	115		(32 cm) 6.26	270	Standard	1.20		0.33				Brain and Neck CTA [factory]						None		Br36								
	CTA Brain and Neck 1.50 Bv36 I3																						1.50	1.00			None	None		Bv36	On	3								
Brain and Neck Dynamic CTA [factory] (Adult)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the carotid and vertebral arteries, COW, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, DynamicSpiralAdultNeckAngio >Contrast Protocol: Brain and Neck CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard						7.0																		
			Topogram 0.60 Tr20 sag																																					
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																			
			Premonitoring 10.00 Br36																					10.00					None		Br36									
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Brain and Neck CTA [factory]																		
			Monitoring 10.00 Br36																					10.00					None		Br36									
		Dynamic CTA Brain and Neck		Neck	DynamicSpiralAdultNeckAngio		64 x 0.60	Full	145	Auto (Vascular)	90	120		(32 cm) 61.3	1670	Standard	1.00		0.33				Brain and Neck CTA [factory]										None		Br36					
	Dynamic CTA Brain and Neck 3.00 Bv36 I3 20.95s - 48.22s																						3.00	3.00						Bv36	On	3								
Neck CTA [factory] (Adult)	>Indications: stenosis, occlusion, plaques, abnormalities of the carotid arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioNeck >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						7.0																		
			Topogram 0.60 Tr20 cor																																					
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																			
			Premonitoring 10.00 Br36																					10.00					None		Br36									
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]																		
			Monitoring 10.00 Br36																					10.00					None		Br36									
		CTA Neck		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	145	Auto (Vascular)	100	115		(32 cm) 6.26	207	Standard	1.20		0.33				CTA [factory]									None		Br36						
			CTA Neck 3.00 Bv36 I3 sag																					3.00	3.00			None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3							
			CTA Neck 3.00 Bv36 I3 cor																																					
			CTA Neck 3.00 Bv36 I3 ax																						3.00	3.00			Parallel Ranges	None	MIP Thin	Bv36	On	3						
			CTA Neck 3.00 Bv36 I3 cor																																					
	CTA Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00			Parallel Vascular Ranges	None	MPR	Br40	On	3								
	CTA Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00			Parallel Vascular Ranges	None	MPR	Br40	On	3								
Neck DE CTA [factory] (Adult)	>Indications: stenosis, occlusion, plaques, abnormalities of the carotid arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioNeckTwinbeam >Decision Tree: Neck CTA TwinBeam DE possible [factory] >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						7.0																		
			Topogram 0.60 Tr20 cor																																					
		Premonitoring		Vascular			1 x	Manual	25	Soft Tissue	100	26		(32 cm)	1.06	Narrow		0.0	0.33	Full	2.0																			
			Premonitoring 10.00 Br36																					10.00						None		Br36								
		Monitoring		Vascular			1 x	Manual	25	Soft Tissue	100	26		(32 cm)	31.8	Narrow		0.0	0.33	Full	10.0	CTA																		
			Monitoring 10.00 Br36																					10.00						None		Br36								
		CTA Neck		Neck	RoutineSpiralAdultAngioNeckTwinbeam	Neck CTA	64 x	Full	145	Auto	AuSn	353		(32 cm)	336	Standard	0.45		0.33				CTA																	
			CTA Neck 3.00 Bv36 I3 sag																					3.00	3.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE; Neck CTA TwinBeam DE possible [factory].Single Energy	None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3						
			CTA Neck 3.00 Bv36 I3 cor																																					
			CTA Neck 3.00 Bv36 I3 ax																																					
			CTA Neck 3.00 Bv36 I3 cor																						3.00	3.00			Neck CTA TwinBeam DE possible [factory].TwinBeam DE	Parallel	None	MIP	Bv36	On	3					
			CTA Neck 3.00 Qr36 cor																						3.00	3.00			Parallel	None	MIP	Qr36	Off	3						
			CTA Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00			Parallel Vascular Ranges	None	MPR	Br40	On	3						
	CTA Neck 3.00 Br40 I3 Parallel Vascular																							3.00	3.00			Parallel Vascular Ranges	None	MPR	Br40	On	3							
	CTA Neck 1.50 Qr40 I3																							1.50	1.00			SPP	None		Qr40	On	3							

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Charact eristic	Scan Mode	Decision Trees	Acquisi tion [mm]	CARE Dose4D & CARE kV	CAR E kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy* cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickn ess [mm]	Incre ment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Stren gth	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Neck DSA [factory] (Adult)	>Indications: stenosis, occlusion, plaques, abnormalities of the carotid arteries, w and wo CM >Scan modes: RoutineSpiralAdultAngioNeck, PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioNeck >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard						7.0															
			Topogram 0.60 Tr20 cor																											Tr20							
		Non Contrast Neck		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	46	Auto (Non-Contrast)	100	54		(32 cm) 2.94	97.2	Standard	1.20		0.33		7.0																
			Non Contrast Neck 1.50 Bv36 I3																				1.50	1.00			None	None		Bv36	On	3					
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																
			Premonitoring 10.00 Br36																					10.00					None		Br36						
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]															
			Monitoring 10.00 Br36		Neck	RoutineSpiralAdultAngioNeck		64 x 0.60	Full	145	Auto (Vascular)	100	115		(32 cm) 6.26	207	Standard	1.20		0.33				CTA [factory]				None		Br36							
	CTA Neck 1.50 Bv36 I3																					1.50	1.00			None	None		Bv36	On	3						
Thorax CTA [factory] (Adult)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the thoracic aorta, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioBody >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						6.0															
			Topogram 0.60 Tr20 cor																											Tr20							
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																
			Premonitoring 10.00 Br36																				10.00					None		Br36							
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]															
			Monitoring 10.00 Br36		Vascular Body	RoutineSpiralAdultAngioBody		64 x 0.60	Full	70	Auto (Vascular)	100	62		(32 cm) 3.38	112	Standard	1.50		0.33				CTA [factory]				None		Br36							
			CTA Thorax 5.00 Bv36 I3 sag																					5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
			CTA Thorax 5.00 Bv36 I3 cor																																		
			CTA Thorax 5.00 Bv36 I3 ax																																		
			CTA Thorax 5.00 BI60 I3 ax																						5.00	5.00		None	Impression Multi Recons (2)	MIP Thin	BI60	On	3				
			CTA Thorax 5.00 Br40 I3 ax																											Br40	On	3					
	CTA Thorax 3.00 Bv36 I3 Radial Vascular																						3.00			Radial Vascular Ranges	None	MIP Thin	Bv36	On	3						
Thorax Dynamic CTA [factory] (Adult)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the thoracic aorta, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, DynamicSpiralAdultBodyAngio >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						15.0															
			Topogram 0.60 Tr20 cor																											Tr20							
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																
			Premonitoring 10.00 Br36																				10.00					None		Br36							
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]															
			Monitoring 10.00 Br36																					10.00					None		Br36						
	Dynamic CTA Thorax		Vascular Body	DynamicSpiralAdultBodyAngio		64 x 0.60	Full	70	Auto (Vascular)	90	70		(32 cm) 26.6	873	Standard	1.00		0.33				CTA [factory]															
			Dynamic CTA Thorax 5.00 Bv36 I3 28.09s - 55.09s																				5.00	3.00					Bv36	On	3						

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Thorax Coronary CTA [factory] (Adult)	>Indications: chest pain, triple rule out, vascular anomalies, aneurysms, dissection, ruptures of the thoracic aorta, pulmonary embolism, stenosis of the coronary arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, CardiacSpiralAdultAngioFlexdose >Decision Tree: Thorax Coronary CTA HR variability and HR [factory] >Contrast Protocol: Coronary CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					10.0																
		Topogram 0.60 Tr20 cor																																			
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																
		Premonitoring 10.00 Br36					1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Coronary CTA [factory]		10.00				None									
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Coronary CTA [factory]		10.00				None									
		Monitoring 10.00 Br36		Cardio	CardiacSpiralAdultAngioFlexdose	Thorax Coronary CTA HR variability and HR [factory]	64 x 0.60	Full	64	Auto (Vascular)	100	42		(32 cm) 11.3	397	Narrow	0.31		0.33			Coronary CTA [factory]		10.00				None									
		CTA Thorax Coronary																				Coronary CTA [factory]															
		CTA Thorax Coronary 5.00 Bv36 I3 BestDiast long axis																					5.00	5.00	Thorax Coronary CTA HR variability and HR [factory].standard; Thorax Coronary CTA HR variability and HR [factory].low dose	Parallel Ranges	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3	BestPhase (diastolic)		75	%		
		CTA Thorax Coronary 5.00 Bv36 I3 BestDiast short axis																																			
		CTA Thorax Coronary 5.00 Bv36 I3 BestDiast four chamber																																			
		CTA Thorax Coronary 3.00 Bv40 I3 BestDiast Radial Vascular RCA																					3.00			Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%		
		CTA Thorax Coronary 3.00 Bv40 I3 BestDiast Radial Vascular LAD																					3.00			Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%		
		CTA Thorax Coronary 3.00 Bv40 I3 BestDiast Radial Vascular CX																					3.00			Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%		
		CTA Thorax Coronary 0.80 Bv40 I3 BestDiast																					0.80	0.50		None	None		Bv40	On	3	BestPhase		75	%		
		CTA Thorax Coronary 3.00 Bv36 I3 BestDiast sag PE																					3.00	3.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3	BestPhase (diastolic)		75	%		
		CTA Thorax Coronary 3.00 Bv36 I3 BestDiast cor PE																																			
		CTA Thorax Coronary 3.00 Bv36 I3 BestDiast ax PE																																			
		CTA Thorax Coronary 3.00 BI60 I3 BestDiast ax Lung																																			
		CTA Thorax Coronary 3.00 Br40 I3 BestDiast ax Mediastinum																																			
		CTA Thorax Coronary 5.00 Bv36 I3 BestSyst long axis																																			
		CTA Thorax Coronary 5.00 Bv36 I3 BestSyst short axis																																			
		CTA Thorax Coronary 5.00 Bv36 I3 BestSyst four chamber																																			
		CTA Thorax Coronary 3.00 Bv40 I3 BestSyst Radial Vascular RCA																																			
		CTA Thorax Coronary 3.00 Bv40 I3 BestSyst Radial Vascular LAD																																			
		CTA Thorax Coronary 3.00 Bv40 I3 BestSyst Radial Vascular CX																																			
		CTA Thorax Coronary 0.80 Bv40 I3 BestSyst																																			
Pulmonary CTA [factory] (Adult)	>Indications: pulmonary embolism, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioBody >Contrast Protocol: Pulmonary CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					4.0																
		Topogram 0.60 Tr20 cor																																			
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																
		Premonitoring 10.00 Br36																					10.00					None									
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Pulmonary CTA [factory]		10.00													
		Monitoring 10.00 Br36																					10.00					None									
		CTA Pulmonary		Vascular Body	RoutineSpiralAdultAngioBody		64 x 0.60	Full	70	Auto (Vascular)	100	62		(32 cm) 3.38	112	Standard	1.50		0.33			Pulmonary CTA [factory]															
		CTA Pulmonary 3.00 Bv36 I3 sag																					3.00	3.00			None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
		CTA Pulmonary 3.00 Bv36 I3 cor																																			
		CTA Pulmonary 3.00 Bv36 I3 ax																																			
CTA Pulmonary 3.00 BI60 I3 ax																																					
CTA Pulmonary 3.00 Br40 I3 ax																																					

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Pulmonary DE CTA [factory] (Adult)	>Indications: pulmonary embolism, visualize the iodine enhancement in pulmonary vessels, visualize and quantify the iodine uptake of the lung parenchyma, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioBodyTwinbeam >Decision Tree: Pulmonary and Body CTA TwinBeam DE possible [factory] >Contrast Protocol: Pulmonary CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					4.0																
			Topogram 0.60 Tr20 cor																																		
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																
			Premonitoring 10.00 Br36																					10.00				None		Br36							
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Pulmonary CTA [factory]															
			Monitoring 10.00 Br36																					10.00				None		Br36							
		CTA Pulmonary		Vascular Body	RoutineSpiralAdultAngioBodyTwinbeam	Pulmonary and Body CTA TwinBeam DE possible [factory]	64 x 0.60	Full	70	Auto (Vascular)	AuSn 120	161		(32 cm) 4.47	153	Standard	0.30		0.33				Pulmonary CTA [factory]														
			CTA Pulmonary 3.00 Bv36 I3 sag																					3.00	3.00	Pulmonary and Body CTA TwinBeam DE possible [factory].TwinBeam DE; Pulmonary and Body CTA TwinBeam DE possible [factory].Single Energy	None	Orientation Multi Recons (3)	MPR	Bv36	On	3					
			CTA Pulmonary 3.00 Bv36 I3 cor																																		
			CTA Pulmonary 3.00 Bv36 I3 ax																																		
			CTA Pulmonary 3.00 BI60 I3 ax																					3.00	3.00	Pulmonary and Body CTA TwinBeam DE possible [factory].TwinBeam DE; Pulmonary and Body CTA TwinBeam DE possible [factory].Single Energy	None	Impression Multi Recons (2)	MIP Thin	BI60	On	3					
			CTA Pulmonary 3.00 Br40 I3 ax																											Br40	On	3					
			CTA Pulmonary 3.00 Qr40 I3 sag																					3.00	3.00	Pulmonary and Body CTA TwinBeam DE possible [factory].TwinBeam DE	DE Parallel Ranges	Orientation Multi Recons (3)	MIP Thin	Qr40	On	3					
			CTA Pulmonary 3.00 Qr40 I3 cor																																		
			CTA Pulmonary 3.00 Qr40 I3 ax																																		
			CTA Pulmonary 2.00 Qr40 I3																						2.00	1.50	Pulmonary and Body CTA TwinBeam DE possible [factory].TwinBeam DE	SPP	None		Qr40	On	3				
			CTA Pulmonary 3.00 Bv36 I3 sag																					3.00	3.00												
			CTA Pulmonary 3.00 Bv36 I3 cor																									None	Orientation Multi Recons (3)	MPR	Bv36	On	3				
			CTA Pulmonary 3.00 Bv36 I3 ax																																		
Coronary Calcium Scoring [factory] (Adult)	>Indications: identification, quantification of calcified lesions in the coronary arteries, wo CM >Scan modes: CardiacSequenceAdult >Decision Tree: Coronary Calcium Scoring HR variability [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					10.0																
			Topogram 0.60 Tr20 cor																																		
		Calcium Scoring		Cardio	CardiacSequenceAdult	Coronary Calcium Scoring HR variability [factory]	64 x 0.60	Full	19	Auto (Bone/Calcium)	120	15		(32 cm) 1.76	30.4	Narrow		34.5	0.33	0.23	10.0																
			Calcium Scoring 3.00 Sa36 BestDiast																					3.00	1.50	Coronary Calcium Scoring HR variability [factory].standard	CT CaScoring	None		Sa36		3	BestPhase (diastolic)		75	%	
			Calcium Scoring 3.00 Sa36 BestSyst																						3.00	1.50	Coronary Calcium Scoring HR variability [factory].end systolic arrhythmia	CT CaScoring	None		Sa36		3	BestPhase (systolic)		350	ms

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit				
Coronary CTA [factory] (Adult)	>Indications: identification, quantification of calcified lesions in the coronary arteries, wo CM; stenosis of the coronary arteries, w CM >Scan modes: CardiacSequenceAdult, PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, CardiacSpiralAdultAngio >Decision Tree: Coronary Calcium Scoring HR variability [factory], Coronary CTA HR variability and HR [factory] >Contrast Protocol: Coronary CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						10.0																	
			Topogram 0.60 Tr20 cor																											Tr20									
		Calcium Scoring		Cardio	CardiacSequenceAdult	Coronary Calcium Scoring HR variability [factory]	64 x 0.60	Full	19	Auto (Bone/Calcium)	120	15			(32 cm) 1.76	30.4	Narrow			34.5	0.33	0.23	10.0																
			Calcium Scoring 3.00 Sa36 BestDiast																					3.00	1.50	Coronary Calcium Scoring HR variability [factory].standard	CT CaScoring	None		Sa36		3	BestPhase (diastolic)		75	%			
			Calcium Scoring 3.00 Sa36 BestSyst																					3.00	1.50	Coronary Calcium Scoring HR variability [factory].end systolic arrhythmia	CT CaScoring	None		Sa36		3	BestPhase (systolic)		350	ms			
			Calcium Scoring 3.00 Sa36 I3 BestDiast for CTA planning																					3.00	3.00	Coronary Calcium Scoring HR variability [factory].standard; Coronary Calcium Scoring HR variability [factory].end systolic arrhythmia	None	None		Sa36	On	3	BestPhase (diastolic)		75	%			
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26			(32 cm) 1.06	1.06	Narrow			0.0	0.33	Full 0.33	2.0															
			Premonitoring 10.00 Br36						1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26			(32 cm) 31.8	31.8	Narrow			0.0	0.33	Full 0.33	10.0	Coronary CTA [factory]	10.00				None		Br36						
		Monitoring		Vascular Body					1 x 10.00							(32 cm) 31.8	31.8	Narrow			0.0	0.33	Full 0.33	10.0	Coronary CTA [factory]														
			Monitoring 10.00 Br36																						10.00					None		Br36							
		CTA Coronary		Cardio	CardiacSpiralAdultAngio	Coronary CTA HR variability and HR [factory]	64 x 0.60	Full	64	Auto (Vascular)	100	54				(32 cm) 14.6	205	Narrow	0.31		0.33				Coronary CTA [factory]	10.00				None		Br36							
			CTA Coronary 5.00 Bv36 I3 BestDiast long axis																						5.00	5.00	Coronary CTA HR variability and HR [factory].standard; Coronary CTA HR	Parallel Ranges	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3	BestPhase (diastolic)		75	%		
			CTA Coronary 5.00 Bv36 I3 BestDiast short axis																																				
			CTA Coronary 5.00 Bv36 I3 BestDiast four chamber																																				
			CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular RCA																						3.00		Coronary CTA HR variability and HR [factory].standard; Coronary CTA HR variability and HR [factory].low dose	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%		
			CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular LAD																						3.00			Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%		
			CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular CX																						3.00			Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%		
			CTA Coronary 0.80 Bv40 I3 BestDiast																						0.80	0.50		None	None		Bv40	On	3	BestPhase		75	%		
			CTA Coronary 5.00 Bv36 I3 BestSyst long axis																							5.00	5.00	Coronary CTA HR variability and HR [factory].high heart rate; Coronary CTA HR	Parallel Ranges	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3	BestPhase (systolic)		350	ms	
			CTA Coronary 5.00 Bv36 I3 BestSyst short axis																																				
			CTA Coronary 5.00 Bv36 I3 BestSyst four chamber																																				
			CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular RCA																							3.00		Coronary CTA HR variability and HR [factory].high heart rate; Coronary CTA HR variability and HR [factory].end systolic arrhythmia	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (svstolic)		350	ms	
			CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular LAD																							3.00			Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (systolic)		350	ms	
			CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular CX																							3.00			Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (svstolic)		350	ms	
			CTA Coronary 0.80 Bv40 I3 BestSyst																							0.80	0.50		None	None		Bv40	On	3	BestPhase		350	ms	

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit			
Coronary CTA Function [factory] (Adult)	>Indications: identification, quantification of calcified lesions in the coronary arteries, wo CM; stenosis of the coronary arteries, w CM >Scan modes: CardiacSequenceAdult, PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, CardiacSpiralAdultAngio >Decision Tree: Coronary Calcium Scoring HR variability [factory], Coronary CTA HR variability and HR [factory] >Contrast Protocol: Coronary CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						10.0																
			Topogram 0.60 Tr20 cor																											Tr20								
		Calcium Scoring		Cardio	CardiacSequenceAdult	Coronary Calcium Scoring HR variability	64 x 0.60	Full	19	Auto (Bone/Calcium)	120	15		(32 cm) 1.76	30.4	Narrow		34.5	0.33	0.23	10.0																	
			Calcium Scoring 3.00 Sa36 BestDiast																				3.00	1.50	Coronary Calcium Scoring HR variability [factory].standard	CT CaScoring	None		Sa36		3	BestPhase (diastolic)		75	%			
			Calcium Scoring 3.00 Sa36 BestSyst																				3.00	1.50	Coronary Calcium Scoring HR variability [factory].end systolic arrhythmia	CT CaScoring	None		Sa36		3	BestPhase (systolic)		350	ms			
			Calcium Scoring 3.00 Sa36 I3 BestDiast for CTA planning																				3.00	3.00	Coronary Calcium Scoring HR variability [factory].standard;	None	None		Sa36	On	3	BestPhase (diastolic)		75	%			
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																	
			Premonitoring 10.00 Br36																					10.00				None		Br36								
		Monitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Coronary CTA [factory]																
			Monitoring 10.00 Br36																					10.00				None		Br36								
		CTA Coronary		Cardio	CardiacSpiralAdultAngio	Coronary CTA HR variability and HR [factory]	64 x 0.60	Full	64	Auto (Vascular)	100	54		(32 cm) 14.6	205	Narrow	0.31		0.33			Coronary CTA [factory]																
			CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular RCA																					3.00		Coronary CTA HR variability and HR [factory].standard; Coronary CTA HR variability and HR [factory].low dose	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%		
			CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular LAD																				3.00		Radial Vascular Ranges		None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%			
			CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular CX																				3.00		Radial Vascular Ranges		None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%			
			CTA Coronary 0.80 Bv40 I3 BestDiast																				0.80	0.50	None		None		Bv40	On	3	BestPhase (diastolic)		75	%			
			CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular RCA																					3.00		Coronary CTA HR variability and HR [factory].high heart rate; Coronary CTA HR variability and HR [factory].end systolic arrhythmia	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (systolic)		350	ms		
			CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular LAD																				3.00		Radial Vascular Ranges		None	MIP Thin	Bv40	On	3	BestPhase (systolic)		350	ms			
			CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular CX																				3.00		Radial Vascular Ranges		None	MIP Thin	Bv40	On	3	BestPhase (systolic)		350	ms			
			CTA Coronary 0.80 Bv40 I3 BestSyst																				0.80	0.50	None		None		Bv40	On	3	BestPhase (systolic)		350	ms			
			CTA Coronary 3.00 Bv36 I3 0% - 90%																						3.00	2.00	Coronary CTA HR variability and HR [factory].standard;	None	None		Bv36	On	3	Multi Phase	10	0%; 10%; 20%;	%	

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit					
Coronary CTA HeartFlow [factory] (Adult)	>Indications: identification, quantification of calcified lesions in the coronary arteries, wo CM; stenosis of the coronary arteries, w CM >Scan modes: CardiacSequenceAdult, PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, CardiacSpiralAdultAngio >Decision Tree: Coronary Calcium Scoring HR variability [factory], Coronary CTA HR variability and HR [factory]; Coronary CTA image quality and calcium [factory] >Contrast Protocol: Coronary CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						10.0																		
		Topogram 0.60 Tr20 cor																												Tr20										
		Calcium Scoring		Cardio	CardiacSequenceAdult	Coronary Calcium Scoring HR variability [factory]	64 x 0.60	Full	19	Auto (Bone/Calcium)	120	15		(32 cm) 1.76	30.4	Narrow		34.5	0.33	0.23	10.0																			
		Calcium Scoring 3.00 Sa36 BestDiast																					3.00	1.50	Coronary Calcium Scoring HR variability [factory].standard	CT CaScoring	None		Sa36		3	BestPhase (diastolic)		75	%					
		Calcium Scoring 3.00 Sa36 BestSyst																					3.00	1.50	Coronary Calcium Scoring HR variability [factory].end systolic arrhythmia	CT CaScoring	None		Sa36		3	BestPhase (systolic)		350	ms					
		Calcium Scoring 3.00 Sa36 I3 BestDiast for CTA planning																					3.00	3.00	Coronary Calcium Scoring HR variability [factory].standard; Coronary Calcium Scoring HR variability [factory].end systolic arrhythmia	None	None		Sa36	On	3	BestPhase (diastolic)		75	%					
		Premonitoring		Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0																			
		Monitoring	Premonitoring 10.00 Br36	Vascular Body			1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	Coronary CTA [factory]		10.00				None		Br36										
		Monitoring 10.00 Br36																					10.00				None		Br36											
		CTA Coronary		Cardio	CardiacSpiralAdultAngio	Coronary CTA HR variability and HR [factory]; Coronary CTA image quality and calcium [factory]	64 x 0.60	Full	64	Auto (Vascular)	100	54		(32 cm) 14.6	205	Narrow	0.31		0.33				Coronary CTA [factory]		10.00				None		Br36									
		CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular RCA																					3.00		Coronary CTA HR variability and HR [factory].standard; Coronary CTA HR variability and HR [factory].low dose; Coronary CTA image quality and calcium [factory].standard image quality; Coronary CTA image quality and calcium [factory].increased image quality and calcium [factory].increased image quality (e.g. FFR)	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%					
		CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular LAD																					3.00		Coronary CTA HR variability and HR [factory].standard image quality; Coronary CTA image quality and calcium [factory].increased image quality; Coronary CTA image quality and calcium [factory].increased image quality (e.g. FFR)	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%					
		CTA Coronary 3.00 Bv40 I3 BestDiast Radial Vascular CX																					3.00		Coronary CTA HR variability and HR [factory].increased image quality (e.g. FFR)	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (diastolic)		75	%					
		CTA Coronary 0.80 Bv40 I3 BestDiast																					0.80	0.50	Coronary CTA HR variability and HR [factory].standard; Coronary CTA HR variability and HR [factory].low dose; Coronary CTA image quality and calcium [factory].standard image quality; Coronary CTA image quality and calcium [factory].increased image quality; Coronary CTA image quality and calcium [factory].increased image quality (e.g. FFR)	None	None		Bv40	On	3	BestPhase (diastolic)		75	%					
		CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular RCA																					3.00		Coronary CTA HR variability and HR [factory].high heart rate; Coronary CTA HR variability and HR [factory].end systolic arrhythmia; Coronary CTA image quality and calcium [factory].standard image quality; Coronary CTA image quality and calcium [factory].increased image quality; Coronary CTA image quality and calcium [factory].increased image quality (e.g. FFR)	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (systolic)		350	ms					
		CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular LAD																					3.00		Coronary CTA HR variability and HR [factory].standard image quality; Coronary CTA image quality and calcium [factory].increased image quality; Coronary CTA image quality and calcium [factory].increased image quality (e.g. FFR)	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (systolic)		350	ms					
		CTA Coronary 3.00 Bv40 I3 BestSyst Radial Vascular CX																					3.00		Coronary CTA HR variability and HR [factory].increased image quality and calcium [factory].increased image quality (e.g. FFR)	Radial Vascular Ranges	None	MIP Thin	Bv40	On	3	BestPhase (systolic)		350	ms					
		CTA Coronary 0.80 Bv40 I3 BestSyst																					0.80	0.50	Coronary CTA HR variability and HR [factory].increased image quality (e.g. FFR)	None	None		Bv40	On	3	BestPhase (systolic)		350	ms					
		CTA Coronary 0.60 Bv36 I3 65% - 85%																						0.60	0.40	Coronary CTA HR variability and HR [factory].standard; Coronary CTA HR variability and HR [factory].low dose; Coronary CTA image quality and calcium [factory].increased image quality (e.g. FFR)	None	None		Bv36	On	3	Multi Phase	5	65%; 70%; 75%; 80%; 85%	%				
		CTA Coronary 0.60 Bv36 I3 30% - 45%																						0.60	0.40	Coronary CTA HR variability and HR [factory].high heart rate; Coronary CTA image quality and calcium [factory].increased image quality (e.g. FFR)	None	None		Bv36	On	3	Multi Phase	5	30%; 35%; 40%; 45%	%				
		CTA Coronary 0.60 Bv36 I3 300ms - 400ms																						0.60	0.40	Coronary CTA HR variability and HR [factory].end systolic arrhythmia; Coronary CTA image quality and calcium [factory].increased image quality (e.g. FFR)	None	None		Bv36	On	3	Multi Phase	50	300ms; 350ms; 400ms	ms				

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
TAVI [factory] (Adult)	>Indications: TAVI/TAVR, the evaluation of the access route (peripheral, transapical, subclavian, transaortic), aortic root and aortic annulus dimensions, as well as aortic valve structure and calcification, w CM >Scan modes: TestbolusSequenceAdult2d, CardiacSpiralAdultAngio, RoutineSpiralAdultAngioBodyFast >Decision Tree: Patient Instruction TAVI [factory] >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						10.0															
			Topogram 0.60 Tr20 cor																																		
		Test Bolus		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 15.9	15.9	Narrow		0.0	0.33	Full 0.33	8.0									Tr20						
			Test Bolus 10.00 Br36																					10.00					None			Br36					
		Valve		Cardio	CardiacSpiralAdultAngio	Patient Instruction TAVI [factory]	64 x 0.60	Full	64	Auto (Vascular)	100	54		(32 cm) 14.6	103	Narrow	0.31		0.33			2.0	CTA [factory]						None								
			Valve 0.80 Bv36 I3 BestSyst Aortic Valve																					0.80	0.60	Patient Instruction TAVI [factory].Text_CT_ClinicalDecisionTreeStrategyName_Index_Information_for_Patient_Instruction		None	None		Bv36	On	3	BestPhase (systolic)		30	%
		TAVI Planning		Vascular Body	RoutineSpiralAdultAngioBodyFast		64 x 0.60	Full	70	Auto (Vascular)	100	62		(32 cm) 2.64	141	Narrow	1.50		0.33			10.0	CTA [factory]														
			TAVI Planning 0.80 Bv36 I3																						0.80	0.60			None	None		Bv36	On	3			
			TAVI Planning 3.00 BI60 I3																						3.00	3.00			None	None		BI60	On	3			
	TAVI Planning 3.00 Bv36 I3 Parallel Vascular R																						3.00	3.00			Parallel Vascular Ranges	None	MPR	Bv36	On	3					
	TAVI Planning 3.00 Bv36 I3 Parallel Vascular L																						3.00	3.00				None	MPR	Bv36	On	3					
Abdomen CTA [factory] (Adult)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the abdominal aorta, renal arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioBody >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						6.0															
			Topogram 0.60 Tr20 cor																																		
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0															
			Premonitoring 10.00 Br36																					10.00					None			Br36					
		Monitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]														
			Monitoring 10.00 Br36																					10.00					None			Br36					
		CTA Abdomen		Vascular Body	RoutineSpiralAdultAngioBody		64 x 0.60	Full	115	Auto (Vascular)	100	101		(32 cm) 5.50	238	Standard	1.50		0.33				CTA [factory]														
			CTA Abdomen 5.00 Bv36 I3 sag																					5.00	5.00			None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
			CTA Abdomen 5.00 Bv36 I3 cor																																		
	CTA Abdomen 5.00 Bv36 I3 ax																																				
Abdomen DE CTA [factory] (Adult)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the abdominal aorta, renal arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioBodyTwinbeam >Decision Tree: Pulmonary and Body CTA TwinBeam DE possible [factory] >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						6.0															
			Topogram 0.60 Tr20 cor																																		
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0															
			Premonitoring 10.00 Br36																					10.00					None			Br36					
		Monitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]														
			Monitoring 10.00 Br36																					10.00					None			Br36					
		CTA Abdomen		Vascular Body	RoutineSpiralAdultAngioBodyTwinbeam	Pulmonary and Body CTA TwinBeam DE possible [factory]	64 x 0.60	Full	115	Auto (Vascular)	AuSn 120	262		(32 cm) 7.28	249	Standard	0.45		0.33				CTA [factory]														
			CTA Abdomen 5.00 Bv36 I3 sag																					5.00	5.00	Pulmonary and Body CTA TwinBeam DE possible [factory].TwinBeam DE; Pulmonary and Body CTA TwinBeam DE possible [factory].Single Energy		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
			CTA Abdomen 5.00 Bv36 I3 cor																																		
	CTA Abdomen 5.00 Bv36 I3 ax																																				
	CTA Abdomen 3.00 Bv36 I3 ax																						3.00	3.00	Pulmonary and Body CTA TwinBeam DE possible [factory].TwinBeam DE		Parallel Ranges	None	MIP Thin	Bv36	On	3					
	CTA Abdomen Qr40 I3 cor Radial																										DE Radial Ranges	None		Qr40	On	3					
	CTA Abdomen 1.50 Qr40 I3																						1.50	1.00			SPP	None		Qr40	On	3					
Whole Body CTA [factory] (Adult)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the whole aorta, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioBody >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard						6.0															
			Topogram 0.60 Tr20 cor																																		
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0															
			Premonitoring 10.00 Br36																					10.00					None			Br36					
		Monitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]														
			Monitoring 10.00 Br36																					10.00					None			Br36					
		CTA Whole Body		Vascular Body	RoutineSpiralAdultAngioBody		64 x 0.60	Full	115	Auto (Vascular)	100	101		(32 cm) 5.50	348	Standard	1.50		0.33				CTA [factory]														
			CTA Whole Body 5.00 Bv36 I3 sag																					5.00	5.00			None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
			CTA Whole Body 5.00 Bv36 I3 cor																																		
	CTA Whole Body 5.00 Bv36 I3 ax																																				
	CTA Whole Body 3.00 Bv36 I3 Radial Vascular																						3.00				Radial Vascular Ranges	None		MIP Thin	Bv36	On	3				

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Extremity CTA [factory] (Adult)	>Indications: vascular anomalies, aneurysms of the femoral arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioExtremities >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard						2.0															
			Topogram 0.60 Tr20 cor																																		
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.50	Full 0.50	2.0															
			Premonitoring 10.00 Br36																					10.00				None		Br36							
		Monitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.50	Full 0.50	10.0	CTA [factory]														
			Monitoring 10.00 Br36																						10.00				None		Br36						
		CTA Extremity		Runoff	RoutineSpiralAdultAngioExtremities		64 x 0.60	Full	155	Auto (Vascular)	100	136		(32 cm) 7.41	597	Standard	0.35		0.50				CTA [factory]														
			CTA Extremity 5.00 Bv36 I3 sag																					5.00	5.00			None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
			CTA Extremity 5.00 Bv36 I3 cor																																		
			CTA Extremity 5.00 Bv36 I3 ax																																		
	CTA Extremity 3.00 Bv36 I3 Radial Vascular R																						3.00				Radial Vascular Ranges	None	MIP Thin	Bv36	On	3					
	CTA Extremity 3.00 Bv36 I3 Radial Vascular L																						3.00				Radial Vascular Ranges	None	MIP Thin	Bv36	On	3					
Extremity DE CTA [factory] (Adult)	>Indications: vascular anomalies, aneurysms of the femoral arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, RoutineSpiralAdultAngioExtremitiesTwinbeam >Decision Tree: Extremity CTA TwinBeam DE possible [factory] >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard						2.0															
			Topogram 0.60 Tr20 cor																																		
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.50	Full 0.50	2.0															
			Premonitoring 10.00 Br36																						10.00				None		Br36						
		Monitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.50	Full 0.50	10.0	CTA [factory]														
			Monitoring 10.00 Br36																							10.00				None		Br36					
		CTA Extremity		Runoff	RoutineSpiralAdultAngioExtremitiesTwinbeam	Extremity CTA TwinBeam DE possible [factory]	64 x 0.60	Full	155	Auto (Vascular)	AuSn 120	352		(32 cm) 9.78	824	Standard	0.35		0.50				CTA [factory]														
			CTA Extremity 5.00 Bv36 I3 sag																					5.00	5.00	Extremity CTA TwinBeam DE possible [factory].TwinBeam DE; Extremity CTA TwinBeam	None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
			CTA Extremity 5.00 Bv36 I3 cor																																		
			CTA Extremity 5.00 Bv36 I3 ax																																		
	CTA Extremity 3.00 Bv36 I3 Radial Vascular R																						3.00		Extremity CTA TwinBeam DE possible [factory].TwinBeam DE; Extremity CTA TwinBeam DE possible [factory].Single Energy	Radial Vascular Ranges	None	MIP Thin	Bv36	On	3						
	CTA Extremity 3.00 Bv36 I3 Radial Vascular L																						3.00		Extremity CTA TwinBeam DE possible [factory].TwinBeam DE possible [factory].Single Energy	Radial Vascular Ranges	None	MIP Thin	Bv36	On	3						
	CTA Extremity 3.00 Bv36 I3 ax																						3.00	3.00	Extremity CTA TwinBeam DE possible [factory].TwinBeam DE possible [factory].TwinBeam DE possible [factory].TwinBeam DE	Parallel Ranges	None	MIP Thin	Bv36	On	3						
	CTA Extremity Qr40 I3 cor Radial																								Extremity CTA TwinBeam DE possible [factory].TwinBeam DE possible [factory].TwinBeam DE	DE Radial Ranges	None	MIP	Qr40	On	3						
	CTA Extremity 1.50 Qr40 I3																						1.50	1.00	Extremity CTA TwinBeam DE possible [factory].TwinBeam DE	SPP	None		Qr40	On	3						
Extremity Dynamic CTA [factory] (Adult)	>Indications: stenosis, occlusion, plaques, abnormalities, ruptures of the femoral arteries, w CM >Scan modes: PremonitoringSequenceAdult2d, MonitoringSequenceAdult2d, DynamicSpiralAdultBodyAngio >Contrast Protocol: CTA [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard						2.0															
			Topogram 0.60 Tr20 cor																																		
		Premonitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 1.06	1.06	Narrow		0.0	0.33	Full 0.33	2.0															
			Premonitoring 10.00 Br36																						10.00				None		Br36						
		Monitoring		Vascular Body				1 x 10.00	Manual kV	25	Soft Tissue with Contrast	100	26		(32 cm) 31.8	31.8	Narrow		0.0	0.33	Full 0.33	10.0	CTA [factory]														
			Monitoring 10.00 Br36																							10.00				None		Br36					
Dynamic CTA Extremity		Vascular Body	DynamicSpiralAdultBodyAngio		64 x 0.60	Full	110	Auto (Vascular)	90	109		(32 cm) 48.8	2133	Standard	1.50		0.33				CTA [factory]																
	Dynamic CTA Extremity 5.00 Bv36 I3 18.08s - 45.01s																						5.00	3.00					Bv36	On	3						
Head Trauma [factory] (Adult)	>Indications: emergency head trauma, hemorrhage, wo CM >Scan modes: RoutineSpiralAdultHead	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (16 cm) 0.02	0.56	Standard						2.0															
			Topogram 0.60 Tr20 cor																																		
		Head		Brain	RoutineSpiralAdultHead		64 x 0.60	Off			120	379		(16 cm) 71.1	944	Standard	0.55		0.50			2.0															
			Head 5.00 Hr40 ax																						5.00	5.00		None	Orientation Multi	MPR	Hr40	Off	2				
			Head 5.00 Hr40 sag																																		
			Head 5.00 Hr40 cor																																		
			Head 4.00 Hr40 BrainHem																						4.00	4.00		Brain Hemorrhage	None			Off	2				
			Head 1.50 Hr40 SkullUnf																						1.50	0.60		Skull Unfolding	None		Hr40	Off	3				
			Head 0.80 Hr60 ax																						0.80	0.60		None	Orientation Multi	MPR	Hr60	Off	2				
			Head 0.80 Hr60 sag																																		
	Head 0.80 Hr60 cor																																				
	Head 0.80 Hr60 SkullUnf																						0.80	0.60		Skull Unfolding	None		Hr60	Off	2						

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Charact eristic	Scan Mode	Decision Trees	Acquisi tion [mm]	CARE Dose4D & CARE kV	CAR E kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy* cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickn ess [mm]	Incre ment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Stren gth	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit				
Body Trauma [factory] (Adult)	>Indications: fast screening for emergency whole body trauma, hemorrhage, w/w/o CM >Scan modes: RoutineSpiralAdultAbdo men	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.87	Standard						6.0																	
			Topogram 0.60 Tr20 cor																												Tr20								
		Body		Abdomen	RoutineSpiral AdultAbdomen		64 x 0.60	Full	207	Auto (Non-Contrast)	120	159		(32 cm) 14.2	875	Standard	0.60		0.33		6.0																		
			Body 1.50 Br40 ax																					1.50	1.00			None	Impressi on Multi Recons (3)	MPR	Br40	Off	3						
			Body 1.50 Br60 ax																											Br60	Off	3							
			Body 1.50 BI60 ax Body Br60 Rib																									Radial Rib Ranges	None	MPR	BI60 Br60	Off	3						
Poly Trauma [factory] (Adult)	>Indications: emergency whole body trauma, hemorrhage, w/w/o CM >Scan modes: RoutineSpiralAdultHead, RoutineSpiralAdultNeck, RoutineSpiralAdultAbdo men	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.09	4.67	Standard						2.0																	
			Topogram 0.60 Tr20 sag																												Tr20								
		Head		Brain	RoutineSpiral AdultHead		64 x 0.60	Full	429	Auto (Non-Contrast)	120	330		(16 cm) 61.9	699	Standard	0.55		1.00		2.0																		
			Head 5.00 Hr40 ax																					5.00	5.00			None	Orientatio n Multi Recons (3)	MPR	Hr40	Off	3						
			Head 5.00 Hr40 sag																																				
			Head 5.00 Hr40 cor																																				
			Head 4.00 Hr40 BrainHem																					4.00	4.00			Brain Hemorrhag e	None			Off	3						
			Head 1.50 Hr40 SkullUnf																						1.50	0.60			Skull Unfolding	None		Hr40	Off	3					
			Head 5.00 Hr60																						5.00	5.00			None	None		Hr60	Off	3					
			Head 0.80 Hr60 SkullUnf																						0.80	0.60			Skull Unfolding	None		Hr60	Off	3					
		Neck		Neck	RoutineSpiral AdultNeck		64 x 0.60	Full	222	Auto (Non-Contrast)	120	171		(32 cm) 15.3	276	Standard	0.80		1.00		12.5																		
			Neck 5.00 Br40 ax																						5.00	5.00			None	Orientatio n Multi Recons (3)	MPR	Br40	Off	3					
			Neck 5.00 Br40 sag																																				
			Neck 5.00 Br40 cor																																				
			Neck 5.00 Br60 sag																																				
		Topogram			ScanplanningT opoAdultSn		6 x 0.60					Sn 100		60	SPR (32 cm) 0.01	0.87	Standard							6.0						None	None	MPR	Br60	Off	3				
			Topogram 0.60 Tr20 cor																													Tr20							
		Thorax and Abdomen		Abdomen	RoutineSpiral AdultAbdomen		64 x 0.60	Full	207	Auto (Non-Contrast)	120	159		(32 cm) 14.2	875	Standard	0.60		0.50		6.0																		
			Thorax and Abdomen 5.00 Br40 ax																						5.00	5.00			None	Orientatio n Multi Recons (2)	MPR	Br40	Off	3					
			Thorax and Abdomen 5.00 Br40 cor																																				
			Thorax and Abdomen 5.00 BI60 sag																						5.00	5.00			None	Orientatio n Multi Recons (3)	MIP Thin	BI60	Off	3					
			Thorax and Abdomen 5.00 BI60 cor																																				
			Thorax and Abdomen 5.00 BI60 ax																																				
			Thorax and Abdomen 5.00 Br60 sag																						5.00	5.00			None	Orientatio n Multi Recons (2)	MPR	Br60	Off	3					
			Thorax and Abdomen 5.00 Br60 cor																																				
	Thorax and Abdomen Br60 Rib																										Radial Rib Ranges	None	MPR	Br60	Off	3							
RTP Head [factory] (Adult)	>Indications: radiotherapy planning of the head region, w/w/o CM >Scan modes: RoutineSpiralAdultHead >Decision Tree: RTP Brain [factory]	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.09	4.67	Standard						2.0																	
		Head		Brain	RoutineSpiral AdultHead	RTP Brain [factory]	64 x 0.60	Off			120	265		(16 cm) 49.7	809	Standard	0.55		1.00		2.0										Tr20								
			Head 2.00 Qr40 I2																					2.00	2.00			RTP Brain [factory].standard	None	None		Qr40	On	2					
			Head 1.00 Qr40 I2																						1.00	1.00			[factory].Text_CT_ClinicalD ecisionTreeStrategyName_I ndex_stereotactic	None	None		Qr40	On	2				
RTP Head DE [factory] (Adult)	>Indications: radiotherapy planning of the head region for dose calculation and in addition monoenergetic plus images for contouring wo CM >Scan modes: RoutineSpiralAdultHead, RoutineSpiralAdultHeadT winspiral (max. 100 cm) >Decision Tree: TwinSpiral contrast medium [factory]	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 140		60	SPR (16 cm) 0.09	4.67	Standard						2.0																	
			Topogram 0.60 Tr20 sag																												Tr20								
		Head for dose calculation		Brain	RoutineSpiral AdultHead		64 x 0.60	Off			120	265		(16 cm) 49.7	809	Standard	0.55		1.00		2.0																		
		Head for contouring		Brain	RoutineSpiral AdultHeadTwi nspiral (max. 100 cm)	TwinSpiral contrast medium [factory]	64 x 0.60	Off						(16 cm) 48.8	847	Standard	0.60		1.00		2.0																		
			Head for contouring 1.50 Qr40 I2																						1.50	1.50			TwinSpiral contrast medium [factory].Information for TwinSpiral	None	None		Qr40	On	2				
RTP Head and Neck [factory] (Adult)	>Indications: radiotherapy planning of the head, neck ,shoulder region, w/w/o CM >Scan modes: RoutineSpiralAdultNeck	Topogram			ScanplanningT opoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						2.0																	
			Topogram 0.60 Tr20 cor																												Tr20								
		Head and Neck		Neck	RoutineSpiral AdultNeck		64 x 0.60	Manual kV	155	Auto (Non-Contrast)	120	119		(32 cm) 10.7	341	Standard	0.80		1.00		2.0																		
			Head and Neck 1.50 Qr40 I3																						1.50	1.50			None	None		Qr40	On	3					

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
RTP Head and Neck DE [factory] (Adult)	>Indications: radiotherapy planning of the head, neck and shoulder region dose calculation and in addition monoenergetic plus images for contouring w/o CM >Scan modes: RoutineSpiralAdultNeck, RoutineSpiralAdultNeckTwinspiral (max. 100 cm) >Decision Tree: TwinSpiral contrast medium [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0																
			Topogram 0.60 Tr20 cor																																		
		Head and Neck for dose calculation		Neck	RoutineSpiralAdultNeck		64 x 0.60	Manual kV	155	Auto (Non-Contrast)	120	119		(32 cm) 10.7	341	Standard	0.80		1.00		2.0																
		Head and Neck for dose calculation 1.50 Qr40 I3																					1.50	1.50		None	None			Qr40	On	3					
		Head and Neck for contouring		Neck	RoutineSpiralAdultNeckTwinspiral (max. 100 cm)	TwinSpiral contrast medium [factory]	64 x 0.60	Manual kV	155	Auto (Non-Contrast)				(32 cm) 10.5	348	Standard	0.80		0.50		2.0																
			Head and Neck for contouring 1.50 Qr40 I3																				1.50	1.50		TwinSpiral contrast medium [factory].Information for TwinSpiral	None	None		Qr40	On	3					
RTP Thorax [factory] (Adult)	>Indications: radiotherapy planning of the thorax region, w/w/o CM >Scan modes: RoutineSpiralAdultThorax	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0																
			Topogram 0.60 Tr20 cor																																		
		Thorax		Thorax	RoutineSpiralAdultThorax		64 x 0.60	Manual kV	80	Auto (Non-Contrast)	120	62		(32 cm) 5.55	178	Standard	0.80		0.33		2.0																
			Thorax 1.50 Qr40 I3																					1.50	1.50		None	None		Qr40	On	3					
RTP Respiratory Motion Management [factory] (Adult)	>Indications: motion management for radiotherapy planning for moving tumors, w/w/o CM >Scan modes: RespiratorySpiralAdult	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0																
			Topogram 0.60 Tr20 cor																																		
		Respiratory Motion		Respiratory	RespiratorySpiralAdult		64 x 0.60	Manual kV	26	Auto (Non-Contrast)	120	20		(32 cm) 39.0	1238	Standard	0.06		0.33		2.0																
			Respiratory Motion 1.50 Qr40 I3 Average CT Respiratory Motion 1.50 Qr40 I3 100% Ex - 80% In																					1.50	1.50		None	None		Qr40	On	3	Average CT				
																							1.50	1.50		None	None		Qr40	On	3	Multi Phase	10	100% Ex; 80% Ex; 60% Ex; 40% Ex; 20% Ex; 0% Ex; 20% In; 40% In; 60% In; 80% In	Amplitude (% Ex, % In)		
RTP Abdomen [factory] (Adult)	>Indications: radiotherapy planning of the abdomen region, w/w/o CM >Scan modes: RoutineSpiralAdultAbdomen	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0																
			Topogram 0.60 Tr20 cor																																		
		Abdomen		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Manual kV	145	Auto (Non-Contrast)	120	112		(32 cm) 10.0	321	Standard	0.80		0.50		2.0																
			Abdomen 1.50 Qr40 I3																				1.50	1.50		None	None		Qr40	On	3						
RTP Abdomen DE [factory] (Adult)	>Indications: radiotherapy planning of the abdomen region, w/o CM >Scan modes: RoutineSpiralAdultAbdomenTwinspiral (max. 100 cm)	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0																
			Topogram 0.60 Tr20 cor																																		
		Abdomen		Abdomen	RoutineSpiralAdultAbdomenTwinspiral (max. 100 cm)		64 x 0.60	Manual kV	145	Auto (Non-Contrast)				(32 cm) 9.62	317	Standard	0.80		0.33		2.0																
			Abdomen 1.50 Qr40 I3																				1.50	1.50		None	None		Qr40	On	3						
RTP Pelvis [factory] (Adult)	>Indications: radiotherapy planning of the pelvis region, w/w/o CM >Scan modes: RoutineSpiralAdultPelvis	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard					2.0																
			Topogram 0.60 Tr20 cor																																		
		Pelvis		Pelvis	RoutineSpiralAdultPelvis		64 x 0.60	Manual kV	170	Auto (Non-Contrast)	120	131		(32 cm) 11.7	258	Standard	0.80		1.00		2.0																
			Pelvis 1.50 Qr40 I3																				1.50	1.50		None	None		Qr40	On	3						
RTP Extremity [factory] (Adult)	>Indications: radiotherapy planning of the extremities region, w/w/o CM >Scan modes: RoutineSpiralAdultExtremities	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard					2.0																
			Topogram 0.60 Tr20 cor																																		
		Extremity		Extremities	RoutineSpiralAdultExtremities		64 x 0.60	Manual kV	75	Auto (Bone/Calcium)	120	58		(32 cm) 5.20	62.4	Standard	0.80		1.00		2.0																
			Extremity 1.50 Qr40 I3																				1.50	1.50		None	None		Qr40	On	3						

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
RTP Direct Density commissioning [factory] (Adult)	>Indications: commissioning of Direct Density for radiotherapy planning scan protocols >Scan modes: RoutineSpiralAdultAbdomen	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		120 kV		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Off			120	168		(32 cm) 15.1	461	Standard	0.35		0.50		2.0																
			120 kV 1.50 Qr40 I3																					1.50	1.00			None	None		Qr40	On	3				
			120 kV 1.50 Sd40 I3																					1.50	1.00			None	None		Sd40	On	3				
			120 kV 1.50 Sm40 I3																					1.50	1.00			None	None		Sm40	On	3				
		70 kV		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Off			70	868		(32 cm) 15.0	460	Standard	0.35		1.00		2.0																
			70 kV 1.50 Qr40 I3																					1.50	1.00			None	None		Qr40	On	3				
			70 kV 1.50 Sd40 I3																					1.50	1.00			None	None		Sd40	On	3				
			70 kV 1.50 Sm40 I3																					1.50	1.00			None	None		Sm40	On	3				
		80 kV		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Off			80	556		(32 cm) 15.0	460	Standard	0.35		0.50		2.0																
			80 kV 1.50 Qr40 I3																					1.50	1.00			None	None		Qr40	On	3				
			80 kV 1.50 Sd40 I3																					1.50	1.00			None	None		Sd40	On	3				
			80 kV 1.50 Sm40 I3																					1.50	1.00			None	None		Sm40	On	3				
		100 kV		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Off			100	276		(32 cm) 15.0	460	Standard	0.35		0.50		2.0																
			100 kV 1.50 Qr40 I3																					1.50	1.00			None	None		Qr40	On	3				
			100 kV 1.50 Sd40 I3																					1.50	1.00			None	None		Sd40	On	3				
			100 kV 1.50 Sm40 I3																					1.50	1.00			None	None		Sm40	On	3				
RTP Twinbeam DE commissioning [factory] (Adult)	>Indications: commissioning of Dual Energy for radiotherapy planning >Scan modes: RoutineSpiralAdultAbdomenTwinbeam	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		AuSn120		Abdomen	RoutineSpiralAdultAbdomenTwinbeam		64 x 0.60	Off			AuSn 120	541		(32 cm) 15.0	514	Standard	0.35		0.50		2.0																
			AuSn120 1.50 Qr40 I3																					1.50	1.00			L&H	None		Qr40	On	3				
			AuSn120 1.50 Qr40 I3 MP 050																					1.50	1.00			None			Qr40	On	3				
			AuSn120 1.50 Qr40 I3 MP 070																					1.50	1.00			None			Qr40	On	3				
			AuSn120 1.50 Qr40 I3 MP 100																				1.50	1.00			None			Qr40	On	3					
			AuSn120 1.50 Qr40 I3 ax																				1.50	1.00			DE Parallel Ranges		MPR	Qr40	On	3					
		AuSn140		Abdomen	RoutineSpiralAdultAbdomenTwinbeam		64 x 0.60	Off			AuSn 140	329		(32 cm) 15.0	514	Standard	0.35		0.50		2.0																
			AuSn140 1.50 Qr40 I3																					1.50	1.00			L&H	None		Qr40	On	3				
			AuSn140 1.50 Qr40 I3 MP 050																					1.50	1.00			None			Qr40	On	3				
			AuSn140 1.50 Qr40 I3 MP 070																					1.50	1.00			None			Qr40	On	3				
			AuSn140 1.50 Qr40 I3 MP 100																					1.50	1.00			None			Qr40	On	3				
			AuSn140 1.50 Qr40 I3 ax																					1.50	1.00			DE Parallel Ranges		MPR	Qr40	On	3				
		RTP Twinspiral DE commissioning [factory] (Adult)	>Indications: commissioning of Dual Energy for radiotherapy planning >Scan modes: RoutineSpiralAdultAbdomenTwinspiral (max. 100 cm)	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					2.0														
					Topogram 0.60 Tr20 cor																											Tr20					
				80/Sn140		Abdomen	RoutineSpiralAdultAbdomenTwinspiral (max. 100 cm)		64 x 0.60	Off						(32 cm) 17.4	546	Standard	0.35		0.50		2.0														
					80/Sn140 1.50 Qr40 I3																					1.50	1.00			L&H	None		Qr40	On	3		
	80/Sn140 1.50 Qr40 I3 MP 050																							1.50	1.00			None	None		Qr40	On	3				
	80/Sn140 1.50 Qr40 I3 MP 070																							1.50	1.00			None	None		Qr40	On	3				
	80/Sn140 1.50 Qr40 I3 MP 100																						1.50	1.00			None	None		Qr40	On	3					
	80/Sn140 1.50 Qr40 I3 ax																						1.50	1.00			DE Parallel Ranges	None	MPR	Qr40	On	3					
100/Sn140				Abdomen	RoutineSpiralAdultAbdomenTwinspiral (max. 100 cm)		64 x 0.60	Off							(32 cm) 17.8	556	Standard	0.35		0.50		2.0															
	100/Sn140 1.50 Qr40 I3																							1.50	1.00			None	None		Qr40	On	3				
	100/Sn140 1.50 Qr40 I3 MP 050																							1.50	1.00			None	None		Qr40	On	3				
	100/Sn140 1.50 Qr40 I3 MP 070																							1.50	1.00			None	None		Qr40	On	3				
	100/Sn140 1.50 Qr40 I3 MP 100																					1.50	1.00			None	None		Qr40	On	3						
	100/Sn140 1.50 Qr40 I3 ax																					1.50	1.00														

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit					
CTDI measurement Head [factory] (Adult)	>Indications: special head protocol for ACR accreditation only; don't use for clinical cases! >Scan modes: RoutineSequenceAdultHead	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (16 cm) 0.02	0.56	Standard						2.0																		
			Topogram 0.60 Tr20 cor																												Tr20									
		Head		Brain	RoutineSequenceAdultHead		64 x 0.60	Off			120	100		(16 cm) 20.9	216	Standard		34.5	1.00	Full 1.00	2.0																			
			Head 5.00 Hr40																					5.00	5.00			None	None		Hr40	Off	3							
CTDI measurement Body [factory] (Adult)	>Indications: special body protocol for ACR accreditation only; don't use for clinical cases! >Scan modes: RoutineSequenceAdultAbdomen	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.30	Standard						2.0																		
			Topogram 0.60 Tr20 cor																												Tr20									
		Body		Abdomen	RoutineSequenceAdultAbdomen		64 x 0.60	Off			120	100		(32 cm) 9.97	103	Standard		34.5	1.00	Full 1.00	2.0																			
			Body 5.00 Br40																					5.00	5.00			None	None		Br40	Off	3							
Thorax Intervention on 3D [factory] (Adult)	>Indications: biopsy, abscess in the thorax w/w/o CM >Scan modes: InterventionSpiralAdultThorax, InterventionSequenceAdultThorax, InterventionFluoroAdultThorax2d	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						6.0																		
			Topogram 0.60 Tr20 cor																												Tr20									
		i-Spiral Thorax		Thorax	InterventionSpiralAdultThorax		64 x 0.60	Full	80	Auto (Non-Contrast)	120	62		(32 cm) 4.40	28.0	Narrow	0.80		0.50		6.0																			
		i-Sequence Thorax		Thorax	InterventionSequenceAdultThorax		64 x 0.60	Off			120	21		(32 cm) 1.49	5.73	Narrow		0.0	0.50	Full 0.50				5.00	4.80			None	None		BI60	Off	3							
			i-Sequence Thorax 5.00 Br40																					5.00	4.80			None	None		Br40	Off	3							
		i-Fluoro Thorax		Thorax	InterventionFluoroAdultThorax2d		3 x 5.00	Off			120	21				Narrow			0.50																					
			i-Fluoro Thorax 5.00 Br40																					5.00	5.00				None		Br40									
Thorax Intervention [factory] (Adult)	>Indications: biopsy, abscess in the thorax w/w/o CM >Scan modes: RoutineSpiralAdultThorax, InterventionSequenceAdultThorax2d, InterventionFluoroAdultThorax2d, RoutineSpiralAdultThorax >Decision Tree: Thorax FoV [factory]	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						6.0																		
			Topogram 0.60 Tr20 cor																												Tr20									
		Planning Thorax		Thorax	RoutineSpiralAdultThorax	Thorax FoV [factory]	64 x 0.60	Full	80	Auto (Non-Contrast)	120	62		(32 cm) 5.55	178	Standard	0.80		0.50		6.0																			
			Planning Thorax 5.00 BI60 I3																					5.00	5.00			Thorax FoV [factory].fast acquisition	None	None		BI60	On	3						
			Planning Thorax 5.00 Br40 I3																						5.00	5.00			Thorax FoV [factory].fast acquisition	None	None		Br40	On	3					
			Planning Thorax 5.00 BI60 I3																						5.00	5.00			Thorax FoV [factory].standard acquisition	None	None		BI60	On	3					
			Planning Thorax 5.00 Br40 I3																						5.00	5.00			Thorax FoV [factory].standard acquisition	None	None		Br40	On	3					
		i-Sequence Thorax		Thorax	InterventionSequenceAdultThorax2d		3 x 5.00	Off			120	21		(32 cm) 1.44	2.16	Narrow		0.0	0.50	Full 0.50																				
			i-Sequence Thorax 5.00 Br40																						5.00					None		Br40								
		i-Fluoro Thorax		Thorax	InterventionFluoroAdultThorax2d		3 x 5.00	Off			120	21				Narrow			0.50																					
			i-Fluoro Thorax 5.00 Br40																						5.00	5.00				None		Br40								
		Control Thorax		Thorax	RoutineSpiralAdultThorax	Thorax FoV [factory]	64 x 0.60	Full	80	Auto (Non-Contrast)	120	62		(32 cm) 5.55	178	Standard	0.80		0.50		6.0																			
			Control Thorax 5.00 BI60 I3																							5.00	5.00			Thorax FoV [factory].fast acquisition	None	None		BI60	On	3				
	Control Thorax 5.00 Br40 I3																							5.00	5.00			Thorax FoV [factory].fast acquisition	None	None		Br40	On	3						
	Control Thorax 5.00 BI60 I3																							5.00	5.00			Thorax FoV [factory].standard acquisition	None	None		BI60	On	3						
	Control Thorax 5.00 Br40 I3																							5.00	5.00			Thorax FoV [factory].standard acquisition	None	None		Br40	On	3						
Abdomen Intervention on 3D [factory] (Adult)	>Indications: biopsy, abscess in the abdomen, w/w/o CM >Scan modes: InterventionSpiralAdultAbdomen, InterventionSequenceAdultAbdomen, InterventionFluoroAdultAbdomen2d	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard						6.0																		
			Topogram 0.60 Tr20 cor																												Tr20									
		i-Spiral Abdomen		Abdomen	InterventionSpiralAdultAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	120	112		(32 cm) 7.95	50.5	Narrow	0.80		0.50		6.0																			
			i-Spiral Abdomen 5.00 Br40																						5.00	4.80			None	None		Br40	Off	3						
		i-Sequence Abdomen		Abdomen	InterventionSequenceAdultAbdomen		64 x 0.60	Off			120	37		(32 cm) 2.63	10.1	Narrow		0.0	0.50	Full 0.50																				
			i-Sequence Abdomen 5.00 Br40																						5.00	4.80			None	None		Br40	Off	3						
		i-Fluoro Abdomen		Abdomen	InterventionFluoroAdultAbdomen2d		3 x 5.00	Off			120	37				Narrow			0.50																					
	i-Fluoro Abdomen 5.00 Br40																						5.00	5.00				None		Br40										

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit			
Abdomen Intervention [factory] (Adult)	>Indications: biopsy, abscess in the abdomen, w/wo CM >Scan modes: RoutineSpiralAdultAbdomen, InterventionSequenceAdultAbdomen2d, InterventionFluoroAdultAbdomen2d	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 100		60	SPR (32 cm) 0.01	0.49	Standard					6.0																	
			Topogram 0.60 Tr20 cor																																			
		Planning Abdomen		Abdomen	RoutineSpiralAdultAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	120	112		(32 cm) 10.0	321	Standard	0.80		0.50		6.0										Tr20							
			Planning Abdomen 5.00 Br40 I3																					5.00	5.00		None	None		Br40	On	3						
		i-Sequence Abdomen		Abdomen	InterventionSequenceAdultAbdomen2d		3 x 5.00	Off			120	37		(32 cm) 2.53	3.80	Narrow		0.0	0.50	Full 0.50																		
			i-Sequence Abdomen 5.00 Br40																						5.00				None		Br40							
		i-Fluoro Abdomen		Abdomen	InterventionFluoroAdultAbdomen2d		3 x 5.00	Off			120	37				Narrow				0.50																		
			i-Fluoro Abdomen 5.00 Br40																					5.00	5.00			None		Br40								
Spine Pain Therapy 3D [factory] (Adult)	>Indications: periradicular therapy, nerve root block, facet joint injection, arthrogram in the spine region, w/wo CM >Scan modes: InterventionSpiralAdultSpineSn, InterventionSequenceAdultSpineSn, InterventionFluoroAdultSpineSn2d	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard					2.0																	
			Topogram 0.60 Tr20 sag																													Tr20						
		i-Spiral Spine		Spine	InterventionSpiralAdultSpineSn		64 x 0.60	Full	70	Auto (Bone/Calcium)	Sn 130	247		(32 cm) 3.96	31.2	Narrow	0.60		0.50		6.0																	
			i-Spiral Spine 3.00 Br60								Sn 130	85		(32 cm) 1.36	5.23	Narrow		0.0	0.50	Full 0.50					3.00	2.80		None	None		Br60	Off	3					
		i-Sequence Spine		Spine	InterventionSequenceAdultSpineSn		64 x 0.60	Off			Sn 130																											
			i-Sequence Spine 3.20 Br60																						3.20	3.00		None	None		Br60	Off	3					
		i-Fluoro Spine		Spine	InterventionFluoroAdultSpineSn2d		3 x 3.00	Off			Sn 130	85				Narrow				0.50																		
			i-Fluoro Spine 3.00 Br60																					3.00	3.00			None		Br60								
Spine Pain Therapy [factory] (Adult)	>Indications: periradicular therapy, nerve root block, facet joint injection, arthrogram in the spine region, w/wo CM >Scan modes: RoutineSpiralAdultSpineSn, InterventionSequenceAdultSpineSn2d, InterventionFluoroAdultSpineSn2d	Topogram			ScanplanningTopoAdultSn		6 x 0.60				Sn 140		60	SPR (32 cm) 0.05	2.64	Standard					2.0																	
			Topogram 0.60 Tr20 sag																													Tr20						
		Planning Spine		Spine	RoutineSpiralAdultSpineSn		64 x 0.60	Full	70	Auto (Bone/Calcium)	Sn 130	247		(32 cm) 4.77	93.7	Standard	0.60		0.50		6.0																	
			Planning Spine 3.00 Br60 I3																						3.00	3.00		None	None		Br60	On	3					
			Planning Spine 3.00 Br40 I3																						3.00	3.00		None	None		Br40	On	3					
		i-Sequence Spine		Spine	InterventionSequenceAdultSpineSn2d		3 x 3.00	Off			Sn 130	85		(32 cm) 1.45	1.31	Narrow		0.0	0.50	Full 0.50																		
			i-Sequence Spine 3.00 Br60																						3.00				None		Br60							
		i-Fluoro Spine		Spine	InterventionFluoroAdultSpineSn2d		3 x 3.00	Off			Sn 130	85				Narrow				0.50																		
			i-Fluoro Spine 3.00 Br60																					3.00	3.00			None		Br60								

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy* cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Brain [factory] (Child)	>Indications: head trauma, hemorrhage, hydrocephalus, abnormalities, inflammation, wo CM >Scan modes: RoutineSpiralChildHeadXcare	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (16 cm) 0.04	1.10	Narrow					2.0																
			Topogram 0.60 Tr20 sag																											Tr20							
		Brain		Brain	RoutineSpiralChildHeadXcare		64 x 0.60	Full	345	Auto (Non-Contrast)	100	394		(16 cm) 46.5	524	Standard	0.55		1.00		2.0																
			Brain 5.00 Hr40 I2 Brain 5.00 Hr60 I2																				5.00	5.00		None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2						
Brain Multi Phase [factory] (Child)	>Indications: tumors, hydrocephalus, abnormalities, inflammation, w and wo CM >Scan modes: RoutineSpiralChildHeadXcare	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (16 cm) 0.04	1.10	Narrow					2.0																
			Topogram 0.60 Tr20 sag																											Tr20							
		Non Contrast Brain		Brain	RoutineSpiralChildHeadXcare		64 x 0.60	Full	345	Auto (Non-Contrast)	100	394		(16 cm) 46.5	524	Standard	0.55		1.00		2.0																
			Non Contrast Brain 5.00 Hr40 I2 Non Contrast Brain 5.00 Hr60 I2																				5.00	5.00		None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2						
		Venous Phase Brain		Brain	RoutineSpiralChildHeadXcare		64 x 0.60	Full	345	Soft Tissue with Contrast	100	312		(16 cm) 36.8	415	Standard	0.55		1.00		100.0																
			Venous Phase Brain 5.00 Hr40 I2 Venous Phase Brain 5.00 Hr60 I2																				5.00	5.00		None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2						
Inner Ear [factory] (Child)	>Indications: inflammation, tumorous processes of pyramids, acoustic neuroma, pathologies of the mastoid process, cerebellopontine angle tumors, cochlea implants, trauma, wo CM >Scan modes: RoutineSpiralChildHeadSn	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (16 cm) 0.04	1.10	Narrow					2.0																
			Topogram 0.60 Tr20 sag																											Tr20							
		Inner Ear		Brain	RoutineSpiralChildHeadSn		64 x 0.60	Full	110	Auto (Bone/Calcium)	Sn 100	1,227		(16 cm) 11.6	97.5	Standard	0.55		1.00		2.0																
			Inner Ear 1.50 Hr60 I3 ax Inner Ear 1.50 Hr40 I3 ax Inner Ear 0.80 Hr60 I3 ax																				1.50	1.50		None	Impression Multi Recons (2)	MPR	Hr60 Hr40 Hr60	On On On	3 3 3						
Sinus [factory] (Child)	>Indications: sinusitis, mucocoele, pneumatization, polyposis, tumor, corrections, wo CM >Scan modes: RoutineSpiralChildHeadSnXcare	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (16 cm) 0.04	1.10	Narrow					2.0																
			Topogram 0.60 Tr20 sag																											Tr20							
		Sinus		Brain	RoutineSpiralChildHeadSnXcare		64 x 0.60	Full	50	Auto (Bone/Calcium)	Sn 100	581		(16 cm) 5.51	62.7	Standard	0.55		1.00		2.0																
			Sinus 3.00 Hr60 I3 ax Sinus 3.00 Hr40 I3 ax Sinus 3.00 Hr60 I3 ax																				3.00	3.00		None	Impression Multi Recons (2)	MPR	Hr60 Hr40 Hr60	On On On	3 3 3						
Orbit [factory] (Child)	>Indications: fracture, tumor, trauma, wo CM >Scan modes: RoutineSpiralChildHeadSnXcare	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (16 cm) 0.04	1.10	Narrow					2.0																
			Topogram 0.60 Tr20 sag Topogram Recon 1																											Tr20							
		Orbit		Brain	RoutineSpiralChildHeadSnXcare		64 x 0.60	Full	110	Auto (Bone/Calcium)	Sn 100	1,227		(16 cm) 11.6	97.5	Standard	0.55		1.00		2.0																
			Orbit 2.00 Hr60 I3 ax Orbit 2.00 Hr40 I3 ax Orbit 2.00 Hr60 I3 ax Orbit Hr40 I3 cor Radial																				2.00	2.00		None	Impression Multi Recons (2)	MPR	Hr60 Hr40 Hr60	On On On	3 3 3						
																							2.00	2.00		None	None	MPR	Hr60	On	3						
																										Radial Ranges	None	VRT	Hr40	On	3						
Neck [factory] (Child)	>Indications: masses, tumor, metastases, lymphoma, lymph nodes in the larynx or pharynx region, w/wo CM >Scan modes: RoutineSpiralChildNeck	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (32 cm) 0.02	0.56	Narrow					7.0																
			Topogram 0.60 Tr20 sag																											Tr20							
		Neck		Neck	RoutineSpiralChildNeck		64 x 0.60	Full	155	Auto (Non-Contrast)	100	182		(32 cm) 7.75	140	Narrow	1.20		1.00		7.0																
			Neck 5.00 Br40 I3 sag Neck 5.00 Br40 I3 cor																				5.00	5.00		None	Orientation Multi Recons (3)	MPR	Br40	On	3						
			Neck 5.00 Br40 I3 ax																																		
Shoulder [factory] (Child)	>Indications: evaluation of joint cavities, trauma, dislocations, orthopedic indications, shoulder joint, wo CM >Scan modes: RoutineSpiralChildShoulder	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.10	Narrow					2.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Shoulder		Shoulder	RoutineSpiralChildShoulder		64 x 0.60	Full	130	Auto (Bone/Calcium)	100	155		(32 cm) 6.60	107	Narrow	1.50		1.00		2.0																
			Shoulder 5.00 Br60 I3 ax																				5.00	5.00		None	None	MPR	Br60	On	3						
			Shoulder 3.00 Br60 I3 sag Shoulder 3.00 Br60 I3 cor Shoulder 3.00 Br60 I3 ax R																				3.00	3.00		Parallel Ranges	Orientation Multi Recons (3)	MPR	Br60	On	3						
			Shoulder 3.00 Br40 I3 sag Shoulder 3.00 Br40 I3 cor Shoulder 3.00 Br40 I3 ax R																				3.00	3.00		Parallel Ranges	Orientation Multi Recons (3)	MPR	Br40	On	3						
			Shoulder 3.00 Br60 I3 sag Shoulder 3.00 Br60 I3 cor Shoulder 3.00 Br60 I3 ax L																				3.00	3.00		Parallel Ranges	Orientation Multi Recons (3)	MPR	Br60	On	3						
			Shoulder 3.00 Br40 I3 sag Shoulder 3.00 Br40 I3 cor Shoulder 3.00 Br40 I3 ax L																				3.00	3.00		Parallel Ranges	Orientation Multi Recons (3)	MPR	Br40	On	3						

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Thorax [factory] (Child)	>Indications: masses, tumor, metastases, lymphoma, lymph nodes, tuberculosis, empyema, infection, early visualization of pulmonary nodules, screening of the lung, w/wo CM >Scan modes: RoutineSpiralChildThorax	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Thorax		Thorax	RoutineSpiralChildThorax		64 x 0.60	Full	80	Auto (Non-Contrast)	100	102		(32 cm) 4.34	144	Narrow	1.50		0.33		6.0																
			Thorax 5.00 Br40 I3 cor																				5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Thorax 5.00 Br40 I3 ax																				5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3						
			Thorax 5.00 BI60 I3 cor																																		
			Thorax 5.00 BI60 I3 ax																																		
			Thorax 10.00 BI60 I3 cor																				10.00	10.00		None	None	MIP Thin	BI60	On	3						
	Thorax 1.00 BI60 I3																					1.00	0.70		Lung CAD	None		BI60	On	3							
Thorax and Abdomen [factory] (Child)	>Indications: masses, tumor, lymphoma, lymph nodes of the lung, liver, pancreas, gall bladder, spleen, kidneys, w CM, 2 phases preconfigured >Scan modes: RoutineSpiralChildAbdomenFlexdose	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Thorax and Abdomen		Abdomen	RoutineSpiralChildAbdomenFlexdose		64 x 0.60	Full	145	Auto (Non-Contrast)	100	151		(32 cm) 6.41	346	Narrow	1.20		0.33		6.0																
			Thorax and Abdomen 5.00 Br40 I3 cor																				5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Thorax and Abdomen 5.00 Br40 I3 ax																																		
			Thorax and Abdomen 5.00 BI60 I3 sag Lung																				5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3						
			Thorax and Abdomen 5.00 BI60 I3 cor Lung																																		
			Thorax and Abdomen 5.00 BI60 I3 ax Lung																																		
	Thorax and Abdomen 1.00 BI60 I3																					1.00	0.70		Lung CAD	None		BI60	On	3							
Abdomen [factory] (Child)	>Indications: masses, tumor, lymphoma, lymph nodes of the liver, pancreas, gall bladder, spleen, kidneys, intestine, w/wo CM >Scan modes: RoutineSpiralChildAbdomen	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Abdomen		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	100	188		(32 cm) 8.01	265	Narrow	1.20		0.50		6.0																
			Abdomen 5.00 Br40 I3 cor																				5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3						
		Abdomen 5.00 Br40 I3 ax																																			
Abdomen Multi Phase [factory] (Child)	>Indications: masses, tumor, lymphoma, lymph nodes of the liver, pancreas, gall bladder, spleen, kidneys, intestine, w and wo CM, 3 phases preconfigured >Scan modes: RoutineSpiralChildAbdomen, PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAbdomen	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Non Contrast Liver		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	100	188		(32 cm) 8.01	185	Narrow	1.20		0.50		6.0																
			Non Contrast Liver 5.00 Br40 I3 cor																				5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Non Contrast Liver 5.00 Br40 I3 ax																																		
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	50		(32 cm) 1.01	0.50	Narrow		0.0	0.50	Full 0.50	2.0																
			Premonitoring 5.00 Br36																				5.00					None		Br36							
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	50		(32 cm) 30.2	15.1	Narrow		0.0	0.50	Full 0.50	10.0																
			Monitoring 5.00 Br36																					5.00					None		Br36						
		Arterial Phase Abdomen		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Soft Tissue with Contrast)	100	149		(32 cm) 6.35	210	Narrow	1.20		0.50																		
			Arterial Phase Abdomen 5.00 Br40 I3 cor																					5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3					
			Arterial Phase Abdomen 5.00 Br40 I3 ax																																		
Venous Phase Abdomen		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Soft Tissue with Contrast)	100	149		(32 cm) 6.35	146	Narrow	1.20		0.50		64.5																		
	Venous Phase Abdomen 5.00 Br40 I3 cor																					5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3							
	Venous Phase Abdomen 5.00 Br40 I3 ax																																				
Whole Body [factory] (Child)	>Indications: masses, tumor, lymphoma, lymph nodes of the lung, liver, pancreas, gall bladder, spleen, kidneys, w/wo CM >Scan modes: RoutineSpiralChildAbdomen	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Whole Body		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	100	188		(32 cm) 8.01	425	Narrow	1.20		0.50		6.0																
			Whole Body 5.00 Br40 I3 cor																				5.00	5.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3						
			Whole Body 5.00 Br40 I3 ax																																		
			Whole Body 5.00 BI60 I3 sag Lung																				5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	BI60	On	3						
			Whole Body 5.00 BI60 I3 cor Lung																																		
			Whole Body 5.00 BI60 I3 ax Lung																																		
			Whole Body 1.00 BI60 I3																					1.00	0.70		Lung CAD	None		BI60	On	3					
			Whole Body 5.00 Br60 I3 sag																					5.00	5.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
	Whole Body 5.00 Br60 I3 cor																																				
	Whole Body 5.00 Br60 I3 ax																																				
	Whole Body 5.00 Br40 I3 ax																					5.00	5.00		None	None	MPR	Br40	On	3							

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit	
Elbow [factory] (Child)	>Indications: trauma, disorders of the elbow joint, wo CM >Scan modes: RoutineSpiralChildExtremitiesSn	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.10	Narrow					2.0															
			Topogram 0.60 Tr20 cor																											Tr20						
		Elbow		Extremities	RoutineSpiralChildExtremitiesSn		64 x 0.60	Full	75	Auto (Bone/Calcium)	Sn 100	844		(32 cm) 3.47	51.4	Narrow	0.80		1.00		2.0															
			Elbow 2.00 Br60 I3 sag R																				2.00	2.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
			Elbow 2.00 Br60 I3 cor R																																	
			Elbow 2.00 Br60 I3 ax R																																	
			Elbow 2.00 Br40 I3 ax R																				2.00	2.00		None	None	MPR	Br40	On	3					
			Elbow 2.00 Br60 I3 sag L																				2.00	2.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
	Elbow 2.00 Br60 I3 cor L																																			
	Elbow 2.00 Br60 I3 ax L																																			
	Elbow 2.00 Br40 I3 ax L																						2.00	2.00		None	None	MPR	Br40	On	3					
Foot [factory] (Child)	>Indications: trauma, disorders of the foot bones, wo CM >Scan modes: RoutineSpiralChildExtremitiesSn	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.10	Narrow					2.0															
			Topogram 0.60 Tr20 cor																											Tr20						
		Foot		Extremities	RoutineSpiralChildExtremitiesSn		64 x 0.60	Full	75	Auto (Bone/Calcium)	Sn 100	844		(32 cm) 3.47	51.4	Narrow	0.80		1.00		2.0															
			Foot 3.00 Br60 I3 sag R																				3.00	3.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
			Foot 3.00 Br60 I3 cor R																																	
			Foot 3.00 Br60 I3 ax R																																	
			Foot 3.00 Br40 I3 ax R																				3.00	3.00		None	None	MPR	Br40	On	3					
			Foot 3.00 Br60 I3 sag L																				3.00	3.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
	Foot 3.00 Br60 I3 cor L																																			
	Foot 3.00 Br60 I3 ax L																																			
	Foot 3.00 Br40 I3 ax L																						3.00	3.00		None	None	MPR	Br40	On	3					
Leg or Ankle [factory] (Child)	>Indications: trauma, disorders of the femur, fibula, tibia bones, ankle joint, wo CM >Scan modes: RoutineSpiralChildExtremitiesSn	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (32 cm) 0.02	0.56	Narrow					2.0															
			Topogram 0.60 Tr20 sag																											Tr20						
		Leg or Ankle		Extremities	RoutineSpiralChildExtremitiesSn		64 x 0.60	Full	75	Auto (Bone/Calcium)	Sn 100	844		(32 cm) 3.47	51.4	Narrow	0.80		1.00		2.0															
			Leg or Ankle 2.00 Br60 I3 sag R																				2.00	2.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
			Leg or Ankle 2.00 Br60 I3 cor R																																	
			Leg or Ankle 2.00 Br60 I3 ax R																																	
			Leg or Ankle 2.00 Br60 I3 ax R																				2.00	2.00		None	None	MPR	Br40	On	3					
			Leg or Ankle 2.00 Br60 I3 sag L																				2.00	2.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
	Leg or Ankle 2.00 Br60 I3 cor L																																			
	Leg or Ankle 2.00 Br60 I3 ax L																																			
	Leg or Ankle 2.00 Br40 I3 ax L																						3.00	3.00		None	None	MPR	Br40	On	3					
Knee [factory] (Child)	>Indications: trauma, disorders of the knee joint, wo CM >Scan modes: RoutineSpiralChildExtremitiesSn	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.10	Narrow					2.0															
			Topogram 0.60 Tr20 cor																											Tr20						
		Knee		Extremities	RoutineSpiralChildExtremitiesSn		64 x 0.60	Full	75	Auto (Bone/Calcium)	Sn 100	844		(32 cm) 3.47	61.9	Narrow	0.80		1.00		2.0															
			Knee 3.00 Br60 I3 sag R																				3.00	3.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3					
			Knee 3.00 Br60 I3 cor R																																	
			Knee 3.00 Br60 I3 ax R																																	
			Knee 3.00 Br40 I3 ax R																																	
			Knee 3.00 Br60 I3 cor R																				3.00	3.00		None	None	MPR	Br40	On	3					
	Knee 3.00 Br60 I3 sag L																				3.00	3.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3							
	Knee 3.00 Br60 I3 cor L																																			
	Knee 3.00 Br60 I3 ax L																																			
	Knee 3.00 Br40 I3 ax L																																			
	Knee 3.00 Br60 I3 cor L																						3.00	3.00		None	None	MPR	Br40	On	3					
	Knee 3.00 Br60 I3 cor L																						3.00	3.00		Parallel Ranges	None	None	MPR	Br60	On	3				
Brain CTA [factory] (Child)	>Indications: stenosis, aneurysm, occlusion, plaques, abnormalities of the vertebral arteries, COW, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAngioHead	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (32 cm) 0.02	0.56	Narrow					2.0															
			Topogram 0.60 Tr20 sag																											Tr20						
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 1.03	0.51	Narrow		0.0	0.33	Full 0.33	2.0															
			Premonitoring 5.00 Br36																									None		Br36						
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 30.8	15.4	Narrow		0.0	0.33	Full 0.33	10.0			5.00												
			Monitoring 5.00 Br36																																	
		CTA Brain		Head/Vascular Head	RoutineSpiralChildAngioHead		64 x 0.60	Full	105	Auto (Vascular)	80	94		(16 cm) 5.84	77.2	Standard	1.50		0.33																	
			CTA Brain 3.00 Hv36 I3 cor																																	
	CTA Brain 3.00 Hv36 I3 sag																																			
	CTA Brain 3.00 Hv36 I3 ax																																			

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Neck CTA [factory] (Child)	>Indications: stenosis, occlusion, plaques, abnormalities of the carotid arteries, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAngioNeck	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.10	Narrow					7.0																
		Topogram 0.60 Tr20 cor																												Tr20							
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 1.03	0.51	Narrow		0.0	0.33	Full 0.33	2.0																
		Premonitoring 5.00 Br36																					5.00					None		Br36							
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 30.8	15.4	Narrow		0.0	0.33	Full 0.33	10.0			5.00					None		Br36						
		Monitoring 5.00 Br36																						5.00					None		Br36						
		CTA Neck		Neck	RoutineSpiralChildAngioNeck		64 x 0.60	Full	145	Auto (Vascular)	80	140		(32 cm) 2.90	67.2	Narrow	1.50		0.33					3.00	3.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
		CTA Neck 3.00 Bv36 I3 sag																						3.00	3.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
		CTA Neck 3.00 Bv36 I3 cor																																			
		CTA Neck 3.00 Bv36 I3 ax																																			
		CTA Neck 2.00 Bv36 I3 cor																						2.00	2.00		Parallel Ranges	None	MIP Thin	Bv36	On	3					
CTA Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00		Parallel Ranges	None	MPR	Br40	On	3							
CTA Neck 3.00 Br40 I3 Parallel Vascular																						3.00	3.00		Parallel Vascular Ranges	None	MPR	Br40	On	3							
Thorax CTA [factory] (Child)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the thoracic aorta, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAngioBody	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
		Topogram 0.60 Tr20 cor																												Tr20							
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 1.03	0.51	Narrow		0.0	0.33	Full 0.33	2.0																
		Premonitoring 5.00 Br36																					5.00					None		Br36							
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 30.8	15.4	Narrow		0.0	0.33	Full 0.33	10.0																
		Monitoring 5.00 Br36																						5.00					None		Br36						
		CTA Thorax		Vascular Body	RoutineSpiralChildAngioBody		64 x 0.60	Full	70	Auto (Vascular)	80	96		(32 cm) 1.99	46.1	Narrow	1.50		0.33					5.00					None		Br36						
		CTA Thorax 5.00 Bv36 I3 sag																						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
		CTA Thorax 5.00 Bv36 I3 cor																																			
		CTA Thorax 5.00 Bv36 I3 ax																																			
		CTA Thorax 5.00 BI60 I3 ax																						5.00	5.00		None	Impression Multi Recons (2)	MIP Thin	BI60	On	3					
CTA Thorax 5.00 Br40 I3 ax																												Br40	On	3							
Pulmonary CTA [factory] (Child)	>Indications: pulmonary embolism, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAngioBody	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					4.0																
		Topogram 0.60 Tr20 cor																												Tr20							
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 1.03	0.51	Narrow		0.0	0.33	Full 0.33	2.0																
		Premonitoring 5.00 Br36																					5.00					None		Br36							
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 30.8	15.4	Narrow		0.0	0.33	Full 0.33	10.0																
		Monitoring 5.00 Br36																						5.00					None		Br36						
		CTA Pulmonary		Vascular Body	RoutineSpiralChildAngioBody		64 x 0.60	Full	70	Auto (Vascular)	80	96		(32 cm) 1.99	46.1	Narrow	1.50		0.33										None		Br36						
		CTA Pulmonary 3.00 Bv36 I3 sag																						3.00	3.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3					
		CTA Pulmonary 3.00 Bv36 I3 cor																																			
		CTA Pulmonary 3.00 Bv36 I3 ax																																			
		CTA Pulmonary 3.00 BI60 I3 ax																						3.00	3.00		None	Impression Multi Recons (2)	MIP Thin	BI60	On	3					
CTA Pulmonary 3.00 Br40 I3 ax																												Br40	On	3							
Cardiac CTA [factory] (Child)	>Indications: vascular anomalies, complex congenital heart disease, kawasaki, septal defect, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, CardiacSpiralChildAngio	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					10.0																
		Topogram 0.60 Tr20 cor																												Tr20							
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 1.03	0.51	Narrow		0.0	0.33	Full 0.33	2.0																
		Premonitoring 5.00 Br36																						5.00					None		Br36						
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 30.8	15.4	Narrow		0.0	0.33	Full 0.33	10.0																
		Monitoring 5.00 Br36																						5.00					None		Br36						
		CTA Cardiac		Cardio	CardiacSpiralChildAngio		64 x 0.60	Full	64	Auto (Vascular)	80	75		(32 cm) 9.87	138	Narrow	0.31		0.33																		
		CTA Cardiac 3.00 Bv36 I3 BestSyst long axis																						3.00	3.00		Parallel Ranges	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3	BestPhase (systolic)		75	%	
		CTA Cardiac 3.00 Bv36 I3 BestSyst short axis																																			
		CTA Cardiac 3.00 Bv36 I3 BestSyst four chamber																																			
		CTA Cardiac 0.80 Bv40 I3 BestSyst																						0.80	0.50		None	None			Bv40	On	3	BestPhase		30	%

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy* cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit		
Abdomen CTA [factory] (Child)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the abdominal aorta, renal arteries, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAngioBody	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																											Tr20							
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 1.03	0.51	Narrow		0.0	0.33	Full 0.33	2.0																
			Premonitoring 5.00 Br36																					5.00				None			Br36						
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 30.8	15.4	Narrow		0.0	0.33	Full 0.33	10.0																
			Monitoring 5.00 Br36																						5.00				None			Br36					
		CTA Abdomen		Vascular Body	RoutineSpiralChildAngioBody		64 x 0.60	Full	115	Auto (Vascular)	80	152		(32 cm) 3.14	104	Narrow	1.50		0.33						5.00				None			Br36					
			CTA Abdomen 5.00 Bv36 I3 sag CTA Abdomen 5.00 Bv36 I3 cor CTA Abdomen 5.00 Bv36																						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
Whole Body CTA [factory] (Child)	>Indications: vascular anomalies, aneurysms, dissection, ruptures of the whole aorta, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAngioBody	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																												Tr20						
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 1.03	0.51	Narrow		0.0	0.33	Full 0.33	2.0																
			Premonitoring 5.00 Br36																					5.00				None			Br36						
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	51		(32 cm) 30.8	15.4	Narrow		0.0	0.33	Full 0.33	10.0																
			Monitoring 5.00 Br36																						5.00				None			Br36					
		CTA Whole Body		Vascular Body	RoutineSpiralChildAngioBody		64 x 0.60	Full	115	Auto (Vascular)	80	152		(32 cm) 3.14	167	Narrow	1.50		0.33						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
			CTA Whole Body 5.00 Bv36 I3 sag CTA Whole Body 5.00 Bv36 I3 cor CTA Whole Body 5.00 Bv36 I3 ax																						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
Extremity CTA [factory] (Child)	>Indications: vascular anomalies, aneurysms of the femoral arteries, w CM >Scan modes: PremonitoringSequenceChild2d, MonitoringSequenceChild2d, RoutineSpiralChildAngioExtremities	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0																
			Topogram 0.60 Tr20 cor																												Tr20						
		Premonitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	50		(32 cm) 1.01	0.50	Narrow		0.0	0.50	Full 0.50	2.0																
			Premonitoring 5.00 Br36																					5.00				None			Br36						
		Monitoring		Vascular Body			1 x 5.00	Manual kV	25	Soft Tissue with Contrast	80	50		(32 cm) 30.2	15.1	Narrow		0.0	0.50	Full 0.50	10.0																
			Monitoring 5.00 Br36																						5.00				None			Br36					
		CTA Extremity		Runoff	RoutineSpiralChildAngioExtremities		64 x 0.60	Full	155	Auto (Vascular)	80	200		(32 cm) 4.14	211	Narrow	0.45		0.50						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
			CTA Extremity 5.00 Bv36 I3 sag CTA Extremity 5.00 Bv36 I3 cor CTA Extremity 5.00 Bv36 I3 ax																						5.00	5.00		None	Orientation Multi Recons (3)	MIP Thin	Bv36	On	3				
Head Trauma [factory] (Child)	>Indications: emergency head trauma, hemorrhage, wo CM >Scan modes: RoutineSpiralChildHead	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (16 cm) 0.01	0.21	Narrow					2.0																
			Topogram 0.60 Tr20 cor																												Tr20						
		Head		Brain	RoutineSpiralChildHead		64 x 0.60	Full	493	Auto (Non-Contrast)	110	449		(16 cm) 67.8	901	Standard	0.55		0.50		2.0																
			Head 5.00 Hr40 ax Head 5.00 Hr40 sag Head 5.00 Hr40 cor																					5.00	5.00		None	Orientation Multi Recons (3)	MPR	Hr40	Off	2					
			Head 0.80 Hr60 ax Head 0.80 Hr60 sag Head 0.80 Hr60 cor																					0.80	0.60		None	Orientation Multi Recons (3)	MPR	Hr60	Off	2					
Body Trauma [factory] (Child)	>Indications: fast screening for emergency whole body trauma, hemorrhage, wo CM >Scan modes: RoutineSpiralChildAbdomen	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					6.0																
			Topogram 0.60 Tr20 cor																												Tr20						
		Body		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	207	Auto (Non-Contrast)	100	269		(32 cm) 11.5	591	Narrow	0.65		0.33		6.0																
			Body 1.50 Br40 ax Body 1.50 Br60 ax Body 1.50 BI60 ax																					1.50	1.00		None	Impression Multi Recons (3)	MPR	Br40 Br60 BI60	Off Off Off	3 3 3					
RTP Head [factory] (Child)	>Indications: radiotherapy planning of the head region, w/wo CM >Scan modes: RoutineSpiralChildHead	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 140		25	SPR (16 cm) 0.04	1.79	Narrow					2.0																
			Topogram 0.60 Tr20 sag																												Tr20						
		Head		Brain	RoutineSpiralChildHead		64 x 0.60	Manual kV	345	Auto (Non-Contrast)	120	265		(16 cm) 49.7	561	Standard	0.55		1.00		2.0																
RTP Head and Neck [factory] (Child)	>Indications: radiotherapy planning of the head, neck and shoulder region, w/wo CM >Scan modes: RoutineSpiralChildNeck	Head 1.50 Qr40 I2																					1.50	1.50		None	None			Qr40	On	2					
		Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0											Tr2					

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy*cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit
RTP Thorax [factory] (Child)	>Indications: radiotherapy planning of the thorax region, w/w/o CM >Scan modes: RoutineSpiralChildThorax	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		Thorax		Thorax	RoutineSpiralChildThorax		64 x 0.60	Manual kV	80	Auto (Non-Contrast)	120	62		(32 cm) 4.40	96.8	Narrow	0.80		0.33		2.0														
RTP Respiratory Motion Management [factory] (Child)	>Indications: motion management for radiotherapy planning for moving tumors, w/w/o CM >Scan modes: RespiratorySpiralChild		Thorax 1.50 Qr40 I3																				1.50	1.50		None	None		Qr40	On	3				
		Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		Respiratory Motion		Respiratory	RespiratorySpiralChild		64 x 0.60	Manual kV	26	Auto (Non-Contrast)	120	20		(32 cm) 30.9	672	Narrow	0.06		0.33		2.0														
RTP Abdomen [factory] (Child)	>Indications: radiotherapy planning of the abdomen region, w/w/o CM >Scan modes: RoutineSpiralChildAbdomen	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		Abdomen		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Manual kV	145	Auto (Non-Contrast)	120	112		(32 cm) 7.95	254	Narrow	0.80		0.50		2.0														
RTP Extremity [factory] (Child)	>Indications: radiotherapy planning of the extremities region, w/w/o CM >Scan modes: RoutineSpiralChildExtremities		Abdomen 1.50 Qr40 I3																				1.50	1.50		None	None		Qr40	On	3				
		Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.10	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		Extremity		Extremities	RoutineSpiralChildExtremities		64 x 0.60	Manual kV	77	Non-Contrast	120	59		(32 cm) 4.19	50.3	Narrow	0.80		1.00		2.0														
CTDI measurement Head [factory] (Child)	>Indications: special head protocol for ACR accreditation only; don't use for clinical cases! >Scan modes: RoutineSequenceChildHead	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (16 cm) 0.01	0.21	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		Head		Brain	RoutineSequenceChildHead		64 x 0.60	Off			120	50		(16 cm) 10.4	108	Standard		34.5	1.00	Full 1.00	2.0														
CTDI measurement Body [factory] (Child)	>Indications: special body protocol for ACR accreditation only; don't use for clinical cases! >Scan modes: RoutineSequenceChildAbdomen	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.10	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		Body		Abdomen	RoutineSequenceChildAbdomen		64 x 0.60	Off			120	50		(32 cm) 3.95	40.9	Narrow		34.5	1.00	Full 1.00	2.0														
Brain Baby [factory] (Child)	>Indications: head trauma, hemorrhage, hydrocephalus, abnormalities, inflammation, w/w/o CM >Scan modes: RoutineSpiralChildHead	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (16 cm) 0.01	0.21	Narrow					2.0														
			Topogram 0.60 Tr20 sag																										Tr20						
		Brain		Brain	RoutineSpiralChildHead		64 x 0.60	Full	345	Auto (Non-Contrast)	100	225		(16 cm) 26.6	298	Standard	1.50		0.50		2.0														
			Brain 3.00 Hr40 I2 Brain 3.00 Hr60 I2																				3.00	3.00		None	Impression Multi Recons (2)		Hr40 Hr60	On On	2 2				
Body Baby [factory] (Child)	>Indications: masses, tumor, lymphoma of the thorax,liver, pancreas, kidneys, gastrointestinal studies, w/w/o CM >Scan modes: RoutineSpiralChildAbdomen	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		Body		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	100	149		(32 cm) 6.32	147	Narrow	1.50		0.33		2.0														
			Body 3.00 Br40 I3 cor Body 3.00 Br40 I3 ax																				3.00	3.00		None	Orientation Multi Recons (2)	MPR	Br40	On	3				
			Body 3.00 Br60 I3 sag Body 3.00 Br60 I3 cor Body 3.00 Br60 I3 ax																				3.00	3.00		None	Orientation Multi Recons (3)	MPR	Br60	On	3				
Thorax Intervention 3D [factory] (Child)	>Indications: biopsy, abscess in the thorax w/w/o CM >Scan modes: InterventionSpiralChildThorax, InterventionSequenceChildThorax, InterventionFluoroChildThorax2d	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0														
			Topogram 0.60 Tr20 cor																										Tr20						
		i-Spiral Thorax		Thorax	InterventionSpiralChildThorax		64 x 0.60	Full	80	Auto (Non-Contrast)	100	102		(32 cm) 4.34	31.1	Narrow	1.50		0.50		6.0														
			i-Spiral Thorax 3.20 BI60																				3.20	3.00		None	None		BI60	Off	3				
		i-Sequence Thorax		Thorax	InterventionSequenceChildThorax		64 x 0.60	Off			100	17		(32 cm) 0.72	2.78	Narrow		0.0	0.50	Full 0.50			3.20	3.00		None	None		Br40	Off	3				
		i-Fluoro Thorax		Thorax	InterventionFluoroChildThorax2d		3 x 3.00	Off			100	17				Narrow			0.50																
			i-Fluoro Thorax 3.00 Br40																				3.00	3.00			None		Br40						

Protocol Name	Protocol Description	Range Name	Recon Name	Organ Characteristic	Scan Mode	Decision Trees	Acquisition [mm]	CARE Dose4D & CARE kV	CARE kV IQ level	CARE kV optimized for	kV	Eff. mAs	mA	CTDIvol [mGy]	DLP [mGy* cm]	Shaped Filter	Pitch	Feed [mm]	Rot. Time [s]	Scan Time [s]	Absolute Start Delay [s]	Contrast Protocol	Slice Thickness [mm]	Increment [mm]	Recon Strategies	Inline Results	Multi Recons	Image Type	Kernel	Iterative Recon	IR Strength	Recon Type	Recon Phase Step	Recon Phase (s)	Recon Phase Unit	
Thorax Intervention [factory] (Child)	>Indications: biopsy, abscess in the thorax w/w/o CM >Scan modes: RoutineSpiralChildThorax, InterventionSequenceChildThorax2d, InterventionFluoroChildThorax2d, RoutineSpiralChildThorax	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0															
			Topogram 0.60 Tr20 cor																											Tr20						
		Planning Thorax		Thorax	RoutineSpiralChildThorax		64 x 0.60	Full	80	Auto (Non-Contrast)	100	102		(32 cm) 4.34	101	Narrow	1.50		0.50		6.0															
			Planning Thorax 3.00 BI60 I3																					3.00	3.00		None	None		BI60	On	3				
			Planning Thorax 3.00 Br40 I3																					3.00	3.00		None	None		Br40	On	3				
		i-Sequence Thorax		Thorax	InterventionSequenceChildThorax2d		3 x 3.00	Off				100	17		(32 cm) 0.77	0.69	Narrow		0.0	0.50	Full 0.50															
			i-Sequence Thorax 3.00 Br40																					3.00				None		Br40						
		i-Fluoro Thorax		Thorax	InterventionFluoroChildThorax2d		3 x 3.00	Off				100	17				Narrow			0.50																
			i-Fluoro Thorax 3.00 BI60																					3.00	3.00			None		BI60						
		Control Thorax		Thorax	RoutineSpiralChildThorax		64 x 0.60	Full	80	Auto (Non-Contrast)	100	102		(32 cm) 4.34	101	Narrow	1.50		0.50		6.0															
	Control Thorax 3.00 BI60 I3																						3.00	3.00		None	None		BI60	On	3					
	Control Thorax 3.00 Br40 I3																						3.00	3.00		None	None		Br40	On	3					
Abdomen Intervention 3D [factory] (Child)	>Indications: biopsy, abscess in the abdomen, w/w/o CM >Scan modes: InterventionSpiralChildAbdomen, InterventionSequenceChildAbdomen, InterventionFluoroChildAbdomen2d	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0															
			Topogram 0.60 Tr20 cor																											Tr20						
		i-Spiral Abdomen		Abdomen	InterventionSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	100	188		(32 cm) 8.01	56.1	Narrow	1.20		0.50		6.0															
			i-Spiral Abdomen 3.20 Br40																					3.20	3.00		None	None		Br40	Off	3				
		i-Sequence Abdomen		Abdomen	InterventionSequenceChildAbdomen		64 x 0.60	Off				100	31		(32 cm) 1.32	5.07	Narrow		0.0	0.50	Full 0.50															
			i-Sequence Abdomen 3.20 Br40																					3.20	3.00		None	None		Br40	Off	3				
		i-Fluoro Abdomen		Abdomen	InterventionFluoroChildAbdomen2d		3 x 3.00	Off				100	31				Narrow			0.50																
			i-Fluoro Abdomen 3.00 Br40																					3.00	3.00			None		Br40						
Abdomen Intervention [factory] (Child)	>Indications: biopsy, abscess in the abdomen, w/w/o CM >Scan modes: RoutineSpiralChildAbdomen, InterventionSequenceChildAbdomen2d, InterventionFluoroChildAbdomen2d	Topogram			ScanplanningTopoChildSn		6 x 0.60				Sn 100		25	SPR (32 cm) 0.01	0.17	Narrow					2.0															
			Topogram 0.60 Tr20 cor																											Tr20						
		Planning Abdomen		Abdomen	RoutineSpiralChildAbdomen		64 x 0.60	Full	145	Auto (Non-Contrast)	100	188		(32 cm) 8.01	265	Narrow	1.20		0.50		6.0															
			Planning Abdomen 3.00 Br40 I3																					3.00	3.00		None	None		Br40	On	3				
		i-Sequence Abdomen		Abdomen	InterventionSequenceChildAbdomen2d		3 x 3.00	Off				100	31		(32 cm) 1.41	1.27	Narrow		0.0	0.50	Full 0.50															
			i-Sequence Abdomen 3.00 Br40																					3.00				None		Br40						
		i-Fluoro Abdomen		Abdomen	InterventionFluoroChildAbdomen2d		3 x 3.00	Off				100	31				Narrow			0.50																
	i-Fluoro Abdomen 3.00 Br40																					3.00	3.00			None		Br40								

Brain contrast medium necessary [factory]			
Scan Strategy			
	wo CM	w CM	
InfoLevel	None	None	
InfoText			
Scan Mode	RoutineSpiralAdultHead	RoutineSpiralAdultHead	
Injector Coupling	Off	On	
Contrast Protocol		Brain [factory]	
Absolute Start Delay	2	180	
Decision Tree			
	wo CM		
	Contrast medium?		
yes	no		
w CM	wo CM		
Side [factory]			
Scan Strategy			
	recon both	recon right	recon left
InfoLevel	None	None	None
InfoText			
Decision Tree			
	recon both		
	Side?		
both	right	left	
recon both	recon right	recon left	
Metal implant [factory]			
Scan Strategy			
	wo iMAR	w iMAR	
InfoLevel	None	None	
InfoText			
Decision Tree			
	wo iMAR		
	Metal implant?		
yes	no		
w iMAR	wo iMAR		
Sinus sinusitis or soft tissue [factory]			
Scan Strategy			
	low dose	standard dose	
InfoLevel	None	None	
InfoText			
Scan Mode	RoutineSpiralAdultHeadSnXcare	RoutineSpiralAdultHeadXcare	
CARE Dose4D & CARE kV	Full	Full	
CARE kV optimized for	Bone/Calcium	Non-Contrast	
CARE kV IQ level	50	100	
Decision Tree			
	low dose		
	Sinus Sinusitis/soft tissue?		
sinusitis	soft tissue		
low dose	standard dose		
RTP Brain [factory]			
Scan Strategy			
	standard	stereotactic	
InfoLevel	None	None	
InfoText			
Decision Tree			
	standard		
	Stereotaxic?		
yes	no		
stereotactic	standard		
Neck FoV [factory]			
Scan Strategy			
	standard acquisition	fast acquisition	
InfoLevel	Information	None	
InfoText	Please check scan time!		
Scan Mode	RoutineSpiralAdultNeck	RoutineSpiralAdultNeck	
Pitch	0.8	1.2	
Decision Tree			
	standard acquisition		
	FoV?		
yes	no		
fast acquisition	standard acquisition		

Neck contrast medium necessary [factory]		
Scan Strategy		
	wo CM	w CM
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultNeck	RoutineSpiralAdultNeck
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Soft Tissue with Contrast
Injector Coupling	Off	On
Contrast Protocol		Neck [factory]
Absolute Start Delay	2	15
Decision Tree		
wo CM		
Contrast medium?		
yes	no	
w CM	wo CM	

Neck VNC TwinBeam DE possible [factory]		
Scan Strategy		
	TwinBeam DE	Single Energy
InfoLevel	None	Important Information
InfoText		no Dual Energy information!
Scan Mode	RoutineSpiralAdultNeckTwinbeam	RoutineSpiralAdultNeck
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	AuSn 120	70
CARE kV max. kV	AuSn 140	140
Decision Tree		
TwinBeam DE		
Dual Energy Recommendation		
Dual Energy	Single Energy	
TwinBeam DE	Single Energy	

Thorax multi [factory]						
Scan Strategy						
	wo CM wo X-CARE	wo CM w X-CARE	wo CM fast acquisition wo X-CARE	w CM wo X-CARE	w CM w X-CARE	w CM fast acquisition wo X-CARE
InfoLevel	None	None	None	None	None	None
InfoText						
Scan Mode	RoutineSpiralAdultThorax	RoutineSpiralAdultThoraxXcare	RoutineSpiralAdultThorax	RoutineSpiralAdultThorax	RoutineSpiralAdultThoraxXcare	RoutineSpiralAdultThorax
CARE Dose4D & CARE kV	Full	Full	Full	Full	Full	Full
CARE kV optimized for	Non-Contrast	Non-Contrast	Non-Contrast	Soft Tissue with Contrast	Soft Tissue with Contrast	Soft Tissue with Contrast
Injector Coupling	Off	Off	Off	On	On	On
Contrast Protocol				Thorax [factory]	Thorax [factory]	Thorax [factory]
Absolute Start Delay	2	2	2	25	25	25
Pitch	1.2	0.6	1.5	1.2	0.6	1.5
Decision Tree						
wo CM wo X-CARE						
Breathhold >5 s?						
above 5 s					below 5 s	
wo CM wo X-CARE					wo CM fast acquisition wo X-CARE	
Contrast medium?					Contrast medium?	
yes			no			
w CM wo X-CARE			wo CM wo X-CARE			w CM fast acquisition wo X-CARE
Gender			Gender			wo CM fast acquisition wo X-CARE
Male	Female	Other	Male	Female	Other	
w CM wo X-CARE	w CM w X-CARE	w CM w X-CARE	wo CM wo X-CARE	wo CM w X-CARE	wo CM w X-CARE	

Lung screening [factory]				
Scan Strategy				
	low dose wo X-CARE	low dose w X-CARE	high resolution wo X-CARE	high resolution w X-CARE
InfoLevel	None	None	None	None
InfoText				
Scan Mode	RoutineSpiralAdultThoraxSn	RoutineSpiralAdultThoraxSnXcare	RoutineSpiralAdultThorax	RoutineSpiralAdultThoraxSnXcare
CARE Dose4D & CARE kV	Full	Full	Full	Full
CARE kV IQ level	15	15	100	100
Pitch	0.8	0.6	0.8	0.6
Decision Tree				
low dose wo X-CARE				
Lung low dose/high resolution?				
lung cancer screening			high resolution	
low dose wo X-CARE			high resolution wo X-CARE	
Gender			Gender	
Male	Female	Other	Male	Female
low dose wo X-CARE	low dose w X-CARE	low dose w X-CARE	high resolution wo X-CARE	high resolution w X-CARE
				high resolution w X-CARE

Lung Low Dose breathhold and X-CARE [factory]			
Scan Strategy			
	standard acquisition wo X-CARE	standard aquisition w X-CARE	fast acquisition wo X-CARE
InfoLevel	None	None	None
InfoText			
Scan Mode	RoutineSpiralAdultThoraxSn	RoutineSpiralAdultThoraxSnXcare	RoutineSpiralAdultThoraxSn
Pitch	1.2	0.6	1.5
Decision Tree			
standard acquisition wo X-CARE			
Breathhold >5 s?			
above 5 s		below 5 s	
standard acquisition wo X-CARE		fast acquisition wo X-CARE	
Gender			
Male	Female	Other	
standard acquisition wo X-CARE	standard aquisition w X-CARE	standard aquisition w X-CARE	
Thorax breathhold and X-CARE [factory]			
Scan Strategy			
	standard acquisition wo X-CARE	standard aquisition w X-CARE	fast acquisition wo X-CARE
InfoLevel	None	None	None
InfoText			
Scan Mode	RoutineSpiralAdultThorax	RoutineSpiralAdultThoraxXcare	RoutineSpiralAdultThorax
Pitch	1.2	0.6	1.5
Decision Tree			
standard acquisition wo X-CARE			
Breathhold >5 s?			
above 5 s		below 5 s	
standard acquisition wo X-CARE		fast acquisition wo X-CARE	
Gender			
Male	Female	Other	
standard acquisition wo X-CARE	standard aquisition w X-CARE	standard aquisition w X-CARE	
Thorax contrast medium necessary [factory]			
Scan Strategy			
	wo CM	w CM	
InfoLevel	None	None	
InfoText			
Scan Mode	RoutineSpiralAdultThorax	RoutineSpiralAdultThorax	
CARE Dose4D & CARE kV	Full	Full	
CARE kV optimized for	Non-Contrast	Soft Tissue with Contrast	
Injector Coupling	Off	On	
Contrast Protocol		Thorax [factory]	
Absolute Start Delay	2	25	
Decision Tree			
wo CM			
Contrast medium?			
yes	no		
w CM	wo CM		
Thorax FoV [factory]			
Scan Strategy			
	standard acquisition	fast acquisition	
InfoLevel	Information	None	
InfoText	Please check scan time!		
Scan Mode	RoutineSpiralAdultThorax	RoutineSpiralAdultThorax	
Pitch	0.8	1.2	
Decision Tree			
standard acquisition			
FoV?			
yes	no		
fast acquisition	standard acquisition		
Thorax and Abdomen FoV [factory]			
Scan Strategy			
	standard acquisition	fast acquisition	
InfoLevel	Information	None	
InfoText	Please check scan time!		
Scan Mode	RoutineSpiralAdultAbdomenFlexdose	RoutineSpiralAdultAbdomenFlexdose	
Pitch	0.8	1.2	
Decision Tree			
standard acquisition			
FoV?			
yes	no		
fast acquisition	standard acquisition		

Thorax and Abdomen contrast medium necessary [factory]		
Scan Strategy		
	wo CM	w CM
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultAbdomenFlexdose	RoutineSpiralAdultAbdomenFlexdose
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Soft Tissue with Contrast
Injector Coupling	Off	On
Contrast Protocol		Abdomen [factory]
Absolute Start Delay	2	20
Decision Tree		
	wo CM	
	Contrast medium?	
yes	no	
w CM	wo CM	

Thorax and Abdomen VNC TwinBeam DE possible [factory]		
Scan Strategy		
	TwinBeam DE	Single Energy
InfoLevel	None	Important Information
InfoText		No Dual Energy information!
Scan Mode	RoutineSpiralAdultAbdomenTwinbeamFlexdose	RoutineSpiralAdultAbdomenFlexdose
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	AuSn 120	70
CARE kV max. kV	AuSn 140	140
Decision Tree		
	TwinBeam DE	
	Dual Energy Recommendation	
Dual Energy	Single Energy	
TwinBeam DE	Single Energy	

Abdomen contrast medium necessary [factory]		
Scan Strategy		
	wo CM	w CM
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultAbdomen	RoutineSpiralAdultAbdomen
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Soft Tissue with Contrast
Injector Coupling	Off	On
Contrast Protocol		Abdomen [factory]
Absolute Start Delay	2	75
Decision Tree		
	wo CM	
	Contrast medium?	
yes	no	
w CM	wo CM	

Abdomen FoV [factory]		
Scan Strategy		
	standard acquisition	fast acquisition
InfoLevel	Information	None
InfoText	Please check scan time!	
Scan Mode	RoutineSpiralAdultAbdomen	RoutineSpiralAdultAbdomen
Pitch	0.8	1.2
Decision Tree		
	standard acquisition	
	FoV?	
yes	no	
fast acquisition	standard acquisition	

Abdomen VNC TwinBeam DE possible [factory]		
Scan Strategy		
	TwinBeam DE	Single Energy
InfoLevel	None	Important Information
InfoText		No Dual Energy information!
Scan Mode	RoutineSpiralAdultAbdomenTwinbeam	RoutineSpiralAdultAbdomen
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	AuSn 120	70
CARE kV max. kV	AuSn 140	140
Decision Tree		
	TwinBeam DE	
	Dual Energy Recommendation	
Dual Energy	Single Energy	
TwinBeam DE	Single Energy	

Kidney TwinSpiral DE possible [factory]		
Scan Strategy		
	TwinSpiral DE	Single Energy
InfoLevel	None	Important Information
InfoText		No Dual Energy information. Please check patient instruction!
Scan Mode	RoutineSpiralAdultAbdomenTwinspiral	RoutineSpiralAdultAbdomenSn
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	80	Sn 100
CARE kV max. kV	100	Sn 140
Decision Tree		
TwinSpiral DE		
Dual Energy Recommendation		
Dual Energy	Single Energy	
TwinSpiral DE	Single Energy	
Whole Body soft tissue or bone [factory]		
Scan Strategy		
	standard dose	low dose
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultAbdomen	RoutineSpiralAdultAbdomenSn
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Bone/Calcium
CARE kV IQ level	145	72
Decision Tree		
standard dose		
Soft tissue/bone?		
bone	soft tissue	
low dose	standard dose	
Pelvis soft tissue or bone [factory]		
Scan Strategy		
	standard dose	low dose
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultPelvis	RoutineSpiralAdultPelvisSn
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Bone/Calcium
CARE kV IQ level	170	85
Decision Tree		
standard dose		
Soft tissue/bone?		
bone	soft tissue	
low dose	standard dose	
Pelvis contrast medium necessary [factory]		
Scan Strategy		
	wo CM	w CM
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultPelvis	RoutineSpiralAdultPelvis
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Soft Tissue with Contrast
Injector Coupling	Off	On
Contrast Protocol		Abdomen [factory]
Absolute Start Delay	2	85
Decision Tree		
wo CM		
Contrast medium?		
yes	no	
w CM	wo CM	
Pelvis TwinSpiral DE possible [factory]		
Scan Strategy		
	TwinSpiral DE	Single Energy
InfoLevel	None	Important Information
InfoText		No Dual Energy information. Please check patient instruction!
Scan Mode	RoutineSpiralAdultPelvisTwinspiral	RoutineSpiralAdultPelvis
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	80	70
CARE kV max. kV	100	140
Decision Tree		
TwinSpiral DE		
Dual Energy Recommendation		
Dual Energy	Single Energy	
TwinSpiral DE	Single Energy	

C-Spine disc or bone [factory]		
Scan Strategy		
	standard dose	low dose
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultNeck	RoutineSpiralAdultNeckSn
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Bone/Calcium
CARE kV IQ level	195	97
Decision Tree		
	standard dose	
	Spine disc/fracture?	
spine disc	fracture	
standard dose	low dose	
T- and L-Spine disc or bone [factory]		
Scan Strategy		
	standard dose	low dose
InfoLevel	None	None
InfoText		
Scan Mode	RoutineSpiralAdultSpine	RoutineSpiralAdultSpineSn
CARE Dose4D & CARE kV	Full	Full
CARE kV optimized for	Non-Contrast	Bone/Calcium
CARE kV IQ level	285	142
Decision Tree		
	standard dose	
	Spine disc/fracture?	
spine disc	fracture	
standard dose	low dose	
Spine TwinSpiral DE possible [factory]		
Scan Strategy		
	TwinSpiral DE	Single Energy
InfoLevel	None	Important Information
InfoText		No Dual Energy information. Please check patient instruction!
Scan Mode	RoutineSpiralAdultSpineTwinspiral	RoutineSpiralAdultSpine
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	80	70
CARE kV max. kV	100	140
Decision Tree		
	TwinSpiral DE	
	Dual Energy Recommendation	
Dual Energy	Single Energy	
TwinSpiral DE	Single Energy	
Spine TwinSpiral metal DE possible [factory]		
Scan Strategy		
	TwinSpiral DE	Single Energy
InfoLevel	None	Important Information
InfoText		No Dual Energy information. Please check patient instruction!
Scan Mode	RoutineSpiralAdultSpineTwinspiralMetal	RoutineSpiralAdultSpine
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	80	70
CARE kV max. kV	100	140
Decision Tree		
	TwinSpiral DE	
	Dual Energy Recommendation	
Dual Energy	Single Energy	
TwinSpiral DE	Single Energy	
Brain CTA TwinBeam DE possible [factory]		
Scan Strategy		
	TwinBeam DE	Single Energy
InfoLevel	None	Important Information
InfoText		No Dual Energy information!
Scan Mode	RoutineSpiralAdultAngioHeadTwinbeam	RoutineSpiralAdultAngioHead
CARE Dose4D & CARE kV	Full	Full
CARE kV min. kV	AuSn 120	70
CARE kV max. kV	AuSn 140	140
Decision Tree		
	TwinBeam DE	
	Dual Energy Recommendation	
Dual Energy	Single Energy	
TwinBeam DE	Single Energy	

Neck CTA TwinBeam DE possible [factory]					
Scan Strategy					
	TwinBeam DE		Single Energy		
InfoLevel	None		Important Information		
InfoText			No Dual Energy information!		
Scan Mode	RoutineSpiralAdultAngioNeckTwinbeam		RoutineSpiralAdultAngioNeck		
CARE Dose4D & CARE kV	Full		Full		
CARE kV min. kV	AuSn 120		70		
CARE kV max. kV	AuSn 140		140		
Decision Tree					
TwinBeam DE					
Dual Energy Recommendation					
Dual Energy	Single Energy				
TwinBeam DE	Single Energy				
Pulmonary and Body CTA TwinBeam DE possible [factory]					
Scan Strategy					
	TwinBeam DE		Single Energy		
InfoLevel	None		Important Information		
InfoText			no Dual Energy information!		
Scan Mode	RoutineSpiralAdultAngioBodyTwinbeam		RoutineSpiralAdultAngioBody		
CARE Dose4D & CARE kV	Full		Full		
CARE kV min. kV	AuSn 120		70		
CARE kV max. kV	AuSn 140		140		
Injector Coupling	On		On		
Post Bolus Delay (BT)	6		2		
Decision Tree					
TwinBeam DE					
Dual Energy Recommendation					
Dual Energy	Single Energy				
TwinBeam DE	Single Energy				
Thorax Coronary CTA HR variability and HR [factory]					
Scan Strategy					
	standard	low dose	high heart rate	end systolic arrhythmia	
InfoLevel	None	None	None	None	
InfoText					
Scan Mode	CardiacSpiralAdultAngioFlexdose	CardiacSpiralAdultAngioFlexdose	CardiacSpiralAdultAngioFlexdose	CardiacSpiralAdultAngioFlexdose	
Phase Start	70	75	35	300	
Phase End	80	75	45	400	
Phase Unit	%	%	%	ms	
Decision Tree					
standard					
Heart Rate Variability					
Regular			Unstable		Arrhythmic
standard			standard		
Heart Rate			Heart Rate		end systolic arrhythmia
Low	Moderate	High	Low	Moderate	High
low dose	standard	high heart rate	low dose	standard	high heart rate
Coronary Calcium Scoring HR variability [factory]					
Scan Strategy					
	standard	end systolic arrhythmia			
InfoLevel	None	None			
InfoText					
Scan Mode	CardiacSequenceAdult	CardiacSequenceAdult			
Phase Start	75	350			
Phase End	75	350			
Phase Unit	%	ms			
Decision Tree					
standard					
Heart Rate Variability					
Regular	Unstable	Arrhythmic			
standard	standard	end systolic arrhythmia			

Coronary CTA HR variability and HR [factory]						
Scan Strategy						
	standard		low dose		high heart rate	
InfoLevel	None		None		None	
InfoText						
Scan Mode	CardiacSpiralAdultAngio		CardiacSpiralAdultAngio		CardiacSpiralAdultAngioBisegment	
Phase Start	70		75		35	
Phase End	80		75		45	
Phase Unit	%		%		ms	
Decision Tree						
standard						
Heart Rate Variability						
Regular standard			Unstable standard		Arrhythmic end systolic arrhythmia	
Heart Rate			Heart Rate			
Moderate			Moderate			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate			high heart rate			
low dose			low dose			
standard			standard			
High			High			
high heart rate						

Patient Instruction TAVI [factory]

Scan Strategy

Text_CT_ClinicalDecisionTreeStrategyName_Index_Information_for_Patient_Instruction

InfoLevel

Important Information

InfoText

Please provide a manual audio instruction for combined Valve and TAVI planning ranges!

Decision Tree

Text_CT_ClinicalDecisionTreeStrategyName_Index_Information_for_Patient_Instruction

Twin Spiral contrast medium [factory]

Scan Strategy

Information for TwinSpiral

InfoLevel

Important Information

InfoText

To ensure constant contrast distribution, the delay needs to be larger than 75 seconds.

Decision Tree

Information for TwinSpiral

Info template [factory]

Scan Strategy

template 1

InfoLevel

None

InfoText

Decision Tree

template 1

Extremity DE template [factory]

Scan Strategy

template 1

InfoLevel

None

InfoText

Scan Mode

RoutineSpiralAdultExtremities

Decision Tree

template 1

Brain CTA template [factory]

Scan Strategy

template 1

InfoLevel

None

InfoText

Scan Mode

RoutineSpiralAdultAngioHead

Decision Tree

template 1

Neck CTA template [factory]

Scan Strategy

template 1

InfoLevel

None

InfoText

Scan Mode

RoutineSpiralAdultAngioNeck

Decision Tree

template 1

Body CTA template [factory]

Scan Strategy

template 1

InfoLevel

None

InfoText

Scan Mode

RoutineSpiralAdultAngioBody

Decision Tree

template 1

Extremity CTA template [factory]

Scan Strategy

template 1

InfoLevel

None

InfoText

Scan Mode

RoutineSpiralAdultAngioExtremities

Decision Tree

template 1

Contrast Protocol Name	CM Name	Iodine Concentration [mg/ml]	Phase Type	Flow [ml/s]	Volume [ml]	Duration [s]	CM Ratio [%]	License
Factory Contrast Protocols								
Brain [factory]		300						
Contrast			Contrast	2.5	60	24	100	
Saline			Saline	2.5	40	16	0	
Brain Perfusion [factory]		370						
Contrast			Contrast	6.0	41	7	100	
Saline			Saline	6.0	40	7	0	
Neck [factory]		300						
Contrast			Contrast	2.5	70	28	100	
Saline			Saline	2.5	40	16	0	
Thorax [factory]		300						
Contrast			Contrast	3.0	70	24	100	
Saline			Saline	3.0	40	14	0	
Abdomen [factory]		300						
Contrast			Contrast	4.0	100	25	100	
Saline			Saline	4.0	40	10	0	
Liver Perfusion [factory]		370						
Contrast			Contrast	5.5	50	10	100	
Saline			Saline	5.5	40	8	0	
CTA [factory]		300						
Contrast			Contrast	4.0	80	20	100	
Saline			Saline	4	40	10	0	
Brain and Neck CTA [factory]		370						
Contrast			Contrast	5.0	80	16	100	
Saline			Saline	5.0	40	8	0	
Pulmonary CTA [factory]		370						
Contrast			Contrast	5.0	65	13	100	
Saline			Saline	5.0	40	8	0	
Coronary CTA [factory]		320						Cardiac
Contrast			Contrast	6.0	60	10	100	
Saline			Saline	6.0	40	7	0	



Manufacturer's note:

The CE marking applies only to Medical Devices which have been put on the market according to the EC Regulation.

Unauthorized changes to this product are not covered by the CE mark and the related Declaration of Conformity.

Caution: Federal law restricts this device to sale by or on the order of a physician, dentist, or veterinarian (21 CFR 801.109(b)(1)).

The original language of this document is English.

Siemens Healthineers Headquarters

Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen
Germany
Phone: +49 9131 84-0
siemens-healthineers.com

Manufacturer

Siemens Shanghai Medical Equipment Ltd.
278 Zhou Zhu Road
201318 Shanghai
P.R. China

EU Authorized Representative

Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen
Germany