

FCC Compliance for Network Analyzers

All test and measurement equipment is exempt from FCC compliance standards because it is used to work on open, unshielded equipment. So, Keysight network analyzers are not referenced to an FCC compliance number. However, our CISPR Pub 11 Level A compliance is roughly equivalent to FCC Part 15 Level A compliance.

The reference to the FCC exclusion is in CFR 47, Part 15, para 15.103(c), as follows:

"15.103 Exempted Devices.

(c) A digital device used exclusively as industrial, commercial, or medical test equipment."

The FCC published this exemption in April 1980.

Was this helpful? Yes No

Didn't find what you're looking for?

Contact Support

DISCOVER

Products +
Services
Solutions
Industries
Events
Keysight
University

INSIGHTS

Discover
Insights
Success Stories
Resources
Blog
Community

PARTNERS

SUPPORT

Keysight Product
Support
Ixia Product
Support
Manage
Software
Licenses
Product Order
Status
Parts

ABOUT KEYSIGHT

Newsroom
Investor

Relations

Quality and
Security

Corporate Social
Responsibility

Diversity, Equity,
and Inclusion

Modern Slavery

Act

Transparency
Statement

Careers

© Keysight Technologies 2000–2023 | Privacy | Sitemap | Terms
| Trademark Acknowledgements | Feedback