

ART-LBL-01 Unit label

PKG-EEG-015 Rev B
generated by tools/artwork_gen.py
2026-09-02 CC BY-SA 4.0

Size	50 x 25 mm
Substrate	Matt white polyester, 50 µm
Print	Thermal transfer, resin ribbon, black
Finish	Clear over-laminate, 25 µm
Adhesive	Permanent acrylic, service -40 to +90 °C
Placement	POD-P1 lid outer face, on the flat 55 x 30 mm keep-out, clear of the gasket line
Data Matrix	ECC200, 32 x 32 modules in 12.0 mm, module 0.375 mm, quiet zone 2 modules = 0.75 mm. At 300 dpi a module is 4.4 dots.



Variable data -- these fields are NOT artwork and must be supplied before print

<<UNIT_SERIAL>>	Allocated by the programme with the purchase order; PKG-EEG-015 section 5.
<<ATECC_FACTORY_SERIAL_18_HEX>>	Read at TST-EEG-004 Rev C T00 and T6.
<<KEY_FINGERPRINT_16_HEX>>	Computed as FW-EEG-001 Rev C section 7 defines it, at T6.

Notes

- The Data Matrix payload is ``<serial>|<hw rev>|<18 hex ATECC serial>|<16 hex fingerprint>``, pipe-delimited with no spaces, per PKG-EEG-015 section 4.1. It cannot be drawn as a placeholder, so the template shows the reserved area and the specimen below carries a real symbol.
- The specimen is drawn for TIOV-B-0000. PKG-EEG-015 section 5 allocates 0001 to 0009, 0010 to 0099 and 0100 to 0999, so 0000 is in no block and a specimen print can never be mistaken for the label of a unit that exists. That reservation is proposed to the programme in PKG-EEG-015 section 9 and is not yet ruled.
- The fingerprint on the specimen is the illustrative string PKG-EEG-015 section 4.1 prints, which that section marks as illustrative. It is not the fingerprint of any key.
- The symbol has been checked by encoding and decoding it in this tool and against the ISO/IEC 16022 worked example. It has NOT been read by a barcode verifier. PKG-EEG-015 section 4.1 requires ISO/IEC 15415 grade C or better after the IPA test, and that first-article verification is what releases this artwork.