Lean Six Sigma Green Belt Certification Course



Learning Objectives

By the end of this lesson, you will be able to:

- Interpret measurement system analysis
- List the measurement system characteristics and system concepts
- Outline the measurement system properties
- Describe G R&R study



Problems at Java House Coffee

Java House Coffee customers had been complaining about the coffee temperature. However, investigation showed that the temperature was within the customers' desired range.







The investigation team discovered that the device used to take and read coffee temperature was malfunctioning. A new device showed coffee temperature was colder than what they had previously thought.

DIGITAL

Calculate and Interpret Measurement System Capability using GR&R



Measurement System Analysis is a technique that identifies measurement error or variation and its sources to reduce variation.




Obtain information

Establish criteria


Compare measurement methods

Form basis for evaluating a measurement method

Observed value = True value ± Measurement error



True variability = Process variability + Measurement variability



Measurement System Properties

System Characteristics

Accuracy

 $\mu_{\text{total}} = \mu_{\text{process}} \pm \mu_{\text{measurement}}$

Precision

$$\sigma^2_{\text{total}} = \sigma^2_{\text{process}} + \sigma^2_{\text{measurement}}$$




To have stable measurement system, it is important to focus on the accuracy, first by addressing measurement issues and then getting accurate results.

Measurement System Properties

Accuracy

Degree of conformity of measured or calculated value to its actual or true value







Precision

Degree to which repeated measurements, under unchanged conditions, show the same results





Measurement System Properties










DIGITAL PERATIONS


Understand Measurement System Characteristics

Measurement System Characteristics



Measurement Characteristics: Resolution

Resolution is the smallest reading the measurement scale can detect.






Resolution is the first characteristic that needs to be tested in an MSA.

Measurement Characteristics: Bias

Bias is the measure of the distance between the measured value and the true or actual value.






Possible causes for bias are need for equipment calibration, improper calibration, or the wrong gage used for the job.

Measurement Characteristics: Linearity

Linearity is a measure of consistency of bias over a range of measurement from the smallest number to a higher number and vice-versa.






Possible causes for linearity issues could be poor maintenance, worn instrument or equipment, or application measurement method.

Measurement Characteristics: Stability

Stability refers to the ability of a measurement system to show the same values, every time the same item is measured.






Possible causes for instability issues are normal aging and worn instrument.



Measurement Characteristics: Repeatability

Repeatability is the variation in measurements obtained with one measurement instrument, when it is used several times by one appraiser while measuring the same characteristic on the same part, repeatedly.






Possible causes for repeatability variation could be appraiser technique, position on part where measurement is taken, and lack of training.

Measurement Characteristics: Reproducibility

Reproducibility is variation in the average of measurements made by different operators using the same gage when measuring identical characteristics of the same part.





Possible causes for reproducibility variation could be difference in operator training, technique, or experience.

DIGITAL

Measurement System Concepts

Measurement Correlation

Measurement Correlation is the strength of relationship of multiple measurement systems for use by quantifying accuracy relative to a Gold Standard or Gold Tester.




A Gold Standard is the measurement system used to determine and assign specifications values.



Precision / Tolerance (P/T)

Precision/Tolerance (P/T) addresses what percent of the tolerance is taken up by measurement error.



A measurement system is

acceptable if P/T is less than 30%.

GAGE R&R and %GRR

Gage RR or Gage Repeatability and Reproducibility is a measure of measurement system variation.


$$\sigma_{MS} = \sqrt{\sigma_{EV}^2 + \sigma_{AV}^2}$$

Percent Gage R and R is a measurement system capability index.


$$\%GRR = \frac{\sigma_{MS}}{\sigma_{Process}} * 100$$

Target: %GRR< 30%

Sources of Variation



GRR Study: Components



GRR Study: Guidelines

GRR studies should be performed over the range of expected observations

Actual equipment should be used

Written procedures or approved practices should be followed

Measurement variability should be presented as is

Measurement variability should be separated into causal components, prioritized, and targeted for action



GRR Study: Variable and Attribute Gage R&R

Variable Gage R&R

Analyzes measurement systems with the help of variable or continuous data

Measurement system typically involves a physical gage.

The result of this is quantification of the percentage of variation contributed by the measurement system.

Attribute Gage R&R

Analyzes measurement systems using attribute or discrete data

Measurement system typically utilizes manual or automated counting/monitoring.

The result of this is quantification of the proportion of defective measurements in DPMO, % Agreement, or Sigma Level.

Variable Gage R&R: Data Collection

The number of operators are usually 3.

The number of units to measure is usually 10.

General sampling techniques are used to represent the population.


The number of trials for each operator is 2 to 3.

The gage is checked for calibration and resolution.

The units are measured by the first operator in random order, and the same order is followed by the other operators.

Each trial is repeated.

Variable Gage R&R: Four Criteria





It is important to assess these values to determine whether each value meets the requirement to ensure that the measurement system can reliably measure the process.

Variable Gage R&R

Average and Range Method

- For years, the most used method for Gage R&R was the Average and Range Method
- Many people refer to the AIAG's
 Measurement Systems Analysis Manual
 when doing Gage R&R studies
 - AIAG stands for Automotive Industry Action Group
- Calculates the % of the total variation, using standard deviation, that is due to the Gage R&R

ANOVA Method

- ANOVA is considered the best method for analyzing GRR studies due to the following reasons:
 - It separates equipment and operator variation and provides insight on the combined effect of the two.
 - It uses variance instead of standard deviation as a measure of variation and therefore gives a better estimate of the measurement system variation.
- Allows you to estimate the variance of the components

Variable Gage R&R: Continuous Gage R&R Example

Characteristics:	Outside Diameter			Gage Number:		1234			Performed by:		AJH	
Specifications:	25 p	25 plus or minus 5		G	age Type:		Indicator					
Upper Spec	30		of Trials =	3	K ₁ =			Kbar diff =	1.05	D ₄ =	2.58	
Lower Spec	20	•	praisers =	3	K ₂ =	0.5231		Rbarbar =	1.40951	R,=	6.94476	
Total Tol	10	#	of parts =	10	K3=	0.3146						
	Part											
Appraiser/Trial#	1	2	3	4	5	- 6	7	8	9	10	Average	
Α 1	25.2853	27.2206	24.9934			25.6739	28.1287	22.6879	27.5386	27.6878	26.10238349	
2	24.3149	26.3773	26.3198	29.7438		24.5878	29.2963	22.988	25.1735	27.4172	25.86167087	
3	23.9702	27.5079	25.2571	29.5507	23.0531	26.3858	28.8585	22.9091	25.0981	26.3628	25.8953406	
Average	24.5235	27.0353	25.5235	29.4879	22.6965	25.5492	28.7612	22.8617	25.9367	27.1559	Xbar _a = 25.9531	
Range	1.31504	1.13068	1.32638	0.57474	0.6551	1.79798	1.16761	0.30011	2.44047	1.32507	Rbar a = 1.20332	
B 1	24.2795	26.4859	24.3112	29.4141	22.4786	24.9739	28.5552	22.9627	26.1618	26.9581	25.6580863	
2	24.6539	26.1579	25.1468	29.7305	23.5527	25.1368	27.5264	21.1795	25.6621	26.6376	25.53839788	
3	24.6816	25.7614	24.2967	29.7714	23.4755	25.5896	28.9718	21.3315	26.5696	26.7966	25.72457364	
Average	24.5383	26,1351	24.5849	29.6386	23.1689	25.2334	28.3511	21.8246	26,1312	26.7974	Xbar _b = 25.6404	
Range	0.40219	0.72453	0.85009	0.35728	1.07411	0.61569	1.4454	1.78315	0.90758	0.3205	Rbar _k = 0.84805	
C 1	25.5592	26.49	25.2298	31.759	25.0551	28.2445	30.1261	23.5133	27.1405	28.0329	27.11504426	
2	24.5037	27.094	25.1838	29.3502	22.7241	27.1026	27.5052	22.8021	25.379	26.9182	25.85627872	
3	27.7377	25.6571	24.1554	28.4375	25.8523	28.7592	29,9997	24.0493	27.2547	29.0944	27.09974051	
Average	25.9335	26.4137	24.8563	29.8489	24.5438	28.0354	29.2103	23,4549	26.5914	28.0151	Xbar _s = 26.6904	
Range	3.23396	1.43681	1.07436	3.32146	3.12815	1.6566	2.62097	1.24724	1.87571	2.17621	Rbar _s = 2.17715	
Part Average	24.9985	26.528	24.9882	29.6585	23.4698	26.2727	28.7742	22.7137	26.2198	27.3228	Xbarbar = 26.0946 R _p = 6.94476	
[(Rbar a = 1.20331882984201 + Rbar b = 0.848052350183393 + Rbar c = 2.17714672507304) / # of appraisers = 3] = Rbarbar									= Rbarbar	Rbarbar = 1.40951		
(Max Xbar = 26.69	03544987	033) - Min	Xbar = 25.6	64035260	6906) = Xb	ar diff					Xbardiff = 1.05	
(Rbarbar = 1,40950	059683661	15) x (D4 = 2	2.58) = UCL	_R							UCLR = 3.63653	

Variable Gage R&R: Continuous Gage R&R Example

Part No. & Name	· 123	3456, Shaft	Gage Name:			Shaft Gage Date:			March 30th		
Characteristics		ide Diameter	Gage Number:		1234		Perform		AJH		
Specifications		us or minus 5	Gage Type:			Indicator		/			
	. <u>20 p</u> .	33 OT 11111133 O	_	-gyp		II Iaioavoi					
Upper Spe	c 30	# of Trials =	3] K ₁ =	0.5908	l Xbar diff = I	1.05	D _a =	2.58		
Lower Spe		# of appraisers =	3	K ₂ =		Rbarbar =	1.40951	· -	5.94476		
Total To		# of parts =		K ₃ =							
]							
	Me	asurement Unit Ana	alysis				% Total Var	riation			
Repeatability - Equipment Variation (EV)						Percent Equipment \					
EV	= (Rbarbar) x	(K ₁)					100 [EV / TV]]			
	= (1.4095059	16836615) × (0.5908	3)	Trials	K _t	=	100[0.83277	2.3969]			
	= 0.8327			2	0.8862	=	34.741 %				
				3	0.5908						
Reproducibility -	Appraiser Va	riation (AV)				Percent Appraiser Variation					
AV	= SQRT{(Xba	ar diff x K 2)^2 - (EV/[(# parts) x (# trials)])}		%AV = 100[AV/TV]					
	= SQRT(0.30	117 - 0.0231)				= 100[0.5278/2.3969]					
	= 0.5278		Appraisers		3	=	22.020 %				
			K2	0.7071	0.5231						
Repeatability & F	-					Percent Gage Repe			y Variation		
	= SQRT[(EV)						100 [GRR / T'				
		934 + 0.2786]					100 [0.98597	2.3969]			
	= 0.9859					=	41.132 %				
Part Variation (P	Vì			Parts	K ₃	Percent Part Variation					
•	-, = R,×K3			2	0.7071		 100[PV/TV]			
	= 2.1848			3	0.5231		100 [2.1848]	_			
				4	0.4467		91.151 %	_			
Total Variation (T	'V)			5	0.403						
TV	= SQRT[(GR	R^2) + (PV^2)]		6	0.3742	Number of Distinct C	ategories that	can be Di	stinguished		
	= SQRT [0.97	72 + 4.7734]		7	0.3534		1.41(PV/GRI				
	= 2.3969			8	0.3375	=	1.41(2.18487	0.9859)			
				9	0.3249	=	3.125 or	approxima	ately: 3		
				10	0.3146						



Variable Gage R&R: Continuous Gage R&R Example

Part No. & Name:	123	3456, Shaft	l Ga	ge Name:		Shaft Gage	Da	te: March 30th			
Characteristics:		ide Diameter		e Number:			Performed b				
Specifications:		us or minus 5	Gage Type:		Indicator						
орсоновного.		33 011111100 0] _	-ga .ypa.		i raioasor	J				
Upper Spec	30	# of Trials =	3	K ₁ =	0.5908	Xbar diff =	1.05) ₄ = 2.58			
Lower Spec		# of appraisers =	3	K ₂ =		Rbarbar =		R _s = 6.94476			
Total Tol		# of parts =	10	K ₃ =		i ibaibai	1. 10001	, C.OTTIO			
10(01101	اڭــــــــــــــــــــــــــــــــــــ	or paids		'`3	0.0110						
	Me	asurement Unit Ana	lysis	% Total Tolerance							
Repeatability - Equipment Variation (EV)						Percent Equipment ¹	Variation				
		(K ₁) x (5.1Sigma)					100 [EV / Total To	ol]			
=	(1.4095059	i6836615) x (0.5908) x (5.1)	Trials	K _t	=	100 [4.247 / 10]				
=	4.247			2	0.8862	=	42.470 %				
				3	0.5908						
Reproducibility - A	Appraiser Va	riation (AV)				Percent Appraiser Variation					
AV =	(5.1Sigma)	$\times SQRT\{(Xbardiff\times$	K 2)°2 - (E\	//[(# parts)	x (# trials)]])}					
=	(5.1) x SQR	T(0.3017 - 0.0231)				=	100 [2.6918 / 10]				
=	2.6918		Appraisers		3	=	26.918 %				
			K2	0.7071	0.5231						
Repeatability & Re	-					Percent Gage Repe	· · · · · · · · · · · · · · · · · · ·				
		'2) + (AV''2)]					100 [GRR / Total	Tol]			
		037 + 7.2458]				=	100 [5.0282710]				
=	5.0282					=	50.282 %				
Part Variation (PV)	1			Parts	K ₃	Percent Part Variation	NE.				
	, (R₅)×(K₃)×	(5.1Sigms)		Pails 2	0.7071		100 [PV / Total To	-J.1			
	• •	(3. r 3igma) 14409127) x (0.3146) o (5.1)	3	0.5231		100 [11.14248 / 10				
	11.1425	14403121) / (0.3140) A (O. I)	4	0.4467	=	111.425 %				
_	II. ITEU			5	0.403	_	111.420 /1				
Total Tolerance (T	otal Toli			6	0.3742	Number of Distinct C	ategories that can	be Distinguished			
•	· ·	o - Lower Spec		7	0.3534		1.41(PV/GRR)				
	30 - 20	- -		8	0.3375		1.41(11.1424875.)	0282)			
	10			9	0.3249			roximately: 3			
				10	0.3146			,			



Part to Part

Total Variation

11.04723

12.27501

Variable Gage R&R: Continuous Gage R&R Example

Total Gage R&R 1.039055 1.079634 19.00% Repeatability 0.867345 0.752287 13.24% Reproducibility 0.572143 0.327347 5.76% Reproducibility 0.572143 0.327347 5.76% Reproducibility 0.5278 0.278573	G	age R&R Study	- ANOVA Method		G	age R&R Study -)	Kbar/Range Method	
Source St. Dev. Variance Variance Variance Cotal Gage R&R 1.039055 1.079634 19.00% Repeatability 0.867345 0.752267 13.24% Repeatability 0.8327 0.693389 Reproducibility 0.572143 0.327347 5.76% Reproducibility 0.572143 0.327347 5.76% Reproducibility 0.5278 0.278573 0.278573 Operator*Part 0.262011 0.06865 1.21% Operator*Part 0.262011 0.06865 1.21% Operator*Part 0.262011 0.06865 1.21% Operator*Part 0.262011 0.06865 1.21% Operator*Data 2.383497 5.681056 100.00% Operator*Data	Variance and Stand	lard Deviation C	Components		Variance and Sta	ndard Deviation C	Components	
Total Gage R&R 1.039055 1.079634 19.00% Repeatability 0.867345 0.752287 13.24% Repeatability 0.867345 0.752287 13.24% Repeatability 0.572143 0.327347 5.76% Repeatability 0.572143 0.327347 5.76% Reproducibility 0.572143 0.258697 4.55% Part to Part 2.1848 4.773351 Total Variation 2.38693 4.601422 81.00% Reproducibility 2.383497 5.681056 100.00% Repeatability 2.383497 2.681056 100.00% Repeatability 2.383497 Reproducibility 2.383497 3.681056 100.00% Repeatability 3.55 Standard Deviations 3.681056 3.639% 3				<u>% of</u>				
Repeatability	Source	St. Dev.	<u>Variance</u>	<u>Variance</u>	Source	St. Dev.	<u>Variance</u>	
Reproducibility 0.572143 0.327347 5.76% Operator 0.508623 0.258697 4.55% Operator 0.508623 0.258697 4.55% Operator 0.262011 0.06865 1.21% Operator 0.262011 0.06865 1.21% Operator 0.282011 0.06865 1.21% Operator 0.283497 0.681056 100.00% Operator 0.2383497 0.681056 100.00% Operator 0.2383497 0.681056 0.000% Operator 0.2383497 0.681056 0.000% Operator 0.2383497 0.681056 0.000% Operator 0.2383497 0.681056 0.000% Operator 0.2888888888888888888888888888888888888	Total Gage R&R	1.039055	1.079634	19.00%		0.9859	0.971999	
Operator 0.508623 0.258697 4.55% Operator 2.1848 4.773351	Repeatability	0.867345	0.752287	13.24%	Repeatability	0.8327	0.693389	
Total Variation 2.3969 5.74513	Reproducibility	0.572143	0.327347	5.76%	Reproducibility	0.5278	0.278573	
Part to Part 2.145093 4.601422 81.00% Total Variation 2.383497 5.681056 100.00% Process Tolerance = 10 Frocess Tolerance = 10	Operator	0.508623	0.258697	4.55%	Part to Part	2.1848	4.773351	
Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process Tolerance = 10 Process To	Operator*Part	0.262011	0.06865	1.21%	Total Variation	2.3969	5.74513	
Process Tolerance = 10 Gage R&R Using 5.15 Standard Deviations (99%) Study	Part to Part	2.145093	4.601422	81.00%				
Gage R&R Using 5.15 Standard Deviations (99%) Study % Study % of Deviation Source Variation Variation Tolerance Total Gage R&R 5.351131 43.59% 53.51% Total Gage R&R 5.077385 41.13% Repeatability 4.466826 36.39% 44.67% Repeatability 4.288405 34.74% Reproducibility 2.946535 24.00% 29.47% Reproducibility 2.71817 22.02% Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%	Total Variation	2.383497	5.681056	100.00%	Process Tolerance	:= 10		
Gage R&R Using 5.15 Standard Deviations (99%) Study % Study % of Deviation Source Variation Variation Tolerance Total Gage R&R 5.351131 43.59% 53.51% Total Gage R&R 5.077385 41.13% Repeatability 4.466826 36.39% 44.67% Repeatability 4.288405 34.74% Reproducibility 2.946535 24.00% 29.47% Reproducibility 2.71817 22.02% Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%								
Study % Study % of Study % Study % Study Source Variation Variation Total Gage R&R 5.351131 43.59% 53.51% Repeatability 4.466826 36.39% 44.67% Repeatability 4.288405 34.74% Reproducibility 2.946535 24.00% 29.47% Reproducibility 2.71817 22.02% Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%	Process Tolerance =	- 10						
Source Variation Variation Tolerance Source Variation Variation Total Gage R&R 5.351131 43.59% 53.51% Total Gage R&R 5.077385 41.13% Repeatability 4.466826 36.39% 44.67% Repeatability 4.288405 34.74% Reproducibility 2.946535 24.00% 29.47% Reproducibility 2.71817 22.02% Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%	Trococo Totoranos	10						
Total Gage R&R 5.351131 43.59% 53.51% Total Gage R&R 5.077385 41.13% Repeatability 4.466826 36.39% 44.67% Repeatability 4.288405 34.74% Reproducibility 2.946535 24.00% 29.47% Reproducibility 2.71817 22.02% Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%			iations (99%)		Gage R&R Using	5.15 Standard Dev	iations (99%)	
Repeatability 4.466826 36.39% 44.67% Repeatability 4.288405 34.74% Reproducibility 2.946535 24.00% 29.47% Reproducibility 2.71817 22.02% Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%		15 Standard Dev		<u>% of</u>	Gage R&R Using			
Reproducibility 2.946535 24.00% 29.47% Reproducibility 2.71817 22.02% Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%		15 Standard Dev	% Study			<u>Study</u>	% Study	
Operator 2.619409 21.34% 26.19% Part to Part 11.25172 91.15%	Gage R&R Using 5.	15 Standard Dev Study Variation	% Study Variation	Tolerance	<u>Source</u>	Study Variation	% Study Variation	
	Gage R&R Using 5.* Source Total Gage R&R	Study Variation 5.351131	% Study Variation 43.59%	Tolerance 53.51%	Source Total Gage R&R	Study Variation 5.077385	% Study Variation 41.13%	
Operator*Part 1.349357 10.99% 13.49% Total Variation 12.34404 100.00%	Gage R&R Using 5.* Source Total Gage R&R Repeatability	Study Variation 5.351131 4.466826	% Study Variation 43.59% 36.39%	Tolerance 53.51% 44.67%	Source Total Gage R&R Repeatability	Study Variation 5.077385 4.288405	% Study Variation 41.13% 34.74%	
	Gage R&R Using 5.* Source Total Gage R&R Repeatability Reproducibility	15 Standard Dev Study Variation 5.351131 4.466826 2.946535	% Study Variation 43.59% 36.39% 24.00%	Tolerance 53.51% 44.67% 29.47%	Source Total Gage R&R Repeatability Reproducibility	Study Variation 5.077385 4.288405 2.71817	% Study Variation 41.13% 34.74% 22.02%	

ndc = 2

110.47%

122.75%

90.00%

100.00%



Variable Gage R&R: Interpretation

Check the value of %GRR.

If %GRR < 30, Gage Variation is acceptable, and thus the gage is acceptable. If %GRR > 30, the gage is not acceptable.

Check EV first.

If EV = 0, the MS is reliable and the variation in the gage is contributed by different operators. If AV = 0, the MS is precise.

If EV = 0, resolve AV by providing operators with training.

Attribute Data: Data Collection in GR&R

Select between 20 to 30 samples that represent full range of variation

50:50 mix of good/bad parts is recommended

Have master appraiser categorize each of the test sample as good or bad

Take two or three inspectors and have them assess each test sample

Randomize samples and repeat assessment




Attribute Data: Example

Test Samples	Master/Expert	Operator1_1	Operator1_2	Operator2_1	Operator2_2	Operator3_1	Operator3_2
1	okay	okay	okay	okay	okay	okay	okay
2	defective	defective	defective	defective	defective	defective	defective
3	okay	okay	defective		okay	okay	okay
4	okay	okay	okay	okay	okay	okay	okay
5	okay	okay	defective	okay	okay	okay	okay
6	okay	okay	okay	okay	okay	okay	okay
7	okay	okay	okay	okay	okay	okay	okay
22	defective	defective	defective	defective	defective	defective	defective
23	defective	defective	defective	defective	defective	defective	defective
24	okay	okay	okay	okay	okay	okay	okay
25	okay	defective	defective	defective	defective	defective	defective
26	defective	defective	defective	defective	defective	defective	defective
27	okay	okay	okay	okay	okay	okay	okay
28	okay	okay	okay	okay	okay	okay	okay
29	okay	defective	defective	okay	okay	okay	defective
30	okay	defective	okay	okay	okay	okay	okay

Attribute Data: Example

Operator1_1	Operator1_2	Within Op1	Op1 with Standard	
okay	okay	1	1	
defective	defective	1	1	
okay	defective	0	0	
okay	okay	1	1	
okay	defective	0	0	
okay	okay	1	1	
okay	okay	1	1	
defective	defective	1	1	
defective	defective	1	1	
okay	okay	1	1	
defective	defective	1	0	
defective	defective	1	1	
okay	okay	1	1	
okay	okay	1	1	
defective	defective	1	0	
defective	okay	0	0	
	#Matched	27	27	
	#Inspected	30	30	
	%Agree	90.00%	80.00%	






Attribute Data: Example


Decision	Effectiveness	Miss Rate	False Alarm Rate
Acceptable for the appraiser	> 90%	< 2%	< 5%
Marginally acceptable for the appraiser	80% to 90%	2% to 5%	5%to 10%
Unacceptable for the appraiser	< 80%	> 5%	> 10%

Interpretation of MSA



Key Takeaways

- A Measurement System Analysis (MSA) is needed prior to any data collection activities.
- There are two major system characteristics: accuracy and precision.
- An MSA will help determine if the measurement system characteristics exist.
- The measurement system concepts are Measurement correlation, gage RR and % GRR, and Precision/Tolerance (P/T).
- Gage R&R studies can be conducted for variable and attribute data.
- Average and Range Method and ANOVA Method are the two methods for analyzing GRR Studies.



DIGITAL



Knowledge Check

A team studies the results from measurement system analysis of plate thickness and determines that the average values observed is 3.5 cm, with values close together, and the true value is 4.0 cm. How can the measurement system be described?

- A. Low Accuracy, High Precision
- B. Low Accuracy, Low Precision
- C. High Accuracy, High Precision
- D. High Accuracy, Low Precision





1

A team studies the results from measurement system analysis of plate thickness and determines that the average values observed is 3.5 cm, with values close together, and the true value is 4.0 cm. How can the measurement system be described?

- A. Low Accuracy, High Precision
- B. Low Accuracy, Low Precision
- C. High Accuracy, High Precision
- D. High Accuracy, Low Precision



The correct answer is A

The scenario described here is of low accuracy (average observed value does not equal target value) and high precision (the values are close together).



2

The change or lack of change of a measurement system readings of the same part over time is an example of _____.

- A. Linearity
- B. Stability
- C. Repeatability
- D. Reproducibility



2

The change or lack of change of a measurement system readings of the same part over time is an example of _____.

- A. Linearity
- B. Stability
- C. Repeatability
- D. Reproducibility



The correct answer is **B**

The change or lack of change of a measurement system readings of the same part over time is an example of stability.

3

A team has measured the thickness of a part over and over and has compared the average observed value to the actual value. The difference between the average observed value and the actual value is called:

- A. Linearity
- B. Bias
- C. Stability
- D. Gage R&R



3

A team has measured the thickness of a part over and over and has compared the average observed value to the actual value. The difference between the average observed value and the actual value is called:

- A. Linearity
- B. Bias
- C. Stability
- D. Gage R&R



The correct answer is **B**

Bias is the measurement characteristic described as the difference between the average observed value and a reference value.



What is the major difference between the ANOVA and Average and Range method for Gage R&R study?

- A. The results are significantly different
- B. ANOVA is based on standard deviation and the Average and Range method is based on variance
- C. There is no difference
- D. ANOVA is based on variance and the Average and Range method is based on standard deviation





What is the major difference between the ANOVA and Average and Range method for Gage R&R study?

- A. The results are significantly different
- B. ANOVA is based on standard deviation and the Average and Range method is based on variance
- C. There is no difference
- D. ANOVA is based on variance and the Average and Range method is based on standard deviation



The correct answer is **D**

ANOVA is based on variance and the Average and Range method is based on standard deviation which causes a slight difference in calculations.



5

To conduct a continuous Gage R&R study, how many sample parts are needed?

- A. 10
- B. 20
- C. 25
- D. 3





5

To conduct a continuous Gage R&R study, how many sample parts are needed?

- A. 10
- B. 20
- C. 25
- D. 3



The correct answer is A

For a continuous Gage R&R study, ideally 10 parts are recommended that represent the range of output possibilities from the process.

