

Fe PLS_kfold – External Validation

R^2 pred = 0.797

RMSE pred = 12700 ppm

Predicted XRF-calibrated (mg kg^{-1})

$2\text{e}+05$

$1\text{e}+05$

$0\text{e}+00$

$0\text{e}+00$

$1\text{e}+05$

$2\text{e}+05$

Observed ICP-MS (mg kg^{-1})

Site

- BI10
- HER42PB
- KER1
- KER3
- PB1

