

Ti RF – External Validation

R^2 pred = 0.974

RMSE pred = 914.6 ppm

Predicted XRF-calibrated (mg kg^{-1})

2e+04

1e+04

0e+00

0e+00

1e+04

2e+04

Observed ICP-MS (mg kg^{-1})

Site

● BI10

● HER42PB

● KER1

● KER3

● PB1

