IQMS User Letter 38 – Guidance on complaint handling for Breelib

Version: 11.0
Effective Date: 2025-03-03
Applicable to: Breelib Medical Device; CQU/GBS of Argentina, Austria, Chile, Germany, Italy, Poland, Portugal, UK and Colombia

Contents

1. Introduction 2
2. Tasks & Responsibilities of the Patient Support Program (PSP) 2
3. Tasks & Responsibilities of CQU/GBS 2
4. Guidance on PTC / AE cases related to patients under 18 Years 3
5. Guidance on Breelib Devices returned due to treatment stop 3
6. Guidance on Complaint Description & Error Coding 4

Table 1: Guidance on AE cases related to patients under 18 Years 3
Table 2: Proposal for Key Terms for "Title" to Complaint Description and Error Coding 4

INTERNAL

1. Introduction

This User Letter provides guidance on how to process PTCs or PTC + AEs related to Breelib a “Bayer branded device” in IQMS.

2. Tasks & Responsibilities of the Patient Support Program (PSP)

PSP receives complaints from Health Care Professional (HCP), nurse, patient etc.

Option 1: PSP documents Product Technical Complaints (PTC) or PTC + AE information in the “Breelib Report Form for PTC or PTC + AE” (Breelib Form).

Option 2: PSP includes PTC or PTC + AE information in Safe Track (in English) and adds additional relevant information in the section „**Additional context**”.

• PSP forwards filled out “Breelib Form” to Country Quality Unit (CQU)/Global Business Services (GBS) or enters respective PTC or PTC plus AE information in Safe Track

• PSP contacts the complainant if further information is requested by CQU/GBS

• PSP receives the corresponding IQMS PTC number from CQU/GBS

• **PSP sends the device to Vectura** and informs CQU/GBS about the shipment date.

To ensure that every returned sample is clearly identifiable and traceable upon receipt:

- Devices must be individually packaged (no loose bags) to prevent transport damage
- **All returned devices must be accompanied by the IQMS PTC number**

Remark: Device must be sent to Vectura within two weeks to enable timely investigation

• PSP informs CQU/GBS about no shipment e.g. if the device was not defective or if PSP was able to solve the issue. PSP provides information on the solution/steps to CQU/GBS.

Responsible Quality Unit/RQU responsible: SC Heredia (CR02)/see RQU Responsible list)

Shipping Address of the device	Vectura Delivery Devices Limited; “Breelib complaints”, 205 Cambridge Science Park Milton Rd, Milton; Cambridge CB4 0GZ Cambridge; United Kingdom

3. Tasks & Responsibilities of CQU/GBS

CQU/GBS enter all PTC or PTC + AE in IQMS incl. “Safe Track Printout” or “Breelib Form”.

Note: When using the "Breelib form", the form will be forwarded as a PDF from the RQU to Vectura. If Safe Track is used, only a IQMS printout (Formatted output) is sent to Vectura by the RQU. Therefore, all relevant information added in Safe Track under the section “additional context” has to be transferred to the field “Description”.

INTERNAL

• CQU/GBS exchange information with PV for reported Adverse Events and enter received Global Safety Database Number in IQMS

• CQU/GBS report back to PSP:

Option 1: "Breelib Form" (updated with IQMS PTC number)

Option 2: IQMS CMPL Identifier

CQU/GBS include information on the shipment or no shipment (should contain reason or further information) in IQMS and notify the RQU via Communication protocol via Notification. In addition:

• Include only one complaint per serial number/device in IQMS.

• If a new complaint (new error) is reported for the same device/serial number at a later timepoint enter a new PTC into IQMS.

• Do not close complaints (including use errors) locally to enable reporting to Vectura (see exception 5).

4. Guidance on PTC / AE cases related to patients under 18 Years

Table 1: Guidance on AE cases related to patients under 18 Years

Description	Comment	IQMS	Argus
Is the report related to a PTC of the device?	PTC only	Yes	No
Reports related to a PTC of the Breelib device and AEs e.g. missed dose, headache, dizziness etc. are reported?	PTC+AE	Yes	Yes
Special scenario: reports related to a PTC for the Breelib device and off-label use for Ventavis (patient < 18 years old) as this is the only trigger for a PV case	PTC only in IQMS and PV only case in Argus - **cases not LINKED**	yes	yes

5. Guidance on Breelib Devices returned due to treatment stop

If the Breelib device is returned not due to a Product Technical Complaint (PTC), but for reasons unrelated to the device—such as the stop of treatment due to the patient's death or other underlying health issues—please note the following:

No PTC Entry Required: If a PTC meets the criteria (AE = stop of treatment due to patient's death or other health issues), it should be canceled with the justification: "No PTC reported; device deactivation only."

- **Device Disposal – No Manufacturer Notification or Return Required:** There is no requirement to inform the Legal Manufacturer or to return the device for investigation; the device can be disposed of.

INTERNAL

6. Guidance on Complaint Description & Error Coding

Include the original description provided by the complainant, reporter, or PSP in the "Description" field in IQMS. If the information was submitted via Safe Track, transfer the details from the "Additional Context" section into the “Description” field.

Additionally, please review the “Complaint as Reported / Examples & Explanation” in Table 2, Column 1 and add the key terms as proposed in Column 2, along with the error coding proposed in Column 3 to the Complaint “Title”.

Example: Breelib, SNXXXXXXXXXXXX (if known), **Tried to turn on, but no response, Administration Device - Malfunction - Power Up Fault** SSFXXXXXXXX

Table 2: Proposal for Key Terms for “Title” to Complaint Description and Error Coding

Complaint as reported / Examples & Explanation	Key Terms for the “Title” in the Intake section	Level 1/Level 2/Level 3 (Object type/Error code/Error subcode
-Power button seems to stick -Power button seems to be trapped -Device could not be charged -Power button failed	Tried to turn on, but no response	Administration Device - Malfunction - Power Up Fault
-Device can't be charged -Device could not be charged regularly -Device does not charge with dedicated (Starter Pack) power adaptor	Device does not charge	Administration Device - Malfunction - Charge / discharge
-Breelib always turns red and switches off when using this nebulizer head -Breelib switches off shortly after starting inhalations -Switched often off during inhalation but could be started again -Device base unit flashes short green and turns off -Device switches off during inhalations, can't be charged regularly	Unexpected turn-off	Administration Device - Malfunction – Unexpected Turn-off
-Device base unit shines in all possible colors and switch off -Device base unit flashes short green and turns off -Device base unit blinks orange and switches off with beeping -Breelib always turns red and switches off when using a specific nebulizer head -After turning on all lights are on, then one acoustic signal and device turns off	Unexpected signals (blinking lights or beep), not identified	Administration Device - Malfunction – Unexpected Signals
Device blocks during inhalation Drug is not nebulized	No release of the product	Administration Device - Malfunction – Product release impossible / incomplete
-Sluggish inhalations, valve seems to run roughly	Inhalation takes longer than normal	Administration Device - Malfunction – Application time

INTERNAL

Complaint as reported / Examples & Explanation	Key Terms for the "Title" in the Intake section	Level 1/Level 2/Level 3 (Object type/Error code/Error subcode
-Valve does not react -Higher resistance by taking breath from beginning	Inhalation requires more effort than normal	Administration Device - Malfunction - Inhalation Difficult
-Drug flows directly through the membrane of nebulizer head from beginning on -After some days of regular use Ventavis ran through the mesh	Ventavis runs through the mesh	Administration Device - Malfunction – Membrane
-Inhalations are possible, but the device does not switch off after inhalations -Device turns to a permanent bluetooth mode and tries to pair itself to the breeconnect	Device does not switch-off	Administration Device - Malfunction - Device does not Switch-off
-E.g. missing mouthpiece or IFU or QRC, or neb head	Device parts missing	-Administration Device - Quantity - Finished Package incomplete (parts missing) -Administration Device - Quantity - Final Product missing in shipping container

Abbreviations: AE = Adverse Event; CQU = Country Quality Unit; HCP = Health Care Professional; PSP = Patient Support Program; GBS=Global Business Services; PTC = Product Technical Complaint; Breelib form= Breelib Report Form for PTC with or without related AE

This User Letter remains valid until further notification

Support: Global Complaint Management
PH: Helene Janzen <helene.janzen1@bayer.com

INTERNAL