# IQMS User Letter 73 – Guidance on Complaint specifics for Radiology Medical Devices

Version: 1.0
Effective Date: 2025-12-08
Applicable to: Radiology medical devices only

## 1. Immediate escalation to RQU for complaints involving Radiology Medical Devices

GSOP 115 outlines cases requiring immediate escalation in section 4.1.2.
With this user letter **an immediate escalation (send complaint to the owner) to RQU** is required for complaints involving Radiology Medical Devices if complaints include allegations of Death, Air Injection or Extravasation.

Rationale: Radiology complaints are not sent to Argus for assessment of reportability to the authorities. These activities are subsumed by the RQU within IQMS. Thus, it is critical that such complaints associated with Radiology medical devices with allegations of Death, Air Injection or Extravasation are assigned and sent to RQU within 24 hours as indicated in section 4.1.2 of GSOP 115.

## 2. Field safety corrective action process

In case of a complaint is reported by a customer regarding missing documents (e.g. Field safety corrective action-related documents), a complaint should be logged within the IQMS system. This complaint may be logged with Feedback Type / Error Codes = Documentation. The initiator can then provide the missing documentation to the customer as needed and close the complaint. In case the missing documents are not available or cannot be sent, the initiator may utilize Immediate Escalation task and assign it to the relevant Quality Team member who can provide the missing documents. Once the missing documents are provided to the customer, the complaint may be closed **locally**.



INTERNAL