

CLiTICAL アプリの規制上の位置づけに関する見解

Statement on the Regulatory Status of the CLiTICAL Application

発行者 特定非営利活動法人 日本血管外科学会 JCLIMB 委員会
対象 iOS アプリケーション「CLiTICAL」(Bundle ID: org.studiomexx.critical-ios)
日付 _____

1. 目的

本アプリは、包括的高度慢性下肢虚血（CLTI）に対する血行再建術の術後転帰について、査読付き論文として公表済みの予測モデルに基づく参考情報を、医療従事者に提示することを目的とする。

2. 対象利用者

医師その他の医療従事者。患者本人またはその家族による使用を意図しない。

3. 機能

利用者が手入力した患者情報を、下記文献に記載された回帰式に代入し、集団レベルの推定値を表示する。センサー・HealthKit・画像・診断機器からの入力は受け取らず、測定・信号取得・画像解析の機能を有しない。診断名の提示、術式の提案、治療の実施可否に関する助言を行わない。

4. 出典

(説明変数・回帰係数・分類基準はすべて原著記載のとおりであり、独自のモデルを含まない)

Miyata T, et al. *Br J Surg*. 2022;109(11):1123. doi:10.1093/bjs/znab036

Miyata T, et al. *Eur J Vasc Endovasc Surg*. 2022. doi:10.1016/j.ejvs.2022.05.038

5. 規制上の位置づけ

上記の機能および使用目的に鑑み、本アプリは、既に公表・周知された知見を医療従事者に提示するとともに、診断および治療方針の決定は担当医師が行うものであることから、医薬品医療機器等法に定める医療機器プログラムに該当しないものと判断する。したがって、医療機器としての承認・認証・届出は行っておらず、これに対応する承認書等は存在しない。

6. 配信地域

予測モデルの導出集団に鑑み、当初より日本国内のみを配信対象としている。

署名 _____

役職 _____

連絡先 _____

Statement on the Regulatory Status of the CLiTICAL Application

English translation of the Japanese original

Issued by Japanese Society for Vascular Surgery, JCLIMB Committee
Subject iOS application “CLiTICAL” (Bundle ID: org.studiomexx.critical-ios)
Date _____

1. Purpose

The application presents healthcare professionals with reference information on outcomes after revascularisation for chronic limb-threatening ischaemia (CLTI), based on prediction models already published in peer-reviewed journals.

2. Intended users

Physicians and other healthcare professionals. It is not intended for use by patients or their families.

3. Function

The application applies patient information entered manually by the user to the regression equations published in the articles below, and displays population-level estimates. It receives no input from sensors, HealthKit, images or diagnostic instruments, and performs no measurement, signal acquisition or image analysis. It does not name a condition, propose a procedure, or advise for or against an intervention.

4. Sources (every explanatory variable, regression coefficient and classification threshold is as published; the application adds no model of its own)

Miyata T, et al. *Br J Surg*. 2022;109(11):1123. doi:10.1093/bjs/znab036

Miyata T, et al. *Eur J Vasc Endovasc Surg*. 2022. doi:10.1016/j.ejvs.2022.05.038

5. Regulatory status

In view of the function and intended purpose set out above — the application presents already-published knowledge to a healthcare professional, and diagnosis and treatment decisions are made by the attending physician — the Committee considers that the application is not a medical device program under Japan’s Act on Securing Quality, Efficacy and Safety of Pharmaceuticals and Medical Devices (the Pharmaceuticals and Medical Devices Act). It has therefore not been approved, certified or notified as a medical device, and no such approval document exists.

6. Availability

In view of the population from which the prediction models were derived, the application has been distributed in Japan only since its first release.

Signature _____

Title _____

Contact _____