

MEDICAL DEVICE RISK ASSESSMENT REPORT

ISO 14971 Inspired Risk Management System

Prepared By : G. Subiksha

Department : Biomedical Engineering

Version : 1.0

Generated : 10-07-2026 19:04

Executive Summary

Metric	Value
Total Hazards	6
Total FMEA Records	3
Critical Risks	2
High Risks	4
Medium Risks	0
Low Risks	0

Hazard Analysis

Hazard	Severity	Probability	Risk Score	Risk Level
battery failure	5	4	20	Critical
battery failure	4	3	12	High
battery failure	4	3	12	High
battery explosion	4	3	12	High
battery explosion	5	4	20	Critical
battery failure	4	3	12	High

Failure Mode and Effects Analysis (FMEA)

Failure Mode	Effect	Cause	RPN
battery failure	shutdown	battery shortage	160
			1

battery failure	shutdown	battery aging	160
-----------------	----------	---------------	-----

Regulatory Compliance

- ✓ Risk assessment performed following ISO 14971 principles.
- ✓ Failure Mode and Effects Analysis (FMEA) completed.
- ✓ Risk prioritization based on Severity × Probability.
- ✓ Mitigation measures documented for identified hazards.
- ✓ Report generated using Medical Device Risk Assessment Tool.

Approval

Prepared By	Reviewed By	Approved By
_____	_____	_____

This report is generated automatically by the Medical Device Risk Assessment Tool. It is intended for educational purposes and follows ISO 14971-inspired risk management principles.