

## Calibration Certificate

CC-4115

Format No-SCL/333/CL/F

ULR NO : CC41152500001365F	Certificate No.: 126741
Date of Receipt of UUC: 04-05-2025	Calibration Completion Date: 05-05-2025
Date of Issue of Certificate : 05-05-2025	Suggested Next Calibration Date : 05-05-2027

10-10-2025

1	Customer Details :
	Name : Torrent pharmaceuticals ltd Near Indrad village SH 41,
	Address : Mehsana 380009



SCAN TO VALIDATE  
CERTIFICATE

2	IUC Description :			
	Name	MEASURING CYLINDER CLASS A	Serial No	CBXS000407
	Make	SUPERTEK	Range/Capacity	10 ML ml
	Catalogue Number	72.504.0010IS	Accuracy	±0.10 ML ml
	Product Std	ISO 4788	Least Count	0.2 ML ml
	Calibration Std used	ISO 4787	Location	In Lab

3	IUC Condition : Good
---	----------------------

4	STANDARD EQUIPMENT USED AND THEIR TRACEABILITY					
	SL	Name	Identification	Traceble To	Certificate no	Validity
	1	BALANCE	METTLER	NIIRT	C4CAT/2081/C0873/MMV (1)	27-09-2025
	2	THERMOMETER	JRM	AUTOCAL	C2523650/T/1	16-02-2026

5	Environmental Conditions	Temperature : 26°C	RH : 48%	atmospheric pressure: 990.1 - 5 hPa
---	--------------------------	--------------------	----------	-------------------------------------

6	Calibration Results:			
	SL	Nominal Value(ml)	Observed Volume (ml) at 27 °C	Expanded Uncertainty at (95% C.L.K=2)
	1	10 <span>ML</span> <span>ml</span>	9.9378 <span>ML</span> <span>ml</span>	0.003 <span>ML</span>

### Notes:

- Above result are only related to the items calibrated.
- The above calibration result is valid at the time of calibration under lab conditions
- The Certificate shall not be reproduced except in full and without written permission from Supertek Calibration Laboratory Ambala, INDIA saha and add one more point
- The IUC was calibrated with standard calibrated measuring equipment's. The calibration of the same are traceable to Nation/International standards.
- The reported uncertainty in measurements is an expanded uncertainty at coverage factor "K=2", which corresponds to coverage probability and level of confidence of approximately 95%.
- Statement of conformity will be mentioned in this report only when requested by the customer in Customer Service Request Form and the decision rule is as per the Simple Acceptance Decision Rule.

### Calibrated By:

*Signature*

gurpreet

### Approved By:

*Signature*

ketan Sood

Address: Plot No. - 333HSIIC, Sector-2, Saha, Haryana, 133104

Address: 333 HSIIC, Saha Ambala, India