

Appendix 2 Summary of Responses to Fishtail Chimney Questionnaire

Participant	F/tail ?	Impinger?	flow-rate (L/min)	SS pad diameter (mm)	cigs per pad	ports per run	f/tail- plate distance (mm)
PM Europe	Y	Y	2.5	44	3	4	3
BAT	Y	N	2.5	47	3	4	30-38
JTI	Y	N	3	44	3	4	10
KGTRI	Y	N	2.3	92	3	4	0(?)
PDM	Y	N	2	44	3	4	8
ITL Montréal	Y	N	3	44	5	4	37*
Wattens	Y	N	2.5	44	3	1 (RM1)	6
PM USA		N (2 pads)	2.5	44	3	4	3
RJRT	Y	Y	3	44	3	20	9
Labstat	Y	Y	3	44	2(+2)	4	6

2505627034