

Net F/T
Force/Torque Sensor

Welcome	System Status: Healthy
Snapshot	
Demo	
Settings	Configurations User-defined configurations are displayed on this page. Use the <i>View Configuration</i> drop-down list and the <i>Go</i> button to display another configuration.
Thresholding	
Configurations	Each configuration loads a transducer calibration. A configuration can select the measurement system used for Force Units and Torque Units. A configuration can also apply a tool transformation to the output data.
Communications	
System Info	After you have created a configuration, you can enable it on the Settings page.
ATI Web Site	To save the values, you must click <i>Apply</i> .

View Configuration: #16 - empty ▼ Go

Configuration #16 (Active configuration)

Configuration Name: Maximum of 32 characters

Calibration Select: #2 - FT32404 ▼

Calibration Type: SI-12-0.12

Force Units: N ▼

Torque Units: Nm ▼

Counts per Force: 1000000

Counts per Torque: 1000000000

Calibrated Sensing Range (Units):

Fx	Fy	Fz	Tx	Ty	Tz
12	12	17	0.11999	0.11999	0.11999

Calibrated sensing range values apply to the factory origin (without tool transformation).

Scaling Factor for DeviceNet and CAN:

Fx	Fy	Fz	Tx	Ty	Tz
519	519	519	3663	3663	3663

Tool Transform Distance Units: in ▼

Tool Transform Angle Units: degrees ▼

Tool Transform:

Dx	Dy	Dz	Rx	Ry	Rz
0	0	0	0	0	0

Using a tool transformation will change how transducer readings are reported and change the apparent sensing ranges and apparent resolutions.

User-defined Field #1: Maximum of 16 characters

User-defined Field #2: Maximum of 16 characters

Apply Cancel