

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Date Of Issue: 02-03-2023

Sheet : 2 of 2

Format No. : VI-FRM-ME-006 ULR No.: CC247323100007699F Report No: VI/22-23/8113-02

CALIBRATED INSTRUMENT / EQUIPMENT DETAILS

Nomenclature	Setting Rods	Make	-----
Range / Resolution	50 mm	Sl. No.	-----
Calibration Done At	VI Mechanical Lab	Temperature / Humidity	20.2°C 50%RH
Calibrated on	02- 03- 2023	Calibration due on	01 - 03 - 2024
Discipline	Mechanical (Dimensional)	-	-

MASTER EQUIPMENT TRACEABILITY DETAILS

Sl.No.	Nomenclature	Make & Model	Sl. No./ID No.	Traceable Cert. No.	Traceable to	Validity
1	Digital Linear Height Master	Mitutoyo	300471007 / VI/ME/004	VI/22-23/INT-03	VI -Bangalore	15 - 07 - 2023

The master equipments used are traceable to National Standards Ref. Doc. Comparison Method and SOP-16-04

CALIBRATION RESULTS

All values are in mm

Sl.No.	Std. Values	Actual Values	Error
1	50 mm @ $\pm 0\mu\text{m}$	50.0042	+0.0042

Note :

- Determination of step sizes, parallelism and flatness of measuring faces of setting rod by direct method using Digital Linear Height Master.

Conclusion :

- Uncertainty of calibration at 95.45 % Confidence level and Coverage Factor $K = 2 : \pm 10\mu\text{m}$
- The Reported Results are valid only for the conditions of the received Instruments /gauges at the time of and under the stated conditions of the calibration

Calibrated By

Nisarga A
(Calibration Engineer)

Checked By

P.Santhosh kumar
(Lab In-Charge)

