



COMMITTED TO THE  
CUSTOMER SINCE - 1996

# Vaidyanatheshwara INSTRUMENTS

## CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Date of Issue: 20 - 03 - 2023

Sheet: 1 OF 1

Format No. : VI-FRM-ME-041		ULR No.: CC247323100010601F		Report No: VI/22-23/8525-19	
Customer Name and Address		M/s. MAG ENGINEERING Unit A, 46A, 3 <sup>rd</sup> Main, 2 <sup>nd</sup> Phase Peenya, Bangalore, Karnataka - 560058.			
Customer Ref. No. and Date		DC NO: SIA/RGP21-22/0365 & 13-03-2023		Received Condition	Satisfactory
SRF NO		8525		Date of Receipt	14-03-2023
CALIBRATED INSTRUMENT / EQUIPMENT DETAILS					
Nomenclature	Thread Plug Gauge	Make	HIP		
Size	M10-6H	Customer Instrument Reference No.	M023		
Calibration Done At	VI Mechanical Lab	Temperature / Humidity	20.0 - 20.2°C	52% - 54% (RH)	
Calibrated on	20 - 03 - 2023	Calibration due on	19 - 03 - 2024		
Discipline	Mechanical (Dimensional)				
MASTER EQUIPMENT TRACEABILITY DETAILS					
Nomenclature	Make	Range / Resolution	ID No.	Certificate No.	Traceable To
Universal Length Machine	Octagon	Upto300 mm / 0.1 µm	VI-ME-015	VI/22-23/INT-ME-171-01	VI-Bangalore
CALIBRATION PROCEDURE					
Used Wire Size: 0.895 mm			Measuring Method: Three Wire Method		
Standard Used: Metric threads IS 4218-2001 & 2334-2001			SOP NO: - 16-37		
The Three Wire Method of Measuring Pitch Diameter					
Gauge Nominal Values AS PER IS STANDARD					
Gauge Nominal value	Go Side (mm)		No-Go Side (mm)		
Major Diameter Max Value	10.0230		9.5225		
Major Diameter Min Value	10.0010		9.5005		
Effective Diameter Max Value	9.0435		9.2170		
Effective Diameter Min Value	9.0325		9.2060		
Effective Diameter Wear Limit	9.0205		9.2000		
Minor Diameter Max Value	8.1595		8.1595		

### Calibration Results of Go side

Sl.no	Effective dia (mm)	Deviation (µm)
01	9.0209	-11.6
02	9.0212	-11.3

### Calibration Results of No-Go side

Sl.no	Effective dia (mm)	Deviation (µm)
01	9.2068	-
02	9.2065	-

### Conclusion:

Uncertainty of calibration at 95.45% Confidence level and Coverage Factor K = 2:  $\pm 3.2 \mu\text{m}$   
The Reported Results are valid only for the conditions of the received Instruments/gauges at the time of and under the stated conditions of the calibration.

Calibrated By

Syed Sikander  
(Calibration Engineer)

Checked By

P. Santhosh Kumar  
(Lab-In Charge)

Authorized Signatory

