
VisionX Pro | Imaging evidence


RESEARCH OUTPUT - NOT A CLINICAL REPORT

2026-10-16T23:36:38.588084+00:00

A pretrained chest X-ray model evaluated the input and restored images. Scores are uncalibrated sigmoid outputs, not disease probabilities. No diagnosis, treatment, or triage recommendation is made.



Input



Restored: nlm

Finding responses

Finding	Input score	Restored score	Change
Lung Opacity	0.8028	0.2284	-0.5744
Edema	0.5279	0.0071	-0.5207
Pneumothorax	0.4588	0.0092	-0.4497
Effusion	0.4165	0.0933	-0.3232
Cardiomegaly	0.3850	0.2014	-0.1836
Atelectasis	0.2865	0.0979	-0.1886
Fracture	0.2694	0.0293	-0.2402
Lung Lesion	0.1866	0.0257	-0.1609
Consolidation	0.1772	0.0240	-0.1533
Pneumonia	0.1757	0.0076	-0.1681
Enlarged Cardiomediastinum	0.0798	0.0547	-0.0251
Infiltration	0.0734	0.1073	+0.0338
Nodule	0.0674	0.0398	-0.0276
Fibrosis	0.0350	0.0961	+0.0611
Mass	0.0212	0.0192	-0.0020
Pleural Thickening	0.0099	0.0521	+0.0423
Emphysema	0.0030	0.0145	+0.0115
Hernia	0.0017	0.0021	+0.0004

Provenance & interpretation

Model: torchxrayvision/densenet121-res224-all

Restoration: nlm

Compute device: cpu

Restoration time: 0.070 seconds


Paired classification time: 0.428 seconds

Largest score shift: Lung Opacity: 0.5744

Synthetic noise sigma: 0.05

Working pixel SHA-256: 05ad969342cb2e0d334f21c75a70bc9511529bfd1a3025149df2202445604e0d

Class activation: Lung Opacity



Left: input; right: restored. Shared display scale. Maps are derived from 7 x 7 features on the 224 x 224 center crop. They are class contributions, not lesion boundaries or proof of disease.

Controlled experiment: input SSIM 0.5374; restored SSIM 0.9083. Mean score drift from reference: input 0.1777; restored 0.0146. Similarity and score consistency do not establish clinical benefit.

Research prototype; not a clinical diagnosis or triage system.

No local training, clinical validation, or decision threshold calibration.

CAMs are coarse class-contribution maps, not lesion segmentations.

Public demo cases are not held-out test data; no accuracy is inferred.

Sources: mlmed.org/torchxrayvision | arxiv.org/abs/1512.04150 | github.com/JingyunLiang/SwinIR. Full configuration, model hashes and any sensitivity results are included in the JSON evidence export.