


OCT Fiber Scanner
Endoscope

PROBE
A

2016 Sergio Vilches

Length [cm]	104
Resonant Frequency [Hz]	765
Coupling Efficiency	0,613
Backreflectivity [ppm]	43
Max Field [mm]	1,2




OCT Fiber Scanner
Endoscope

PROBE
B

2016 Sergio Vilches

Length [cm]	102
Resonant Frequency [Hz]	842
Coupling Efficiency	0,58
Backreflectivity [ppm]	180

CH3 Shorted to GND!!!



OCT Fiber Scanner
Endoscope

PROBE
C

2016 Sergio Vilches

Length [cm]	104
Resonant Frequency [Hz]	744
Coupling Efficiency	0,53
Backreflectivity [ppm]	111