






I'm a Certified Prosthetist





I'm a Patient

3:12

Connected

Device Info

Battery 37%

Medium

Battery level is moderate. Consider charging your knee as the battery level decreases.

Right

Steps 2K

Beep On/Off

Modes

Current Standard

Quick Access

3:12

Devices

Device name: BR4B-7FCB4

Connected 37%

SN: BCBR4B094D2400001

MAC address: 5894B207FCB4

Firmware version: 0.0.8

Click to update

First pairing time: 2024-05-10

Steps: 2320 steps

Right

3:12

Devices

Device name: BR4B-7FCB4

Connected 37%

Upgrade

Connected 37%

A new firmware version v1.0.2 has been detected. Update to the latest firmware to get all the new features and improvements.

Please ensure that the battery level is above 20%, and keep Kneuro powered on and connected to your app until the update completes. Interrupting the update may cause device malfunction.

Later Update

2:54

Firmware update

Device name: BR4B-7FCB4

Connected 37%

SN: BCBR4B094D2400001

Firmware version: 0.0.8

Upgrading firmware 4%

Cancel