

Fluke Biomedical Ansur

Test and Inspection Procedure

Copyright © 2000 - 2012 Fluke Biomedical

Registro test

TEST CORRECTO

Test realizado		Componentes utilizados del Ansur	
Fecha:	05/9/2024	Ansur	Versión 2.9.7
Registro:	IEC 62353.mtr	'Plug-In': ESA620	Versión 1.1.11
Plantilla:	IEC 62353 Alternative Leakage - CL1.mtt		

Configuración test

Selecciones

Pruebas de servicio realizadas	Normas realizadas
	IEC 62353 (CL1)

Equipo bajo test

Serial number	8019771	Type	Electrocardiografo
Appliance code	.	Model	AT-101
Group	Swiss Medical	Location	Centro Medico Barrio Parque
Status	Montado	Address 1	Consultorio 307, 1piso 3
Manufacturer	Schiller	Address 2	.

Datos MTI

Instrumento de test	No de serie	Versión Firmware
ESA 620	052	v2.10

Firmas

Resultado test

Elemento de test		Tipo de test		Fall o	
IEC 62353 Alternative Leakage - Class I		Auto Sequence			
Module setup					
#	Module info		Class		Leads
2	Module code Serial number Type	R	CF		10
3	Module code Serial number Type	F	CF		10
4	Module code Serial number Type	L	CF		10
5	Module code Serial number Type	N	CF		10
6	Module code Serial number Type	C1	CF		10
7	Module code Serial number Type	C2	CF		10
8	Module code Serial number Type	C3	CF		10
9	Module code Serial number Type	C4	CF		10
10	Module code Serial number Type	C5	CF		10
11	Module code Serial number Type	C6	CF		10
Mains Voltage					
Mains Voltage		Mains Voltage			
Live to Neutral		Mains Voltage Live to Neutral			
Resultado:	Valor	Unidad	Lím.superio r Lim.inferior	Standard	
Live to Neutral	224,3	V		IEC 62353 (CL1)	
Neutral to Earth		Mains Voltage Neutral to Earth			
Resultado:	Valor	Unidad	Lím.superio r Lim.inferior	Standard	
Neutral to Earth	128,0	V		IEC 62353 (CL1)	
Live to Earth		Mains Voltage Live to Earth			
Resultado:	Valor	Unidad	Lím.superio r Lim.inferior	Standard	
Live to Earth	87,7	V		IEC 62353 (CL1)	

Elemento de test		Tipo de test				Fallo
Protective Earth Resistance		Protective Earth Resistance				o
Configuración:						
Test Current: 200 mA						
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
PE Resistance1	,300	Ohm	r 0,3		IEC 62353 (CL1)	
Patient Leakage Current		Patient Leakage Current				
Configuración:						
Total Leakage: No						
Applied Parts: open						
Normal Condition		Patient Leakage Current Normal Condition				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
R	0,0	uAAC+DC	r 10		IEC 62353 (CL1)	
F	0,0	uAAC+DC	10		IEC 62353 (CL1)	
L	0,0	uAAC+DC	10		IEC 62353 (CL1)	
N	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C1	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C2	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C3	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C4	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C5	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C6	0,0	uAAC+DC	10		IEC 62353 (CL1)	
Open Neutral		Patient Leakage Current Open Neutral				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
R	0,0	uAAC+DC	r 10		IEC 62353 (CL1)	
F	0,0	uAAC+DC	10		IEC 62353 (CL1)	
L	0,0	uAAC+DC	10		IEC 62353 (CL1)	
N	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C1	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C2	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C3	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C4	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C5	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C6	0,0	uAAC+DC	10		IEC 62353 (CL1)	
Open Earth		Patient Leakage Current Open Earth				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
R	0,1	uAAC+DC	r 10		IEC 62353 (CL1)	
F	0,1	uAAC+DC	10		IEC 62353 (CL1)	
L	0,1	uAAC+DC	10		IEC 62353 (CL1)	
N	0,1	uAAC+DC	10		IEC 62353 (CL1)	
C1	0,1	uAAC+DC	10		IEC 62353 (CL1)	
C2	0,1	uAAC+DC	10		IEC 62353 (CL1)	
C3	0,1	uAAC+DC	10		IEC 62353 (CL1)	
C4	0,1	uAAC+DC	10		IEC 62353 (CL1)	
C5	0,1	uAAC+DC	10		IEC 62353 (CL1)	
C6	0,1	uAAC+DC	10		IEC 62353 (CL1)	
Normal Condition, Reversed Mains		Patient Leakage Current Normal Condition, Reversed Mains				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
R	0,0	uAAC+DC	r 10		IEC 62353 (CL1)	
F	0,0	uAAC+DC	10		IEC 62353 (CL1)	
L	0,0	uAAC+DC	10		IEC 62353 (CL1)	
N	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C1	0,0	uAAC+DC	10		IEC 62353 (CL1)	
C2	0,0	uAAC+DC	10		IEC 62353 (CL1)	

Elemento de test		Tipo de test				Fall o
Resultado:		Valor	Unidad	Lim.superio r	Lim.inferior	Standard
C3		0,0	uAAC+DC	10		IEC 62353 (CL1)
C4		0,0	uAAC+DC	10		IEC 62353 (CL1)
C5		0,0	uAAC+DC	10		IEC 62353 (CL1)
C6		0,0	uAAC+DC	10		IEC 62353 (CL1)
Open Neutral, Reversed Mains		Patient Leakage Current Open Neutral, Reversed Mains				
Resultado:		Valor	Unidad	Lim.superio r	Lim.inferior	Standard
R		0,0	uAAC+DC			IEC 62353 (CL1)
F		0,0	uAAC+DC			IEC 62353 (CL1)
L		0,0	uAAC+DC			IEC 62353 (CL1)
N		0,0	uAAC+DC			IEC 62353 (CL1)
C1		0,0	uAAC+DC			IEC 62353 (CL1)
C2		0,0	uAAC+DC			IEC 62353 (CL1)
C3		0,0	uAAC+DC			IEC 62353 (CL1)
C4		0,0	uAAC+DC			IEC 62353 (CL1)
C5		0,0	uAAC+DC			IEC 62353 (CL1)
C6		0,0	uAAC+DC			IEC 62353 (CL1)
Open Earth, Reversed Mains		Patient Leakage Current Open Earth, Reversed Mains				
Resultado:		Valor	Unidad	Lim.superio r	Lim.inferior	Standard
R		0,0	uAAC+DC			IEC 62353 (CL1)
F		0,0	uAAC+DC			IEC 62353 (CL1)
L		0,0	uAAC+DC			IEC 62353 (CL1)
N		0,0	uAAC+DC			IEC 62353 (CL1)
C1		0,0	uAAC+DC			IEC 62353 (CL1)
C2		0,0	uAAC+DC			IEC 62353 (CL1)
C3		0,0	uAAC+DC			IEC 62353 (CL1)
C4		0,0	uAAC+DC			IEC 62353 (CL1)
C5		0,0	uAAC+DC			IEC 62353 (CL1)
C6		0,0	uAAC+DC			IEC 62353 (CL1)
Earth Leakage Current Configuración: Applied Parts: open		Earth Leakage Current				
Normal Condition		Earth Leakage Current Normal Condition				
Resultado:		Valor	Unidad	Lim.superio r	Lim.inferior	Standard
Normal Condition		3,2	uAAC+DC			IEC 62353 (CL1)
Open Neutral		Earth Leakage Current Open Neutral				
Resultado:		Valor	Unidad	Lim.superio r	Lim.inferior	Standard
Open Neutral		14,3	uAAC+DC			IEC 62353 (CL1)
Normal Condition, Reversed Mains		Earth Leakage Current Normal Condition, Reversed Mains				
Resultado:		Valor	Unidad	Lim.superio r	Lim.inferior	Standard
Normal Condition, Reversed Mains		2,7	uAAC+DC			IEC 62353 (CL1)
Open Neutral, Reversed Mains		Earth Leakage Current Open Neutral, Reversed Mains				
Resultado:		Valor	Unidad	Lim.superio r	Lim.inferior	Standard
Open Neutral, Reversed Mains		14,3	uAAC+DC			IEC 62353 (CL1)

Elemento de test		Tipo de test				Fall o
Enclosure Leakage Current <i>Configuración:</i> Applied Parts: open		Enclosure Leakage Current				
Normal Condition	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Normal Condition	0,2	uAAC+DC			IEC 62353 (CL1)
Open Neutral		Enclosure Leakage Current Open Neutral				
Open Neutral	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Open Neutral	0,2	uAAC+DC			IEC 62353 (CL1)
Open Earth		Enclosure Leakage Current Open Earth				
Open Earth	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Open Earth	3,4	uAAC+DC			IEC 62353 (CL1)
Normal Condition, Reversed Mains		Enclosure Leakage Current Normal Condition, Reversed Mains				
Normal Condition, Reversed Mains	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Normal Condition, Reversed Mains	0,2	uAAC+DC			IEC 62353 (CL1)
Open Neutral, Reversed Mains		Enclosure Leakage Current Open Neutral, Reversed Mains				
Open Neutral, Reversed Mains	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Open Neutral, Reversed Mains		uAAC+DC			IEC 62353 (CL1)
Open Earth, Reversed Mains		Enclosure Leakage Current Open Earth, Reversed Mains				
Open Earth, Reversed Mains	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Open Earth, Reversed Mains	2,8	uAAC+DC			IEC 62353 (CL1)
Alternative Equipment Leakage		Alternative Equipment Leakage				
Closed Earth		Alternative Equipment Leakage Closed Earth				
Closed Earth	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Closed Earth	28,7	uAAC	1000		IEC 62353 (CL1)
Open Earth		Alternative Equipment Leakage Open Earth				
Open Earth	<i>Resultado:</i>	<i>Valor</i>	<i>Unidad</i>	<i>Lím.superio r</i>	<i>Lim.inferior</i>	<i>Standard</i>
	Open Earth	28,7	uAAC	1000		IEC 62353 (CL1)
Alternative Applied Part Leakage		Alternative Applied Part Leakage				

Elemento de test		Tipo de test				Fall o
Shorted Mains		Alternative Applied Part Leakage Normal Condition				
Resultado:	Valor	Unidad	Lím.superio r	Lim.inferior	Standard	
R	0,7	uAAC	50		IEC 62353 (CL1)	
F	0,7	uAAC	50		IEC 62353 (CL1)	
L	0,7	uAAC	50		IEC 62353 (CL1)	
N	0,7	uAAC	50		IEC 62353 (CL1)	
C1	0,7	uAAC	50		IEC 62353 (CL1)	
C2	0,7	uAAC	50		IEC 62353 (CL1)	
C3	0,7	uAAC	50		IEC 62353 (CL1)	
C4	0,7	uAAC	50		IEC 62353 (CL1)	
C5	0,7	uAAC	50		IEC 62353 (CL1)	
C6	0,7	uAAC	50		IEC 62353 (CL1)	
Insulation Resistance		Insulation Resistance				
Configuración: Test Voltage: 500V						
Mains to Protective Earth		Insulation Resistance Mains to Protective Earth				
Resultado:	Valor	Unidad	Lím.superio r	Lim.inferior	Standard	
Mains to Protective Earth	52,1	MOhm			IEC 62353 (CL1)	
Applied Parts to Protective Earth		Insulation Resistance Applied Parts to Protective Earth				
Resultado:	Valor	Unidad	Lím.superio r	Lim.inferior	Standard	
Applied Parts to Protective Earth	99999,0	MOhm				
Mains to Applied Parts		Insulation Resistance Mains to Applied Parts				
Resultado:		Unidad	Lím.superio r	Lim.inferior	Standard	
Mains to Applied Parts		MOhm			IEC 62353 (CL1)	
Differential Leakage		Differential Leakage				
Normal Condition		Differential Leakage Normal Condition				
Resultado:	Valor	Unidad	Lím.superio r	Lim.inferior	Standard	
Normal Condition	0,0	uA	500		IEC 62353 (CL1)	
Open Earth		Differential Leakage Open Earth				
Resultado:	Valor	Unidad	Lím.superio r	Lim.inferior	Standard	
Open Earth	0,0	uA	500		IEC 62353 (CL1)	
Normal Condition, Reversed Mains		Differential Leakage Normal Condition, Reversed Mains				
Resultado:	Valor	Unidad	Lím.superio r	Lim.inferior	Standard	
Normal Condition, Reversed Mains	0,0	uA	500		IEC 62353 (CL1)	
Open Earth, Reversed Mains		Differential Leakage Open Earth, Reversed Mains				

Elemento de test		Tipo de test				Fall o
Resultado:	Valor	Unidad	Lím.superio r	Lim.inferior	Standard	
Open Earth, Reversed Mains	0,0	uA	500		IEC 62353 (CL1)	

