

Fluke Biomedical Ansur

Test and Inspection Procedure

Copyright © 2000 - 2012 Fluke Biomedical

Registro test

TEST CORRECTO

Test realizado		Componentes utilizados del Ansur	
Fecha:	12/3/2025	Ansur	Versión 2.9.7
Registro:	IEC 62353 Alternative Leakage - CL1.mtr	'Plug-In': ESA620	Versión 1.1.11
Plantilla:	IEC 62353 Alternative Leakage - CL1.mtt		

Configuración test

Selecciones

Pruebas de servicio realizadas	Normas realizadas
	IEC 62353 (CL1)

Equipo bajo test

Serial number	53602	Type	Centrifuga
Appliance code		Model	LC - 25
Group	Laboratorio	Location	Catamarca
Status	MT	Address 1	2do Piso
Manufacturer	Luguimac	Address 2	

Datos MTI

Instrumento de test	Número de serie	Versión Firmware
ESA 620	2982052	v2.10

Firmas

Resultado test

Elemento de test		Tipo de test			Fall o
IEC 62353 Alternative Leakage - Class I		Auto Sequence			
Mains Voltage		Mains Voltage			
Live to Neutral	Mains Voltage Live to Neutral	Mains Voltage Live to Neutral			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard				
	Live to Neutral 223,9 V IEC 62353 (CL1)				
Neutral to Earth	Mains Voltage Neutral to Earth	Mains Voltage Neutral to Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard				
	Neutral to Earth 224,2 V IEC 62353 (CL1)				
Live to Earth	Mains Voltage Live to Earth	Mains Voltage Live to Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard				
	Live to Earth 0,5 V IEC 62353 (CL1)				
Protective Earth Resistance		Protective Earth Resistance			
Configuración: Test Current: 200 mA					
PE Resistance1	Resultado: Valor Unidad Lim.superior Lim.inferior Standard				
	PE Resistance1 0,208 Ohm 0,3 IEC 62353 (CL1)				
Alternative Equipment Leakage		Alternative Equipment Leakage			
Closed Earth	Alternative Equipment Leakage Closed Earth	Alternative Equipment Leakage Closed Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard				
	Closed Earth 99,4 uAAC 1000 IEC 62353 (CL1)				
Open Earth	Alternative Equipment Leakage Open Earth	Alternative Equipment Leakage Open Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard				
	Open Earth 99,1 uAAC 1000 IEC 62353 (CL1)				
Insulation Resistance		Insulation Resistance			
Configuración: Test Voltage: 500V					
Mains to Protective Earth	Insulation Resistance Mains to Protective Earth	Insulation Resistance Mains to Protective Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard				
	Mains to Protective Earth 99999,0 MOhm IEC 62353 (CL1)				