

Realizado por CMA - ING. CLINICA
Melina Montalto - Enrique Despaigne

Registro test

TEST CORRECTO

Test realizado

Fecha: 17/8/2023
Registro: IEC 62353 Alternative
Leakage - CL1.mtr
Plantilla: IEC 62353 Alternative
Leakage - CL1.mtt

Componentes utilizados del Ansur

Ansur Versión 2.9.7
'Plug-In': ESA620 Versión 1.1.11

Configuración test

Selecciones

Pruebas de servicio realizadas

Normas realizadas

IEC 62353 (CL1)

Equipo bajo test

Serial number	08019771	Type	Electrocardiografo
Appliance code		Model	AT-101
Group	Electrocardiografia	Location	Barrio Parque
Status	MT	Address 1	2 do piso
Manufacturer	schiller	Address 2	

Datos MTI

Instrumento de test	Número de serie	Versión Firmware
ESA 620	2982052	v2.10

Firmas

Resultado test

Elemento de test		Tipo de test			Fall o
IEC 62353 Alternative Leakage - Class I		Auto Sequence			
Mains Voltage		Mains Voltage			
Live to Neutral	Mains Voltage Live to Neutral	Mains Voltage Live to Neutral			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard	r			
	Live to Neutral 222,7 V IEC 62353 (CL1)				
Neutral to Earth	Mains Voltage Neutral to Earth	Mains Voltage Neutral to Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard	r			
	Neutral to Earth 0,6 V IEC 62353 (CL1)				
Live to Earth	Mains Voltage Live to Earth	Mains Voltage Live to Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard	r			
	Live to Earth 223,4 V IEC 62353 (CL1)				
Protective Earth Resistance		Protective Earth Resistance			
Configuración: Test Current: 200 mA					
PE Resistance1	Resultado: Valor Unidad Lim.superior Lim.inferior Standard	r			
	0,279 Ohm 0,3 IEC 62353 (CL1)				
Alternative Equipment Leakage		Alternative Equipment Leakage			
Closed Earth	Alternative Equipment Leakage Closed Earth	Alternative Equipment Leakage Closed Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard	r			
	Closed Earth 37,7 uAAC 1000 IEC 62353 (CL1)				
Open Earth	Alternative Equipment Leakage Open Earth	Alternative Equipment Leakage Open Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard	r			
	Open Earth 37,7 uAAC 1000 IEC 62353 (CL1)				
Insulation Resistance		Insulation Resistance			
Configuración: Test Voltage: 500V					
Mains to Protective Earth	Insulation Resistance Mains to Protective Earth	Insulation Resistance Mains to Protective Earth			
	Resultado: Valor Unidad Lim.superior Lim.inferior Standard	r			
	Mains to Protective Earth 99999,0 MOhm IEC 62353 (CL1)				