

Fluke Biomedical Ansur

Test and Inspection Procedure

Copyright © 2000 - 2012 Fluke Biomedical

Registro test

TEST CORRECTO

Test realizado

Fecha: 12/3/2025
Registro: IEC 62353 Alternative Leakage - CL1.mtr
Plantilla: IEC 62353 Alternative Leakage - CL1.mtt

Componentes utilizados del Ansur

Ansur Versión 2.9.7
'Plug-In': ESA620 Versión 1.1.11

Configuración test

Selecciones

Pruebas de servicio realizadas	Normas realizadas
	IEC 62353 (CL1)

Equipo bajo test

Serial number	24084	Type	Centrifuga
Appliance code		Model	DCA 300 RVT
Group	Laboratorio	Location	Catamarca
Status	MT	Address 1	2do Piso
Manufacturer	Presvac	Address 2	

Datos MTI

Instrumento de test	Número de serie	Versión Firmware
ESA 620	2982052	v2.10

Firmas

Resultado test

Elemento de test			Tipo de test			Fall o
IEC 62353 Alternative Leakage - Class I			Auto Sequence			
Mains Voltage			Mains Voltage			
Live to Neutral			Mains Voltage Live to Neutral			
Resultado:	Valor	Unidad	Lim.superior r	Lim.inferior	Standard	
Live to Neutral	224,2	V			IEC 62353 (CL1)	
Neutral to Earth			Mains Voltage Neutral to Earth			
Resultado:	Valor	Unidad	Lim.superior r	Lim.inferior	Standard	
Neutral to Earth	225,3	V			IEC 62353 (CL1)	
Live to Earth			Mains Voltage Live to Earth			
Resultado:	Valor	Unidad	Lim.superior r	Lim.inferior	Standard	
Live to Earth	0,7	V			IEC 62353 (CL1)	
Protective Earth Resistance			Protective Earth Resistance			
Configuración: Test Current: 200 mA						
Resultado:	Valor	Unidad	Lim.superior r	Lim.inferior	Standard	
PE Resistance1	0,261	Ohm	0,3		IEC 62353 (CL1)	
Alternative Equipment Leakage			Alternative Equipment Leakage			
Closed Earth			Alternative Equipment Leakage Closed Earth			
Resultado:	Valor	Unidad	Lim.superior r	Lim.inferior	Standard	
Closed Earth	641,0	uAAC	1000		IEC 62353 (CL1)	
Open Earth			Alternative Equipment Leakage Open Earth			
Resultado:	Valor	Unidad	Lim.superior r	Lim.inferior	Standard	
Open Earth	645,0	uAAC	1000		IEC 62353 (CL1)	
Insulation Resistance			Insulation Resistance			
Configuración: Test Voltage: 500V						
Mains to Protective Earth			Insulation Resistance Mains to Protective Earth			
Resultado:	Valor	Unidad	Lim.superior r	Lim.inferior	Standard	
Mains to Protective Earth	99999,0	MOhm			IEC 62353 (CL1)	