

Fluke Biomedical Ansur

Test and Inspection Procedure

Copyright © 2000 - 2012 Fluke Biomedical

Registro test

TEST CORRECTO

Test realizado		Componentes utilizados del Ansur	
Fecha:	12/3/2025	Ansur	Versión 2.9.7
Registro:	IEC 62353.mtr	'Plug-In': ESA620	Versión 1.1.11
Plantilla:	IEC 62353 Alternative Leakage - CL1.mtt		

Configuración test

Selecciones

Pruebas de servicio realizadas	Normas realizadas
	IEC 62353 (CL1)

Equipo bajo test

Serial number	26355	Type	Baño termico
Appliance code	.	Model	Masson
Group	CMA	Location	Catamarca
Status	MT	Address 1	Piso 2
Manufacturer	Vicking	Address 2	.

Datos MTI

Instrumento de test	Número de serie	Versión Firmware
ESA 620	2982052	v2.10

Firmas

Resultado test

Elemento de test		Tipo de test			Fall o
IEC 62353 Alternative Leakage - Class I		Auto Sequence			
Mains Voltage		Mains Voltage			
Live to Neutral		Mains Voltage Live to Neutral			
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard
Live to Neutral	224,2	V			IEC 62353 (CL1)
Neutral to Earth		Mains Voltage Neutral to Earth			
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard
Neutral to Earth	224,4	V			IEC 62353 (CL1)
Live to Earth		Mains Voltage Live to Earth			
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard
Live to Earth	0,4	V			IEC 62353 (CL1)
Protective Earth Resistance		Protective Earth Resistance			
Configuración: Test Current: 200 mA					
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard
PE Resistance1	0,185	Ohm	0,3		IEC 62353 (CL1)
Earth Leakage Current		Earth Leakage Current			
Configuración: Applied Parts: open					
Normal Condition		Earth Leakage Current Normal Condition			
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard
Normal Condition	15,6	uAAC+DC			IEC 62353 (CL1)
Open Neutral		Earth Leakage Current Open Neutral			
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard
Open Neutral	27,7	uAAC+DC			IEC 62353 (CL1)
Normal Condition, Reversed Mains		Earth Leakage Current Normal Condition, Reversed Mains			
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard
Normal Condition, Reversed Mains	7,9	uAAC+DC			IEC 62353 (CL1)
Open Neutral, Reversed Mains		Earth Leakage Current Open Neutral, Reversed Mains			

Elemento de test		Tipo de test				Fall o
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Open Neutral, Reversed Mains	22,8	uAAC+DC			IEC 62353 (CL1)	
Enclosure Leakage Current		Enclosure Leakage Current				
Configuración: Applied Parts: open						
Normal Condition		Enclosure Leakage Current Normal Condition				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Normal Condition	0,2	uAAC+DC			IEC 62353 (CL1)	
Open Neutral		Enclosure Leakage Current Open Neutral				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Open Neutral	0,2	uAAC+DC			IEC 62353 (CL1)	
Open Earth		Enclosure Leakage Current Open Earth				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Open Earth	15,7	uAAC+DC			IEC 62353 (CL1)	
Normal Condition, Reversed Mains		Enclosure Leakage Current Normal Condition, Reversed Mains				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Normal Condition, Reversed Mains	0,2	uAAC+DC			IEC 62353 (CL1)	
Open Neutral, Reversed Mains		Enclosure Leakage Current Open Neutral, Reversed Mains				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Open Neutral, Reversed Mains	0,2	uAAC+DC			IEC 62353 (CL1)	
Open Earth, Reversed Mains		Enclosure Leakage Current Open Earth, Reversed Mains				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Open Earth, Reversed Mains	7,9	uAAC+DC			IEC 62353 (CL1)	
Alternative Equipment Leakage		Alternative Equipment Leakage				
Closed Earth		Alternative Equipment Leakage Closed Earth				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Closed Earth	35,3	uAAC	1000		IEC 62353 (CL1)	
Open Earth		Alternative Equipment Leakage Open Earth				
Resultado:	Valor	Unidad	Lím.superior	Lim.inferior	Standard	
Open Earth	34,8	uAAC	1000		IEC 62353 (CL1)	

Elemento de test		Tipo de test			Fall o
Insulation Resistance <i>Configuración:</i> Test Voltage: 500V		Insulation Resistance			
Mains to Protective Earth <i>Resultado:</i>		Insulation Resistance Mains to Protective Earth			
	Valor	Unidad	Lim.superior	Lim.inferior	Standard
Mains to Protective Earth	99999,0	MOhm			IEC 62353 (CL1)
Differential Leakage		Differential Leakage			
Normal Condition <i>Resultado:</i>		Differential Leakage Normal Condition			
	Valor	Unidad	Lim.superior	Lim.inferior	Standard
Normal Condition	55,0	uA	500		IEC 62353 (CL1)
Open Earth <i>Resultado:</i>		Differential Leakage Open Earth			
	Valor	Unidad	Lim.superior	Lim.inferior	Standard
Open Earth	70,0	uA	500		IEC 62353 (CL1)
Normal Condition, Reversed Mains <i>Resultado:</i>		Differential Leakage Normal Condition, Reversed Mains			
	Valor	Unidad	Lim.superior	Lim.inferior	Standard
Normal Condition, Reversed Mains	17,0	uA	500		IEC 62353 (CL1)
Open Earth, Reversed Mains <i>Resultado:</i>		Differential Leakage Open Earth, Reversed Mains			
	Valor	Unidad	Lim.superior	Lim.inferior	Standard
Open Earth, Reversed Mains	19,0	uA	500		IEC 62353 (CL1)