


## Instrument Calibration Certificate

Model : ToxiRae 3  
Customer : J. de Jonge flowsystems B.V.  
Serial Number : G011256875 / 53H2SM021  
Date of Calibration : 28-03-2019  
Calibration due : 24-09-2019  
  
Calibration Tank : CG-Q58-5      H2S

GAS	Due Date	Span	LOW	HIGH	TWA	STEL
H2S	180days	25 ppm	5ppm	5ppm	1.6ppm	3.2ppm

Signature Controller: P. van der Jagt



This instrument has been factory inspected, tested and calibrated  
in accordance with the and requirements of our registered Quality System ISO 9001,  
Operating Standards and Sales' Agreements.  
All calibration gases are traceable to a known national standard.  
Refer to your product user manual to determine recommended calibration frequency.

certificate