

Calibration Test Certificate

2019-03-27 18:09:47

Device

Serial Number: **KA419-1010421**
Next Cal Due: **2019-09-23**
Manufacturer: **BW Technologies**
Detector Type: **GasAlertMicroClip**
Device User: **J. de Jonge Lease BV / 53VGDTM092**

Test Result **Pass**

Sensors

Type:	H2S	CO	LEL	O2
Result:	Pass	Pass	Pass	Pass
Final Reading:				
Next Calibration Due:	2019-09-23	2019-09-23	2019-09-23	2019-09-23

Set Points

Type:	H2S	CO	LEL	O2
High Alarm:	10.0 ppm	50.0 ppm	20.0 %LEL	23.5 %vol
Low Alarm:	5.0 ppm	25.0 ppm	10.0 %LEL	19.5 %vol
TWA Alarm:	1.6 ppm	25.0 ppm		
STEL Alarm:	5.0 ppm	50.0 ppm		

Options

Datalog Interval:	15 seconds	Unit Programmed:	N/A
H2S STEL Period:	15 minutes	CO STEL Period:	15 minutes

Test Station

Dock Serial Number:	5345DOX01161200027	Dock Location:	Dock 2
Inlet 1:		Inlet 2:	
Used:	Yes	No	No
Concentration:	50.0 %LEL, 18.0 %vol, 25.0 ppm, 100.0 ppm	20.9 %vol	
Type:	CH4, O2, H2S, CO	O2	
Notes:	---	---	

Test Notes:

Test Notes: