

Introduction to Diagnostic x-ray Imaging (Part I)

Yulei Jiang

Department of Radiology, The University of Chicago

Outline

- **Brief historical survey**
- **Linear-system model**
- **Focal spot and magnification**
- **Scatter**
- **Image detector**

A Brief History of Diagnostic x-ray Imaging

- **1895: German physicist Wilhelm Conrad Roentgen discovered x-ray, and produced the first x-ray picture of the human body—his wife's hand.**


The First Medical Image



Source: Wikipedia


Early 20th Century

- **1900: widespread use of chest x-ray made early detection of tuberculosis a reality.**
- **1906: x-ray contrast medium use began.**
- **1912: Marie Curie published theory of radioactivity; investigation of x-ray radiation for patient therapy (e.g. treatment of cancer).**
- **1913: William Coolidge invented the hot cathode x-ray tube**




Source: Wikipedia

Coolidge Hot Cathode x-ray Tube



Source: Wikipedia

Rotating Anode x-ray Tube



Source: Wikipedia

Middle and Late 20th Century

- **1950: Nuclear medicine imaging begins.**
- **1955: image intensifier TV allowed dynamic x-ray imaging of beating heart and blood vessels.**
- **1960: Ultrasound imaging began.**
- **1970: x-ray mammography began.**

Shoe-fitting Fluoroscopy






CERTIFICATE

SHOE-FITTING TEST DATA FOR _____

1. ANKLE ROLL GOOD ☐ FAIR ☐ POOR ☐

2. WEIGHT DISTRIBUTION



RIGHT WAY

LEFT RIGHT
 _____% BALL _____%
 _____% OUTER _____%
 _____% HEEL _____%

WRONG WAY

3. X-RAY FITTING TEST



LEFT RIGHT
☐ GOOD ☐
☐ FAIR ☐
☐ POOR ☐

WRONG WAY

This scientific way of approaching the problem of poorly-fitted shoes eliminates guesswork. Now you can see for yourself!



Source:
 Oak
 Ridge
 Associated
 University

Late 20th Century

- **1972: Godfrey Hounsfield and Allan Cormack invented CT.**
- **1978: Digital radiography began.**
- **1980: Paul Lauterbur and others developed MRI.**
- **1985: University of California scientists developed PET.**

Common x-ray Procedures

- **Chest x-ray (TB, pneumonia, cancer, etc.)**
- **Orthopedic evaluations (bone fracture, bone age of children)**
- **Mammography (breast cancer)**
- **Abdomen (contrast fluoroscopy)**
- **Verification of surgical markers (radiation therapy verification)**
- **Dental examination**
- **Chiropractic examinations**

Medical Imaging as Applied Science

- **Imaging theory**
- **Practical aspects**
- **Physicists, radiologists, technologists, and engineers**

Outline

- Brief historical survey
- **Linear-system model**
- Focal spot and magnification
- Scatter
- Image detector


Linear-System Model of Projection (Planar) x-ray Imaging

- **Based on first principle (geometry, physics)**
- **Math not difficult but somewhat tedious**
- **Idealized in several ways**
- **Provides a way of thinking of the imaging system**
- **Was not necessarily the source of system development**

Summary of Objective

- Describe the image as a convolution of the object with the x-ray source

Simplified Representation of an x-ray System



Idealizations

- **3D source** → **2D planar source**
- **Anisotropic source** → **isotropic source**
- **3D object** → **2D planar object**

Geometry: From the Image Plane Perspective of the Source Plane

- A patch of the image plane expressed as solid angle to a patch in the source plane:

$$d\Omega = \frac{d^2 r'' \cos \theta}{R^2}$$


- x-ray photon density at the patch in the image plane from the patch in the source plane:

$$f(\vec{r}) d^2 r \frac{d\Omega}{4\pi} = f(\vec{r}) \frac{\cos^3 \theta}{4\pi (s_1 + s_2)^2} d^2 r d^2 r''$$

- Total x-ray photon density at the patch in the image plane:

$$h(\vec{r}'') d^2 r'' = \frac{T d^2 r''}{4\pi (s_1 + s_2)^2} \int_{source} d^2 r \cos^3 \theta f(\vec{r}) g(\vec{r}')$$

Simplified Representation of an x-ray System



General Description


$$h(\vec{r}'') \, d^2r'' = \frac{T \, d^2r''}{4\pi(s_1 + s_2)^2} \int_{source} d^2r \cos^3\theta \, f(\vec{r}) \, g(\vec{r}')$$

$$\text{Let: } C = \frac{T}{4\pi(s_1 + s_2)^2}$$

and consider source far from the imaging plane:


$$h(\vec{r}'') \approx C \int_{source} d^2r \, f_{source}(\vec{r}) \, g_{object}(\vec{r}')$$

Geometry: Determine the Object-Plane Vector by Line of Sight



Geometry: The Object Plane

Determination by the Line of Sight



Geometry: Line-of-Sight Determination of Object-Plane Vector

$$\frac{\vec{r}' - \vec{r}}{s_1} = \frac{\vec{r}'' - \vec{r}}{s_1 + s_2} \quad \rightarrow \quad \vec{r}' = \frac{s_1 \vec{r}'' + s_2 \vec{r}}{s_1 + s_2} = a\vec{r}'' + b\vec{r}$$


Define: $a = \frac{s_1}{s_1 + s_2} \quad b = \frac{s_2}{s_1 + s_2} = 1 - a$

Elimination of the Object-Plane Vector


$$h(\vec{r}'') \approx C \int_{\text{source}} d^2r f_{\text{source}}(\vec{r}) g_{\text{object}}(\vec{r}')$$

$$h(\vec{r}'') d^2r'' \approx C d^2r'' \int_{\text{source}} d^2r f_{\text{source}}(\vec{r}) g_{\text{object}}(a\vec{r}'' + b\vec{r})$$


Geometry: Project the Source through the Origin of the Object Plane to the Image Plane



source





object



image

$$\vec{r}_0'' = -\frac{b}{a}\vec{r}$$

Geometry: Projection of the Source through the Object Plane to the Image Plane



Geometry: Evaluation of Source and Object at Image-Plane Coordinates

$$f_{image}(\vec{r}_0'') = f_{source}(\vec{r}) = f_{source}\left(-\frac{a}{b}\vec{r}_0''\right)$$

$$g_{image}(\vec{r}_0'') = g_{object}(a\vec{r}_0'')$$

$$g_{object}(a\vec{r}'' + b\vec{r}) = g_{object}(a\vec{r}'' - a\vec{r}_0'') = g_{image}(\vec{r}'' - \vec{r}_0'')$$

Important Results

$$h(\vec{r}'') \, d^2r'' \approx C \, d^2r'' \int_{\text{source}} d^2r \, f_{\text{source}}(\vec{r}) g_{\text{object}}(a\vec{r}'' + b\vec{r})$$

$$\begin{aligned} h(\vec{r}'') &= \left(\frac{a}{b}\right)^2 C \int_{\text{image}} d^2r_0'' \, f_{\text{image}}(\vec{r}_0'') g_{\text{image}}(\vec{r}'' - \vec{r}_0'') \\ &= \left(a/b\right)^2 C f_{\text{image}}(\vec{r}'') \otimes g_{\text{image}}(\vec{r}'') \end{aligned}$$

$$\begin{aligned} \mathbf{H}_{f, \text{image}}(\vec{\rho}'') &= \left(a/b\right)^2 C \mathbf{F}_{f, \text{image}}(\vec{\rho}'') \mathbf{G}_{f, \text{image}}(\vec{\rho}'') \\ &= \left(C/a^2\right) \mathbf{F}_{f, \text{source}}(-b\vec{\rho}''/a) \mathbf{G}_{f, \text{object}}(\vec{\rho}''/a) \end{aligned}$$

Review of Assumptions

- **Source far away from the imaging plane**
- **Isotropy of x-ray source**
- **2D planar x-ray source**
- **2D planar object**

Have Not Considered

- Scatter
- Image detector
- Patient motion

Remarks

- **x-ray source is “in” every image**
- **x-ray source is an important part of the imaging chain**
- **Linear system analysis becomes possible**
- **Based on system-geometry consideration alone**

Several Forms of the Convolution Equation

$$\mathbf{H}_{f, image}(\bar{\rho}'') = (C/a^2) \mathbf{F}_{f, source}(-b\bar{\rho}''/a) \mathbf{G}_{f, object}(\bar{\rho}''/a)$$

$$\mathbf{H}_{f, image}(a\bar{\rho}') = (C/a^2) \mathbf{F}_{f, source}(-b\bar{\rho}') \mathbf{G}_{f, object}(\bar{\rho}')$$

$$\mathbf{H}_{f, image}(-a\bar{\rho}/b) = (C/a^2) \mathbf{G}_{f, object}(-\bar{\rho}/b) \mathbf{F}_{f, source}(\bar{\rho})$$

Outline

- Brief historical survey
- Linear-system model
- **Focal spot and magnification**
- Scatter
- Image detector

Focal Spot Consideration

- **Simple uniform-disk focal spot**
- **To derive the PSF of the focal spot**


Focal Spot Imaged Through a Point Object

$$f_{source}(\vec{r}) = f_0 \text{ circ}\left(\frac{2r}{d_{source}^{\text{focal spot}}}\right)$$

$$g_{object}(\vec{r}') = \delta(\vec{r}' - \vec{r}_1')$$

$$\begin{aligned} h(\vec{r}'') &= C \int_{source} d^2r f_0 \text{ circ}\left(\frac{2r}{d_{source}^{\text{focal spot}}}\right) \delta(a\vec{r}'' + b\vec{r} - \vec{r}') \\ &= \frac{C}{b^2} f_0 \text{ circ}\left(\frac{2|\vec{r}'' - \vec{r}_1'/a|}{b d_{source}^{\text{focal spot}}/a}\right) \end{aligned}$$

Geometric Consideration



$$d_{\text{focal spot image}} = \frac{b}{a} d_{\text{focal spot source}} = \frac{s_2}{s_1} d_{\text{focal spot source}}$$

Remarks

- **Consideration of this simple hypothetical focal spot helps to convince one that the image is a convolution of the object with the focal spot**
- **Note importance of spatial (and hence spatial frequency) scaling**

Total System PSF

Point object

$$\begin{aligned}\text{PSF}_{tot} &= \left(C/a^2\right) \text{Fourier}^{-1} \left\{ \mathbf{D}_{f, \text{image}} \left(a\bar{\rho}' \right) \mathbf{F}_{f, \text{source}} \left(-b\bar{\rho}' \right) \right\} \\ &= \text{PSF}_{focal \ spot} \left(\bar{\mathbf{r}}' \right) \otimes \text{PSF}_{detector} \left(\bar{\mathbf{r}}' \right)\end{aligned}$$

$$\text{PSF}_{focal \ spot} \left(\bar{\mathbf{r}}' \right) = \frac{C}{a^2} \text{Fourier}^{-1} \left\{ \mathbf{F}_{f, \text{source}} \left(-b\bar{\rho}' \right) \right\} = \frac{C}{a^2 b^2} f_{\text{source}} \left(-\frac{\bar{\mathbf{r}}'}{b} \right)$$

$$\text{PSF}_{detector} \left(\bar{\mathbf{r}}' \right) = \text{Fourier}^{-1} \left\{ \mathbf{D}_{f, \text{image}} \left(a\bar{\rho}' \right) \right\} = \frac{1}{a^2} d_{\text{detector}} \left(\frac{\bar{\mathbf{r}}'}{a} \right)$$


Optimal Magnification

$$\mathbf{F}_{f, source}(\vec{\rho}) \propto e^{-\pi \left(\frac{\rho}{\rho_{focal\ spot}} \right)^2} \quad \mathbf{D}_{f, image}(\vec{\rho}'') \propto e^{-\pi \left(\frac{\rho''}{\rho_{detector}} \right)^2}$$

$$\text{MTF}_{tot} \propto e^{-\pi \rho'^2 \left[\frac{1}{(m_t \rho_{detector})^2} + \frac{(m_t - 1)^2}{(m_t \rho_{focal\ spot})^2} \right]}$$

$$m_t^{optimal} = 1 + \left(\frac{\rho_{focal\ spot}}{\rho_{detector}} \right)^2$$

Geometric Consideration



$$d_{\text{focal spot image}} = \frac{b}{a} d_{\text{focal spot source}} = \frac{s_2}{s_1} d_{\text{focal spot source}}$$

Summary on Focal Spot and Magnification

- **Large focal spot ok for contact imaging**
- **Small focal spot required for magnification**

Outline

- Brief historical survey
- Linear-system model
- Focal spot and magnification
- **Scatter**
- Image detector

Linear System Analysis

$$\text{PSF}_{total}(\vec{r}) = \text{PSF}_{primary}(\vec{r}) + \text{PSF}_{scatter}(\vec{r})$$

Modeling of Scatter PSF

$$\text{PSF}_{\text{scatter}}(\vec{r}) =$$

$$\int_0^L dz \left[\Phi_0 e^{-\mu(L-z)} \right] \left[\frac{r_0^2}{2} (1 + \cos^2 \theta) \right] [n_e A] \left[\frac{\cos^3 \theta}{(s+z)^2} \right] \left[e^{-\mu z \sec \theta} \right]$$

$$\text{PSF}_{\text{primary}}(\vec{r}) = \Phi_0 e^{-\mu L}$$

Modeling Scatter at Gaussian


$$\begin{aligned} \text{PSF}_{scatter}(\vec{r}) &= A_s T e^{-\pi \beta_s^2 r^2} & \mathbf{P}_{scatter}(\bar{\rho}) &= \frac{A_s T}{\beta_s^2} e^{-\pi \rho^2 / \beta_s^2} \\ \text{PSF}_{primary}(\vec{r}) &= A_p T e^{-\pi \beta_p^2 r^2} & \mathbf{P}_{primary}(\bar{\rho}) &= \frac{A_p T}{\beta_p^2} e^{-\pi \rho^2 / \beta_p^2} \end{aligned}$$

$$\begin{aligned} \text{MTF}_{total} &= \frac{\mathbf{P}_{primary}(\bar{\rho}) + \mathbf{P}_{scatter}(\bar{\rho})}{\text{PSF}_{primary}(0) + \text{PSF}_{scatter}(0)} \\ &= \frac{e^{-\pi \rho^2 / \beta_p^2} + \text{SPR} e^{-\pi \rho^2 / \beta_s^2}}{1 + \text{SPR}} \end{aligned}$$

Important Results

$$\frac{\mathbf{P}_{total}(\bar{\rho})}{\mathbf{P}_{primary}(\bar{\rho})} \approx \frac{1}{1 + \text{SPR}} \quad \beta_s \ll \bar{\rho} \ll \beta_p$$

Effect of Scatter on Radiation Contrast



$$C_{NS} = \frac{\Delta\phi}{\bar{\phi}} = \frac{\phi_1 - \phi_2}{(\phi_1 + \phi_2)/2} = \frac{\phi_{1P} - \phi_{2P}}{(\phi_{1P} + \phi_{2P})/2} = \frac{\phi_{1P} - \phi_{2P}}{P}$$


$$C_S = \frac{(\phi_{1P} + \phi_{1S}) - (\phi_{2P} + \phi_{2S})}{[(\phi_{1P} + \phi_{1S}) + (\phi_{2P} + \phi_{2S})]/2}$$

$$= \frac{(\phi_{1P} - \phi_{2P}) + (\phi_{1S} - \phi_{2S})}{[(\phi_{1P} + \phi_{2P}) + (\phi_{1S} + \phi_{2S})]/2} = \frac{\phi_{1P} - \phi_{2P}}{P + S}$$

$$P = \frac{(\phi_{1P} + \phi_{2P})}{2}, S = \phi_S$$

$$\phi_{1S} \approx \phi_{2S} = \phi_S$$

Effect of Scatter on Radiation Contrast



$$C_{NS} = \frac{\Delta\phi}{\bar{\phi}} = \frac{\phi_1 - \phi_2}{(\phi_1 + \phi_2)/2} = \frac{\phi_{1p} - \phi_{2p}}{(\phi_{1p} + \phi_{2p})/2} = \frac{\phi_{1p} - \phi_{2p}}{P}$$

$$C_S = \frac{\phi_{1p} - \phi_{2p}}{P} \frac{P}{P + S} = C_{NS} \frac{P}{P + S} = C_{NS} \left(1 - \frac{S}{P + S} \right)$$

$$= C_{NS} (1 - SF) = C_{NS} \frac{1}{1 + S/P}$$


$$P = \frac{(\phi_{1p} + \phi_{2p})}{2}, S = \phi_s$$

$$\phi_{1s} \approx \phi_{2s} = \phi_s$$



Define scatter fraction
& scatter-to-primary ratio:

$$SF = \frac{S}{P + S} = \frac{S/P}{1 + S/P}$$

Example Images of Reduced Scatter




Example Images of Reduced Scatter



Anti-scatter Grid Basics

Antiscatter grid design





Grid characteristics:

t = thickness of lead strips
 h = height of lead strips
 D = distance between lead strips

$$\text{Grid Ratio} = \frac{h}{D}$$

$$\text{Grid Frequency} = \frac{1}{t + D}$$

Focal range: determined by geometry of lead strips



Summary

- **A general linear-system model that describes planar x-ray imaging**
- **Image is a convolution of the object with the focal spot**
- **Magnification important for this convolution**
- **Scatter degrades image by inserting a broad PSF into the linear-system model**