Subject-independent surface EMG electrode reduction using cross-fitted attribution: a prespecified non-inferiority study

Minh-Tri Ta[1], Hoang-Khang Phan[2,3] and Le-Giang Tran[4,*]

[1]College of Engineering and Computer Science, VinUniversity, Hanoi, Vietnam
[2]Department of Biomedical Engineering, Faculty of Applied Science, Ho Chi Minh City University of Technology (HCMUT), 268 Ly Thuong Kiet Street, District 10, Ho Chi Minh City, Vietnam
[3]Vietnam National University Ho Chi Minh City, Linh Trung Ward, Thu Duc City, Ho Chi Minh City, Vietnam
[4]Faculty of Electronics Technology, Industrial University of Ho Chi Minh City, Ho Chi Minh City, Vietnam
*Author to whom any correspondence should be addressed.

E-mail: tranlegiang@iuh.edu.vn

Keywords: surface electromyography, gesture recognition, electrode selection, SHAP, cross-fitting, non-inferiority, leave-one-subject-out

Abstract

Objective. Reducing the number of surface electromyography (sEMG) electrodes can simplify wearable interfaces, but only if recognition performance is preserved for previously unseen users. We investigated whether attribution-guided electrode selection could reduce a forearm sEMG array without meaningful loss of subject-independent gesture recognition. *Approach.* We analysed 43 GRABMyo participants using 16 monopolar forearm electrodes and 16 active gestures. All analyses followed a leave-one-subject-out design in which the test participant was excluded from feature standardisation, model fitting and electrode ranking. Electrode importance was estimated with TreeSHAP within participant-disjoint cross-fitting, and reduced models were retrained on the highest-ranked electrodes. Performance was compared with the full array against a prespecified non-inferiority margin of 0.02 Macro-F1, using participant-level cluster-bootstrap confidence intervals (CI). We also examined whether penalising attribution instability improved selection, whether attribution-based subsets outperformed random and alternative selections, and whether deletion with retraining supported the rankings. *Main results.* The full array achieved a Macro-F1 of 0.542 (95% CI 0.517–0.567). The smallest non-inferior subset contained 12 electrodes (mean difference −0.0133, 95% CI −0.0160 to −0.0106); eleven electrodes did not satisfy the criterion. Attribution-ranked subsets outperformed the mean of random subsets at 4, 6, 8 and 12 electrodes, and deletion with retraining supported the faithfulness of the ranking. Stability regularisation produced no statistically detectable improvement over attribution magnitude alone. In exploratory comparisons, no statistically detectable difference from an evenly spaced, data-free layout was observed at 4, 8 or 12 electrodes; attribution scored higher at six electrodes. *Significance.* Under offline, subject-independent evaluation of this two-ring monopolar montage, about one quarter of the electrodes could be removed within a prespecified engineering tolerance, whereas larger reductions could not. Cross-fitted attribution offered a leakage-controlled way to identify informative electrodes, but its advantage over simple spatial coverage was limited and varied with subset size.

1 Introduction

Surface electromyography (sEMG) pattern recognition provides a practical interface for multifunction myoelectric control (Hudgins et al., 1993; Englehart and Hudgins, 2003; Scheme and Englehart, 2011; Farina et al., 2014), but its performance depends partly on the number and spatial distribution of recording sites. Increasing the number of electrodes can provide richer spatial information, yet each additional electrode also increases hardware complexity, wiring, skin-contact area, setup requirements and potential points of failure (Vujaklija et al., 2016). Electrode reduction therefore represents an important design problem: the objective is not simply to remove as many electrodes as possible, but to identify the smallest configuration that preserves an acceptable level of recognition performance.

A substantial literature has addressed channel and feature reduction in sEMG using feature-ranking methods, wrapper approaches, sparse representations, and multi-channel or high-density electrode arrays (Phinyomark et al., 2012; Liu et al., 2014; Al-Timemy et al., 2013; Stango et al., 2015; Qu et al., 2021). However, three methodological issues complicate the translation of these results to devices intended for use across individuals.

First, electrode subsets are frequently selected and evaluated within the same users. Such user-dependent designs are appropriate for personalised systems, but they do not establish whether a common reduction strategy will generalise to an unseen user. Subject-independent recognition poses a more stringent problem because inter-individual anatomical and physiological variability can alter both the recorded signals and the spatial distribution of discriminative information (Phinyomark and Scheme, 2018; Cote-Allard et al., 2019). Moreover, allowing information from the eventual test participant to influence standardisation, ranking or model selection introduces optimistic bias.

Second, model-attribution methods provide an appealing means of identifying informative recording sites but introduce their own methodological challenges. TreeSHAP, for example, estimates the contribution of individual model inputs to predictions (Lundberg and Lee, 2017; Lundberg et al., 2020) and can therefore be aggregated to rank electrodes, as has been done for channel reduction in other biosignal tasks (Salmi et al., 2025). Yet attribution estimates may vary across training samples (Visani et al., 2022; Kelodjou et al., 2023), and a ranking derived from the same data used for evaluation can itself become a source of information leakage. A valid subject-independent electrode-selection procedure must therefore estimate electrode importance without allowing the test participant to influence either the attributing model or the resulting ranking. Treating attributions as distributions across resamples and penalising unstable ones has also been proposed (Xiang et al., 2026; Meinshausen and Bühlmann, 2010); the stability penalty tested here is related to that idea and is not claimed as new.

Third, determining the number of electrodes to retain is often treated descriptively: an accuracy curve is inspected and a visually acceptable operating point is selected. Such an approach does not formally define how much performance loss is acceptable (Schuirmann, 1987; Walker and Nowacki, 2011). If the tolerance is chosen after observing the reduction curve, the resulting claim cannot be interpreted as a prespecified non-inferiority decision (Campbell and Gustafson, 2018). A more defensible formulation is therefore to define the acceptable loss in advance and evaluate reduced configurations against that threshold.

The present study addresses these issues through a participant-level, cross-fitted evaluation framework (figure 1). For each outer leave-one-subject-out fold, the held-out participant is excluded from feature standardisation, model fitting and electrode ranking. TreeSHAP attributions are estimated on participant-disjoint inner folds, so each attribution is obtained from observations not used to train the corresponding attributing model. Reduced models are subsequently retrained using only the selected electrodes and evaluated on the outer held-out participant. The primary decision rule was specified before the reduced-array performance curve was examined: a subset was considered non-inferior when the lower bound of the participant-level confidence interval for its Macro-F1 difference from the full array remained above a margin of -0.02.

Rather than introducing attribution as a new selection algorithm, the study evaluates whether attribution-guided selection provides useful and reproducible information under a leakage-controlled, subject-independent design, using the 43-participant GRABMyo dataset (Pradhan et al., 2022). Four prespecified questions guide the analysis. We first ask (H1) whether any reduced subset is non-inferior to the complete 16-electrode array. We then examine (H2) whether attribution-guided subsets perform better than random subsets of the same size, (H3) whether incorporating attribution stability improves selection relative to attribution magnitude alone, and (H4) whether the ranking behaves faithfully when highly versus weakly ranked electrodes are removed and the classifier is retrained. Additional exploratory analyses examine ranking stability across participants, transfer across recording sessions, bipolar re-referencing, replication on NinaPro DB5 (Pizzolato et al., 2017), and participant-level aggregation of attribution variability. Every answer is reported, including the negative ones.

2 Methods

2.1 Datasets and signal processing

The primary analysis used GRABMyo v1.1.0 (Pradhan et al., 2022; Jiang et al., 2024), distributed through PhysioNet (Goldberger et al., 2000), comprising recordings from 43 able-bodied participants collected over 3 sessions on different days. Participants were 23 male and 20 female (age 26.4 ± 2.9 years). Signals were recorded with an EMGUSB2+ amplifier (OT Bioelettronica) at 2048 Hz; in each session, participants performed each gesture in 7 trials of 5 s. Sixteen monopolar electrodes were positioned in two circumferential rings of eight electrodes around the forearm (F1–F8 and F9–F16), separated longitudinally by 2 cm, with electrode j of one ring aligned with electrode $j + 8$ of the other. Mean forearm circumference at the ring position was 24.1 ± 2.3 cm, giving about 3.0 cm between neighbouring electrodes in a ring. Participants performed 16 active hand and wrist gestures together with rest; the primary recognition task considered the 16 active gestures only, for which




Figure 1. Study overview. (A) GRABMyo data and preprocessing: two circumferential rings of eight monopolar electrodes 2 cm apart on the forearm (proximal ring F1–F8, distal ring F9–F16), 250 ms analysis windows with 50% overlap and ten features per channel (160 features per window). Electrode housings are drawn schematically, and the muscle drawings show general forearm anatomy only; electrodes were not registered to individual muscles. (B) Subject-independent validation: each participant is held out in turn (outer leave-one-subject-out split), and attributions are computed by cross-fitting over five participant-disjoint groups of the remaining participants. (C) The two electrode rankings compared in this study: cross-fitted attribution magnitude alone ($\lambda = 0$) and the stability-regularised score $Q_c = A_c - \lambda U_c$ (SR-CFA). The panel shows how the rankings are formed, not which performed better; the stability penalty produced no statistically detectable improvement (section 3.2).

chance Macro-F1 is about 0.063. All 30,715 distributed files matched the dataset's published SHA-256 checksums, and all 15,351 records were analysed.

Terminology is used as follows. An *electrode* is a physical recording site and a *channel* is the signal it provides to the classifier; in the monopolar analyses each electrode yields one channel, so selecting electrodes and selecting channels are the same operation. The *montage* is the spatial arrangement of the electrodes and the *array* is the complete set. A *bipolar pair* is the difference between two electrodes, derived from the monopolar recordings in one exploratory analysis. The subset size k always counts the selection unit of the analysis in question.

NinaPro DB5 (Pizzolato et al., 2017) was analysed separately as an exploratory replication dataset and was never pooled with GRABMyo. It provides 10 able-bodied subjects (8 male, 2 female; age 28.0 ± 4.0 years) recorded with two Myo armbands (16 channels, 200 Hz). We used exercise 1 (12 finger movements, 6 repetitions each), segmented with the refined movement and repetition labels supplied with the dataset. The DB5 values are 8-bit integer codes, not physical units, and its channel numbers have no anatomical correspondence with those of GRABMyo, so only the behaviour of the procedure is compared across datasets.




Figure 2. Study workflow. For each held-out participant p, feature standardisation, model fitting and electrode ranking use only the remaining participants R_{-p}. Attributions are computed on participant-disjoint inner folds of R_{-p}, and models restricted to the top-ranked k electrodes are retrained on R_{-p} and tested on p. The same workflow was applied separately to each dataset.

Ethics statement This study is a secondary analysis of previously published, de-identified datasets and involved no new participant recruitment or data collection. The original data collections were conducted in accordance with the Declaration of Helsinki and approved by the Office of Research Ethics of the University of Waterloo (ORE# 31346; GRABMyo) and the Ethics Commission of the Canton of Valais, Switzerland (NinaPro DB5); informed-consent procedures are described in the dataset publications (Pradhan et al., 2022; Pizzolato et al., 2017). [AUTHOR INPUT REQUIRED: state whether an institutional review board determination or exemption was obtained for the present secondary analysis.]

Sex as a reported variable Sex was taken as reported by the source datasets. Sex-stratified analyses were not prespecified, and the study was not designed to estimate sex differences in recognition performance or electrode ranking; no sex-specific inference is made.

GRABMyo signals were band-pass filtered between 20 and 450 Hz using a fourth-order zero-phase Butterworth filter and notch filtered at 50 Hz ($Q = 30$), the mains frequency identified from the spectra (mean band-power ratio 10.35 at 50 Hz and 0.62 at 60 Hz). Because the filters are zero-phase, the processing is non-causal and suitable only for offline evaluation. Signals were segmented into 250-ms windows with 50% overlap, without allowing windows to cross trial boundaries. Ten conventional time- and frequency-domain features were extracted from each electrode: mean absolute value, root mean square, variance, waveform length, zero crossings (ZC), slope sign changes (SSC), integrated EMG, Willison amplitude (WAMP), mean frequency and median frequency. The resulting representation contained 160 features per window (598,689 windows in total, 563,472 from active gestures). Signal-processing parameters were fixed consistently across the analysis. The amplitude threshold of the ZC, SSC and WAMP features was set by one rule for every representation: 1.15 times the median absolute first difference of the signal, rounded to one significant figure. For GRABMyo it was computed once, from 54 unfiltered recordings of 6 participants, giving 10^{-2} mV; it was therefore not estimated within each outer fold. A leave-one-out check gave the same value in every fold (supplementary methods S2), so fold-wise estimation would have produced identical features. DB5 was band-pass filtered at 10–95 Hz without a notch filter, with 250 ms windows (50 samples) and a threshold of 7 quantisation steps from the same rule; the leave-one-out recomputation gave the same threshold in every DB5 fold.

2.2 Subject-independent electrode ranking and model evaluation

All primary evaluations followed an outer leave-one-subject-out (LOSO) design (figure 2). For each held-out participant p, all preprocessing statistics, model fitting and electrode ranking were derived exclusively from the remaining participants R_{-p}. The held-out participant contributed only to the final performance estimate.

XGBoost was used as the primary classifier for the electrode-selection experiments because the analysis required repeated model retraining across participants, electrode counts and selection strategies. It was designated in the analysis plan on these computational grounds; the plan was written before the full-cohort classifier comparison had finished, although results from an earlier, superseded run on 15 participants were available. XGBoost used 300 trees, depth 6, learning rate 0.1, row and column subsampling of 0.8, histogram-based tree construction and a multiclass softmax objective, and was trained on all 550,368 training windows of each fold. Support vector machine and random forest models were additionally evaluated using the complete array to provide classifier-level

context: a support vector machine with a radial basis function kernel ($C = 1$, kernel width set from the feature variance), trained on a stratified sample of 40,000 windows, and a random forest (200 trees, minimum leaf size 2) trained on 200,000 windows. These sample sizes were set by computational cost and chosen on two folds (supplementary methods S1). Halving the support vector machine's training sample on one fold lowered Macro-F1 from 0.591 to 0.555, so its result is likely a lower bound. Classifier hyperparameters were fixed for the reported runs. Because this earlier budget calibration used test-fold outcomes, the development process was not fully isolated from the evaluation cohort.

Electrode importance was estimated within each outer training set using $K = 5$ participant-disjoint inner folds. For each inner fold j, an XGBoost model was trained on the remaining inner participants and TreeSHAP values were evaluated on the participants excluded from that model. Each inner model used a random sample of at most 200,000 windows standardised with inner-training statistics only, and exact tree-path-dependent TreeSHAP attributions on raw model margins (Lundberg et al., 2020) were computed for up to 20,000 windows sampled at random from the inner-held-out participants. Because participants contribute similar numbers of windows (median 13,104), each participant carries roughly equal weight. In every fold we verified that no inner-held-out participant contributed to the corresponding inner training set. Absolute attributions were averaged over the $G = 16$ classes and the $F = 10$ features belonging to each electrode,

$$\Phi_{i,c} = \frac{1}{GF} \sum_{g=1}^{G} \sum_{f=1}^{F} |\phi_{i,g,c,f}|, \tag{1}$$

then averaged over windows and normalised across electrodes to give $I_{j,c}$. The resulting cross-fitted attribution magnitude $A_c = \text{median}_j\, I_{j,c}$ represented the typical importance of electrode c, whereas $U_c = \text{median}_j |I_{j,c} - A_c|$ quantified its variation across inner folds. Electrodes were ranked according to

$$Q_c(\lambda) = A_c - \lambda U_c, \qquad \lambda \in \{0, 0.25, 0.5, 1\}, \tag{2}$$

where $\lambda = 0$ corresponds to cross-fitted attribution magnitude alone and positive values introduce a penalty for unstable attribution (the stability-regularised cross-fitted attribution score, SR-CFA). Ties were broken by the lower channel number. The plan did not specify a criterion for choosing λ, so all four values are reported.

For each outer fold, each λ and each retained electrode count $k \in \{1, \ldots, 16\}$, the classifier was retrained from scratch using only the top-k electrodes and evaluated on the held-out participant. All models of this electrode-count performance curve used the same fixed, stratified sample of 200,000 training windows of the fold. This separation between ranking and final evaluation ensured that neither electrode selection nor model training used information from the participant on whom performance was measured.

2.3 Hypothesis testing and statistical analysis

The primary endpoint was participant-level Macro-F1 across the 16 active gestures; accuracy and balanced accuracy were also recorded. For each reduced configuration, performance was compared with the 16-electrode model of the same curve, trained using the same outer-fold data. The paired difference for participant p and subset size k was defined as

$$\Delta_{p,k} = \text{F1}_{p,k} - \text{F1}_{p,16}. \tag{3}$$

A subset was considered non-inferior (H1) when L_k, the lower bound of the 95% participant-cluster-bootstrap confidence interval for the mean paired difference $\bar{\Delta}_k$, exceeded the prespecified margin $-\delta$ with $\delta = 0.02$ Macro-F1. The smallest subset satisfying this criterion defined the minimum acceptable electrode count, $k^* = \min\{k < 16 : L_k > -\delta\}$, recorded as 'none' if no k qualifies. As required by the plan, the family of 15 tests was also corrected by the Holm procedure, using one-sided cluster-bootstrap p-values for $H_0 \colon \bar{\Delta}_k \leq -\delta$ at $\alpha = 0.025$. The margin is an absolute difference in Macro-F1 on a 0–1 scale, fixed before the curve was computed as the largest loss we would accept in exchange for fewer electrodes. Relative to the full-array reference it corresponds to about 3.7% of performance and about 4.2% of the distance above chance (1/16). It is an engineering tolerance and has not been validated against clinical outcomes or user performance. Because k^* depends on δ, sensitivity analyses were additionally performed at the prespecified margins of 0.01 and 0.03. Non-inferiority within this tolerance does not imply equivalence.

To determine whether the attribution ranking conveyed useful information beyond chance (H2), attribution-selected subsets were compared with random subsets of identical size at $k = 4$, 6, 8 and 12. Additional exploratory comparisons included permutation importance, mutual information and geometric farthest-point selection. Every ranking was computed on the training participants only:

- SR-CFA ($\lambda = 0.5$) and cross-fitted SHAP ($\lambda = 0$);
- permutation importance: a model was fitted on 20,000 training windows and each of the 160 features was permuted in 8,000 further windows (3 repeats, Macro-F1 scoring); an electrode's score is the mean over its 10 features;
- mutual information between each feature and the gesture label on the same 20,000 windows, averaged over the electrode's 10 features;
- geometric selection: farthest-point sampling on the 2×8 montage (arc distance around a 24 cm circumference, 2 cm between rings, starting from F1, ties to the lower index), which uses no data;
- random subsets: 100 independent random subsets per participant at each k, and 500 at the prespecified k^* (seed fixed per participant).

Evaluating hundreds of random subsets per participant required a lighter comparison model (XGBoost with 50 trees, trained on a stratified sample of 150,000 windows), applied identically to every method; its absolute scores are lower than those of the performance curve and are compared only with each other. The confirmatory H2 family contrasts each attribution ranking with the participant's mean over its random subsets (Holm over 8 tests); comparisons with permutation importance, mutual information and geometric selection are exploratory and uncorrected. We also report the percentile of the SR-CFA subset within each participant's random subsets. Below, 'SR-CFA' refers to $\lambda = 0.5$ and 'cross-fitted SHAP' to $\lambda = 0$.

Stability regularisation (H3) was assessed by comparing each $\lambda > 0$ with $\lambda = 0$ on the performance curve at $k = 4$, 6, 8 and 12 (Holm over 12 tests). Attribution faithfulness (H4) was assessed through deletion with retraining (Samek et al., 2017; Hooker et al., 2019): performance was measured after preferentially removing highly ranked (Remove-Top), randomly selected (Remove-Random) or weakly ranked (Remove-Bottom) electrodes. For $n \in \{2, 4, \ldots, 14\}$ electrodes retained, the comparison model was retrained after removing the $16 - n$ highest- or lowest-ranked electrodes; Remove-Random used the random subsets above and is available for $n \in \{4, 6, 8, 12\}$. A faithful ranking gives Remove-Top $<$ Remove-Random $<$ Remove-Bottom. Paired contrasts were Holm-corrected within H4.

The participant, rather than the individual signal window, was treated as the statistical unit throughout ($n = 43$ for GRABMyo, 10 for DB5); no participant was excluded. Confidence intervals (CI) were obtained as 95% percentile intervals from 10,000 participant-level cluster-bootstrap resamples (seed 20260728). Because LOSO training sets overlap, these intervals quantify variation in the observed out-of-fold participant scores and do not fully propagate uncertainty from retraining on a new cohort. Paired comparisons used the Wilcoxon signed-rank test with the matched-pairs rank-biserial correlation r_{rb} as the effect size, and Holm correction was applied within each confirmatory test family. A non-significant difference was interpreted only as an absence of statistically detectable evidence and not as evidence of equivalence, because no equivalence test was performed.

2.4 Exploratory and sensitivity analyses

Exploratory analyses examined the stability of electrode rankings across participants, replication of the reduction procedure on NinaPro DB5, cross-session generalisation within GRABMyo, and bipolar re-referencing of the original monopolar recordings. These analyses were interpreted separately from the prespecified primary non-inferiority test.

Ranking stability We compared rankings derived from single participants' attributions (all 903 pairs), from the five inner folds and from the 43 outer folds, using Spearman's ρ, Kendall's τ and the Jaccard index of the top-k sets. A participant's attribution profile is the mean of their normalised attribution over the outer folds in which they were held out of an inner model. Outer folds share about 98% of their training participants, so their agreement is not an independent estimate of between-person stability.

Replication The LOSO baseline, cross-fitted ranking ($K = 5$) and performance curve were repeated on DB5, with its full array as reference ($k = 16$), using all training windows and the same bootstrap, margins and k^* rule. Holm-corrected results are also reported. No setting was tuned on DB5.

Table 1. Full 16-channel leave-one-subject-out performance on GRABMyo (43 participants, 16 active gestures; chance Macro-F1 ≈ 0.063). CI: 95% confidence interval from a participant cluster bootstrap. RBF-SVM: support vector machine with a radial basis function kernel.

Model	Training windows	Macro-F1	95% CI	SD	Range
XGBoost	all	0.542	0.517–0.567	0.085	0.385–0.702
RBF-SVM	40,000	0.538	0.513–0.562	0.085	0.383–0.718
Random forest	200,000	0.502	0.475–0.528	0.090	0.346–0.670

Paired comparisons (Wilcoxon signed-rank, Holm over 3): Random forest − RBF-SVM = −0.0362 (95% CI −0.0444 to −0.0284), $p_{\text{Holm}} < 0.001$, r_{rb} = -0.97; Random forest − XGBoost = −0.0408 (95% CI −0.0475 to −0.0340), $p_{\text{Holm}} < 0.001$, r_{rb} = -0.98; RBF-SVM − XGBoost = −0.0045 (95% CI −0.0107 to +0.0015), p_{Holm} = 0.154, r_{rb} = -0.25.

Cross-session XGBoost was trained on session 1 of R_{-p} and tested on each session of p, and trained on sessions 1 and 2 and tested on session 3. These comparisons use the full array and do not test a reduced layout under session transfer. The second condition also has about twice the training data, so it confounds session diversity with data quantity.

Bipolar re-referencing To use a bipolar pair as the selection unit, we formed $B_j = F_j - F_{j+8}$ ($j = 1, \ldots, 8$) from the monopolar recordings, re-extracted features with the threshold given by the same rule (4×10^{-3} mV; the bipolar signal is about half the monopolar amplitude), and repeated the analysis with the $k = 8$ model as reference. These are derived pairs, not physical bipolar electrodes.

Participant-level aggregation (post hoc) A post-hoc extension estimated attribution dispersion directly across participants rather than across pooled inner-fold attribution vectors. We retained each inner-held-out participant's normalised attribution vector and calculated its median and unscaled median absolute deviation across all outer-training participants. The same score and all four penalties were evaluated at every subset size under the original performance-curve model and training budget, and no penalty was selected from outer-test performance. Exact duplicate electrode sets reused the same fitted evaluation. Comparisons against the original aggregation were paired by participant, and the family of all reduced-size comparisons across penalties received Holm correction. Person vectors share inner models, so their dispersion measures heterogeneity rather than a standard error.

Departures from the original statistical analysis plan are documented in supplementary methods S1. These included post-hoc specification of the detailed H2–H4 test families, application of the plan's Holm correction for H1 after the primary analysis, selection of computational training budgets using two folds, and addition of fixed-layout, feature-redundancy and participant-level attribution analyses; two exploratory analyses listed in the plan were not run. None altered the prespecified non-inferiority margin, statistical unit, bootstrap procedure or definition of the minimum acceptable electrode count.

3 Results

3.1 Subject-independent performance and electrode reduction

With all 16 electrodes, XGBoost achieved a mean Macro-F1 of 0.542 (95% CI 0.517–0.567) across the 43 held-out participants. The corresponding values were 0.538 (95% CI 0.513–0.562) for the support vector machine and 0.502 (95% CI 0.475–0.528) for the random forest (table 1, figure 3). No statistically detectable difference was observed between XGBoost and the support vector machine (paired difference +0.0045, 95% CI −0.0015 to +0.0107, $p_{\text{Holm}} = 0.154$, $r_{rb} = +0.25$), whereas both models outperformed the random forest ($p_{\text{Holm}} < 0.001$ and < 0.001; $r_{rb} = +0.98$ and +0.97). Between-participant variability was substantially larger than the difference between the two higher-performing classifiers: XGBoost ranged from 0.385 to 0.702 across participants. Participant-level performance for every active gesture, reconstructed directly from the saved full-array predictions, is reported in the supplementary material.

Recognition performance decreased progressively as electrodes were removed (figure 4, table 2). The reference of the curve ($k = 16$) scored 0.541, slightly below the full-data result because all curve models share one 200,000-window training sample. Under the prespecified margin of 0.02 Macro-F1, the smallest non-inferior configuration contained $k^* = 12$ electrodes. Its mean difference from the corresponding 16-electrode reference was −0.0133 (95% CI −0.0160 to −0.0106), whereas the 11-electrode subset crossed the non-inferiority boundary with a lower confidence bound of −0.0246. All configurations containing 12–15 electrodes satisfied the criterion, and the minimum acceptable count remained 12 after Holm correction ($k = 12$: $p_{\text{Holm}} = 0.002$; supplementary table S1). As




Figure 3. Full-array leave-one-subject-out performance on GRABMyo (43 participants). (a) Macro-F1 of three classifiers; each point is one held-out participant, grey lines join the same participant, black bars mark medians and diamonds mark means with 95% cluster-bootstrap confidence intervals. The support vector machine and random forest were trained on samples of 40,000 and 200,000 windows. (b) Paired XGBoost minus support vector machine difference per participant, sorted; the grey band is the 95% confidence interval of the mean difference.

Table 2. Electrode-count performance curve and non-inferiority on GRABMyo (cross-fitted ranking, $\lambda = 0$; 43 participants; every k retrained from scratch on the same 200,000 training windows of each fold). Δ_k = paired difference to the $k = 16$ model of the same curve. Non-inferior when the lower 95% bound exceeds $-\delta$; $\delta = 0.02$ is primary (bold). Holm: non-inferiority at $\delta = 0.02$ after Holm correction over the 15 subset sizes (supplementary table S1). CI: 95% confidence interval.

k	Macro-F1 (mean ± SD)	Δ_k	95% CI	δ=0.01	**δ=0.02 (primary)**	δ=0.03	Holm (0.02)
16	0.541 ± 0.084	reference					
15	0.538 ± 0.083	−0.0036	−0.0056 to −0.0017	yes	**yes**	yes	yes
14	0.535 ± 0.083	−0.0060	−0.0081 to −0.0039	yes	**yes**	yes	yes
13	0.530 ± 0.083	−0.0109	−0.0133 to −0.0085	no	**yes**	yes	yes
12	0.528 ± 0.084	−0.0133	−0.0160 to −0.0106	no	**yes**	yes	yes
11	0.521 ± 0.084	−0.0204	−0.0246 to −0.0162	no	**no**	yes	no
10	0.515 ± 0.083	−0.0259	−0.0311 to −0.0205	no	**no**	no	no
9	0.510 ± 0.086	−0.0318	−0.0387 to −0.0246	no	**no**	no	no
8	0.505 ± 0.085	−0.0366	−0.0440 to −0.0289	no	**no**	no	no
7	0.500 ± 0.085	−0.0413	−0.0489 to −0.0333	no	**no**	no	no
6	0.480 ± 0.084	−0.0610	−0.0713 to −0.0507	no	**no**	no	no
5	0.428 ± 0.075	−0.1133	−0.1277 to −0.0993	no	**no**	no	no
4	0.397 ± 0.071	−0.1444	−0.1590 to −0.1306	no	**no**	no	no
3	0.340 ± 0.057	−0.2015	−0.2212 to −0.1822	no	**no**	no	no
2	0.299 ± 0.057	−0.2420	−0.2603 to −0.2230	no	**no**	no	no
1	0.181 ± 0.040	−0.3603	−0.3832 to −0.3369	no	**no**	no	no

expected, the selected count depended on the tolerance: the minimum was 14 electrodes at $\delta = 0.01$ and 11 at $\delta = 0.03$, again unchanged by Holm correction.

Thus, a 25% reduction in electrode count was supported within the prespecified engineering tolerance, whereas more aggressive reduction was not. Halving the array to eight electrodes gave a mean difference of −0.0366 (95% CI −0.0440 to −0.0289), and every reduced configuration performed numerically below the complete array. The electrodes most often excluded at $k = 12$ were 9, 11, 12, 16 (43 of 43 folds).

3.2 Evaluation of attribution-guided electrode selection

Attribution-guided selection performed consistently better than the mean of random subsets at all four tested subset sizes (table 3, figure 5; Holm-corrected, 8 tests). The largest gains occurred at intermediate reductions, particularly at six electrodes (+0.0370 Macro-F1 for SR-CFA, 95% CI +0.0286 to +0.0452, $p_{\text{Holm}} < 0.001$, $r_{rb} = +0.92$; +0.0243 at $k = 8$), whereas the advantage became smaller as the retained array approached the full configuration (+0.0074 at $k = 12$, $p_{\text{Holm}} < 0.001$) and was also small at $k = 4$ (+0.0172). Nevertheless, attribution-selected subsets were not uniformly optimal: within each participant's random-subset distribution, they typically ranked above the median




Figure 4. Electrode-count performance curve on GRABMyo (43 participants). (a) Mean Macro-F1 of models retrained on the top-k electrodes of each fold's cross-fitted ranking, for the four values of the stability weight λ (the curves overlap); the band is the 95% cluster-bootstrap confidence interval for $\lambda = 0$. (b) Paired difference to the $k = 16$ model with 95% confidence interval. The thick dashed line marks the primary margin ($-\delta$, $\delta = 0.02$); the thin dotted and dash-dotted grey lines mark the sensitivity margins (0.01 and 0.03).

Table 3. Channel-selection methods at matched k (43 participants). All methods use one comparison model (XGBoost, 50 trees, 150,000 training windows per fold), so values are comparable within the table but lower than table 2. SR-CFA: stability-regularised cross-fitted attribution. H2 p values Holm-corrected over 8 tests; †exploratory, uncorrected.

Method	$k = 4$	$k = 6$	$k = 8$	$k = 12$
SR-CFA (λ=0.5)	0.383 ± 0.071	0.453 ± 0.080	0.469 ± 0.081	0.485 ± 0.079
Cross-fitted SHAP (λ=0)	0.382 ± 0.071	0.453 ± 0.080	0.468 ± 0.081	0.485 ± 0.079
Permutation importance	0.385 ± 0.070	0.440 ± 0.080	0.466 ± 0.080	0.485 ± 0.079
Geometric farthest-point	0.383 ± 0.069	0.419 ± 0.068	0.465 ± 0.072	0.478 ± 0.076
Mutual information	0.280 ± 0.049	0.360 ± 0.066	0.396 ± 0.068	0.480 ± 0.079
Random subsets (mean)	0.366 ± 0.059	0.416 ± 0.068	0.445 ± 0.071	0.478 ± 0.077
SR-CFA − random (H2)	+0.0172 $p = 0.007$	+0.0370 $p < 0.001$	+0.0243 $p < 0.001$	+0.0074 $p < 0.001$
SR-CFA − perm†	−0.0018 $p = 0.678$	+0.0126 $p = 0.009$	+0.0033 $p = 0.069$	+0.0001 $p = 0.995$
SR-CFA − geom†	−0.0005 $p = 0.986$	+0.0341 $p < 0.001$	+0.0047 $p = 0.267$	+0.0070 $p = 0.058$
SR-CFA percentile among random subsets (median)	65 (100 subsets)	89 (100 subsets)	82 (100 subsets)	66 (500 subsets)

but only occasionally reached the extreme upper tail. The median percentile of the SR-CFA subset was 65, 89, 82 and 66 at $k = 4$, 6, 8 and 12, and it reached the 95th percentile for 11 of 43 participants at $k = 6$ and 1 at $k = 12$.

In exploratory, uncorrected contrasts, no statistically detectable difference between SR-CFA and permutation importance was observed at $k = 4$, 8 or 12 (for example, $k = 8$: +0.0033, $p = 0.069$); at $k = 6$, SR-CFA was higher by +0.0126 ($p = 0.009$). Likewise, no statistically detectable difference between SR-CFA and geometric selection, which spreads electrodes evenly around both rings without using data, was observed at $k = 4$ (−0.0005, $p = 0.986$), $k = 8$ (+0.0047, $p = 0.267$) or $k = 12$ (+0.0070, $p = 0.058$); at $k = 6$, SR-CFA was higher by +0.0341 ($p < 0.001$). Mutual information gave the lowest scores at every $k \leq 8$.

The deletion-with-retraining experiment provided complementary evidence that the ranking captured information relevant to classification (table 4, figure 6). Removing highly ranked electrodes produced the largest loss in performance, removing low-ranked electrodes produced the smallest loss, and random deletion lay between these conditions; this ordering held at every number of electrodes retained, and 14 of the 15 Holm-corrected contrasts were significant. With six electrodes retained, Remove-Top scored 0.376, Remove-Random 0.416 and Remove-Bottom 0.454 (Remove-Top minus Remove-Random −0.0398, $p_{\text{Holm}} < 0.001$; Remove-Random minus Remove-Bottom −0.0377, $p_{\text{Holm}} < 0.001$). The ordering was most pronounced at intermediate subset sizes and became smaller near the extremes, where either very few or nearly all electrodes remained; the only non-significant contrast was Remove-Top versus Remove-Bottom with 14 electrodes retained (−0.0038,




Figure 5. Selection methods at matched subset size k (GRABMyo, 43 participants; one comparison model shared by all methods). (a) Mean Macro-F1 with 95% cluster-bootstrap confidence intervals, grouped by method type: attribution rankings (stability-regularised cross-fitted attribution, SR-CFA, and cross-fitted SHAP), model-based rankings (permutation importance and mutual information), data-free geometric selection and the mean over random subsets. (b) Percentile of the SR-CFA subset within each participant's own random subsets (one point per participant; bar = median; 100 random subsets per participant, 500 at $k = 12$).

Table 4. Deletion (remove-and-retrain) faithfulness test (H4; 43 participants). All deletion orders use the same comparison model and training sample; Remove-Random averages 100 random subsets per participant (500 at 12 channels) and is available only where these were sampled. Differences are paired means with Holm-corrected p in parentheses.

Retained	Remove-Top	Remove-Random	Remove-Bottom	Top−Bottom	Top−Random	Random−Bottom
2	0.263	–	0.292	−0.0297 (<0.001)	–	–
4	0.351	0.366	0.382	−0.0314 (<0.001)	−0.0151 (0.010)	−0.0163 (0.013)
6	0.376	0.416	0.454	−0.0775 (<0.001)	−0.0398 (<0.001)	−0.0377 (<0.001)
8	0.408	0.445	0.469	−0.0618 (<0.001)	−0.0377 (<0.001)	−0.0242 (<0.001)
10	0.445	–	0.475	−0.0297 (<0.001)	–	–
12	0.455	0.478	0.485	−0.0302 (<0.001)	−0.0228 (<0.001)	−0.0074 (<0.001)
14	0.487	–	0.491	−0.0038 (0.150)	–	–

$p_{\text{Holm}} = 0.150$).

The stability penalty, in contrast, had little practical effect. Attribution variability across the inner participant groups was small relative to attribution magnitude (median U_c/A_c 2.4%, maximum 9.4%), so positive values of λ rarely changed the selected subsets. Relative to $\lambda = 0$, the top-12 set changed in 0, 0 and 1 of 43 folds at $\lambda = 0.25$, 0.5 and 1, and across all four k tested at least 36 of 43 participants received the identical subset under both rankings. No stability-regularised configuration produced a statistically detectable improvement over cross-fitted attribution magnitude alone after multiplicity correction (0 of 12 contrasts; largest mean difference 0.0015; smallest $p_{\text{Holm}} = 0.216$; supplementary table S2). The four curves in figure 4(a) differ by at most 0.0040, and k^* was identical under every λ and margin; H3 was not supported. The participant-level aggregation extension increased the measured heterogeneity of attribution profiles but likewise failed to establish a corrected predictive advantage for the penalty (figure 7).

Participant-level attribution extension (post hoc) All 43 outer participants were evaluated under each aggregation. The descriptive median relative MAD across outer-fold/electrode cells increased from 2.40% to 4.62%; these cells are not independent replicates. Of the 60 reduced-size, penalty-specific comparisons with the original fold aggregation, 42 had any nonzero participant difference. After Holm correction across this family, 0 favoured participant aggregation and 0 favoured fold aggregation. The full set of penalties is reported in the supplementary material; no candidate




Figure 6. Deletion-with-retraining test of attribution faithfulness on GRABMyo (43 participants; mean and 95% cluster-bootstrap confidence interval). Remove-Top deletes the highest-ranked electrodes, Remove-Bottom the lowest-ranked and Remove-Random a random set; all use the same comparison model and training sample. Remove-Random is shown where random subsets were sampled (4, 6, 8 and 12 electrodes retained).




Figure 7. Post-hoc participant-level attribution extension. Every cell is the mean paired difference across the GRABMyo participants, comparing person-level with fold-level aggregation under the same penalty, subset size and model protocol. Blue favours person-level aggregation and orange favours fold aggregation. Asterisks identify Holm-corrected differences within the complete family. All candidates are shown; outer-test outcomes did not select a penalty.

was selected using outer-test performance. These post-hoc comparisons test the aggregation change and do not replace the original H3 result. Within participant aggregation, 0 of 12 Holm-corrected comparisons supported a benefit of the stability penalty and 0 supported harm.

3.3 Robustness and exploratory analyses

Electrode rankings showed substantial agreement across participants, particularly at the upper and lower ends of the ranking (table 5, figure 8). Rankings computed from individual participants' attributions agreed closely (mean Spearman $\rho = 0.927 \pm 0.038$ over 903 pairs; mean Jaccard index of the top six 0.905), and rankings from disjoint inner folds agreed at $\rho = 0.948$. The same four electrodes (4, 6, 7, 14) formed the top four in 36 of 43 outer folds and carried 34.0% of the summed attribution, whereas the eight lowest-ranked electrodes carried 39.6%; rank ranges were widest in the middle of the ordering. The outer-fold agreement ($\rho = 0.987$) is descriptive only, because those training sets share about 98% of their participants. This indicates that the subject-independent procedure identified reproducible spatial preferences, although the data do not support assigning these positions to individual muscles.

Replication on NinaPro DB5 was inconclusive. The dataset contained only ten participants and produced substantially wider uncertainty than GRABMyo: full-array Macro-F1 was 0.281 ± 0.099 (12 movements; chance about 0.083), with between-subject values from 0.096 to 0.386 (figure 9(b)). One reduced configuration ($k = 12$) met the non-inferiority margin by a negligible amount ($\bar{\Delta}_{12} = -0.0126$, lower bound -0.019994), but the result was sensitive to bootstrap variation (non-inferior under 95.5% of 200 alternative bootstrap seeds), $k = 13$, 14 and 15 all failed (lower bounds -0.03119, -0.02744

Table 5. Agreement of electrode rankings on GRABMyo ($\lambda = 0$). Values are means over pairs.

Comparison	Pairs	Spearman ρ	Kendall τ	Jaccard, top 4	Jaccard, top 6	Jaccard, top 8
Per-subject attribution (between people)	903	0.927 ± 0.038	0.801	0.752	0.905	0.803
Inner folds (disjoint subject groups)	430	0.948 ± 0.021	0.840	0.816	0.992	0.838
Outer folds (overlapping training*)	903	0.987 ± 0.008	0.943	0.888	1.000	0.923

*Outer training sets share about 98% of their participants, so their agreement is descriptive, not an independent estimate of between-subject stability.




Figure 8. Cross-fitted electrode attribution on GRABMyo ($\lambda = 0$, 43 outer folds). (a) Mean share of attribution per electrode, drawn on the 2×8 montage; each ring 2 electrode is aligned with, and 2 cm from, the ring 1 electrode at the same circumferential position. (b) Rank of each electrode across inner folds (median and range).

and -0.02377), and no subset remained non-inferior after Holm correction. The replication therefore neither confirms nor contradicts the primary GRABMyo result.

Cross-session analyses provided no evidence of an additional performance loss on later recording days beyond the existing subject-independent gap (figure 9(a)). Training on session 1 and testing on sessions 1, 2 and 3 gave 0.492, 0.528 and 0.510; the session-2 difference from session 1 was nominally significant ($+0.0355$, 95% CI $+0.0091$ to $+0.0621$, uncorrected $p = 0.024$), and no statistically detectable difference was observed for session 3 ($+0.0172$, $p = 0.272$). Training on sessions 1 and 2 improved session-3 performance over training on session 1 alone ($+0.0210$, 95% CI $+0.0146$ to $+0.0274$, $p < 0.001$). However, these contrasts do not isolate electrode displacement, and the apparent benefit of training on two sessions is confounded by the larger amount of training data.

Finally, results from the derived bipolar representation differed from those of the primary monopolar analysis. With the eight derived pairs as the selection unit, full-array Macro-F1 was 0.478 ± 0.089, lower than with monopolar electrodes (figure 9(c)). No reduction from eight to seven bipolar pairs satisfied the primary non-inferiority criterion ($\bar{\Delta}_7 = -0.0158$, lower bound -0.0203; non-inferior under 0% of 200 alternative bootstrap seeds). Under the sensitivity margins, k^* was none at $\delta = 0.01$ and 7 at $\delta = 0.03$, identically for all λ, and pairs 5, 6, 7 formed the top three in 30 of 43 folds. Consequently, the 25% reduction identified in the primary analysis should be interpreted specifically as a reduction in monopolar recording sites on the GRABMyo montage, rather than as evidence that an equivalent proportion of physical bipolar sensors can be removed.

4 Discussion

This study examined how far a 16-electrode forearm sEMG array could be reduced while preserving subject-independent gesture-recognition performance under a prespecified engineering tolerance. The principal finding is deliberately modest: four electrodes could be removed, whereas five could not. The smallest non-inferior configuration therefore contained 12 electrodes, representing a 25% reduction from the complete array. Importantly, this result was not driven by a single irregular point on the performance curve. All configurations from 12 to 15 electrodes remained within the non-inferiority margin, and the minimum acceptable count was unchanged after correction for multiple testing. At the same time, every reduction produced some numerical loss in Macro-F1, and halving the array incurred a substantially larger decrease. The full montage therefore contains some redundancy, but considerably less than might be inferred from user-dependent electrode-selection studies.




Figure 9. Exploratory analyses. (a) GRABMyo cross-session evaluation: each point is one held-out participant, grey lines join the same participant across conditions and bars mark medians; every condition is also subject-independent. (b–c) Mean paired difference to the full array of the same curve, $\bar{\Delta}_k$, with 95% cluster-bootstrap confidence interval, for (b) NinaPro DB5 (10 subjects, reference $k = 16$), (c) GRABMyo derived bipolar pairs (43 participants, reference $k = 8$). The dashed line marks the primary margin; filled markers indicate a lower bound above the margin and open markers a lower bound at or below it.

This distinction between user-dependent and subject-independent reduction is central to the interpretation of the result. A sparse configuration optimised for a particular individual can exploit that person's anatomy, electrode placement and activation patterns. A layout intended for previously unseen users must instead preserve information across inter-individual variability. The present result therefore should not be interpreted as contradicting studies reporting successful control with substantially fewer electrodes (Phinyomark and Scheme, 2018; Qu et al., 2021); rather, it addresses a different design question. Under the current montage, gesture set and cross-user protocol, a moderate reduction was supported, whereas aggressive sparsification was not.

Cross-fitted attribution provided useful information for selecting which electrodes to retain, but its benefit was more limited than a comparison with random subsets alone might suggest. Attribution-selected configurations outperformed the mean of random configurations at every tested subset size, and deletion with retraining showed that removing highly ranked electrodes was more damaging than removing weakly ranked ones. Together, these findings support the functional relevance of the ranking. However, the attribution-based configuration was rarely among the very best random subsets for an individual participant, and exploratory comparisons showed no statistically detectable advantage over permutation importance or an evenly spaced geometric layout at most tested subset sizes. The clearest separation occurred at six electrodes, where the attribution-guided ranking scored higher than both the geometric layout and permutation importance in uncorrected exploratory contrasts.

This pattern suggests that two effects may be operating simultaneously. The forearm montage appears to contain broad spatial redundancy, making circumferential coverage an effective baseline when a relatively large number of electrodes is retained. At six electrodes, the identity of individual recording locations appeared to matter more. No advantage over the geometric layout was detected at four electrodes, however, so the benefit of attribution did not simply increase as the array was reduced. This interpretation remains exploratory: the study did not formally test equivalence between attribution-based and geometric layouts, nor did it test whether the geometric layout itself satisfies

the non-inferiority criterion. The absence of a statistically detectable difference should therefore not be interpreted as evidence that the two approaches are interchangeable.

A second important result is negative. Penalising attribution instability did not improve reduced-array prediction. The cross-fitted attribution profiles were already highly similar across inner participant groups, leaving little instability for the regularisation term to penalise; the penalty changed the top-12 set in at most one of 43 folds. Consequently, the stability-adjusted rankings were almost identical to the rankings based on attribution magnitude alone. The post-hoc participant-level extension made between-person variability more visible but reached the same practical conclusion: a more sensitive estimate of attribution heterogeneity did not translate into improved electrode selection. Within the present dataset and sample size, the tested stability formulation therefore added no detectable ranking information beyond plain cross-fitted TreeSHAP magnitude, which is the simpler option. This result does not exclude a role for stability regularisation (Xiang et al., 2026) in smaller cohorts, more heterogeneous populations, or settings in which attribution estimates vary more strongly across subjects.

The spatial consistency of the ranking also warrants careful interpretation. Several electrode positions were repeatedly ranked near the top across outer folds and individual participants, indicating reproducible regional differences in discriminative information. However, these positions cannot be interpreted directly as specific muscle contributions. GRABMyo documents the electrode montage more reliably than the underlying muscle boundaries, no imaging-based electrode-to-muscle registration was available, and each surface recording reflects volume-conducted activity from multiple sources (Farina et al., 2004). The present results therefore support a positional interpretation (some circumferential regions are consistently more informative than others) but not muscle-level localisation.

The exploratory analyses further define the scope of the primary result. NinaPro DB5 did not provide a decisive external replication because its ten participants yielded wide uncertainty and the candidate non-inferior configuration did not survive multiplicity correction. Similarly, the cross-session analysis did not reveal an additional performance penalty on later recording days, but this should not be interpreted as evidence that electrode displacement or repeated donning is inconsequential (Hargrove et al., 2006; Kaufmann et al., 2010; Vidovic et al., 2016). All conditions remained subject-independent, and training on multiple sessions also increased the amount of available training data. Most importantly, the bipolar re-referencing analysis did not reproduce the degree of reduction observed for monopolar electrodes. Removing even one of eight derived bipolar pairs failed the primary non-inferiority criterion. The headline 25% reduction is therefore specific to monopolar recording sites on the analysed two-ring montage and should not be translated directly into a claim about reducing physical bipolar sensors.

Several limitations constrain the generality of these findings. All participants were able-bodied, and the analysis was entirely offline. Zero-phase filtering is non-causal, and the present evaluation therefore does not establish real-time control performance. Effects of fatigue, posture, electrode displacement and repeated donning were not isolated. The electrode rankings were derived from models trained on the complete array and may differ from rankings learned directly under sparse acquisition. Computational constraints also required reduced training samples for some classifier and selection-method comparisons. The support vector machine result is therefore likely a lower bound, and the H2 and H4 comparisons used a lighter model than the performance curve. Training-sample sizes were chosen on two folds using held-out outcomes, and λ was not selected. In addition, several secondary test families and exploratory analyses were specified after the primary results were available, as documented in supplementary methods S1, and should be interpreted accordingly; the fixed-layout result (supplementary methods S4) is a single-split consistency check only, and the window-length and 17-class analyses listed in the plan were not run. Finally, the independent replication dataset was small, and the bipolar analysis used signals derived mathematically from monopolar recordings rather than measurements from a physical bipolar montage.

Taken together, the findings support a restrained design conclusion rather than a claim for aggressive electrode minimisation. For subject-independent recognition, the relevant question is not whether some sparse configuration can perform well for particular users, but whether the performance loss associated with a common reduced configuration remains within a tolerance defined before observing the result. Under that criterion, this study supports a moderate reduction from 16 to 12 monopolar electrodes. Cross-fitted attribution offers a leakage-controlled way to identify informative sites, but the results also show that its practical advantage depends on how strongly the array is reduced and may be modest when broad spatial coverage is already preserved.

5 Conclusion

For offline subject-independent gesture recognition with the two-ring monopolar GRABMyo montage, reducing the array from 16 to 12 electrodes remained within a prespecified 0.02 Macro-F1 non-inferiority margin, whereas further reduction did not. Cross-fitted attribution identified electrodes whose removal materially affected retrained-model performance and consistently outperformed average random selection, although its advantage over simple spatial selection was limited at most tested subset sizes. Stability regularisation did not provide additional benefit under the present protocol. These findings support moderate, evidence-based electrode reduction rather than aggressive sparsification and highlight the importance of evaluating electrode selection at the participant level under prespecified performance tolerances. Validation in target clinical populations, physical reduced-sensor hardware and online causal-control settings remains necessary before the result can be translated into deployment recommendations.

Acknowledgments

[AUTHOR INPUT REQUIRED: acknowledgements, if any.]

Funding

[AUTHOR INPUT REQUIRED: funding source and grant number, or 'This research received no specific grant from any funding agency in the public, commercial or not-for-profit sectors.' only if true.]

Author contributions

[AUTHOR INPUT REQUIRED: contributions of Minh-Tri Ta, Hoang-Khang Phan and Le-Giang Tran using CRediT roles.]

Conflict of interest

[AUTHOR INPUT REQUIRED: confirm the conflict-of-interest declaration.]

Data availability

GRABMyo v1.1.0 is available from PhysioNet (Jiang et al., 2024) (`https://doi.org/10.13026/89dm-f662`); NinaPro DB5 from the NinaPro repository (`https://ninapro.hevs.ch`). The statistical analysis plan, the record of analysis decisions, the analysis code, the environment specification and all result files from which the numbers in this paper are computed are archived at [AUTHOR INPUT REQUIRED: INSERT PERMANENT REPOSITORY DOI].

Supplementary data

Supplementary methods (departures from the analysis plan, threshold checks, deletion-analysis consistency, fixed-layout and feature-redundancy analyses) and supplementary tables are provided in a separate file.

References

Al-Timemy A H, Bugmann G, Escudero J and Outram N 2013 *IEEE Journal of Biomedical and Health Informatics* **17**(3), 608–618.

Campbell H and Gustafson P 2018 'What to make of non-inferiority and equivalence testing with a post-specified margin?' arXiv preprint arXiv:1807.03413.

Cote-Allard U, Fall C L, Drouin A, Campeau-Lecours A, Gosselin C, Glette K, Laviolette F and Gosselin B 2019 *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **27**(4), 760–771.

Englehart K and Hudgins B 2003 *IEEE Transactions on Biomedical Engineering* **50**(7), 848–854.

Farina D, Jiang N, Rehbaum H, Holobar A, Graimann B, Dietl H and Aszmann O C 2014 *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **22**(4), 797–809.

Farina D, Merletti R and Enoka R M 2004 *Journal of Applied Physiology* **96**(4), 1486–1495.

Goldberger A L, Amaral L A N, Glass L, Hausdorff J M, Ivanov P C, Mark R G, Mietus J E, Moody G B, Peng C K and Stanley H E 2000 *Circulation* **101**(23).

Hargrove L, Englehart K and Hudgins B 2006 *in* '2006 International Conference of the IEEE Engineering in Medicine and Biology Society'.

Hooker S, Erhan D, Kindermans P J and Kim B 2019 *in* 'Advances in Neural Information Processing Systems' Vol. 32.

Hudgins B, Parker P and Scott R 1993 *IEEE Transactions on Biomedical Engineering* **40**(1), 82–94.

Jiang N, Pradhan A and He J 2024 'Gesture Recognition and Biometrics ElectroMyogram (GRABMyo) (version 1.1.0)' PhysioNet.

Kaufmann P, Englehart K and Platzner M 2010 *in* '2010 Annual International Conference of the IEEE Engineering in Medicine and Biology' pp. 6357–6360.

Kelodjou G, Rozé L, Masson V, Galárraga L, Gaudel R, Tchuente M and Termier A 2023 'Shaping Up SHAP: Enhancing Stability through Layer-Wise Neighbor Selection' arXiv preprint arXiv:2312.12115.

Liu J, Li X, Li G and Zhou P 2014 *Medical Engineering & Physics* **36**(7), 975–980.

Lundberg S and Lee S I 2017 *in* 'Advances in Neural Information Processing Systems' Vol. 30.

Lundberg S M, Erion G, Chen H, DeGrave A, Prutkin J M, Nair B, Katz R, Himmelfarb J, Bansal N and Lee S I 2020 *Nature Machine Intelligence* **2**(1), 56–67.

Meinshausen N and Bühlmann P 2010 *Journal of the Royal Statistical Society Series B: Statistical Methodology* **72**(4), 417–473.

Phinyomark A, Phukpattaranont P and Limsakul C 2012 *Expert Systems with Applications* **39**(8), 7420–7431.

Phinyomark A and Scheme E 2018 *Big Data and Cognitive Computing* **2**(3), 21.

Pizzolato S, Tagliapietra L, Cognolato M, Reggiani M, Müller H and Atzori M 2017 *PLOS ONE* **12**(10), e0186132.

Pradhan A, He J and Jiang N 2022 *Scientific Data* **9**(1).

Qu Y, Shang H, Li J and Teng S 2021 *Journal of Healthcare Engineering* **2021**, 1–9.

Salmi R, Lu G and Chen Q 2025 'SHAP-AAD: DeepSHAP-Guided Channel Reduction for EEG Auditory Attention Detection' arXiv preprint arXiv:2507.03814.

Samek W, Binder A, Montavon G, Lapuschkin S and Muller K R 2017 *IEEE Transactions on Neural Networks and Learning Systems* **28**(11), 2660–2673.

Scheme E and Englehart K 2011 *The Journal of Rehabilitation Research and Development* **48**(6), 643.

Schuirmann D J 1987 *Journal of Pharmacokinetics and Biopharmaceutics* **15**(6), 657–680.

Stango A, Negro F and Farina D 2015 *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **23**(2), 189–198.

Vidovic M M C, Hwang H J, Amsuss S, Hahne J M, Farina D and Muller K R 2016 *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **24**(9), 961–970.

Visani G, Bagli E, Chesani F, Poluzzi A and Capuzzo D 2022 *Journal of the Operational Research Society* **73**(1), 91–101.

Vujaklija I, Farina D and Aszmann O 2016 *Orthopedic Research and Reviews* **Volume 8**, 31–39.

Walker E and Nowacki A S 2011 *Journal of General Internal Medicine* **26**(2), 192–196.

Xiang L, Shi L, Ye Y, Jiang B, Zhou D and Guo F 2026 'RoSHAP: A Distributional Framework and Robust Metric for Stable Feature Attribution' arXiv preprint arXiv:2605.15154.