


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200119-381  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 19-Jan-20 10:45:17 AM  
Ended: 19-Jan-20 2:51:09 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

### Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	10:45:17 AM	314	0.69	OK	
CC-2	17.5	10:46:17 AM	6393	17.8	OK	

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE L	e1	1	10:49:16 AM	1243	2.89 kUA/I	CZ4AN	OK	
slgE Neg	d1	1	10:50:17 AM	22	0.01 kUA/I	BL1CP	OK	
slgE M	t3	1	11:06:16 AM	4603	12.0 kUA/I	CZ1AR	OK	
IQC271119	d1	1	11:09:16 AM	6488	18.1 kUA/I	<lab>	----	
			Expected value:	----				
			Expected SD:	----				
			SD limit:	----				
			Range min:	----				
			Range max:	----				
			Limit CV%:	----				

## Laboratory Report

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000503	phad	5	2:45:08 PM	36	0.23 PAU/I	0	Negative	
011990000506	phad	5	2:00:10 PM	50	0.39 PAU/I	1	Positive	
011990000507	phad	5	1:52:10 PM	174	1.83 PAU/I	2	Positive	
011990000513	phad	5	2:32:08 PM	50	0.39 PAU/I	1	Positive	
011990000515	phad	5	1:44:10 PM	36	0.23 PAU/I	0	Negative	
011990000516	phad	5	1:36:10 PM	45	0.33 PAU/I	0	Negative	
011990000517	phad	5	10:56:16 AM	102	0.99 PAU/I	2	Positive	
011990000518	phad	5	12:31:13 PM	53	0.43 PAU/I	1	Positive	
011990000522	phad	5	2:16:09 PM	42	0.29 PAU/I	0	Negative	
011990000523	phad	5	2:24:09 PM	160	1.67 PAU/I	2	Positive	
011990000525	phad	5	1:28:11 PM	34	0.20 PAU/I	0	Negative	
011990000526	phad	5	11:07:16 AM	102	0.99 PAU/I	2	Positive	
011990000527	phad	5	11:40:15 AM	35	0.22 PAU/I	0	Negative	
011990000528	phad	5	12:23:13 PM	36	0.23 PAU/I	0	Negative	
011990000534	phad	5	12:47:12 PM	35	0.21 PAU/I	0	Negative	
011990000535	phad	5	1:20:11 PM	37	0.23 PAU/I	0	Negative	
011990000536	phad	5	2:08:09 PM	48	0.37 PAU/I	1	Positive	
011990000537	phad	5	11:32:15 AM	34	0.20 PAU/I	0	Negative	
011990000538	phad	5	12:15:13 PM	30	0.16 PAU/I	0	Negative	
011990000544	phad	5	1:09:11 PM	40	0.28 PAU/I	0	Negative	
011990000545	phad	5	12:55:12 PM	31	0.17 PAU/I	0	Negative	
011990000546	phad	5	12:39:12 PM	33	0.20 PAU/I	0	Negative	
011990000547	phad	5	11:24:15 AM	167	1.75 PAU/I	2	Positive	
011990000548	phad	5	11:16:15 AM	31	0.17 PAU/I	0	Negative	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000503	fx5	5	2:46:08 PM	20	0.04 kUA/I	0	Negative	
011990000503	fx1	5	2:47:08 PM	19	0.03 kUA/I	0	Negative	
011990000503	fx2	5	2:48:08 PM	15	Low RU	0	Negative	
011990000503	fx22	5	2:49:08 PM	29	0.14 kUA/I	0	Negative	
011990000503	gx2	5	2:50:08 PM	19	0.03 kUA/I	0	Negative	
011990000506	fx5	5	2:01:09 PM	24	0.09 kUA/I	0	Negative	
011990000506	fx1	5	2:02:09 PM	16	Low RU	0	Negative	
011990000506	fx2	5	2:03:09 PM	22	0.06 kUA/I	0	Negative	
011990000506	fx22	5	2:04:09 PM	27	0.12 kUA/I	0	Negative	
011990000506	gx2	5	2:05:09 PM	21	0.05 kUA/I	0	Negative	
011990000506	wx1	5	2:06:09 PM	31	0.17 kUA/I	0	Negative	
011990000506	t20	5	2:07:09 PM	19	0.02 kUA/I	0		
011990000507	fx5	5	1:53:10 PM	59	0.50 kUA/I	1	Positive	
011990000507	fx1	5	1:54:10 PM	23	0.07 kUA/I	0	Negative	
011990000507	fx2	5	1:55:09 PM	136	1.39 kUA/I	2	Positive	
011990000507	fx22	5	1:56:10 PM	30	0.15 kUA/I	0	Negative	
011990000507	gx2	5	1:57:10 PM	25	0.10 kUA/I	0	Negative	
011990000507	wx1	5	1:58:10 PM	53	0.43 kUA/I	1	Positive	
011990000507	t20	5	1:59:09 PM	28	0.14 kUA/I	0		
011990000513	fx5	5	2:33:08 PM	24	0.09 kUA/I	0	Negative	
011990000513	fx1	5	2:34:09 PM	18	0.01 kUA/I	0	Negative	
011990000513	fx2	5	2:35:08 PM	14	Low RU	0	Negative	
011990000513	fx22	5	2:41:08 PM	29	0.15 kUA/I	0	Negative	
011990000513	gx2	5	2:42:08 PM	23	0.08 kUA/I	0	Negative	
011990000513	wx1	5	2:43:08 PM	28	0.13 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000513	t20	5	2:44:08 PM	16	Low RU	0		
011990000515	fx5	5	1:45:10 PM	24	0.08 kUA/l	0	Negative	
011990000515	fx1	5	1:46:10 PM	27	0.12 kUA/l	0	Negative	
011990000515	fx2	5	1:47:10 PM	12	Low RU	0	Negative	
011990000515	fx22	5	1:48:10 PM	28	0.13 kUA/l	0	Negative	
011990000515	gx2	5	1:49:10 PM	20	0.04 kUA/l	0	Negative	
011990000515	wx1	5	1:50:10 PM	70	0.63 kUA/l	1	Positive	
011990000515	t20	5	1:51:10 PM	18	0.01 kUA/l	0		
011990000516	fx5	5	1:37:10 PM	22	0.07 kUA/l	0	Negative	
011990000516	fx1	5	1:38:10 PM	19	0.03 kUA/l	0	Negative	
011990000516	fx2	5	1:39:10 PM	21	0.06 kUA/l	0	Negative	
011990000516	fx22	5	1:40:10 PM	29	0.15 kUA/l	0	Negative	
011990000516	gx2	5	1:41:10 PM	24	0.09 kUA/l	0	Negative	
011990000516	wx1	5	1:42:10 PM	30	0.15 kUA/l	0	Negative	
011990000516	t20	5	1:43:10 PM	18	0.02 kUA/l	0		
011990000517	fx5	5	10:57:16 AM	22	0.06 kUA/l	0	Negative	
011990000517	fx1	5	10:58:16 AM	25	0.09 kUA/l	0	Negative	
011990000517	fx2	5	10:59:16 AM	40	0.27 kUA/l	0	Negative	
011990000517	fx22	5	11:00:16 AM	33	0.19 kUA/l	0	Negative	
011990000517	gx2	5	11:01:16 AM	25	0.10 kUA/l	0	Negative	
011990000517	wx1	5	11:02:16 AM	32	0.18 kUA/l	0	Negative	
011990000517	t20	5	11:03:16 AM	19	0.03 kUA/l	0		
011990000518	fx5	5	12:32:13 PM	20	0.04 kUA/l	0	Negative	
011990000518	fx1	5	12:33:13 PM	23	0.07 kUA/l	0	Negative	
011990000518	fx2	5	12:34:12 PM	460	5.15 kUA/l	3	Positive	
011990000518	fx22	5	12:35:13 PM	32	0.17 kUA/l	0	Negative	
011990000518	gx2	5	12:36:12 PM	28	0.14 kUA/l	0	Negative	
011990000518	wx1	5	12:37:12 PM	34	0.20 kUA/l	0	Negative	
011990000518	t20	5	12:38:12 PM	20	0.04 kUA/l	0		
011990000522	fx5	5	2:17:09 PM	47	0.36 kUA/l	1	Positive	
011990000522	fx1	5	2:18:09 PM	22	0.06 kUA/l	0	Negative	
011990000522	fx2	5	2:19:09 PM	20	0.04 kUA/l	0	Negative	
011990000522	fx22	5	2:20:09 PM	33	0.20 kUA/l	0	Negative	
011990000522	gx2	5	2:21:09 PM	24	0.09 kUA/l	0	Negative	
011990000522	wx1	5	2:22:09 PM	33	0.19 kUA/l	0	Negative	
011990000522	t20	5	2:23:09 PM	18	0.01 kUA/l	0		
011990000523	fx5	5	2:25:09 PM	67	0.59 kUA/l	1	Positive	
011990000523	fx1	5	2:26:09 PM	22	0.06 kUA/l	0	Negative	
011990000523	fx2	5	2:27:09 PM	16	Low RU	0	Negative	
011990000523	fx22	5	2:28:08 PM	31	0.16 kUA/l	0	Negative	
011990000523	gx2	5	2:29:08 PM	25	0.09 kUA/l	0	Negative	
011990000523	wx1	5	2:30:08 PM	31	0.17 kUA/l	0	Negative	
011990000523	t20	5	2:31:08 PM	17	Low RU	0		
011990000525	fx5	5	1:29:11 PM	19	0.02 kUA/l	0	Negative	
011990000525	fx1	5	1:30:11 PM	19	0.02 kUA/l	0	Negative	
011990000525	fx2	5	1:31:11 PM	13	Low RU	0	Negative	
011990000525	fx22	5	1:32:11 PM	28	0.14 kUA/l	0	Negative	
011990000525	gx2	5	1:33:10 PM	22	0.06 kUA/l	0	Negative	
011990000525	wx1	5	1:34:10 PM	28	0.13 kUA/l	0	Negative	
011990000525	t20	5	1:35:10 PM	17	0.01 kUA/l	0		
011990000526	fx5	5	11:08:16 AM	22	0.06 kUA/l	0	Negative	
011990000526	fx1	5	11:10:16 AM	19	0.02 kUA/l	0	Negative	
011990000526	fx2	5	11:11:16 AM	155	1.61 kUA/l	2	Positive	
011990000526	fx22	5	11:12:16 AM	31	0.17 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000526	gx2	5	11:13:16 AM	23	0.07 kUA/l	0	Negative	
011990000526	wx1	5	11:14:16 AM	32	0.18 kUA/l	0	Negative	
011990000526	t20	5	11:15:15 AM	21	0.05 kUA/l	0		
011990000527	fx5	5	11:41:15 AM	62	0.54 kUA/l	1	Positive	
011990000527	fx1	5	11:51:14 AM	23	0.07 kUA/l	0	Negative	
011990000527	fx2	5	11:52:14 AM	16	Low RU	0	Negative	
011990000527	fx22	5	11:53:14 AM	36	0.23 kUA/l	0	Negative	
011990000527	gx2	5	11:54:14 AM	35	0.21 kUA/l	0	Negative	
011990000527	wx1	5	12:13:13 PM	31	0.17 kUA/l	0	Negative	
011990000527	t20	5	12:14:14 PM	20	0.04 kUA/l	0		
011990000528	fx5	5	12:24:13 PM	19	0.03 kUA/l	0	Negative	
011990000528	fx1	5	12:25:13 PM	22	0.06 kUA/l	0	Negative	
011990000528	fx2	5	12:26:13 PM	16	Low RU	0	Negative	
011990000528	fx22	5	12:27:13 PM	31	0.17 kUA/l	0	Negative	
011990000528	gx2	5	12:28:13 PM	22	0.06 kUA/l	0	Negative	
011990000528	wx1	5	12:29:13 PM	268	2.91 kUA/l	2	Positive	
011990000528	t20	5	12:30:13 PM	29	0.14 kUA/l	0		
011990000534	fx5	5	12:48:12 PM	24	0.08 kUA/l	0	Negative	
011990000534	fx1	5	12:49:12 PM	24	0.09 kUA/l	0	Negative	
011990000534	fx2	5	12:50:12 PM	15	Low RU	0	Negative	
011990000534	fx22	5	12:51:12 PM	32	0.18 kUA/l	0	Negative	
011990000534	gx2	5	12:52:12 PM	23	0.08 kUA/l	0	Negative	
011990000534	wx1	5	12:53:12 PM	39	0.26 kUA/l	0	Negative	
011990000534	t20	5	12:54:12 PM	17	0.01 kUA/l	0		
011990000535	fx5	5	1:21:11 PM	24	0.08 kUA/l	0	Negative	
011990000535	fx1	5	1:22:11 PM	18	0.02 kUA/l	0	Negative	
011990000535	fx2	5	1:23:11 PM	14	Low RU	0	Negative	
011990000535	fx22	5	1:24:11 PM	28	0.14 kUA/l	0	Negative	
011990000535	gx2	5	1:25:11 PM	21	0.05 kUA/l	0	Negative	
011990000535	wx1	5	1:26:11 PM	29	0.14 kUA/l	0	Negative	
011990000535	t20	5	1:27:11 PM	18	0.02 kUA/l	0		
011990000536	fx5	5	2:09:09 PM	27	0.12 kUA/l	0	Negative	
011990000536	fx1	5	2:10:09 PM	23	0.08 kUA/l	0	Negative	
011990000536	fx2	5	2:11:09 PM	42	0.30 kUA/l	0	Negative	
011990000536	fx22	5	2:12:09 PM	39	0.27 kUA/l	0	Negative	
011990000536	gx2	5	2:13:09 PM	44	0.32 kUA/l	0	Negative	
011990000536	wx1	5	2:14:09 PM	38	0.25 kUA/l	0	Negative	
011990000536	t20	5	2:15:09 PM	22	0.06 kUA/l	0		
011990000537	fx5	5	11:33:15 AM	19	0.03 kUA/l	0	Negative	
011990000537	fx1	5	11:34:15 AM	16	Low RU	0	Negative	
011990000537	fx2	5	11:35:15 AM	19	0.02 kUA/l	0	Negative	
011990000537	fx22	5	11:36:15 AM	29	0.14 kUA/l	0	Negative	
011990000537	gx2	5	11:37:15 AM	22	0.06 kUA/l	0	Negative	
011990000537	wx1	5	11:38:15 AM	28	0.14 kUA/l	0	Negative	
011990000537	t20	5	11:39:15 AM	18	0.01 kUA/l	0		
011990000538	fx5	5	12:16:13 PM	20	0.03 kUA/l	0	Negative	
011990000538	fx1	5	12:17:13 PM	16	Low RU	0	Negative	
011990000538	fx2	5	12:18:13 PM	12	Low RU	0	Negative	
011990000538	fx22	5	12:19:13 PM	27	0.12 kUA/l	0	Negative	
011990000538	gx2	5	12:20:13 PM	23	0.07 kUA/l	0	Negative	
011990000538	wx1	5	12:21:13 PM	26	0.11 kUA/l	0	Negative	
011990000538	t20	5	12:22:13 PM	16	Low RU	0		
011990000544	fx5	5	1:10:12 PM	23	0.08 kUA/l	0	Negative	
011990000544	fx1	5	1:11:11 PM	19	0.03 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000544	fx2	5	1:12:11 PM	13	Low RU	0	Negative	
011990000544	fx22	5	1:13:11 PM	32	0.18 kUA/I	0	Negative	
011990000544	gx2	5	1:14:11 PM	21	0.05 kUA/I	0	Negative	
011990000544	wx1	5	1:15:11 PM	28	0.14 kUA/I	0	Negative	
011990000544	t20	5	1:16:11 PM	17	0.01 kUA/I	0		
011990000545	fx5	5	12:56:12 PM	20	0.04 kUA/I	0	Negative	
011990000545	fx1	5	12:57:12 PM	16	Low RU	0	Negative	
011990000545	fx2	5	1:04:11 PM	13	Low RU	0	Negative	
011990000545	fx22	5	1:05:11 PM	30	0.16 kUA/I	0	Negative	
011990000545	gx2	5	1:06:11 PM	21	0.05 kUA/I	0	Negative	
011990000545	wx1	5	1:07:11 PM	28	0.13 kUA/I	0	Negative	
011990000545	t20	5	1:08:11 PM	17	Low RU	0		
011990000546	fx5	5	12:40:12 PM	23	0.07 kUA/I	0	Negative	
011990000546	fx1	5	12:41:12 PM	17	0.01 kUA/I	0	Negative	
011990000546	fx2	5	12:42:12 PM	15	Low RU	0	Negative	
011990000546	fx22	5	12:43:12 PM	32	0.18 kUA/I	0	Negative	
011990000546	gx2	5	12:44:12 PM	21	0.05 kUA/I	0	Negative	
011990000546	wx1	5	12:45:12 PM	29	0.14 kUA/I	0	Negative	
011990000546	t20	5	12:46:12 PM	16	Low RU	0		
011990000547	fx5	5	11:25:15 AM	21	0.05 kUA/I	0	Negative	
011990000547	fx1	5	11:26:15 AM	17	Low RU	0	Negative	
011990000547	fx2	5	11:27:16 AM	99	0.96 kUA/I	2	Positive	
011990000547	fx22	5	11:28:15 AM	34	0.21 kUA/I	0	Negative	
011990000547	gx2	5	11:29:15 AM	31	0.17 kUA/I	0	Negative	
011990000547	wx1	5	11:30:15 AM	370	4.11 kUA/I	3	Positive	
011990000547	t20	5	11:31:15 AM	50	0.39 kUA/I	1		
011990000548	fx5	5	11:17:15 AM	20	0.03 kUA/I	0	Negative	
011990000548	fx1	5	11:18:16 AM	23	0.08 kUA/I	0	Negative	
011990000548	fx2	5	11:19:16 AM	17	Low RU	0	Negative	
011990000548	fx22	5	11:20:15 AM	33	0.19 kUA/I	0	Negative	
011990000548	gx2	5	11:21:15 AM	22	0.06 kUA/I	0	Negative	
011990000548	wx1	5	11:22:15 AM	38	0.25 kUA/I	0	Negative	
011990000548	t20	5	11:23:15 AM	16	Low RU	0		
011990000918	fx1	5	12:58:12 PM	22	0.06 kUA/I	0	Negative	
011990000918	fx2	5	12:59:12 PM	13	Low RU	0	Negative	
011990000918	fx22	5	1:00:12 PM	32	0.18 kUA/I	0	Negative	
011990000918	gx2	5	1:01:12 PM	20	0.04 kUA/I	0	Negative	
011990000918	wx1	5	1:02:12 PM	86	0.81 kUA/I	2	Positive	
011990000918	t20	5	1:03:12 PM	17	Low RU	0		
012001400615A	f233	1	2:38:08 PM	191	0.41 kUA/I			
012001500339A	f202	1	2:40:08 PM	109	0.21 kUA/I			
012001500339A	f203	1	2:37:08 PM	131	0.27 kUA/I			
012001500838A	d1	1	2:36:08 PM	21	0.01 kUA/I			

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	



# Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	10:45:17 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	10:46:17 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE L	e1	10:49:16 AM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06!		Exp
slgE Neg	d1	10:50:17 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12!		Exp
slgE M	t3	11:06:16 AM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
IQC271119	d1	11:09:16 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000918	fx1	12:58:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000918	fx2	12:59:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000918	fx22	1:00:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000918	gx2	1:01:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000918	wx1	1:02:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000918	t20	1:03:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000503	phad	2:45:08 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000503	fx5	2:46:08 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000503	fx1	2:47:08 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000503	fx2	2:48:08 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000503	fx22	2:49:08 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000503	gx2	2:50:08 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	phad	2:32:08 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	fx5	2:33:08 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	fx1	2:34:09 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	fx2	2:35:08 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	fx22	2:41:08 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	gx2	2:42:08 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	wx1	2:43:08 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000513	t20	2:44:08 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	phad	2:24:09 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	fx5	2:25:09 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	fx1	2:26:09 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	fx2	2:27:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	fx22	2:28:08 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	gx2	2:29:08 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	wx1	2:30:08 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000523	t20	2:31:08 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	phad	2:16:09 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	fx5	2:17:09 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	fx1	2:18:09 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	fx2	2:19:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	fx22	2:20:09 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	gx2	2:21:09 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	wx1	2:22:09 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000522	t20	2:23:09 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	phad	2:08:09 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	fx5	2:09:09 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	fx1	2:10:09 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	fx2	2:11:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	fx22	2:12:09 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	gx2	2:13:09 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	wx1	2:14:09 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000536	t20	2:15:09 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000506	phad	2:00:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000506	fx5	2:01:09 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000506	fx1	2:02:09 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000506	fx2	2:03:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000506	fx22	2:04:09 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000506	gx2	2:05:09 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000506	wx1	2:06:09 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000506	t20	2:07:09 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	phad	1:52:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	fx5	1:53:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	fx1	1:54:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	fx2	1:55:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	fx22	1:56:10 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	gx2	1:57:10 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	wx1	1:58:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000507	t20	1:59:09 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	phad	1:44:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	fx5	1:45:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	fx1	1:46:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	fx2	1:47:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	fx22	1:48:10 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	gx2	1:49:10 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	wx1	1:50:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000515	t20	1:51:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	phad	1:36:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	fx5	1:37:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	fx1	1:38:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	fx2	1:39:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	fx22	1:40:10 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	gx2	1:41:10 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	wx1	1:42:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000516	t20	1:43:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	phad	1:28:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	fx5	1:29:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	fx1	1:30:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	fx2	1:31:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	fx22	1:32:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	gx2	1:33:10 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	wx1	1:34:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000525	t20	1:35:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	phad	1:20:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	fx5	1:21:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	fx1	1:22:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	fx2	1:23:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	fx22	1:24:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	gx2	1:25:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	wx1	1:26:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000535	t20	1:27:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	phad	1:09:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	fx5	1:10:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	fx1	1:11:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	fx2	1:12:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	fx22	1:13:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	gx2	1:14:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	wx1	1:15:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000544	t20	1:16:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	phad	12:55:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	fx5	12:56:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	fx1	12:57:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	fx2	1:04:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	fx22	1:05:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	gx2	1:06:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	wx1	1:07:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000545	t20	1:08:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000534	phad	12:47:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000534	fx5	12:48:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000534	fx1	12:49:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000534	fx2	12:50:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000534	fx22	12:51:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000534	gx2	12:52:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000534	wx1	12:53:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000534	t20	12:54:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	phad	12:39:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	fx5	12:40:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	fx1	12:41:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	fx2	12:42:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	fx22	12:43:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	gx2	12:44:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	wx1	12:45:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000546	t20	12:46:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	phad	12:31:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	fx5	12:32:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	fx1	12:33:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	fx2	12:34:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	fx22	12:35:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	gx2	12:36:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	wx1	12:37:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000518	t20	12:38:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	phad	12:23:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	fx5	12:24:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	fx1	12:25:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	fx2	12:26:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	fx22	12:27:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	gx2	12:28:13 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	wx1	12:29:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000528	t20	12:30:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	phad	12:15:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	fx5	12:16:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	fx1	12:17:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	fx2	12:18:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	fx22	12:19:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	gx2	12:20:13 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	wx1	12:21:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000538	t20	12:22:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	phad	11:40:15 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	fx5	11:41:15 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	fx1	11:51:14 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	fx2	11:52:14 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	fx22	11:53:14 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	gx2	11:54:14 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	wx1	12:13:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000527	t20	12:14:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	phad	11:32:15 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	fx5	11:33:15 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	fx1	11:34:15 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	fx2	11:35:15 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	fx22	11:36:15 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	gx2	11:37:15 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	wx1	11:38:15 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000537	t20	11:39:15 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	phad	11:24:15 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	fx5	11:25:15 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	fx1	11:26:15 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	fx2	11:27:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	fx22	11:28:15 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	gx2	11:29:15 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	wx1	11:30:15 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000547	t20	11:31:15 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000548	phad	11:16:15 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000548	fx5	11:17:15 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000548	fx1	11:18:16 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000548	fx2	11:19:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000548	fx22	11:20:15 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000548	gx2	11:21:15 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000548	wx1	11:22:15 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000548	t20	11:23:15 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	phad	11:07:16 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	fx5	11:08:16 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	fx1	11:10:16 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	fx2	11:11:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	fx22	11:12:16 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	gx2	11:13:16 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	wx1	11:14:16 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000526	t20	11:15:15 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	phad	10:56:16 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	fx5	10:57:16 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	fx1	10:58:16 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	fx2	10:59:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	fx22	11:00:16 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	gx2	11:01:16 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	wx1	11:02:16 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000517	t20	11:03:16 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
012001400615A	f233	2:38:08 PM	904LK 21/03	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500339A	f202	2:40:08 PM	997KA 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500339A	f203	2:37:08 PM	998HG 20/11	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500838A	d1	2:36:08 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			

### Errors

#### Expired lot numbers used for Quality Control

CZ4AN	Quality control vial
BL1CP	Quality control vial
352D0	ImmunoCAP


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200119-381  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 19-Jan-20 2:51:08 PM  
Ended: 19-Jan-20 9:04:14 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

### Curve Controls

No curve controls in this run

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	5:31:02 PM	6129	16.9 kUA/l	<lab>	----	
			Expected value:	---				
			Expected SD:	---				
			SD limit:	---				
			Range min:	---				
			Range max:	---				
			Limit CV%:	---				

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
----------	------	-----	----------	------	------	-------	---------	---------

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000143	phad	5	4:45:03 PM	36	0.23 PAU/I	0	Negative	
011990000144	phad	5	4:53:03 PM	217	2.33 PAU/I	2	Positive	
011990000145	phad	5	4:29:04 PM	32	0.17 PAU/I	0	Negative	
011990000146	phad	5	4:37:04 PM	42	0.30 PAU/I	0	Negative	
011990000147	phad	5	4:13:05 PM	43	0.31 PAU/I	0	Negative	
011990000148	phad	5	4:21:05 PM	38	0.25 PAU/I	0	Negative	
011990000149	phad	5	3:57:05 PM	52	0.42 PAU/I	1	Positive	
011990000150	phad	5	4:05:05 PM	219	2.35 PAU/I	2	Positive	
011990000504	phad	5	3:17:07 PM	34	0.21 PAU/I	0	Negative	
011990000505	phad	5	3:25:06 PM	87	0.82 PAU/I	2	Positive	
011990000514	phad	5	3:09:07 PM	35	0.21 PAU/I	0	Negative	
011990000524	phad	5	3:33:06 PM	447	5.00 PAU/I	3	Positive	
011990000533	phad	5	2:53:08 PM	39	0.26 PAU/I	0	Negative	
011990000543	phad	5	3:01:07 PM	31	0.17 PAU/I	0	Negative	
011990000553	phad	5	3:41:06 PM	82	0.77 PAU/I	2	Positive	
011990000554	phad	5	3:49:06 PM	32	0.18 PAU/I	0	Negative	
011990000558	phad	5	5:29:02 PM	607	6.86 PAU/I	3	Positive	
011990000559	phad	5	5:13:03 PM	40	0.28 PAU/I	0	Negative	
011990000560	phad	5	5:46:01 PM	973	11.2 PAU/I	3	Positive	
011990000566	phad	5	5:21:02 PM	61	0.52 PAU/I	1	Positive	
011990000568	phad	5	5:38:01 PM	30	0.16 PAU/I	0	Negative	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011934500583A	t3	1	5:55:01 PM	7980	23.7 kUA/I			
011934900953A	t3	1	5:54:01 PM	8562	26.0 kUA/I			
011935100949A	t3	1	5:56:01 PM	21222	>100			
011990000143	fx5	5	4:46:04 PM	24	0.09 kUA/I	0	Negative	
011990000143	fx1	5	4:47:04 PM	15	Low RU	0	Negative	
011990000143	fx2	5	4:48:03 PM	12	Low RU	0	Negative	
011990000143	fx22	5	4:49:03 PM	28	0.14 kUA/I	0	Negative	
011990000143	gx2	5	4:50:03 PM	188	1.99 kUA/I	2	Positive	
011990000143	wx1	5	4:51:03 PM	30	0.16 kUA/I	0	Negative	
011990000143	t20	5	4:52:03 PM	18	0.01 kUA/I	0		
011990000144	fx5	5	4:54:03 PM	48	0.37 kUA/I	1	Positive	
011990000144	fx1	5	4:55:04 PM	19	0.02 kUA/I	0	Negative	
011990000144	fx2	5	4:56:03 PM	30	0.16 kUA/I	0	Negative	
011990000144	fx22	5	4:57:03 PM	34	0.20 kUA/I	0	Negative	
011990000144	gx2	5	4:58:03 PM	24	0.09 kUA/I	0	Negative	
011990000144	wx1	5	4:59:03 PM	216	2.32 kUA/I	2	Positive	
011990000144	t20	5	5:00:03 PM	19	0.02 kUA/I	0		
011990000145	fx5	5	4:30:04 PM	22	0.06 kUA/I	0	Negative	
011990000145	fx1	5	4:31:04 PM	19	0.02 kUA/I	0	Negative	
011990000145	fx2	5	4:32:04 PM	13	Low RU	0	Negative	
011990000145	fx22	5	4:33:04 PM	28	0.13 kUA/I	0	Negative	
011990000145	gx2	5	4:34:04 PM	22	0.06 kUA/I	0	Negative	
011990000145	wx1	5	4:35:04 PM	28	0.13 kUA/I	0	Negative	
011990000145	t20	5	4:36:04 PM	16	Low RU	0		
011990000146	fx5	5	4:38:04 PM	52	0.42 kUA/I	1	Positive	
011990000146	fx1	5	4:39:04 PM	25	0.10 kUA/I	0	Negative	
011990000146	fx2	5	4:40:04 PM	20	0.03 kUA/I	0	Negative	
011990000146	fx22	5	4:41:04 PM	33	0.20 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000146	gx2	5	4:42:04 PM	25	0.10 kUA/l	0	Negative	
011990000146	wx1	5	4:43:04 PM	32	0.18 kUA/l	0	Negative	
011990000146	t20	5	4:44:03 PM	20	0.04 kUA/l	0		
011990000147	fx5	5	4:14:05 PM	46	0.35 kUA/l	0	Negative	
011990000147	fx1	5	4:15:05 PM	24	0.09 kUA/l	0	Negative	
011990000147	fx2	5	4:16:05 PM	23	0.07 kUA/l	0	Negative	
011990000147	fx22	5	4:17:05 PM	30	0.16 kUA/l	0	Negative	
011990000147	gx2	5	4:18:05 PM	22	0.07 kUA/l	0	Negative	
011990000147	wx1	5	4:19:05 PM	27	0.12 kUA/l	0	Negative	
011990000147	t20	5	4:20:05 PM	18	0.01 kUA/l	0		
011990000148	fx5	5	4:22:04 PM	19	0.03 kUA/l	0	Negative	
011990000148	fx1	5	4:23:05 PM	18	0.01 kUA/l	0	Negative	
011990000148	fx2	5	4:24:04 PM	19	0.02 kUA/l	0	Negative	
011990000148	fx22	5	4:25:04 PM	30	0.15 kUA/l	0	Negative	
011990000148	gx2	5	4:26:04 PM	21	0.05 kUA/l	0	Negative	
011990000148	wx1	5	4:27:05 PM	26	0.11 kUA/l	0	Negative	
011990000148	t20	5	4:28:04 PM	18	0.01 kUA/l	0		
011990000149	fx5	5	3:58:05 PM	36	0.23 kUA/l	0	Negative	
011990000149	fx1	5	3:59:05 PM	19	0.02 kUA/l	0	Negative	
011990000149	fx2	5	4:00:06 PM	23	0.07 kUA/l	0	Negative	
011990000149	fx22	5	4:01:05 PM	32	0.18 kUA/l	0	Negative	
011990000149	gx2	5	4:02:05 PM	22	0.06 kUA/l	0	Negative	
011990000149	wx1	5	4:03:05 PM	41	0.29 kUA/l	0	Negative	
011990000149	t20	5	4:04:05 PM	25	0.10 kUA/l	0		
011990000150	fx5	5	4:06:05 PM	23	0.07 kUA/l	0	Negative	
011990000150	fx1	5	4:07:05 PM	18	0.01 kUA/l	0	Negative	
011990000150	fx2	5	4:08:05 PM	16	Low RU	0	Negative	
011990000150	fx22	5	4:09:05 PM	33	0.19 kUA/l	0	Negative	
011990000150	gx2	5	4:10:05 PM	25	0.10 kUA/l	0	Negative	
011990000150	wx1	5	4:11:05 PM	34	0.20 kUA/l	0	Negative	
011990000150	t20	5	4:12:05 PM	18	0.02 kUA/l	0		
011990000503	wx1	5	2:51:08 PM	30	0.16 kUA/l	0	Negative	
011990000503	t20	5	2:52:08 PM	18	0.02 kUA/l	0		
011990000504	fx5	5	3:18:07 PM	20	0.04 kUA/l	0	Negative	
011990000504	fx1	5	3:19:06 PM	19	0.03 kUA/l	0	Negative	
011990000504	fx2	5	3:20:07 PM	27	0.12 kUA/l	0	Negative	
011990000504	fx22	5	3:21:07 PM	29	0.14 kUA/l	0	Negative	
011990000504	gx2	5	3:22:07 PM	22	0.06 kUA/l	0	Negative	
011990000504	wx1	5	3:23:07 PM	26	0.10 kUA/l	0	Negative	
011990000504	t20	5	3:24:06 PM	17	Low RU	0		
011990000505	fx5	5	3:26:06 PM	30	0.16 kUA/l	0	Negative	
011990000505	fx1	5	3:27:06 PM	23	0.07 kUA/l	0	Negative	
011990000505	fx2	5	3:28:06 PM	159	1.66 kUA/l	2	Positive	
011990000505	fx22	5	3:29:06 PM	38	0.25 kUA/l	0	Negative	
011990000505	gx2	5	3:30:06 PM	30	0.16 kUA/l	0	Negative	
011990000505	wx1	5	3:31:07 PM	35	0.21 kUA/l	0	Negative	
011990000505	t20	5	3:32:06 PM	21	0.05 kUA/l	0		
011990000514	fx5	5	3:10:07 PM	23	0.07 kUA/l	0	Negative	
011990000514	fx1	5	3:11:07 PM	18	0.02 kUA/l	0	Negative	
011990000514	fx2	5	3:12:07 PM	12	Low RU	0	Negative	
011990000514	fx22	5	3:13:07 PM	27	0.13 kUA/l	0	Negative	
011990000514	gx2	5	3:14:07 PM	21	0.06 kUA/l	0	Negative	
011990000514	wx1	5	3:15:07 PM	30	0.16 kUA/l	0	Negative	
011990000514	t20	5	3:16:07 PM	17	Low RU	0		



## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000524	fx5	5	3:34:06 PM	98	0.95 kUA/l	2	Positive	
011990000524	fx1	5	3:35:06 PM	20	0.04 kUA/l	0	Negative	
011990000524	fx2	5	3:36:06 PM	15	Low RU	0	Negative	
011990000524	fx22	5	3:37:06 PM	35	0.22 kUA/l	0	Negative	
011990000524	gx2	5	3:38:06 PM	24	0.09 kUA/l	0	Negative	
011990000524	wx1	5	3:39:06 PM	133	1.35 kUA/l	2	Positive	
011990000524	t20	5	3:40:06 PM	18	0.01 kUA/l	0		
011990000533	fx5	5	2:54:08 PM	19	0.03 kUA/l	0	Negative	
011990000533	fx1	5	2:55:08 PM	18	0.01 kUA/l	0	Negative	
011990000533	fx2	5	2:56:07 PM	13	Low RU	0	Negative	
011990000533	fx22	5	2:57:07 PM	29	0.15 kUA/l	0	Negative	
011990000533	gx2	5	2:58:07 PM	18	0.01 kUA/l	0	Negative	
011990000533	wx1	5	2:59:07 PM	31	0.17 kUA/l	0	Negative	
011990000533	t20	5	3:00:07 PM	17	Low RU	0		
011990000543	fx5	5	3:02:07 PM	21	0.04 kUA/l	0	Negative	
011990000543	fx1	5	3:03:07 PM	18	0.01 kUA/l	0	Negative	
011990000543	fx2	5	3:04:07 PM	14	Low RU	0	Negative	
011990000543	fx22	5	3:05:07 PM	27	0.12 kUA/l	0	Negative	
011990000543	gx2	5	3:06:07 PM	19	0.03 kUA/l	0	Negative	
011990000543	wx1	5	3:07:07 PM	26	0.11 kUA/l	0	Negative	
011990000543	t20	5	3:08:07 PM	15	Low RU	0		
011990000553	fx5	5	3:42:06 PM	45	0.33 kUA/l	0	Negative	
011990000553	fx1	5	3:43:06 PM	38	0.24 kUA/l	0	Negative	
011990000553	fx2	5	3:44:06 PM	61	0.52 kUA/l	1	Positive	
011990000553	fx22	5	3:45:06 PM	53	0.43 kUA/l	1	Positive	
011990000553	gx2	5	3:46:06 PM	51	0.41 kUA/l	1	Positive	
011990000553	wx1	5	3:47:06 PM	56	0.46 kUA/l	1	Positive	
011990000553	t20	5	3:48:05 PM	32	0.18 kUA/l	0		
011990000554	fx5	5	3:50:06 PM	21	0.05 kUA/l	0	Negative	
011990000554	fx1	5	3:51:06 PM	17	Low RU	0	Negative	
011990000554	fx2	5	3:52:05 PM	13	Low RU	0	Negative	
011990000554	fx22	5	3:53:05 PM	27	0.12 kUA/l	0	Negative	
011990000554	gx2	5	3:54:05 PM	30	0.15 kUA/l	0	Negative	
011990000554	wx1	5	3:55:06 PM	30	0.15 kUA/l	0	Negative	
011990000554	t20	5	3:56:05 PM	17	Low RU	0		
011990000558	fx5	5	5:30:02 PM	174	1.83 kUA/l	2	Positive	
011990000558	fx1	5	5:32:02 PM	194	2.06 kUA/l	2	Positive	
011990000558	fx2	5	5:33:02 PM	108	1.07 kUA/l	2	Positive	
011990000558	fx22	5	5:34:02 PM	159	1.65 kUA/l	2	Positive	
011990000558	gx2	5	5:35:02 PM	201	2.15 kUA/l	2	Positive	
011990000558	wx1	5	5:36:02 PM	197	2.10 kUA/l	2	Positive	
011990000558	t20	5	5:37:02 PM	169	1.78 kUA/l	2		
011990000559	fx5	5	5:14:03 PM	69	0.62 kUA/l	1	Positive	
011990000559	fx1	5	5:15:03 PM	21	0.04 kUA/l	0	Negative	
011990000559	fx2	5	5:16:02 PM	14	Low RU	0	Negative	
011990000559	fx22	5	5:17:02 PM	30	0.16 kUA/l	0	Negative	
011990000559	gx2	5	5:18:02 PM	24	0.08 kUA/l	0	Negative	
011990000559	wx1	5	5:19:02 PM	29	0.15 kUA/l	0	Negative	
011990000559	t20	5	5:20:02 PM	19	0.03 kUA/l	0		
011990000560	fx5	5	5:47:01 PM	23	0.07 kUA/l	0	Negative	
011990000560	fx1	5	5:48:01 PM	21	0.05 kUA/l	0	Negative	
011990000560	fx2	5	5:49:01 PM	113	1.12 kUA/l	2	Positive	
011990000560	fx22	5	5:50:01 PM	32	0.19 kUA/l	0	Negative	
011990000560	gx2	5	5:51:01 PM	30	0.16 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000560	wx1	5	5:52:01 PM	44	0.32 kUA/l	0	Negative	
011990000560	t20	5	5:53:01 PM	22	0.07 kUA/l	0		
011990000566	fx5	5	5:22:02 PM	20	0.04 kUA/l	0	Negative	
011990000566	fx1	5	5:23:03 PM	16	Low RU	0	Negative	
011990000566	fx2	5	5:24:02 PM	12	Low RU	0	Negative	
011990000566	fx22	5	5:25:02 PM	29	0.15 kUA/l	0	Negative	
011990000566	gx2	5	5:26:02 PM	23	0.08 kUA/l	0	Negative	
011990000566	wx1	5	5:27:02 PM	27	0.12 kUA/l	0	Negative	
011990000566	t20	5	5:28:02 PM	17	Low RU	0		
011990000568	fx5	5	5:39:01 PM	21	0.05 kUA/l	0	Negative	
011990000568	fx1	5	5:40:01 PM	17	Low RU	0	Negative	
011990000568	fx2	5	5:41:01 PM	13	Low RU	0	Negative	
011990000568	fx22	5	5:42:02 PM	33	0.19 kUA/l	0	Negative	
011990000568	gx2	5	5:43:01 PM	29	0.14 kUA/l	0	Negative	
011990000568	wx1	5	5:44:01 PM	26	0.11 kUA/l	0	Negative	
011990000568	t20	5	5:45:01 PM	17	0.01 kUA/l	0		
012001900752A	m6	1	5:01:03 PM	18	0.00 kUA/l			
012001900752A	m3	1	5:02:03 PM	14	0.00 kUA/l			
012001900752A	d2	1	5:03:03 PM	17	0.00 kUA/l			
012001900752A	d1	1	5:10:03 PM	19	0.00 kUA/l			
012001900752A	ex72	1	5:11:03 PM	16			Negative	
012001900752A	gx2	1	5:12:02 PM	21	0.01 kUA/l	0	Negative	

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	5:31:02 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000144	phad	4:53:03 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000144	fx5	4:54:03 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000144	fx1	4:55:04 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000144	fx2	4:56:03 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000144	fx22	4:57:03 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000144	gx2	4:58:03 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000144	wx1	4:59:03 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000144	t20	5:00:03 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000143	phad	4:45:03 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000143	fx5	4:46:04 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000143	fx1	4:47:04 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000143	fx2	4:48:03 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000143	fx22	4:49:03 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000143	gx2	4:50:03 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000143	wx1	4:51:03 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000143	t20	4:52:03 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	phad	4:37:04 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	fx5	4:38:04 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	fx1	4:39:04 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	fx2	4:40:04 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	fx22	4:41:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	gx2	4:42:04 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	wx1	4:43:04 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000146	t20	4:44:03 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	phad	4:29:04 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	fx5	4:30:04 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	fx1	4:31:04 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	fx2	4:32:04 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	fx22	4:33:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	gx2	4:34:04 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	wx1	4:35:04 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000145	t20	4:36:04 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	phad	4:21:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	fx5	4:22:04 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	fx1	4:23:05 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	fx2	4:24:04 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	fx22	4:25:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	gx2	4:26:04 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	wx1	4:27:05 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000148	t20	4:28:04 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	phad	4:13:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	fx5	4:14:05 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	fx1	4:15:05 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	fx2	4:16:05 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	fx22	4:17:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	gx2	4:18:05 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	wx1	4:19:05 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000147	t20	4:20:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	phad	4:05:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	fx5	4:06:05 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	fx1	4:07:05 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	fx2	4:08:05 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	fx22	4:09:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	gx2	4:10:05 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	wx1	4:11:05 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000150	t20	4:12:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	phad	3:57:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	fx5	3:58:05 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	fx1	3:59:05 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	fx2	4:00:06 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	fx22	4:01:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	gx2	4:02:05 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	wx1	4:03:05 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000149	t20	4:04:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	phad	3:49:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	fx5	3:50:06 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	fx1	3:51:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	fx2	3:52:05 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	fx22	3:53:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	gx2	3:54:05 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	wx1	3:55:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000554	t20	3:56:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	phad	3:41:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	fx5	3:42:06 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	fx1	3:43:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	fx2	3:44:06 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	fx22	3:45:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	gx2	3:46:06 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	wx1	3:47:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000553	t20	3:48:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000524	phad	3:33:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000524	fx5	3:34:06 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000524	fx1	3:35:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000524	fx2	3:36:06 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000524	fx22	3:37:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000524	gx2	3:38:06 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000524	wx1	3:39:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000524	t20	3:40:06 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	phad	3:25:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	fx5	3:26:06 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	fx1	3:27:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	fx2	3:28:06 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	fx22	3:29:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	gx2	3:30:06 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	wx1	3:31:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000505	t20	3:32:06 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	phad	3:17:07 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	fx5	3:18:07 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	fx1	3:19:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	fx2	3:20:07 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	fx22	3:21:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	gx2	3:22:07 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	wx1	3:23:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000504	t20	3:24:06 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	phad	3:09:07 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	fx5	3:10:07 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	fx1	3:11:07 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	fx2	3:12:07 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	fx22	3:13:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	gx2	3:14:07 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	wx1	3:15:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000514	t20	3:16:07 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	phad	3:01:07 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	fx5	3:02:07 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	fx1	3:03:07 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	fx2	3:04:07 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	fx22	3:05:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	gx2	3:06:07 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	wx1	3:07:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000543	t20	3:08:07 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	phad	2:53:08 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	fx5	2:54:08 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	fx1	2:55:08 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	fx2	2:56:07 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	fx22	2:57:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	gx2	2:58:07 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	wx1	2:59:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000533	t20	3:00:07 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000503	wx1	2:51:08 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000503	t20	2:52:08 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
012001900752A	m6	5:01:03 PM	356FF 21/03	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900752A	m3	5:02:03 PM	368K6 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900752A	d2	5:03:03 PM	358E2 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900752A	d1	5:10:03 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900752A	ex72	5:11:03 PM	521G9 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900752A	gx2	5:12:02 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000559	phad	5:13:03 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000559	fx5	5:14:03 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000559	fx1	5:15:03 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000559	fx2	5:16:02 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000559	fx22	5:17:02 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000559	gx2	5:18:02 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000559	wx1	5:19:02 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000559	t20	5:20:02 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000566	phad	5:21:02 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000566	fx5	5:22:02 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000566	fx1	5:23:03 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000566	fx2	5:24:02 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000566	fx22	5:25:02 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000566	gx2	5:26:02 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000566	wx1	5:27:02 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000566	t20	5:28:02 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	phad	5:29:02 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	fx5	5:30:02 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	fx1	5:32:02 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	fx2	5:33:02 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	fx22	5:34:02 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	gx2	5:35:02 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	wx1	5:36:02 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000558	t20	5:37:02 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	phad	5:38:01 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	fx5	5:39:01 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	fx1	5:40:01 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	fx2	5:41:01 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	fx22	5:42:02 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	gx2	5:43:01 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	wx1	5:44:01 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000568	t20	5:45:01 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	phad	5:46:01 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	fx5	5:47:01 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	fx1	5:48:01 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	fx2	5:49:01 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	fx22	5:50:01 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	gx2	5:51:01 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	wx1	5:52:01 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000560	t20	5:53:01 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011934900953A	t3	5:54:01 PM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			Exp
011934500583A	t3	5:55:01 PM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			Exp
011935100949A	t3	5:56:01 PM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			Exp

### Errors

#### Expired lot numbers used for Sample

352D0 ImmunoCAP