


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200122-384
Instrument name:	Phadia-A
Instrument type:	Phadia 250
Status:	Ready
Started:	22-Jan-20 2:49:25 PM
Ended:	22-Jan-20 6:15:20 PM
Approved:	Approved (faziz)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-A
Instrument type:	Phadia 250
Measured:	05-Jan-20 1:26:01 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	4:21:21 PM	6664	18.8 kUA/l	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000469	phad	5	5:21:19 PM	297	3.25 PAU/l	2	Positive	
011990000470	phad	5	4:41:21 PM	38	0.26 PAU/l	0	Negative	
011990000472	phad	5	5:37:18 PM	34	0.20 PAU/l	0	Negative	
011990000481	phad	5	5:13:20 PM	155	1.61 PAU/l	2	Positive	
011990000482	phad	5	4:33:21 PM	570	6.43 PAU/l	3	Positive	
011990000484	phad	5	5:29:19 PM	62	0.53 PAU/l	1	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000493	phad	5	5:05:20 PM	36	0.22 PAU/I	0	Negative	
011990000494	phad	5	4:25:21 PM	47	0.35 PAU/I	1	Positive	
011990000496	phad	5	5:53:18 PM	42	0.30 PAU/I	0	Negative	
011990000902	phad	5	4:03:22 PM	42	0.30 PAU/I	0	Negative	
011990000903	phad	5	6:09:18 PM	65	0.57 PAU/I	1	Positive	
011990000905	phad	5	4:49:20 PM	32	0.18 PAU/I	0	Negative	
011990000909	phad	5	5:45:18 PM	34	0.20 PAU/I	0	Negative	
011990000910	phad	5	6:01:18 PM	57	0.47 PAU/I	1	Positive	
011990000914	phad	5	4:11:22 PM	39	0.26 PAU/I	0	Negative	
011990000925	phad	5	4:57:20 PM	93	0.89 PAU/I	2	Positive	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000469	fx5	5	5:22:20 PM	47	0.35 kUA/I	1	Positive	
011990000469	fx1	5	5:23:19 PM	22	0.07 kUA/I	0	Negative	
011990000469	fx2	5	5:24:19 PM	353	3.90 kUA/I	3	Positive	
011990000469	fx22	5	5:25:19 PM	53	0.42 kUA/I	1	Positive	
011990000469	gx2	5	5:26:20 PM	30	0.16 kUA/I	0	Negative	
011990000469	wx1	5	5:27:19 PM	62	0.53 kUA/I	1	Positive	
011990000469	t20	5	5:28:19 PM	29	0.15 kUA/I	0		
011990000470	fx5	5	4:42:21 PM	25	0.09 kUA/I	0	Negative	
011990000470	fx1	5	4:43:20 PM	17	0.01 kUA/I	0	Negative	
011990000470	fx2	5	4:44:21 PM	13	Low RU	0	Negative	
011990000470	fx22	5	4:45:21 PM	35	0.21 kUA/I	0	Negative	
011990000470	gx2	5	4:46:21 PM	27	0.12 kUA/I	0	Negative	
011990000470	wx1	5	4:47:20 PM	124	1.25 kUA/I	2	Positive	
011990000470	t20	5	4:48:21 PM	21	0.04 kUA/I	0		
011990000472	fx5	5	5:38:19 PM	20	0.04 kUA/I	0	Negative	
011990000472	fx1	5	5:39:18 PM	15	Low RU	0	Negative	
011990000472	fx2	5	5:40:19 PM	20	0.03 kUA/I	0	Negative	
011990000472	fx22	5	5:41:19 PM	34	0.20 kUA/I	0	Negative	
011990000472	gx2	5	5:42:18 PM	21	0.06 kUA/I	0	Negative	
011990000472	wx1	5	5:43:18 PM	31	0.16 kUA/I	0	Negative	
011990000472	t20	5	5:44:19 PM	17	Low RU	0		
011990000481	fx5	5	5:14:20 PM	29	0.14 kUA/I	0	Negative	
011990000481	fx1	5	5:15:19 PM	22	0.07 kUA/I	0	Negative	
011990000481	fx2	5	5:16:20 PM	165	1.72 kUA/I	2	Positive	
011990000481	fx22	5	5:17:19 PM	36	0.23 kUA/I	0	Negative	
011990000481	gx2	5	5:18:20 PM	27	0.12 kUA/I	0	Negative	
011990000481	wx1	5	5:19:19 PM	108	1.07 kUA/I	2	Positive	
011990000481	t20	5	5:20:20 PM	39	0.26 kUA/I	0		
011990000482	fx5	5	4:34:21 PM	24	0.08 kUA/I	0	Negative	
011990000482	fx1	5	4:35:21 PM	18	0.02 kUA/I	0	Negative	
011990000482	fx2	5	4:36:21 PM	14	Low RU	0	Negative	
011990000482	fx22	5	4:37:21 PM	32	0.18 kUA/I	0	Negative	
011990000482	gx2	5	4:38:21 PM	22	0.06 kUA/I	0	Negative	
011990000482	wx1	5	4:39:21 PM	31	0.17 kUA/I	0	Negative	
011990000482	t20	5	4:40:21 PM	18	0.01 kUA/I	0		
011990000484	fx5	5	5:30:19 PM	24	0.09 kUA/I	0	Negative	
011990000484	fx1	5	5:31:18 PM	23	0.08 kUA/I	0	Negative	
011990000484	fx2	5	5:32:19 PM	14	Low RU	0	Negative	
011990000484	fx22	5	5:33:19 PM	33	0.19 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000484	gx2	5	5:34:19 PM	21	0.05 kUA/l	0	Negative	
011990000484	wx1	5	5:35:18 PM	32	0.18 kUA/l	0	Negative	
011990000484	t20	5	5:36:19 PM	18	0.02 kUA/l	0		
011990000493	fx5	5	5:06:20 PM	28	0.14 kUA/l	0	Negative	
011990000493	fx1	5	5:07:20 PM	18	0.02 kUA/l	0	Negative	
011990000493	fx2	5	5:08:20 PM	12	Low RU	0	Negative	
011990000493	fx22	5	5:09:20 PM	33	0.19 kUA/l	0	Negative	
011990000493	gx2	5	5:10:20 PM	23	0.08 kUA/l	0	Negative	
011990000493	wx1	5	5:11:19 PM	34	0.20 kUA/l	0	Negative	
011990000493	t20	5	5:12:20 PM	18	0.01 kUA/l	0		
011990000494	fx5	5	4:26:22 PM	22	0.06 kUA/l	0	Negative	
011990000494	fx1	5	4:27:21 PM	18	0.02 kUA/l	0	Negative	
011990000494	fx2	5	4:28:21 PM	62	0.53 kUA/l	1	Positive	
011990000494	fx22	5	4:29:21 PM	35	0.22 kUA/l	0	Negative	
011990000494	gx2	5	4:30:21 PM	21	0.06 kUA/l	0	Negative	
011990000494	wx1	5	4:31:21 PM	29	0.14 kUA/l	0	Negative	
011990000494	t20	5	4:32:22 PM	17	Low RU	0		
011990000496	fx5	5	5:54:18 PM	22	0.06 kUA/l	0	Negative	
011990000496	fx1	5	5:55:18 PM	16	Low RU	0	Negative	
011990000496	fx2	5	5:56:18 PM	14	Low RU	0	Negative	
011990000496	fx22	5	5:57:17 PM	34	0.20 kUA/l	0	Negative	
011990000496	gx2	5	5:58:18 PM	23	0.07 kUA/l	0	Negative	
011990000496	wx1	5	5:59:17 PM	30	0.15 kUA/l	0	Negative	
011990000496	t20	5	6:00:18 PM	20	0.04 kUA/l	0		
011990000902	fx5	5	4:04:23 PM	21	0.06 kUA/l	0	Negative	
011990000902	fx1	5	4:05:22 PM	18	0.01 kUA/l	0	Negative	
011990000902	fx2	5	4:06:22 PM	15	Low RU	0	Negative	
011990000902	fx22	5	4:07:22 PM	32	0.19 kUA/l	0	Negative	
011990000902	gx2	5	4:08:22 PM	23	0.08 kUA/l	0	Negative	
011990000902	wx1	5	4:09:22 PM	33	0.19 kUA/l	0	Negative	
011990000902	t20	5	4:10:22 PM	18	0.01 kUA/l	0		
011990000903	fx5	5	6:10:17 PM	22	0.06 kUA/l	0	Negative	
011990000903	fx1	5	6:11:17 PM	29	0.14 kUA/l	0	Negative	
011990000903	fx2	5	6:12:17 PM	52	0.41 kUA/l	1	Positive	
011990000903	fx22	5	6:13:17 PM	37	0.24 kUA/l	0	Negative	
011990000903	gx2	5	6:14:17 PM	32	0.18 kUA/l	0	Negative	
011990000905	fx5	5	4:50:21 PM	21	0.05 kUA/l	0	Negative	
011990000905	fx1	5	4:51:20 PM	16	Low RU	0	Negative	
011990000905	fx2	5	4:52:21 PM	13	Low RU	0	Negative	
011990000905	fx22	5	4:53:20 PM	31	0.17 kUA/l	0	Negative	
011990000905	gx2	5	4:54:21 PM	23	0.07 kUA/l	0	Negative	
011990000905	wx1	5	4:55:20 PM	29	0.14 kUA/l	0	Negative	
011990000905	t20	5	4:56:21 PM	17	Low RU	0		
011990000909	fx5	5	5:46:18 PM	21	0.04 kUA/l	0	Negative	
011990000909	fx1	5	5:47:18 PM	17	Low RU	0	Negative	
011990000909	fx2	5	5:48:18 PM	14	Low RU	0	Negative	
011990000909	fx22	5	5:49:18 PM	31	0.17 kUA/l	0	Negative	
011990000909	gx2	5	5:50:18 PM	33	0.19 kUA/l	0	Negative	
011990000909	wx1	5	5:51:18 PM	38	0.25 kUA/l	0	Negative	
011990000909	t20	5	5:52:18 PM	17	Low RU	0		
011990000910	fx5	5	6:02:17 PM	27	0.12 kUA/l	0	Negative	
011990000910	fx1	5	6:03:17 PM	32	0.18 kUA/l	0	Negative	
011990000910	fx2	5	6:04:17 PM	17	Low RU	0	Negative	
011990000910	fx22	5	6:05:17 PM	45	0.33 kUA/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000910	gx2	5	6:06:17 PM	29	0.14 kUA/I	0	Negative	
011990000910	wx1	5	6:07:17 PM	147	1.52 kUA/I	2	Positive	
011990000910	t20	5	6:08:17 PM	136	1.39 kUA/I	2		
011990000914	fx5	5	4:12:22 PM	32	0.18 kUA/I	0	Negative	
011990000914	fx1	5	4:17:22 PM	17	0.01 kUA/I	0	Negative	
011990000914	fx2	5	4:18:22 PM	28	0.13 kUA/I	0	Negative	
011990000914	fx22	5	4:19:21 PM	37	0.24 kUA/I	0	Negative	
011990000914	gx2	5	4:22:22 PM	22	0.06 kUA/I	0	Negative	
011990000914	wx1	5	4:23:21 PM	30	0.15 kUA/I	0	Negative	
011990000914	t20	5	4:24:22 PM	18	0.01 kUA/I	0		
011990000925	fx5	5	4:58:21 PM	24	0.08 kUA/I	0	Negative	
011990000925	fx1	5	4:59:20 PM	16	Low RU	0	Negative	
011990000925	fx2	5	5:00:20 PM	12	Low RU	0	Negative	
011990000925	fx22	5	5:01:20 PM	33	0.19 kUA/I	0	Negative	
011990000925	gx2	5	5:02:20 PM	25	0.10 kUA/I	0	Negative	
011990000925	wx1	5	5:03:20 PM	31	0.16 kUA/I	0	Negative	
011990000925	t20	5	5:04:20 PM	18	0.01 kUA/I	0		
012002000355A	f92	1	4:00:22 PM	112	0.22 kUA/I			
012002000355A	f209	1	4:01:22 PM	66	0.12 kUA/I			
012002000355A	f14	1	4:02:22 PM	90	0.17 kUA/I			
012002100642A	e1	1	2:55:24 PM	2047	4.89 kUA/I			
012002100642A	d2	1	2:56:25 PM	18	0.00 kUA/I			
012002100642A	e5	1	2:57:24 PM	1463	3.43 kUA/I			
012002100642A	f245	1	2:58:25 PM	10	0.00 kUA/I			
012002100642A	d1	1	2:59:24 PM	20	0.01 kUA/I			
012002100642A	f2	1	3:00:25 PM	16	0.00 kUA/I			
012002100642A	f44	1	3:01:24 PM	19	0.00 kUA/I			
012002100642A	f4	1	3:02:25 PM	19	0.00 kUA/I			
012002100642A	gx2	1	3:03:24 PM	21	0.01 kUA/I	0	Negative	
012002100642A	f17	1	3:04:24 PM	13	0.00 kUA/I			
012002100642A	f13	1	3:05:24 PM	16	0.00 kUA/I			
012002100696A	f256	1	2:49:25 PM	74	0.14 kUA/I			
012002100696A	ex72	1	2:50:25 PM	38			Negative	
012002100696A	mx2	1	2:51:24 PM	69			Negative	
012002100696A	gx2	1	2:52:25 PM	93	0.18 kUA/I	0	Negative	
012002100696A	i6	1	2:53:25 PM	1610	3.79 kUA/I			
012002100696A	f13	1	2:54:25 PM	102	0.20 kUA/I			
012002100753A	f18	1	3:06:25 PM	93	0.18 kUA/I			
012002100753A	f202	1	3:07:24 PM	1017	2.35 kUA/I			
012002100753A	f1	1	3:08:25 PM	309	0.68 kUA/I			
012002100753A	f345	1	3:09:24 PM	70	0.13 kUA/I			
012002100753A	f233	1	3:10:24 PM	146	0.30 kUA/I			
012002100753A	f201	1	3:11:24 PM	30	0.03 kUA/I			
012002100753A	f203	1	3:12:24 PM	1013	2.34 kUA/I			
012002100753A	f10	1	3:13:24 PM	261	0.57 kUA/I			
012002100753A	f256	1	3:14:24 PM	105	0.21 kUA/I			
012002100782A	f245	1	3:15:23 PM	34	0.04 kUA/I			
012002100782A	f2	1	3:16:24 PM	501	1.13 kUA/I			
012002100782A	f14	1	3:17:24 PM	371	0.82 kUA/I			
012002100782A	f4	1	3:18:24 PM	54	0.09 kUA/I			
012002101008A	m3	1	3:19:23 PM	23	0.01 kUA/I			
012002101008A	g2	1	3:20:24 PM	61	0.10 kUA/I			
012002101008A	e1	1	3:21:24 PM	15	0.00 kUA/I			
012002101008A	i206	1	3:22:24 PM	41	0.06 kUA/I			

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
012002101008A	e5	1	3:23:23 PM	19	0.01 kUA/I			
012002101008A	hx2	1	3:24:23 PM	41			Negative	
012002101008A	d1	1	3:25:23 PM	48	0.07 kUA/I			
012002101008A	ex72	1	3:26:23 PM	24			Negative	
012002101008A	mx2	1	3:27:23 PM	36			Negative	
012002101008A	i6	1	3:28:23 PM	53	0.09 kUA/I			
012002101055A	f20	2	3:29:23 PM	635	2.88 kUA/I			
012002101055A	f92	2	3:30:23 PM	80	0.30 kUA/I			
012002101055A	f27	2	3:31:23 PM	26	0.04 kUA/I			
012002101055A	f78	2	3:32:23 PM	22	0.02 kUA/I			
012002101055A	f202	2	3:33:23 PM	238	1.03 kUA/I			
012002101055A	f81	2	3:34:23 PM	19	0.01 kUA/I			
012002101055A	f82	2	3:35:23 PM	18	0.01 kUA/I			
012002101055A	f245	2	3:36:23 PM	1637	7.72 kUA/I			
012002101055A	f79	2	3:37:23 PM	62	0.21 kUA/I			
012002101055A	f17	2	3:38:23 PM	471	2.11 kUA/I			
012002101055A	hx2	2	3:39:23 PM	95			Positive	
012002101055A	d1	2	3:40:23 PM	63	0.22 kUA/I			
012002101055A	f2	2	3:41:23 PM	71	0.25 kUA/I			
012002101055A	f10	2	3:42:23 PM	2616	12.7 kUA/I			
012002101055A	f44	2	3:43:23 PM	44	0.13 kUA/I			
012002101055A	f256	2	3:44:23 PM	46	0.14 kUA/I			
012002101055A	f4	2	3:45:22 PM	95	0.37 kUA/I			
012002101055A	mx2	2	3:46:23 PM	28			Negative	
012002101055A	f13	2	3:47:22 PM	473	2.12 kUA/I			
012002101154B	f78	1	3:58:22 PM	15	0.00 kUA/I			
012002101154B	f2	1	3:59:22 PM	17	0.00 kUA/I			
012002200300A	m3	1	3:48:23 PM	16	0.00 kUA/I			
012002200300A	g2	1	3:49:23 PM	17	0.00 kUA/I			
012002200300A	e1	1	3:50:23 PM	11	0.00 kUA/I			
012002200300A	i206	1	3:51:22 PM	15	0.00 kUA/I			
012002200300A	e5	1	3:52:22 PM	12	0.00 kUA/I			
012002200300A	d1	1	3:53:22 PM	19	0.01 kUA/I			
012002200300A	e215	1	3:54:22 PM	12	0.00 kUA/I			
012002200300A	ex72	1	3:55:22 PM	16			Negative	
012002200300A	mx2	1	3:56:22 PM	27			Negative	
012002200300A	gx2	1	3:57:22 PM	20	0.01 kUA/I	0	Negative	

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
-----------------	------	----------	----------	-----------	---------	----------	----------	-----------	---------	---------

Laboratory Report

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	4:21:21 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002000355A	f92	4:00:22 PM	561KS 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			Exp
012002000355A	f209	4:01:22 PM	680GA 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002000355A	f14	4:02:22 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100696A	f256	2:49:25 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100696A	ex72	2:50:25 PM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100696A	mx2	2:51:24 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100696A	gx2	2:52:25 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100696A	i6	2:53:25 PM	465D3 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100696A	f13	2:54:25 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f18	3:06:25 PM	421BJ 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f202	3:07:24 PM	997KA 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f1	3:08:25 PM	361EN 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f345	3:09:24 PM	BTXGP 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f233	3:10:24 PM	904LM 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f201	3:11:24 PM	906J3 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f203	3:12:24 PM	998HG 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f10	3:13:24 PM	418DE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100753A	f256	3:14:24 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100782A	f245	3:15:23 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100782A	f2	3:16:24 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100782A	f14	3:17:24 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100782A	f4	3:18:24 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	m3	3:19:23 PM	368KB 21/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	g2	3:20:24 PM	379CG 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	e1	3:21:24 PM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	i206	3:22:24 PM	C07GX 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	e5	3:23:23 PM	360F3 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	hx2	3:24:23 PM	446BR 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	d1	3:25:23 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	ex72	3:26:23 PM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	mx2	3:27:23 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101008A	i6	3:28:23 PM	465D3 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101055A	f20	3:29:23 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f92	3:30:23 PM	561KS 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f27	3:31:23 PM	450CC 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f78	3:32:23 PM	CTSB0 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f202	3:33:23 PM	997KA 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f81	3:34:23 PM	527GX 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f82	3:35:23 PM	528GB 19/11!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f245	3:36:23 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	Exp
012002101055A	f79	3:37:23 PM	525JR 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f17	3:38:23 PM	D2FBX 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	hx2	3:39:23 PM	446BR 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	d1	3:40:23 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f2	3:41:23 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f10	3:42:23 PM	418DE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f44	3:43:23 PM	457BA 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f256	3:44:23 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f4	3:45:22 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	mx2	3:46:23 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002101055A	f13	3:47:22 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002100642A	e1	2:55:24 PM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	d2	2:56:25 PM	358E2 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	e5	2:57:24 PM	360F3 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	f245	2:58:25 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	d1	2:59:24 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	f2	3:00:25 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	f44	3:01:24 PM	457BA 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	f4	3:02:25 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	gx2	3:03:24 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	m3	3:48:23 PM	368KB 21/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	g2	3:49:23 PM	379CG 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002200300A	e1	3:50:23 PM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	i206	3:51:22 PM	C07GX 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	e5	3:52:22 PM	360F3 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	d1	3:53:22 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	e215	3:54:22 PM	BP5GE 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	ex72	3:55:22 PM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	mx2	3:56:22 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200300A	gx2	3:57:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	f17	3:04:24 PM	D2FBX 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002100642A	f13	3:05:24 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101154B	f78	3:58:22 PM	CTSB0 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002101154B	f2	3:59:22 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
011990000484	phad	5:29:19 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000484	fx5	5:30:19 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000484	fx1	5:31:18 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000484	fx2	5:32:19 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000484	fx22	5:33:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000484	gx2	5:34:19 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000484	wx1	5:35:18 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000484	t20	5:36:19 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	phad	5:37:18 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	fx5	5:38:19 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	fx1	5:39:18 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	fx2	5:40:19 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	fx22	5:41:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	gx2	5:42:18 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	wx1	5:43:18 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000472	t20	5:44:19 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	phad	5:45:18 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	fx5	5:46:18 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	fx1	5:47:18 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	fx2	5:48:18 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	fx22	5:49:18 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	gx2	5:50:18 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	wx1	5:51:18 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000909	t20	5:52:18 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	phad	5:53:18 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	fx5	5:54:18 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	fx1	5:55:18 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	fx2	5:56:18 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	fx22	5:57:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	gx2	5:58:18 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	wx1	5:59:17 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000496	t20	6:00:18 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	phad	6:01:18 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	fx5	6:02:17 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	fx1	6:03:17 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	fx2	6:04:17 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	fx22	6:05:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	gx2	6:06:17 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	wx1	6:07:17 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000910	t20	6:08:17 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000903	phad	6:09:18 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000903	fx5	6:10:17 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000903	fx1	6:11:17 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000903	fx2	6:12:17 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000903	fx22	6:13:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000903	gx2	6:14:17 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000902	phad	4:03:22 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000902	fx5	4:04:23 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000902	fx1	4:05:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000902	fx2	4:06:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000902	fx22	4:07:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000902	gx2	4:08:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000902	wx1	4:09:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000902	t20	4:10:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	phad	4:11:22 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	fx5	4:12:22 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	fx1	4:17:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	fx2	4:18:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	fx22	4:19:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	gx2	4:22:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	wx1	4:23:21 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000914	t20	4:24:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	phad	4:25:21 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	fx5	4:26:22 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	fx1	4:27:21 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	fx2	4:28:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	fx22	4:29:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	gx2	4:30:21 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	wx1	4:31:21 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000494	t20	4:32:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	phad	4:33:21 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	fx5	4:34:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	fx1	4:35:21 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	fx2	4:36:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	fx22	4:37:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	gx2	4:38:21 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	wx1	4:39:21 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000482	t20	4:40:21 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	phad	4:41:21 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	fx5	4:42:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	fx1	4:43:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	fx2	4:44:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	fx22	4:45:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	gx2	4:46:21 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	wx1	4:47:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000470	t20	4:48:21 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	phad	4:49:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	fx5	4:50:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	fx1	4:51:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	fx2	4:52:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	fx22	4:53:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	gx2	4:54:21 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	wx1	4:55:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000905	t20	4:56:21 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	phad	4:57:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	fx5	4:58:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	fx1	4:59:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	fx2	5:00:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	fx22	5:01:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	gx2	5:02:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	wx1	5:03:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000925	t20	5:04:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	phad	5:05:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	fx5	5:06:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	fx1	5:07:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	fx2	5:08:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	fx22	5:09:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	gx2	5:10:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	wx1	5:11:19 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000493	t20	5:12:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	phad	5:13:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	fx5	5:14:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	fx1	5:15:19 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	fx2	5:16:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	fx22	5:17:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	gx2	5:18:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	wx1	5:19:19 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000481	t20	5:20:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000469	phad	5:21:19 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000469	fx5	5:22:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000469	fx1	5:23:19 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000469	fx2	5:24:19 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000469	fx22	5:25:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000469	gx2	5:26:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000469	wx1	5:27:19 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000469	t20	5:28:19 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

Errors

Expired lot numbers used for Sample

680GA	ImmunoCAP
528GB	ImmunoCAP