


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200109-124  
Instrument name: Phadia-B  
Instrument type: Phadia 250  
Status: Ready  
Started: 09-Jan-20 11:02:55 AM  
Ended: 09-Jan-20 3:17:35 PM  
Approved: Approved (Ludy)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-B  
Instrument type: Phadia 250  
Measured: 23-Dec-19 3:51:37 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	18	0.00			
		3:51:37 PM	19	0.00	OK		
		3:52:36 PM	18	0.00	OK		
		%CV	4.5	>99			
CAL-0.35	0.35	Mean	147	0.35			
		3:53:37 PM	146	0.35	OK		
		3:54:36 PM	148	0.35	OK		
		%CV	0.6	0.7			
CAL-0.70	0.70	Mean	271	0.69			
		3:55:37 PM	274	0.70	OK		
		3:56:36 PM	268	0.68	OK		
		%CV	1.6	1.7			
CAL-3.50	3.50	Mean	1279	3.51			
		3:57:37 PM	1283	3.52	OK		
		3:58:36 PM	1275	3.50	OK		
		%CV	0.5	0.5			
CAL-17.5	17.5	Mean	5583	17.5			
		3:59:37 PM	5591	17.5	OK		
		4:00:36 PM	5575	17.5	OK		
		%CV	0.2	0.2			
CAL-100	100	Mean	17182	100			
		4:01:37 PM	17224	101	OK		
		4:02:36 PM	17140	99.4	OK		
		%CV	0.3	0.9			

Curve data	Value	Logged mean	Difference
ED-20	10.1	9.67	+3.97
ED-50	30.1	29.6	+1.76
ED-80	62.3	62.2	+0.20
Slope	211	216	-2.31

Number of curves included in mean: 52

### Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	11:02:55 AM	352	0.91	High*	
		Expected conc:	0.70			
		Expected SD:	0.08			
		Inner limit:	2.200			
		Outer limit:	3.500			
CC-2	17.5	11:03:56 AM	7048	23.3	High*	
		Expected conc:	17.5			
		Expected SD:	1.75			
		Inner limit:	2.200			
		Outer limit:	3.500			
CC-1	0.70	12:48:52 PM	321	0.83	OK	
CC-2	17.5	12:49:52 PM	6372	20.5	OK	

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
----------	------	-----	----------	------	-----------	-----	-----------	---------

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	11:20:55 AM	3740	11.0 kUA/I	CZ1AR	OK	
IQC271119	d1	1	11:25:55 AM	6969	22.9 kUA/I	<lab>	----	
			Expected value:	---				
			Expected SD:	---				
			SD limit:	---				
			Range min:	---				
			Range max:	---				
			Limit CV%:	---				
slgE L	e1	1	11:31:55 AM	1244	3.41 kUA/I	CZ4AN	OK	
slgE Neg	d1	1	11:32:54 AM	21	0.01 kUA/I	BL1CP	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000701	phad	5	2:34:48 PM	257	3.27 PAU/I	---	Positive	
011990000707	phad	5	2:26:48 PM	33	0.21 PAU/I	---	Negative	
011990000708	phad	5	2:58:47 PM	48	0.41 PAU/I	---	Positive	
011990000709	phad	5	1:06:51 PM	35	0.23 PAU/I	---	Negative	
011990000713	phad	5	1:14:51 PM	40	0.30 PAU/I	---	Negative	
011990000714	phad	5	2:02:49 PM	2554	36.4 PAU/I	---	Positive	
011990000715	phad	5	1:54:49 PM	86	0.93 PAU/I	---	Positive	
011990000717	phad	5	2:50:47 PM	33	0.21 PAU/I	---	Negative	
011990000718	phad	5	1:30:50 PM	34	0.23 PAU/I	---	Negative	
011990000719	phad	5	1:22:51 PM	36	0.24 PAU/I	---	Negative	
011990000720	phad	5	2:10:48 PM	32	0.20 PAU/I	---	Negative	
011990000722	phad	5	2:42:47 PM	32	0.19 PAU/I	---	Negative	
011990000723	phad	5	1:46:49 PM	2160	30.4 PAU/I	---	Positive	
011990000724	phad	5	1:38:49 PM	32	0.20 PAU/I	---	Negative	
011990000725	phad	5	2:18:48 PM	162	1.97 PAU/I	---	Positive	
011990000751	phad	5	12:06:53 PM	146	1.76 PAU/I	---	Positive	
011990000752	phad	5	12:22:53 PM	33	0.21 PAU/I	---	Negative	
011990000753	phad	5	12:14:53 PM	33	0.21 PAU/I	---	Negative	
011990000756	phad	5	11:58:53 AM	36	0.25 PAU/I	---	Negative	
011990000757	phad	5	12:30:52 PM	37	0.27 PAU/I	---	Negative	
011990000761	phad	5	11:50:54 AM	34	0.22 PAU/I	---	Negative	
011990000762	phad	5	12:56:51 PM	31	0.18 PAU/I	---	Negative	
011990000766	phad	5	11:42:54 AM	42	0.33 PAU/I	---	Negative	
011990000771	phad	5	11:34:54 AM	41	0.32 PAU/I	---	Negative	
011990000772	phad	5	11:18:54 AM	89	0.97 PAU/I	---	Positive	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000701	fx5	5	2:35:49 PM	33	0.21 kUA/I	---	Negative	
011990000701	fx1	5	2:36:48 PM	19	0.02 kUA/I	---	Negative	
011990000701	fx2	5	2:37:48 PM	116	1.35 kUA/I	---	Positive	
011990000701	fx22	5	2:38:47 PM	31	0.19 kUA/I	---	Negative	
011990000701	gx2	5	2:39:48 PM	31	0.18 kUA/I	---	Negative	
011990000701	wx1	5	2:40:47 PM	57	0.53 kUA/I	---	Positive	
011990000701	t20	5	2:41:48 PM	34	0.22 kUA/I	---		
011990000707	fx5	5	2:27:48 PM	26	0.11 kUA/I	---	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000707	fx1	5	2:28:48 PM	15	---	---	Negative	
011990000707	fx2	5	2:29:49 PM	16	---	---	Negative	
011990000707	fx22	5	2:30:48 PM	33	0.21 kUA/I	---	Negative	
011990000707	gx2	5	2:31:48 PM	22	0.06 kUA/I	---	Negative	
011990000707	wx1	5	2:32:47 PM	26	0.12 kUA/I	---	Negative	
011990000707	t20	5	2:33:48 PM	17	---	---		
011990000708	fx5	5	2:59:47 PM	24	0.08 kUA/I	---	Negative	
011990000708	fx1	5	3:00:47 PM	16	---	---	Negative	
011990000708	fx2	5	3:01:48 PM	18	0.01 kUA/I	---	Negative	
011990000708	fx22	5	3:02:47 PM	30	0.16 kUA/I	---	Negative	
011990000708	gx2	5	3:03:47 PM	34	0.22 kUA/I	---	Negative	
011990000708	wx1	5	3:04:46 PM	76	0.79 kUA/I	---	Positive	
011990000708	t20	5	3:05:47 PM	17	---	---		
011990000709	fx5	5	1:07:51 PM	22	0.05 kUA/I	---	Negative	
011990000709	fx1	5	1:08:51 PM	16	---	---	Negative	
011990000709	fx2	5	1:09:51 PM	15	---	---	Negative	
011990000709	fx22	5	1:10:51 PM	31	0.17 kUA/I	---	Negative	
011990000709	gx2	5	1:11:50 PM	20	0.03 kUA/I	---	Negative	
011990000709	wx1	5	1:12:50 PM	32	0.20 kUA/I	---	Negative	
011990000709	t20	5	1:13:51 PM	15	---	---		
011990000713	fx5	5	1:15:51 PM	19	0.02 kUA/I	---	Negative	
011990000713	fx1	5	1:16:50 PM	16	---	---	Negative	
011990000713	fx2	5	1:17:50 PM	14	---	---	Negative	
011990000713	fx22	5	1:18:50 PM	31	0.18 kUA/I	---	Negative	
011990000713	gx2	5	1:19:50 PM	19	0.01 kUA/I	---	Negative	
011990000713	wx1	5	1:20:50 PM	29	0.15 kUA/I	---	Negative	
011990000713	t20	5	1:21:51 PM	17	---	---		
011990000714	fx5	5	2:03:49 PM	93	1.02 kUA/I	---	Positive	
011990000714	fx1	5	2:04:49 PM	97	1.09 kUA/I	---	Positive	
011990000714	fx2	5	2:05:49 PM	91	0.99 kUA/I	---	Positive	
011990000714	fx22	5	2:06:48 PM	244	3.09 kUA/I	---	Positive	
011990000714	gx2	5	2:07:49 PM	514	6.80 kUA/I	---	Positive	
011990000714	wx1	5	2:08:49 PM	3049	44.1 kUA/I	---	Positive	
011990000714	t20	5	2:09:49 PM	1170	16.0 kUA/I	---		
011990000715	fx5	5	1:55:50 PM	25	0.10 kUA/I	---	Negative	
011990000715	fx1	5	1:56:49 PM	18	---	---	Negative	
011990000715	fx2	5	1:57:50 PM	17	---	---	Negative	
011990000715	fx22	5	1:58:49 PM	31	0.18 kUA/I	---	Negative	
011990000715	gx2	5	1:59:49 PM	78	0.83 kUA/I	---	Positive	
011990000715	wx1	5	2:00:49 PM	49	0.42 kUA/I	---	Positive	
011990000715	t20	5	2:01:49 PM	18	---	---		
011990000717	fx5	5	2:51:48 PM	21	0.04 kUA/I	---	Negative	
011990000717	fx1	5	2:52:47 PM	16	---	---	Negative	
011990000717	fx2	5	2:53:47 PM	33	0.21 kUA/I	---	Negative	
011990000717	fx22	5	2:54:47 PM	32	0.20 kUA/I	---	Negative	
011990000717	gx2	5	2:55:47 PM	19	0.02 kUA/I	---	Negative	
011990000717	wx1	5	2:56:47 PM	25	0.10 kUA/I	---	Negative	
011990000717	t20	5	2:57:47 PM	15	---	---		
011990000718	fx5	5	1:31:50 PM	21	0.04 kUA/I	---	Negative	
011990000718	fx1	5	1:32:50 PM	17	---	---	Negative	
011990000718	fx2	5	1:33:50 PM	13	---	---	Negative	
011990000718	fx22	5	1:34:50 PM	30	0.17 kUA/I	---	Negative	
011990000718	gx2	5	1:35:50 PM	19	0.02 kUA/I	---	Negative	
011990000718	wx1	5	1:36:50 PM	28	0.14 kUA/I	---	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000718	t20	5	1:37:50 PM	17	---	---		
011990000719	fx5	5	1:23:50 PM	20	0.02 kUA/I	---	Negative	
011990000719	fx1	5	1:24:50 PM	17	---	---	Negative	
011990000719	fx2	5	1:25:50 PM	13	---	---	Negative	
011990000719	fx22	5	1:26:50 PM	30	0.17 kUA/I	---	Negative	
011990000719	gx2	5	1:27:50 PM	20	0.03 kUA/I	---	Negative	
011990000719	wx1	5	1:28:50 PM	31	0.18 kUA/I	---	Negative	
011990000719	t20	5	1:29:50 PM	17	---	---		
011990000720	fx5	5	2:11:49 PM	20	0.03 kUA/I	---	Negative	
011990000720	fx1	5	2:12:48 PM	15	---	---	Negative	
011990000720	fx2	5	2:13:49 PM	14	---	---	Negative	
011990000720	fx22	5	2:14:48 PM	31	0.19 kUA/I	---	Negative	
011990000720	gx2	5	2:15:49 PM	22	0.06 kUA/I	---	Negative	
011990000720	wx1	5	2:16:48 PM	28	0.14 kUA/I	---	Negative	
011990000720	t20	5	2:17:49 PM	17	---	---		
011990000722	fx5	5	2:43:48 PM	19	0.02 kUA/I	---	Negative	
011990000722	fx1	5	2:44:47 PM	13	---	---	Negative	
011990000722	fx2	5	2:45:48 PM	14	---	---	Negative	
011990000722	fx22	5	2:46:47 PM	30	0.16 kUA/I	---	Negative	
011990000722	gx2	5	2:47:48 PM	25	0.10 kUA/I	---	Negative	
011990000722	wx1	5	2:48:47 PM	26	0.12 kUA/I	---	Negative	
011990000722	t20	5	2:49:47 PM	17	---	---		
011990000723	fx5	5	1:47:50 PM	28	0.14 kUA/I	---	Negative	
011990000723	fx1	5	1:48:49 PM	92	1.01 kUA/I	---	Positive	
011990000723	fx2	5	1:49:50 PM	1273	17.5 kUA/I	---	Positive	
011990000723	fx22	5	1:50:49 PM	38	0.27 kUA/I	---	Negative	
011990000723	gx2	5	1:51:50 PM	431	5.66 kUA/I	---	Positive	
011990000723	wx1	5	1:52:49 PM	164	2.00 kUA/I	---	Positive	
011990000723	t20	5	1:53:50 PM	46	0.38 kUA/I	---		
011990000724	fx5	5	1:39:50 PM	20	0.03 kUA/I	---	Negative	
011990000724	fx1	5	1:40:50 PM	18	---	---	Negative	
011990000724	fx2	5	1:41:50 PM	14	---	---	Negative	
011990000724	fx22	5	1:42:49 PM	30	0.17 kUA/I	---	Negative	
011990000724	gx2	5	1:43:50 PM	23	0.07 kUA/I	---	Negative	
011990000724	wx1	5	1:44:49 PM	45	0.37 kUA/I	---	Positive	
011990000724	t20	5	1:45:50 PM	18	---	---		
011990000725	fx5	5	2:19:49 PM	25	0.10 kUA/I	---	Negative	
011990000725	fx1	5	2:20:48 PM	17	---	---	Negative	
011990000725	fx2	5	2:21:49 PM	19	0.02 kUA/I	---	Negative	
011990000725	fx22	5	2:22:48 PM	35	0.23 kUA/I	---	Negative	
011990000725	gx2	5	2:23:48 PM	27	0.12 kUA/I	---	Negative	
011990000725	wx1	5	2:24:48 PM	37	0.26 kUA/I	---	Negative	
011990000725	t20	5	2:25:49 PM	22	0.06 kUA/I	---		
011990000751	fx5	5	12:07:54 PM	23	0.07 kUA/I	---	Negative	
011990000751	fx1	5	12:08:53 PM	23	0.07 kUA/I	---	Negative	
011990000751	fx2	5	12:09:54 PM	17	---	---	Negative	
011990000751	fx22	5	12:10:53 PM	35	0.23 kUA/I	---	Negative	
011990000751	gx2	5	12:11:53 PM	24	0.08 kUA/I	---	Negative	
011990000751	wx1	5	12:12:53 PM	191	2.36 kUA/I	---	Positive	
011990000751	t20	5	12:13:53 PM	29	0.15 kUA/I	---		
011990000752	fx5	5	12:23:53 PM	19	0.02 kUA/I	---	Negative	
011990000752	fx1	5	12:24:52 PM	18	---	---	Negative	
011990000752	fx2	5	12:25:53 PM	13	---	---	Negative	
011990000752	fx22	5	12:26:52 PM	29	0.16 kUA/I	---	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000752	gx2	5	12:27:53 PM	20	0.03 kUA/I	---	Negative	
011990000752	wx1	5	12:28:52 PM	32	0.19 kUA/I	---	Negative	
011990000752	t20	5	12:29:52 PM	17	---	---		
011990000753	fx5	5	12:15:53 PM	18	0.01 kUA/I	---	Negative	
011990000753	fx1	5	12:16:53 PM	17	---	---	Negative	
011990000753	fx2	5	12:17:54 PM	12	---	---	Negative	
011990000753	fx22	5	12:18:52 PM	29	0.15 kUA/I	---	Negative	
011990000753	gx2	5	12:19:53 PM	22	0.06 kUA/I	---	Negative	
011990000753	wx1	5	12:20:52 PM	33	0.21 kUA/I	---	Negative	
011990000753	t20	5	12:21:53 PM	16	---	---		
011990000756	fx5	5	11:59:54 AM	20	0.03 kUA/I	---	Negative	
011990000756	fx1	5	12:00:53 PM	16	---	---	Negative	
011990000756	fx2	5	12:01:54 PM	13	---	---	Negative	
011990000756	fx22	5	12:02:53 PM	31	0.18 kUA/I	---	Negative	
011990000756	gx2	5	12:03:54 PM	19	0.01 kUA/I	---	Negative	
011990000756	wx1	5	12:04:53 PM	30	0.17 kUA/I	---	Negative	
011990000756	t20	5	12:05:53 PM	16	---	---		
011990000757	fx5	5	12:31:52 PM	22	0.06 kUA/I	---	Negative	
011990000757	fx1	5	12:50:51 PM	17	---	---	Negative	
011990000757	fx2	5	12:51:51 PM	14	---	---	Negative	
011990000757	fx22	5	12:52:51 PM	30	0.16 kUA/I	---	Negative	
011990000757	gx2	5	12:53:51 PM	20	0.04 kUA/I	---	Negative	
011990000757	wx1	5	12:54:51 PM	33	0.20 kUA/I	---	Negative	
011990000757	t20	5	12:55:51 PM	16	---	---		
011990000761	fx5	5	11:51:54 AM	22	0.06 kUA/I	---	Negative	
011990000761	fx1	5	11:52:53 AM	17	---	---	Negative	
011990000761	fx2	5	11:53:54 AM	13	---	---	Negative	
011990000761	fx22	5	11:54:53 AM	30	0.16 kUA/I	---	Negative	
011990000761	gx2	5	11:55:54 AM	19	0.01 kUA/I	---	Negative	
011990000761	wx1	5	11:56:53 AM	32	0.19 kUA/I	---	Negative	
011990000761	t20	5	11:57:54 AM	16	---	---		
011990000762	fx5	5	12:59:51 PM	23	0.07 kUA/I	---	Negative	
011990000762	fx1	5	1:00:51 PM	17	---	---	Negative	
011990000762	fx2	5	1:01:51 PM	13	---	---	Negative	
011990000762	fx22	5	1:02:51 PM	28	0.14 kUA/I	---	Negative	
011990000762	gx2	5	1:03:51 PM	21	0.04 kUA/I	---	Negative	
011990000762	wx1	5	1:04:51 PM	30	0.16 kUA/I	---	Negative	
011990000762	t20	5	1:05:51 PM	16	---	---		
011990000766	fx5	5	11:43:55 AM	21	0.05 kUA/I	---	Negative	
011990000766	fx1	5	11:44:54 AM	16	---	---	Negative	
011990000766	fx2	5	11:45:54 AM	23	0.08 kUA/I	---	Negative	
011990000766	fx22	5	11:46:54 AM	30	0.17 kUA/I	---	Negative	
011990000766	gx2	5	11:47:54 AM	18	0.01 kUA/I	---	Negative	
011990000766	wx1	5	11:48:54 AM	32	0.20 kUA/I	---	Negative	
011990000766	t20	5	11:49:54 AM	17	---	---		
011990000771	fx5	5	11:35:55 AM	29	0.15 kUA/I	---	Negative	
011990000771	fx1	5	11:36:54 AM	18	---	---	Negative	
011990000771	fx2	5	11:37:55 AM	16	---	---	Negative	
011990000771	fx22	5	11:38:54 AM	31	0.18 kUA/I	---	Negative	
011990000771	gx2	5	11:39:55 AM	33	0.21 kUA/I	---	Negative	
011990000771	wx1	5	11:40:54 AM	29	0.15 kUA/I	---	Negative	
011990000771	t20	5	11:41:55 AM	16	---	---		
011990000772	fx5	5	11:19:55 AM	23	0.07 kUA/I	---	Negative	
011990000772	fx1	5	11:26:54 AM	19	0.02 kUA/I	---	Negative	



# Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000772	fx2	5	11:27:55 AM	23	0.07 kUA/l	---	Negative	
011990000772	fx22	5	11:28:55 AM	33	0.21 kUA/l	---	Negative	
011990000772	gx2	5	11:29:55 AM	28	0.14 kUA/l	---	Negative	
011990000772	wx1	5	11:30:54 AM	29	0.16 kUA/l	---	Negative	
011990000772	t20	5	11:33:55 AM	19	0.01 kUA/l	---		

## Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	3:51:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	3:52:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:53:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:54:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:55:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:56:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:57:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:58:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	3:59:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	4:00:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:01:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:02:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	11:02:55 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	11:03:56 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-1	a_IgE	12:48:52 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	12:49:52 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	11:20:55 AM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
IQC271119	d1	11:25:55 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		
slgE L	e1	11:31:55 AM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06!		Exp
slgE Neg	d1	11:32:54 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12!		Exp

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000772	phad	11:18:54 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000772	fx5	11:19:55 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000772	fx1	11:26:54 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000772	fx2	11:27:55 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000772	fx22	11:28:55 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000772	gx2	11:29:55 AM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000772	wx1	11:30:54 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000772	t20	11:33:55 AM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	phad	11:34:54 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	fx5	11:35:55 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	fx1	11:36:54 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	fx2	11:37:55 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	fx22	11:38:54 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	gx2	11:39:55 AM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	wx1	11:40:54 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000771	t20	11:41:55 AM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	phad	11:42:54 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	fx5	11:43:55 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	fx1	11:44:54 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	fx2	11:45:54 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	fx22	11:46:54 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	gx2	11:47:54 AM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	wx1	11:48:54 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000766	t20	11:49:54 AM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000761	phad	11:50:54 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000761	fx5	11:51:54 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000761	fx1	11:52:53 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000761	fx2	11:53:54 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000761	fx22	11:54:53 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000761	gx2	11:55:54 AM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000761	wx1	11:56:53 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000761	t20	11:57:54 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	phad	11:58:53 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	fx5	11:59:54 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	fx1	12:00:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	fx2	12:01:54 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	fx22	12:02:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	gx2	12:03:54 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	wx1	12:04:53 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000756	t20	12:05:53 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	phad	12:06:53 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	fx5	12:07:54 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	fx1	12:08:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	fx2	12:09:54 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	fx22	12:10:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	gx2	12:11:53 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	wx1	12:12:53 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000751	t20	12:13:53 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	phad	12:14:53 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	fx5	12:15:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	fx1	12:16:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	fx2	12:17:54 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	fx22	12:18:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	gx2	12:19:53 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	wx1	12:20:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000753	t20	12:21:53 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	phad	12:22:53 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	fx5	12:23:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	fx1	12:24:52 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	fx2	12:25:53 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	fx22	12:26:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	gx2	12:27:53 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	wx1	12:28:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000752	t20	12:29:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	phad	12:30:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	fx5	12:31:52 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	fx1	12:50:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	fx2	12:51:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	fx22	12:52:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	gx2	12:53:51 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	wx1	12:54:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000757	t20	12:55:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	phad	12:56:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	fx5	12:59:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	fx1	1:00:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	fx2	1:01:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	fx22	1:02:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	gx2	1:03:51 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	wx1	1:04:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000762	t20	1:05:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	phad	1:06:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	fx5	1:07:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	fx1	1:08:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	fx2	1:09:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	fx22	1:10:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	gx2	1:11:50 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	wx1	1:12:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000709	t20	1:13:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000713	phad	1:14:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000713	fx5	1:15:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000713	fx1	1:16:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000713	fx2	1:17:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000713	fx22	1:18:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000713	gx2	1:19:50 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000713	wx1	1:20:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000713	t20	1:21:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	phad	1:22:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	fx5	1:23:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	fx1	1:24:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	fx2	1:25:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	fx22	1:26:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	gx2	1:27:50 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	wx1	1:28:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000719	t20	1:29:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	phad	1:30:50 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	fx5	1:31:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	fx1	1:32:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	fx2	1:33:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	fx22	1:34:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	gx2	1:35:50 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	wx1	1:36:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000718	t20	1:37:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	phad	1:38:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	fx5	1:39:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	fx1	1:40:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	fx2	1:41:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	fx22	1:42:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	gx2	1:43:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	wx1	1:44:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000724	t20	1:45:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	phad	1:46:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	fx5	1:47:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	fx1	1:48:49 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	fx2	1:49:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	fx22	1:50:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	gx2	1:51:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	wx1	1:52:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000723	t20	1:53:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	phad	1:54:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	fx5	1:55:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	fx1	1:56:49 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	fx2	1:57:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	fx22	1:58:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	gx2	1:59:49 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	wx1	2:00:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000715	t20	2:01:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	phad	2:02:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	fx5	2:03:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	fx1	2:04:49 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	fx2	2:05:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	fx22	2:06:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	gx2	2:07:49 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	wx1	2:08:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000714	t20	2:09:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	phad	2:10:48 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	fx5	2:11:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	fx1	2:12:48 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	fx2	2:13:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	fx22	2:14:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	gx2	2:15:49 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	wx1	2:16:48 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000720	t20	2:17:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000725	phad	2:18:48 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000725	fx5	2:19:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000725	fx1	2:20:48 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000725	fx2	2:21:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000725	fx22	2:22:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000725	gx2	2:23:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000725	wx1	2:24:48 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000725	t20	2:25:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	phad	2:26:48 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	fx5	2:27:48 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	fx1	2:28:48 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	fx2	2:29:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	fx22	2:30:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	gx2	2:31:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	wx1	2:32:47 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000707	t20	2:33:48 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	phad	2:34:48 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	fx5	2:35:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	fx1	2:36:48 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	fx2	2:37:48 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	fx22	2:38:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	gx2	2:39:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	wx1	2:40:47 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000701	t20	2:41:48 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	phad	2:42:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	fx5	2:43:48 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	fx1	2:44:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	fx2	2:45:48 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	fx22	2:46:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	gx2	2:47:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	wx1	2:48:47 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000722	t20	2:49:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	phad	2:50:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	fx5	2:51:48 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	fx1	2:52:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	fx2	2:53:47 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	fx22	2:54:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	gx2	2:55:47 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	wx1	2:56:47 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000717	t20	2:57:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	phad	2:58:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	fx5	2:59:47 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	fx1	3:00:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	fx2	3:01:48 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	fx22	3:02:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	gx2	3:03:47 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	wx1	3:04:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000708	t20	3:05:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

### Errors

#### Expired lot numbers used for Quality Control

352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200109-124  
Instrument name: Phadia-B  
Instrument type: Phadia 250  
Status: Ready  
Started: 09-Jan-20 3:06:46 PM  
Ended: 09-Jan-20 6:18:45 PM  
Approved: Approved (Isabella)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-B  
Instrument type: Phadia 250  
Measured: 23-Dec-19 3:51:37 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	18	0.00			
		3:51:37 PM	19	0.00	OK		
		3:52:36 PM	18	0.00	OK		
		%CV	4.5	>99			
CAL-0.35	0.35	Mean	147	0.35			
		3:53:37 PM	146	0.35	OK		
		3:54:36 PM	148	0.35	OK		
		%CV	0.6	0.7			
CAL-0.70	0.70	Mean	271	0.69			
		3:55:37 PM	274	0.70	OK		
		3:56:36 PM	268	0.68	OK		
		%CV	1.6	1.7			
CAL-3.50	3.50	Mean	1279	3.51			
		3:57:37 PM	1283	3.52	OK		
		3:58:36 PM	1275	3.50	OK		
		%CV	0.5	0.5			
CAL-17.5	17.5	Mean	5583	17.5			
		3:59:37 PM	5591	17.5	OK		
		4:00:36 PM	5575	17.5	OK		
		%CV	0.2	0.2			
CAL-100	100	Mean	17182	100			
		4:01:37 PM	17224	101	OK		
		4:02:36 PM	17140	99.4	OK		
		%CV	0.3	0.9			

Curve data	Value	Logged mean	Difference
ED-20	10.1	9.67	+3.97
ED-50	30.1	29.6	+1.76
ED-80	62.3	62.2	+0.20
Slope	211	216	-2.31

Number of curves included in mean: 52

### Curve Controls

No curve controls in this run

### Quality Controls

No quality controls in this run

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000703	phad	5	3:22:46 PM	36	0.24 PAU/l	---	Negative	
011990000704	phad	5	3:14:46 PM	134	1.59 PAU/l	---	Positive	
011990000705	phad	5	3:30:45 PM	32	0.19 PAU/l	---	Negative	
011990000706	phad	5	4:30:44 PM	63	0.62 PAU/l	---	Positive	
011990000710	phad	5	3:41:45 PM	1105	15.1 PAU/l	---	Positive	
011990000711	phad	5	4:38:43 PM	90	0.98 PAU/l	---	Positive	
011990000712	phad	5	3:06:46 PM	30	0.16 PAU/l	---	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000716	phad	5	4:46:43 PM	36	0.25 PAU/I	---	Negative	
011990000721	phad	5	4:54:42 PM	33	0.20 PAU/I	---	Negative	
011990000728	phad	5	4:22:44 PM	279	3.57 PAU/I	---	Positive	
011990000732	phad	5	4:14:44 PM	87	0.94 PAU/I	---	Positive	
011990000733	phad	5	4:06:44 PM	35	0.23 PAU/I	---	Negative	
011990000738	phad	5	3:58:45 PM	34	0.22 PAU/I	---	Negative	
011990000740	phad	5	5:02:42 PM	43	0.35 PAU/I	---	Negative	
011990000743	phad	5	3:49:45 PM	41	0.32 PAU/I	---	Negative	
011990000754	phad	5	5:19:41 PM	52	0.46 PAU/I	---	Positive	
011990000755	phad	5	5:43:40 PM	41	0.32 PAU/I	---	Negative	
011990000758	phad	5	5:10:42 PM	36	0.25 PAU/I	---	Negative	
011990000759	phad	5	5:49:40 PM	31	0.19 PAU/I	---	Negative	
011990000760	phad	5	5:54:58 PM	---	---	---	---	Err
011990000770	phad	5	5:31:41 PM	105	1.19 PAU/I	---	Positive	
011990000774	phad	5	5:37:40 PM	35	0.23 PAU/I	---	Negative	
011990000775	phad	5	5:25:41 PM	65	0.65 PAU/I	---	Positive	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000703	fx5	5	3:23:46 PM	25	0.10 KUA/I	---	Negative	
011990000703	fx1	5	3:24:46 PM	15	---	---	Negative	
011990000703	fx2	5	3:25:46 PM	18	---	---	Negative	
011990000703	fx22	5	3:26:46 PM	33	0.21 KUA/I	---	Negative	
011990000703	gx2	5	3:27:46 PM	20	0.03 KUA/I	---	Negative	
011990000703	wx1	5	3:28:46 PM	29	0.15 KUA/I	---	Negative	
011990000703	t20	5	3:29:46 PM	15	---	---		
011990000704	fx5	5	3:15:46 PM	29	0.15 KUA/I	---	Negative	
011990000704	fx1	5	3:16:46 PM	24	0.08 KUA/I	---	Negative	
011990000704	fx2	5	3:17:46 PM	49	0.43 KUA/I	---	Positive	
011990000704	fx22	5	3:18:46 PM	38	0.28 KUA/I	---	Negative	
011990000704	gx2	5	3:19:46 PM	26	0.11 KUA/I	---	Negative	
011990000704	wx1	5	3:20:46 PM	118	1.36 KUA/I	---	Positive	
011990000704	t20	5	3:21:46 PM	52	0.47 KUA/I	---		
011990000705	fx5	5	3:31:46 PM	19	0.02 KUA/I	---	Negative	
011990000705	fx1	5	3:32:45 PM	14	---	---	Negative	
011990000705	fx2	5	3:33:45 PM	36	0.25 KUA/I	---	Negative	
011990000705	fx22	5	3:34:45 PM	31	0.18 KUA/I	---	Negative	
011990000705	gx2	5	3:38:45 PM	25	0.09 KUA/I	---	Negative	
011990000705	wx1	5	3:39:45 PM	29	0.15 KUA/I	---	Negative	
011990000705	t20	5	3:40:45 PM	16	---	---		
011990000706	fx5	5	4:31:43 PM	57	0.53 KUA/I	---	Positive	
011990000706	fx1	5	4:32:43 PM	18	---	---	Negative	
011990000706	fx2	5	4:33:43 PM	18	---	---	Negative	
011990000706	fx22	5	4:34:43 PM	33	0.21 KUA/I	---	Negative	
011990000706	gx2	5	4:35:43 PM	20	0.03 KUA/I	---	Negative	
011990000706	wx1	5	4:36:43 PM	56	0.52 KUA/I	---	Positive	
011990000706	t20	5	4:37:43 PM	18	---	---		
011990000710	fx5	5	3:42:45 PM	31	0.19 KUA/I	---	Negative	
011990000710	fx1	5	3:43:45 PM	41	0.32 KUA/I	---	Negative	
011990000710	fx2	5	3:44:45 PM	61	0.59 KUA/I	---	Positive	
011990000710	fx22	5	3:45:45 PM	55	0.51 KUA/I	---	Positive	
011990000710	gx2	5	3:46:45 PM	49	0.42 KUA/I	---	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000710	wx1	5	3:47:45 PM	223	2.80 kUA/l	---	Positive	
011990000710	t20	5	3:48:45 PM	30	0.17 kUA/l	---		
011990000711	fx5	5	4:39:43 PM	21	0.04 kUA/l	---	Negative	
011990000711	fx1	5	4:40:43 PM	16	---	---	Negative	
011990000711	fx2	5	4:41:43 PM	32	0.19 kUA/l	---	Negative	
011990000711	fx22	5	4:42:43 PM	35	0.23 kUA/l	---	Negative	
011990000711	gx2	5	4:43:43 PM	25	0.09 kUA/l	---	Negative	
011990000711	wx1	5	4:44:43 PM	37	0.26 kUA/l	---	Negative	
011990000711	t20	5	4:45:43 PM	20	0.03 kUA/l	---		
011990000712	fx5	5	3:07:46 PM	25	0.10 kUA/l	---	Negative	
011990000712	fx1	5	3:08:46 PM	16	---	---	Negative	
011990000712	fx2	5	3:09:47 PM	16	---	---	Negative	
011990000712	fx22	5	3:10:46 PM	31	0.18 kUA/l	---	Negative	
011990000712	gx2	5	3:11:46 PM	24	0.08 kUA/l	---	Negative	
011990000712	wx1	5	3:12:46 PM	26	0.12 kUA/l	---	Negative	
011990000712	t20	5	3:13:46 PM	16	---	---		
011990000716	fx5	5	4:47:43 PM	43	0.35 kUA/l	---	Negative	
011990000716	fx1	5	4:48:43 PM	15	---	---	Negative	
011990000716	fx2	5	4:49:43 PM	14	---	---	Negative	
011990000716	fx22	5	4:50:42 PM	29	0.15 kUA/l	---	Negative	
011990000716	gx2	5	4:51:42 PM	26	0.11 kUA/l	---	Negative	
011990000716	wx1	5	4:52:42 PM	30	0.16 kUA/l	---	Negative	
011990000716	t20	5	4:53:42 PM	17	---	---		
011990000721	fx5	5	4:55:43 PM	23	0.07 kUA/l	---	Negative	
011990000721	fx1	5	4:56:42 PM	15	---	---	Negative	
011990000721	fx2	5	4:57:42 PM	14	---	---	Negative	
011990000721	fx22	5	4:58:42 PM	34	0.22 kUA/l	---	Negative	
011990000721	gx2	5	4:59:42 PM	20	0.02 kUA/l	---	Negative	
011990000721	wx1	5	5:00:42 PM	29	0.15 kUA/l	---	Negative	
011990000721	t20	5	5:01:42 PM	15	---	---		
011990000728	fx5	5	4:23:44 PM	41	0.32 kUA/l	---	Negative	
011990000728	fx1	5	4:24:44 PM	23	0.07 kUA/l	---	Negative	
011990000728	fx2	5	4:25:44 PM	59	0.56 kUA/l	---	Positive	
011990000728	fx22	5	4:26:43 PM	44	0.36 kUA/l	---	Positive	
011990000728	gx2	5	4:27:43 PM	38	0.28 kUA/l	---	Negative	
011990000728	wx1	5	4:28:44 PM	103	1.16 kUA/l	---	Positive	
011990000728	t20	5	4:29:44 PM	37	0.26 kUA/l	---		
011990000732	fx5	5	4:15:44 PM	33	0.21 kUA/l	---	Negative	
011990000732	fx1	5	4:16:44 PM	20	0.03 kUA/l	---	Negative	
011990000732	fx2	5	4:17:44 PM	13	---	---	Negative	
011990000732	fx22	5	4:18:44 PM	32	0.20 kUA/l	---	Negative	
011990000732	gx2	5	4:19:44 PM	22	0.06 kUA/l	---	Negative	
011990000732	wx1	5	4:20:44 PM	296	3.80 kUA/l	---	Positive	
011990000732	t20	5	4:21:44 PM	41	0.31 kUA/l	---		
011990000733	fx5	5	4:07:44 PM	21	0.04 kUA/l	---	Negative	
011990000733	fx1	5	4:08:44 PM	14	---	---	Negative	
011990000733	fx2	5	4:09:44 PM	12	---	---	Negative	
011990000733	fx22	5	4:10:44 PM	28	0.14 kUA/l	---	Negative	
011990000733	gx2	5	4:11:44 PM	21	0.04 kUA/l	---	Negative	
011990000733	wx1	5	4:12:44 PM	30	0.17 kUA/l	---	Negative	
011990000733	t20	5	4:13:44 PM	16	---	---		
011990000738	fx5	5	3:59:44 PM	20	0.03 kUA/l	---	Negative	
011990000738	fx1	5	4:00:44 PM	18	---	---	Negative	
011990000738	fx2	5	4:01:45 PM	17	---	---	Negative	



## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000738	fx22	5	4:02:45 PM	31	0.17 kUA/I	---	Negative	
011990000738	gx2	5	4:03:45 PM	20	0.04 kUA/I	---	Negative	
011990000738	wx1	5	4:04:44 PM	28	0.14 kUA/I	---	Negative	
011990000738	t20	5	4:05:44 PM	18	---	---		
011990000740	fx5	5	5:03:42 PM	23	0.08 kUA/I	---	Negative	
011990000740	fx1	5	5:04:42 PM	16	---	---	Negative	
011990000740	fx2	5	5:05:42 PM	22	0.06 kUA/I	---	Negative	
011990000740	fx22	5	5:06:42 PM	31	0.18 kUA/I	---	Negative	
011990000740	gx2	5	5:07:42 PM	26	0.11 kUA/I	---	Negative	
011990000740	wx1	5	5:08:42 PM	33	0.21 kUA/I	---	Negative	
011990000740	t20	5	5:09:42 PM	20	0.03 kUA/I	---		
011990000743	fx5	5	3:50:45 PM	32	0.19 kUA/I	---	Negative	
011990000743	fx1	5	3:51:45 PM	22	0.05 kUA/I	---	Negative	
011990000743	fx2	5	3:53:45 PM	14	---	---	Negative	
011990000743	fx22	5	3:54:45 PM	38	0.28 kUA/I	---	Negative	
011990000743	gx2	5	3:55:45 PM	23	0.08 kUA/I	---	Negative	
011990000743	wx1	5	3:56:44 PM	114	1.31 kUA/I	---	Positive	
011990000743	t20	5	3:57:44 PM	49	0.43 kUA/I	---		
011990000754	fx1	5	5:20:41 PM	18	---	---	Negative	
011990000754	fx2	5	5:21:42 PM	17	---	---	Negative	
011990000754	gx2	5	5:22:41 PM	24	0.09 kUA/I	---	Negative	
011990000754	wx1	5	5:23:41 PM	31	0.17 kUA/I	---	Negative	
011990000754	t20	5	5:24:41 PM	17	---	---		
011990000755	fx1	5	5:44:40 PM	18	0.01 kUA/I	---	Negative	
011990000755	fx2	5	5:45:40 PM	15	---	---	Negative	
011990000755	gx2	5	5:46:40 PM	23	0.07 kUA/I	---	Negative	
011990000755	wx1	5	5:47:41 PM	32	0.19 kUA/I	---	Negative	
011990000755	t20	5	5:48:40 PM	18	0.01 kUA/I	---		
011990000758	fx5	5	5:11:42 PM	19	0.02 kUA/I	---	Negative	
011990000758	fx1	5	5:12:42 PM	15	---	---	Negative	
011990000758	fx2	5	5:13:42 PM	12	---	---	Negative	
011990000758	gx2	5	5:15:42 PM	25	0.10 kUA/I	---	Negative	
011990000758	wx1	5	5:17:41 PM	27	0.13 kUA/I	---	Negative	
011990000758	t20	5	5:18:41 PM	15	---	---		
011990000759	fx1	5	5:49:58 PM	---	---	---	---	Err
011990000759	fx2	5	5:50:58 PM	---	---	---	---	Err
011990000759	gx2	5	5:51:58 PM	---	---	---	---	Err
011990000759	wx1	5	5:52:57 PM	---	---	---	---	Err
011990000759	t20	5	5:53:58 PM	---	---	---	---	Err
011990000760	fx1	5	5:55:57 PM	---	---	---	---	Err
011990000760	fx2	5	5:56:57 PM	---	---	---	---	Err
011990000760	gx2	5	5:58:40 PM	26	0.11 kUA/I	---	Negative	
011990000760	wx1	5	5:59:40 PM	31	0.18 kUA/I	---	Negative	
011990000760	t20	5	6:00:40 PM	18	---	---		
011990000770	fx1	5	5:32:41 PM	21	0.04 kUA/I	---	Negative	
011990000770	fx2	5	5:33:41 PM	13	---	---	Negative	
011990000770	gx2	5	5:34:41 PM	20	0.03 kUA/I	---	Negative	
011990000770	wx1	5	5:35:41 PM	31	0.18 kUA/I	---	Negative	
011990000770	t20	5	5:36:41 PM	16	---	---		
011990000774	fx1	5	5:38:40 PM	15	---	---	Negative	
011990000774	fx2	5	5:39:41 PM	12	---	---	Negative	
011990000774	gx2	5	5:40:41 PM	20	0.03 kUA/I	---	Negative	
011990000774	wx1	5	5:41:40 PM	30	0.16 kUA/I	---	Negative	
011990000774	t20	5	5:42:40 PM	16	---	---		

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000775	fx1	5	5:26:41 PM	25	0.10 kUA/I	---	Negative	
011990000775	fx2	5	5:27:41 PM	21	0.05 kUA/I	---	Negative	
011990000775	gx2	5	5:28:41 PM	41	0.32 kUA/I	---	Negative	
011990000775	wx1	5	5:29:41 PM	45	0.37 kUA/I	---	Positive	
011990000775	t20	5	5:30:41 PM	26	0.11 kUA/I	---		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	3:51:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	3:52:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:53:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:54:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:55:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:56:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:57:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:58:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	3:59:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	4:00:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:01:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:02:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000760	phad	5:54:58 PM	610EL 21/07	A1HGR 21/06					
011990000760	fx1	5:55:57 PM	447C1 21/08	A1HGR 21/06					
011990000760	fx2	5:56:57 PM	448BT 21/07	A1HGR 21/06					
011990000760	gx2	5:58:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000760	wx1	5:59:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000760	t20	6:00:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000759	phad	5:49:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000759	fx1	5:49:58 PM	447C1 21/08	A1HGR 21/06					
011990000759	fx2	5:50:58 PM	448BT 21/07	A1HGR 21/06					
011990000759	gx2	5:51:58 PM	433BL 21/08	A1HGR 21/06					
011990000759	wx1	5:52:57 PM	436BB 21/05	A1HGR 21/06					
011990000759	t20	5:53:58 PM	475GG 21/07	A1HGR 21/06					
011990000755	phad	5:43:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000755	fx1	5:44:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000755	fx2	5:45:40 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000755	gx2	5:46:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000755	wx1	5:47:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000755	t20	5:48:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000774	phad	5:37:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000774	fx1	5:38:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000774	fx2	5:39:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000774	gx2	5:40:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000774	wx1	5:41:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000774	t20	5:42:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000770	phad	5:31:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000770	fx1	5:32:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000770	fx2	5:33:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000770	gx2	5:34:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000770	wx1	5:35:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000770	t20	5:36:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000775	phad	5:25:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000775	fx1	5:26:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000775	fx2	5:27:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000775	gx2	5:28:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000775	wx1	5:29:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000775	t20	5:30:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000754	phad	5:19:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000754	fx1	5:20:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000754	fx2	5:21:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000754	gx2	5:22:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000754	wx1	5:23:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000754	t20	5:24:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000758	phad	5:10:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000758	fx5	5:11:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000758	fx1	5:12:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000758	fx2	5:13:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000758	gx2	5:15:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000758	wx1	5:17:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000758	t20	5:18:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	phad	3:06:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	fx5	3:07:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	fx1	3:08:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	fx2	3:09:47 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	fx22	3:10:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	gx2	3:11:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	wx1	3:12:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000712	t20	3:13:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	phad	3:14:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	fx5	3:15:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	fx1	3:16:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	fx2	3:17:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	fx22	3:18:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	gx2	3:19:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	wx1	3:20:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000704	t20	3:21:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	phad	3:22:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	fx5	3:23:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	fx1	3:24:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	fx2	3:25:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	fx22	3:26:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	gx2	3:27:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	wx1	3:28:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000703	t20	3:29:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	phad	3:30:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	fx5	3:31:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	fx1	3:32:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	fx2	3:33:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	fx22	3:34:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	gx2	3:38:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	wx1	3:39:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000705	t20	3:40:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	phad	3:41:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	fx5	3:42:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	fx1	3:43:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	fx2	3:44:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	fx22	3:45:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	gx2	3:46:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	wx1	3:47:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000710	t20	3:48:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	phad	5:02:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	fx5	5:03:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	fx1	5:04:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	fx2	5:05:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	fx22	5:06:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	gx2	5:07:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	wx1	5:08:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000740	t20	5:09:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000721	phad	4:54:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000721	fx5	4:55:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000721	fx1	4:56:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000721	fx2	4:57:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000721	fx22	4:58:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000721	gx2	4:59:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000721	wx1	5:00:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000721	t20	5:01:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	phad	4:46:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	fx5	4:47:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	fx1	4:48:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	fx2	4:49:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	fx22	4:50:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	gx2	4:51:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	wx1	4:52:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000716	t20	4:53:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	phad	4:38:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	fx5	4:39:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	fx1	4:40:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	fx2	4:41:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	fx22	4:42:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	gx2	4:43:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	wx1	4:44:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000711	t20	4:45:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	phad	4:30:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	fx5	4:31:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	fx1	4:32:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	fx2	4:33:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	fx22	4:34:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	gx2	4:35:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	wx1	4:36:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000706	t20	4:37:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	phad	4:22:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	fx5	4:23:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	fx1	4:24:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	fx2	4:25:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	fx22	4:26:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	gx2	4:27:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	wx1	4:28:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000728	t20	4:29:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	phad	4:14:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	fx5	4:15:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	fx1	4:16:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	fx2	4:17:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	fx22	4:18:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	gx2	4:19:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	wx1	4:20:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000732	t20	4:21:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	phad	4:06:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	fx5	4:07:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	fx1	4:08:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	fx2	4:09:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	fx22	4:10:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	gx2	4:11:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	wx1	4:12:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000733	t20	4:13:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	phad	3:58:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	fx5	3:59:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	fx1	4:00:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	fx2	4:01:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	fx22	4:02:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	gx2	4:03:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	wx1	4:04:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000738	t20	4:05:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000743	phad	3:49:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000743	fx5	3:50:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000743	fx1	3:51:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000743	fx2	3:53:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000743	fx22	3:54:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000743	gx2	3:55:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000743	wx1	3:56:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000743	t20	3:57:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

### Errors

Identity	Test	Measured	Error
011990000760	phad	5:54:58 PM	410-Error in stop solution pipetting 1 module
011990000760	fx1	5:55:57 PM	306-Shortage of stop solution
011990000760	fx2	5:56:57 PM	306-Shortage of stop solution
011990000759	fx1	5:49:58 PM	306-Shortage of stop solution
011990000759	fx2	5:50:58 PM	412-Error in stop solution pipetting 3 module
011990000759	gx2	5:51:58 PM	306-Shortage of stop solution
011990000759	wx1	5:52:57 PM	411-Error in stop solution pipetting 2 module
011990000759	t20	5:53:58 PM	306-Shortage of stop solution