


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200116-130  
Instrument name: Phadia-B  
Instrument type: Phadia 250  
Status: Ready  
Started: 16-Jan-20 2:41:41 PM  
Ended: 16-Jan-20 3:53:54 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-B  
Instrument type: Phadia 250  
Measured: 23-Dec-19 3:51:37 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	18	0.00			
		3:51:37 PM	19	0.00	OK		
		3:52:36 PM	18	0.00	OK		
		%CV	4.5	>99			
CAL-0.35	0.35	Mean	147	0.35			
		3:53:37 PM	146	0.35	OK		
		3:54:36 PM	148	0.35	OK		
		%CV	0.6	0.7			
CAL-0.70	0.70	Mean	271	0.69			
		3:55:37 PM	274	0.70	OK		
		3:56:36 PM	268	0.68	OK		
		%CV	1.6	1.7			
CAL-3.50	3.50	Mean	1279	3.51			
		3:57:37 PM	1283	3.52	OK		
		3:58:36 PM	1275	3.50	OK		
		%CV	0.5	0.5			
CAL-17.5	17.5	Mean	5583	17.5			
		3:59:37 PM	5591	17.5	OK		
		4:00:36 PM	5575	17.5	OK		
		%CV	0.2	0.2			
CAL-100	100	Mean	17182	100			
		4:01:37 PM	17224	101	OK		
		4:02:36 PM	17140	99.4	OK		
		%CV	0.3	0.9			

Curve data	Value	Logged mean	Difference
ED-20	10.1	9.67	+3.97
ED-50	30.1	29.6	+1.76
ED-80	62.3	62.2	+0.20
Slope	211	216	-2.31

Number of curves included in mean: 52

### Curve Controls

No curve controls in this run

### Quality Controls

No quality controls in this run

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000369	phad	5	2:46:41 PM	50	0.44 PAU/l	1	Positive	
011990000370	phad	5	2:54:40 PM	115	1.33 PAU/l	2	Positive	
011990000380	phad	5	3:02:40 PM	256	3.25 PAU/l	2	Positive	
011990000390	phad	5	3:10:40 PM	37	0.26 PAU/l	0	Negative	
011990000918	phad	5	3:34:39 PM	28	0.14 PAU/l	0	Negative	
011990000923	phad	5	3:18:40 PM	29	0.16 PAU/l	0	Negative	
011990000924	phad	5	3:26:39 PM	69	0.70 PAU/l	2	Positive	

## Laboratory Report

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000360	fx2	5	2:41:41 PM	3004	43.4 kUA/l	4	Positive	
011990000360	fx22	5	2:42:41 PM	105	1.19 kUA/l	2	Positive	
011990000360	gx2	5	2:43:41 PM	164	1.99 kUA/l	2	Positive	
011990000360	wx1	5	2:44:41 PM	357	4.64 kUA/l	3	Positive	
011990000360	t20	5	2:45:41 PM	108	1.24 kUA/l	2		
011990000369	fx5	5	2:47:41 PM	23	0.06 kUA/l	0	Negative	
011990000369	fx1	5	2:48:40 PM	18	Low RU	0	Negative	
011990000369	fx2	5	2:49:41 PM	24	0.08 kUA/l	0	Negative	
011990000369	fx22	5	2:50:41 PM	32	0.20 kUA/l	0	Negative	
011990000369	gx2	5	2:51:41 PM	23	0.07 kUA/l	0	Negative	
011990000369	wx1	5	2:52:41 PM	44	0.36 kUA/l	1	Positive	
011990000369	t20	5	2:53:40 PM	17	Low RU	0		
011990000370	fx5	5	2:55:40 PM	37	0.26 kUA/l	0	Negative	
011990000370	fx1	5	2:56:40 PM	42	0.33 kUA/l	0	Negative	
011990000370	fx2	5	2:57:40 PM	56	0.52 kUA/l	1	Positive	
011990000370	fx22	5	2:58:40 PM	37	0.27 kUA/l	0	Negative	
011990000370	gx2	5	2:59:40 PM	29	0.16 kUA/l	0	Negative	
011990000370	wx1	5	3:00:40 PM	38	0.28 kUA/l	0	Negative	
011990000370	t20	5	3:01:40 PM	19	0.02 kUA/l	0		
011990000380	fx5	5	3:03:40 PM	27	0.12 kUA/l	0	Negative	
011990000380	fx1	5	3:04:40 PM	15	Low RU	0	Negative	
011990000380	fx2	5	3:05:40 PM	27	0.13 kUA/l	0	Negative	
011990000380	fx22	5	3:06:40 PM	32	0.19 kUA/l	0	Negative	
011990000380	gx2	5	3:07:40 PM	245	3.10 kUA/l	2	Positive	
011990000380	wx1	5	3:08:40 PM	28	0.14 kUA/l	0	Negative	
011990000380	t20	5	3:09:40 PM	16	Low RU	0		
011990000390	fx5	5	3:11:40 PM	83	0.89 kUA/l	2	Positive	
011990000390	fx1	5	3:12:39 PM	18	Low RU	0	Negative	
011990000390	fx2	5	3:13:40 PM	56	0.52 kUA/l	1	Positive	
011990000390	fx22	5	3:14:39 PM	31	0.18 kUA/l	0	Negative	
011990000390	gx2	5	3:15:40 PM	17	---	---	Negative	
011990000390	wx1	5	3:16:40 PM	31	0.19 kUA/l	0	Negative	
011990000390	t20	5	3:17:40 PM	16	Low RU	0		
011990000918	fx5	5	3:35:39 PM	23	0.06 kUA/l	0	Negative	
011990000923	fx5	5	3:19:39 PM	18	Low RU	0	Negative	
011990000923	fx1	5	3:20:39 PM	14	Low RU	0	Negative	
011990000923	fx2	5	3:21:39 PM	12	Low RU	0	Negative	
011990000923	fx22	5	3:22:39 PM	27	0.12 kUA/l	0	Negative	
011990000923	gx2	5	3:23:39 PM	18	---	---	Negative	
011990000923	wx1	5	3:24:39 PM	26	0.11 kUA/l	0	Negative	
011990000923	t20	5	3:25:39 PM	16	Low RU	0		
011990000924	fx5	5	3:27:39 PM	20	0.03 kUA/l	0	Negative	
011990000924	fx1	5	3:28:39 PM	16	Low RU	0	Negative	
011990000924	fx2	5	3:29:39 PM	12	Low RU	0	Negative	
011990000924	fx22	5	3:30:39 PM	25	0.10 kUA/l	0	Negative	
011990000924	gx2	5	3:31:39 PM	18	---	---	Negative	
011990000924	wx1	5	3:32:39 PM	27	0.12 kUA/l	0	Negative	
011990000924	t20	5	3:33:39 PM	16	Low RU	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	3:51:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

## Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	3:52:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:53:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:54:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:55:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:56:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:57:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:58:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	3:59:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	4:00:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:01:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:02:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000360	fx2	2:41:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000360	fx22	2:42:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000360	gx2	2:43:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000360	wx1	2:44:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000360	t20	2:45:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	phad	2:46:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	fx5	2:47:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	fx1	2:48:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	fx2	2:49:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	fx22	2:50:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	gx2	2:51:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	wx1	2:52:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000369	t20	2:53:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	phad	2:54:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	fx5	2:55:40 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	fx1	2:56:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	fx2	2:57:40 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	fx22	2:58:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	gx2	2:59:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	wx1	3:00:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000370	t20	3:01:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	phad	3:02:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	fx5	3:03:40 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	fx1	3:04:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	fx2	3:05:40 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	fx22	3:06:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	gx2	3:07:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	wx1	3:08:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000380	t20	3:09:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	phad	3:10:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	fx5	3:11:40 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	fx1	3:12:39 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	fx2	3:13:40 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	fx22	3:14:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	gx2	3:15:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	wx1	3:16:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000390	t20	3:17:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	phad	3:18:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	fx5	3:19:39 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	fx1	3:20:39 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	fx2	3:21:39 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	fx22	3:22:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	gx2	3:23:39 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	wx1	3:24:39 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000923	t20	3:25:39 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000924	phad	3:26:39 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000924	fx5	3:27:39 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000924	fx1	3:28:39 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000924	fx2	3:29:39 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000924	fx22	3:30:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000924	gx2	3:31:39 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000924	wx1	3:32:39 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000924	t20	3:33:39 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000918	phad	3:34:39 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000918	fx5	3:35:39 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	