


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200116-380
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 16-Jan-20 10:11:24 AM
Ended: 16-Jan-20 2:12:17 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	10:11:24 AM	319	0.70	OK	
CC-2	17.5	10:12:24 AM	6374	17.7	OK	

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	10:15:24 AM	4356	11.3 kUA/l	CZ1AR	OK	
slgE L	e1	1	10:16:24 AM	1171	2.72 kUA/l	CZ4AN	OK	
slgE Neg	d1	1	10:17:24 AM	20	0.01 kUA/l	BL1CP	OK	
IQC271119	d1	1	10:18:24 AM	6248	17.3 kUA/l	<lab>	----	
			Expected value:	----				
			Expected SD:	----				
			SD limit:	----				
			Range min:	----				
			Range max:	----				
			Limit CV%:	----				

Laboratory Report

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000351	phad	5	12:39:18 PM	34	0.20 PAU/I	0	Negative	
011990000352	phad	5	12:03:20 PM	113	1.12 PAU/I	2	Positive	
011990000356	phad	5	1:34:16 PM	58	0.48 PAU/I	1	Positive	
011990000361	phad	5	12:11:19 PM	38	0.26 PAU/I	0	Negative	
011990000362	phad	5	11:55:20 AM	35	0.22 PAU/I	0	Negative	
011990000366	phad	5	1:41:16 PM	95	0.91 PAU/I	2	Positive	
011990000371	phad	5	1:19:17 PM	33	0.20 PAU/I	0	Negative	
011990000372	phad	5	12:47:18 PM	32	0.18 PAU/I	0	Negative	
011990000373	phad	5	11:47:20 AM	35	0.21 PAU/I	0	Negative	
011990000376	phad	5	1:48:16 PM	568	6.41 PAU/I	3	Positive	
011990000377	phad	5	1:27:17 PM	37	0.24 PAU/I	0	Negative	
011990000381	phad	5	11:39:21 AM	36	0.23 PAU/I	0	Negative	
011990000382	phad	5	12:55:18 PM	30	0.16 PAU/I	0	Negative	
011990000386	phad	5	1:55:16 PM	32	0.18 PAU/I	0	Negative	
011990000391	phad	5	1:03:18 PM	53	0.43 PAU/I	1	Positive	
011990000392	phad	5	1:11:17 PM	38	0.25 PAU/I	0	Negative	
011990000396	phad	5	2:02:15 PM	48	0.36 PAU/I	1	Positive	
011990000726	phad	5	11:31:21 AM	126	1.28 PAU/I	2	Positive	
011990000727	phad	5	11:23:21 AM	35	0.22 PAU/I	0	Negative	
011990000731	phad	5	11:07:22 AM	32	0.18 PAU/I	0	Negative	
011990000736	phad	5	11:15:22 AM	104	1.01 PAU/I	2	Positive	
011990000737	phad	5	10:21:24 AM	1511	17.7 PAU/I	4	Positive	
011990000741	phad	5	10:45:23 AM	34	0.20 PAU/I	0	Negative	
011990000742	phad	5	10:29:24 AM	26	0.11 PAU/I	0	Negative	
011990000746	phad	5	10:53:22 AM	32	0.18 PAU/I	0	Negative	
011990000747	phad	5	10:37:23 AM	27	0.12 PAU/I	0	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000351	fx5	5	12:40:19 PM	18	0.02 kUA/I	0	Negative	
011990000351	fx1	5	12:41:18 PM	16	Low RU	0	Negative	
011990000351	fx2	5	12:42:18 PM	13	Low RU	0	Negative	
011990000351	fx22	5	12:43:18 PM	28	0.13 kUA/I	0	Negative	
011990000351	gx2	5	12:44:18 PM	20	0.04 kUA/I	0	Negative	
011990000351	wx1	5	12:45:18 PM	27	0.13 kUA/I	0	Negative	
011990000351	t20	5	12:46:18 PM	17	Low RU	0		
011990000352	fx5	5	12:04:20 PM	23	0.07 kUA/I	0	Negative	
011990000352	fx1	5	12:05:20 PM	22	0.07 kUA/I	0	Negative	
011990000352	fx2	5	12:06:19 PM	74	0.67 kUA/I	1	Positive	
011990000352	fx22	5	12:07:20 PM	33	0.20 kUA/I	0	Negative	
011990000352	gx2	5	12:08:19 PM	21	0.05 kUA/I	0	Negative	
011990000352	wx1	5	12:09:19 PM	72	0.65 kUA/I	1	Positive	
011990000352	t20	5	12:10:19 PM	29	0.15 kUA/I	0		
011990000356	fx5	5	1:35:16 PM	24	0.09 kUA/I	0	Negative	
011990000356	fx1	5	1:36:17 PM	19	0.03 kUA/I	0	Negative	
011990000356	fx2	5	1:37:17 PM	24	0.09 kUA/I	0	Negative	
011990000356	gx2	5	1:38:16 PM	32	0.18 kUA/I	0	Negative	
011990000356	wx1	5	1:39:16 PM	35	0.22 kUA/I	0	Negative	
011990000356	t20	5	1:40:16 PM	23	0.07 kUA/I	0		
011990000361	fx5	5	12:12:20 PM	21	0.05 kUA/I	0	Negative	
011990000361	fx1	5	12:13:19 PM	21	0.06 kUA/I	0	Negative	
011990000361	fx2	5	12:14:19 PM	25	0.10 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000361	fx22	5	12:35:19 PM	33	0.19 kUA/l	0	Negative	
011990000361	gx2	5	12:36:18 PM	25	0.10 kUA/l	0	Negative	
011990000361	wx1	5	12:37:18 PM	29	0.15 kUA/l	0	Negative	
011990000361	t20	5	12:38:18 PM	17	Low RU	0		
011990000362	fx5	5	11:56:20 AM	21	0.06 kUA/l	0	Negative	
011990000362	fx1	5	11:57:20 AM	15	Low RU	0	Negative	
011990000362	fx2	5	11:58:20 AM	13	Low RU	0	Negative	
011990000362	fx22	5	11:59:20 AM	27	0.13 kUA/l	0	Negative	
011990000362	gx2	5	12:00:20 PM	20	0.04 kUA/l	0	Negative	
011990000362	wx1	5	12:01:20 PM	28	0.13 kUA/l	0	Negative	
011990000362	t20	5	12:02:20 PM	17	0.01 kUA/l	0		
011990000366	fx5	5	1:42:16 PM	20	0.03 kUA/l	0	Negative	
011990000366	fx1	5	1:43:16 PM	20	0.03 kUA/l	0	Negative	
011990000366	fx2	5	1:44:16 PM	25	0.10 kUA/l	0	Negative	
011990000366	gx2	5	1:45:16 PM	22	0.06 kUA/l	0	Negative	
011990000366	wx1	5	1:46:16 PM	29	0.14 kUA/l	0	Negative	
011990000366	t20	5	1:47:16 PM	17	Low RU	0		
011990000371	fx5	5	1:20:17 PM	26	0.11 kUA/l	0	Negative	
011990000371	fx1	5	1:21:17 PM	18	0.01 kUA/l	0	Negative	
011990000371	fx2	5	1:22:17 PM	14	Low RU	0	Negative	
011990000371	fx22	5	1:23:17 PM	29	0.14 kUA/l	0	Negative	
011990000371	gx2	5	1:24:17 PM	23	0.07 kUA/l	0	Negative	
011990000371	wx1	5	1:25:17 PM	27	0.12 kUA/l	0	Negative	
011990000371	t20	5	1:26:17 PM	19	0.03 kUA/l	0		
011990000372	fx5	5	12:48:18 PM	23	0.08 kUA/l	0	Negative	
011990000372	fx1	5	12:49:18 PM	17	Low RU	0	Negative	
011990000372	fx2	5	12:50:18 PM	12	Low RU	0	Negative	
011990000372	fx22	5	12:51:18 PM	24	0.09 kUA/l	0	Negative	
011990000372	gx2	5	12:52:18 PM	23	0.08 kUA/l	0	Negative	
011990000372	wx1	5	12:53:18 PM	29	0.14 kUA/l	0	Negative	
011990000372	t20	5	12:54:18 PM	16	Low RU	0		
011990000373	fx5	5	11:48:20 AM	37	0.23 kUA/l	0	Negative	
011990000373	fx1	5	11:49:20 AM	19	0.02 kUA/l	0	Negative	
011990000373	fx2	5	11:50:20 AM	16	Low RU	0	Negative	
011990000373	fx22	5	11:51:20 AM	35	0.22 kUA/l	0	Negative	
011990000373	gx2	5	11:52:20 AM	35	0.22 kUA/l	0	Negative	
011990000373	wx1	5	11:53:20 AM	27	0.13 kUA/l	0	Negative	
011990000373	t20	5	11:54:20 AM	15	Low RU	0		
011990000376	fx5	5	1:49:16 PM	47	0.36 kUA/l	1	Positive	
011990000376	fx1	5	1:50:16 PM	27	0.12 kUA/l	0	Negative	
011990000376	fx2	5	1:51:16 PM	126	1.28 kUA/l	2	Positive	
011990000376	gx2	5	1:52:16 PM	237	2.56 kUA/l	2	Positive	
011990000376	wx1	5	1:53:16 PM	89	0.85 kUA/l	2	Positive	
011990000376	t20	5	1:54:16 PM	25	0.09 kUA/l	0		
011990000377	fx5	5	1:28:17 PM	24	0.09 kUA/l	0	Negative	
011990000377	fx1	5	1:29:17 PM	16	Low RU	0	Negative	
011990000377	fx2	5	1:30:17 PM	15	Low RU	0	Negative	
011990000377	gx2	5	1:31:17 PM	21	0.05 kUA/l	0	Negative	
011990000377	wx1	5	1:32:16 PM	27	0.12 kUA/l	0	Negative	
011990000377	t20	5	1:33:16 PM	17	Low RU	0		
011990000381	fx5	5	11:40:20 AM	20	0.04 kUA/l	0	Negative	
011990000381	fx1	5	11:41:20 AM	17	Low RU	0	Negative	
011990000381	fx2	5	11:42:20 AM	11	Low RU	0	Negative	
011990000381	fx22	5	11:43:20 AM	30	0.16 kUA/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000381	gx2	5	11:44:21 AM	21	0.05 kUA/l	0	Negative	
011990000381	wx1	5	11:45:20 AM	28	0.13 kUA/l	0	Negative	
011990000381	t20	5	11:46:20 AM	18	0.01 kUA/l	0		
011990000382	fx5	5	12:56:18 PM	20	0.04 kUA/l	0	Negative	
011990000382	fx1	5	12:57:17 PM	17	Low RU	0	Negative	
011990000382	fx2	5	12:58:18 PM	12	Low RU	0	Negative	
011990000382	fx22	5	12:59:18 PM	29	0.15 kUA/l	0	Negative	
011990000382	gx2	5	1:00:18 PM	24	0.09 kUA/l	0	Negative	
011990000382	wx1	5	1:01:18 PM	34	0.20 kUA/l	0	Negative	
011990000382	t20	5	1:02:17 PM	16	Low RU	0		
011990000386	fx5	5	1:56:16 PM	24	0.08 kUA/l	0	Negative	
011990000386	fx1	5	1:57:16 PM	16	Low RU	0	Negative	
011990000386	fx2	5	1:58:16 PM	12	Low RU	0	Negative	
011990000386	gx2	5	1:59:16 PM	20	0.04 kUA/l	0	Negative	
011990000386	wx1	5	2:00:15 PM	28	0.13 kUA/l	0	Negative	
011990000386	t20	5	2:01:15 PM	18	0.02 kUA/l	0		
011990000391	fx5	5	1:04:18 PM	21	0.05 kUA/l	0	Negative	
011990000391	fx1	5	1:05:17 PM	18	0.02 kUA/l	0	Negative	
011990000391	fx2	5	1:06:17 PM	21	0.05 kUA/l	0	Negative	
011990000391	fx22	5	1:07:17 PM	30	0.15 kUA/l	0	Negative	
011990000391	gx2	5	1:08:18 PM	22	0.06 kUA/l	0	Negative	
011990000391	wx1	5	1:09:18 PM	32	0.18 kUA/l	0	Negative	
011990000391	t20	5	1:10:17 PM	17	Low RU	0		
011990000392	fx5	5	1:12:17 PM	34	0.21 kUA/l	0	Negative	
011990000392	fx1	5	1:13:17 PM	22	0.06 kUA/l	0	Negative	
011990000392	fx2	5	1:14:17 PM	13	Low RU	0	Negative	
011990000392	fx22	5	1:15:17 PM	29	0.15 kUA/l	0	Negative	
011990000392	gx2	5	1:16:17 PM	23	0.07 kUA/l	0	Negative	
011990000392	wx1	5	1:17:18 PM	29	0.15 kUA/l	0	Negative	
011990000392	t20	5	1:18:17 PM	17	Low RU	0		
011990000726	fx5	5	11:32:21 AM	75	0.68 kUA/l	1	Positive	
011990000726	fx1	5	11:33:21 AM	18	0.01 kUA/l	0	Negative	
011990000726	fx2	5	11:34:21 AM	108	1.07 kUA/l	2	Positive	
011990000726	fx22	5	11:35:21 AM	36	0.23 kUA/l	0	Negative	
011990000726	gx2	5	11:36:21 AM	32	0.18 kUA/l	0	Negative	
011990000726	wx1	5	11:37:21 AM	51	0.40 kUA/l	1	Positive	
011990000726	t20	5	11:38:20 AM	32	0.18 kUA/l	0		
011990000727	fx5	5	11:24:22 AM	21	0.05 kUA/l	0	Negative	
011990000727	fx1	5	11:25:21 AM	16	Low RU	0	Negative	
011990000727	fx2	5	11:26:21 AM	13	Low RU	0	Negative	
011990000727	fx22	5	11:27:21 AM	33	0.19 kUA/l	0	Negative	
011990000727	gx2	5	11:28:21 AM	25	0.10 kUA/l	0	Negative	
011990000727	wx1	5	11:29:21 AM	207	2.21 kUA/l	2	Positive	
011990000727	t20	5	11:30:21 AM	18	0.02 kUA/l	0		
011990000731	fx5	5	11:08:22 AM	18	0.02 kUA/l	0	Negative	
011990000731	fx1	5	11:09:22 AM	15	Low RU	0	Negative	
011990000731	fx2	5	11:10:22 AM	11	Low RU	0	Negative	
011990000731	fx22	5	11:11:22 AM	33	0.19 kUA/l	0	Negative	
011990000731	gx2	5	11:12:22 AM	21	0.05 kUA/l	0	Negative	
011990000731	wx1	5	11:13:21 AM	28	0.13 kUA/l	0	Negative	
011990000731	t20	5	11:14:21 AM	17	Low RU	0		
011990000736	fx5	5	11:16:22 AM	26	0.11 kUA/l	0	Negative	
011990000736	fx1	5	11:17:21 AM	18	0.01 kUA/l	0	Negative	
011990000736	fx2	5	11:18:21 AM	55	0.45 kUA/l	1	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000736	fx22	5	11:19:21 AM	33	0.20 kUA/l	0	Negative	
011990000736	gx2	5	11:20:21 AM	23	0.07 kUA/l	0	Negative	
011990000736	wx1	5	11:21:21 AM	30	0.16 kUA/l	0	Negative	
011990000736	t20	5	11:22:21 AM	19	0.03 kUA/l	0		
011990000737	fx5	5	10:22:24 AM	65	0.57 kUA/l	1	Positive	
011990000737	fx1	5	10:23:23 AM	43	0.31 kUA/l	0	Negative	
011990000737	fx2	5	10:24:23 AM	588	6.64 kUA/l	3	Positive	
011990000737	fx22	5	10:25:24 AM	52	0.42 kUA/l	1	Positive	
011990000737	gx2	5	10:26:23 AM	211	2.26 kUA/l	2	Positive	
011990000737	wx1	5	10:27:23 AM	135	1.38 kUA/l	2	Positive	
011990000737	t20	5	10:28:23 AM	99	0.96 kUA/l	2		
011990000741	fx5	5	10:46:23 AM	20	0.04 kUA/l	0	Negative	
011990000741	fx1	5	10:47:23 AM	16	Low RU	0	Negative	
011990000741	fx2	5	10:48:23 AM	13	Low RU	0	Negative	
011990000741	fx22	5	10:49:23 AM	30	0.15 kUA/l	0	Negative	
011990000741	gx2	5	10:50:23 AM	20	0.04 kUA/l	0	Negative	
011990000741	wx1	5	10:51:22 AM	89	0.84 kUA/l	2	Positive	
011990000741	t20	5	10:52:22 AM	17	Low RU	0		
011990000742	fx5	5	10:30:24 AM	19	0.03 kUA/l	0	Negative	
011990000742	fx1	5	10:31:23 AM	15	Low RU	0	Negative	
011990000742	fx2	5	10:32:23 AM	11	Low RU	0	Negative	
011990000742	fx22	5	10:33:23 AM	28	0.13 kUA/l	0	Negative	
011990000742	gx2	5	10:34:23 AM	23	0.08 kUA/l	0	Negative	
011990000742	wx1	5	10:35:23 AM	28	0.13 kUA/l	0	Negative	
011990000742	t20	5	10:36:23 AM	16	Low RU	0		
011990000746	fx5	5	11:00:22 AM	19	0.02 kUA/l	0	Negative	
011990000746	fx1	5	11:01:22 AM	16	Low RU	0	Negative	
011990000746	fx2	5	11:02:22 AM	11	Low RU	0	Negative	
011990000746	fx22	5	11:03:22 AM	29	0.14 kUA/l	0	Negative	
011990000746	gx2	5	11:04:22 AM	19	0.03 kUA/l	0	Negative	
011990000746	wx1	5	11:05:22 AM	26	0.11 kUA/l	0	Negative	
011990000746	t20	5	11:06:22 AM	16	Low RU	0		
011990000747	fx5	5	10:38:23 AM	18	0.02 kUA/l	0	Negative	
011990000747	fx1	5	10:39:23 AM	15	Low RU	0	Negative	
011990000747	fx2	5	10:40:23 AM	11	Low RU	0	Negative	
011990000747	fx22	5	10:41:23 AM	29	0.14 kUA/l	0	Negative	
011990000747	gx2	5	10:42:23 AM	19	0.02 kUA/l	0	Negative	
011990000747	wx1	5	10:43:23 AM	27	0.13 kUA/l	0	Negative	
011990000747	t20	5	10:44:23 AM	18	0.01 kUA/l	0		
012000500914A	f37	1	2:08:15 PM	166	0.35 kUA/l			
012000500914A	f23	1	2:09:15 PM	4137	10.6 kUA/l			
012000500914A	f41	1	2:10:15 PM	52	0.08 kUA/l			
012000500914A	f40	1	2:11:15 PM	149	0.31 kUA/l			

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	10:11:24 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	10:12:24 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	10:15:24 AM	352D0 19/12/	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
slgE L	e1	10:16:24 AM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06/		Exp
slgE Neg	d1	10:17:24 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12/		Exp
IQC271119	d1	10:18:24 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012000500914A	f37	2:08:15 PM	454BN 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012000500914A	f23	2:09:15 PM	423CK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012000500914A	f41	2:10:15 PM	456CN 20/03	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012000500914A	f40	2:11:15 PM	455DT 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000737	phad	10:21:24 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	fx5	10:22:24 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	fx1	10:23:23 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	fx2	10:24:23 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	fx22	10:25:24 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	gx2	10:26:23 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	wx1	10:27:23 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	t20	10:28:23 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000737	phad	10:29:24 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000742	fx5	10:30:24 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000742	fx1	10:31:23 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000742	fx2	10:32:23 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000742	fx22	10:33:23 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000742	gx2	10:34:23 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000742	wx1	10:35:23 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000742	t20	10:36:23 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	phad	10:37:23 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	fx5	10:38:23 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	fx1	10:39:23 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	fx2	10:40:23 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	fx22	10:41:23 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	gx2	10:42:23 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	wx1	10:43:23 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000747	t20	10:44:23 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	phad	10:45:23 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	fx5	10:46:23 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	fx1	10:47:23 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	fx2	10:48:23 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	fx22	10:49:23 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	gx2	10:50:23 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	wx1	10:51:22 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000741	t20	10:52:22 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	phad	10:53:22 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	fx5	11:00:22 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	fx1	11:01:22 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	fx2	11:02:22 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	fx22	11:03:22 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	gx2	11:04:22 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	wx1	11:05:22 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000746	t20	11:06:22 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000731	phad	11:07:22 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000731	fx5	11:08:22 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000731	fx1	11:09:22 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000731	fx2	11:10:22 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000731	fx22	11:11:22 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000731	gx2	11:12:22 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000731	wx1	11:13:21 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000731	t20	11:14:21 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	phad	11:15:22 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	fx5	11:16:22 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	fx1	11:17:21 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	fx2	11:18:21 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	fx22	11:19:21 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	gx2	11:20:21 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	wx1	11:21:21 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000736	t20	11:22:21 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	phad	11:23:21 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	fx5	11:24:22 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	fx1	11:25:21 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	fx2	11:26:21 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	fx22	11:27:21 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	gx2	11:28:21 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	wx1	11:29:21 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000727	t20	11:30:21 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	phad	11:31:21 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	fx5	11:32:21 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	fx1	11:33:21 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	fx2	11:34:21 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	fx22	11:35:21 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	gx2	11:36:21 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	wx1	11:37:21 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000726	t20	11:38:20 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	phad	11:39:21 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	fx5	11:40:20 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	fx1	11:41:20 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	fx2	11:42:20 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	fx22	11:43:20 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	gx2	11:44:21 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	wx1	11:45:20 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000381	t20	11:46:20 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	phad	11:47:20 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	fx5	11:48:20 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	fx1	11:49:20 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	fx2	11:50:20 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	fx22	11:51:20 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	gx2	11:52:20 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	wx1	11:53:20 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000373	t20	11:54:20 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	phad	11:55:20 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	fx5	11:56:20 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	fx1	11:57:20 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	fx2	11:58:20 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	fx22	11:59:20 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	gx2	12:00:20 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	wx1	12:01:20 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000362	t20	12:02:20 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	phad	12:03:20 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	fx5	12:04:20 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	fx1	12:05:20 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	fx2	12:06:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	fx22	12:07:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	gx2	12:08:19 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	wx1	12:09:19 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000352	t20	12:10:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000361	phad	12:11:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000361	fx5	12:12:20 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000361	fx1	12:13:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000361	fx2	12:14:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000361	fx22	12:35:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000361	gx2	12:36:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000361	wx1	12:37:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000361	t20	12:38:18 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	phad	12:39:18 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	fx5	12:40:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	fx1	12:41:18 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	fx2	12:42:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	fx22	12:43:18 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	gx2	12:44:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	wx1	12:45:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000351	t20	12:46:18 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	phad	12:47:18 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	fx5	12:48:18 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	fx1	12:49:18 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	fx2	12:50:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	fx22	12:51:18 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	gx2	12:52:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	wx1	12:53:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000372	t20	12:54:18 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	phad	12:55:18 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	fx5	12:56:18 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	fx1	12:57:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	fx2	12:58:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	fx22	12:59:18 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	gx2	1:00:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	wx1	1:01:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000382	t20	1:02:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	phad	1:03:18 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	fx5	1:04:18 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	fx1	1:05:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	fx2	1:06:17 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	fx22	1:07:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	gx2	1:08:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	wx1	1:09:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000391	t20	1:10:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	phad	1:11:17 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	fx5	1:12:17 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	fx1	1:13:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	fx2	1:14:17 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	fx22	1:15:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	gx2	1:16:17 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	wx1	1:17:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000392	t20	1:18:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	phad	1:19:17 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	fx5	1:20:17 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	fx1	1:21:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	fx2	1:22:17 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	fx22	1:23:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	gx2	1:24:17 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	wx1	1:25:17 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000371	t20	1:26:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000377	phad	1:27:17 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000377	fx5	1:28:17 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000377	fx1	1:29:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000377	fx2	1:30:17 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000377	gx2	1:31:17 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000377	wx1	1:32:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000377	t20	1:33:16 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000356	phad	1:34:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000356	fx5	1:35:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000356	fx1	1:36:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000356	fx2	1:37:17 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000356	gx2	1:38:16 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000356	wx1	1:39:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000356	t20	1:40:16 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	phad	1:41:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	fx5	1:42:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	fx1	1:43:16 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	fx2	1:44:16 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	gx2	1:45:16 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	wx1	1:46:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	t20	1:47:16 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	phad	1:48:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	fx5	1:49:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	fx1	1:50:16 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	fx2	1:51:16 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	gx2	1:52:16 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	wx1	1:53:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	t20	1:54:16 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	phad	1:55:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	fx5	1:56:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	fx1	1:57:16 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	fx2	1:58:16 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	gx2	1:59:16 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	wx1	2:00:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	t20	2:01:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	phad	2:02:15 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	

Errors

Expired lot numbers used for Quality Control

352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200116-380
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 16-Jan-20 2:12:15 PM
Ended: 16-Jan-20 4:25:31 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	3:21:13 PM	5957	16.3 kUA/l	<lab>	----	
			Expected value:	---				
			Expected SD:	---				
			SD limit:	---				
			Range min:	---				
			Range max:	---				
			Limit CV%:	---				

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
----------	------	-----	----------	------	------	-------	---------	---------

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000355	phad	5	2:24:15 PM	38	0.25 PAU/I	0	Negative	
011990000365	phad	5	2:32:14 PM	34	0.20 PAU/I	0	Negative	
011990000375	phad	5	2:40:14 PM	227	2.44 PAU/I	2	Positive	
011990000385	phad	5	2:48:14 PM	42	0.30 PAU/I	0	Negative	
011990000508	phad	5	2:56:13 PM	30	0.16 PAU/I	0	Negative	
011990000509	phad	5	3:31:12 PM	65	0.56 PAU/I	1	Positive	
011990000519	phad	5	3:23:12 PM	438	4.89 PAU/I	3	Positive	
011990000520	phad	5	4:03:11 PM	41	0.29 PAU/I	0	Negative	
011990000529	phad	5	3:39:12 PM	33	0.19 PAU/I	0	Negative	
011990000530	phad	5	4:11:10 PM	103	1.01 PAU/I	2	Positive	
011990000539	phad	5	3:14:13 PM	45	0.33 PAU/I	0	Negative	
011990000540	phad	5	3:47:11 PM	43	0.31 PAU/I	0	Negative	
011990000549	phad	5	3:04:13 PM	33	0.19 PAU/I	0	Negative	
011990000550	phad	5	3:55:11 PM	32	0.18 PAU/I	0	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000355	fx5	5	2:25:15 PM	22	0.06 kUA/I	0	Negative	
011990000355	fx1	5	2:26:15 PM	18	0.01 kUA/I	0	Negative	
011990000355	fx2	5	2:27:15 PM	15	Low RU	0	Negative	
011990000355	fx22	5	2:28:15 PM	29	0.14 kUA/I	0	Negative	
011990000355	gx2	5	2:29:15 PM	21	0.05 kUA/I	0	Negative	
011990000355	wx1	5	2:30:14 PM	28	0.14 kUA/I	0	Negative	
011990000355	t20	5	2:31:14 PM	16	Low RU	0		
011990000356	fx22	5	2:13:16 PM	38	0.25 kUA/I	0	Negative	
011990000365	fx5	5	2:33:15 PM	20	0.03 kUA/I	0	Negative	
011990000365	fx1	5	2:34:14 PM	27	0.12 kUA/I	0	Negative	
011990000365	fx2	5	2:35:14 PM	15	Low RU	0	Negative	
011990000365	fx22	5	2:36:14 PM	28	0.14 kUA/I	0	Negative	
011990000365	gx2	5	2:37:14 PM	21	0.05 kUA/I	0	Negative	
011990000365	wx1	5	2:38:14 PM	36	0.22 kUA/I	0	Negative	
011990000365	t20	5	2:39:14 PM	17	Low RU	0		
011990000366	fx22	5	2:14:15 PM	29	0.15 kUA/I	0	Negative	
011990000375	fx5	5	2:41:14 PM	26	0.11 kUA/I	0	Negative	
011990000375	fx1	5	2:42:14 PM	20	0.04 kUA/I	0	Negative	
011990000375	fx2	5	2:43:14 PM	170	1.79 kUA/I	2	Positive	
011990000375	fx22	5	2:44:14 PM	30	0.15 kUA/I	0	Negative	
011990000375	gx2	5	2:45:14 PM	50	0.39 kUA/I	1	Positive	
011990000375	wx1	5	2:46:14 PM	94	0.91 kUA/I	2	Positive	
011990000375	t20	5	2:47:14 PM	48	0.37 kUA/I	1		
011990000376	fx22	5	2:15:15 PM	43	0.31 kUA/I	0	Negative	
011990000377	fx22	5	2:12:15 PM	31	0.17 kUA/I	0	Negative	
011990000385	fx5	5	2:49:14 PM	22	0.06 kUA/I	0	Negative	
011990000385	fx1	5	2:50:14 PM	17	Low RU	0	Negative	
011990000385	fx2	5	2:51:14 PM	23	0.08 kUA/I	0	Negative	
011990000385	fx22	5	2:52:14 PM	32	0.19 kUA/I	0	Negative	
011990000385	gx2	5	2:53:14 PM	22	0.06 kUA/I	0	Negative	
011990000385	wx1	5	2:54:14 PM	28	0.13 kUA/I	0	Negative	
011990000385	t20	5	2:55:14 PM	16	Low RU	0		
011990000386	fx22	5	2:16:15 PM	27	0.12 kUA/I	0	Negative	
011990000396	fx5	5	2:17:15 PM	21	0.05 kUA/I	0	Negative	
011990000396	fx1	5	2:18:15 PM	23	0.07 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000396	fx2	5	2:19:15 PM	29	0.15 kUA/l	0	Negative	
011990000396	fx22	5	2:20:15 PM	29	0.15 kUA/l	0	Negative	
011990000396	gx2	5	2:21:15 PM	20	0.04 kUA/l	0	Negative	
011990000396	wx1	5	2:22:15 PM	155	1.62 kUA/l	2	Positive	
011990000396	t20	5	2:23:15 PM	64	0.55 kUA/l	1		
011990000508	fx5	5	2:57:13 PM	22	0.07 kUA/l	0	Negative	
011990000508	fx1	5	2:58:13 PM	16	Low RU	0	Negative	
011990000508	fx2	5	2:59:13 PM	14	Low RU	0	Negative	
011990000508	fx22	5	3:00:14 PM	28	0.14 kUA/l	0	Negative	
011990000508	gx2	5	3:01:14 PM	22	0.07 kUA/l	0	Negative	
011990000508	wx1	5	3:02:13 PM	28	0.13 kUA/l	0	Negative	
011990000508	t20	5	3:03:13 PM	17	Low RU	0		
011990000509	fx5	5	3:32:12 PM	31	0.17 kUA/l	0	Negative	
011990000509	fx1	5	3:33:12 PM	19	0.03 kUA/l	0	Negative	
011990000509	fx2	5	3:34:12 PM	27	0.12 kUA/l	0	Negative	
011990000509	fx22	5	3:35:12 PM	29	0.14 kUA/l	0	Negative	
011990000509	gx2	5	3:36:12 PM	28	0.14 kUA/l	0	Negative	
011990000509	wx1	5	3:37:12 PM	48	0.37 kUA/l	1	Positive	
011990000509	t20	5	3:38:12 PM	23	0.08 kUA/l	0		
011990000519	fx5	5	3:24:12 PM	19	0.02 kUA/l	0	Negative	
011990000519	fx1	5	3:25:12 PM	17	Low RU	0	Negative	
011990000519	fx2	5	3:26:12 PM	31	0.17 kUA/l	0	Negative	
011990000519	fx22	5	3:27:12 PM	29	0.14 kUA/l	0	Negative	
011990000519	gx2	5	3:28:12 PM	21	0.05 kUA/l	0	Negative	
011990000519	wx1	5	3:29:12 PM	26	0.11 kUA/l	0	Negative	
011990000519	t20	5	3:30:12 PM	16	Low RU	0		
011990000520	fx5	5	4:04:11 PM	28	0.13 kUA/l	0	Negative	
011990000520	fx1	5	4:05:10 PM	21	0.05 kUA/l	0	Negative	
011990000520	fx2	5	4:06:11 PM	24	0.09 kUA/l	0	Negative	
011990000520	fx22	5	4:07:11 PM	32	0.18 kUA/l	0	Negative	
011990000520	gx2	5	4:08:11 PM	24	0.08 kUA/l	0	Negative	
011990000520	wx1	5	4:09:10 PM	31	0.17 kUA/l	0	Negative	
011990000520	t20	5	4:10:10 PM	18	0.01 kUA/l	0		
011990000529	fx5	5	3:40:12 PM	19	0.03 kUA/l	0	Negative	
011990000529	fx1	5	3:41:12 PM	21	0.05 kUA/l	0	Negative	
011990000529	fx2	5	3:42:12 PM	13	Low RU	0	Negative	
011990000529	fx22	5	3:43:12 PM	30	0.16 kUA/l	0	Negative	
011990000529	gx2	5	3:44:12 PM	23	0.07 kUA/l	0	Negative	
011990000529	wx1	5	3:45:11 PM	28	0.13 kUA/l	0	Negative	
011990000529	t20	5	3:46:11 PM	17	Low RU	0		
011990000530	fx5	5	4:12:10 PM	22	0.06 kUA/l	0	Negative	
011990000530	fx1	5	4:13:10 PM	16	Low RU	0	Negative	
011990000530	fx2	5	4:14:10 PM	15	Low RU	0	Negative	
011990000530	fx22	5	4:15:10 PM	33	0.20 kUA/l	0	Negative	
011990000530	gx2	5	4:16:10 PM	22	0.06 kUA/l	0	Negative	
011990000530	wx1	5	4:17:10 PM	143	1.47 kUA/l	2	Positive	
011990000530	t20	5	4:18:10 PM	21	0.04 kUA/l	0		
011990000539	fx5	5	3:15:13 PM	21	0.05 kUA/l	0	Negative	
011990000539	fx1	5	3:16:13 PM	16	Low RU	0	Negative	
011990000539	fx2	5	3:17:13 PM	25	0.09 kUA/l	0	Negative	
011990000539	fx22	5	3:18:12 PM	28	0.14 kUA/l	0	Negative	
011990000539	gx2	5	3:19:12 PM	22	0.06 kUA/l	0	Negative	
011990000539	wx1	5	3:20:13 PM	30	0.16 kUA/l	0	Negative	
011990000539	t20	5	3:22:12 PM	17	Low RU	0		

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000540	fx5	5	3:48:12 PM	20	0.03 kUA/l	0	Negative	
011990000540	fx1	5	3:49:11 PM	17	Low RU	0	Negative	
011990000540	fx2	5	3:50:12 PM	39	0.26 kUA/l	0	Negative	
011990000540	fx22	5	3:51:11 PM	25	0.10 kUA/l	0	Negative	
011990000540	gx2	5	3:52:11 PM	20	0.04 kUA/l	0	Negative	
011990000540	wx1	5	3:53:11 PM	29	0.14 kUA/l	0	Negative	
011990000540	t20	5	3:54:11 PM	18	0.01 kUA/l	0		
011990000549	fx5	5	3:05:13 PM	20	0.03 kUA/l	0	Negative	
011990000549	fx1	5	3:06:13 PM	17	Low RU	0	Negative	
011990000549	fx2	5	3:09:13 PM	15	Low RU	0	Negative	
011990000549	fx22	5	3:10:13 PM	27	0.13 kUA/l	0	Negative	
011990000549	gx2	5	3:11:13 PM	24	0.08 kUA/l	0	Negative	
011990000549	wx1	5	3:12:13 PM	27	0.13 kUA/l	0	Negative	
011990000549	t20	5	3:13:13 PM	17	Low RU	0		
011990000550	fx5	5	3:56:11 PM	20	0.04 kUA/l	0	Negative	
011990000550	fx1	5	3:57:11 PM	15	Low RU	0	Negative	
011990000550	fx2	5	3:58:11 PM	13	Low RU	0	Negative	
011990000550	fx22	5	3:59:11 PM	26	0.11 kUA/l	0	Negative	
011990000550	gx2	5	4:00:11 PM	23	0.07 kUA/l	0	Negative	
011990000550	wx1	5	4:01:11 PM	30	0.15 kUA/l	0	Negative	
011990000550	t20	5	4:02:11 PM	18	0.01 kUA/l	0		
012000500914A	f3	1	3:08:13 PM	40	0.05 kUA/l			

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	3:21:13 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012000500914A	f3	3:08:13 PM	404CG 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000377	fx22	2:12:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000356	fx22	2:13:16 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000366	fx22	2:14:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000376	fx22	2:15:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000386	fx22	2:16:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	fx5	2:17:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	fx1	2:18:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	fx2	2:19:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	fx22	2:20:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	gx2	2:21:15 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	wx1	2:22:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000396	t20	2:23:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000355	phad	2:24:15 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000355	fx5	2:25:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000355	fx1	2:26:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000355	fx2	2:27:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000355	fx22	2:28:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000355	gx2	2:29:15 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000355	wx1	2:30:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000355	t20	2:31:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	phad	2:32:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	fx5	2:33:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	fx1	2:34:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	fx2	2:35:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	fx22	2:36:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	gx2	2:37:14 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	wx1	2:38:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000365	t20	2:39:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	phad	2:40:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	fx5	2:41:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	fx1	2:42:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	fx2	2:43:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	fx22	2:44:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	gx2	2:45:14 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	wx1	2:46:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000375	t20	2:47:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	phad	2:48:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	fx5	2:49:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	fx1	2:50:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	fx2	2:51:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	fx22	2:52:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	gx2	2:53:14 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	wx1	2:54:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000385	t20	2:55:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	phad	2:56:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	fx5	2:57:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	fx1	2:58:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	fx2	2:59:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	fx22	3:00:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	gx2	3:01:14 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	wx1	3:02:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000508	t20	3:03:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	phad	3:04:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	fx5	3:05:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	fx1	3:06:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	fx2	3:09:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	fx22	3:10:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	gx2	3:11:13 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	wx1	3:12:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000549	t20	3:13:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	phad	3:14:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	fx5	3:15:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	fx1	3:16:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	fx2	3:17:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	fx22	3:18:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	gx2	3:19:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	wx1	3:20:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000539	t20	3:22:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	phad	3:23:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	fx5	3:24:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	fx1	3:25:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	fx2	3:26:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	fx22	3:27:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	gx2	3:28:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	wx1	3:29:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000519	t20	3:30:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000509	phad	3:31:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000509	fx5	3:32:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000509	fx1	3:33:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000509	fx2	3:34:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000509	fx22	3:35:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000509	gx2	3:36:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000509	wx1	3:37:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000509	t20	3:38:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	phad	3:39:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	fx5	3:40:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	fx1	3:41:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	fx2	3:42:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	fx22	3:43:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	gx2	3:44:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	wx1	3:45:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000529	t20	3:46:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	phad	3:47:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	fx5	3:48:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	fx1	3:49:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	fx2	3:50:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	fx22	3:51:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	gx2	3:52:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	wx1	3:53:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000540	t20	3:54:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	phad	3:55:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	fx5	3:56:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	fx1	3:57:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	fx2	3:58:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	fx22	3:59:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	gx2	4:00:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	wx1	4:01:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000550	t20	4:02:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	phad	4:03:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	fx5	4:04:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	fx1	4:05:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	fx2	4:06:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	fx22	4:07:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	gx2	4:08:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	wx1	4:09:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000520	t20	4:10:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	phad	4:11:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	fx5	4:12:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	fx1	4:13:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	fx2	4:14:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	fx22	4:15:10 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	gx2	4:16:10 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	wx1	4:17:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000530	t20	4:18:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	