


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200123-385  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 23-Jan-20 10:05:28 AM  
Ended: 23-Jan-20 2:06:20 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	10:05:28 AM	314	0.69	OK	
CC-2	17.5	10:06:28 AM	6570	18.4	OK	

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	10:07:28 AM	3350	8.36 kUA/l	CZ1AR	OK	
slgE L	e1	1	10:08:28 AM	1292	3.01 kUA/l	CZ4AN	OK	
slgE Neg	d1	1	10:09:28 AM	21	0.01 kUA/l	BL1CP	OK	
IQC271119	d1	1	10:10:28 AM	6569	18.4 kUA/l	<lab>	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000464	phad	5	1:54:19 PM	33	0.20 PAU/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000465	phad	5	1:01:21 PM	79	0.73 PAU/I	2	Positive	
011990000466	phad	5	12:21:23 PM	42	0.30 PAU/I	0	Negative	
011990000467	phad	5	11:37:24 AM	36	0.23 PAU/I	0	Negative	
011990000468	phad	5	10:51:26 AM	33	0.19 PAU/I	0	Negative	
011990000476	phad	5	1:46:20 PM	83	0.77 PAU/I	2	Positive	
011990000477	phad	5	12:53:21 PM	32	0.18 PAU/I	0	Negative	
011990000478	phad	5	12:13:23 PM	39	0.26 PAU/I	0	Negative	
011990000479	phad	5	11:23:25 AM	237	2.56 PAU/I	2	Positive	
011990000480	phad	5	10:43:26 AM	52	0.42 PAU/I	1	Positive	
011990000488	phad	5	1:38:20 PM	41	0.28 PAU/I	0	Negative	
011990000489	phad	5	12:45:22 PM	402	4.47 PAU/I	3	Positive	
011990000490	phad	5	12:05:23 PM	95	0.91 PAU/I	2	Positive	
011990000491	phad	5	11:15:25 AM	35	0.22 PAU/I	0	Negative	
011990000492	phad	5	10:27:27 AM	32	0.18 PAU/I	0	Negative	
011990000901	phad	5	11:07:25 AM	97	0.93 PAU/I	2	Positive	
011990000904	phad	5	10:19:27 AM	272	2.97 PAU/I	2	Positive	
011990000906	phad	5	11:57:23 AM	737	8.40 PAU/I	3	Positive	
011990000907	phad	5	1:30:20 PM	34	0.20 PAU/I	0	Negative	
011990000908	phad	5	12:37:23 PM	35	0.21 PAU/I	0	Negative	
011990000920	phad	5	10:35:27 AM	1883	22.4 PAU/I	4	Positive	
011990000926	phad	5	12:29:23 PM	35	0.21 PAU/I	0	Negative	
011990000928	phad	5	1:12:21 PM	330	3.64 PAU/I	3	Positive	
011990000931	phad	5	10:59:26 AM	58	0.48 PAU/I	1	Positive	
011990000932	phad	5	11:49:24 AM	42	0.30 PAU/I	0	Negative	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000464	fx5	5	1:55:19 PM	23	0.07 kUA/I	0	Negative	
011990000464	fx1	5	1:56:19 PM	18	0.02 kUA/I	0	Negative	
011990000464	fx2	5	1:57:19 PM	14	Low RU	0	Negative	
011990000464	fx22	5	1:58:19 PM	30	0.15 kUA/I	0	Negative	
011990000464	gx2	5	2:03:19 PM	19	0.02 kUA/I	0	Negative	
011990000464	wx1	5	2:04:19 PM	36	0.23 kUA/I	0	Negative	
011990000464	t20	5	2:05:19 PM	16	Low RU	0		
011990000465	fx5	5	1:02:21 PM	22	0.06 kUA/I	0	Negative	
011990000465	fx1	5	1:03:21 PM	19	0.03 kUA/I	0	Negative	
011990000465	fx2	5	1:04:21 PM	19	0.03 kUA/I	0	Negative	
011990000465	fx22	5	1:05:21 PM	35	0.21 kUA/I	0	Negative	
011990000465	gx2	5	1:06:21 PM	27	0.12 kUA/I	0	Negative	
011990000465	wx1	5	1:07:21 PM	279	3.04 kUA/I	2	Positive	
011990000465	t20	5	1:08:21 PM	22	0.07 kUA/I	0		
011990000466	fx5	5	12:22:23 PM	23	0.07 kUA/I	0	Negative	
011990000466	fx1	5	12:23:23 PM	25	0.10 kUA/I	0	Negative	
011990000466	fx2	5	12:24:23 PM	21	0.05 kUA/I	0	Negative	
011990000466	fx22	5	12:25:22 PM	31	0.16 kUA/I	0	Negative	
011990000466	gx2	5	12:26:23 PM	27	0.12 kUA/I	0	Negative	
011990000466	wx1	5	12:27:23 PM	33	0.20 kUA/I	0	Negative	
011990000466	t20	5	12:28:23 PM	20	0.03 kUA/I	0		
011990000467	fx5	5	11:42:25 AM	19	0.03 kUA/I	0	Negative	
011990000467	fx1	5	11:43:24 AM	18	0.02 kUA/I	0	Negative	
011990000467	fx2	5	11:44:24 AM	13	Low RU	0	Negative	
011990000467	fx22	5	11:45:24 AM	29	0.15 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000467	gx2	5	11:46:24 AM	30	0.16 kUA/l	0	Negative	
011990000467	wx1	5	11:47:24 AM	32	0.18 kUA/l	0	Negative	
011990000467	t20	5	11:48:24 AM	17	Low RU	0		
011990000468	fx5	5	10:52:26 AM	19	0.03 kUA/l	0	Negative	
011990000468	fx1	5	10:53:26 AM	19	0.03 kUA/l	0	Negative	
011990000468	fx2	5	10:54:26 AM	14	Low RU	0	Negative	
011990000468	fx22	5	10:55:26 AM	27	0.12 kUA/l	0	Negative	
011990000468	gx2	5	10:56:26 AM	19	0.03 kUA/l	0	Negative	
011990000468	wx1	5	10:57:26 AM	27	0.12 kUA/l	0	Negative	
011990000468	t20	5	10:58:26 AM	17	Low RU	0		
011990000476	fx5	5	1:47:20 PM	705	8.03 kUA/l	3	Positive	
011990000476	fx1	5	1:48:20 PM	45	0.33 kUA/l	0	Negative	
011990000476	fx2	5	1:49:19 PM	24	0.09 kUA/l	0	Negative	
011990000476	fx22	5	1:50:19 PM	59	0.50 kUA/l	1	Positive	
011990000476	gx2	5	1:51:19 PM	48	0.37 kUA/l	1	Positive	
011990000476	wx1	5	1:52:20 PM	60	0.51 kUA/l	1	Positive	
011990000476	t20	5	1:53:20 PM	42	0.30 kUA/l	0		
011990000477	fx5	5	12:54:21 PM	21	0.05 kUA/l	0	Negative	
011990000477	fx1	5	12:55:22 PM	21	0.05 kUA/l	0	Negative	
011990000477	fx2	5	12:56:22 PM	14	Low RU	0	Negative	
011990000477	fx22	5	12:57:22 PM	30	0.16 kUA/l	0	Negative	
011990000477	gx2	5	12:58:21 PM	36	0.23 kUA/l	0	Negative	
011990000477	wx1	5	12:59:21 PM	32	0.18 kUA/l	0	Negative	
011990000477	t20	5	1:00:21 PM	17	0.01 kUA/l	0		
011990000478	fx5	5	12:14:23 PM	26	0.11 kUA/l	0	Negative	
011990000478	fx1	5	12:15:23 PM	22	0.06 kUA/l	0	Negative	
011990000478	fx2	5	12:16:23 PM	18	0.02 kUA/l	0	Negative	
011990000478	fx22	5	12:17:23 PM	30	0.15 kUA/l	0	Negative	
011990000478	gx2	5	12:18:23 PM	24	0.08 kUA/l	0	Negative	
011990000478	wx1	5	12:19:23 PM	29	0.15 kUA/l	0	Negative	
011990000478	t20	5	12:20:23 PM	19	0.03 kUA/l	0		
011990000479	fx5	5	11:24:25 AM	20	0.04 kUA/l	0	Negative	
011990000479	fx1	5	11:25:25 AM	17	Low RU	0	Negative	
011990000479	fx2	5	11:26:25 AM	17	Low RU	0	Negative	
011990000479	fx22	5	11:27:25 AM	31	0.17 kUA/l	0	Negative	
011990000479	gx2	5	11:28:25 AM	26	0.11 kUA/l	0	Negative	
011990000479	wx1	5	11:29:25 AM	179	1.89 kUA/l	2	Positive	
011990000479	t20	5	11:30:25 AM	33	0.19 kUA/l	0		
011990000480	fx5	5	10:44:27 AM	20	0.04 kUA/l	0	Negative	
011990000480	fx1	5	10:45:27 AM	17	Low RU	0	Negative	
011990000480	fx2	5	10:46:26 AM	25	0.10 kUA/l	0	Negative	
011990000480	fx22	5	10:47:26 AM	33	0.19 kUA/l	0	Negative	
011990000480	gx2	5	10:48:26 AM	25	0.09 kUA/l	0	Negative	
011990000480	wx1	5	10:49:26 AM	31	0.17 kUA/l	0	Negative	
011990000480	t20	5	10:50:26 AM	20	0.03 kUA/l	0		
011990000488	fx5	5	1:39:20 PM	24	0.08 kUA/l	0	Negative	
011990000488	fx1	5	1:40:20 PM	20	0.04 kUA/l	0	Negative	
011990000488	fx2	5	1:41:20 PM	19	0.03 kUA/l	0	Negative	
011990000488	fx22	5	1:42:20 PM	31	0.17 kUA/l	0	Negative	
011990000488	gx2	5	1:43:20 PM	24	0.09 kUA/l	0	Negative	
011990000488	wx1	5	1:44:20 PM	34	0.20 kUA/l	0	Negative	
011990000488	t20	5	1:45:20 PM	18	0.01 kUA/l	0		
011990000489	fx5	5	12:46:22 PM	32	0.17 kUA/l	0	Negative	
011990000489	fx1	5	12:47:22 PM	31	0.17 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000489	fx2	5	12:48:22 PM	23	0.08 kUA/l	0	Negative	
011990000489	fx22	5	12:49:22 PM	38	0.25 kUA/l	0	Negative	
011990000489	gx2	5	12:50:22 PM	1071	12.4 kUA/l	3	Positive	
011990000489	wx1	5	12:51:22 PM	57	0.48 kUA/l	1	Positive	
011990000489	t20	5	12:52:22 PM	26	0.11 kUA/l	0		
011990000490	fx5	5	12:06:23 PM	23	0.07 kUA/l	0	Negative	
011990000490	fx1	5	12:07:23 PM	19	0.02 kUA/l	0	Negative	
011990000490	fx2	5	12:08:24 PM	16	Low RU	0	Negative	
011990000490	fx22	5	12:09:23 PM	31	0.17 kUA/l	0	Negative	
011990000490	gx2	5	12:10:23 PM	25	0.09 kUA/l	0	Negative	
011990000490	wx1	5	12:11:23 PM	34	0.20 kUA/l	0	Negative	
011990000490	t20	5	12:12:23 PM	20	0.04 kUA/l	0		
011990000491	fx5	5	11:16:25 AM	24	0.08 kUA/l	0	Negative	
011990000491	fx1	5	11:17:25 AM	17	0.01 kUA/l	0	Negative	
011990000491	fx2	5	11:18:25 AM	17	0.01 kUA/l	0	Negative	
011990000491	fx22	5	11:19:25 AM	29	0.15 kUA/l	0	Negative	
011990000491	gx2	5	11:20:25 AM	22	0.06 kUA/l	0	Negative	
011990000491	wx1	5	11:21:25 AM	31	0.17 kUA/l	0	Negative	
011990000491	t20	5	11:22:25 AM	18	0.01 kUA/l	0		
011990000492	fx5	5	10:28:27 AM	21	0.04 kUA/l	0	Negative	
011990000492	fx1	5	10:29:27 AM	19	0.02 kUA/l	0	Negative	
011990000492	fx2	5	10:30:27 AM	14	Low RU	0	Negative	
011990000492	fx22	5	10:31:27 AM	28	0.13 kUA/l	0	Negative	
011990000492	gx2	5	10:32:27 AM	25	0.09 kUA/l	0	Negative	
011990000492	wx1	5	10:33:27 AM	30	0.15 kUA/l	0	Negative	
011990000492	t20	5	10:34:27 AM	17	0.01 kUA/l	0		
011990000901	fx5	5	11:08:25 AM	23	0.07 kUA/l	0	Negative	
011990000901	fx1	5	11:09:25 AM	17	0.01 kUA/l	0	Negative	
011990000901	fx2	5	11:10:25 AM	39	0.26 kUA/l	0	Negative	
011990000901	fx22	5	11:11:25 AM	30	0.16 kUA/l	0	Negative	
011990000901	gx2	5	11:12:26 AM	26	0.11 kUA/l	0	Negative	
011990000901	wx1	5	11:13:26 AM	34	0.20 kUA/l	0	Negative	
011990000901	t20	5	11:14:25 AM	18	0.01 kUA/l	0		
011990000904	fx5	5	10:20:27 AM	54	0.44 kUA/l	1	Positive	
011990000904	fx1	5	10:21:27 AM	24	0.08 kUA/l	0	Negative	
011990000904	fx2	5	10:22:27 AM	22	0.06 kUA/l	0	Negative	
011990000904	fx22	5	10:23:27 AM	30	0.16 kUA/l	0	Negative	
011990000904	gx2	5	10:24:27 AM	1584	18.6 kUA/l	4	Positive	
011990000904	wx1	5	10:25:27 AM	34	0.21 kUA/l	0	Negative	
011990000904	t20	5	10:26:27 AM	19	0.03 kUA/l	0		
011990000906	fx5	5	11:58:23 AM	26	0.11 kUA/l	0	Negative	
011990000906	fx1	5	11:59:23 AM	18	0.01 kUA/l	0	Negative	
011990000906	fx2	5	12:00:24 PM	20	0.04 kUA/l	0	Negative	
011990000906	fx22	5	12:01:24 PM	42	0.29 kUA/l	0	Negative	
011990000906	gx2	5	12:02:23 PM	28	0.13 kUA/l	0	Negative	
011990000906	wx1	5	12:03:23 PM	37	0.23 kUA/l	0	Negative	
011990000906	t20	5	12:04:23 PM	21	0.05 kUA/l	0		
011990000907	fx5	5	1:31:20 PM	21	0.05 kUA/l	0	Negative	
011990000907	fx1	5	1:32:20 PM	21	0.05 kUA/l	0	Negative	
011990000907	fx2	5	1:33:20 PM	14	Low RU	0	Negative	
011990000907	fx22	5	1:34:20 PM	30	0.16 kUA/l	0	Negative	
011990000907	gx2	5	1:35:20 PM	22	0.07 kUA/l	0	Negative	
011990000907	wx1	5	1:36:20 PM	32	0.19 kUA/l	0	Negative	
011990000907	t20	5	1:37:20 PM	19	0.03 kUA/l	0		

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000908	fx5	5	12:38:22 PM	21	0.06 kUA/l	0	Negative	
011990000908	fx1	5	12:39:22 PM	19	0.03 kUA/l	0	Negative	
011990000908	fx2	5	12:40:22 PM	15	Low RU	0	Negative	
011990000908	fx22	5	12:41:22 PM	30	0.16 kUA/l	0	Negative	
011990000908	gx2	5	12:42:22 PM	31	0.17 kUA/l	0	Negative	
011990000908	wx1	5	12:43:22 PM	34	0.20 kUA/l	0	Negative	
011990000908	t20	5	12:44:22 PM	18	0.01 kUA/l	0		
011990000920	fx5	5	10:36:27 AM	70	0.62 kUA/l	1	Positive	
011990000920	fx1	5	10:37:27 AM	64	0.55 kUA/l	1	Positive	
011990000920	fx2	5	10:38:26 AM	16	Low RU	0	Negative	
011990000920	fx22	5	10:39:27 AM	70	0.62 kUA/l	1	Positive	
011990000920	gx2	5	10:40:27 AM	2291	27.6 kUA/l	4	Positive	
011990000920	wx1	5	10:41:26 AM	189	2.01 kUA/l	2	Positive	
011990000920	t20	5	10:42:26 AM	236	2.54 kUA/l	2		
011990000926	fx5	5	12:30:23 PM	42	0.30 kUA/l	0	Negative	
011990000926	fx1	5	12:31:22 PM	19	0.03 kUA/l	0	Negative	
011990000926	fx2	5	12:32:22 PM	14	Low RU	0	Negative	
011990000926	fx22	5	12:33:22 PM	28	0.13 kUA/l	0	Negative	
011990000926	gx2	5	12:34:22 PM	22	0.06 kUA/l	0	Negative	
011990000926	wx1	5	12:35:22 PM	32	0.18 kUA/l	0	Negative	
011990000926	t20	5	12:36:23 PM	19	0.03 kUA/l	0		
011990000928	fx5	5	1:13:21 PM	54	0.44 kUA/l	1	Positive	
011990000928	fx1	5	1:24:21 PM	21	0.05 kUA/l	0	Negative	
011990000928	fx2	5	1:25:21 PM	367	4.07 kUA/l	3	Positive	
011990000928	fx22	5	1:26:20 PM	40	0.27 kUA/l	0	Negative	
011990000928	gx2	5	1:27:20 PM	37	0.24 kUA/l	0	Negative	
011990000928	wx1	5	1:28:20 PM	101	0.99 kUA/l	2	Positive	
011990000928	t20	5	1:29:20 PM	39	0.26 kUA/l	0		
011990000931	fx5	5	11:00:26 AM	21	0.04 kUA/l	0	Negative	
011990000931	fx1	5	11:01:26 AM	16	Low RU	0	Negative	
011990000931	fx2	5	11:02:26 AM	20	0.03 kUA/l	0	Negative	
011990000931	fx22	5	11:03:26 AM	29	0.14 kUA/l	0	Negative	
011990000931	gx2	5	11:04:26 AM	20	0.04 kUA/l	0	Negative	
011990000931	wx1	5	11:05:26 AM	29	0.15 kUA/l	0	Negative	
011990000931	t20	5	11:06:26 AM	17	Low RU	0		
011990000932	fx5	5	11:50:24 AM	48	0.37 kUA/l	1	Positive	
011990000932	fx1	5	11:51:24 AM	38	0.25 kUA/l	0	Negative	
011990000932	fx2	5	11:52:24 AM	15	Low RU	0	Negative	
011990000932	fx22	5	11:53:24 AM	32	0.18 kUA/l	0	Negative	
011990000932	gx2	5	11:54:24 AM	22	0.06 kUA/l	0	Negative	
011990000932	wx1	5	11:55:24 AM	32	0.18 kUA/l	0	Negative	
011990000932	t20	5	11:56:24 AM	20	0.03 kUA/l	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	



## Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	10:05:28 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	10:06:28 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	10:07:28 AM	352D0 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ1AR 21/08		Exp
slgE L	e1	10:08:28 AM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ4AN 19/06!		Exp
slgE Neg	d1	10:09:28 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		BL1CP 19/12!		Exp
IQC271119	d1	10:10:28 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000904	phad	10:19:27 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000904	fx5	10:20:27 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000904	fx1	10:21:27 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000904	fx2	10:22:27 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000904	fx22	10:23:27 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000904	gx2	10:24:27 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000904	wx1	10:25:27 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000904	t20	10:26:27 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	phad	10:27:27 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	fx5	10:28:27 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	fx1	10:29:27 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	fx2	10:30:27 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	fx22	10:31:27 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	gx2	10:32:27 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	wx1	10:33:27 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000492	t20	10:34:27 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	phad	10:35:27 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	fx5	10:36:27 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	fx1	10:37:27 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	fx2	10:38:26 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	fx22	10:39:27 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	gx2	10:40:27 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	wx1	10:41:26 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000920	t20	10:42:26 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	phad	10:43:26 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	fx5	10:44:27 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	fx1	10:45:27 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	fx2	10:46:26 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	fx22	10:47:26 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	gx2	10:48:26 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	wx1	10:49:26 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000480	t20	10:50:26 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	phad	10:51:26 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	fx5	10:52:26 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	fx1	10:53:26 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	fx2	10:54:26 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	fx22	10:55:26 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	gx2	10:56:26 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	wx1	10:57:26 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000468	t20	10:58:26 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	phad	10:59:26 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	fx5	11:00:26 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	fx1	11:01:26 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	fx2	11:02:26 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	fx22	11:03:26 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	gx2	11:04:26 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	wx1	11:05:26 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000931	t20	11:06:26 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000901	phad	11:07:25 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000901	fx5	11:08:25 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000901	fx1	11:09:25 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000901	fx2	11:10:25 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000901	fx22	11:11:25 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000901	gx2	11:12:26 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000901	wx1	11:13:26 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000901	t20	11:14:25 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	phad	11:15:25 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	fx5	11:16:25 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	fx1	11:17:25 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	fx2	11:18:25 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	fx22	11:19:25 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	gx2	11:20:25 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	wx1	11:21:25 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000491	t20	11:22:25 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	phad	11:23:25 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	fx5	11:24:25 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	fx1	11:25:25 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	fx2	11:26:25 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	fx22	11:27:25 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	gx2	11:28:25 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	wx1	11:29:25 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000479	t20	11:30:25 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	phad	11:37:24 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	fx5	11:42:25 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	fx1	11:43:24 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	fx2	11:44:24 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	fx22	11:45:24 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	gx2	11:46:24 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	wx1	11:47:24 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000467	t20	11:48:24 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	phad	11:49:24 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	fx5	11:50:24 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	fx1	11:51:24 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	fx2	11:52:24 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	fx22	11:53:24 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	gx2	11:54:24 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	wx1	11:55:24 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000932	t20	11:56:24 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	phad	11:57:23 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	fx5	11:58:23 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	fx1	11:59:23 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	fx2	12:00:24 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	fx22	12:01:24 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	gx2	12:02:23 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	wx1	12:03:23 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000906	t20	12:04:23 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	phad	12:05:23 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	fx5	12:06:23 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	fx1	12:07:23 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	fx2	12:08:24 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	fx22	12:09:23 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	gx2	12:10:23 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	wx1	12:11:23 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000490	t20	12:12:23 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	phad	12:13:23 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	fx5	12:14:23 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	fx1	12:15:23 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	fx2	12:16:23 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	fx22	12:17:23 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	gx2	12:18:23 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	wx1	12:19:23 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000478	t20	12:20:23 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000466	phad	12:21:23 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000466	fx5	12:22:23 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000466	fx1	12:23:23 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000466	fx2	12:24:23 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000466	fx22	12:25:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000466	gx2	12:26:23 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000466	wx1	12:27:23 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000466	t20	12:28:23 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	phad	12:29:23 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	fx5	12:30:23 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	fx1	12:31:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	fx2	12:32:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	fx22	12:33:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	gx2	12:34:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	wx1	12:35:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000926	t20	12:36:23 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	phad	12:37:23 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	fx5	12:38:22 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	fx1	12:39:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	fx2	12:40:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	fx22	12:41:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	gx2	12:42:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	wx1	12:43:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000908	t20	12:44:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	phad	12:45:22 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	fx5	12:46:22 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	fx1	12:47:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	fx2	12:48:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	fx22	12:49:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	gx2	12:50:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	wx1	12:51:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000489	t20	12:52:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	phad	12:53:21 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	fx5	12:54:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	fx1	12:55:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	fx2	12:56:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	fx22	12:57:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	gx2	12:58:21 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	wx1	12:59:21 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000477	t20	1:00:21 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	phad	1:01:21 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	fx5	1:02:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	fx1	1:03:21 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	fx2	1:04:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	fx22	1:05:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	gx2	1:06:21 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	wx1	1:07:21 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000465	t20	1:08:21 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	phad	1:12:21 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	fx5	1:13:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	fx1	1:24:21 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	fx2	1:25:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	fx22	1:26:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	gx2	1:27:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	wx1	1:28:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000928	t20	1:29:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	phad	1:30:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	fx5	1:31:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	fx1	1:32:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	fx2	1:33:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	fx22	1:34:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	gx2	1:35:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	wx1	1:36:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000907	t20	1:37:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000488	phad	1:38:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000488	fx5	1:39:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000488	fx1	1:40:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000488	fx2	1:41:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000488	fx22	1:42:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000488	gx2	1:43:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000488	wx1	1:44:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000488	t20	1:45:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	phad	1:46:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	fx5	1:47:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	fx1	1:48:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	fx2	1:49:19 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	fx22	1:50:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	gx2	1:51:19 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	wx1	1:52:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000476	t20	1:53:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000464	phad	1:54:19 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000464	fx5	1:55:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000464	fx1	1:56:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000464	fx2	1:57:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000464	fx22	1:58:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000464	gx2	2:03:19 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000464	wx1	2:04:19 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000464	t20	2:05:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	

### Errors

#### Expired lot numbers used for Quality Control

352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200123-385  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 23-Jan-20 2:06:19 PM  
Ended: 23-Jan-20 4:12:35 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

No curve controls in this run

### Quality Controls

No quality controls in this run

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000461	phad	5	3:50:15 PM	30	0.16 PAU/l	0	Negative	
011990000462	phad	5	3:18:16 PM	55	0.45 PAU/l	1	Positive	
011990000463	phad	5	2:38:18 PM	32	0.18 PAU/l	0	Negative	
011990000473	phad	5	3:58:14 PM	36	0.23 PAU/l	0	Negative	
011990000474	phad	5	3:02:17 PM	34	0.20 PAU/l	0	Negative	
011990000475	phad	5	2:30:18 PM	45	0.33 PAU/l	0	Negative	
011990000485	phad	5	3:34:16 PM	46	0.35 PAU/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000486	phad	5	3:10:17 PM	32	0.18 PAU/I	0	Negative	
011990000487	phad	5	2:22:19 PM	36	0.23 PAU/I	0	Negative	
011990000911	phad	5	3:42:15 PM	44	0.33 PAU/I	0	Negative	
011990000912	phad	5	2:54:17 PM	49	0.38 PAU/I	1	Positive	
011990000913	phad	5	2:14:19 PM	35	0.21 PAU/I	0	Negative	
011990000927	phad	5	2:06:19 PM	64	0.55 PAU/I	1	Positive	
011990000929	phad	5	3:26:16 PM	40	0.28 PAU/I	0	Negative	
011990000930	phad	5	2:46:17 PM	35	0.21 PAU/I	0	Negative	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000461	fx5	5	3:51:14 PM	19	0.02 kUA/I	0	Negative	
011990000461	fx1	5	3:52:15 PM	19	0.02 kUA/I	0	Negative	
011990000461	fx2	5	3:53:15 PM	12	Low RU	0	Negative	
011990000461	fx22	5	3:54:15 PM	31	0.16 kUA/I	0	Negative	
011990000461	gx2	5	3:55:14 PM	25	0.10 kUA/I	0	Negative	
011990000461	wx1	5	3:56:14 PM	28	0.13 kUA/I	0	Negative	
011990000461	t20	5	3:57:14 PM	18	0.01 kUA/I	0		
011990000462	fx5	5	3:19:16 PM	26	0.11 kUA/I	0	Negative	
011990000462	fx1	5	3:20:16 PM	20	0.04 kUA/I	0	Negative	
011990000462	fx2	5	3:21:16 PM	15	Low RU	0	Negative	
011990000462	fx22	5	3:22:16 PM	28	0.14 kUA/I	0	Negative	
011990000462	gx2	5	3:23:16 PM	22	0.06 kUA/I	0	Negative	
011990000462	wx1	5	3:24:16 PM	39	0.26 kUA/I	0	Negative	
011990000462	t20	5	3:25:16 PM	20	0.04 kUA/I	0		
011990000463	fx5	5	2:39:18 PM	18	0.02 kUA/I	0	Negative	
011990000463	fx1	5	2:40:18 PM	18	0.02 kUA/I	0	Negative	
011990000463	fx2	5	2:41:18 PM	12	Low RU	0	Negative	
011990000463	fx22	5	2:42:18 PM	28	0.14 kUA/I	0	Negative	
011990000463	gx2	5	2:43:18 PM	20	0.04 kUA/I	0	Negative	
011990000463	wx1	5	2:44:18 PM	29	0.15 kUA/I	0	Negative	
011990000463	t20	5	2:45:17 PM	16	Low RU	0		
011990000473	fx5	5	3:59:14 PM	32	0.18 kUA/I	0	Negative	
011990000473	fx1	5	4:00:14 PM	30	0.15 kUA/I	0	Negative	
011990000473	fx2	5	4:01:14 PM	30	0.15 kUA/I	0	Negative	
011990000473	fx22	5	4:02:14 PM	42	0.29 kUA/I	0	Negative	
011990000473	gx2	5	4:03:14 PM	38	0.25 kUA/I	0	Negative	
011990000473	wx1	5	4:04:14 PM	33	0.19 kUA/I	0	Negative	
011990000473	t20	5	4:05:14 PM	21	0.06 kUA/I	0		
011990000474	fx5	5	3:03:17 PM	22	0.06 kUA/I	0	Negative	
011990000474	fx1	5	3:04:17 PM	17	Low RU	0	Negative	
011990000474	fx2	5	3:05:17 PM	13	Low RU	0	Negative	
011990000474	fx22	5	3:06:17 PM	29	0.14 kUA/I	0	Negative	
011990000474	gx2	5	3:07:17 PM	18	0.02 kUA/I	0	Negative	
011990000474	wx1	5	3:08:17 PM	29	0.15 kUA/I	0	Negative	
011990000474	t20	5	3:09:17 PM	17	Low RU	0		
011990000475	fx5	5	2:31:18 PM	136	1.39 kUA/I	2	Positive	
011990000475	fx1	5	2:32:18 PM	26	0.11 kUA/I	0	Negative	
011990000475	fx2	5	2:33:18 PM	23	0.07 kUA/I	0	Negative	
011990000475	fx22	5	2:34:18 PM	31	0.17 kUA/I	0	Negative	
011990000475	gx2	5	2:35:18 PM	28	0.13 kUA/I	0	Negative	
011990000475	wx1	5	2:36:18 PM	40	0.27 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000475	t20	5	2:37:18 PM	21	0.05 kUA/l	0		
011990000485	fx5	5	3:35:16 PM	22	0.06 kUA/l	0	Negative	
011990000485	fx1	5	3:36:15 PM	17	Low RU	0	Negative	
011990000485	fx2	5	3:37:15 PM	50	0.40 kUA/l	1	Positive	
011990000485	fx22	5	3:38:15 PM	33	0.19 kUA/l	0	Negative	
011990000485	gx2	5	3:39:15 PM	24	0.09 kUA/l	0	Negative	
011990000485	wx1	5	3:40:15 PM	40	0.28 kUA/l	0	Negative	
011990000485	t20	5	3:41:15 PM	38	0.25 kUA/l	0		
011990000486	fx5	5	3:11:16 PM	19	0.02 kUA/l	0	Negative	
011990000486	fx1	5	3:12:16 PM	17	Low RU	0	Negative	
011990000486	fx2	5	3:13:16 PM	12	Low RU	0	Negative	
011990000486	fx22	5	3:14:16 PM	31	0.16 kUA/l	0	Negative	
011990000486	gx2	5	3:15:16 PM	23	0.08 kUA/l	0	Negative	
011990000486	wx1	5	3:16:16 PM	30	0.16 kUA/l	0	Negative	
011990000486	t20	5	3:17:16 PM	16	Low RU	0		
011990000487	fx5	5	2:23:18 PM	24	0.09 kUA/l	0	Negative	
011990000487	fx1	5	2:24:18 PM	23	0.07 kUA/l	0	Negative	
011990000487	fx2	5	2:25:18 PM	22	0.07 kUA/l	0	Negative	
011990000487	fx22	5	2:26:18 PM	32	0.18 kUA/l	0	Negative	
011990000487	gx2	5	2:27:18 PM	22	0.07 kUA/l	0	Negative	
011990000487	wx1	5	2:28:18 PM	30	0.15 kUA/l	0	Negative	
011990000487	t20	5	2:29:19 PM	18	0.01 kUA/l	0		
011990000911	fx5	5	3:43:15 PM	22	0.06 kUA/l	0	Negative	
011990000911	fx1	5	3:44:15 PM	31	0.17 kUA/l	0	Negative	
011990000911	fx2	5	3:45:15 PM	14	Low RU	0	Negative	
011990000911	fx22	5	3:46:15 PM	31	0.17 kUA/l	0	Negative	
011990000911	gx2	5	3:47:15 PM	23	0.08 kUA/l	0	Negative	
011990000911	wx1	5	3:48:15 PM	449	5.02 kUA/l	3	Positive	
011990000911	t20	5	3:49:15 PM	23	0.07 kUA/l	0		
011990000912	fx5	5	2:55:17 PM	29	0.14 kUA/l	0	Negative	
011990000912	fx1	5	2:56:17 PM	18	0.02 kUA/l	0	Negative	
011990000912	fx2	5	2:57:17 PM	28	0.14 kUA/l	0	Negative	
011990000912	fx22	5	2:58:17 PM	32	0.18 kUA/l	0	Negative	
011990000912	gx2	5	2:59:17 PM	23	0.08 kUA/l	0	Negative	
011990000912	wx1	5	3:00:17 PM	34	0.20 kUA/l	0	Negative	
011990000912	t20	5	3:01:17 PM	17	Low RU	0		
011990000913	fx5	5	2:15:19 PM	19	0.03 kUA/l	0	Negative	
011990000913	fx1	5	2:16:19 PM	19	0.02 kUA/l	0	Negative	
011990000913	fx2	5	2:17:19 PM	15	Low RU	0	Negative	
011990000913	fx22	5	2:18:19 PM	28	0.14 kUA/l	0	Negative	
011990000913	gx2	5	2:19:18 PM	21	0.05 kUA/l	0	Negative	
011990000913	wx1	5	2:20:19 PM	30	0.16 kUA/l	0	Negative	
011990000913	t20	5	2:21:19 PM	17	Low RU	0		
011990000927	fx5	5	2:07:19 PM	24	0.09 kUA/l	0	Negative	
011990000927	fx1	5	2:08:19 PM	18	0.01 kUA/l	0	Negative	
011990000927	fx2	5	2:09:19 PM	22	0.07 kUA/l	0	Negative	
011990000927	fx22	5	2:10:19 PM	31	0.17 kUA/l	0	Negative	
011990000927	gx2	5	2:11:19 PM	22	0.07 kUA/l	0	Negative	
011990000927	wx1	5	2:12:19 PM	34	0.20 kUA/l	0	Negative	
011990000927	t20	5	2:13:19 PM	23	0.07 kUA/l	0		
011990000929	fx5	5	3:27:16 PM	32	0.17 kUA/l	0	Negative	
011990000929	fx1	5	3:28:16 PM	18	0.01 kUA/l	0	Negative	
011990000929	fx2	5	3:29:16 PM	21	0.04 kUA/l	0	Negative	
011990000929	fx22	5	3:30:16 PM	33	0.19 kUA/l	0	Negative	



## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000929	gx2	5	3:31:15 PM	35	0.21 KUA/I	0	Negative	
011990000929	wx1	5	3:32:15 PM	36	0.23 KUA/I	0	Negative	
011990000929	t20	5	3:33:15 PM	18	0.01 KUA/I	0		
011990000930	fx5	5	2:47:17 PM	20	0.04 KUA/I	0	Negative	
011990000930	fx1	5	2:48:18 PM	18	0.01 KUA/I	0	Negative	
011990000930	fx2	5	2:49:18 PM	15	Low RU	0	Negative	
011990000930	fx22	5	2:50:17 PM	31	0.17 KUA/I	0	Negative	
011990000930	gx2	5	2:51:17 PM	24	0.08 KUA/I	0	Negative	
011990000930	wx1	5	2:52:17 PM	30	0.16 KUA/I	0	Negative	
011990000930	t20	5	2:53:17 PM	17	Low RU	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000927	phad	2:06:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000927	fx5	2:07:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000927	fx1	2:08:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000927	fx2	2:09:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000927	fx22	2:10:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000927	gx2	2:11:19 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000927	wx1	2:12:19 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000927	t20	2:13:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	phad	2:14:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	fx5	2:15:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	fx1	2:16:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	fx2	2:17:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	fx22	2:18:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	gx2	2:19:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	wx1	2:20:19 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000913	t20	2:21:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	phad	2:22:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	fx5	2:23:18 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	fx1	2:24:18 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	fx2	2:25:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	fx22	2:26:18 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	gx2	2:27:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	wx1	2:28:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000487	t20	2:29:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	phad	2:30:18 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	fx5	2:31:18 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	fx1	2:32:18 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	fx2	2:33:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	fx22	2:34:18 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	gx2	2:35:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	wx1	2:36:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000475	t20	2:37:18 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000463	phad	2:38:18 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000463	fx5	2:39:18 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000463	fx1	2:40:18 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000463	fx2	2:41:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000463	fx22	2:42:18 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000463	gx2	2:43:18 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000463	wx1	2:44:18 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000463	t20	2:45:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	phad	2:46:17 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	fx5	2:47:17 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	fx1	2:48:18 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	fx2	2:49:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	fx22	2:50:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	gx2	2:51:17 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	wx1	2:52:17 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000930	t20	2:53:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	phad	2:54:17 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	fx5	2:55:17 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	fx1	2:56:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	fx2	2:57:17 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	fx22	2:58:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	gx2	2:59:17 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	wx1	3:00:17 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000912	t20	3:01:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	phad	3:10:17 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	fx5	3:11:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	fx1	3:12:16 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	fx2	3:13:16 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	fx22	3:14:16 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	gx2	3:15:16 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	wx1	3:16:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000486	t20	3:17:16 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	phad	3:02:17 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	fx5	3:03:17 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	fx1	3:04:17 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	fx2	3:05:17 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	fx22	3:06:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	gx2	3:07:17 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	wx1	3:08:17 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000474	t20	3:09:17 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	phad	3:18:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	fx5	3:19:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	fx1	3:20:16 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	fx2	3:21:16 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	fx22	3:22:16 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	gx2	3:23:16 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	wx1	3:24:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000462	t20	3:25:16 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	phad	3:26:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	fx5	3:27:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	fx1	3:28:16 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	fx2	3:29:16 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	fx22	3:30:16 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	gx2	3:31:15 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	wx1	3:32:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000929	t20	3:33:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	phad	3:42:15 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	fx5	3:43:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	fx1	3:44:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	fx2	3:45:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	fx22	3:46:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	gx2	3:47:15 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	wx1	3:48:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000911	t20	3:49:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000485	phad	3:34:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000485	fx5	3:35:16 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000485	fx1	3:36:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000485	fx2	3:37:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000485	fx22	3:38:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000485	gx2	3:39:15 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000485	wx1	3:40:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000485	t20	3:41:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	phad	3:58:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	fx5	3:59:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	fx1	4:00:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	fx2	4:01:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	fx22	4:02:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	gx2	4:03:14 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	wx1	4:04:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000473	t20	4:05:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	phad	3:50:15 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	fx5	3:51:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	fx1	3:52:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	fx2	3:53:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	fx22	3:54:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	gx2	3:55:14 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	wx1	3:56:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000461	t20	3:57:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	