


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200128-388  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 28-Jan-20 9:57:04 AM  
Ended: 28-Jan-20 1:57:56 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	9:57:04 AM	299	0.66	OK	
CC-2	17.5	9:58:04 AM	6081	16.7	OK	

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	9:59:04 AM	3861	9.81 kUA/l	CZ1AR	OK	
slgE Neg	d1	1	10:00:04 AM	22	0.01 kUA/l	BL1CP	OK	
slgE L	e1	1	10:01:04 AM	1439	3.37 kUA/l	CZ4AS	OK	
IQC271119	d1	1	10:02:04 AM	6268	17.4 kUA/l	<lab>	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000626	phad	5	1:12:56 PM	30	0.16 PAU/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000627	phad	5	1:04:57 PM	84	0.79 PAU/I	2	Positive	
011990000628	phad	5	1:28:56 PM	29	0.14 PAU/I	0	Negative	
011990000629	phad	5	1:20:56 PM	308	3.38 PAU/I	2	Positive	
011990000676	phad	5	10:32:03 AM	41	0.29 PAU/I	0	Negative	
011990000677	phad	5	10:40:02 AM	31	0.16 PAU/I	0	Negative	
011990000678	phad	5	10:08:04 AM	94	0.90 PAU/I	2	Positive	
011990000679	phad	5	10:16:03 AM	52	0.42 PAU/I	1	Positive	
011990000680	phad	5	10:24:03 AM	82	0.77 PAU/I	2	Positive	
011990000681	phad	5	10:48:02 AM	736	8.38 PAU/I	3	Positive	
011990000682	phad	5	11:20:01 AM	28	0.13 PAU/I	0	Negative	
011990000683	phad	5	11:12:01 AM	113	1.12 PAU/I	2	Positive	
011990000684	phad	5	10:56:02 AM	25	0.10 PAU/I	0	Negative	
011990000685	phad	5	11:04:02 AM	30	0.16 PAU/I	0	Negative	
011990000686	phad	5	12:08:59 PM	74	0.67 PAU/I	1	Positive	
011990000687	phad	5	11:42:00 AM	173	1.82 PAU/I	2	Positive	
011990000688	phad	5	12:00:59 PM	95	0.92 PAU/I	2	Positive	
011990000689	phad	5	11:28:01 AM	29	0.15 PAU/I	0	Negative	
011990000690	phad	5	11:35:00 AM	5531	74.6 PAU/I	5	Positive	
011990000691	phad	5	12:56:57 PM	720	8.20 PAU/I	3	Positive	
011990000692	phad	5	12:32:58 PM	601	6.80 PAU/I	3	Positive	
011990000693	phad	5	12:44:57 PM	111	1.11 PAU/I	2	Positive	
011990000694	phad	5	12:16:59 PM	48	0.37 PAU/I	1	Positive	
011990000695	phad	5	12:24:58 PM	463	5.18 PAU/I	3	Positive	
011990000696	phad	5	1:49:55 PM	2125	25.5 PAU/I	4	Positive	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000626	fx5	5	1:13:56 PM	19	0.02 kUA/I	0	Negative	
011990000626	fx1	5	1:14:57 PM	16	Low RU	0	Negative	
011990000626	fx2	5	1:15:56 PM	13	Low RU	0	Negative	
011990000626	fx22	5	1:16:56 PM	26	0.10 kUA/I	0	Negative	
011990000626	gx2	5	1:17:56 PM	21	0.05 kUA/I	0	Negative	
011990000626	wx1	5	1:18:56 PM	29	0.14 kUA/I	0	Negative	
011990000626	t20	5	1:19:56 PM	15	Low RU	0		
011990000627	fx5	5	1:05:57 PM	28	0.13 kUA/I	0	Negative	
011990000627	fx1	5	1:06:57 PM	31	0.17 kUA/I	0	Negative	
011990000627	fx2	5	1:07:57 PM	14	Low RU	0	Negative	
011990000627	fx22	5	1:08:57 PM	40	0.27 kUA/I	0	Negative	
011990000627	gx2	5	1:09:57 PM	20	0.04 kUA/I	0	Negative	
011990000627	wx1	5	1:10:57 PM	540	6.08 kUA/I	3	Positive	
011990000627	t20	5	1:11:56 PM	164	1.72 kUA/I	2		
011990000628	fx5	5	1:29:56 PM	44	0.32 kUA/I	0	Negative	
011990000628	fx1	5	1:30:56 PM	20	0.03 kUA/I	0	Negative	
011990000628	fx2	5	1:31:56 PM	14	Low RU	0	Negative	
011990000628	fx22	5	1:32:56 PM	30	0.15 kUA/I	0	Negative	
011990000628	gx2	5	1:33:56 PM	19	0.03 kUA/I	0	Negative	
011990000628	wx1	5	1:34:56 PM	28	0.14 kUA/I	0	Negative	
011990000628	t20	5	1:35:56 PM	18	0.01 kUA/I	0		
011990000629	fx5	5	1:21:56 PM	388	4.31 kUA/I	3	Positive	
011990000629	fx1	5	1:22:56 PM	26	0.11 kUA/I	0	Negative	
011990000629	fx2	5	1:23:56 PM	555	6.26 kUA/I	3	Positive	
011990000629	fx22	5	1:24:56 PM	42	0.29 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000629	gx2	5	1:25:56 PM	38	0.25 kUA/I	0	Negative	
011990000629	wx1	5	1:26:56 PM	48	0.36 kUA/I	1	Positive	
011990000629	t20	5	1:27:56 PM	25	0.10 kUA/I	0		
011990000676	fx5	5	10:33:02 AM	21	0.05 kUA/I	0	Negative	
011990000676	fx1	5	10:34:02 AM	15	Low RU	0	Negative	
011990000676	fx2	5	10:35:02 AM	20	0.03 kUA/I	0	Negative	
011990000676	fx22	5	10:36:03 AM	26	0.11 kUA/I	0	Negative	
011990000676	gx2	5	10:37:03 AM	20	0.03 kUA/I	0	Negative	
011990000676	wx1	5	10:38:02 AM	31	0.16 kUA/I	0	Negative	
011990000676	t20	5	10:39:02 AM	16	Low RU	0		
011990000677	fx5	5	10:41:02 AM	22	0.06 kUA/I	0	Negative	
011990000677	fx1	5	10:42:02 AM	15	Low RU	0	Negative	
011990000677	fx2	5	10:43:02 AM	31	0.17 kUA/I	0	Negative	
011990000677	fx22	5	10:44:02 AM	26	0.11 kUA/I	0	Negative	
011990000677	gx2	5	10:45:02 AM	21	0.05 kUA/I	0	Negative	
011990000677	wx1	5	10:46:02 AM	34	0.20 kUA/I	0	Negative	
011990000677	t20	5	10:47:02 AM	17	Low RU	0		
011990000678	fx5	5	10:09:04 AM	59	0.49 kUA/I	1	Positive	
011990000678	fx1	5	10:10:04 AM	35	0.22 kUA/I	0	Negative	
011990000678	fx2	5	10:11:03 AM	77	0.70 kUA/I	2	Positive	
011990000678	fx22	5	10:12:03 AM	37	0.24 kUA/I	0	Negative	
011990000678	gx2	5	10:13:03 AM	33	0.19 kUA/I	0	Negative	
011990000678	wx1	5	10:14:03 AM	40	0.27 kUA/I	0	Negative	
011990000678	t20	5	10:15:03 AM	22	0.07 kUA/I	0		
011990000679	fx5	5	10:17:04 AM	72	0.65 kUA/I	1	Positive	
011990000679	fx1	5	10:18:03 AM	35	0.22 kUA/I	0	Negative	
011990000679	fx2	5	10:19:03 AM	16	Low RU	0	Negative	
011990000679	fx22	5	10:20:03 AM	32	0.18 kUA/I	0	Negative	
011990000679	gx2	5	10:21:03 AM	21	0.05 kUA/I	0	Negative	
011990000679	wx1	5	10:22:03 AM	31	0.16 kUA/I	0	Negative	
011990000679	t20	5	10:23:03 AM	19	0.03 kUA/I	0		
011990000680	fx5	5	10:25:03 AM	33	0.19 kUA/I	0	Negative	
011990000680	fx1	5	10:26:03 AM	19	0.03 kUA/I	0	Negative	
011990000680	fx2	5	10:27:03 AM	166	1.74 kUA/I	2	Positive	
011990000680	fx22	5	10:28:03 AM	28	0.13 kUA/I	0	Negative	
011990000680	gx2	5	10:29:03 AM	22	0.06 kUA/I	0	Negative	
011990000680	wx1	5	10:30:03 AM	40	0.27 kUA/I	0	Negative	
011990000680	t20	5	10:31:03 AM	22	0.06 kUA/I	0		
011990000681	fx5	5	10:49:02 AM	26	0.11 kUA/I	0	Negative	
011990000681	fx1	5	10:50:02 AM	38	0.25 kUA/I	0	Negative	
011990000681	fx2	5	10:51:02 AM	58	0.48 kUA/I	1	Positive	
011990000681	fx22	5	10:52:02 AM	43	0.31 kUA/I	0	Negative	
011990000681	gx2	5	10:53:02 AM	32	0.18 kUA/I	0	Negative	
011990000681	wx1	5	10:54:02 AM	169	1.77 kUA/I	2	Positive	
011990000681	t20	5	10:55:02 AM	66	0.58 kUA/I	1		
011990000682	fx5	5	11:21:01 AM	22	0.06 kUA/I	0	Negative	
011990000682	fx1	5	11:22:01 AM	16	Low RU	0	Negative	
011990000682	fx2	5	11:23:01 AM	14	Low RU	0	Negative	
011990000682	fx22	5	11:24:01 AM	28	0.13 kUA/I	0	Negative	
011990000682	gx2	5	11:25:01 AM	22	0.06 kUA/I	0	Negative	
011990000682	wx1	5	11:26:01 AM	31	0.16 kUA/I	0	Negative	
011990000682	t20	5	11:27:00 AM	16	Low RU	0		
011990000683	fx5	5	11:13:01 AM	25	0.10 kUA/I	0	Negative	
011990000683	fx1	5	11:14:01 AM	27	0.13 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000683	fx2	5	11:15:01 AM	68	0.60 kUA/l	1	Positive	
011990000683	fx22	5	11:16:01 AM	29	0.14 kUA/l	0	Negative	
011990000683	gx2	5	11:17:01 AM	25	0.10 kUA/l	0	Negative	
011990000683	wx1	5	11:18:01 AM	42	0.30 kUA/l	0	Negative	
011990000683	t20	5	11:19:01 AM	23	0.07 kUA/l	0		
011990000684	fx5	5	10:57:02 AM	19	0.03 kUA/l	0	Negative	
011990000684	fx1	5	10:58:02 AM	16	Low RU	0	Negative	
011990000684	fx2	5	10:59:02 AM	10	Low RU	0	Negative	
011990000684	fx22	5	11:00:02 AM	28	0.14 kUA/l	0	Negative	
011990000684	gx2	5	11:01:01 AM	19	0.03 kUA/l	0	Negative	
011990000684	wx1	5	11:02:01 AM	31	0.16 kUA/l	0	Negative	
011990000684	t20	5	11:03:01 AM	16	Low RU	0		
011990000685	fx5	5	11:05:02 AM	19	0.02 kUA/l	0	Negative	
011990000685	fx1	5	11:06:01 AM	15	Low RU	0	Negative	
011990000685	fx2	5	11:07:01 AM	12	Low RU	0	Negative	
011990000685	fx22	5	11:08:01 AM	29	0.14 kUA/l	0	Negative	
011990000685	gx2	5	11:09:01 AM	22	0.06 kUA/l	0	Negative	
011990000685	wx1	5	11:10:01 AM	29	0.15 kUA/l	0	Negative	
011990000685	t20	5	11:11:01 AM	16	Low RU	0		
011990000686	fx5	5	12:09:59 PM	47	0.36 kUA/l	1	Positive	
011990000686	fx1	5	12:10:59 PM	18	0.01 kUA/l	0	Negative	
011990000686	fx2	5	12:11:59 PM	12	Low RU	0	Negative	
011990000686	fx22	5	12:12:59 PM	28	0.13 kUA/l	0	Negative	
011990000686	gx2	5	12:13:59 PM	25	0.10 kUA/l	0	Negative	
011990000686	wx1	5	12:14:59 PM	33	0.19 kUA/l	0	Negative	
011990000686	t20	5	12:15:59 PM	23	0.07 kUA/l	0		
011990000687	fx5	5	11:43:00 AM	23	0.08 kUA/l	0	Negative	
011990000687	fx1	5	11:44:00 AM	18	0.02 kUA/l	0	Negative	
011990000687	fx2	5	11:45:00 AM	12	Low RU	0	Negative	
011990000687	fx22	5	11:57:59 AM	28	0.13 kUA/l	0	Negative	
011990000687	gx2	5	11:46:00 AM	21	0.05 kUA/l	0	Negative	
011990000687	wx1	5	11:58:59 AM	225	2.42 kUA/l	2	Positive	
011990000687	t20	5	12:00:00 PM	16	Low RU	0		
011990000688	fx5	5	12:01:59 PM	19	0.03 kUA/l	0	Negative	
011990000688	fx1	5	12:02:59 PM	18	0.02 kUA/l	0	Negative	
011990000688	fx2	5	12:03:59 PM	12	Low RU	0	Negative	
011990000688	fx22	5	12:04:59 PM	27	0.12 kUA/l	0	Negative	
011990000688	gx2	5	12:05:59 PM	19	0.03 kUA/l	0	Negative	
011990000688	wx1	5	12:06:59 PM	28	0.13 kUA/l	0	Negative	
011990000688	t20	5	12:07:59 PM	16	Low RU	0		
011990000689	fx5	5	11:29:00 AM	36	0.22 kUA/l	0	Negative	
011990000689	fx1	5	11:30:00 AM	21	0.05 kUA/l	0	Negative	
011990000689	fx2	5	11:31:00 AM	16	Low RU	0	Negative	
011990000689	fx22	5	11:55:59 AM	29	0.15 kUA/l	0	Negative	
011990000689	gx2	5	11:32:01 AM	27	0.12 kUA/l	0	Negative	
011990000689	wx1	5	11:33:00 AM	33	0.19 kUA/l	0	Negative	
011990000689	t20	5	11:34:00 AM	19	0.02 kUA/l	0		
011990000690	fx5	5	11:36:00 AM	24	0.08 kUA/l	0	Negative	
011990000690	fx1	5	11:37:00 AM	19	0.02 kUA/l	0	Negative	
011990000690	fx2	5	11:38:00 AM	232	2.50 kUA/l	2	Positive	
011990000690	fx22	5	11:56:59 AM	29	0.14 kUA/l	0	Negative	
011990000690	gx2	5	11:39:00 AM	24	0.09 kUA/l	0	Negative	
011990000690	wx1	5	11:40:00 AM	36	0.23 kUA/l	0	Negative	
011990000690	t20	5	11:41:00 AM	20	0.03 kUA/l	0		

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000691	fx5	5	12:57:57 PM	24	0.08 kUA/l	0	Negative	
011990000691	fx1	5	12:58:57 PM	23	0.07 kUA/l	0	Negative	
011990000691	fx2	5	12:59:57 PM	16	Low RU	0	Negative	
011990000691	fx22	5	1:00:57 PM	32	0.18 kUA/l	0	Negative	
011990000691	gx2	5	1:01:57 PM	29	0.15 kUA/l	0	Negative	
011990000691	wx1	5	1:02:57 PM	58	0.48 kUA/l	1	Positive	
011990000691	t20	5	1:03:57 PM	54	0.43 kUA/l	1		
011990000692	fx5	5	12:33:58 PM	27	0.13 kUA/l	0	Negative	
011990000692	fx1	5	12:34:58 PM	17	Low RU	0	Negative	
011990000692	fx2	5	12:35:58 PM	1255	14.6 kUA/l	3	Positive	
011990000692	fx22	5	12:36:58 PM	45	0.34 kUA/l	0	Negative	
011990000692	gx2	5	12:41:58 PM	252	2.73 kUA/l	2	Positive	
011990000692	wx1	5	12:42:58 PM	93	0.90 kUA/l	2	Positive	
011990000692	t20	5	12:43:58 PM	35	0.22 kUA/l	0		
011990000693	fx5	5	12:45:58 PM	33	0.19 kUA/l	0	Negative	
011990000693	fx1	5	12:46:58 PM	35	0.22 kUA/l	0	Negative	
011990000693	fx2	5	12:51:57 PM	14	Low RU	0	Negative	
011990000693	fx22	5	12:52:57 PM	27	0.12 kUA/l	0	Negative	
011990000693	gx2	5	12:53:57 PM	20	0.03 kUA/l	0	Negative	
011990000693	wx1	5	12:54:57 PM	54	0.44 kUA/l	1	Positive	
011990000693	t20	5	12:55:57 PM	18	0.01 kUA/l	0		
011990000694	fx5	5	12:17:59 PM	22	0.07 kUA/l	0	Negative	
011990000694	fx1	5	12:18:59 PM	17	Low RU	0	Negative	
011990000694	fx2	5	12:19:59 PM	17	Low RU	0	Negative	
011990000694	fx22	5	12:20:58 PM	26	0.11 kUA/l	0	Negative	
011990000694	gx2	5	12:21:58 PM	24	0.08 kUA/l	0	Negative	
011990000694	wx1	5	12:22:58 PM	29	0.14 kUA/l	0	Negative	
011990000694	t20	5	12:23:58 PM	17	Low RU	0		
011990000695	fx5	5	12:25:58 PM	27	0.12 kUA/l	0	Negative	
011990000695	fx1	5	12:26:58 PM	36	0.23 kUA/l	0	Negative	
011990000695	fx2	5	12:27:58 PM	1344	15.7 kUA/l	3	Positive	
011990000695	fx22	5	12:28:58 PM	56	0.46 kUA/l	1	Positive	
011990000695	gx2	5	12:29:58 PM	86	0.81 kUA/l	2	Positive	
011990000695	wx1	5	12:30:58 PM	112	1.12 kUA/l	2	Positive	
011990000695	t20	5	12:31:58 PM	59	0.50 kUA/l	1		
011990000696	fx5	5	1:50:55 PM	66	0.57 kUA/l	1	Positive	
011990000696	fx1	5	1:51:55 PM	29	0.15 kUA/l	0	Negative	
011990000696	fx2	5	1:52:55 PM	78	0.72 kUA/l	2	Positive	
011990000696	fx22	5	1:53:55 PM	44	0.32 kUA/l	0	Negative	
011990000696	gx2	5	1:54:55 PM	53	0.42 kUA/l	1	Positive	
011990000696	wx1	5	1:55:55 PM	123	1.24 kUA/l	2	Positive	
011990000696	t20	5	1:56:55 PM	114	1.14 kUA/l	2		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	



# Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	9:57:04 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	9:58:04 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	9:59:04 AM	352D0 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ1AR 21/08		Exp
slgE Neg	d1	10:00:04 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		BL1CP 19/12!		Exp
slgE L	e1	10:01:04 AM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ4AS 20/03		
IQC271119	d1	10:02:04 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000680	phad	10:24:03 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000680	fx5	10:25:03 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000680	fx1	10:26:03 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000680	fx2	10:27:03 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000680	fx22	10:28:03 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000680	gx2	10:29:03 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000680	wx1	10:30:03 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000680	t20	10:31:03 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	phad	10:16:03 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	fx5	10:17:04 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	fx1	10:18:03 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	fx2	10:19:03 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	fx22	10:20:03 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	gx2	10:21:03 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	wx1	10:22:03 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000679	t20	10:23:03 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	phad	10:08:04 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	fx5	10:09:04 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	fx1	10:10:04 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	fx2	10:11:03 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	fx22	10:12:03 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	gx2	10:13:03 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	wx1	10:14:03 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000678	t20	10:15:03 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	phad	10:40:02 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	fx5	10:41:02 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	fx1	10:42:02 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	fx2	10:43:02 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	fx22	10:44:02 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	gx2	10:45:02 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	wx1	10:46:02 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000677	t20	10:47:02 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	phad	10:32:03 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	fx5	10:33:02 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	fx1	10:34:02 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	fx2	10:35:02 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	fx22	10:36:03 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	gx2	10:37:03 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	wx1	10:38:02 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000676	t20	10:39:02 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	phad	11:04:02 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	fx5	11:05:02 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	fx1	11:06:01 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	fx2	11:07:01 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	fx22	11:08:01 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	gx2	11:09:01 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	wx1	11:10:01 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000685	t20	11:11:01 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000684	phad	10:56:02 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000684	fx5	10:57:02 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000684	fx1	10:58:02 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000684	fx2	10:59:02 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000684	fx22	11:00:02 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000684	gx2	11:01:01 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000684	wx1	11:02:01 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000684	t20	11:03:01 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	phad	11:12:01 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	fx5	11:13:01 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	fx1	11:14:01 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	fx2	11:15:01 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	fx22	11:16:01 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	gx2	11:17:01 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	wx1	11:18:01 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000683	t20	11:19:01 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	phad	11:20:01 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	fx5	11:21:01 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	fx1	11:22:01 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	fx2	11:23:01 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	fx22	11:24:01 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	gx2	11:25:01 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	wx1	11:26:01 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000682	t20	11:27:00 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	phad	10:48:02 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	fx5	10:49:02 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	fx1	10:50:02 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	fx2	10:51:02 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	fx22	10:52:02 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	gx2	10:53:02 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	wx1	10:54:02 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000681	t20	10:55:02 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	phad	11:35:00 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	fx5	11:36:00 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	fx1	11:37:00 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	fx2	11:38:00 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	fx22	11:56:59 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	gx2	11:39:00 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	wx1	11:40:00 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000690	t20	11:41:00 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	phad	11:28:01 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	fx5	11:29:00 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	fx1	11:30:00 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	fx2	11:31:00 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	fx22	11:55:59 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	gx2	11:32:01 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	wx1	11:33:00 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000689	t20	11:34:00 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	phad	12:00:59 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	fx5	12:01:59 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	fx1	12:02:59 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	fx2	12:03:59 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	fx22	12:04:59 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	gx2	12:05:59 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	wx1	12:06:59 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000688	t20	12:07:59 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	phad	11:42:00 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	fx5	11:43:00 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	fx1	11:44:00 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	fx2	11:45:00 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	fx22	11:57:59 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	gx2	11:46:00 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	wx1	11:58:59 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000687	t20	12:00:00 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000686	phad	12:08:59 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000686	fx5	12:09:59 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000686	fx1	12:10:59 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000686	fx2	12:11:59 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000686	fx22	12:12:59 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000686	gx2	12:13:59 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000686	wx1	12:14:59 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000686	t20	12:15:59 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	phad	12:24:58 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	fx5	12:25:58 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	fx1	12:26:58 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	fx2	12:27:58 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	fx22	12:28:58 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	gx2	12:29:58 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	wx1	12:30:58 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000695	t20	12:31:58 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	phad	12:16:59 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	fx5	12:17:59 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	fx1	12:18:59 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	fx2	12:19:59 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	fx22	12:20:58 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	gx2	12:21:58 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	wx1	12:22:58 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000694	t20	12:23:58 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	phad	12:44:57 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	fx5	12:45:58 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	fx1	12:46:58 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	fx2	12:51:57 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	fx22	12:52:57 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	gx2	12:53:57 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	wx1	12:54:57 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000693	t20	12:55:57 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	phad	12:32:58 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	fx5	12:33:58 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	fx1	12:34:58 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	fx2	12:35:58 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	fx22	12:36:58 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	gx2	12:41:58 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	wx1	12:42:58 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000692	t20	12:43:58 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	phad	12:56:57 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	fx5	12:57:57 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	fx1	12:58:57 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	fx2	12:59:57 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	fx22	1:00:57 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	gx2	1:01:57 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	wx1	1:02:57 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000691	t20	1:03:57 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	phad	1:04:57 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	fx5	1:05:57 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	fx1	1:06:57 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	fx2	1:07:57 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	fx22	1:08:57 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	gx2	1:09:57 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	wx1	1:10:57 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000627	t20	1:11:56 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	phad	1:12:56 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	fx5	1:13:56 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	fx1	1:14:57 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	fx2	1:15:56 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	fx22	1:16:56 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	gx2	1:17:56 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	wx1	1:18:56 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000626	t20	1:19:56 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000629	phad	1:20:56 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000629	fx5	1:21:56 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000629	fx1	1:22:56 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000629	fx2	1:23:56 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000629	fx22	1:24:56 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000629	gx2	1:25:56 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000629	wx1	1:26:56 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000629	t20	1:27:56 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	phad	1:28:56 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	fx5	1:29:56 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	fx1	1:30:56 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	fx2	1:31:56 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	fx22	1:32:56 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	gx2	1:33:56 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	wx1	1:34:56 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000628	t20	1:35:56 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	phad	1:49:55 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	fx5	1:50:55 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	fx1	1:51:55 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	fx2	1:52:55 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	fx22	1:53:55 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	gx2	1:54:55 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	wx1	1:55:55 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000696	t20	1:56:55 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

### Errors

#### Expired lot numbers used for Quality Control

352D0	ImmunoCAP
BL1CP	Quality control vial


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200128-388  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 28-Jan-20 1:57:55 PM  
Ended: 28-Jan-20 5:34:50 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

No curve controls in this run

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	3:31:51 PM	5928	16.2 kUA/l	<lab>	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000801	phad	5	2:29:54 PM	47	0.35 PAU/l	1	Positive	
011990000802	phad	5	2:21:54 PM	53	0.42 PAU/l	1	Positive	
011990000803	phad	5	2:13:54 PM	30	0.16 PAU/l	0	Negative	
011990000804	phad	5	2:05:55 PM	112	1.11 PAU/l	2	Positive	
011990000805	phad	5	1:57:55 PM	199	2.12 PAU/l	2	Positive	
011990000806	phad	5	3:09:52 PM	209	2.24 PAU/l	2	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000807	phad	5	3:01:53 PM	33	0.19 PAU/I	0	Negative	
011990000808	phad	5	2:53:53 PM	29	0.14 PAU/I	0	Negative	
011990000809	phad	5	2:45:53 PM	108	1.06 PAU/I	2	Positive	
011990000810	phad	5	2:37:53 PM	136	1.39 PAU/I	2	Positive	
011990000811	phad	5	3:34:52 PM	47	0.35 PAU/I	1	Positive	
011990000812	phad	5	3:42:51 PM	38	0.25 PAU/I	0	Negative	
011990000813	phad	5	5:02:48 PM	30	0.15 PAU/I	0	Negative	
011990000814	phad	5	4:06:50 PM	198	2.10 PAU/I	2	Positive	
011990000815	phad	5	3:58:50 PM	78	0.72 PAU/I	2	Positive	
011990000816	phad	5	3:50:51 PM	35	0.21 PAU/I	0	Negative	
011990000817	phad	5	5:10:48 PM	36	0.22 PAU/I	0	Negative	
011990000818	phad	5	4:54:48 PM	28	0.14 PAU/I	0	Negative	
011990000819	phad	5	4:38:49 PM	29	0.14 PAU/I	0	Negative	
011990000820	phad	5	4:14:50 PM	1652	19.5 PAU/I	4	Positive	
011990000821	phad	5	5:18:48 PM	47	0.36 PAU/I	1	Positive	
011990000822	phad	5	4:46:49 PM	31	0.17 PAU/I	0	Negative	
011990000823	phad	5	4:30:50 PM	33	0.19 PAU/I	0	Negative	
011990000824	phad	5	4:22:50 PM	311	3.42 PAU/I	2	Positive	
011990000825	phad	5	5:26:47 PM	205	2.18 PAU/I	2	Positive	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000801	fx5	5	2:30:54 PM	21	0.05 kUA/I	0	Negative	
011990000801	fx1	5	2:31:54 PM	23	0.07 kUA/I	0	Negative	
011990000801	fx2	5	2:32:54 PM	29	0.14 kUA/I	0	Negative	
011990000801	fx22	5	2:33:54 PM	29	0.14 kUA/I	0	Negative	
011990000801	gx2	5	2:34:54 PM	19	0.02 kUA/I	0	Negative	
011990000801	wx1	5	2:35:53 PM	31	0.16 kUA/I	0	Negative	
011990000801	t20	5	2:36:53 PM	18	0.01 kUA/I	0		
011990000802	fx5	5	2:22:54 PM	23	0.07 kUA/I	0	Negative	
011990000802	fx1	5	2:23:54 PM	20	0.04 kUA/I	0	Negative	
011990000802	fx2	5	2:24:54 PM	13	Low RU	0	Negative	
011990000802	fx22	5	2:25:54 PM	28	0.13 kUA/I	0	Negative	
011990000802	gx2	5	2:26:54 PM	34	0.20 kUA/I	0	Negative	
011990000802	wx1	5	2:27:54 PM	31	0.16 kUA/I	0	Negative	
011990000802	t20	5	2:28:54 PM	26	0.11 kUA/I	0		
011990000803	fx5	5	2:14:54 PM	21	0.04 kUA/I	0	Negative	
011990000803	fx1	5	2:15:54 PM	20	0.04 kUA/I	0	Negative	
011990000803	fx2	5	2:16:54 PM	16	Low RU	0	Negative	
011990000803	fx22	5	2:17:54 PM	27	0.12 kUA/I	0	Negative	
011990000803	gx2	5	2:18:54 PM	19	0.02 kUA/I	0	Negative	
011990000803	wx1	5	2:19:54 PM	28	0.13 kUA/I	0	Negative	
011990000803	t20	5	2:20:54 PM	16	Low RU	0		
011990000804	fx5	5	2:06:54 PM	20	0.04 kUA/I	0	Negative	
011990000804	fx1	5	2:07:54 PM	15	Low RU	0	Negative	
011990000804	fx2	5	2:08:54 PM	40	0.28 kUA/I	0	Negative	
011990000804	fx22	5	2:09:54 PM	30	0.15 kUA/I	0	Negative	
011990000804	gx2	5	2:10:55 PM	22	0.06 kUA/I	0	Negative	
011990000804	wx1	5	2:11:55 PM	30	0.15 kUA/I	0	Negative	
011990000804	t20	5	2:12:54 PM	15	Low RU	0		
011990000805	fx5	5	1:58:55 PM	32	0.19 kUA/I	0	Negative	
011990000805	fx1	5	1:59:55 PM	20	0.04 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000805	fx2	5	2:00:55 PM	405	4.51 kUA/l	3	Positive	
011990000805	fx22	5	2:01:55 PM	37	0.24 kUA/l	0	Negative	
011990000805	gx2	5	2:02:55 PM	43	0.31 kUA/l	0	Negative	
011990000805	wx1	5	2:03:55 PM	49	0.38 kUA/l	1	Positive	
011990000805	t20	5	2:04:54 PM	41	0.29 kUA/l	0		
011990000806	fx5	5	3:10:52 PM	31	0.17 kUA/l	0	Negative	
011990000806	fx1	5	3:11:52 PM	25	0.10 kUA/l	0	Negative	
011990000806	fx2	5	3:12:52 PM	86	0.81 kUA/l	2	Positive	
011990000806	fx22	5	3:13:52 PM	39	0.26 kUA/l	0	Negative	
011990000806	gx2	5	3:14:52 PM	61	0.52 kUA/l	1	Positive	
011990000806	wx1	5	3:32:51 PM	48	0.37 kUA/l	1	Positive	
011990000806	t20	5	3:33:51 PM	30	0.15 kUA/l	0		
011990000807	fx5	5	3:02:52 PM	19	0.03 kUA/l	0	Negative	
011990000807	fx1	5	3:03:52 PM	14	Low RU	0	Negative	
011990000807	fx2	5	3:04:52 PM	12	Low RU	0	Negative	
011990000807	fx22	5	3:05:52 PM	27	0.12 kUA/l	0	Negative	
011990000807	gx2	5	3:06:53 PM	19	0.03 kUA/l	0	Negative	
011990000807	wx1	5	3:07:52 PM	29	0.14 kUA/l	0	Negative	
011990000807	t20	5	3:08:52 PM	16	Low RU	0		
011990000808	fx5	5	2:54:53 PM	27	0.12 kUA/l	0	Negative	
011990000808	fx1	5	2:55:53 PM	17	Low RU	0	Negative	
011990000808	fx2	5	2:56:52 PM	12	Low RU	0	Negative	
011990000808	fx22	5	2:57:53 PM	28	0.13 kUA/l	0	Negative	
011990000808	gx2	5	2:58:53 PM	19	0.03 kUA/l	0	Negative	
011990000808	wx1	5	2:59:53 PM	28	0.14 kUA/l	0	Negative	
011990000808	t20	5	3:00:53 PM	17	Low RU	0		
011990000809	fx5	5	2:46:53 PM	22	0.07 kUA/l	0	Negative	
011990000809	fx1	5	2:47:53 PM	20	0.03 kUA/l	0	Negative	
011990000809	fx2	5	2:48:53 PM	25	0.10 kUA/l	0	Negative	
011990000809	fx22	5	2:49:53 PM	30	0.16 kUA/l	0	Negative	
011990000809	gx2	5	2:50:53 PM	24	0.09 kUA/l	0	Negative	
011990000809	wx1	5	2:51:53 PM	39	0.27 kUA/l	0	Negative	
011990000809	t20	5	2:52:53 PM	18	0.01 kUA/l	0		
011990000810	fx5	5	2:38:54 PM	28	0.13 kUA/l	0	Negative	
011990000810	fx1	5	2:39:53 PM	21	0.04 kUA/l	0	Negative	
011990000810	fx2	5	2:40:53 PM	73	0.66 kUA/l	1	Positive	
011990000810	fx22	5	2:41:53 PM	39	0.26 kUA/l	0	Negative	
011990000810	gx2	5	2:42:53 PM	48	0.37 kUA/l	1	Positive	
011990000810	wx1	5	2:43:53 PM	66	0.57 kUA/l	1	Positive	
011990000810	t20	5	2:44:53 PM	31	0.16 kUA/l	0		
011990000811	fx5	5	3:35:52 PM	23	0.07 kUA/l	0	Negative	
011990000811	fx1	5	3:36:51 PM	16	Low RU	0	Negative	
011990000811	fx2	5	3:37:51 PM	20	0.03 kUA/l	0	Negative	
011990000811	fx22	5	3:38:51 PM	28	0.13 kUA/l	0	Negative	
011990000811	gx2	5	3:39:51 PM	32	0.18 kUA/l	0	Negative	
011990000811	wx1	5	3:40:51 PM	31	0.17 kUA/l	0	Negative	
011990000811	t20	5	3:41:51 PM	17	Low RU	0		
011990000812	fx5	5	3:43:51 PM	19	0.02 kUA/l	0	Negative	
011990000812	fx1	5	3:44:51 PM	16	Low RU	0	Negative	
011990000812	fx2	5	3:45:51 PM	15	Low RU	0	Negative	
011990000812	fx22	5	3:46:51 PM	28	0.14 kUA/l	0	Negative	
011990000812	gx2	5	3:47:51 PM	21	0.05 kUA/l	0	Negative	
011990000812	wx1	5	3:48:51 PM	28	0.14 kUA/l	0	Negative	
011990000812	t20	5	3:49:51 PM	16	Low RU	0		



## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000813	fx5	5	5:03:48 PM	22	0.06 kUA/l	0	Negative	
011990000813	fx1	5	5:04:48 PM	24	0.08 kUA/l	0	Negative	
011990000813	fx2	5	5:05:48 PM	14	Low RU	0	Negative	
011990000813	fx22	5	5:06:48 PM	30	0.15 kUA/l	0	Negative	
011990000813	gx2	5	5:07:48 PM	20	0.04 kUA/l	0	Negative	
011990000813	wx1	5	5:08:48 PM	31	0.17 kUA/l	0	Negative	
011990000813	t20	5	5:09:48 PM	17	Low RU	0		
011990000814	fx5	5	4:07:50 PM	171	1.80 kUA/l	2	Positive	
011990000814	fx1	5	4:08:50 PM	518	5.83 kUA/l	3	Positive	
011990000814	fx2	5	4:09:50 PM	24	0.08 kUA/l	0	Negative	
011990000814	fx22	5	4:10:50 PM	605	6.85 kUA/l	3	Positive	
011990000814	gx2	5	4:11:50 PM	68	0.60 kUA/l	1	Positive	
011990000814	wx1	5	4:12:50 PM	212	2.27 kUA/l	2	Positive	
011990000814	t20	5	4:13:50 PM	216	2.32 kUA/l	2		
011990000815	fx5	5	3:59:50 PM	44	0.32 kUA/l	0	Negative	
011990000815	fx1	5	4:00:50 PM	23	0.08 kUA/l	0	Negative	
011990000815	fx2	5	4:01:50 PM	44	0.32 kUA/l	0	Negative	
011990000815	fx22	5	4:02:51 PM	36	0.23 kUA/l	0	Negative	
011990000815	gx2	5	4:03:51 PM	28	0.13 kUA/l	0	Negative	
011990000815	wx1	5	4:04:50 PM	39	0.26 kUA/l	0	Negative	
011990000815	t20	5	4:05:50 PM	26	0.10 kUA/l	0		
011990000816	fx5	5	3:51:51 PM	23	0.08 kUA/l	0	Negative	
011990000816	fx1	5	3:52:51 PM	18	0.01 kUA/l	0	Negative	
011990000816	fx2	5	3:53:51 PM	14	Low RU	0	Negative	
011990000816	fx22	5	3:54:51 PM	25	0.10 kUA/l	0	Negative	
011990000816	gx2	5	3:55:51 PM	21	0.04 kUA/l	0	Negative	
011990000816	wx1	5	3:56:50 PM	30	0.16 kUA/l	0	Negative	
011990000816	t20	5	3:57:50 PM	17	0.01 kUA/l	0		
011990000817	fx5	5	5:11:48 PM	26	0.11 kUA/l	0	Negative	
011990000817	fx1	5	5:12:48 PM	16	Low RU	0	Negative	
011990000817	fx2	5	5:13:48 PM	16	Low RU	0	Negative	
011990000817	fx22	5	5:14:48 PM	32	0.19 kUA/l	0	Negative	
011990000817	gx2	5	5:15:48 PM	21	0.04 kUA/l	0	Negative	
011990000817	wx1	5	5:16:48 PM	30	0.15 kUA/l	0	Negative	
011990000817	t20	5	5:17:48 PM	16	Low RU	0		
011990000818	fx5	5	4:55:49 PM	18	0.02 kUA/l	0	Negative	
011990000818	fx1	5	4:56:48 PM	17	0.01 kUA/l	0	Negative	
011990000818	fx2	5	4:57:48 PM	14	Low RU	0	Negative	
011990000818	fx22	5	4:58:49 PM	32	0.18 kUA/l	0	Negative	
011990000818	gx2	5	4:59:49 PM	19	0.02 kUA/l	0	Negative	
011990000818	wx1	5	5:00:48 PM	30	0.16 kUA/l	0	Negative	
011990000818	t20	5	5:01:48 PM	16	Low RU	0		
011990000819	fx5	5	4:39:49 PM	20	0.04 kUA/l	0	Negative	
011990000819	fx1	5	4:40:49 PM	16	Low RU	0	Negative	
011990000819	fx2	5	4:41:49 PM	14	Low RU	0	Negative	
011990000819	fx22	5	4:42:49 PM	31	0.16 kUA/l	0	Negative	
011990000819	gx2	5	4:43:49 PM	24	0.08 kUA/l	0	Negative	
011990000819	wx1	5	4:44:49 PM	55	0.45 kUA/l	1	Positive	
011990000819	t20	5	4:45:49 PM	15	Low RU	0		
011990000820	fx5	5	4:15:50 PM	32	0.18 kUA/l	0	Negative	
011990000820	fx1	5	4:16:50 PM	23	0.08 kUA/l	0	Negative	
011990000820	fx2	5	4:17:50 PM	19	0.03 kUA/l	0	Negative	
011990000820	fx22	5	4:18:50 PM	40	0.28 kUA/l	0	Negative	
011990000820	gx2	5	4:19:50 PM	117	1.17 kUA/l	2	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000820	wx1	5	4:20:50 PM	111	1.10 kUA/l	2	Positive	
011990000820	t20	5	4:21:50 PM	122	1.23 kUA/l	2		
011990000821	fx5	5	5:19:48 PM	21	0.05 kUA/l	0	Negative	
011990000821	fx1	5	5:20:48 PM	17	Low RU	0	Negative	
011990000821	fx2	5	5:21:47 PM	28	0.13 kUA/l	0	Negative	
011990000821	fx22	5	5:22:47 PM	30	0.16 kUA/l	0	Negative	
011990000821	gx2	5	5:23:47 PM	19	0.03 kUA/l	0	Negative	
011990000821	wx1	5	5:24:47 PM	33	0.20 kUA/l	0	Negative	
011990000821	t20	5	5:25:47 PM	18	0.02 kUA/l	0		
011990000822	fx5	5	4:47:49 PM	32	0.18 kUA/l	0	Negative	
011990000822	fx1	5	4:48:49 PM	17	Low RU	0	Negative	
011990000822	fx2	5	4:49:49 PM	56	0.46 kUA/l	1	Positive	
011990000822	fx22	5	4:50:49 PM	29	0.14 kUA/l	0	Negative	
011990000822	gx2	5	4:51:49 PM	22	0.06 kUA/l	0	Negative	
011990000822	wx1	5	4:52:49 PM	28	0.14 kUA/l	0	Negative	
011990000822	t20	5	4:53:48 PM	17	Low RU	0		
011990000823	fx5	5	4:31:50 PM	88	0.83 kUA/l	2	Positive	
011990000823	fx1	5	4:32:49 PM	20	0.04 kUA/l	0	Negative	
011990000823	fx2	5	4:33:49 PM	20	0.04 kUA/l	0	Negative	
011990000823	fx22	5	4:34:49 PM	31	0.17 kUA/l	0	Negative	
011990000823	gx2	5	4:35:49 PM	24	0.09 kUA/l	0	Negative	
011990000823	wx1	5	4:36:49 PM	30	0.16 kUA/l	0	Negative	
011990000823	t20	5	4:37:49 PM	19	0.02 kUA/l	0		
011990000824	fx5	5	4:23:50 PM	25	0.10 kUA/l	0	Negative	
011990000824	fx1	5	4:24:50 PM	18	0.01 kUA/l	0	Negative	
011990000824	fx2	5	4:25:49 PM	90	0.86 kUA/l	2	Positive	
011990000824	fx22	5	4:26:50 PM	40	0.27 kUA/l	0	Negative	
011990000824	gx2	5	4:27:49 PM	30	0.16 kUA/l	0	Negative	
011990000824	wx1	5	4:28:49 PM	53	0.43 kUA/l	1	Positive	
011990000824	t20	5	4:29:49 PM	43	0.30 kUA/l	0		
011990000825	fx5	5	5:27:48 PM	28	0.13 kUA/l	0	Negative	
011990000825	fx1	5	5:28:47 PM	33	0.19 kUA/l	0	Negative	
011990000825	fx2	5	5:29:47 PM	36	0.22 kUA/l	0	Negative	
011990000825	fx22	5	5:30:47 PM	28	0.13 kUA/l	0	Negative	
011990000825	gx2	5	5:31:47 PM	22	0.06 kUA/l	0	Negative	
011990000825	wx1	5	5:32:47 PM	30	0.15 kUA/l	0	Negative	
011990000825	t20	5	5:33:47 PM	18	0.01 kUA/l	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
-----------------	------	----------	----------	-----------	---------	----------	----------	-----------	---------	---------

# Laboratory Report

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	3:31:51 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000805	phad	1:57:55 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000805	fx5	1:58:55 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000805	fx1	1:59:55 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000805	fx2	2:00:55 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000805	fx22	2:01:55 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000805	gx2	2:02:55 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000805	wx1	2:03:55 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000805	t20	2:04:54 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	phad	2:05:55 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	fx5	2:06:54 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	fx1	2:07:54 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	fx2	2:08:54 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	fx22	2:09:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	gx2	2:10:55 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	wx1	2:11:55 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000804	t20	2:12:54 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	phad	2:13:54 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	fx5	2:14:54 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	fx1	2:15:54 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	fx2	2:16:54 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	fx22	2:17:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	gx2	2:18:54 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	wx1	2:19:54 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000803	t20	2:20:54 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	phad	2:21:54 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	fx5	2:22:54 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	fx1	2:23:54 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	fx2	2:24:54 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	fx22	2:25:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	gx2	2:26:54 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	wx1	2:27:54 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000802	t20	2:28:54 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	phad	2:29:54 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	fx5	2:30:54 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	fx1	2:31:54 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	fx2	2:32:54 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	fx22	2:33:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	gx2	2:34:54 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	wx1	2:35:53 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000801	t20	2:36:53 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	phad	2:37:53 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	fx5	2:38:54 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	fx1	2:39:53 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	fx2	2:40:53 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	fx22	2:41:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	gx2	2:42:53 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	wx1	2:43:53 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000810	t20	2:44:53 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	phad	2:45:53 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	fx5	2:46:53 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	fx1	2:47:53 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	fx2	2:48:53 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	fx22	2:49:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	gx2	2:50:53 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	wx1	2:51:53 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000809	t20	2:52:53 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000808	phad	2:53:53 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000808	fx5	2:54:53 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000808	fx1	2:55:53 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000808	fx2	2:56:52 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000808	fx22	2:57:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000808	gx2	2:58:53 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000808	wx1	2:59:53 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000808	t20	3:00:53 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	phad	3:01:53 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	fx5	3:02:52 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	fx1	3:03:52 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	fx2	3:04:52 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	fx22	3:05:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	gx2	3:06:53 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	wx1	3:07:52 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000807	t20	3:08:52 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	phad	3:09:52 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	fx5	3:10:52 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	fx1	3:11:52 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	fx2	3:12:52 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	fx22	3:13:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	gx2	3:14:52 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	wx1	3:32:51 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000806	t20	3:33:51 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	phad	3:58:50 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	fx5	3:59:50 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	fx1	4:00:50 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	fx2	4:01:50 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	fx22	4:02:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	gx2	4:03:51 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	wx1	4:04:50 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000815	t20	4:05:50 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	phad	4:06:50 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	fx5	4:07:50 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	fx1	4:08:50 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	fx2	4:09:50 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	fx22	4:10:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	gx2	4:11:50 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	wx1	4:12:50 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000814	t20	4:13:50 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	phad	5:02:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	fx5	5:03:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	fx1	5:04:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	fx2	5:05:48 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	fx22	5:06:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	gx2	5:07:48 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	wx1	5:08:48 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000813	t20	5:09:48 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	phad	3:42:51 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	fx5	3:43:51 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	fx1	3:44:51 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	fx2	3:45:51 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	fx22	3:46:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	gx2	3:47:51 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	wx1	3:48:51 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000812	t20	3:49:51 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	phad	3:34:52 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	fx5	3:35:52 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	fx1	3:36:51 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	fx2	3:37:51 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	fx22	3:38:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	gx2	3:39:51 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	wx1	3:40:51 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000811	t20	3:41:51 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000820	phad	4:14:50 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000820	fx5	4:15:50 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000820	fx1	4:16:50 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000820	fx2	4:17:50 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000820	fx22	4:18:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000820	gx2	4:19:50 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000820	wx1	4:20:50 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	



# Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000820	t20	4:21:50 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	phad	4:38:49 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	fx5	4:39:49 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	fx1	4:40:49 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	fx2	4:41:49 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	fx22	4:42:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	gx2	4:43:49 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	wx1	4:44:49 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000819	t20	4:45:49 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	phad	4:54:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	fx5	4:55:49 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	fx1	4:56:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	fx2	4:57:48 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	fx22	4:58:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	gx2	4:59:49 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	wx1	5:00:48 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000818	t20	5:01:48 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	phad	5:10:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	fx5	5:11:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	fx1	5:12:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	fx2	5:13:48 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	fx22	5:14:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	gx2	5:15:48 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	wx1	5:16:48 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000817	t20	5:17:48 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	phad	3:50:51 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	fx5	3:51:51 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	fx1	3:52:51 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	fx2	3:53:51 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	fx22	3:54:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	gx2	3:55:51 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	wx1	3:56:50 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000816	t20	3:57:50 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	phad	5:26:47 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	fx5	5:27:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	fx1	5:28:47 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	fx2	5:29:47 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	fx22	5:30:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	gx2	5:31:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	wx1	5:32:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000825	t20	5:33:47 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	phad	4:22:50 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	fx5	4:23:50 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	fx1	4:24:50 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	fx2	4:25:49 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	fx22	4:26:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	gx2	4:27:49 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	wx1	4:28:49 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000824	t20	4:29:49 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	phad	4:30:50 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	fx5	4:31:50 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	fx1	4:32:49 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	fx2	4:33:49 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	fx22	4:34:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	gx2	4:35:49 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	wx1	4:36:49 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000823	t20	4:37:49 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	phad	4:46:49 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	fx5	4:47:49 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	fx1	4:48:49 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	fx2	4:49:49 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	fx22	4:50:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	gx2	4:51:49 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	wx1	4:52:49 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000822	t20	4:53:48 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000821	phad	5:18:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000821	fx5	5:19:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000821	fx1	5:20:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000821	fx2	5:21:47 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000821	fx22	5:22:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000821	gx2	5:23:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000821	wx1	5:24:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000821	t20	5:25:47 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	




## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200128-388  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 28-Jan-20 5:34:47 PM  
Ended: 28-Jan-20 10:01:58 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

No curve controls in this run

### Quality Controls

No quality controls in this run

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000645	phad	5	5:34:47 PM	157	1.63 PAU/l	2	Positive	
011990000650	phad	5	5:42:47 PM	311	3.41 PAU/l	2	Positive	
011990000851	phad	5	6:22:45 PM	132	1.35 PAU/l	2	Positive	
011990000852	phad	5	6:14:46 PM	626	7.10 PAU/l	3	Positive	
011990000853	phad	5	6:06:46 PM	49	0.38 PAU/l	1	Positive	
011990000854	phad	5	5:58:46 PM	31	0.16 PAU/l	0	Negative	
011990000855	phad	5	5:50:46 PM	144	1.48 PAU/l	2	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000856	phad	5	6:30:45 PM	815	9.32 PAU/I	3	Positive	
011990000857	phad	5	6:38:44 PM	150	1.55 PAU/I	2	Positive	
011990000859	phad	5	6:46:44 PM	102	0.99 PAU/I	2	Positive	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000645	fx5	5	5:35:47 PM	22	0.06 kUA/I	0	Negative	
011990000645	fx1	5	5:36:47 PM	25	0.09 kUA/I	0	Negative	
011990000645	fx2	5	5:37:47 PM	15	Low RU	0	Negative	
011990000645	fx22	5	5:38:47 PM	34	0.21 kUA/I	0	Negative	
011990000645	gx2	5	5:39:47 PM	22	0.07 kUA/I	0	Negative	
011990000645	wx1	5	5:40:47 PM	28	0.14 kUA/I	0	Negative	
011990000645	t20	5	5:41:47 PM	17	Low RU	0		
011990000650	fx5	5	5:43:47 PM	26	0.11 kUA/I	0	Negative	
011990000650	fx1	5	5:44:46 PM	22	0.07 kUA/I	0	Negative	
011990000650	fx2	5	5:45:47 PM	240	2.59 kUA/I	2	Positive	
011990000650	fx22	5	5:46:47 PM	35	0.22 kUA/I	0	Negative	
011990000650	gx2	5	5:47:47 PM	43	0.31 kUA/I	0	Negative	
011990000650	wx1	5	5:48:46 PM	91	0.87 kUA/I	2	Positive	
011990000650	t20	5	5:49:46 PM	112	1.11 kUA/I	2		
011990000851	fx5	5	6:23:45 PM	21	0.05 kUA/I	0	Negative	
011990000851	fx1	5	6:24:45 PM	18	0.01 kUA/I	0	Negative	
011990000851	fx2	5	6:25:45 PM	18	0.02 kUA/I	0	Negative	
011990000851	fx22	5	6:26:45 PM	29	0.15 kUA/I	0	Negative	
011990000851	gx2	5	6:27:45 PM	24	0.08 kUA/I	0	Negative	
011990000851	wx1	5	6:28:45 PM	30	0.15 kUA/I	0	Negative	
011990000851	t20	5	6:29:45 PM	18	0.02 kUA/I	0		
011990000852	fx5	5	6:15:45 PM	154	1.60 kUA/I	2	Positive	
011990000852	fx1	5	6:16:45 PM	96	0.93 kUA/I	2	Positive	
011990000852	fx2	5	6:17:45 PM	37	0.24 kUA/I	0	Negative	
011990000852	fx22	5	6:18:45 PM	92	0.88 kUA/I	2	Positive	
011990000852	gx2	5	6:19:45 PM	49	0.38 kUA/I	1	Positive	
011990000852	wx1	5	6:20:45 PM	326	3.59 kUA/I	3	Positive	
011990000852	t20	5	6:21:45 PM	399	4.44 kUA/I	3		
011990000853	fx5	5	6:07:46 PM	28	0.13 kUA/I	0	Negative	
011990000853	fx1	5	6:08:46 PM	30	0.15 kUA/I	0	Negative	
011990000853	fx2	5	6:09:45 PM	18	0.01 kUA/I	0	Negative	
011990000853	fx22	5	6:10:45 PM	45	0.34 kUA/I	0	Negative	
011990000853	gx2	5	6:11:45 PM	23	0.07 kUA/I	0	Negative	
011990000853	wx1	5	6:12:45 PM	123	1.24 kUA/I	2	Positive	
011990000853	t20	5	6:13:45 PM	105	1.03 kUA/I	2		
011990000854	fx5	5	5:59:46 PM	22	0.06 kUA/I	0	Negative	
011990000854	fx1	5	6:00:46 PM	22	0.06 kUA/I	0	Negative	
011990000854	fx2	5	6:01:46 PM	16	Low RU	0	Negative	
011990000854	fx22	5	6:02:46 PM	32	0.18 kUA/I	0	Negative	
011990000854	gx2	5	6:03:46 PM	24	0.08 kUA/I	0	Negative	
011990000854	wx1	5	6:04:46 PM	28	0.14 kUA/I	0	Negative	
011990000854	t20	5	6:05:46 PM	17	Low RU	0		
011990000855	fx5	5	5:51:46 PM	139	1.42 kUA/I	2	Positive	
011990000855	fx1	5	5:52:46 PM	139	1.43 kUA/I	2	Positive	
011990000855	fx2	5	5:53:46 PM	69	0.62 kUA/I	1	Positive	
011990000855	fx22	5	5:54:46 PM	129	1.30 kUA/I	2	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000855	gx2	5	5:55:46 PM	136	1.39 kUA/l	2	Positive	
011990000855	wx1	5	5:56:46 PM	136	1.39 kUA/l	2	Positive	
011990000855	t20	5	5:57:46 PM	119	1.19 kUA/l	2		
011990000856	fx5	5	6:31:45 PM	50	0.39 kUA/l	1	Positive	
011990000856	fx1	5	6:32:45 PM	30	0.16 kUA/l	0	Negative	
011990000856	fx2	5	6:33:45 PM	653	7.41 kUA/l	3	Positive	
011990000856	fx22	5	6:34:45 PM	40	0.28 kUA/l	0	Negative	
011990000856	gx2	5	6:35:45 PM	60	0.50 kUA/l	1	Positive	
011990000856	wx1	5	6:36:45 PM	117	1.17 kUA/l	2	Positive	
011990000856	t20	5	6:37:44 PM	68	0.60 kUA/l	1		
011990000857	fx5	5	6:39:44 PM	26	0.11 kUA/l	0	Negative	
011990000857	fx1	5	6:40:44 PM	31	0.16 kUA/l	0	Negative	
011990000857	fx2	5	6:41:45 PM	17	Low RU	0	Negative	
011990000857	fx22	5	6:42:44 PM	35	0.22 kUA/l	0	Negative	
011990000857	gx2	5	6:43:45 PM	311	3.42 kUA/l	2	Positive	
011990000857	wx1	5	6:44:44 PM	29	0.14 kUA/l	0	Negative	
011990000857	t20	5	6:45:44 PM	17	0.01 kUA/l	0		
011990000859	fx5	5	6:47:44 PM	26	0.11 kUA/l	0	Negative	
011990000859	fx1	5	6:48:44 PM	20	0.04 kUA/l	0	Negative	
011990000859	fx2	5	6:49:44 PM	18	0.01 kUA/l	0	Negative	
011990000859	fx22	5	6:50:44 PM	28	0.14 kUA/l	0	Negative	
011990000859	gx2	5	6:51:44 PM	35	0.22 kUA/l	0	Negative	
011990000859	wx1	5	6:52:44 PM	76	0.70 kUA/l	1	Positive	
011990000859	t20	5	6:53:44 PM	19	0.02 kUA/l	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000855	phad	5:50:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000855	fx5	5:51:46 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000855	fx1	5:52:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000855	fx2	5:53:46 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000855	fx22	5:54:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000855	gx2	5:55:46 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000855	wx1	5:56:46 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000855	t20	5:57:46 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	phad	5:58:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	fx5	5:59:46 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	fx1	6:00:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	fx2	6:01:46 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	fx22	6:02:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	gx2	6:03:46 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	wx1	6:04:46 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000854	t20	6:05:46 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000853	phad	6:06:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000853	fx5	6:07:46 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000853	fx1	6:08:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000853	fx2	6:09:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000853	fx22	6:10:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000853	gx2	6:11:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000853	wx1	6:12:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000853	t20	6:13:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	phad	6:14:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	fx5	6:15:45 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	fx1	6:16:45 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	fx2	6:17:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	fx22	6:18:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	gx2	6:19:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	wx1	6:20:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000852	t20	6:21:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	phad	6:22:45 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	fx5	6:23:45 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	fx1	6:24:45 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	fx2	6:25:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	fx22	6:26:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	gx2	6:27:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	wx1	6:28:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000851	t20	6:29:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	phad	6:46:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	fx5	6:47:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	fx1	6:48:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	fx2	6:49:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	fx22	6:50:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	gx2	6:51:44 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	wx1	6:52:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000859	t20	6:53:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	phad	6:38:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	fx5	6:39:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	fx1	6:40:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	fx2	6:41:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	fx22	6:42:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	gx2	6:43:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	wx1	6:44:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000857	t20	6:45:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	phad	6:30:45 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	fx5	6:31:45 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	fx1	6:32:45 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	fx2	6:33:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	fx22	6:34:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	gx2	6:35:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	wx1	6:36:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000856	t20	6:37:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	phad	5:42:47 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	fx5	5:43:47 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	fx1	5:44:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	fx2	5:45:47 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	fx22	5:46:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	gx2	5:47:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	wx1	5:48:46 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000650	t20	5:49:46 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	phad	5:34:47 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	fx5	5:35:47 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	fx1	5:36:47 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	fx2	5:37:47 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	fx22	5:38:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	gx2	5:39:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	wx1	5:40:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000645	t20	5:41:47 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	