


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200113-377
Instrument name:	Phadia-A
Instrument type:	Phadia 250
Status:	Ready
Started:	13-Jan-20 12:54:30 PM
Ended:	13-Jan-20 5:14:22 PM
Approved:	Approved (faziz)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-A
Instrument type:	Phadia 250
Measured:	05-Jan-20 1:26:01 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	12:54:30 PM	305	0.67	OK	
CC-2	17.5	12:55:30 PM	6324	17.6	OK	

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	1:00:30 PM	3653	9.21 kUA/I	CZ1AR	OK	
slgE L	e1	1	1:01:30 PM	1123	2.60 kUA/I	CZ4AN	OK	
slgE Neg	d1	1	1:02:29 PM	24	0.02 kUA/I	BL1CP	OK	
IQC271119	d1	1	1:06:29 PM	6954	19.8 kUA/I	<lab>	----	
			Expected value:	----				
			Expected SD:	----				
			SD limit:	----				
			Range min:	----				
			Range max:	----				
			Limit CV%:	----				

Laboratory Report

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000101	phad	5	2:32:26 PM	32	0.18 PAU/I	---	Negative	
011990000105	phad	5	5:06:21 PM	153	1.59 PAU/I	---	Positive	
011990000107	phad	5	4:50:21 PM	33	0.19 PAU/I	---	Negative	
011990000108	phad	5	4:58:21 PM	127	1.28 PAU/I	---	Positive	
011990000109	phad	5	3:53:23 PM	144	1.48 PAU/I	---	Positive	
011990000110	phad	5	4:42:22 PM	38	0.24 PAU/I	---	Negative	
011990000111	phad	5	1:36:28 PM	127	1.28 PAU/I	---	Positive	
011990000301	phad	5	3:29:24 PM	203	2.16 PAU/I	---	Positive	
011990000302	phad	5	3:21:24 PM	271	2.96 PAU/I	---	Positive	
011990000303	phad	5	2:16:27 PM	33	0.20 PAU/I	---	Negative	
011990000304	phad	5	2:08:27 PM	205	2.19 PAU/I	---	Positive	
011990000311	phad	5	3:13:25 PM	121	1.21 PAU/I	---	Positive	
011990000312	phad	5	4:05:23 PM	46	0.35 PAU/I	---	Negative	
011990000313	phad	5	2:00:28 PM	247	2.67 PAU/I	---	Positive	
011990000314	phad	5	3:45:24 PM	35	0.21 PAU/I	---	Negative	
011990000321	phad	5	2:48:26 PM	37	0.24 PAU/I	---	Negative	
011990000322	phad	5	3:37:24 PM	36	0.22 PAU/I	---	Negative	
011990000323	phad	5	1:52:28 PM	39	0.26 PAU/I	---	Negative	
011990000324	phad	5	2:24:27 PM	51	0.41 PAU/I	---	Positive	
011990000331	phad	5	2:56:26 PM	35	0.21 PAU/I	---	Negative	
011990000332	phad	5	1:20:29 PM	363	4.02 PAU/I	---	Positive	
011990000333	phad	5	1:44:28 PM	85	0.79 PAU/I	---	Positive	
011990000341	phad	5	2:40:26 PM	2073	24.8 PAU/I	---	Positive	
011990000342	phad	5	1:12:30 PM	419	4.67 PAU/I	---	Positive	
011990000343	phad	5	1:28:29 PM	35	0.22 PAU/I	---	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000101	fx5	5	2:33:26 PM	20	0.04 kUA/I	---	Negative	
011990000101	fx1	5	2:34:26 PM	17	---	---	Negative	
011990000101	fx2	5	2:35:26 PM	19	0.03 kUA/I	---	Negative	
011990000101	fx22	5	2:36:26 PM	30	0.16 kUA/I	---	Negative	
011990000101	gx2	5	2:37:26 PM	22	0.06 kUA/I	---	Negative	
011990000101	wx1	5	2:38:26 PM	29	0.14 kUA/I	---	Negative	
011990000101	t20	5	2:39:26 PM	19	0.03 kUA/I	---		
011990000105	fx5	5	5:07:21 PM	28	0.13 kUA/I	---	Negative	
011990000105	fx1	5	5:08:21 PM	23	0.07 kUA/I	---	Negative	
011990000105	fx2	5	5:09:21 PM	14	---	---	Negative	
011990000105	fx22	5	5:10:21 PM	31	0.17 kUA/I	---	Negative	
011990000105	gx2	5	5:11:20 PM	23	0.07 kUA/I	---	Negative	
011990000105	wx1	5	5:12:20 PM	56	0.46 kUA/I	---	Positive	
011990000105	t20	5	5:13:20 PM	50	0.39 kUA/I	---		
011990000107	fx5	5	4:51:21 PM	25	0.09 kUA/I	---	Negative	
011990000107	fx1	5	4:52:21 PM	18	0.01 kUA/I	---	Negative	
011990000107	fx2	5	4:53:21 PM	12	---	---	Negative	
011990000107	fx22	5	4:54:21 PM	32	0.18 kUA/I	---	Negative	
011990000107	gx2	5	4:55:22 PM	26	0.11 kUA/I	---	Negative	
011990000107	wx1	5	4:56:21 PM	30	0.16 kUA/I	---	Negative	
011990000107	t20	5	4:57:21 PM	18	0.01 kUA/I	---		
011990000108	fx5	5	4:59:21 PM	28	0.13 kUA/I	---	Negative	
011990000108	fx1	5	5:00:21 PM	19	0.02 kUA/I	---	Negative	
011990000108	fx2	5	5:01:21 PM	330	3.64 kUA/I	---	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000108	fx22	5	5:02:21 PM	37	0.24 kUA/l	---	Negative	
011990000108	gx2	5	5:03:21 PM	29	0.14 kUA/l	---	Negative	
011990000108	wx1	5	5:04:21 PM	37	0.24 kUA/l	---	Negative	
011990000108	t20	5	5:05:21 PM	20	0.04 kUA/l	---		
011990000109	fx5	5	3:54:23 PM	31	0.17 kUA/l	---	Negative	
011990000109	fx1	5	3:55:23 PM	20	0.04 kUA/l	---	Negative	
011990000109	fx2	5	3:56:23 PM	79	0.73 kUA/l	---	Positive	
011990000109	fx22	5	3:57:23 PM	33	0.19 kUA/l	---	Negative	
011990000109	gx2	5	3:58:23 PM	35	0.21 kUA/l	---	Negative	
011990000109	wx1	5	3:59:23 PM	48	0.36 kUA/l	---	Positive	
011990000109	t20	5	4:00:23 PM	49	0.38 kUA/l	---		
011990000110	fx5	5	4:43:21 PM	26	0.11 kUA/l	---	Negative	
011990000110	fx1	5	4:44:21 PM	24	0.09 kUA/l	---	Negative	
011990000110	fx2	5	4:45:21 PM	23	0.08 kUA/l	---	Negative	
011990000110	fx22	5	4:46:22 PM	32	0.18 kUA/l	---	Negative	
011990000110	gx2	5	4:47:22 PM	22	0.06 kUA/l	---	Negative	
011990000110	wx1	5	4:48:21 PM	28	0.14 kUA/l	---	Negative	
011990000110	t20	5	4:49:21 PM	19	0.03 kUA/l	---		
011990000111	fx5	5	1:37:28 PM	31	0.17 kUA/l	---	Negative	
011990000111	fx1	5	1:38:28 PM	20	0.04 kUA/l	---	Negative	
011990000111	fx2	5	1:39:28 PM	86	0.82 kUA/l	---	Positive	
011990000111	fx22	5	1:40:29 PM	30	0.15 kUA/l	---	Negative	
011990000111	gx2	5	1:41:28 PM	23	0.07 kUA/l	---	Negative	
011990000111	wx1	5	1:42:28 PM	54	0.44 kUA/l	---	Positive	
011990000111	t20	5	1:43:28 PM	25	0.10 kUA/l	---		
011990000301	fx5	5	3:30:24 PM	32	0.18 kUA/l	---	Negative	
011990000301	fx1	5	3:31:24 PM	25	0.10 kUA/l	---	Negative	
011990000301	fx2	5	3:32:24 PM	92	0.88 kUA/l	---	Positive	
011990000301	fx22	5	3:33:24 PM	33	0.19 kUA/l	---	Negative	
011990000301	gx2	5	3:34:24 PM	32	0.17 kUA/l	---	Negative	
011990000301	wx1	5	3:35:24 PM	70	0.62 kUA/l	---	Positive	
011990000301	t20	5	3:36:24 PM	77	0.70 kUA/l	---		
011990000302	fx5	5	3:22:24 PM	19	0.03 kUA/l	---	Negative	
011990000302	fx1	5	3:23:25 PM	19	0.03 kUA/l	---	Negative	
011990000302	fx2	5	3:24:24 PM	15	---	---	Negative	
011990000302	fx22	5	3:25:24 PM	32	0.18 kUA/l	---	Negative	
011990000302	gx2	5	3:26:24 PM	24	0.09 kUA/l	---	Negative	
011990000302	wx1	5	3:27:24 PM	28	0.13 kUA/l	---	Negative	
011990000302	t20	5	3:28:24 PM	18	0.02 kUA/l	---		
011990000303	fx5	5	2:17:27 PM	23	0.07 kUA/l	---	Negative	
011990000303	fx1	5	2:18:27 PM	29	0.14 kUA/l	---	Negative	
011990000303	fx2	5	2:19:27 PM	16	---	---	Negative	
011990000303	fx22	5	2:20:27 PM	27	0.13 kUA/l	---	Negative	
011990000303	gx2	5	2:21:27 PM	22	0.06 kUA/l	---	Negative	
011990000303	wx1	5	2:22:27 PM	29	0.15 kUA/l	---	Negative	
011990000303	t20	5	2:23:27 PM	19	0.03 kUA/l	---		
011990000304	fx5	5	2:09:27 PM	157	1.63 kUA/l	---	Positive	
011990000304	fx1	5	2:10:27 PM	183	1.94 kUA/l	---	Positive	
011990000304	fx2	5	2:11:27 PM	84	0.79 kUA/l	---	Positive	
011990000304	fx22	5	2:12:27 PM	184	1.94 kUA/l	---	Positive	
011990000304	gx2	5	2:13:27 PM	193	2.04 kUA/l	---	Positive	
011990000304	wx1	5	2:14:27 PM	285	3.11 kUA/l	---	Positive	
011990000304	t20	5	2:15:27 PM	180	1.90 kUA/l	---		
011990000311	fx5	5	3:14:25 PM	29	0.15 kUA/l	---	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000311	fx1	5	3:15:25 PM	22	0.06 kUA/l	---	Negative	
011990000311	fx2	5	3:16:24 PM	60	0.51 kUA/l	---	Positive	
011990000311	fx22	5	3:17:25 PM	33	0.19 kUA/l	---	Negative	
011990000311	gx2	5	3:18:24 PM	30	0.16 kUA/l	---	Negative	
011990000311	wx1	5	3:19:24 PM	263	2.86 kUA/l	---	Positive	
011990000311	t20	5	3:20:24 PM	41	0.28 kUA/l	---		
011990000312	fx5	5	4:06:23 PM	31	0.17 kUA/l	---	Negative	
011990000312	fx1	5	4:07:23 PM	26	0.11 kUA/l	---	Negative	
011990000312	fx2	5	4:08:22 PM	19	0.03 kUA/l	---	Negative	
011990000312	fx22	5	4:39:22 PM	41	0.29 kUA/l	---	Negative	
011990000312	gx2	5	4:09:23 PM	41	0.28 kUA/l	---	Negative	
011990000312	wx1	5	4:40:22 PM	251	2.72 kUA/l	---	Positive	
011990000312	t20	5	4:41:22 PM	31	0.17 kUA/l	---		
011990000313	fx5	5	2:01:28 PM	22	0.06 kUA/l	---	Negative	
011990000313	fx1	5	2:02:28 PM	19	0.03 kUA/l	---	Negative	
011990000313	fx2	5	2:03:27 PM	13	---	---	Negative	
011990000313	fx22	5	2:04:27 PM	28	0.13 kUA/l	---	Negative	
011990000313	gx2	5	2:05:27 PM	23	0.08 kUA/l	---	Negative	
011990000313	wx1	5	2:06:27 PM	37	0.24 kUA/l	---	Negative	
011990000313	t20	5	2:07:27 PM	17	---	---		
011990000314	fx5	5	3:46:23 PM	22	0.06 kUA/l	---	Negative	
011990000314	fx1	5	3:47:23 PM	18	0.02 kUA/l	---	Negative	
011990000314	fx2	5	3:48:23 PM	15	---	---	Negative	
011990000314	fx22	5	3:49:23 PM	33	0.19 kUA/l	---	Negative	
011990000314	gx2	5	3:50:23 PM	25	0.09 kUA/l	---	Negative	
011990000314	wx1	5	3:51:24 PM	31	0.17 kUA/l	---	Negative	
011990000314	t20	5	3:52:24 PM	19	0.03 kUA/l	---		
011990000321	fx5	5	2:49:26 PM	38	0.25 kUA/l	---	Negative	
011990000321	fx1	5	2:50:26 PM	24	0.09 kUA/l	---	Negative	
011990000321	fx2	5	2:51:26 PM	14	---	---	Negative	
011990000321	fx22	5	2:52:25 PM	39	0.26 kUA/l	---	Negative	
011990000321	gx2	5	2:53:25 PM	19	0.02 kUA/l	---	Negative	
011990000321	wx1	5	2:54:25 PM	43	0.31 kUA/l	---	Negative	
011990000321	t20	5	2:55:26 PM	22	0.06 kUA/l	---		
011990000322	fx5	5	3:38:24 PM	22	0.07 kUA/l	---	Negative	
011990000322	fx1	5	3:39:24 PM	19	0.03 kUA/l	---	Negative	
011990000322	fx2	5	3:40:23 PM	16	---	---	Negative	
011990000322	fx22	5	3:41:24 PM	36	0.22 kUA/l	---	Negative	
011990000322	gx2	5	3:42:24 PM	37	0.24 kUA/l	---	Negative	
011990000322	wx1	5	3:43:24 PM	31	0.17 kUA/l	---	Negative	
011990000322	t20	5	3:44:23 PM	17	---	---		
011990000323	fx5	5	1:53:28 PM	21	0.04 kUA/l	---	Negative	
011990000323	fx1	5	1:54:28 PM	18	0.02 kUA/l	---	Negative	
011990000323	fx2	5	1:55:28 PM	21	0.05 kUA/l	---	Negative	
011990000323	fx22	5	1:56:28 PM	29	0.15 kUA/l	---	Negative	
011990000323	gx2	5	1:57:28 PM	19	0.03 kUA/l	---	Negative	
011990000323	wx1	5	1:58:28 PM	35	0.21 kUA/l	---	Negative	
011990000323	t20	5	1:59:28 PM	20	0.04 kUA/l	---		
011990000324	fx5	5	2:25:27 PM	19	0.03 kUA/l	---	Negative	
011990000324	fx1	5	2:26:27 PM	16	---	---	Negative	
011990000324	fx2	5	2:27:27 PM	21	0.05 kUA/l	---	Negative	
011990000324	fx22	5	2:28:27 PM	30	0.16 kUA/l	---	Negative	
011990000324	gx2	5	2:29:27 PM	26	0.11 kUA/l	---	Negative	
011990000324	wx1	5	2:30:26 PM	185	1.95 kUA/l	---	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000324	t20	5	2:31:26 PM	18	0.02 kUA/I	---		
011990000331	fx5	5	2:57:25 PM	20	0.04 kUA/I	---	Negative	
011990000331	fx1	5	2:58:25 PM	18	0.01 kUA/I	---	Negative	
011990000331	fx2	5	2:59:25 PM	17	---	---	Negative	
011990000331	fx22	5	3:00:25 PM	33	0.19 kUA/I	---	Negative	
011990000331	gx2	5	3:01:25 PM	27	0.12 kUA/I	---	Negative	
011990000331	wx1	5	3:02:25 PM	30	0.16 kUA/I	---	Negative	
011990000331	t20	5	3:03:25 PM	20	0.04 kUA/I	---		
011990000332	fx5	5	1:21:29 PM	27	0.13 kUA/I	---	Negative	
011990000332	fx1	5	1:22:29 PM	19	0.03 kUA/I	---	Negative	
011990000332	fx2	5	1:23:29 PM	225	2.42 kUA/I	---	Positive	
011990000332	fx22	5	1:24:29 PM	33	0.19 kUA/I	---	Negative	
011990000332	gx2	5	1:25:29 PM	28	0.14 kUA/I	---	Negative	
011990000332	wx1	5	1:26:29 PM	74	0.67 kUA/I	---	Positive	
011990000332	t20	5	1:27:29 PM	40	0.27 kUA/I	---		
011990000333	fx5	5	1:45:28 PM	21	0.05 kUA/I	---	Negative	
011990000333	fx1	5	1:46:28 PM	73	0.66 kUA/I	---	Positive	
011990000333	fx2	5	1:47:28 PM	78	0.72 kUA/I	---	Positive	
011990000333	fx22	5	1:48:28 PM	29	0.15 kUA/I	---	Negative	
011990000333	gx2	5	1:49:28 PM	25	0.10 kUA/I	---	Negative	
011990000333	wx1	5	1:50:28 PM	251	2.73 kUA/I	---	Positive	
011990000333	t20	5	1:51:28 PM	24	0.08 kUA/I	---		
011990000341	fx5	5	2:41:26 PM	26	0.11 kUA/I	---	Negative	
011990000341	fx1	5	2:42:26 PM	21	0.05 kUA/I	---	Negative	
011990000341	fx2	5	2:43:26 PM	21	0.05 kUA/I	---	Negative	
011990000341	fx22	5	2:44:26 PM	37	0.24 kUA/I	---	Negative	
011990000341	gx2	5	2:45:26 PM	28	0.14 kUA/I	---	Negative	
011990000341	wx1	5	2:46:26 PM	952	11.0 kUA/I	---	Positive	
011990000341	t20	5	2:47:26 PM	70	0.62 kUA/I	---		
011990000342	fx5	5	1:13:29 PM	60	0.50 kUA/I	---	Positive	
011990000342	fx1	5	1:14:29 PM	82	0.76 kUA/I	---	Positive	
011990000342	fx2	5	1:15:29 PM	57	0.48 kUA/I	---	Positive	
011990000342	fx22	5	1:16:29 PM	78	0.72 kUA/I	---	Positive	
011990000342	gx2	5	1:17:29 PM	108	1.07 kUA/I	---	Positive	
011990000342	wx1	5	1:18:29 PM	331	3.65 kUA/I	---	Positive	
011990000342	t20	5	1:19:29 PM	47	0.36 kUA/I	---		
011990000343	fx5	5	1:29:29 PM	26	0.11 kUA/I	---	Negative	
011990000343	fx1	5	1:30:29 PM	22	0.06 kUA/I	---	Negative	
011990000343	fx2	5	1:31:29 PM	14	---	---	Negative	
011990000343	fx22	5	1:32:29 PM	34	0.21 kUA/I	---	Negative	
011990000343	gx2	5	1:33:29 PM	21	0.05 kUA/I	---	Negative	
011990000343	wx1	5	1:34:29 PM	32	0.18 kUA/I	---	Negative	
011990000343	t20	5	1:35:28 PM	19	0.03 kUA/I	---		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	12:54:30 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	12:55:30 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	1:00:30 PM	352D0 19/12/	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
slgE L	e1	1:01:30 PM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06/		Exp
slgE Neg	d1	1:02:29 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12/		Exp
IQC271119	d1	1:06:29 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000342	phad	1:12:30 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000342	fx5	1:13:29 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000342	fx1	1:14:29 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000342	fx2	1:15:29 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000342	fx22	1:16:29 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000342	gx2	1:17:29 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000342	wx1	1:18:29 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000342	t20	1:19:29 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	phad	1:20:29 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	fx5	1:21:29 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	fx1	1:22:29 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	fx2	1:23:29 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	fx22	1:24:29 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	gx2	1:25:29 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	wx1	1:26:29 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000332	t20	1:27:29 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	phad	1:28:29 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	fx5	1:29:29 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	fx1	1:30:29 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	fx2	1:31:29 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	fx22	1:32:29 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	gx2	1:33:29 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	wx1	1:34:29 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000343	t20	1:35:28 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	phad	1:44:28 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	fx5	1:45:28 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	fx1	1:46:28 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	fx2	1:47:28 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	fx22	1:48:28 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	gx2	1:49:28 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	wx1	1:50:28 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000333	t20	1:51:28 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	phad	1:52:28 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	fx5	1:53:28 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	fx1	1:54:28 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	fx2	1:55:28 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	fx22	1:56:28 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	gx2	1:57:28 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	wx1	1:58:28 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000323	t20	1:59:28 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000313	phad	2:00:28 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000313	fx5	2:01:28 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000313	fx1	2:02:28 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000313	fx2	2:03:27 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000313	fx22	2:04:27 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000313	gx2	2:05:27 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000313	wx1	2:06:27 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000313	t20	2:07:27 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	phad	2:08:27 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	fx5	2:09:27 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	fx1	2:10:27 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	fx2	2:11:27 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	fx22	2:12:27 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	gx2	2:13:27 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	wx1	2:14:27 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000304	t20	2:15:27 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	phad	2:16:27 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	fx5	2:17:27 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	fx1	2:18:27 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	fx2	2:19:27 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	fx22	2:20:27 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	gx2	2:21:27 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	wx1	2:22:27 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000303	t20	2:23:27 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	phad	2:24:27 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	fx5	2:25:27 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	fx1	2:26:27 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	fx2	2:27:27 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	fx22	2:28:27 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	gx2	2:29:27 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	wx1	2:30:26 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000324	t20	2:31:26 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	phad	3:45:24 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	fx5	3:46:23 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	fx1	3:47:23 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	fx2	3:48:23 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	fx22	3:49:23 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	gx2	3:50:23 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	wx1	3:51:24 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000314	t20	3:52:24 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	phad	1:36:28 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	fx5	1:37:28 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	fx1	1:38:28 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	fx2	1:39:28 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	fx22	1:40:29 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	gx2	1:41:28 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	wx1	1:42:28 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000111	t20	1:43:28 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	phad	2:32:26 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	fx5	2:33:26 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	fx1	2:34:26 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	fx2	2:35:26 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	fx22	2:36:26 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	gx2	2:37:26 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	wx1	2:38:26 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000101	t20	2:39:26 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	phad	2:40:26 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	fx5	2:41:26 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	fx1	2:42:26 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	fx2	2:43:26 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	fx22	2:44:26 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	gx2	2:45:26 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	wx1	2:46:26 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000341	t20	2:47:26 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	phad	2:56:26 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	fx5	2:57:25 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	fx1	2:58:25 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	fx2	2:59:25 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	fx22	3:00:25 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	gx2	3:01:25 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	wx1	3:02:25 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000331	t20	3:03:25 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000321	phad	2:48:26 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000321	fx5	2:49:26 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000321	fx1	2:50:26 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000321	fx2	2:51:26 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000321	fx22	2:52:25 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000321	gx2	2:53:25 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000321	wx1	2:54:25 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000321	t20	2:55:26 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	phad	3:13:25 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	fx5	3:14:25 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	fx1	3:15:25 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	fx2	3:16:24 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	fx22	3:17:25 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	gx2	3:18:24 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	wx1	3:19:24 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000311	t20	3:20:24 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	phad	3:21:24 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	fx5	3:22:24 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	fx1	3:23:25 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	fx2	3:24:24 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	fx22	3:25:24 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	gx2	3:26:24 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	wx1	3:27:24 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000302	t20	3:28:24 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	phad	3:29:24 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	fx5	3:30:24 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	fx1	3:31:24 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	fx2	3:32:24 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	fx22	3:33:24 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	gx2	3:34:24 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	wx1	3:35:24 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000301	t20	3:36:24 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	phad	3:37:24 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	fx5	3:38:24 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	fx1	3:39:24 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	fx2	3:40:23 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	fx22	3:41:24 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	gx2	3:42:24 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	wx1	3:43:24 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000322	t20	3:44:23 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	phad	3:53:23 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	fx5	3:54:23 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	fx1	3:55:23 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	fx2	3:56:23 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	fx22	3:57:23 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	gx2	3:58:23 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	wx1	3:59:23 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000109	t20	4:00:23 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	phad	4:05:23 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	fx5	4:06:23 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	fx1	4:07:23 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	fx2	4:08:22 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	fx22	4:39:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	gx2	4:09:23 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	wx1	4:40:22 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000312	t20	4:41:22 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	phad	4:42:22 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	fx5	4:43:21 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	fx1	4:44:21 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	fx2	4:45:21 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	fx22	4:46:22 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	gx2	4:47:22 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	wx1	4:48:21 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000110	t20	4:49:21 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000107	phad	4:50:21 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000107	fx5	4:51:21 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000107	fx1	4:52:21 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000107	fx2	4:53:21 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000107	fx22	4:54:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000107	gx2	4:55:22 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000107	wx1	4:56:21 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000107	t20	4:57:21 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	phad	4:58:21 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	fx5	4:59:21 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	fx1	5:00:21 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	fx2	5:01:21 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	fx22	5:02:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	gx2	5:03:21 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	wx1	5:04:21 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000108	t20	5:05:21 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	phad	5:06:21 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	fx5	5:07:21 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	fx1	5:08:21 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	fx2	5:09:21 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	fx22	5:10:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	gx2	5:11:20 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	wx1	5:12:20 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000105	t20	5:13:20 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Errors

Expired lot numbers used for Quality Control

352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200113-377
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 13-Jan-20 5:14:21 PM
Ended: 13-Jan-20 8:37:16 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	7:15:16 PM	6103	16.8 kUA/l	<lab>	----	
			Expected value:	---				
			Expected SD:	---				
			SD limit:	---				
			Range min:	---				
			Range max:	---				
			Limit CV%:	---				

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
----------	------	-----	----------	------	------	-------	---------	---------

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000102	phad	5	5:46:19 PM	44	0.32 PAU/I	---	Negative	
011990000103	phad	5	5:22:20 PM	136	1.40 PAU/I	---	Positive	
011990000104	phad	5	5:30:20 PM	34	0.21 PAU/I	---	Negative	
011990000106	phad	5	5:14:21 PM	62	0.53 PAU/I	---	Positive	
011990000112	phad	5	5:38:19 PM	34	0.20 PAU/I	---	Negative	
011990000113	phad	5	6:58:17 PM	33	0.19 PAU/I	---	Negative	
011990000114	phad	5	7:06:16 PM	34	0.21 PAU/I	---	Negative	
011990000115	phad	5	6:42:17 PM	33	0.19 PAU/I	---	Negative	
011990000116	phad	5	6:50:17 PM	1561	18.4 PAU/I	---	Positive	
011990000117	phad	5	6:10:18 PM	36	0.22 PAU/I	---	Negative	
011990000118	phad	5	6:18:18 PM	74	0.67 PAU/I	---	Positive	
011990000119	phad	5	5:54:19 PM	43	0.31 PAU/I	---	Negative	
011990000120	phad	5	6:02:19 PM	44	0.32 PAU/I	---	Negative	
011990000121	phad	5	6:26:18 PM	36	0.23 PAU/I	---	Negative	
011990000122	phad	5	6:34:18 PM	33	0.20 PAU/I	---	Negative	
011990000133	phad	5	8:21:14 PM	135	1.37 PAU/I	---	Positive	
011990000134	phad	5	8:29:14 PM	31	0.17 PAU/I	---	Negative	
011990000135	phad	5	8:05:15 PM	34	0.20 PAU/I	---	Negative	
011990000136	phad	5	8:13:14 PM	35	0.22 PAU/I	---	Negative	
011990000137	phad	5	7:47:15 PM	66	0.57 PAU/I	---	Positive	
011990000138	phad	5	7:55:15 PM	33	0.19 PAU/I	---	Negative	
011990000139	phad	5	7:14:16 PM	32	0.18 PAU/I	---	Negative	
011990000140	phad	5	7:23:16 PM	53	0.43 PAU/I	---	Positive	
011990000141	phad	5	7:31:16 PM	509	5.73 PAU/I	---	Positive	
011990000142	phad	5	7:39:15 PM	34	0.20 PAU/I	---	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000102	fx5	5	5:47:19 PM	24	0.09 kUA/I	---	Negative	
011990000102	fx1	5	5:48:19 PM	15	---	---	Negative	
011990000102	fx2	5	5:49:19 PM	23	0.07 kUA/I	---	Negative	
011990000102	fx22	5	5:50:19 PM	28	0.13 kUA/I	---	Negative	
011990000102	gx2	5	5:51:19 PM	22	0.06 kUA/I	---	Negative	
011990000102	wx1	5	5:52:19 PM	34	0.21 kUA/I	---	Negative	
011990000102	t20	5	5:53:19 PM	18	0.02 kUA/I	---		
011990000103	fx5	5	5:23:20 PM	62	0.53 kUA/I	---	Positive	
011990000103	fx1	5	5:24:20 PM	47	0.36 kUA/I	---	Positive	
011990000103	fx2	5	5:25:20 PM	154	1.60 kUA/I	---	Positive	
011990000103	fx22	5	5:26:20 PM	51	0.40 kUA/I	---	Positive	
011990000103	gx2	5	5:27:20 PM	41	0.28 kUA/I	---	Negative	
011990000103	wx1	5	5:28:20 PM	60	0.50 kUA/I	---	Positive	
011990000103	t20	5	5:29:20 PM	28	0.13 kUA/I	---		
011990000104	fx5	5	5:31:20 PM	27	0.12 kUA/I	---	Negative	
011990000104	fx1	5	5:32:20 PM	20	0.04 kUA/I	---	Negative	
011990000104	fx2	5	5:33:20 PM	12	---	---	Negative	
011990000104	fx22	5	5:34:20 PM	29	0.15 kUA/I	---	Negative	
011990000104	gx2	5	5:35:20 PM	25	0.09 kUA/I	---	Negative	
011990000104	wx1	5	5:36:20 PM	30	0.16 kUA/I	---	Negative	
011990000104	t20	5	5:37:19 PM	16	---	---		
011990000106	fx5	5	5:15:21 PM	21	0.05 kUA/I	---	Negative	
011990000106	fx1	5	5:16:20 PM	17	---	---	Negative	
011990000106	fx2	5	5:17:20 PM	13	---	---	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000106	fx22	5	5:18:20 PM	34	0.20 kUA/l	---	Negative	
011990000106	gx2	5	5:19:20 PM	24	0.08 kUA/l	---	Negative	
011990000106	wx1	5	5:20:20 PM	32	0.18 kUA/l	---	Negative	
011990000106	t20	5	5:21:20 PM	17	---	---		
011990000112	fx5	5	5:39:19 PM	28	0.13 kUA/l	---	Negative	
011990000112	fx1	5	5:40:19 PM	17	---	---	Negative	
011990000112	fx2	5	5:41:19 PM	15	---	---	Negative	
011990000112	fx22	5	5:42:20 PM	32	0.18 kUA/l	---	Negative	
011990000112	gx2	5	5:43:20 PM	22	0.06 kUA/l	---	Negative	
011990000112	wx1	5	5:44:20 PM	34	0.20 kUA/l	---	Negative	
011990000112	t20	5	5:45:19 PM	18	0.01 kUA/l	---		
011990000113	fx5	5	6:59:17 PM	22	0.06 kUA/l	---	Negative	
011990000113	fx1	5	7:00:17 PM	20	0.03 kUA/l	---	Negative	
011990000113	fx2	5	7:01:17 PM	17	---	---	Negative	
011990000113	fx22	5	7:02:17 PM	27	0.12 kUA/l	---	Negative	
011990000113	gx2	5	7:03:17 PM	24	0.09 kUA/l	---	Negative	
011990000113	wx1	5	7:04:16 PM	31	0.17 kUA/l	---	Negative	
011990000113	t20	5	7:05:16 PM	16	---	---		
011990000114	fx5	5	7:07:17 PM	29	0.14 kUA/l	---	Negative	
011990000114	fx1	5	7:08:17 PM	18	0.02 kUA/l	---	Negative	
011990000114	fx2	5	7:09:16 PM	17	---	---	Negative	
011990000114	fx22	5	7:10:16 PM	29	0.14 kUA/l	---	Negative	
011990000114	gx2	5	7:11:16 PM	29	0.14 kUA/l	---	Negative	
011990000114	wx1	5	7:12:16 PM	45	0.33 kUA/l	---	Negative	
011990000114	t20	5	7:13:16 PM	16	---	---		
011990000115	fx5	5	6:43:17 PM	25	0.10 kUA/l	---	Negative	
011990000115	fx1	5	6:44:17 PM	21	0.04 kUA/l	---	Negative	
011990000115	fx2	5	6:45:17 PM	13	---	---	Negative	
011990000115	fx22	5	6:46:17 PM	29	0.14 kUA/l	---	Negative	
011990000115	gx2	5	6:47:17 PM	23	0.07 kUA/l	---	Negative	
011990000115	wx1	5	6:48:17 PM	31	0.16 kUA/l	---	Negative	
011990000115	t20	5	6:49:17 PM	18	0.01 kUA/l	---		
011990000116	fx5	5	6:51:17 PM	26	0.11 kUA/l	---	Negative	
011990000116	fx1	5	6:52:17 PM	19	0.02 kUA/l	---	Negative	
011990000116	fx2	5	6:53:17 PM	62	0.53 kUA/l	---	Positive	
011990000116	fx22	5	6:54:17 PM	32	0.18 kUA/l	---	Negative	
011990000116	gx2	5	6:55:17 PM	25	0.10 kUA/l	---	Negative	
011990000116	wx1	5	6:56:17 PM	41	0.29 kUA/l	---	Negative	
011990000116	t20	5	6:57:17 PM	21	0.04 kUA/l	---		
011990000117	fx5	5	6:11:19 PM	23	0.07 kUA/l	---	Negative	
011990000117	fx1	5	6:12:18 PM	20	0.04 kUA/l	---	Negative	
011990000117	fx2	5	6:13:18 PM	17	---	---	Negative	
011990000117	fx22	5	6:14:18 PM	33	0.20 kUA/l	---	Negative	
011990000117	gx2	5	6:15:18 PM	25	0.10 kUA/l	---	Negative	
011990000117	wx1	5	6:16:18 PM	32	0.18 kUA/l	---	Negative	
011990000117	t20	5	6:17:18 PM	17	---	---		
011990000118	fx5	5	6:19:18 PM	25	0.10 kUA/l	---	Negative	
011990000118	fx1	5	6:20:18 PM	20	0.03 kUA/l	---	Negative	
011990000118	fx2	5	6:21:18 PM	18	0.02 kUA/l	---	Negative	
011990000118	fx22	5	6:22:18 PM	30	0.15 kUA/l	---	Negative	
011990000118	gx2	5	6:23:18 PM	21	0.05 kUA/l	---	Negative	
011990000118	wx1	5	6:24:18 PM	29	0.15 kUA/l	---	Negative	
011990000118	t20	5	6:25:18 PM	19	0.03 kUA/l	---		
011990000119	fx5	5	5:55:19 PM	26	0.11 kUA/l	---	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000119	fx1	5	5:56:19 PM	16	---	---	Negative	
011990000119	fx2	5	5:57:19 PM	19	0.02 kUA/I	---	Negative	
011990000119	fx22	5	5:58:19 PM	31	0.17 kUA/I	---	Negative	
011990000119	gx2	5	5:59:19 PM	27	0.12 kUA/I	---	Negative	
011990000119	wx1	5	6:00:19 PM	32	0.18 kUA/I	---	Negative	
011990000119	t20	5	6:01:19 PM	17	---	---		
011990000120	fx5	5	6:03:19 PM	67	0.59 kUA/I	---	Positive	
011990000120	fx1	5	6:04:19 PM	49	0.37 kUA/I	---	Positive	
011990000120	fx2	5	6:05:18 PM	18	0.01 kUA/I	---	Negative	
011990000120	fx22	5	6:06:19 PM	34	0.20 kUA/I	---	Negative	
011990000120	gx2	5	6:07:19 PM	33	0.19 kUA/I	---	Negative	
011990000120	wx1	5	6:08:18 PM	35	0.22 kUA/I	---	Negative	
011990000120	t20	5	6:09:18 PM	22	0.06 kUA/I	---		
011990000121	fx5	5	6:27:18 PM	24	0.09 kUA/I	---	Negative	
011990000121	fx1	5	6:28:18 PM	23	0.07 kUA/I	---	Negative	
011990000121	fx2	5	6:29:18 PM	15	---	---	Negative	
011990000121	fx22	5	6:30:18 PM	30	0.16 kUA/I	---	Negative	
011990000121	gx2	5	6:31:18 PM	21	0.05 kUA/I	---	Negative	
011990000121	wx1	5	6:32:18 PM	30	0.16 kUA/I	---	Negative	
011990000121	t20	5	6:33:18 PM	19	0.02 kUA/I	---		
011990000122	fx5	5	6:35:18 PM	21	0.05 kUA/I	---	Negative	
011990000122	fx1	5	6:36:17 PM	20	0.04 kUA/I	---	Negative	
011990000122	fx2	5	6:37:18 PM	15	---	---	Negative	
011990000122	fx22	5	6:38:17 PM	32	0.18 kUA/I	---	Negative	
011990000122	gx2	5	6:39:18 PM	22	0.06 kUA/I	---	Negative	
011990000122	wx1	5	6:40:18 PM	30	0.16 kUA/I	---	Negative	
011990000122	t20	5	6:41:17 PM	16	---	---		
011990000133	fx5	5	8:22:14 PM	26	0.11 kUA/I	---	Negative	
011990000133	fx1	5	8:23:14 PM	22	0.06 kUA/I	---	Negative	
011990000133	fx2	5	8:24:14 PM	160	1.67 kUA/I	---	Positive	
011990000133	fx22	5	8:25:14 PM	43	0.31 kUA/I	---	Negative	
011990000133	gx2	5	8:26:14 PM	35	0.22 kUA/I	---	Negative	
011990000133	wx1	5	8:27:14 PM	51	0.40 kUA/I	---	Positive	
011990000133	t20	5	8:28:14 PM	34	0.20 kUA/I	---		
011990000134	fx5	5	8:30:13 PM	22	0.07 kUA/I	---	Negative	
011990000134	fx1	5	8:31:14 PM	19	0.02 kUA/I	---	Negative	
011990000134	fx2	5	8:32:14 PM	14	---	---	Negative	
011990000134	fx22	5	8:33:14 PM	30	0.16 kUA/I	---	Negative	
011990000134	gx2	5	8:34:13 PM	20	0.04 kUA/I	---	Negative	
011990000134	wx1	5	8:35:13 PM	31	0.17 kUA/I	---	Negative	
011990000134	t20	5	8:36:13 PM	17	---	---		
011990000135	fx5	5	8:06:14 PM	21	0.05 kUA/I	---	Negative	
011990000135	fx1	5	8:07:14 PM	19	0.02 kUA/I	---	Negative	
011990000135	fx2	5	8:08:14 PM	18	0.02 kUA/I	---	Negative	
011990000135	fx22	5	8:09:14 PM	31	0.17 kUA/I	---	Negative	
011990000135	gx2	5	8:10:14 PM	26	0.12 kUA/I	---	Negative	
011990000135	wx1	5	8:11:14 PM	31	0.16 kUA/I	---	Negative	
011990000135	t20	5	8:12:14 PM	18	0.02 kUA/I	---		
011990000136	fx5	5	8:14:14 PM	38	0.25 kUA/I	---	Negative	
011990000136	fx1	5	8:15:14 PM	25	0.09 kUA/I	---	Negative	
011990000136	fx2	5	8:16:14 PM	14	---	---	Negative	
011990000136	fx22	5	8:17:14 PM	31	0.17 kUA/I	---	Negative	
011990000136	gx2	5	8:18:14 PM	27	0.12 kUA/I	---	Negative	
011990000136	wx1	5	8:19:14 PM	33	0.19 kUA/I	---	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000136	t20	5	8:20:14 PM	16	---	---		
011990000137	fx5	5	7:48:15 PM	22	0.07 kUA/l	---	Negative	
011990000137	fx1	5	7:49:15 PM	20	0.04 kUA/l	---	Negative	
011990000137	fx2	5	7:50:15 PM	59	0.49 kUA/l	---	Positive	
011990000137	fx22	5	7:51:15 PM	39	0.26 kUA/l	---	Negative	
011990000137	gx2	5	7:52:15 PM	27	0.12 kUA/l	---	Negative	
011990000137	wx1	5	7:53:15 PM	37	0.24 kUA/l	---	Negative	
011990000137	t20	5	7:54:15 PM	22	0.07 kUA/l	---		
011990000138	fx5	5	7:56:15 PM	22	0.06 kUA/l	---	Negative	
011990000138	fx1	5	7:57:15 PM	21	0.05 kUA/l	---	Negative	
011990000138	fx2	5	7:58:15 PM	16	---	---	Negative	
011990000138	fx22	5	7:59:15 PM	31	0.17 kUA/l	---	Negative	
011990000138	gx2	5	8:02:14 PM	23	0.07 kUA/l	---	Negative	
011990000138	wx1	5	8:03:14 PM	33	0.19 kUA/l	---	Negative	
011990000138	t20	5	8:04:15 PM	16	---	---		
011990000139	fx5	5	7:16:16 PM	28	0.14 kUA/l	---	Negative	
011990000139	fx1	5	7:17:16 PM	20	0.04 kUA/l	---	Negative	
011990000139	fx2	5	7:18:16 PM	17	---	---	Negative	
011990000139	fx22	5	7:19:16 PM	29	0.15 kUA/l	---	Negative	
011990000139	gx2	5	7:20:16 PM	22	0.07 kUA/l	---	Negative	
011990000139	wx1	5	7:21:16 PM	31	0.17 kUA/l	---	Negative	
011990000139	t20	5	7:22:16 PM	17	0.01 kUA/l	---		
011990000140	fx5	5	7:24:16 PM	25	0.09 kUA/l	---	Negative	
011990000140	fx1	5	7:25:16 PM	19	0.03 kUA/l	---	Negative	
011990000140	fx2	5	7:26:16 PM	123	1.24 kUA/l	---	Positive	
011990000140	fx22	5	7:27:16 PM	29	0.14 kUA/l	---	Negative	
011990000140	gx2	5	7:28:16 PM	33	0.19 kUA/l	---	Negative	
011990000140	wx1	5	7:29:16 PM	35	0.22 kUA/l	---	Negative	
011990000140	t20	5	7:30:16 PM	40	0.27 kUA/l	---		
011990000141	fx5	5	7:32:16 PM	114	1.14 kUA/l	---	Positive	
011990000141	fx1	5	7:33:16 PM	63	0.55 kUA/l	---	Positive	
011990000141	fx2	5	7:34:15 PM	84	0.78 kUA/l	---	Positive	
011990000141	fx22	5	7:35:16 PM	66	0.58 kUA/l	---	Positive	
011990000141	gx2	5	7:36:16 PM	75	0.68 kUA/l	---	Positive	
011990000141	wx1	5	7:37:16 PM	111	1.10 kUA/l	---	Positive	
011990000141	t20	5	7:38:15 PM	67	0.59 kUA/l	---		
011990000142	fx5	5	7:40:15 PM	20	0.04 kUA/l	---	Negative	
011990000142	fx1	5	7:41:15 PM	23	0.07 kUA/l	---	Negative	
011990000142	fx2	5	7:42:15 PM	15	---	---	Negative	
011990000142	fx22	5	7:43:15 PM	31	0.17 kUA/l	---	Negative	
011990000142	gx2	5	7:44:15 PM	29	0.14 kUA/l	---	Negative	
011990000142	wx1	5	7:45:15 PM	32	0.19 kUA/l	---	Negative	
011990000142	t20	5	7:46:15 PM	17	---	---		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	7:15:16 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000106	phad	5:14:21 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000106	fx5	5:15:21 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000106	fx1	5:16:20 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000106	fx2	5:17:20 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000106	fx22	5:18:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000106	gx2	5:19:20 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000106	wx1	5:20:20 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000106	t20	5:21:20 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	phad	5:22:20 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	fx5	5:23:20 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	fx1	5:24:20 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	fx2	5:25:20 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	fx22	5:26:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	gx2	5:27:20 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	wx1	5:28:20 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000103	t20	5:29:20 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	phad	5:30:20 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	fx5	5:31:20 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	fx1	5:32:20 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	fx2	5:33:20 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	fx22	5:34:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	gx2	5:35:20 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	wx1	5:36:20 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000104	t20	5:37:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	phad	5:38:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	fx5	5:39:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	fx1	5:40:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	fx2	5:41:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	fx22	5:42:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	gx2	5:43:20 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	wx1	5:44:20 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000112	t20	5:45:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	phad	5:46:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	fx5	5:47:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	fx1	5:48:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	fx2	5:49:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	fx22	5:50:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	gx2	5:51:19 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	wx1	5:52:19 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000102	t20	5:53:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	phad	5:54:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	fx5	5:55:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	fx1	5:56:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	fx2	5:57:19 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	fx22	5:58:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	gx2	5:59:19 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	wx1	6:00:19 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000119	t20	6:01:19 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000120	phad	6:02:19 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000120	fx5	6:03:19 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000120	fx1	6:04:19 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000120	fx2	6:05:18 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000120	fx22	6:06:19 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000120	gx2	6:07:19 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	

21-Jan-20 10:24:05 AM

21-Jan-20 10:24:05 AM

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000133	phad	8:21:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000133	fx5	8:22:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000133	fx1	8:23:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000133	fx2	8:24:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000133	fx22	8:25:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000133	gx2	8:26:14 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000133	wx1	8:27:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000133	t20	8:28:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	phad	8:29:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	fx5	8:30:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	fx1	8:31:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	fx2	8:32:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	fx22	8:33:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	gx2	8:34:13 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	wx1	8:35:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000134	t20	8:36:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200113-377
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 13-Jan-20 8:37:13 PM
Ended: 13-Jan-20 10:03:31 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

No quality controls in this run

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000123	phad	5	9:25:11 PM	32	0.19 PAU/l	---	Negative	
011990000124	phad	5	9:33:11 PM	37	0.24 PAU/l	---	Negative	
011990000125	phad	5	9:09:12 PM	37	0.24 PAU/l	---	Negative	
011990000126	phad	5	9:17:11 PM	38	0.25 PAU/l	---	Negative	
011990000127	phad	5	8:53:13 PM	514	5.78 PAU/l	---	Positive	
011990000128	phad	5	9:01:13 PM	36	0.22 PAU/l	---	Negative	
011990000129	phad	5	8:37:13 PM	490	5.50 PAU/l	---	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000130	phad	5	8:45:13 PM	35	0.22 PAU/I	---	Negative	
011990000131	phad	5	9:41:11 PM	1341	15.7 PAU/I	---	Positive	
011990000132	phad	5	9:49:10 PM	39	0.26 PAU/I	---	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000123	fx5	5	9:26:11 PM	22	0.07 kUA/I	---	Negative	
011990000123	fx1	5	9:27:11 PM	18	0.02 kUA/I	---	Negative	
011990000123	fx2	5	9:28:11 PM	15	---	---	Negative	
011990000123	fx22	5	9:29:11 PM	30	0.16 kUA/I	---	Negative	
011990000123	gx2	5	9:30:11 PM	24	0.08 kUA/I	---	Negative	
011990000123	wx1	5	9:31:11 PM	26	0.11 kUA/I	---	Negative	
011990000123	t20	5	9:32:11 PM	18	0.02 kUA/I	---		
011990000124	fx5	5	9:34:11 PM	26	0.11 kUA/I	---	Negative	
011990000124	fx1	5	9:35:11 PM	17	---	---	Negative	
011990000124	fx2	5	9:36:10 PM	15	---	---	Negative	
011990000124	fx22	5	9:37:11 PM	32	0.18 kUA/I	---	Negative	
011990000124	gx2	5	9:38:11 PM	32	0.18 kUA/I	---	Negative	
011990000124	wx1	5	9:39:11 PM	32	0.18 kUA/I	---	Negative	
011990000124	t20	5	9:40:11 PM	18	0.02 kUA/I	---		
011990000125	fx5	5	9:10:12 PM	31	0.17 kUA/I	---	Negative	
011990000125	fx1	5	9:11:12 PM	21	0.05 kUA/I	---	Negative	
011990000125	fx2	5	9:12:12 PM	14	---	---	Negative	
011990000125	fx22	5	9:13:12 PM	33	0.19 kUA/I	---	Negative	
011990000125	gx2	5	9:14:12 PM	24	0.08 kUA/I	---	Negative	
011990000125	wx1	5	9:15:12 PM	37	0.23 kUA/I	---	Negative	
011990000125	t20	5	9:16:12 PM	18	0.01 kUA/I	---		
011990000126	fx5	5	9:18:11 PM	24	0.08 kUA/I	---	Negative	
011990000126	fx1	5	9:19:12 PM	23	0.07 kUA/I	---	Negative	
011990000126	fx2	5	9:20:12 PM	14	---	---	Negative	
011990000126	fx22	5	9:21:11 PM	33	0.19 kUA/I	---	Negative	
011990000126	gx2	5	9:22:11 PM	27	0.12 kUA/I	---	Negative	
011990000126	wx1	5	9:23:11 PM	30	0.16 kUA/I	---	Negative	
011990000126	t20	5	9:24:11 PM	19	0.03 kUA/I	---		
011990000127	fx5	5	8:54:13 PM	39	0.27 kUA/I	---	Negative	
011990000127	fx1	5	8:55:13 PM	37	0.24 kUA/I	---	Negative	
011990000127	fx2	5	8:56:12 PM	576	6.51 kUA/I	---	Positive	
011990000127	fx22	5	8:57:12 PM	49	0.38 kUA/I	---	Positive	
011990000127	gx2	5	8:58:12 PM	1079	12.5 kUA/I	---	Positive	
011990000127	wx1	5	8:59:12 PM	58	0.49 kUA/I	---	Positive	
011990000127	t20	5	9:00:13 PM	34	0.20 kUA/I	---		
011990000128	fx5	5	9:02:12 PM	22	0.06 kUA/I	---	Negative	
011990000128	fx1	5	9:03:12 PM	19	0.02 kUA/I	---	Negative	
011990000128	fx2	5	9:04:12 PM	14	---	---	Negative	
011990000128	fx22	5	9:05:12 PM	36	0.23 kUA/I	---	Negative	
011990000128	gx2	5	9:06:12 PM	30	0.16 kUA/I	---	Negative	
011990000128	wx1	5	9:07:12 PM	31	0.17 kUA/I	---	Negative	
011990000128	t20	5	9:08:12 PM	17	---	---		
011990000129	fx5	5	8:38:13 PM	74	0.67 kUA/I	---	Positive	
011990000129	fx1	5	8:39:13 PM	33	0.19 kUA/I	---	Negative	
011990000129	fx2	5	8:40:13 PM	56	0.46 kUA/I	---	Positive	
011990000129	fx22	5	8:41:13 PM	39	0.27 kUA/I	---	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000129	gx2	5	8:42:13 PM	264	2.87 kUA/I	---	Positive	
011990000129	wx1	5	8:43:13 PM	70	0.63 kUA/I	---	Positive	
011990000129	t20	5	8:44:13 PM	28	0.13 kUA/I	---		
011990000130	fx5	5	8:46:13 PM	24	0.08 kUA/I	---	Negative	
011990000130	fx1	5	8:47:13 PM	20	0.04 kUA/I	---	Negative	
011990000130	fx2	5	8:48:13 PM	18	0.01 kUA/I	---	Negative	
011990000130	fx22	5	8:49:13 PM	31	0.17 kUA/I	---	Negative	
011990000130	gx2	5	8:50:13 PM	23	0.08 kUA/I	---	Negative	
011990000130	wx1	5	8:51:13 PM	29	0.14 kUA/I	---	Negative	
011990000130	t20	5	8:52:13 PM	16	---	---		
011990000131	fx5	5	9:42:10 PM	26	0.11 kUA/I	---	Negative	
011990000131	fx1	5	9:43:11 PM	20	0.04 kUA/I	---	Negative	
011990000131	fx2	5	9:44:10 PM	1579	18.6 kUA/I	---	Positive	
011990000131	fx22	5	9:45:11 PM	50	0.39 kUA/I	---	Positive	
011990000131	gx2	5	9:46:10 PM	84	0.78 kUA/I	---	Positive	
011990000131	wx1	5	9:47:11 PM	129	1.30 kUA/I	---	Positive	
011990000131	t20	5	9:48:10 PM	43	0.30 kUA/I	---		
011990000132	fx5	5	9:50:10 PM	83	0.78 kUA/I	---	Positive	
011990000132	fx1	5	9:51:11 PM	38	0.25 kUA/I	---	Negative	
011990000132	fx2	5	9:52:10 PM	16	---	---	Negative	
011990000132	fx22	5	9:53:11 PM	34	0.21 kUA/I	---	Negative	
011990000132	gx2	5	9:54:10 PM	24	0.09 kUA/I	---	Negative	
011990000132	wx1	5	9:55:10 PM	30	0.15 kUA/I	---	Negative	
011990000132	t20	5	9:56:10 PM	18	0.01 kUA/I	---		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000129	phad	8:37:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000129	fx5	8:38:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000129	fx1	8:39:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000129	fx2	8:40:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000129	fx22	8:41:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000129	gx2	8:42:13 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000129	wx1	8:43:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000129	t20	8:44:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	phad	8:45:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	fx5	8:46:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	fx1	8:47:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	fx2	8:48:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	fx22	8:49:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	gx2	8:50:13 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	wx1	8:51:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000130	t20	8:52:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000127	phad	8:53:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000127	fx5	8:54:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000127	fx1	8:55:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000127	fx2	8:56:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000127	fx22	8:57:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000127	gx2	8:58:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000127	wx1	8:59:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000127	t20	9:00:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	phad	9:01:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	fx5	9:02:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	fx1	9:03:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	fx2	9:04:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	fx22	9:05:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	gx2	9:06:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	wx1	9:07:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000128	t20	9:08:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	phad	9:09:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	fx5	9:10:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	fx1	9:11:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	fx2	9:12:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	fx22	9:13:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	gx2	9:14:12 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	wx1	9:15:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000125	t20	9:16:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	phad	9:17:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	fx5	9:18:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	fx1	9:19:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	fx2	9:20:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	fx22	9:21:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	gx1	9:22:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	wx1	9:23:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000126	t20	9:24:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	phad	9:25:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	fx5	9:26:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	fx1	9:27:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	fx2	9:28:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	fx22	9:29:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	gx2	9:30:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	wx1	9:31:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000123	t20	9:32:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	phad	9:33:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	fx5	9:34:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	fx1	9:35:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	fx2	9:36:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	fx22	9:37:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	gx2	9:38:11 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	wx1	9:39:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000124	t20	9:40:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	phad	9:41:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	fx5	9:42:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	fx1	9:43:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	fx2	9:44:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	fx22	9:45:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	gx2	9:46:10 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	wx1	9:47:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000131	t20	9:48:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	phad	9:49:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	fx5	9:50:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	fx1	9:51:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	fx2	9:52:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	fx22	9:53:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	gx2	9:54:10 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	wx1	9:55:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000132	t20	9:56:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	