


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200126-386
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 26-Jan-20 10:56:57 AM
Ended: 26-Jan-20 3:23:48 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	10:56:57 AM	311	0.68	OK	
CC-2	17.5	10:57:57 AM	6461	18.0	OK	

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	11:33:56 AM	3267	8.13 kUA/I	CZ1AR	OK	
slgE L	e1	1	11:36:56 AM	1205	2.80 kUA/I	CZ4AN	OK	
slgE Neg	d1	1	11:37:56 AM	20	0.01 kUA/I	BL1CP	OK	
IQC271119	d1	1	11:38:56 AM	6229	17.2 kUA/I	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000409	phad	5	1:11:52 PM	1093	12.6 PAU/I	3	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000410	phad	5	12:31:54 PM	33	0.19 PAU/I	0	Negative	
011990000411	phad	5	2:35:49 PM	55	0.45 PAU/I	1	Positive	
011990000412	phad	5	1:51:51 PM	52	0.41 PAU/I	1	Positive	
011990000421	phad	5	1:03:53 PM	318	3.49 PAU/I	2	Positive	
011990000422	phad	5	12:23:54 PM	36	0.23 PAU/I	0	Negative	
011990000423	phad	5	2:27:50 PM	31	0.17 PAU/I	0	Negative	
011990000424	phad	5	1:43:51 PM	36	0.23 PAU/I	0	Negative	
011990000433	phad	5	12:55:53 PM	34	0.21 PAU/I	0	Negative	
011990000434	phad	5	12:15:55 PM	34	0.20 PAU/I	0	Negative	
011990000435	phad	5	2:19:50 PM	517	5.82 PAU/I	3	Positive	
011990000436	phad	5	1:35:52 PM	42	0.30 PAU/I	0	Negative	
011990000445	phad	5	12:47:53 PM	63	0.54 PAU/I	1	Positive	
011990000446	phad	5	12:07:55 PM	125	1.26 PAU/I	2	Positive	
011990000447	phad	5	2:10:51 PM	34	0.21 PAU/I	0	Negative	
011990000448	phad	5	1:27:52 PM	34	0.21 PAU/I	0	Negative	
011990000457	phad	5	12:39:54 PM	43	0.31 PAU/I	0	Negative	
011990000458	phad	5	11:59:55 AM	74	0.67 PAU/I	1	Positive	
011990000459	phad	5	1:59:51 PM	4528	58.9 PAU/I	5	Positive	
011990000460	phad	5	1:19:52 PM	1174	13.6 PAU/I	3	Positive	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000409	fx5	5	1:12:53 PM	38	0.25 kUA/I	0	Negative	
011990000409	fx1	5	1:13:52 PM	28	0.14 kUA/I	0	Negative	
011990000409	fx2	5	1:14:52 PM	1190	13.8 kUA/I	3	Positive	
011990000409	fx22	5	1:15:52 PM	44	0.32 kUA/I	0	Negative	
011990000409	gx2	5	1:16:52 PM	73	0.65 kUA/I	1	Positive	
011990000409	wx1	5	1:17:52 PM	320	3.52 kUA/I	3	Positive	
011990000409	t20	5	1:18:52 PM	121	1.22 kUA/I	2		
011990000410	fx5	5	12:32:54 PM	19	0.03 kUA/I	0	Negative	
011990000410	fx1	5	12:33:54 PM	18	0.02 kUA/I	0	Negative	
011990000410	fx2	5	12:34:54 PM	15	Low RU	0	Negative	
011990000410	fx22	5	12:35:54 PM	28	0.13 kUA/I	0	Negative	
011990000410	gx2	5	12:36:54 PM	23	0.07 kUA/I	0	Negative	
011990000410	wx1	5	12:37:53 PM	29	0.15 kUA/I	0	Negative	
011990000410	t20	5	12:38:54 PM	18	0.01 kUA/I	0		
011990000411	fx5	5	2:36:50 PM	22	0.06 kUA/I	0	Negative	
011990000411	fx1	5	2:37:50 PM	16	Low RU	0	Negative	
011990000411	fx2	5	2:38:49 PM	28	0.13 kUA/I	0	Negative	
011990000411	fx22	5	2:39:49 PM	31	0.17 kUA/I	0	Negative	
011990000411	gx2	5	2:40:49 PM	23	0.07 kUA/I	0	Negative	
011990000411	wx1	5	2:41:49 PM	32	0.18 kUA/I	0	Negative	
011990000411	t20	5	2:42:49 PM	16	Low RU	0		
011990000412	fx5	5	1:52:51 PM	24	0.09 kUA/I	0	Negative	
011990000412	fx1	5	1:53:51 PM	22	0.06 kUA/I	0	Negative	
011990000412	fx2	5	1:54:51 PM	16	Low RU	0	Negative	
011990000412	fx22	5	1:55:51 PM	41	0.28 kUA/I	0	Negative	
011990000412	gx2	5	1:56:51 PM	37	0.24 kUA/I	0	Negative	
011990000412	wx1	5	1:57:51 PM	75	0.68 kUA/I	1	Positive	
011990000412	t20	5	1:58:51 PM	49	0.38 kUA/I	1		
011990000421	fx5	5	1:04:53 PM	26	0.11 kUA/I	0	Negative	
011990000421	fx1	5	1:05:53 PM	20	0.04 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000421	fx2	5	1:06:53 PM	40	0.27 kUA/l	0	Negative	
011990000421	fx22	5	1:07:53 PM	30	0.16 kUA/l	0	Negative	
011990000421	gx2	5	1:08:53 PM	24	0.09 kUA/l	0	Negative	
011990000421	wx1	5	1:09:52 PM	34	0.20 kUA/l	0	Negative	
011990000421	t20	5	1:10:52 PM	19	0.03 kUA/l	0		
011990000422	fx5	5	12:24:54 PM	21	0.05 kUA/l	0	Negative	
011990000422	fx1	5	12:25:54 PM	19	0.03 kUA/l	0	Negative	
011990000422	fx2	5	12:26:54 PM	11	Low RU	0	Negative	
011990000422	fx22	5	12:27:54 PM	29	0.14 kUA/l	0	Negative	
011990000422	gx2	5	12:28:54 PM	25	0.10 kUA/l	0	Negative	
011990000422	wx1	5	12:29:54 PM	31	0.17 kUA/l	0	Negative	
011990000422	t20	5	12:30:54 PM	18	0.02 kUA/l	0		
011990000423	fx5	5	2:28:50 PM	20	0.04 kUA/l	0	Negative	
011990000423	fx1	5	2:29:50 PM	21	0.04 kUA/l	0	Negative	
011990000423	fx2	5	2:30:50 PM	14	Low RU	0	Negative	
011990000423	fx22	5	2:31:50 PM	31	0.16 kUA/l	0	Negative	
011990000423	gx2	5	2:32:50 PM	23	0.08 kUA/l	0	Negative	
011990000423	wx1	5	2:33:49 PM	32	0.18 kUA/l	0	Negative	
011990000423	t20	5	2:34:49 PM	17	Low RU	0		
011990000424	fx5	5	1:44:51 PM	24	0.09 kUA/l	0	Negative	
011990000424	fx1	5	1:45:51 PM	23	0.08 kUA/l	0	Negative	
011990000424	fx2	5	1:46:51 PM	14	Low RU	0	Negative	
011990000424	fx22	5	1:47:51 PM	29	0.15 kUA/l	0	Negative	
011990000424	gx2	5	1:48:51 PM	26	0.11 kUA/l	0	Negative	
011990000424	wx1	5	1:49:52 PM	34	0.20 kUA/l	0	Negative	
011990000424	t20	5	1:50:51 PM	18	0.01 kUA/l	0		
011990000433	fx5	5	12:56:53 PM	23	0.08 kUA/l	0	Negative	
011990000433	fx1	5	12:57:53 PM	19	0.03 kUA/l	0	Negative	
011990000433	fx2	5	12:58:53 PM	12	Low RU	0	Negative	
011990000433	fx22	5	12:59:53 PM	31	0.17 kUA/l	0	Negative	
011990000433	gx2	5	1:00:53 PM	60	0.51 kUA/l	1	Positive	
011990000433	wx1	5	1:01:53 PM	56	0.46 kUA/l	1	Positive	
011990000433	t20	5	1:02:53 PM	17	0.01 kUA/l	0		
011990000434	fx5	5	12:16:54 PM	22	0.06 kUA/l	0	Negative	
011990000434	fx1	5	12:17:54 PM	18	0.01 kUA/l	0	Negative	
011990000434	fx2	5	12:18:54 PM	15	Low RU	0	Negative	
011990000434	fx22	5	12:19:54 PM	31	0.17 kUA/l	0	Negative	
011990000434	gx2	5	12:20:54 PM	59	0.50 kUA/l	1	Positive	
011990000434	wx1	5	12:21:54 PM	31	0.17 kUA/l	0	Negative	
011990000434	t20	5	12:22:54 PM	18	0.02 kUA/l	0		
011990000435	fx5	5	2:20:50 PM	25	0.09 kUA/l	0	Negative	
011990000435	fx1	5	2:21:50 PM	21	0.05 kUA/l	0	Negative	
011990000435	fx2	5	2:22:50 PM	297	3.25 kUA/l	2	Positive	
011990000435	fx22	5	2:23:50 PM	37	0.24 kUA/l	0	Negative	
011990000435	gx2	5	2:24:50 PM	41	0.29 kUA/l	0	Negative	
011990000435	wx1	5	2:25:50 PM	1308	15.2 kUA/l	3	Positive	
011990000435	t20	5	2:26:50 PM	48	0.37 kUA/l	1		
011990000436	fx5	5	1:36:52 PM	21	0.06 kUA/l	0	Negative	
011990000436	fx1	5	1:37:51 PM	22	0.06 kUA/l	0	Negative	
011990000436	fx2	5	1:38:51 PM	23	0.07 kUA/l	0	Negative	
011990000436	fx22	5	1:39:51 PM	34	0.20 kUA/l	0	Negative	
011990000436	gx2	5	1:40:52 PM	27	0.13 kUA/l	0	Negative	
011990000436	wx1	5	1:41:52 PM	30	0.16 kUA/l	0	Negative	
011990000436	t20	5	1:42:52 PM	18	0.01 kUA/l	0		

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000445	fx5	5	12:48:53 PM	20	0.04 kUA/l	0	Negative	
011990000445	fx1	5	12:49:53 PM	18	0.02 kUA/l	0	Negative	
011990000445	fx2	5	12:50:53 PM	56	0.46 kUA/l	1	Positive	
011990000445	fx22	5	12:51:53 PM	29	0.14 kUA/l	0	Negative	
011990000445	gx2	5	12:52:54 PM	22	0.06 kUA/l	0	Negative	
011990000445	wx1	5	12:53:53 PM	32	0.19 kUA/l	0	Negative	
011990000445	t20	5	12:54:53 PM	17	Low RU	0		
011990000446	fx5	5	12:08:54 PM	24	0.08 kUA/l	0	Negative	
011990000446	fx1	5	12:09:55 PM	32	0.18 kUA/l	0	Negative	
011990000446	fx2	5	12:10:55 PM	37	0.24 kUA/l	0	Negative	
011990000446	fx22	5	12:11:54 PM	32	0.17 kUA/l	0	Negative	
011990000446	gx2	5	12:12:54 PM	24	0.08 kUA/l	0	Negative	
011990000446	wx1	5	12:13:54 PM	32	0.18 kUA/l	0	Negative	
011990000446	t20	5	12:14:54 PM	19	0.02 kUA/l	0		
011990000447	fx5	5	2:11:50 PM	25	0.10 kUA/l	0	Negative	
011990000447	fx1	5	2:12:50 PM	25	0.10 kUA/l	0	Negative	
011990000447	fx2	5	2:13:50 PM	14	Low RU	0	Negative	
011990000447	fx22	5	2:14:50 PM	31	0.17 kUA/l	0	Negative	
011990000447	gx2	5	2:15:50 PM	24	0.09 kUA/l	0	Negative	
011990000447	wx1	5	2:17:51 PM	31	0.17 kUA/l	0	Negative	
011990000447	t20	5	2:18:50 PM	15	Low RU	0		
011990000448	fx5	5	1:28:52 PM	27	0.12 kUA/l	0	Negative	
011990000448	fx1	5	1:29:52 PM	18	0.01 kUA/l	0	Negative	
011990000448	fx2	5	1:30:52 PM	15	Low RU	0	Negative	
011990000448	fx22	5	1:31:52 PM	30	0.16 kUA/l	0	Negative	
011990000448	gx2	5	1:32:52 PM	24	0.08 kUA/l	0	Negative	
011990000448	wx1	5	1:33:52 PM	31	0.16 kUA/l	0	Negative	
011990000448	t20	5	1:34:52 PM	16	Low RU	0		
011990000457	fx5	5	12:40:53 PM	22	0.06 kUA/l	0	Negative	
011990000457	fx1	5	12:41:53 PM	22	0.06 kUA/l	0	Negative	
011990000457	fx2	5	12:42:53 PM	12	Low RU	0	Negative	
011990000457	fx22	5	12:43:54 PM	34	0.20 kUA/l	0	Negative	
011990000457	gx2	5	12:44:54 PM	21	0.05 kUA/l	0	Negative	
011990000457	wx1	5	12:45:53 PM	223	2.39 kUA/l	2	Positive	
011990000457	t20	5	12:46:53 PM	47	0.36 kUA/l	1		
011990000458	fx5	5	12:00:55 PM	26	0.11 kUA/l	0	Negative	
011990000458	fx1	5	12:01:55 PM	23	0.08 kUA/l	0	Negative	
011990000458	fx2	5	12:02:55 PM	13	Low RU	0	Negative	
011990000458	fx22	5	12:03:55 PM	33	0.19 kUA/l	0	Negative	
011990000458	gx2	5	12:04:55 PM	75	0.69 kUA/l	1	Positive	
011990000458	wx1	5	12:05:55 PM	39	0.27 kUA/l	0	Negative	
011990000458	t20	5	12:06:55 PM	20	0.04 kUA/l	0		
011990000459	fx5	5	2:00:51 PM	84	0.78 kUA/l	2	Positive	
011990000459	fx1	5	2:01:51 PM	38	0.25 kUA/l	0	Negative	
011990000459	fx2	5	2:02:51 PM	19	0.03 kUA/l	0	Negative	
011990000459	fx22	5	2:03:51 PM	40	0.28 kUA/l	0	Negative	
011990000459	gx2	5	2:04:51 PM	5584	75.4 kUA/l	5	Positive	
011990000459	wx1	5	2:05:51 PM	78	0.72 kUA/l	2	Positive	
011990000459	t20	5	2:06:50 PM	28	0.13 kUA/l	0		
011990000460	fx5	5	1:20:52 PM	43	0.31 kUA/l	0	Negative	
011990000460	fx1	5	1:21:52 PM	27	0.12 kUA/l	0	Negative	
011990000460	fx2	5	1:22:52 PM	98	0.96 kUA/l	2	Positive	
011990000460	fx22	5	1:23:52 PM	30	0.16 kUA/l	0	Negative	
011990000460	gx2	5	1:24:52 PM	30	0.15 kUA/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000460	wx1	5	1:25:52 PM	55	0.45 kUA/I	1	Positive	
011990000460	t20	5	1:26:52 PM	20	0.04 kUA/I	0		
012002200383A	f20	1	3:05:48 PM	26	0.02 kUA/I			
012002200383A	m6	1	3:06:49 PM	602	1.36 kUA/I			
012002200383A	m3	1	3:07:48 PM	523	1.18 kUA/I			
012002200383A	f92	1	3:08:48 PM	86	0.16 kUA/I			
012002200383A	g2	1	3:09:48 PM	49	0.08 kUA/I			
012002200383A	f18	1	3:10:48 PM	22	0.01 kUA/I			
012002200383A	f202	1	3:11:48 PM	23	0.01 kUA/I			
012002200383A	e1	1	3:12:48 PM	25	0.02 kUA/I			
012002200383A	f309	1	3:13:48 PM	46	0.07 kUA/I			
012002200383A	e85	1	3:14:48 PM	37	0.05 kUA/I			
012002200383A	f36	1	3:15:48 PM	70	0.13 kUA/I			
012002200383A	w1	1	3:16:48 PM	70	0.13 kUA/I			
012002200383A	e5	1	3:17:48 PM	48	0.07 kUA/I			
012002200383A	e86	1	3:18:48 PM	21	0.01 kUA/I			
012002200383A	f1	1	3:19:48 PM	41	0.06 kUA/I			
012002200383A	f17	1	3:20:48 PM	73	0.13 kUA/I			
012002200383A	e3	1	3:21:48 PM	31	0.03 kUA/I			
012002200383A	hx2	1	3:22:48 PM	20633			Positive	
012002200978A	f20	1	2:43:49 PM	36	0.04 kUA/I			
012002200978A	f202	1	2:44:49 PM	68	0.12 kUA/I			
012002200978A	e1	1	2:45:49 PM	13	0.00 kUA/I			
012002200978A	d2	1	2:46:49 PM	23	0.01 kUA/I			
012002200978A	e5	1	2:47:49 PM	18	0.00 kUA/I			
012002200978A	f1	1	2:48:49 PM	63	0.11 kUA/I			
012002200978A	f17	1	2:49:49 PM	83	0.16 kUA/I			
012002200978A	d1	1	2:50:49 PM	25	0.02 kUA/I			
012002200978A	f2	1	2:51:49 PM	36	0.04 kUA/I			
012002200978A	f203	1	2:52:49 PM	82	0.15 kUA/I			
012002200978A	f10	1	2:53:49 PM	106	0.21 kUA/I			
012002200978A	f256	1	2:54:49 PM	67	0.12 kUA/I			
012002200978A	f4	1	2:55:49 PM	71	0.13 kUA/I			
012002200978A	f13	1	2:56:49 PM	24	0.02 kUA/I			
012002200978A	ex72	1	2:57:49 PM	20			Negative	
012002300444A	f245	1	2:58:49 PM	16	0.00 kUA/I			
012002300444A	f79	1	2:59:49 PM	15	0.00 kUA/I			
012002300444A	f2	1	3:00:49 PM	43	0.06 kUA/I			
012002300444A	f7	1	3:01:48 PM	42	0.06 kUA/I			
012002300444A	f24	1	3:02:48 PM	15	0.00 kUA/I			
012002300444A	f14	1	3:03:48 PM	18	0.00 kUA/I			
012002300444A	f4	1	3:04:49 PM	33	0.04 kUA/I			

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_lgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_lgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_lgE	10:56:57 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_lgE	10:57:57 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	11:33:56 AM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
slgE L	e1	11:36:56 AM	359EP 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06!		Exp
slgE Neg	d1	11:37:56 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12!		Exp
IQC271119	d1	11:38:56 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002200383A	f20	3:05:48 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	m6	3:06:49 PM	356FF 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	m3	3:07:48 PM	368KB 21/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f92	3:08:48 PM	561KS 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	g2	3:09:48 PM	379CG 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f18	3:10:48 PM	421BJ 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f202	3:11:48 PM	997KA 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e1	3:12:48 PM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f309	3:13:48 PM	BTSG9 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e85	3:14:48 PM	517GH 21/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f36	3:15:48 PM	453BD 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	w1	3:16:48 PM	DWREC 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e5	3:17:48 PM	360F3 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e86	3:18:48 PM	518GE 20/01!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			Exp
012002200383A	f1	3:19:48 PM	361EN 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f17	3:20:48 PM	D2FBX 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e3	3:21:48 PM	371BH 20/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	hx2	3:22:48 PM	446BR 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200978A	f20	2:43:49 PM	422CS 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	f202	2:44:49 PM	997KA 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	e1	2:45:49 PM	359EP 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	d2	2:46:49 PM	358E2 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	e5	2:47:49 PM	360F3 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	f1	2:48:49 PM	361EN 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	f17	2:49:49 PM	D2FBX 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	d1	2:50:49 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002200978A	f2	2:51:49 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200978A	f203	2:52:49 PM	998HG 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200978A	f10	2:53:49 PM	418DE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200978A	f256	2:54:49 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200978A	f4	2:55:49 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200978A	f13	2:56:49 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300444A	f245	2:58:49 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300444A	f79	2:59:49 PM	525JR 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300444A	f2	3:00:49 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300444A	f7	3:01:48 PM	416BC 20/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300444A	f24	3:02:48 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300444A	f14	3:03:48 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300444A	f4	3:04:49 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
011990000460	phad	1:19:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000460	fx5	1:20:52 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000460	fx1	1:21:52 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000460	fx2	1:22:52 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000460	fx22	1:23:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000460	gx2	1:24:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000460	wx1	1:25:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000460	t20	1:26:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000448	phad	1:27:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000448	fx5	1:28:52 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000448	fx1	1:29:52 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000448	fx2	1:30:52 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000448	fx22	1:31:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000448	gx2	1:32:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000448	wx1	1:33:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000448	t20	1:34:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	phad	1:35:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	fx5	1:36:52 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	fx1	1:37:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	fx2	1:38:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	fx22	1:39:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	gx2	1:40:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	wx1	1:41:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000436	t20	1:42:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	phad	1:43:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	fx5	1:44:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	fx1	1:45:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	fx2	1:46:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	fx22	1:47:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	gx2	1:48:51 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	wx1	1:49:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000424	t20	1:50:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	phad	1:51:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	fx5	1:52:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	fx1	1:53:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	fx2	1:54:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	fx22	1:55:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	gx2	1:56:51 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	wx1	1:57:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000412	t20	1:58:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	phad	1:59:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	fx5	2:00:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	fx1	2:01:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	fx2	2:02:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	fx22	2:03:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	gx2	2:04:51 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	wx1	2:05:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000459	t20	2:06:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	phad	2:10:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	fx5	2:11:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	fx1	2:12:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	fx2	2:13:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	fx22	2:14:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	gx2	2:15:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	wx1	2:17:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000447	t20	2:18:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	phad	2:19:50 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	fx5	2:20:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	fx1	2:21:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	fx2	2:22:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	fx22	2:23:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	gx2	2:24:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	wx1	2:25:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000435	t20	2:26:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	phad	2:27:50 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	fx5	2:28:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	fx1	2:29:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	fx2	2:30:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	fx22	2:31:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	gx2	2:32:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	wx1	2:33:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000423	t20	2:34:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000411	phad	2:35:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	
011990000411	fx5	2:36:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000411	fx1	2:37:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000411	fx2	2:38:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000411	fx22	2:39:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000411	gx2	2:40:49 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000411	wx1	2:41:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000411	t20	2:42:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	phad	11:59:55 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	fx5	12:00:55 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	fx1	12:01:55 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	fx2	12:02:55 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	fx22	12:03:55 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	gx2	12:04:55 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	wx1	12:05:55 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000458	t20	12:06:55 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	phad	12:07:55 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	fx5	12:08:54 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	fx1	12:09:55 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	fx2	12:10:55 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	fx22	12:11:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	gx2	12:12:54 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	wx1	12:13:54 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000446	t20	12:14:54 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	phad	12:15:55 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	fx5	12:16:54 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	fx1	12:17:54 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	fx2	12:18:54 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	fx22	12:19:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	gx2	12:20:54 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	wx1	12:21:54 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000434	t20	12:22:54 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	phad	12:23:54 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	fx5	12:24:54 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	fx1	12:25:54 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	fx2	12:26:54 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	fx22	12:27:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	gx2	12:28:54 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	wx1	12:29:54 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000422	t20	12:30:54 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	phad	12:31:54 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	fx5	12:32:54 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	fx1	12:33:54 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	fx2	12:34:54 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	fx22	12:35:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	gx2	12:36:54 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	wx1	12:37:53 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000410	t20	12:38:54 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	phad	12:39:54 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	fx5	12:40:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	fx1	12:41:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	fx2	12:42:53 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	fx22	12:43:54 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	gx2	12:44:54 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	wx1	12:45:53 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000457	t20	12:46:53 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	phad	12:47:53 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	fx5	12:48:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	fx1	12:49:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	fx2	12:50:53 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	fx22	12:51:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	gx2	12:52:54 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	wx1	12:53:53 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000445	t20	12:54:53 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000433	phad	12:55:53 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000433	fx5	12:56:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000433	fx1	12:57:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000433	fx2	12:58:53 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000433	fx22	12:59:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000433	gx2	1:00:53 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000433	wx1	1:01:53 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000433	t20	1:02:53 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	phad	1:03:53 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	fx5	1:04:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	fx1	1:05:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	fx2	1:06:53 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	fx22	1:07:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	gx2	1:08:53 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	wx1	1:09:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000421	t20	1:10:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	phad	1:11:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	fx5	1:12:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	fx1	1:13:52 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	fx2	1:14:52 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	fx22	1:15:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	gx2	1:16:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	wx1	1:17:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000409	t20	1:18:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
012002200978A	ex72	2:57:49 PM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			

Errors

Expired lot numbers used for Sample

518GE ImmunoCAP

Expired lot numbers used for Quality Control

352D0 ImmunoCAP
CZ4AN Quality control vial
BL1CP Quality control vial


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200126-386
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 26-Jan-20 3:23:48 PM
Ended: 26-Jan-20 6:49:42 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

Curve Controls

No curve controls in this run

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	5:15:43 PM	5601	15.1 kUA/l	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000405	phad	5	6:35:41 PM	33	0.19 PAU/l	0	Negative	
011990000406	phad	5	5:55:42 PM	38	0.26 PAU/l	0	Negative	
011990000408	phad	5	5:16:44 PM	263	2.86 PAU/l	2	Positive	
011990000417	phad	5	6:27:41 PM	74	0.67 PAU/l	1	Positive	
011990000418	phad	5	5:47:43 PM	32	0.18 PAU/l	0	Negative	
011990000420	phad	5	5:07:44 PM	118	1.19 PAU/l	2	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000429	phad	5	6:19:41 PM	118	1.18 PAU/I	2	Positive	
011990000430	phad	5	5:39:43 PM	205	2.19 PAU/I	2	Positive	
011990000432	phad	5	4:59:44 PM	55	0.45 PAU/I	1	Positive	
011990000441	phad	5	6:11:42 PM	35	0.22 PAU/I	0	Negative	
011990000442	phad	5	5:31:43 PM	36	0.23 PAU/I	0	Negative	
011990000444	phad	5	4:51:44 PM	39	0.26 PAU/I	0	Negative	
011990000453	phad	5	6:03:42 PM	684	7.77 PAU/I	3	Positive	
011990000454	phad	5	5:19:44 PM	34	0.20 PAU/I	0	Negative	
011990000455	phad	5	6:48:40 PM	33	0.19 PAU/I	0	Negative	
011990000456	phad	5	4:43:45 PM	34	0.20 PAU/I	0	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000405	fx5	5	6:36:41 PM	20	0.04 kUA/I	0	Negative	
011990000405	fx1	5	6:37:41 PM	17	Low RU	0	Negative	
011990000405	fx2	5	6:38:41 PM	13	Low RU	0	Negative	
011990000405	fx22	5	6:39:41 PM	31	0.17 kUA/I	0	Negative	
011990000405	gx2	5	6:40:41 PM	22	0.06 kUA/I	0	Negative	
011990000405	wx1	5	6:41:41 PM	80	0.74 kUA/I	2	Positive	
011990000405	t20	5	6:42:41 PM	17	Low RU	0		
011990000406	fx5	5	5:56:42 PM	22	0.06 kUA/I	0	Negative	
011990000406	fx1	5	5:57:42 PM	19	0.03 kUA/I	0	Negative	
011990000406	fx2	5	5:58:42 PM	16	Low RU	0	Negative	
011990000406	fx22	5	5:59:42 PM	31	0.17 kUA/I	0	Negative	
011990000406	gx2	5	6:00:42 PM	26	0.11 kUA/I	0	Negative	
011990000406	wx1	5	6:01:42 PM	33	0.19 kUA/I	0	Negative	
011990000406	t20	5	6:02:42 PM	17	Low RU	0		
011990000408	fx5	5	5:17:44 PM	20	0.03 kUA/I	0	Negative	
011990000408	fx1	5	5:18:44 PM	17	Low RU	0	Negative	
011990000408	fx2	5	6:43:41 PM	13	Low RU	0	Negative	
011990000408	fx22	5	6:44:41 PM	32	0.18 kUA/I	0	Negative	
011990000408	gx2	5	6:45:40 PM	23	0.07 kUA/I	0	Negative	
011990000408	wx1	5	6:46:41 PM	29	0.14 kUA/I	0	Negative	
011990000408	t20	5	6:47:40 PM	16	Low RU	0		
011990000417	fx5	5	6:28:41 PM	21	0.05 kUA/I	0	Negative	
011990000417	fx1	5	6:29:41 PM	17	Low RU	0	Negative	
011990000417	fx2	5	6:30:41 PM	37	0.24 kUA/I	0	Negative	
011990000417	fx22	5	6:31:41 PM	30	0.15 kUA/I	0	Negative	
011990000417	gx2	5	6:32:41 PM	25	0.10 kUA/I	0	Negative	
011990000417	wx1	5	6:33:41 PM	37	0.23 kUA/I	0	Negative	
011990000417	t20	5	6:34:41 PM	20	0.03 kUA/I	0		
011990000418	fx5	5	5:48:43 PM	24	0.09 kUA/I	0	Negative	
011990000418	fx1	5	5:49:42 PM	18	0.01 kUA/I	0	Negative	
011990000418	fx2	5	5:50:42 PM	12	Low RU	0	Negative	
011990000418	fx22	5	5:51:42 PM	28	0.13 kUA/I	0	Negative	
011990000418	gx2	5	5:52:42 PM	22	0.06 kUA/I	0	Negative	
011990000418	wx1	5	5:53:42 PM	34	0.20 kUA/I	0	Negative	
011990000418	t20	5	5:54:43 PM	16	Low RU	0		
011990000420	fx5	5	5:08:44 PM	47	0.35 kUA/I	1	Positive	
011990000420	fx1	5	5:09:44 PM	27	0.12 kUA/I	0	Negative	
011990000420	fx2	5	5:10:44 PM	50	0.39 kUA/I	1	Positive	
011990000420	fx22	5	5:11:44 PM	46	0.34 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000420	gx2	5	5:12:44 PM	41	0.29 kUA/l	0	Negative	
011990000420	wx1	5	5:13:44 PM	49	0.38 kUA/l	1	Positive	
011990000420	t20	5	5:14:44 PM	29	0.14 kUA/l	0		
011990000429	fx5	5	6:20:41 PM	69	0.62 kUA/l	1	Positive	
011990000429	fx1	5	6:21:41 PM	114	1.14 kUA/l	2	Positive	
011990000429	fx2	5	6:22:42 PM	78	0.72 kUA/l	2	Positive	
011990000429	fx22	5	6:23:41 PM	128	1.29 kUA/l	2	Positive	
011990000429	gx2	5	6:24:41 PM	43	0.31 kUA/l	0	Negative	
011990000429	wx1	5	6:25:41 PM	66	0.58 kUA/l	1	Positive	
011990000429	t20	5	6:26:41 PM	84	0.79 kUA/l	2		
011990000430	fx5	5	5:40:43 PM	31	0.17 kUA/l	0	Negative	
011990000430	fx1	5	5:41:43 PM	24	0.09 kUA/l	0	Negative	
011990000430	fx2	5	5:42:43 PM	169	1.77 kUA/l	2	Positive	
011990000430	fx22	5	5:43:43 PM	43	0.30 kUA/l	0	Negative	
011990000430	gx2	5	5:44:43 PM	41	0.28 kUA/l	0	Negative	
011990000430	wx1	5	5:45:43 PM	92	0.88 kUA/l	2	Positive	
011990000430	t20	5	5:46:43 PM	103	1.01 kUA/l	2		
011990000432	fx5	5	5:00:44 PM	23	0.07 kUA/l	0	Negative	
011990000432	fx1	5	5:01:44 PM	18	0.02 kUA/l	0	Negative	
011990000432	fx2	5	5:02:44 PM	17	Low RU	0	Negative	
011990000432	fx22	5	5:03:44 PM	34	0.21 kUA/l	0	Negative	
011990000432	gx2	5	5:04:44 PM	23	0.08 kUA/l	0	Negative	
011990000432	wx1	5	5:05:44 PM	140	1.43 kUA/l	2	Positive	
011990000432	t20	5	5:06:44 PM	96	0.92 kUA/l	2		
011990000441	fx5	5	6:12:42 PM	19	0.03 kUA/l	0	Negative	
011990000441	fx1	5	6:13:42 PM	18	0.01 kUA/l	0	Negative	
011990000441	fx2	5	6:14:42 PM	13	Low RU	0	Negative	
011990000441	fx22	5	6:15:42 PM	30	0.16 kUA/l	0	Negative	
011990000441	gx2	5	6:16:42 PM	31	0.16 kUA/l	0	Negative	
011990000441	wx1	5	6:17:41 PM	29	0.15 kUA/l	0	Negative	
011990000441	t20	5	6:18:41 PM	17	Low RU	0		
011990000442	fx5	5	5:32:43 PM	25	0.10 kUA/l	0	Negative	
011990000442	fx1	5	5:33:43 PM	16	Low RU	0	Negative	
011990000442	fx2	5	5:34:43 PM	14	Low RU	0	Negative	
011990000442	fx22	5	5:35:43 PM	32	0.18 kUA/l	0	Negative	
011990000442	gx2	5	5:36:43 PM	23	0.08 kUA/l	0	Negative	
011990000442	wx1	5	5:37:43 PM	34	0.20 kUA/l	0	Negative	
011990000442	t20	5	5:38:43 PM	17	Low RU	0		
011990000444	fx5	5	4:52:44 PM	28	0.14 kUA/l	0	Negative	
011990000444	fx1	5	4:53:44 PM	18	0.02 kUA/l	0	Negative	
011990000444	fx2	5	4:54:44 PM	13	Low RU	0	Negative	
011990000444	fx22	5	4:55:44 PM	28	0.13 kUA/l	0	Negative	
011990000444	gx2	5	4:56:44 PM	25	0.10 kUA/l	0	Negative	
011990000444	wx1	5	4:57:44 PM	28	0.13 kUA/l	0	Negative	
011990000444	t20	5	4:58:44 PM	17	0.01 kUA/l	0		
011990000453	fx5	5	6:04:42 PM	42	0.29 kUA/l	0	Negative	
011990000453	fx1	5	6:05:42 PM	28	0.13 kUA/l	0	Negative	
011990000453	fx2	5	6:06:42 PM	593	6.71 kUA/l	3	Positive	
011990000453	fx22	5	6:07:42 PM	47	0.35 kUA/l	1	Positive	
011990000453	gx2	5	6:08:42 PM	96	0.93 kUA/l	2	Positive	
011990000453	wx1	5	6:09:42 PM	238	2.57 kUA/l	2	Positive	
011990000453	t20	5	6:10:42 PM	142	1.46 kUA/l	2		
011990000454	fx5	5	5:20:43 PM	21	0.05 kUA/l	0	Negative	
011990000454	fx1	5	5:21:43 PM	16	Low RU	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000454	fx2	5	5:22:43 PM	13	Low RU	0	Negative	
011990000454	fx22	5	5:23:43 PM	32	0.18 kUA/I	0	Negative	
011990000454	gx2	5	5:24:43 PM	25	0.09 kUA/I	0	Negative	
011990000454	wx1	5	5:25:43 PM	34	0.21 kUA/I	0	Negative	
011990000454	t20	5	5:30:43 PM	16	Low RU	0		
011990000456	fx5	5	4:44:45 PM	21	0.05 kUA/I	0	Negative	
011990000456	fx1	5	4:45:45 PM	22	0.06 kUA/I	0	Negative	
011990000456	fx2	5	4:46:44 PM	15	Low RU	0	Negative	
011990000456	fx22	5	4:47:44 PM	29	0.14 kUA/I	0	Negative	
011990000456	gx2	5	4:48:45 PM	23	0.07 kUA/I	0	Negative	
011990000456	wx1	5	4:49:45 PM	29	0.14 kUA/I	0	Negative	
011990000456	t20	5	4:50:45 PM	16	Low RU	0		
012002200383A	k82	1	3:23:48 PM	25	0.02 kUA/I			
012002200383A	f235	1	3:24:48 PM	62	0.11 kUA/I			
012002200383A	f345	1	3:25:48 PM	39	0.05 kUA/I			
012002200383A	f91	1	3:26:48 PM	55	0.09 kUA/I			
012002200383A	e213	1	3:27:47 PM	57	0.10 kUA/I			
012002200383A	e215	1	3:28:48 PM	32	0.04 kUA/I			
012002200383A	f203	1	3:29:47 PM	55	0.09 kUA/I			
012002200383A	f41	1	3:30:47 PM	22	0.01 kUA/I			
012002200383A	f10	1	3:31:47 PM	91	0.17 kUA/I			
012002200383A	f24	1	3:32:47 PM	53	0.09 kUA/I			
012002200383A	f14	1	3:33:48 PM	26	0.02 kUA/I			
012002200383A	f44	1	3:34:48 PM	40	0.05 kUA/I			
012002200383A	f40	1	3:35:47 PM	29	0.03 kUA/I			
012002200383A	f256	1	3:36:47 PM	47	0.07 kUA/I			
012002200383A	ex72	1	3:37:47 PM	193			Positive	
012002200383A	mx2	1	3:38:47 PM	515			Positive	
012002200383A	gx2	1	3:39:47 PM	64	0.11 kUA/I	0	Negative	
012002200383A	f13	1	3:40:47 PM	44	0.06 kUA/I			
012002200383A	e196	1	3:41:47 PM	294	0.64 kUA/I			
012002300363A	f20	1	4:04:46 PM	13	0.00 kUA/I			
012002300363A	f92	1	4:05:46 PM	18	0.00 kUA/I			
012002300363A	f18	1	4:06:46 PM	9	0.00 kUA/I			
012002300363A	f245	1	4:07:46 PM	10	0.00 kUA/I			
012002300363A	f1	1	4:08:46 PM	11	0.00 kUA/I			
012002300363A	f75	1	4:09:46 PM	13	0.00 kUA/I			
012002300363A	f79	1	4:10:46 PM	16	0.00 kUA/I			
012002300363A	f17	1	4:11:46 PM	11	0.00 kUA/I			
012002300363A	f84	1	4:12:46 PM	10	0.00 kUA/I			
012002300363A	f208	1	4:13:46 PM	17	0.00 kUA/I			
012002300363A	f2	1	4:14:46 PM	15	0.00 kUA/I			
012002300363A	f24	1	4:15:46 PM	14	0.00 kUA/I			
012002300363A	f14	1	4:16:46 PM	11	0.00 kUA/I			
012002300363A	f44	1	4:17:45 PM	16	0.00 kUA/I			
012002300363A	f25	1	4:18:45 PM	12	0.00 kUA/I			
012002300363A	f4	1	4:19:45 PM	18	0.00 kUA/I			
012002300363A	f13	1	4:20:46 PM	15	0.00 kUA/I			
012002300534A	f20	1	3:48:47 PM	13	0.00 kUA/I			
012002300534A	f18	1	3:49:47 PM	10	0.00 kUA/I			
012002300534A	e1	1	3:50:46 PM	10	0.00 kUA/I			
012002300534A	f245	1	3:51:47 PM	13	0.00 kUA/I			
012002300534A	f1	1	3:52:47 PM	16	0.00 kUA/I			
012002300534A	f75	1	3:53:47 PM	15	0.00 kUA/I			

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
012002300534A	f79	1	3:54:46 PM	14	0.00	KUA/I	Positive	
012002300534A	f17	1	3:55:46 PM	12	0.00	KUA/I		
012002300534A	hx2	1	3:56:46 PM	1914				
012002300534A	d1	1	3:57:46 PM	2146	5.15	KUA/I		
012002300534A	f41	1	3:58:46 PM	10	0.00	KUA/I		
012002300534A	f24	1	3:59:46 PM	15	0.00	KUA/I		
012002300534A	f40	1	4:00:46 PM	12	0.00	KUA/I		
012002300534A	f256	1	4:01:47 PM	14	0.00	KUA/I		
012002300534A	f4	1	4:02:47 PM	17	0.00	KUA/I		
012002300534A	f13	1	4:03:46 PM	14	0.00	KUA/I		
012002300828A	f245	1	3:42:47 PM	20	0.01	KUA/I		
012002300828A	f2	1	3:43:47 PM	78	0.14	KUA/I		
012002300828A	f24	1	3:44:47 PM	15	0.00	KUA/I		
012002300828A	f14	1	3:45:47 PM	14	0.00	KUA/I		
012002300828A	f4	1	3:46:47 PM	35	0.04	KUA/I		
012002300828A	f13	1	3:47:47 PM	19	0.01	KUA/I		
012002300840A	f245	1	4:21:46 PM	27	0.02	KUA/I		
012002300840A	f2	1	4:22:46 PM	27	0.02	KUA/I		
012002300840A	f24	1	4:23:45 PM	17	0.00	KUA/I		
012002300840A	f14	1	4:24:45 PM	14	0.00	KUA/I		
012002300840A	f4	1	4:25:45 PM	20	0.01	KUA/I		
012002300840A	f13	1	4:26:45 PM	19	0.01	KUA/I		
012002300896A	f20	1	4:27:45 PM	14	0.00	KUA/I		
012002300896A	f18	1	4:28:45 PM	14	0.00	KUA/I		
012002300896A	f78	1	4:29:45 PM	16	0.00	KUA/I		
012002300896A	f202	1	4:30:45 PM	17	0.00	KUA/I		
012002300896A	f1	1	4:31:45 PM	139	0.28	KUA/I		
012002300896A	f17	1	4:32:45 PM	18	0.00	KUA/I		
012002300896A	f345	1	4:33:45 PM	33	0.04	KUA/I		
012002300896A	f2	1	4:34:45 PM	123	0.25	KUA/I		
012002300896A	f233	1	4:35:45 PM	13	0.00	KUA/I		
012002300896A	f13	1	4:36:45 PM	19	0.01	KUA/I		
012002300896A	f203	1	4:37:45 PM	25	0.02	KUA/I		
012002300896A	f10	1	4:38:45 PM	27	0.02	KUA/I		
012002300896A	f24	1	4:39:45 PM	16	0.00	KUA/I		
012002300896A	f256	1	4:41:45 PM	13	0.00	KUA/I		
012002300896A	f4	1	4:42:45 PM	33	0.04	KUA/I		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
-----------------	------	----------	----------	-----------	---------	----------	----------	-----------	---------	---------

Laboratory Report

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	5:15:43 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002200383A	k82	3:23:48 PM	BJ1KP 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f235	3:24:48 PM	914GA 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f345	3:25:48 PM	BTXGP 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f91	3:26:48 PM	559GG 20/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e213	3:27:47 PM	994GD 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e215	3:28:48 PM	BP5GE 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f203	3:29:47 PM	998HG 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f41	3:30:47 PM	456CN 20/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f10	3:31:47 PM	418DE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f24	3:32:47 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f14	3:33:48 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f44	3:34:48 PM	457BA 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f40	3:35:47 PM	455DT 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f256	3:36:47 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	ex72	3:37:47 PM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	mx2	3:38:47 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	gx2	3:39:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	f13	3:40:47 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002200383A	e196	3:41:47 PM	BY1G5 19/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			Exp
012002300363A	f20	4:04:46 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f92	4:05:46 PM	561KS 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f18	4:06:46 PM	421BJ 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f245	4:07:46 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f1	4:08:46 PM	361EN 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f75	4:09:46 PM	427CE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f79	4:10:46 PM	525JR 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f17	4:11:46 PM	D2FBX 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f84	4:12:46 PM	529KC 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f208	4:13:46 PM	685GA 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f2	4:14:46 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f24	4:15:46 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f14	4:16:46 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f44	4:17:45 PM	457BA 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f25	4:18:45 PM	425BL 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f4	4:19:45 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300363A	f13	4:20:46 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f20	3:48:47 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f18	3:49:47 PM	421BJ 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	e1	3:50:46 PM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f245	3:51:47 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f1	3:52:47 PM	361EN 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f75	3:53:47 PM	427CE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f79	3:54:46 PM	525JR 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f17	3:55:46 PM	D2FBX 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	hx2	3:56:46 PM	446BR 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	d1	3:57:46 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f41	3:58:46 PM	456CN 20/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f24	3:59:46 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f40	4:00:46 PM	455DT 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f256	4:01:47 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f4	4:02:47 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300534A	f13	4:03:46 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300828A	f245	3:42:47 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300828A	f2	3:43:47 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300828A	f24	3:44:47 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300828A	f14	3:45:47 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300828A	f4	3:46:47 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300828A	f13	3:47:47 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300840A	f245	4:21:46 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300840A	f2	4:22:46 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300840A	f24	4:23:45 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300840A	f14	4:24:45 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002300840A	f4	4:25:45 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300840A	f13	4:26:45 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f20	4:27:45 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f18	4:28:45 PM	421BJ 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f78	4:29:45 PM	CTSB0 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f202	4:30:45 PM	997KA 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f1	4:31:45 PM	361EN 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f17	4:32:45 PM	D2FBX 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f345	4:33:45 PM	BTXGP 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f2	4:34:45 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f233	4:35:45 PM	904LM 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f13	4:36:45 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f203	4:37:45 PM	998HG 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f10	4:38:45 PM	418DE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f24	4:39:45 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f256	4:41:45 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002300896A	f4	4:42:45 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
011990000456	phad	4:43:45 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	fx5	4:44:45 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	fx1	4:45:45 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	fx2	4:46:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	fx22	4:47:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	gx2	4:48:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	wx1	4:49:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	t20	4:50:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000456	phad	4:51:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000444	fx5	4:52:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000444	fx1	4:53:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000444	fx2	4:54:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000444	fx22	4:55:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000444	gx2	4:56:44 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000444	wx1	4:57:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000444	t20	4:58:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	phad	4:59:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	fx5	5:00:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	fx1	5:01:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	fx2	5:02:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	fx22	5:03:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	gx2	5:04:44 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	wx1	5:05:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000432	t20	5:06:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	phad	5:07:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	fx5	5:08:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	fx1	5:09:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	fx2	5:10:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	fx22	5:11:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	gx2	5:12:44 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	wx1	5:13:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000420	t20	5:14:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	phad	5:16:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	fx5	5:17:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	fx1	5:18:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	fx2	6:43:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	fx22	6:44:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	gx2	6:45:40 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	wx1	6:46:41 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000408	t20	6:47:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000455	phad	6:48:40 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000454	phad	5:19:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000454	fx5	5:20:43 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000454	fx1	5:21:43 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000454	fx2	5:22:43 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000454	fx22	5:23:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000454	gx2	5:24:43 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000454	wx1	5:25:43 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

27-Jan-20 4:19:06 PM

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000405	phad	6:35:41 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000405	fx5	6:36:41 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000405	fx1	6:37:41 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000405	fx2	6:38:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000405	fx22	6:39:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000405	gx2	6:40:41 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000405	wx1	6:41:41 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000405	t20	6:42:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

Errors

Expired lot numbers used for Sample

BY1G5 ImmunoCAP


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200126-386
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 26-Jan-20 6:49:40 PM
Ended: 26-Jan-20 8:40:57 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

Curve Controls

No curve controls in this run

Quality Controls

No quality controls in this run

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000404	phad	5	7:44:38 PM	32	0.18 PAU/l	0	Negative	
011990000407	phad	5	7:20:39 PM	37	0.24 PAU/l	0	Negative	
011990000416	phad	5	8:00:38 PM	2054	24.6 PAU/l	4	Positive	
011990000419	phad	5	7:12:40 PM	3649	46.0 PAU/l	4	Positive	
011990000428	phad	5	7:52:38 PM	922	10.6 PAU/l	3	Positive	
011990000431	phad	5	7:04:40 PM	49	0.38 PAU/l	1	Positive	
011990000438	phad	5	8:30:37 PM	33	0.19 PAU/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000439	phad	5	8:16:37 PM	115	1.15 PAU/I	2	Positive	
011990000440	phad	5	7:36:39 PM	34	0.21 PAU/I	0	Negative	
011990000443	phad	5	6:56:40 PM	54	0.43 PAU/I	1	Positive	
011990000450	phad	5	8:23:36 PM	35	0.22 PAU/I	0	Negative	
011990000451	phad	5	8:08:37 PM	33	0.19 PAU/I	0	Negative	
011990000452	phad	5	7:28:40 PM	1664	19.6 PAU/I	4	Positive	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000404	fx5	5	7:45:38 PM	20	0.04 kUA/I	0	Negative	
011990000404	fx1	5	7:46:38 PM	17	0.01 kUA/I	0	Negative	
011990000404	fx2	5	7:47:38 PM	13	Low RU	0	Negative	
011990000404	fx22	5	7:48:38 PM	29	0.15 kUA/I	0	Negative	
011990000404	gx2	5	7:49:38 PM	20	0.04 kUA/I	0	Negative	
011990000404	wx1	5	7:50:38 PM	29	0.15 kUA/I	0	Negative	
011990000404	t20	5	7:51:38 PM	16	Low RU	0		
011990000407	fx5	5	7:21:39 PM	50	0.40 kUA/I	1	Positive	
011990000407	fx1	5	7:22:39 PM	18	0.02 kUA/I	0	Negative	
011990000407	fx2	5	7:23:39 PM	15	Low RU	0	Negative	
011990000407	fx22	5	7:24:39 PM	37	0.24 kUA/I	0	Negative	
011990000407	gx2	5	7:25:39 PM	35	0.22 kUA/I	0	Negative	
011990000407	wx1	5	7:26:39 PM	32	0.18 kUA/I	0	Negative	
011990000407	t20	5	7:27:39 PM	18	0.01 kUA/I	0		
011990000416	fx5	5	8:01:38 PM	61	0.52 kUA/I	1	Positive	
011990000416	fx1	5	8:02:38 PM	23	0.08 kUA/I	0	Negative	
011990000416	fx2	5	8:03:37 PM	14	Low RU	0	Negative	
011990000416	fx22	5	8:04:38 PM	34	0.21 kUA/I	0	Negative	
011990000416	gx2	5	8:05:37 PM	305	3.35 kUA/I	2	Positive	
011990000416	wx1	5	8:06:38 PM	45	0.33 kUA/I	0	Negative	
011990000416	t20	5	8:07:37 PM	20	0.04 kUA/I	0		
011990000419	fx5	5	7:13:39 PM	105	1.03 kUA/I	2	Positive	
011990000419	fx1	5	7:14:40 PM	64	0.56 kUA/I	1	Positive	
011990000419	fx2	5	7:15:39 PM	19	0.02 kUA/I	0	Negative	
011990000419	fx22	5	7:16:39 PM	119	1.19 kUA/I	2	Positive	
011990000419	gx2	5	7:17:39 PM	68	0.60 kUA/I	1	Positive	
011990000419	wx1	5	7:18:40 PM	716	8.15 kUA/I	3	Positive	
011990000419	t20	5	7:19:40 PM	195	2.08 kUA/I	2		
011990000428	fx5	5	7:53:38 PM	23	0.07 kUA/I	0	Negative	
011990000428	fx1	5	7:54:38 PM	16	Low RU	0	Negative	
011990000428	fx2	5	7:55:38 PM	13	Low RU	0	Negative	
011990000428	fx22	5	7:56:38 PM	28	0.13 kUA/I	0	Negative	
011990000428	gx2	5	7:57:38 PM	19	0.03 kUA/I	0	Negative	
011990000428	wx1	5	7:58:38 PM	30	0.16 kUA/I	0	Negative	
011990000428	t20	5	7:59:38 PM	16	Low RU	0		
011990000431	fx5	5	7:05:40 PM	21	0.05 kUA/I	0	Negative	
011990000431	fx1	5	7:06:40 PM	18	0.01 kUA/I	0	Negative	
011990000431	fx2	5	7:07:40 PM	28	0.13 kUA/I	0	Negative	
011990000431	fx22	5	7:08:40 PM	35	0.22 kUA/I	0	Negative	
011990000431	gx2	5	7:09:40 PM	23	0.07 kUA/I	0	Negative	
011990000431	wx1	5	7:10:40 PM	33	0.19 kUA/I	0	Negative	
011990000431	t20	5	7:11:40 PM	17	Low RU	0		
011990000438	fx5	5	8:32:37 PM	23	0.07 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000438	fx1	5	8:33:36 PM	19	0.02 kUA/l	0	Negative	
011990000439	fx5	5	8:17:37 PM	23	0.07 kUA/l	0	Negative	
011990000439	fx1	5	8:18:37 PM	19	0.03 kUA/l	0	Negative	
011990000439	fx22	5	8:19:37 PM	34	0.20 kUA/l	0	Negative	
011990000439	gx2	5	8:20:37 PM	43	0.31 kUA/l	0	Negative	
011990000439	wx1	5	8:21:37 PM	62	0.53 kUA/l	1	Positive	
011990000439	t20	5	8:22:37 PM	34	0.21 kUA/l	0		
011990000440	fx5	5	7:37:39 PM	21	0.05 kUA/l	0	Negative	
011990000440	fx1	5	7:38:39 PM	16	Low RU	0	Negative	
011990000440	fx2	5	7:39:39 PM	12	Low RU	0	Negative	
011990000440	fx22	5	7:40:38 PM	30	0.15 kUA/l	0	Negative	
011990000440	gx2	5	7:41:38 PM	27	0.12 kUA/l	0	Negative	
011990000440	wx1	5	7:42:38 PM	35	0.22 kUA/l	0	Negative	
011990000440	t20	5	7:43:38 PM	16	Low RU	0		
011990000443	fx5	5	6:57:40 PM	21	0.05 kUA/l	0	Negative	
011990000443	fx1	5	6:58:40 PM	15	Low RU	0	Negative	
011990000443	fx2	5	6:59:40 PM	22	0.06 kUA/l	0	Negative	
011990000443	fx22	5	7:00:40 PM	31	0.17 kUA/l	0	Negative	
011990000443	gx2	5	7:01:40 PM	24	0.09 kUA/l	0	Negative	
011990000443	wx1	5	7:02:40 PM	37	0.24 kUA/l	0	Negative	
011990000443	t20	5	7:03:40 PM	19	0.03 kUA/l	0		
011990000450	fx5	5	8:24:37 PM	20	0.04 kUA/l	0	Negative	
011990000450	fx1	5	8:25:37 PM	17	Low RU	0	Negative	
011990000450	fx22	5	8:26:37 PM	31	0.17 kUA/l	0	Negative	
011990000450	gx2	5	8:27:36 PM	27	0.13 kUA/l	0	Negative	
011990000450	wx1	5	8:28:37 PM	30	0.16 kUA/l	0	Negative	
011990000450	t20	5	8:29:36 PM	16	Low RU	0		
011990000451	fx5	5	8:09:37 PM	25	0.10 kUA/l	0	Negative	
011990000451	fx1	5	8:10:38 PM	18	0.02 kUA/l	0	Negative	
011990000451	fx2	5	8:11:37 PM	13	Low RU	0	Negative	
011990000451	fx22	5	8:12:38 PM	28	0.14 kUA/l	0	Negative	
011990000451	gx2	5	8:13:37 PM	25	0.10 kUA/l	0	Negative	
011990000451	wx1	5	8:14:37 PM	28	0.13 kUA/l	0	Negative	
011990000451	t20	5	8:15:37 PM	18	0.01 kUA/l	0		
011990000452	fx5	5	7:29:39 PM	159	1.66 kUA/l	2	Positive	
011990000452	fx1	5	7:30:39 PM	194	2.06 kUA/l	2	Positive	
011990000452	fx2	5	7:31:39 PM	15	Low RU	0	Negative	
011990000452	fx22	5	7:32:39 PM	306	3.36 kUA/l	2	Positive	
011990000452	gx2	5	7:33:39 PM	556	6.27 kUA/l	3	Positive	
011990000452	wx1	5	7:34:39 PM	1877	22.3 kUA/l	4	Positive	
011990000452	t20	5	7:35:39 PM	1992	23.8 kUA/l	4		
011990000455	fx5	5	6:49:40 PM	29	0.15 kUA/l	0	Negative	
011990000455	fx1	5	6:50:41 PM	31	0.16 kUA/l	0	Negative	
011990000455	fx2	5	6:51:41 PM	15	Low RU	0	Negative	
011990000455	fx22	5	6:52:41 PM	29	0.14 kUA/l	0	Negative	
011990000455	gx2	5	6:53:40 PM	22	0.06 kUA/l	0	Negative	
011990000455	wx1	5	6:54:40 PM	29	0.15 kUA/l	0	Negative	
011990000455	t20	5	6:55:40 PM	20	0.03 kUA/l	0		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000455	fx5	6:49:40 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000455	fx1	6:50:41 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000455	fx2	6:51:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000455	fx22	6:52:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000455	gx2	6:53:40 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000455	wx1	6:54:40 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000455	t20	6:55:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	phad	6:56:40 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	fx5	6:57:40 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	fx1	6:58:40 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	fx2	6:59:40 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	fx22	7:00:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	gx2	7:01:40 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	wx1	7:02:40 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000443	t20	7:03:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	phad	7:04:40 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	fx5	7:05:40 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	fx1	7:06:40 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	fx2	7:07:40 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	fx22	7:08:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	gx2	7:09:40 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	wx1	7:10:40 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000431	t20	7:11:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	phad	7:12:40 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	fx5	7:13:39 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	fx1	7:14:40 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	fx2	7:15:39 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	fx22	7:16:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	gx2	7:17:39 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	wx1	7:18:40 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000419	t20	7:19:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	phad	7:20:39 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	fx5	7:21:39 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	fx1	7:22:39 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	fx2	7:23:39 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	fx22	7:24:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	gx2	7:25:39 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	wx1	7:26:39 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000407	t20	7:27:39 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	phad	7:28:40 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	fx5	7:29:39 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	fx1	7:30:39 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	fx2	7:31:39 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	fx22	7:32:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	gx2	7:33:39 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	wx1	7:34:39 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000452	t20	7:35:39 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000440	phad	7:36:39 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000440	fx5	7:37:39 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000440	fx1	7:38:39 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000440	fx2	7:39:39 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000440	fx22	7:40:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000440	gx2	7:41:38 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000440	wx1	7:42:38 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000440	t20	7:43:38 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	phad	7:52:38 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	fx5	7:53:38 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	fx1	7:54:38 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	fx2	7:55:38 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	fx22	7:56:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	gx2	7:57:38 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	wx1	7:58:38 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000428	t20	7:59:38 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	phad	8:00:38 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	fx5	8:01:38 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	fx1	8:02:38 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	fx2	8:03:37 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	fx22	8:04:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	gx2	8:05:37 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	wx1	8:06:38 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000416	t20	8:07:37 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	phad	7:44:38 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	fx5	7:45:38 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	fx1	7:46:38 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	fx2	7:47:38 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	fx22	7:48:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	gx2	7:49:38 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	wx1	7:50:38 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000404	t20	7:51:38 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000451	phad	8:08:37 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000451	fx5	8:09:37 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000451	fx1	8:10:38 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000451	fx2	8:11:37 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000451	fx22	8:12:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000451	gx2	8:13:37 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000451	wx1	8:14:37 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000451	t20	8:15:37 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000439	phad	8:16:37 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000439	fx5	8:17:37 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000439	fx1	8:18:37 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000439	fx22	8:19:37 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000439	gx2	8:20:37 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000439	wx1	8:21:37 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000439	t20	8:22:37 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000450	phad	8:23:36 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000450	fx5	8:24:37 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000450	fx1	8:25:37 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000450	fx22	8:26:37 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000450	gx2	8:27:36 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000450	wx1	8:28:37 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000450	t20	8:29:36 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000438	phad	8:30:37 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000438	fx5	8:32:37 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	
011990000438	fx1	8:33:36 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BR 21/05	