Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200112-376
Instrument name:	Phadia-A
Instrument type:	Phadia 250
Status:	Ready
Started:	12-Jan-20 10:47:18 AM
Ended:	12-Jan-20 2:39:13 PM
Approved:	Approved (faziz)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-A
Instrument type:	Phadia 250
Measured:	05-Jan-20 1:26:01 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



thermo scientific

Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	10:47:18 AM	290	0.64	OK	
CC-2	17.5	10:48:18 AM	5862	16.0	OK	

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
sIgE M	t3	1	11:12:17 AM	3438	8.61 kUA/l	CZ1AR	OK	
sIgE L	e1	1	11:13:18 AM	1106	2.56 kUA/l	CZ4AN	OK	
sIgE Neg	d1	1	11:14:18 AM	22	0.01 kUA/l	BL1CP	OK	
IQC271119	d1	1	11:15:17 AM	6304	17.5 kUA/l	<lab>	----	
			Expected value:	---				
			Expected SD:	---				
			SD limit:	---				
			Range min:	---				
			Range max:	---				
			Limit CV%:	---				

thermo scientific

Laboratory Report

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000330	phad	5	1:21:13 PM	276	3.01 PAU/l	---	Positive	
011990000340	phad	5	1:13:13 PM	34	0.20 PAU/l	---	Negative	
011990000350	phad	5	2:38:11 PM	557	6.28 PAU/l	---	Positive	
011990000502	phad	5	2:30:11 PM	1605	18.9 PAU/l	---	Positive	
011990000511	phad	5	1:42:12 PM	82	0.77 PAU/l	---	Positive	
011990000512	phad	5	2:22:11 PM	38	0.24 PAU/l	---	Negative	
011990000521	phad	5	1:34:13 PM	56	0.46 PAU/l	---	Positive	
011990000531	phad	5	1:58:12 PM	82	0.76 PAU/l	---	Positive	
011990000532	phad	5	2:14:11 PM	243	2.63 PAU/l	---	Positive	
011990000541	phad	5	1:50:12 PM	79	0.72 PAU/l	---	Positive	
011990000542	phad	5	2:06:12 PM	75	0.68 PAU/l	---	Positive	
011990000729	phad	5	12:12:15 PM	32	0.18 PAU/l	---	Negative	
011990000730	phad	5	12:20:15 PM	491	5.51 PAU/l	---	Positive	
011990000734	phad	5	12:36:14 PM	81	0.75 PAU/l	---	Positive	
011990000735	phad	5	12:04:16 PM	35	0.21 PAU/l	---	Negative	
011990000739	phad	5	12:28:15 PM	37	0.23 PAU/l	---	Negative	
011990000744	phad	5	11:48:16 AM	44	0.32 PAU/l	---	Negative	
011990000745	phad	5	11:20:17 AM	33	0.19 PAU/l	---	Negative	
011990000748	phad	5	11:56:16 AM	39	0.26 PAU/l	---	Negative	
011990000749	phad	5	11:28:17 AM	50	0.39 PAU/l	---	Positive	
011990000750	phad	5	10:49:18 AM	34	0.20 PAU/l	---	Negative	
011990000760d	phad	5	11:36:16 AM	45	0.33 PAU/l	---	Negative	
011990000764	phad	5	1:05:13 PM	31	0.17 PAU/l	---	Negative	
011990000765	phad	5	12:55:14 PM	52	0.41 PAU/l	---	Positive	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000330	fx5	5	1:22:13 PM	44	0.33 kUA/l	---	Negative	
011990000330	fx1	5	1:23:13 PM	30	0.15 kUA/l	---	Negative	
011990000330	fx2	5	1:24:13 PM	24	0.09 kUA/l	---	Negative	
011990000330	fx22	5	1:25:13 PM	32	0.18 kUA/l	---	Negative	
011990000330	gx2	5	1:26:13 PM	76	0.70 kUA/l	---	Positive	
011990000330	wx1	5	1:27:13 PM	54	0.44 kUA/l	---	Positive	
011990000330	t20	5	1:33:12 PM	26	0.11 kUA/l	---		
011990000340	fx5	5	1:14:13 PM	23	0.07 kUA/l	---	Negative	
011990000340	fx1	5	1:15:13 PM	18	0.02 kUA/l	---	Negative	
011990000340	fx2	5	1:16:14 PM	16	---	---	Negative	
011990000340	fx22	5	1:17:13 PM	31	0.16 kUA/l	---	Negative	
011990000340	gx2	5	1:18:13 PM	21	0.06 kUA/l	---	Negative	
011990000340	wx1	5	1:19:13 PM	29	0.15 kUA/l	---	Negative	
011990000340	t20	5	1:20:13 PM	16	---	---		
011990000502	fx5	5	2:31:11 PM	75	0.68 kUA/l	---	Positive	
011990000502	fx1	5	2:32:11 PM	45	0.33 kUA/l	---	Negative	
011990000502	fx2	5	2:33:10 PM	34	0.21 kUA/l	---	Negative	
011990000502	fx22	5	2:34:11 PM	49	0.38 kUA/l	---	Positive	
011990000502	gx2	5	2:35:10 PM	51	0.40 kUA/l	---	Positive	
011990000502	wx1	5	2:36:11 PM	57	0.47 kUA/l	---	Positive	
011990000502	t20	5	2:37:10 PM	34	0.20 kUA/l	---		
011990000511	fx5	5	1:43:13 PM	39	0.26 kUA/l	---	Negative	
011990000511	fx1	5	1:44:12 PM	25	0.09 kUA/l	---	Negative	
011990000511	fx2	5	1:45:12 PM	30	0.15 kUA/l	---	Negative	
011990000511	fx22	5	1:46:12 PM	34	0.21 kUA/l	---	Negative	

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc		Class	Cut-Off	Remarks
011990000511	gx2	5	1:47:12 PM	28	0.13	kUA/l	---	Negative	
011990000511	wx1	5	1:48:12 PM	34	0.21	kUA/l	---	Negative	
011990000511	t20	5	1:49:12 PM	18	0.02	kUA/l	---		
011990000512	fx5	5	2:23:11 PM	27	0.12	kUA/l	---	Negative	
011990000512	fx1	5	2:24:11 PM	17	0.01	kUA/l	---	Negative	
011990000512	fx2	5	2:25:11 PM	13	---		---	Negative	
011990000512	fx22	5	2:26:11 PM	28	0.13	kUA/l	---	Negative	
011990000512	gx2	5	2:27:11 PM	36	0.23	kUA/l	---	Negative	
011990000512	wx1	5	2:28:11 PM	30	0.16	kUA/l	---	Negative	
011990000512	t20	5	2:29:10 PM	16	---		---		
011990000521	fx5	5	1:35:13 PM	23	0.08	kUA/l	---	Negative	
011990000521	fx1	5	1:36:12 PM	31	0.16	kUA/l	---	Negative	
011990000521	fx2	5	1:37:12 PM	12	---		---	Negative	
011990000521	fx22	5	1:38:12 PM	26	0.11	kUA/l	---	Negative	
011990000521	gx2	5	1:39:12 PM	21	0.05	kUA/l	---	Negative	
011990000521	wx1	5	1:40:12 PM	31	0.17	kUA/l	---	Negative	
011990000521	t20	5	1:41:12 PM	15	---		---		
011990000531	fx5	5	1:59:12 PM	38	0.25	kUA/l	---	Negative	
011990000531	fx1	5	2:00:12 PM	16	---		---	Negative	
011990000531	fx2	5	2:01:11 PM	45	0.33	kUA/l	---	Negative	
011990000531	fx22	5	2:02:12 PM	33	0.20	kUA/l	---	Negative	
011990000531	gx2	5	2:03:12 PM	24	0.08	kUA/l	---	Negative	
011990000531	wx1	5	2:04:12 PM	38	0.25	kUA/l	---	Negative	
011990000531	t20	5	2:05:11 PM	17	0.01	kUA/l	---		
011990000532	fx5	5	2:15:11 PM	43	0.31	kUA/l	---	Negative	
011990000532	fx1	5	2:16:12 PM	21	0.05	kUA/l	---	Negative	
011990000532	fx2	5	2:17:11 PM	12	---		---	Negative	
011990000532	fx22	5	2:18:12 PM	27	0.12	kUA/l	---	Negative	
011990000532	gx2	5	2:19:11 PM	132	1.34	kUA/l	---	Positive	
011990000532	wx1	5	2:20:12 PM	50	0.39	kUA/l	---	Positive	
011990000532	t20	5	2:21:11 PM	17	---		---		
011990000541	fx5	5	1:51:12 PM	35	0.21	kUA/l	---	Negative	
011990000541	fx1	5	1:52:12 PM	38	0.25	kUA/l	---	Negative	
011990000541	fx2	5	1:53:12 PM	38	0.25	kUA/l	---	Negative	
011990000541	fx22	5	1:54:12 PM	51	0.40	kUA/l	---	Positive	
011990000541	gx2	5	1:55:12 PM	47	0.36	kUA/l	---	Positive	
011990000541	wx1	5	1:56:12 PM	51	0.41	kUA/l	---	Positive	
011990000541	t20	5	1:57:12 PM	45	0.33	kUA/l	---		
011990000542	fx5	5	2:07:11 PM	21	0.05	kUA/l	---	Negative	
011990000542	fx1	5	2:08:12 PM	18	0.01	kUA/l	---	Negative	
011990000542	fx2	5	2:09:11 PM	61	0.53	kUA/l	---	Positive	
011990000542	fx22	5	2:10:12 PM	26	0.11	kUA/l	---	Negative	
011990000542	gx2	5	2:11:11 PM	19	0.02	kUA/l	---	Negative	
011990000542	wx1	5	2:12:12 PM	33	0.19	kUA/l	---	Negative	
011990000542	t20	5	2:13:11 PM	17	0.01	kUA/l	---		
011990000729	fx5	5	12:13:15 PM	24	0.08	kUA/l	---	Negative	
011990000729	fx1	5	12:14:15 PM	19	0.02	kUA/l	---	Negative	
011990000729	fx2	5	12:15:15 PM	13	---		---	Negative	
011990000729	fx22	5	12:16:15 PM	33	0.19	kUA/l	---	Negative	
011990000729	gx2	5	12:17:15 PM	20	0.04	kUA/l	---	Negative	
011990000729	wx1	5	12:18:15 PM	28	0.13	kUA/l	---	Negative	
011990000729	t20	5	12:19:15 PM	15	---		---		
011990000730	fx5	5	12:21:15 PM	25	0.10	kUA/l	---	Negative	
011990000730	fx1	5	12:22:15 PM	22	0.07	kUA/l	---	Negative	

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc		Class	Cut-Off	Remarks
011990000730	fx2	5	12:23:15 PM	94	0.90	kUA/l	---	Positive	
011990000730	fx22	5	12:24:15 PM	34	0.20	kUA/l	---	Negative	
011990000730	gx2	5	12:25:15 PM	23	0.08	kUA/l	---	Negative	
011990000730	wx1	5	12:26:15 PM	60	0.50	kUA/l	---	Positive	
011990000730	t20	5	12:27:15 PM	24	0.09	kUA/l	---		
011990000734	fx5	5	12:37:14 PM	19	0.02	kUA/l	---	Negative	
011990000734	fx1	5	12:38:15 PM	19	0.02	kUA/l	---	Negative	
011990000734	fx2	5	12:39:15 PM	17	---		---	Negative	
011990000734	fx22	5	12:40:14 PM	28	0.13	kUA/l	---	Negative	
011990000734	gx2	5	12:41:14 PM	18	0.02	kUA/l	---	Negative	
011990000734	wx1	5	12:42:14 PM	55	0.44	kUA/l	---	Positive	
011990000734	t20	5	12:43:14 PM	15	---		---		
011990000735	fx5	5	12:05:15 PM	25	0.10	kUA/l	---	Negative	
011990000735	fx1	5	12:06:16 PM	17	---		---	Negative	
011990000735	fx2	5	12:07:15 PM	15	---		---	Negative	
011990000735	fx22	5	12:08:15 PM	30	0.16	kUA/l	---	Negative	
011990000735	gx2	5	12:09:15 PM	22	0.07	kUA/l	---	Negative	
011990000735	wx1	5	12:10:16 PM	31	0.16	kUA/l	---	Negative	
011990000735	t20	5	12:11:15 PM	16	---		---		
011990000739	fx5	5	12:29:15 PM	23	0.07	kUA/l	---	Negative	
011990000739	fx1	5	12:30:15 PM	18	0.02	kUA/l	---	Negative	
011990000739	fx2	5	12:31:15 PM	15	---		---	Negative	
011990000739	fx22	5	12:32:14 PM	39	0.26	kUA/l	---	Negative	
011990000739	gx2	5	12:33:15 PM	22	0.07	kUA/l	---	Negative	
011990000739	wx1	5	12:34:15 PM	31	0.17	kUA/l	---	Negative	
011990000739	t20	5	12:35:14 PM	17	---		---		
011990000744	fx5	5	11:49:16 AM	23	0.08	kUA/l	---	Negative	
011990000744	fx1	5	11:50:16 AM	17	---		---	Negative	
011990000744	fx2	5	11:51:16 AM	23	0.07	kUA/l	---	Negative	
011990000744	fx22	5	11:52:16 AM	31	0.17	kUA/l	---	Negative	
011990000744	gx2	5	11:53:16 AM	26	0.11	kUA/l	---	Negative	
011990000744	wx1	5	11:54:16 AM	38	0.25	kUA/l	---	Negative	
011990000744	t20	5	11:55:16 AM	18	0.01	kUA/l	---		
011990000745	fx5	5	11:21:17 AM	21	0.05	kUA/l	---	Negative	
011990000745	fx1	5	11:22:17 AM	16	---		---	Negative	
011990000745	fx2	5	11:23:17 AM	14	---		---	Negative	
011990000745	fx22	5	11:24:17 AM	31	0.17	kUA/l	---	Negative	
011990000745	gx2	5	11:25:17 AM	24	0.09	kUA/l	---	Negative	
011990000745	wx1	5	11:26:17 AM	29	0.15	kUA/l	---	Negative	
011990000745	t20	5	11:27:17 AM	17	---		---		
011990000748	fx5	5	11:57:16 AM	28	0.13	kUA/l	---	Negative	
011990000748	fx1	5	11:58:16 AM	23	0.07	kUA/l	---	Negative	
011990000748	fx2	5	11:59:16 AM	18	0.01	kUA/l	---	Negative	
011990000748	fx22	5	12:00:16 PM	34	0.20	kUA/l	---	Negative	
011990000748	gx2	5	12:01:16 PM	21	0.05	kUA/l	---	Negative	
011990000748	wx1	5	12:02:16 PM	31	0.17	kUA/l	---	Negative	
011990000748	t20	5	12:03:16 PM	21	0.04	kUA/l	---		
011990000749	fx5	5	11:41:17 AM	20	0.04	kUA/l	---	Negative	
011990000749	fx1	5	11:42:17 AM	18	0.01	kUA/l	---	Negative	
011990000749	fx2	5	11:43:16 AM	24	0.08	kUA/l	---	Negative	
011990000749	fx22	5	11:44:16 AM	28	0.13	kUA/l	---	Negative	
011990000749	gx2	5	11:45:16 AM	22	0.06	kUA/l	---	Negative	
011990000749	wx1	5	11:46:16 AM	31	0.16	kUA/l	---	Negative	
011990000749	t20	5	11:47:16 AM	17	---		---		

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc		Class	Cut-Off	Remarks
011990000750	fx5	5	10:50:18 AM	28	0.13	kUA/l	---	Negative	
011990000750	fx1	5	10:51:18 AM	16	---		---	Negative	
011990000750	fx2	5	10:52:18 AM	17	---		---	Negative	
011990000750	fx22	5	11:16:17 AM	28	0.13	kUA/l	---	Negative	
011990000750	gx2	5	11:17:17 AM	19	0.03	kUA/l	---	Negative	
011990000750	wx1	5	11:18:17 AM	30	0.15	kUA/l	---	Negative	
011990000750	t20	5	11:19:17 AM	18	0.01	kUA/l	---		
011990000754	fx5	5	12:45:14 PM	24	0.09	kUA/l	---	Negative	
011990000754	fx22	5	12:46:14 PM	30	0.15	kUA/l	---	Negative	
011990000755	fx5	5	12:53:14 PM	21	0.05	kUA/l	---	Negative	
011990000755	fx22	5	12:54:14 PM	30	0.15	kUA/l	---	Negative	
011990000758	fx22	5	12:44:14 PM	32	0.18	kUA/l	---	Negative	
011990000759d	fx5	5	11:29:17 AM	19	0.03	kUA/l	---	Negative	
011990000759d	fx22	5	11:30:17 AM	28	0.13	kUA/l	---	Negative	
011990000759d	t20	5	11:31:17 AM	16	---		---		
011990000759d	wx1	5	11:32:17 AM	30	0.16	kUA/l	---	Negative	
011990000759d	gx2	5	11:33:17 AM	20	0.03	kUA/l	---	Negative	
011990000759d	fx2	5	11:34:17 AM	13	---		---	Negative	
011990000759d	fx1	5	11:35:16 AM	13	---		---	Negative	
011990000760d	fx5	5	11:37:17 AM	23	0.07	kUA/l	---	Negative	
011990000760d	fx22	5	11:38:16 AM	29	0.15	kUA/l	---	Negative	
011990000760d	fx2	5	11:39:16 AM	21	0.05	kUA/l	---	Negative	
011990000760d	fx1	5	11:40:16 AM	15	---		---	Negative	
011990000764	fx5	5	1:06:14 PM	24	0.09	kUA/l	---	Negative	
011990000764	fx1	5	1:07:14 PM	20	0.04	kUA/l	---	Negative	
011990000764	fx2	5	1:08:14 PM	12	---		---	Negative	
011990000764	fx22	5	1:09:13 PM	29	0.14	kUA/l	---	Negative	
011990000764	gx2	5	1:10:13 PM	18	0.02	kUA/l	---	Negative	
011990000764	wx1	5	1:11:13 PM	26	0.11	kUA/l	---	Negative	
011990000764	t20	5	1:12:13 PM	15	---		---		
011990000765	fx5	5	12:56:14 PM	23	0.07	kUA/l	---	Negative	
011990000765	fx1	5	12:57:14 PM	19	0.02	kUA/l	---	Negative	
011990000765	fx2	5	12:58:14 PM	20	0.03	kUA/l	---	Negative	
011990000765	fx22	5	12:59:14 PM	32	0.18	kUA/l	---	Negative	
011990000765	gx2	5	1:00:14 PM	23	0.07	kUA/l	---	Negative	
011990000765	wx1	5	1:01:14 PM	33	0.19	kUA/l	---	Negative	
011990000765	t20	5	1:04:14 PM	16	---		---		
011990000770	fx5	5	12:49:14 PM	25	0.10	kUA/l	---	Negative	
011990000770	fx22	5	12:50:14 PM	29	0.14	kUA/l	---	Negative	
011990000774	fx5	5	12:51:14 PM	18	0.02	kUA/l	---	Negative	
011990000774	fx22	5	12:52:14 PM	27	0.12	kUA/l	---	Negative	
011990000775	fx5	5	12:47:14 PM	71	0.64	kUA/l	---	Positive	
011990000775	fx22	5	12:48:14 PM	58	0.48	kUA/l	---	Positive	

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

thermo scientific

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	10:47:18 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	10:48:18 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
sIgE M	t3	11:12:17 AM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
sIgE L	e1	11:13:18 AM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06!		Exp
sIgE Neg	d1	11:14:18 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12!		Exp
IQC271119	d1	11:15:17 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000765	phad	12:55:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000765	fx5	12:56:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000765	fx1	12:57:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000765	fx2	12:58:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000765	fx22	12:59:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000765	gx2	1:00:14 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000765	wx1	1:01:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000765	t20	1:04:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	phad	1:05:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	fx5	1:06:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	fx1	1:07:14 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	fx2	1:08:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	fx22	1:09:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	gx2	1:10:13 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	wx1	1:11:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000764	t20	1:12:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000755	fx5	12:53:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000755	fx22	12:54:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000774	fx5	12:51:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000774	fx22	12:52:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000770	fx5	12:49:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000770	fx22	12:50:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000775	fx5	12:47:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000775	fx22	12:48:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000754	fx5	12:45:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000754	fx22	12:46:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	phad	12:36:14 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	fx5	12:37:14 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	fx1	12:38:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	fx2	12:39:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	fx22	12:40:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	gx2	12:41:14 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	wx1	12:42:14 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000734	t20	12:43:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	phad	12:28:15 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	fx5	12:29:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	fx1	12:30:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	fx2	12:31:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	fx22	12:32:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	gx2	12:33:15 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	wx1	12:34:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000739	t20	12:35:14 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000730	phad	12:20:15 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000730	fx5	12:21:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000730	fx1	12:22:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000730	fx2	12:23:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000730	fx22	12:24:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

thermo scientific

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000730	gx2	12:25:15 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000730	wx1	12:26:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000730	t20	12:27:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	phad	12:12:15 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	fx5	12:13:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	fx1	12:14:15 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	fx2	12:15:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	fx22	12:16:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	gx2	12:17:15 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	wx1	12:18:15 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000729	t20	12:19:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	phad	12:04:16 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	fx5	12:05:15 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	fx1	12:06:16 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	fx2	12:07:15 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	fx22	12:08:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	gx2	12:09:15 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	wx1	12:10:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000735	t20	12:11:15 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	phad	11:56:16 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	fx5	11:57:16 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	fx1	11:58:16 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	fx2	11:59:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	fx22	12:00:16 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	gx2	12:01:16 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	wx1	12:02:16 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000748	t20	12:03:16 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	phad	11:48:16 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	fx5	11:49:16 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	fx1	11:50:16 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	fx2	11:51:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	fx22	11:52:16 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	gx2	11:53:16 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	wx1	11:54:16 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000744	t20	11:55:16 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	phad	11:28:17 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	fx5	11:41:17 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	fx1	11:42:17 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	fx2	11:43:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	fx22	11:44:16 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	gx2	11:45:16 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	wx1	11:46:16 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000749	t20	11:47:16 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	phad	11:20:17 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	fx5	11:21:17 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	fx1	11:22:17 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	fx2	11:23:17 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	fx22	11:24:17 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	gx2	11:25:17 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	wx1	11:26:17 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000745	t20	11:27:17 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	phad	10:49:18 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	fx5	10:50:18 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	fx1	10:51:18 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	fx2	10:52:18 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	fx22	11:16:17 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	gx2	11:17:17 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	wx1	11:18:17 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000750	t20	11:19:17 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000759d	fx5	11:29:17 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000759d	fx22	11:30:17 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000759d	t20	11:31:17 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000759d	wx1	11:32:17 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000759d	gx2	11:33:17 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000759d	fx2	11:34:17 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			

thermo scientific

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000759d	fx1	11:35:16 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000760d	phad	11:36:16 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000760d	fx5	11:37:17 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000760d	fx22	11:38:16 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000760d	fx2	11:39:16 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000760d	fx1	11:40:16 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000758	fx22	12:44:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	phad	1:13:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	fx5	1:14:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	fx1	1:15:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	fx2	1:16:14 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	fx22	1:17:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	gx2	1:18:13 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	wx1	1:19:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000340	t20	1:20:13 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	phad	1:21:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	fx5	1:22:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	fx1	1:23:13 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	fx2	1:24:13 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	fx22	1:25:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	gx2	1:26:13 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	wx1	1:27:13 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000330	t20	1:33:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	phad	1:34:13 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	fx5	1:35:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	fx1	1:36:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	fx2	1:37:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	fx22	1:38:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	gx2	1:39:12 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	wx1	1:40:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000521	t20	1:41:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	phad	1:42:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	fx5	1:43:13 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	fx1	1:44:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	fx2	1:45:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	fx22	1:46:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	gx2	1:47:12 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	wx1	1:48:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000511	t20	1:49:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	phad	1:50:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	fx5	1:51:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	fx1	1:52:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	fx2	1:53:12 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	fx22	1:54:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	gx2	1:55:12 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	wx1	1:56:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000541	t20	1:57:12 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	phad	1:58:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	fx5	1:59:12 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	fx1	2:00:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	fx2	2:01:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	fx22	2:02:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	gx2	2:03:12 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	wx1	2:04:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000531	t20	2:05:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	phad	2:06:12 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	fx5	2:07:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	fx1	2:08:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	fx2	2:09:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	fx22	2:10:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	gx2	2:11:11 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	wx1	2:12:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000542	t20	2:13:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000532	phad	2:14:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000532	fx5	2:15:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

thermo scientific

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000532	fx1	2:16:12 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000532	fx2	2:17:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000532	fx22	2:18:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000532	gx2	2:19:11 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000532	wx1	2:20:12 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000532	t20	2:21:11 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	phad	2:22:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	fx5	2:23:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	fx1	2:24:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	fx2	2:25:11 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	fx22	2:26:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	gx2	2:27:11 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	wx1	2:28:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000512	t20	2:29:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	phad	2:30:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	fx5	2:31:11 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	fx1	2:32:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	fx2	2:33:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	fx22	2:34:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	gx2	2:35:10 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	wx1	2:36:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000502	t20	2:37:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000350	phad	2:38:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

Errors

Expired lot numbers used for Sample	
AJ0BM	Diluent

Expired lot numbers used for Quality Control	
352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial

thermo scientific

Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200112-376
Instrument name:	Phadia-A
Instrument type:	Phadia 250
Status:	Ready
Started:	12-Jan-20 2:39:10 PM
Ended:	12-Jan-20 6:15:04 PM
Approved:	Approved (faziz)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-A
Instrument type:	Phadia 250
Measured:	05-Jan-20 1:26:01 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



thermo scientific

Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	3:57:07 PM	49	0.08 kUA/l	<lab>	----	Not
			Expected value:	---				
			Expected SD:	---				
			SD limit:	---				
			Range min:	---				
			Range max:	---				
			Limit CV%:	---				

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000305	phad	5	6:14:02 PM	33	0.19 PAU/l	---	Negative	
011990000306	phad	5	6:06:03 PM	35	0.22 PAU/l	---	Negative	
011990000307	phad	5	3:56:08 PM	37	0.23 PAU/l	---	Negative	
011990000308	phad	5	5:18:05 PM	32	0.18 PAU/l	---	Negative	
011990000309	phad	5	4:38:06 PM	32	0.19 PAU/l	---	Negative	
011990000310	phad	5	3:18:09 PM	41	0.29 PAU/l	---	Negative	
011990000315	phad	5	5:58:03 PM	31	0.16 PAU/l	---	Negative	
011990000316	phad	5	4:54:06 PM	33	0.19 PAU/l	---	Negative	
011990000317	phad	5	4:05:07 PM	236	2.55 PAU/l	---	Positive	
011990000318	phad	5	5:26:05 PM	34	0.20 PAU/l	---	Negative	
011990000319	phad	5	3:10:10 PM	163	1.71 PAU/l	---	Positive	
011990000320	phad	5	4:30:07 PM	80	0.75 PAU/l	---	Positive	
011990000325	phad	5	5:50:04 PM	44	0.32 PAU/l	---	Negative	
011990000326	phad	5	4:46:06 PM	166	1.74 PAU/l	---	Positive	
011990000327	phad	5	4:21:06 PM	1448	17.0 PAU/l	---	Positive	
011990000328	phad	5	3:02:10 PM	63	0.54 PAU/l	---	Positive	
011990000329	phad	5	3:48:08 PM	123	1.24 PAU/l	---	Positive	
011990000334	phad	5	5:10:05 PM	57	0.48 PAU/l	---	Positive	
011990000337	phad	5	4:13:07 PM	31	0.17 PAU/l	---	Negative	
011990000338	phad	5	2:54:11 PM	41	0.29 PAU/l	---	Negative	
011990000339	phad	5	3:40:09 PM	27	0.12 PAU/l	---	Negative	
011990000344	phad	5	5:02:06 PM	30	0.16 PAU/l	---	Negative	
011990000347	phad	5	5:42:04 PM	95	0.91 PAU/l	---	Positive	
011990000348	phad	5	5:34:05 PM	36	0.23 PAU/l	---	Negative	
011990000349	phad	5	2:46:11 PM	31	0.16 PAU/l	---	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000306	fx5	5	6:07:02 PM	30	0.15 kUA/l	---	Negative	
011990000306	fx1	5	6:08:03 PM	20	0.04 kUA/l	---	Negative	
011990000306	fx2	5	6:09:02 PM	18	0.01 kUA/l	---	Negative	
011990000306	fx22	5	6:10:03 PM	28	0.14 kUA/l	---	Negative	
011990000306	gx2	5	6:11:02 PM	26	0.11 kUA/l	---	Negative	
011990000306	wx1	5	6:12:02 PM	41	0.28 kUA/l	---	Negative	
011990000306	t20	5	6:13:02 PM	19	0.03 kUA/l	---		
011990000307	fx5	5	3:58:07 PM	24	0.08 kUA/l	---	Negative	
011990000307	fx1	5	3:59:07 PM	20	0.04 kUA/l	---	Negative	
011990000307	fx2	5	4:00:08 PM	14	---	---	Negative	
011990000307	fx22	5	4:01:07 PM	31	0.17 kUA/l	---	Negative	
011990000307	gx2	5	4:02:08 PM	26	0.11 kUA/l	---	Negative	
011990000307	wx1	5	4:03:07 PM	118	1.18 kUA/l	---	Positive	
011990000307	t20	5	4:04:07 PM	25	0.09 kUA/l	---		
011990000308	fx5	5	5:19:04 PM	21	0.05 kUA/l	---	Negative	
011990000308	fx1	5	5:20:06 PM	18	0.01 kUA/l	---	Negative	
011990000308	fx2	5	5:21:04 PM	13	---	---	Negative	
011990000308	fx22	5	5:22:05 PM	31	0.17 kUA/l	---	Negative	
011990000308	gx2	5	5:23:04 PM	20	0.04 kUA/l	---	Negative	
011990000308	wx1	5	5:24:05 PM	29	0.15 kUA/l	---	Negative	
011990000308	t20	5	5:25:04 PM	17	---	---		
011990000309	fx5	5	4:39:06 PM	25	0.10 kUA/l	---	Negative	
011990000309	fx1	5	4:40:06 PM	23	0.07 kUA/l	---	Negative	
011990000309	fx2	5	4:41:06 PM	13	---	---	Negative	

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000309	fx22	5	4:42:06 PM	34	0.21 kUA/l	---	Negative	
011990000309	gx2	5	4:43:06 PM	22	0.06 kUA/l	---	Negative	
011990000309	wx1	5	4:44:06 PM	28	0.13 kUA/l	---	Negative	
011990000309	t20	5	4:45:06 PM	16	---	---		
011990000310	fx5	5	3:19:09 PM	21	0.05 kUA/l	---	Negative	
011990000310	fx1	5	3:20:09 PM	21	0.05 kUA/l	---	Negative	
011990000310	fx2	5	3:35:08 PM	15	---	---	Negative	
011990000310	fx22	5	3:36:09 PM	26	0.11 kUA/l	---	Negative	
011990000310	gx2	5	3:37:08 PM	50	0.40 kUA/l	---	Positive	
011990000310	wx1	5	3:38:09 PM	38	0.25 kUA/l	---	Negative	
011990000310	t20	5	3:39:08 PM	29	0.15 kUA/l	---		
011990000315	fx5	5	5:59:03 PM	23	0.08 kUA/l	---	Negative	
011990000315	fx1	5	6:00:03 PM	19	0.03 kUA/l	---	Negative	
011990000315	fx2	5	6:01:03 PM	15	---	---	Negative	
011990000315	fx22	5	6:02:03 PM	32	0.18 kUA/l	---	Negative	
011990000315	gx2	5	6:03:02 PM	21	0.05 kUA/l	---	Negative	
011990000315	wx1	5	6:04:03 PM	32	0.18 kUA/l	---	Negative	
011990000315	t20	5	6:05:02 PM	19	0.02 kUA/l	---		
011990000316	fx5	5	4:55:05 PM	21	0.05 kUA/l	---	Negative	
011990000316	fx1	5	4:56:06 PM	16	---	---	Negative	
011990000316	fx2	5	4:57:05 PM	14	---	---	Negative	
011990000316	fx22	5	4:58:06 PM	30	0.15 kUA/l	---	Negative	
011990000316	gx2	5	4:59:05 PM	22	0.06 kUA/l	---	Negative	
011990000316	wx1	5	5:00:06 PM	30	0.15 kUA/l	---	Negative	
011990000316	t20	5	5:01:05 PM	18	0.01 kUA/l	---		
011990000317	fx5	5	4:06:07 PM	38	0.25 kUA/l	---	Negative	
011990000317	fx1	5	4:07:07 PM	23	0.07 kUA/l	---	Negative	
011990000317	fx2	5	4:08:07 PM	147	1.52 kUA/l	---	Positive	
011990000317	fx22	5	4:09:07 PM	36	0.22 kUA/l	---	Negative	
011990000317	gx2	5	4:10:07 PM	28	0.13 kUA/l	---	Negative	
011990000317	wx1	5	4:11:07 PM	52	0.41 kUA/l	---	Positive	
011990000317	t20	5	4:12:07 PM	26	0.11 kUA/l	---		
011990000318	fx5	5	5:27:04 PM	25	0.10 kUA/l	---	Negative	
011990000318	fx1	5	5:28:05 PM	21	0.05 kUA/l	---	Negative	
011990000318	fx2	5	5:29:04 PM	14	---	---	Negative	
011990000318	fx22	5	5:30:05 PM	29	0.14 kUA/l	---	Negative	
011990000318	gx2	5	5:31:04 PM	22	0.06 kUA/l	---	Negative	
011990000318	wx1	5	5:32:04 PM	36	0.22 kUA/l	---	Negative	
011990000318	t20	5	5:33:04 PM	18	0.01 kUA/l	---		
011990000319	fx5	5	3:11:09 PM	32	0.18 kUA/l	---	Negative	
011990000319	fx1	5	3:12:10 PM	46	0.35 kUA/l	---	Negative	
011990000319	fx2	5	3:13:09 PM	305	3.35 kUA/l	---	Positive	
011990000319	fx22	5	3:14:10 PM	42	0.30 kUA/l	---	Negative	
011990000319	gx2	5	3:15:09 PM	45	0.33 kUA/l	---	Negative	
011990000319	wx1	5	3:16:10 PM	44	0.32 kUA/l	---	Negative	
011990000319	t20	5	3:17:09 PM	25	0.10 kUA/l	---		
011990000320	fx5	5	4:31:06 PM	27	0.13 kUA/l	---	Negative	
011990000320	fx1	5	4:32:07 PM	24	0.09 kUA/l	---	Negative	
011990000320	fx2	5	4:33:06 PM	44	0.32 kUA/l	---	Negative	
011990000320	fx22	5	4:34:06 PM	35	0.21 kUA/l	---	Negative	
011990000320	gx2	5	4:35:06 PM	31	0.17 kUA/l	---	Negative	
011990000320	wx1	5	4:36:06 PM	48	0.37 kUA/l	---	Positive	
011990000320	t20	5	4:37:06 PM	27	0.12 kUA/l	---		
011990000325	fx5	5	5:51:03 PM	31	0.17 kUA/l	---	Negative	

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000325	fx1	5	5:52:04 PM	20	0.04 kUA/l	---	Negative	
011990000325	fx2	5	5:53:03 PM	31	0.17 kUA/l	---	Negative	
011990000325	fx22	5	5:54:04 PM	33	0.20 kUA/l	---	Negative	
011990000325	gx2	5	5:55:03 PM	30	0.16 kUA/l	---	Negative	
011990000325	wx1	5	5:56:04 PM	30	0.16 kUA/l	---	Negative	
011990000325	t20	5	5:57:03 PM	21	0.05 kUA/l	---		
011990000326	fx5	5	4:47:06 PM	46	0.35 kUA/l	---	Negative	
011990000326	fx1	5	4:48:06 PM	20	0.03 kUA/l	---	Negative	
011990000326	fx2	5	4:49:05 PM	24	0.09 kUA/l	---	Negative	
011990000326	fx22	5	4:50:06 PM	29	0.15 kUA/l	---	Negative	
011990000326	gx2	5	4:51:06 PM	21	0.05 kUA/l	---	Negative	
011990000326	wx1	5	4:52:06 PM	32	0.18 kUA/l	---	Negative	
011990000326	t20	5	4:53:05 PM	20	0.03 kUA/l	---		
011990000327	fx5	5	4:22:07 PM	20	0.04 kUA/l	---	Negative	
011990000327	fx1	5	4:23:07 PM	19	0.02 kUA/l	---	Negative	
011990000327	fx2	5	4:25:06 PM	324	3.57 kUA/l	---	Positive	
011990000327	fx22	5	4:26:06 PM	33	0.19 kUA/l	---	Negative	
011990000327	gx2	5	4:27:06 PM	22	0.06 kUA/l	---	Negative	
011990000327	wx1	5	4:28:06 PM	30	0.15 kUA/l	---	Negative	
011990000327	t20	5	4:29:06 PM	18	0.02 kUA/l	---		
011990000328	fx5	5	3:03:09 PM	32	0.18 kUA/l	---	Negative	
011990000328	fx1	5	3:04:10 PM	19	0.03 kUA/l	---	Negative	
011990000328	fx2	5	3:05:09 PM	39	0.26 kUA/l	---	Negative	
011990000328	fx22	5	3:06:10 PM	31	0.17 kUA/l	---	Negative	
011990000328	gx2	5	3:07:09 PM	20	0.03 kUA/l	---	Negative	
011990000328	wx1	5	3:08:10 PM	29	0.14 kUA/l	---	Negative	
011990000328	t20	5	3:09:09 PM	17	---	---		
011990000329	fx5	5	3:49:08 PM	31	0.17 kUA/l	---	Negative	
011990000329	fx1	5	3:50:08 PM	19	0.03 kUA/l	---	Negative	
011990000329	fx2	5	3:51:08 PM	156	1.62 kUA/l	---	Positive	
011990000329	fx22	5	3:52:08 PM	33	0.19 kUA/l	---	Negative	
011990000329	gx2	5	3:53:07 PM	22	0.06 kUA/l	---	Negative	
011990000329	wx1	5	3:54:08 PM	107	1.06 kUA/l	---	Positive	
011990000329	t20	5	3:55:08 PM	29	0.14 kUA/l	---		
011990000334	fx5	5	5:11:05 PM	37	0.24 kUA/l	---	Negative	
011990000334	fx1	5	5:12:05 PM	18	0.01 kUA/l	---	Negative	
011990000334	fx2	5	5:13:05 PM	22	0.06 kUA/l	---	Negative	
011990000334	fx22	5	5:14:05 PM	35	0.21 kUA/l	---	Negative	
011990000334	gx2	5	5:15:05 PM	22	0.06 kUA/l	---	Negative	
011990000334	wx1	5	5:16:05 PM	33	0.19 kUA/l	---	Negative	
011990000334	t20	5	5:17:04 PM	22	0.06 kUA/l	---		
011990000337	fx5	5	4:14:07 PM	23	0.07 kUA/l	---	Negative	
011990000337	fx1	5	4:15:07 PM	19	0.03 kUA/l	---	Negative	
011990000337	fx2	5	4:16:07 PM	14	---	---	Negative	
011990000337	fx22	5	4:17:07 PM	32	0.18 kUA/l	---	Negative	
011990000337	gx2	5	4:18:07 PM	20	0.04 kUA/l	---	Negative	
011990000337	wx1	5	4:19:07 PM	40	0.27 kUA/l	---	Negative	
011990000337	t20	5	4:20:07 PM	18	0.01 kUA/l	---		
011990000338	fx5	5	2:55:10 PM	26	0.10 kUA/l	---	Negative	
011990000338	fx1	5	2:56:10 PM	17	---	---	Negative	
011990000338	fx2	5	2:57:09 PM	20	0.04 kUA/l	---	Negative	
011990000338	fx22	5	2:58:10 PM	28	0.14 kUA/l	---	Negative	
011990000338	gx2	5	2:59:10 PM	19	0.02 kUA/l	---	Negative	
011990000338	wx1	5	3:00:10 PM	32	0.18 kUA/l	---	Negative	

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000338	t20	5	3:01:10 PM	16	---	---		
011990000339	fx5	5	3:41:08 PM	20	0.04 kUA/l	---	Negative	
011990000339	fx1	5	3:42:09 PM	17	0.01 kUA/l	---	Negative	
011990000339	fx2	5	3:43:08 PM	14	---	---	Negative	
011990000339	fx22	5	3:44:09 PM	28	0.13 kUA/l	---	Negative	
011990000339	gx2	5	3:45:08 PM	27	0.12 kUA/l	---	Negative	
011990000339	wx1	5	3:46:08 PM	28	0.13 kUA/l	---	Negative	
011990000339	t20	5	3:47:08 PM	15	---	---		
011990000344	fx5	5	5:03:05 PM	19	0.03 kUA/l	---	Negative	
011990000344	fx1	5	5:04:06 PM	17	---	---	Negative	
011990000344	fx2	5	5:05:05 PM	13	---	---	Negative	
011990000344	fx22	5	5:06:05 PM	31	0.17 kUA/l	---	Negative	
011990000344	gx2	5	5:07:05 PM	18	0.01 kUA/l	---	Negative	
011990000344	wx1	5	5:08:06 PM	28	0.13 kUA/l	---	Negative	
011990000344	t20	5	5:09:05 PM	17	---	---		
011990000347	fx5	5	5:43:03 PM	43	0.31 kUA/l	---	Negative	
011990000347	fx1	5	5:44:04 PM	39	0.27 kUA/l	---	Negative	
011990000347	fx2	5	5:45:03 PM	22	0.06 kUA/l	---	Negative	
011990000347	fx22	5	5:46:04 PM	38	0.25 kUA/l	---	Negative	
011990000347	gx2	5	5:47:03 PM	29	0.14 kUA/l	---	Negative	
011990000347	wx1	5	5:48:04 PM	515	5.79 kUA/l	---	Positive	
011990000347	t20	5	5:49:03 PM	66	0.57 kUA/l	---		
011990000348	fx5	5	5:35:04 PM	22	0.06 kUA/l	---	Negative	
011990000348	fx1	5	5:36:04 PM	18	0.01 kUA/l	---	Negative	
011990000348	fx2	5	5:37:04 PM	15	---	---	Negative	
011990000348	fx22	5	5:38:04 PM	34	0.21 kUA/l	---	Negative	
011990000348	gx2	5	5:39:04 PM	25	0.10 kUA/l	---	Negative	
011990000348	wx1	5	5:40:05 PM	28	0.13 kUA/l	---	Negative	
011990000348	t20	5	5:41:03 PM	18	0.01 kUA/l	---		
011990000349	fx5	5	2:47:10 PM	21	0.05 kUA/l	---	Negative	
011990000349	fx1	5	2:48:11 PM	20	0.04 kUA/l	---	Negative	
011990000349	fx2	5	2:49:10 PM	14	---	---	Negative	
011990000349	fx22	5	2:50:11 PM	27	0.12 kUA/l	---	Negative	
011990000349	gx2	5	2:51:10 PM	20	0.04 kUA/l	---	Negative	
011990000349	wx1	5	2:52:10 PM	27	0.12 kUA/l	---	Negative	
011990000349	t20	5	2:53:10 PM	15	---	---		
011990000350	fx5	5	2:39:10 PM	33	0.19 kUA/l	---	Negative	
011990000350	fx1	5	2:40:11 PM	17	---	---	Negative	
011990000350	fx2	5	2:41:10 PM	28	0.13 kUA/l	---	Negative	
011990000350	fx22	5	2:42:11 PM	28	0.14 kUA/l	---	Negative	
011990000350	gx2	5	2:43:10 PM	1104	12.8 kUA/l	---	Positive	
011990000350	wx1	5	2:44:11 PM	36	0.23 kUA/l	---	Negative	
011990000350	t20	5	2:45:10 PM	19	0.02 kUA/l	---		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

thermo scientific

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	3:57:07 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000338	phad	2:54:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000338	fx5	2:55:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000338	fx1	2:56:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000338	fx2	2:57:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000338	fx22	2:58:10 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000338	gx2	2:59:10 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000338	wx1	3:00:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000338	t20	3:01:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	phad	3:02:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	fx5	3:03:09 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	fx1	3:04:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	fx2	3:05:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	fx22	3:06:10 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	gx2	3:07:09 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	wx1	3:08:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000328	t20	3:09:09 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	phad	3:10:10 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	fx5	3:11:09 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	fx1	3:12:10 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	fx2	3:13:09 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	fx22	3:14:10 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	gx2	3:15:09 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	wx1	3:16:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000319	t20	3:17:09 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	phad	3:18:09 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	fx5	3:19:09 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	fx1	3:20:09 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	fx2	3:35:08 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	fx22	3:36:09 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	gx2	3:37:08 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	wx1	3:38:09 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000310	t20	3:39:08 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	phad	3:40:09 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	fx5	3:41:08 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	fx1	3:42:09 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	fx2	3:43:08 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	fx22	3:44:09 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	gx2	3:45:08 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	wx1	3:46:08 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000339	t20	3:47:08 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	phad	3:48:08 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	fx5	3:49:08 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	fx1	3:50:08 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	fx2	3:51:08 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	fx22	3:52:08 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	gx2	3:53:07 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	wx1	3:54:08 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000329	t20	3:55:08 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000350	fx5	2:39:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000350	fx1	2:40:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000350	fx2	2:41:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000350	fx22	2:42:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000350	gx2	2:43:10 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000350	wx1	2:44:11 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

thermo scientific

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000350	t20	2:45:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	phad	2:46:11 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	fx5	2:47:10 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	fx1	2:48:11 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	fx2	2:49:10 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	fx22	2:50:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	gx2	2:51:10 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	wx1	2:52:10 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000349	t20	2:53:10 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	phad	4:30:07 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	fx5	4:31:06 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	fx1	4:32:07 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	fx2	4:33:06 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	fx22	4:34:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	gx2	4:35:06 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	wx1	4:36:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000320	t20	4:37:06 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	phad	4:38:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	fx5	4:39:06 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	fx1	4:40:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	fx2	4:41:06 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	fx22	4:42:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	gx2	4:43:06 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	wx1	4:44:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000309	t20	4:45:06 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	phad	4:46:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	fx5	4:47:06 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	fx1	4:48:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	fx2	4:49:05 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	fx22	4:50:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	gx2	4:51:06 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	wx1	4:52:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000326	t20	4:53:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	phad	4:54:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	fx5	4:55:05 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	fx1	4:56:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	fx2	4:57:05 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	fx22	4:58:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	gx2	4:59:05 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	wx1	5:00:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000316	t20	5:01:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	phad	4:05:07 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	fx5	4:06:07 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	fx1	4:07:07 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	fx2	4:08:07 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	fx22	4:09:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	gx2	4:10:07 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	wx1	4:11:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000317	t20	4:12:07 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	phad	3:56:08 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	fx5	3:58:07 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	fx1	3:59:07 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	fx2	4:00:08 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	fx22	4:01:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	gx2	4:02:08 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	wx1	4:03:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000307	t20	4:04:07 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	phad	4:13:07 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	fx5	4:14:07 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	fx1	4:15:07 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	fx2	4:16:07 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	fx22	4:17:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	gx2	4:18:07 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	wx1	4:19:07 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000337	t20	4:20:07 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

thermo scientific

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000327	phad	4:21:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000327	fx5	4:22:07 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000327	fx1	4:23:07 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000327	fx2	4:25:06 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000327	fx22	4:26:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000327	gx2	4:27:06 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000327	wx1	4:28:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000327	t20	4:29:06 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	phad	5:34:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	fx5	5:35:04 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	fx1	5:36:04 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	fx2	5:37:04 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	fx22	5:38:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	gx2	5:39:04 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	wx1	5:40:05 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000348	t20	5:41:03 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	phad	5:42:04 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	fx5	5:43:03 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	fx1	5:44:04 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	fx2	5:45:03 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	fx22	5:46:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	gx2	5:47:03 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	wx1	5:48:04 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000347	t20	5:49:03 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	phad	5:26:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	fx5	5:27:04 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	fx1	5:28:05 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	fx2	5:29:04 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	fx22	5:30:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	gx2	5:31:04 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	wx1	5:32:04 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000318	t20	5:33:04 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	phad	5:18:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	fx5	5:19:04 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	fx1	5:20:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	fx2	5:21:04 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	fx22	5:22:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	gx2	5:23:04 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	wx1	5:24:05 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000308	t20	5:25:04 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	phad	5:02:06 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	fx5	5:03:05 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	fx1	5:04:06 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	fx2	5:05:05 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	fx22	5:06:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	gx2	5:07:05 PM	433BK 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	wx1	5:08:06 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000344	t20	5:09:05 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	phad	5:10:05 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	fx5	5:11:05 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	fx1	5:12:05 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	fx2	5:13:05 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	fx22	5:14:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	gx2	5:15:05 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	wx1	5:16:05 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000334	t20	5:17:04 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	phad	5:50:04 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	fx5	5:51:03 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	fx1	5:52:04 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	fx2	5:53:03 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	fx22	5:54:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	gx2	5:55:03 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	wx1	5:56:04 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000325	t20	5:57:03 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000315	phad	5:58:03 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

thermo scientific

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000315	fx5	5:59:03 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000315	fx1	6:00:03 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000315	fx2	6:01:03 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000315	fx22	6:02:03 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000315	gx2	6:03:02 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000315	wx1	6:04:03 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000315	t20	6:05:02 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	phad	6:06:03 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	fx5	6:07:02 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	fx1	6:08:03 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	fx2	6:09:02 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	fx22	6:10:03 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	gx2	6:11:02 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	wx1	6:12:02 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000306	t20	6:13:02 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000305	phad	6:14:02 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

Errors

Expired lot numbers used for Sample	
AJ0BM	Diluent

Notes

Identity	Test	Measured	Comment
IQC271119	d1	3:57:07 PM	reject low volume-FA

Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200112-376
Instrument name:	Phadia-A
Instrument type:	Phadia 250
Status:	Ready
Started:	12-Jan-20 6:15:02 PM
Ended:	12-Jan-20 6:39:22 PM
Approved:	Approved (faziz)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-A
Instrument type:	Phadia 250
Measured:	05-Jan-20 1:26:01 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



thermo scientific

Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

No quality controls in this run

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc		Class	Cut-Off	Remarks
011990000335	phad	5	6:30:01 PM	137	1.41	PAU/l	---	Positive	
011990000345	phad	5	6:22:02 PM	37	0.23	PAU/l	---	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks

thermo scientific

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000305	fx5	5	6:15:02 PM	37	0.23 kUA/l	---	Negative	
011990000305	fx1	5	6:16:02 PM	19	0.02 kUA/l	---	Negative	
011990000305	fx2	5	6:17:02 PM	16	---	---	Negative	
011990000305	fx22	5	6:18:02 PM	31	0.17 kUA/l	---	Negative	
011990000305	gx2	5	6:19:02 PM	23	0.07 kUA/l	---	Negative	
011990000305	wx1	5	6:20:02 PM	30	0.16 kUA/l	---	Negative	
011990000305	t20	5	6:21:02 PM	18	0.02 kUA/l	---		
011990000335	fx5	5	6:31:01 PM	23	0.08 kUA/l	---	Negative	
011990000335	fx1	5	6:32:01 PM	20	0.04 kUA/l	---	Negative	
011990000345	fx5	5	6:23:02 PM	26	0.11 kUA/l	---	Negative	
011990000345	fx1	5	6:24:01 PM	21	0.05 kUA/l	---	Negative	
011990000345	fx2	5	6:25:01 PM	15	---	---	Negative	
011990000345	fx22	5	6:26:02 PM	31	0.17 kUA/l	---	Negative	
011990000345	gx2	5	6:27:01 PM	26	0.11 kUA/l	---	Negative	
011990000345	wx1	5	6:28:01 PM	33	0.19 kUA/l	---	Negative	
011990000345	t20	5	6:29:01 PM	19	0.03 kUA/l	---		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000305	fx5	6:15:02 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000305	fx1	6:16:02 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000305	fx2	6:17:02 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000305	fx22	6:18:02 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000305	gx2	6:19:02 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000305	wx1	6:20:02 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000305	t20	6:21:02 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	phad	6:22:02 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	fx5	6:23:02 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	fx1	6:24:01 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	fx2	6:25:01 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	fx22	6:26:02 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	gx2	6:27:01 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	wx1	6:28:01 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000345	t20	6:29:01 PM	475GF 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000335	phad	6:30:01 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000335	fx5	6:31:01 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp
011990000335	fx1	6:32:01 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BM 19/12!	Exp

Errors

Expired lot numbers used for Sample	
AJ0BM	Diluent