


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200202-391  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Processing  
Started: 02-Feb-20 11:41:13 AM  
Ended:  
Approved: Not approved  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	11:41:13 AM	309	0.68	OK	
CC-2	17.5	11:42:12 AM	6105	16.8	OK	

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	11:43:13 AM	3562	8.96 kUA/l	CZ1AR	OK	
slgE Neg	d1	1	11:44:12 AM	23	0.02 kUA/l	BL1CP	OK	
slgE L	e1	1	11:45:13 AM	1327	3.10 kUA/l	CZ4AS	OK	
IQC271119	d1	1	11:46:12 AM	7092	20.3 kUA/l	<lab>	OK	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000657	wx1	5	12:08:11 PM	31	0.17 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000881	fx22	5	12:06:11 PM	31	0.17 kUA/I	0	Negative	
011990000881	wx1	5	12:07:12 PM	33	0.19 kUA/I	0	Negative	
012002800387A	f422	1	12:25:10 PM	18	0.00 kUA/I			
012002800387A	f423	1	12:26:10 PM	248	0.54 kUA/I			
012002800387A	f424	1	12:27:11 PM	15	0.00 kUA/I			
012002800387A	f427	1	12:28:10 PM	14	0.00 kUA/I			
012002800387A	f352	1	12:40:10 PM	13	0.00 kUA/I			
012002800669A	m6	1	11:57:12 AM	41	0.06 kUA/I			
012002800669A	m5	1	11:58:11 AM	40	0.06 kUA/I			
012002800669A	f78	1	11:59:12 AM	11811	41.5 kUA/I			
012002800669A	i206	1	12:38:10 PM	46	0.07 kUA/I			
012002800669A	f245	1	12:00:12 PM	19190	>100			
012002800669A	f1	1	12:01:12 PM	19008	>100			
012002800669A	f75	1	12:02:11 PM	6728	19.0 kUA/I			
012002800669A	hx2	1	12:03:12 PM	203			Positive	
012002800669A	d1	1	12:04:11 PM	67	0.12 kUA/I			
012002800669A	f2	1	12:39:10 PM	12865	47.6 kUA/I			
012002800669A	f233	1	12:05:12 PM	9332	29.3 kUA/I			
012002800669A	f14	1	12:09:12 PM	5656	15.3 kUA/I			
012002800669A	mx2	1	12:10:11 PM	42			Negative	
012002800669A	i6	1	12:11:12 PM	401	0.89 kUA/I			
012002800669A	f13	1	12:12:11 PM	10480	34.6 kUA/I			
012002800669A	w45	1	12:13:12 PM	69	0.12 kUA/I			
012002800723A	hx2	1	11:52:12 AM	22			Negative	
012002800723A	ex72	1	11:53:13 AM	16			Negative	
012002800723A	mx2	1	11:54:12 AM	26			Negative	
012002800723A	gx2	1	11:55:12 AM	19	0.00 kUA/I	0	Negative	
012002800723A	mx4	1	11:56:12 AM	13			Negative	
012002801080A	t19	1	12:14:11 PM	18	0.00 kUA/I			
012002801080A	f309	1	12:15:11 PM	33	0.04 kUA/I			
012002801080A	f3	1	12:16:11 PM	15	0.00 kUA/I			
012002801080A	t3	1	12:17:12 PM	20	0.01 kUA/I			
012002801080A	t14	1	12:18:11 PM	24	0.02 kUA/I			
012002801080A	t20	1	12:19:11 PM	35	0.04 kUA/I	0		
012002801080A	f41	1	12:20:11 PM	16	0.00 kUA/I			
012002801080A	f10	1	12:21:11 PM	339	0.75 kUA/I			
012002801080A	f40	1	12:22:11 PM	21	0.01 kUA/I			
012002900325A	f20	1	12:41:10 PM	383	0.85 kUA/I			
012002900325A	f3	1	12:42:10 PM	2940	7.24 kUA/I			
012002900325A	f17	1	12:43:10 PM	538	1.21 kUA/I			
012002900325A	f203	1	12:44:10 PM	606	1.37 kUA/I			
012002900325A	f41	1	12:45:10 PM	4097	10.5 kUA/I			
012002900325A	f24	1	12:46:10 PM	26	0.02 kUA/I			
012002900325A	f40	1	12:47:10 PM	1123	2.60 kUA/I			
012002900325A	f256	1	12:48:10 PM	2292	5.52 kUA/I			
012002900325A	f13	1	12:49:10 PM	479	1.07 kUA/I			
012002900472A	f20	1	12:50:10 PM	911	2.09 kUA/I			
012002900472A	m6	1	12:51:10 PM	21	0.01 kUA/I			
012002900472A	m3	1	12:52:10 PM	20	0.01 kUA/I			
012002900472A	f202	1	12:53:10 PM	13770	53.4 kUA/I			
012002900472A	i206	1	12:54:09 PM	18	0.00 kUA/I			
012002900472A	e5	1	12:55:10 PM	82	0.15 kUA/I			
012002900472A	f17	1	12:56:10 PM	2808	6.88 kUA/I			
012002900472A	d1	1	12:57:10 PM	25	0.02 kUA/I			

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
012002900472A	f345	1	12:58:09 PM	1036	2.39 kUA/l			
012002900472A	f13	1	12:59:09 PM	1293	3.01 kUA/l			
012002900472A	m1	1	1:00:09 PM	13	0.00 kUA/l			
012002900472A	f203	1	1:01:09 PM	13189	49.6 kUA/l			
012002900472A	f256	1	1:02:09 PM	3947	10.1 kUA/l			
012002900472A	gx2	1	1:03:10 PM	35	0.04 kUA/l	0	Negative	
012002900472A	i6	1	1:04:09 PM	106	0.21 kUA/l			
012002900633A	m6	1	1:05:09 PM	11656	40.6 kUA/l			
012002900633A	m3	1	1:06:09 PM	6080	16.7 kUA/l			
012002900633A	e1	1	1:07:09 PM	3773	9.56 kUA/l			
012002900633A	i206	1	2:24:06 PM	35	0.04 kUA/l			
012002900633A	e5	1	1:08:09 PM	406	0.90 kUA/l			
012002900633A	d1	1	1:09:09 PM	695	1.58 kUA/l			
012002900633A	f8	1	1:10:09 PM	587	1.33 kUA/l			
012002900633A	f7	1	1:11:09 PM	1908	4.54 kUA/l			
012002900633A	f10	1	1:12:09 PM	5355	14.3 kUA/l			
012002900633A	f4	1	1:13:09 PM	655	1.49 kUA/l			
012002900633A	ex72	1	1:14:08 PM	76			Negative	
012002900633A	gx2	1	1:15:09 PM	897	2.06 kUA/l	2	Positive	
012002900633A	i6	1	1:16:09 PM	60	0.10 kUA/l			
012002900791A	t19	1	1:17:09 PM	12	0.00 kUA/l			
012002900791A	m6	1	1:18:08 PM	18	0.00 kUA/l			
012002900791A	m3	1	1:19:09 PM	16	0.00 kUA/l			
012002900791A	g2	1	1:20:08 PM	18	0.00 kUA/l			
012002900791A	m5	1	1:21:09 PM	15	0.00 kUA/l			
012002900791A	m2	1	1:22:08 PM	14	0.00 kUA/l			
012002900791A	i206	1	2:25:06 PM	16	0.00 kUA/l			
012002900791A	t14	1	1:23:09 PM	18	0.00 kUA/l			
012002900791A	t214	1	1:24:08 PM	13	0.00 kUA/l			
012002900791A	hx2	1	1:25:09 PM	16			Negative	
012002900791A	d1	1	1:26:08 PM	19	0.01 kUA/l			
012002900791A	m1	1	1:27:08 PM	11	0.00 kUA/l			
012002900791A	mx2	1	1:28:08 PM	25			Negative	
012002900791A	gx2	1	1:29:08 PM	21	0.01 kUA/l	0	Negative	
012002900791A	i6	1	1:30:08 PM	29	0.03 kUA/l			
012002900791A	w45	1	1:31:08 PM	13	0.00 kUA/l			
012002900791A	w82	1	1:32:08 PM	12	0.00 kUA/l			
012002900791A	gx3	1	1:33:08 PM	17			Negative	
012002900848A	f2	1	1:34:08 PM	22	0.01 kUA/l			
012002900920A	f245	1	1:35:08 PM	45	0.07 kUA/l			
012002900920A	f2	1	1:36:08 PM	82	0.15 kUA/l			
012002900920A	f24	1	1:37:08 PM	16	0.00 kUA/l			
012002900920A	f14	1	1:38:07 PM	16	0.00 kUA/l			
012002900920A	f4	1	1:39:08 PM	45	0.07 kUA/l			
012002900920A	f13	1	1:40:07 PM	22	0.01 kUA/l			
012002900983A	f20	1	1:40:25 PM	---	---			Err
012002900983A	m6	1	1:41:25 PM	---	---			Err
012002900983A	m3	1	1:42:25 PM	---	---			Err
012002900983A	f92	1	1:43:25 PM	---	---			Err
012002900983A	g2	1	1:44:25 PM	---	---			Err
012002900983A	f18	1	1:45:25 PM	---	---			Err
012002900983A	f93	1	1:46:25 PM	---	---			Err
012002900983A	f78	1	1:47:25 PM	---	---			Err
012002900983A	f202	1	1:48:25 PM	---	---			Err

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
012002900983A	e1	1	1:49:25 PM	---	---			Err
012002900983A	f309	1	1:50:25 PM	---	---			Err
012002900983A	f83	1	1:55:25 PM	---	---			Err
012002900983A	f36	1	1:56:24 PM	---	---			Err
012002900983A	w1	1	1:57:25 PM	---	---			Err
012002900983A	d2	1	1:58:25 PM	---	---			Err
012002900983A	e5	1	1:59:25 PM	---	---			Err
012002900983A	f1	1	2:00:24 PM	---	---			Err
012002900983A	f17	1	2:01:24 PM	---	---			Err
012002900983A	d1	1	2:02:24 PM	---	---			Err
012002900983A	f235	1	2:03:24 PM	---	---			Err
012002900983A	f345	1	2:04:24 PM	---	---			Err
012002900983A	f91	1	2:05:24 PM	---	---			Err
012002900983A	f2	1	2:06:24 PM	---	---			Err
012002900983A	f233	1	2:07:24 PM	---	---			Err
012002900983A	f13	1	2:08:24 PM	---	---			Err
012002900983A	f203	1	2:09:24 PM	---	---			Err
012002900983A	f35	1	2:10:24 PM	---	---			Err
012002900983A	f41	1	2:11:24 PM	---	---			Err
012002900983A	f10	1	2:12:24 PM	---	---			Err
012002900983A	f24	1	2:13:24 PM	---	---			Err
012002900983A	f14	1	2:14:24 PM	---	---			Err
012002900983A	f40	1	2:15:24 PM	---	---			Err
012002900983A	f256	1	2:16:24 PM	---	---			Err
012002900983A	f4	1	2:17:24 PM	---	---			Err
012002900983A	ex72	1	2:18:24 PM	---			---	Err
012002900983A	mx2	1	2:19:24 PM	---			---	Err
012002900983A	gx2	1	2:20:24 PM	---	---	---	---	Err
012002900983A	i6	1	2:21:24 PM	---	---			Err
012002901077A	f20	1	2:26:05 PM	13	0.00 kUA/l			
012002901077A	f27	1	2:27:06 PM	28	0.03 kUA/l			
012002901077A	f93	1	2:28:06 PM	13	0.00 kUA/l			

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	11:41:13 AM	865FX 20/02!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	Exp
CC-2	a_IgE	11:42:12 AM	865FX 20/02!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	Exp

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	11:43:13 AM	352D0 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ1AR 21/08		Exp
slgE Neg	d1	11:44:12 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		BL1CP 19/12!		Exp

# Laboratory Report

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE L	e1	11:45:13 AM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ4AS 20/03		
IQC271119	d1	11:46:12 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002800669A	m6	11:57:12 AM	356FF 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	m5	11:58:11 AM	369CM 21/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f78	11:59:12 AM	CTSB0 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	i206	12:38:10 PM	C07GX 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f245	12:00:12 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f1	12:01:12 PM	361EN 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f75	12:02:11 PM	427CE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	hx2	12:03:12 PM	446BR 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	d1	12:04:11 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f2	12:39:10 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f233	12:05:12 PM	904LM 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f14	12:09:12 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	mx2	12:10:11 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	i6	12:11:12 PM	465D3 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	f13	12:12:11 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800669A	w45	12:13:12 PM	CCRGa 21/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800723A	hx2	11:52:12 AM	446BR 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800723A	ex72	11:53:13 AM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800723A	mx2	11:54:12 AM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800723A	gx2	11:55:12 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800723A	mx4	11:56:12 AM	BYDG8 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002801080A	t19	12:14:11 PM	474GG 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			Exp
012002801080A	f309	12:15:11 PM	BTSg9 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002801080A	f3	12:16:11 PM	404CF 20/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002801080A	t3	12:17:12 PM	352D0 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			Exp
012002801080A	t14	12:18:11 PM	470BS 20/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002801080A	t20	12:19:11 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002801080A	f41	12:20:11 PM	456CN 20/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002801080A	f10	12:21:11 PM	418DE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002801080A	f40	12:22:11 PM	455DT 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f20	12:41:10 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f3	12:42:10 PM	404CF 20/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f17	12:43:10 PM	D2FBX 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f203	12:44:10 PM	998HG 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f41	12:45:10 PM	456CN 20/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f24	12:46:10 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f40	12:47:10 PM	455DT 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f256	12:48:10 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900325A	f13	12:49:10 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	f20	12:50:10 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	m6	12:51:10 PM	356FF 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	m3	12:52:10 PM	368KB 21/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	f202	12:53:10 PM	997KA 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	i206	12:54:09 PM	C07GX 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	e5	12:55:10 PM	360F3 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	f17	12:56:10 PM	D2FBV 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	d1	12:57:10 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	f345	12:58:09 PM	BTXGP 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	f13	12:59:09 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	m1	1:00:09 PM	405CP 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	f203	1:01:09 PM	998HG 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	f256	1:02:09 PM	CMEBK 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	gx2	1:03:10 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900472A	i6	1:04:09 PM	465D3 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800387A	f422	12:25:10 PM	CDFH4 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800387A	f423	12:26:10 PM	CDGHN 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800387A	f424	12:27:11 PM	CDHH0 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800387A	f427	12:28:10 PM	CFCH0 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002800387A	f352	12:40:10 PM	CEZH5 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	m6	1:05:09 PM	356FF 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	m3	1:06:09 PM	368KB 21/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002900633A	e1	1:07:09 PM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	i206	2:24:06 PM	C07GX 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	e5	1:08:09 PM	360F3 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	d1	1:09:09 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	f8	1:10:09 PM	373C0 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	f7	1:11:09 PM	416BC 20/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	f10	1:12:09 PM	418DE 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	f4	1:13:09 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	ex72	1:14:08 PM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	gx2	1:15:09 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900633A	i6	1:16:09 PM	465D3 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	t19	1:17:09 PM	474GG 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			Exp
012002900791A	m6	1:18:08 PM	356FF 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	m3	1:19:09 PM	368KB 21/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	g2	1:20:08 PM	379CM 21/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	m5	1:21:09 PM	369CM 21/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	m2	1:22:08 PM	355D5 20/02!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			Exp
012002900791A	i206	2:25:06 PM	C07GX 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	t14	1:23:09 PM	470BS 20/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	t214	1:24:08 PM	C2PGA 20/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	hx2	1:25:09 PM	446BR 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	d1	1:26:08 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	m1	1:27:08 PM	405CP 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	mx2	1:28:08 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	gx2	1:29:08 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	i6	1:30:08 PM	465D3 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	w45	1:31:08 PM	CCRGa 21/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	w82	1:32:08 PM	CCTGA 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900791A	gx3	1:33:08 PM	434B9 20/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900848A	f2	1:34:08 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900920A	f245	1:35:08 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900920A	f2	1:36:08 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900920A	f24	1:37:08 PM	424EF 20/11	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900920A	f14	1:38:07 PM	364DN 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900920A	f4	1:39:08 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900920A	f13	1:40:07 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002900983A	f20	1:40:25 PM	422CS 20/12	A1HGR 21/06					
012002900983A	m6	1:41:25 PM	356FF 21/03	A1HGR 21/06					
012002900983A	m3	1:42:25 PM	368KB 21/10	A1HGR 21/06					
012002900983A	f92	1:43:25 PM	561KS 21/05	A1HGR 21/06					
012002900983A	g2	1:44:25 PM	379CM 21/09	A1HGR 21/06					
012002900983A	f18	1:45:25 PM	421BJ 20/09	A1HGR 21/06					
012002900983A	f93	1:46:25 PM	562JV 20/08	A1HGR 21/06					
012002900983A	f78	1:47:25 PM	CTSB0 20/09	A1HGR 21/06					
012002900983A	f202	1:48:25 PM	997KA 20/12	A1HGR 21/06					
012002900983A	e1	1:49:25 PM	359EP 21/01	A1HGR 21/06					
012002900983A	f309	1:50:25 PM	BTSg9 20/09	A1HGR 21/06					
012002900983A	f83	1:55:25 PM	428C8 20/08	A1HGR 21/06					
012002900983A	f36	1:56:24 PM	453BD 20/09	A1HGR 21/06					
012002900983A	w1	1:57:25 PM	DWREC 21/08	A1HGR 21/06					
012002900983A	d2	1:58:25 PM	358E2 21/01	A1HGR 21/06					
012002900983A	e5	1:59:25 PM	360F3 21/01	A1HGR 21/06					
012002900983A	f1	2:00:24 PM	361EN 21/01	A1HGR 21/06					
012002900983A	f17	2:01:24 PM	D2FBV 20/09	A1HGR 21/06					
012002900983A	d1	2:02:24 PM	357J6 21/01	A1HGR 21/06					
012002900983A	f235	2:03:24 PM	914GA 20/02	A1HGR 21/06					
012002900983A	f345	2:04:24 PM	BTXGP 21/05	A1HGR 21/06					
012002900983A	f91	2:05:24 PM	559GG 20/07	A1HGR 21/06					
012002900983A	f2	2:06:24 PM	362H0 21/01	A1HGR 21/06					
012002900983A	f233	2:07:24 PM	904LM 21/07	A1HGR 21/06					
012002900983A	f13	2:08:24 PM	375MS 20/04	A1HGR 21/06					
012002900983A	f203	2:09:24 PM	998HN 21/09	A1HGR 21/06					
012002900983A	f35	2:10:24 PM	452BL 20/11	A1HGR 21/06					
012002900983A	f41	2:11:24 PM	456CN 20/03	A1HGR 21/06					
012002900983A	f10	2:12:24 PM	418DE 20/10	A1HGR 21/06					



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002900983A	f24	2:13:24 PM	424EF 20/11	A1HGR 21/06					
012002900983A	f14	2:14:24 PM	364DN 20/10	A1HGR 21/06					
012002900983A	f40	2:15:24 PM	455DT 20/10	A1HGR 21/06					
012002900983A	f256	2:16:24 PM	CMEBK 21/01	A1HGR 21/06					
012002900983A	f4	2:17:24 PM	363F1 20/10	A1HGR 21/06					
012002900983A	ex72	2:18:24 PM	521G9 20/12	A1HGR 21/06					
012002900983A	mx2	2:19:24 PM	851J1 21/03	A1HGR 21/06					
012002900983A	gx2	2:20:24 PM	433BL 21/08	A1HGR 21/06					
012002900983A	i6	2:21:24 PM	465D3 20/11	A1HGR 21/06					
012002901077A	f20	2:26:05 PM	422CS 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002901077A	f27	2:27:06 PM	450CC 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
012002901077A	f93	2:28:06 PM	562JV 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			
011990000657	wx1	12:08:11 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	fx22	12:06:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	wx1	12:07:12 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

### Errors

Identity	Test	Measured	Error
012002900983A	f20	1:40:25 PM	402-Error in sample pipetting module
012002900983A	m6	1:41:25 PM	402-Error in sample pipetting module
012002900983A	m3	1:42:25 PM	402-Error in sample pipetting module
012002900983A	f92	1:43:25 PM	402-Error in sample pipetting module
012002900983A	g2	1:44:25 PM	402-Error in sample pipetting module
012002900983A	f18	1:45:25 PM	402-Error in sample pipetting module
012002900983A	f93	1:46:25 PM	402-Error in sample pipetting module
012002900983A	f78	1:47:25 PM	402-Error in sample pipetting module
012002900983A	f202	1:48:25 PM	402-Error in sample pipetting module
012002900983A	e1	1:49:25 PM	402-Error in sample pipetting module
012002900983A	f309	1:50:25 PM	402-Error in sample pipetting module
012002900983A	f83	1:55:25 PM	402-Error in sample pipetting module
012002900983A	f36	1:56:24 PM	402-Error in sample pipetting module
012002900983A	w1	1:57:25 PM	402-Error in sample pipetting module
012002900983A	d2	1:58:25 PM	402-Error in sample pipetting module
012002900983A	e5	1:59:25 PM	402-Error in sample pipetting module
012002900983A	f1	2:00:24 PM	402-Error in sample pipetting module
012002900983A	f17	2:01:24 PM	402-Error in sample pipetting module
012002900983A	d1	2:02:24 PM	402-Error in sample pipetting module
012002900983A	f235	2:03:24 PM	402-Error in sample pipetting module
012002900983A	f345	2:04:24 PM	301-Clogged sample
012002900983A	f91	2:05:24 PM	301-Clogged sample
012002900983A	f2	2:06:24 PM	402-Error in sample pipetting module
012002900983A	f233	2:07:24 PM	402-Error in sample pipetting module
012002900983A	f13	2:08:24 PM	402-Error in sample pipetting module
012002900983A	f203	2:09:24 PM	402-Error in sample pipetting module
012002900983A	f35	2:10:24 PM	402-Error in sample pipetting module
012002900983A	f41	2:11:24 PM	402-Error in sample pipetting module
012002900983A	f10	2:12:24 PM	402-Error in sample pipetting module
012002900983A	f24	2:13:24 PM	301-Clogged sample
012002900983A	f14	2:14:24 PM	301-Clogged sample
012002900983A	f40	2:15:24 PM	402-Error in sample pipetting module
012002900983A	f256	2:16:24 PM	301-Clogged sample
012002900983A	f4	2:17:24 PM	301-Clogged sample



## Laboratory Report

Identity	Test	Measured	Error
012002900983A	ex72	2:18:24 PM	301-Clogged sample
012002900983A	mx2	2:19:24 PM	301-Clogged sample
012002900983A	gx2	2:20:24 PM	301-Clogged sample
012002900983A	i6	2:21:24 PM	402-Error in sample pipetting module

### Expired lot numbers used for Sample

474GG	ImmunoCAP
352D0	ImmunoCAP
355D5	ImmunoCAP

### Expired lot numbers used for Curve Control

865FX	ImmunoCAP
-------	-----------

### Expired lot numbers used for Quality Control

352D0	ImmunoCAP
BL1CP	Quality control vial